



HAL
open science

A Novel Cortisol Immunosensor Based on a Hafnium Oxide/Silicon Structure for Heart Failure Diagnosis

Hamdi Ben Halima, Nadia Zine, Joan Bausells, Nicole Jaffrezic-Renault,
Abdelhamid Errachid

► **To cite this version:**

Hamdi Ben Halima, Nadia Zine, Joan Bausells, Nicole Jaffrezic-Renault, Abdelhamid Errachid. A Novel Cortisol Immunosensor Based on a Hafnium Oxide/Silicon Structure for Heart Failure Diagnosis. *Micromachines*, 2022, 13 (12), pp.2235. 10.3390/mi13122235 . hal-03997539

HAL Id: hal-03997539

<https://hal.science/hal-03997539>

Submitted on 20 Feb 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Article

A Novel Cortisol Immunosensor Based on a Hafnium Oxide/Silicon Structure for Heart Failure Diagnosis

Hamdi Ben Halima ^{1,*}, Nadia Zine ¹, Joan Bausells ², Nicole Jaffrezic-Renault ^{1,*} and Abdelhamid Errachid ¹

¹ Institut de Sciences Analytiques (ISA)-UMR 5280, Université Claude Bernard Lyon 1, 5 rue de la Doua, 69100 Lyon, France; nadia.zine@univ-lyon1.fr (N.Z.); abdelhamid.errachid-el-salhi@univ-lyon1.fr (A.E.)

² Institute of Microelectronics of Barcelona (IMB-CNM, CSIC), Campus UAB, Bellaterra, 08193 Barcelona, Spain; joan.bausells@imb-cnm.csic.es

* Correspondence: hamdi.ben-halima@isa-lyon.fr (H.B.H.); nicole.jaffrezic@univ-lyon1.fr (N.J.-R.); Tel.: +33-751173780 (H.B.H.); +33-437423516 (N.J.-R.)

Abstract: Assessing cortisol levels in human bodies has become essential to diagnose heart failure (HF). In this work, we propose a salivary cortisol detection strategy as part of an easily integrable lab-on-a-chip for detection of HF biomarkers. Our developed capacitive immunosensor based on hafnium oxide (HfO₂)/silicon structure showed good linearity between increasing cortisol concentration and the charge-transfer resistance/capacitance. Moreover, the developed biosensor was demonstrated to be highly selective toward cortisol compared to other HF biomarkers such as tumor necrosis factor (TNF- α) and N-terminal pro-brain natriuretic peptide (NT-proBNP). The precision of our developed biosensor was evaluated, and the difference between the determined cortisol concentration in saliva and its expected one is < 18%.

Keywords: heart failure; cortisol; electrochemical impedance spectroscopy; capacitance; hafnium oxide



Citation: Ben Halima, H.; Zine, N.; Bausells, J.; Jaffrezic-Renault, N.; Errachid, A. A Novel Cortisol Immunosensor Based on a Hafnium Oxide/Silicon Structure for Heart Failure Diagnosis. *Micromachines* **2022**, *13*, 2235. <https://doi.org/10.3390/mi13122235>

Academic Editor: Yasser Zare

Received: 29 November 2022

Accepted: 14 December 2022

Published: 16 December 2022

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

1. Introduction

Heart failure (HF) is a cardiovascular chronic disease caused by structural abnormalities of the heart that make it unable to fill or pump out blood, resulting in lower delivery of oxygen. Many different factors such as cardiac abnormalities (e.g., structural abnormalities including stiffness of the ventricular chambers, functional abnormalities including mitral incompetence, etc.), coexisting conditions (e.g., diabetes, anemia, sleep apnea, etc.), or other factors (e.g., age, genetic influence, lifestyle, etc.) can be the cause of HF [1,2].

Cortisol (also called 11 β , 17 α , 21-trihydroxypregn-4-ene-3,20-dione) is a steroid hormone of the glucocorticoids family, whose chemical formula is C₂₁H₃₀O₅ (Figure 1) and whose molar mass is 362.46 g/mol. Cortisol is a crucial glucocorticoid hormone that is produced by the zona fasciculata of the adrenal gland. It is well known as a “stress hormone”, and it takes part in the regulation of various physiological functions including energy metabolism, electrolyte balance, blood pressure, and cognitive function [3]. In addition, it contributes to the homeostasis of the adrenal [4], cardiovascular [5,6], immune [7], and endocrine systems [8]. Moreover, it plays a key role in brain regions that are important for cognitive learning, retrieval, encoding, and memory consolidation [9]. This steroid hormone thus represents a potential biomarker for numerous pathological conditions and diseases [10], as well as a useful clinical indicator for relapse vulnerability in chronic alcohol use.

In humans, cortisol is secreted from the outside of the adrenal cortex of the adrenal gland. It is mainly released in response to stress and a low glucose concentration in the blood (hypoglycemia) [11–13]. Stress is a condition that releases glucocorticoids, such as cortisol, and catecholamines, such as adrenaline. These two types of molecules can have effects on memory. Indeed, short exposure improves memory, but prolonged exposure can damage the hippocampal cells, which plays an essential role in memory and space

navigation [14–16]. Stress can also lead to high production of cortisol and cause an allostatic load, which can result in various modifications in the human body. Among these are, for example, mood disorders, anxiety disorders, diseases, obesity, or even atrophy of the nerve cells of the brain. An increase in reactive oxygen species (ROS) is linked to several serious pathologies [17], for example, rheumatoid arthritis [18], neurodegenerative diseases [19], atherosclerosis [20], and heart diseases such as heart failure [17,21,22]. In addition, oxidative stress is an important pathophysiological pathway in the development and progression of heart failure [21,23,24], making cortisol a potential biomarker for heart failure. Cortisol plays an essential role in oxidative stress, which is the result of an imbalance between the production of reactive (radical) oxygen species (ROS) and the body's defenses (in terms of antioxidant cellular capacities). Indeed, an increase in cortisol can cause an excess of oxidative stress in the human body and trigger the oxidation of many cells, making them less functional and leading to heart failure in a person [25]. Studies of patients with heart failure have shown that higher cortisol levels indicate a higher risk of death [26,27]. High cortisol levels for a long time can also lead to protein breakdown and muscle loss [28]. At the bone level, cortisol reduces bone formation. It causes osteoporosis in the long term (decreased density of bones), making the bones more fragile, and therefore leads to an increase in bone fracture [29]. Cortisol also has harmful effects on the immune system. In the event of wounds and a high level of stress, an increase in the cortisol level will be observed, and the healing time of the wound will therefore be longer [30]. The amount of cortisol in the blood varies throughout the day, following the circadian cycle. The highest level is early in the morning (around 8 am) and its lowest level from midnight to 4 am [31,32]. In a healthy person, the cortisol concentration can vary from 1 to 12 ng/mL in the morning and from 0.1 to 3 ng/mL in the evening [33,34]. By comparison, a person with heart failure will have greater amounts than a healthy person.




Figure 1. Semi-developed cortisol formula.

However, the monitoring of cortisol levels must be carried out on the same person because the cortisol level is specific to each, depending on their circadian cycle [32], their diet [35], alcohol consumption [36], and quality of sleep, etc. There are different solutions to regulate the cortisol level in the human body. Thus, some of the factors reducing cortisol levels are, for example, taking magnesium after aerobic exercise [37], taking fish oil [38], music therapy [39], massage therapy [40], laughter and humor [41], and regular dancing [42], etc. Nowadays, cortisol is usually determined by enzyme immunoassay (EIA) [43,44], enzyme-linked immunoassay (ELISA), and radioimmunoassay (RIA) [45], as well as chromatographic techniques coupled with mass spectrometry (MS) or tandem MS/MS [46–48]. Commonly, the principal limitations of these steps lie in their high costs, the long run-time, and the requirement of sophisticated technical skills. Herein, a novel and highly selective approach is proposed for detecting cortisol. Similarly, Ben Halima's approach [49–51], a substrate based on hafnium oxide/silicon structure, was biofunctionalized by immobilizing an anti-cortisol antibody onto the ISFET surface after functionalization with 11-triethoxysilyl undecanal (TESUD) by a vapor-phase method in a saturated medium, exploiting the reaction between the aldehyde and the N-terminus of the antibodies. Electrochemical impedance spectroscopy (EIS) was used for analyte detection

due to its ability to detect variations in resistance and capacitance induced by binding events, thus enhancing device sensitivity. The linear range, accuracy, precision, and limit of detection (LOD) of the biosensor were investigated to reach a preliminary validation of the device. Selectivity was confirmed by analyzing TNF- α and the gold standard HF biomarker N-terminal pro-brain natriuretic peptide (NT-proBNP) samples [52]. Finally, our cortisol concentration was quantified with our biosensor using the standard addition method (SAM) in real saliva samples. The obtained results prove our biosensor to be a promising tool for cortisol detection in saliva. To the best of our knowledge, this is the first biosensor that uses HfO₂/silicon structures for cortisol detection. Due to its high κ (dielectric constant), HfO₂ leakage is reduced, and the gate capacitance is enhanced when compared to SiO₂. This type of structure has already been used for the sensitive detection of interleukin-10 [53]. The use of these materials from microelectronics will allow their integration on a silicon lab-on-a-chip.

2. Materials and Methods

2.1. Materials and Chemicals

Easy Drop OCA 20 (DataPhysics Instruments (Filderstadt, Germany) was used to perform the contact angle measurement (CAM). The HfO₂ substrate surface was activated using UV/Ozone Procleaner™ (BioForce, Konstanz, Germany), which created –OH groups on its surface. All experiments were carried out in a Faraday cage at room temperature (20 ± 2 °C). The counter platinum electrode and the reference Ag/AgCl electrode were from BVT Technologies (BVT Technologies, Strážek, Czech Republic). The potentiostat used to carry out the EIS measurements is a VMP3 multichannel (Biologic-EC-Lab Seyssinet-Pariset, France). Data acquisition and modeling were carried out on EC-Lab software (V11.30, BioLogic, Seyssinet-Pariset, France). Phosphate buffer saline solution (PBS) tablets, ethanolamine (purity $\geq 98\%$), and pure ethanol (purity 95.0%) were purchased from Sigma-Aldrich (Saint-Quentin-Fallavier, France), and 11-triethoxysilyl undecanal (TESUD, 90%) from ABCR (Karlsruhe, Germany). Hydrocortisone (cortisol, purity 99%, Cat. No. ab141250) was from Abcam (Cambridge, UK). Anti-cortisol antibody (Cat. No. XM210) was from Novus Biological (Noyal Châtillon sur Seiche, France). NT-proBNP (Cat. No. 8NT2) was from HyTest (Turku, Finland). Recombinant human TNF- α (Cat. No. 210-TA) was supplied by BioTechne (R&DSYSTEMS, noyal chatillon sur seiche, France). Ultrapure water (resistivity > 18 M Ω cm) was produced by the Elga PURELAB Classic system (ELGA Lab-Water, high Whycombe, UK). PBS tablets were used to create PBS buffer by dissolving it in ultrapure water, thus yielding a 0.01 M phosphate buffer (pH 7.4) with 0.0027 M potassium chloride and 0.137 M sodium chloride, as indicated by the supplier.

2.2. Fabrication Technology of HfO₂/Silicon Structures

A thin-film layer of hafnium oxide (HfO₂) high- κ dielectric was deposited by atomic layer deposition (ALD) on top of a thin silicon dioxide (SiO₂) layer grown thermally on silicon wafers. The 4-inch $\langle 100 \rangle$ p-type silicon (p-Si) wafers were first cleaned with a 5% of hydrofluoric acid (HF) solution to remove any remaining organic compounds, right before a controlled dry thermal oxidation process at 800 °C to grow a thin interfacial 30 Å SiO₂ layer. Following the ALD process, a chemical vapor deposition (CVD) technique based on self-terminating gas-substrate reactions was carried out at 200 °C in a Savannah-200 apparatus from Cambridge NanoTech (Cambridge MA, USA), where the wafer surface was exposed to alternating precursor pulses of tetrakis (dimethylamido) hafnium (TDMAH/N₂) and water vapor (H₂O/N₂), forming a single monoatomic layer of HfO₂ per cycle. Therefore, it was considered that the growth per cycle (GPC) reached a thickness of 20 nm. Finally, a post-deposition annealing process (PDA) was carried out by rapid thermal annealing at different temperatures of 400 °C, 500 °C, and 600 °C, studying the effect on the distribution of interface trap charges and its consequent effect in the dielectric constant by characterizing the sample surface roughness with atomic force microscopy (Agilent 5500 AFM) (Agilent Technologies, Palo Alto, CA, USA) in tapping mode, as well as characterizing the electrical

capacitive behavior. Homogenization of the HfO_2 surface was obtained at 600 °C, as shown in Figure 2.




Figure 2. Characterization of the sample surface roughness with atomic force microscopy (AFM) in tapping mode using post-deposition annealing process (PDA) at different temperatures of 400 °C (a), 500 °C (b), and 600 °C (c).

2.3. Functionalization of HfO_2 Surface with Antibodies

For the fabrication of the biosensing platform, HfO_2 was first functionalized using silane aldehyde. For this purpose, bare HfO_2 substrates were cleaned by sonication in acetone, followed by thorough rinsing in ultrapure water. Surface activation of the HfO_2 substrates was performed using a UV/Ozone Procleaner™ in order to create $-\text{OH}$ groups at the hafnium oxide surface for grafting silane aldehyde. Afterwards, the substrates were thoroughly rinsed and sonicated in ultrapure water. The active HfO_2 substrates (with $-\text{OH}$) were functionalized using TESUD using the vapor-phase method [50,51,53,54]. Then, the substrates were placed in an oven at 100 °C. After baking, they were rinsed with absolute ethanol and dried with nitrogen. Subsequently, the functionalized substrate surface was incubated with anti-cortisol antibody (10 $\mu\text{g}/\text{mL}$) already diluted in PBS. Finally, the remaining aldehyde groups were blocked by a treatment with ethanolamine (1% *v/v*) in PBS buffer. This step is crucial to prevent any nonspecific bonding phenomenon at the detection stage of cortisol (see Figure 3).




Figure 3. Functionalization procedure of the HfO_2 surface.

2.4. Magnetic Nanoparticles Biofunctionalization

The magnetic nanoparticles were prepared with a magnetic core of Fe_3O_4 surrounded by styrene/DVB/ACPA polymers that included COOH as the terminal functional group [55]. Initially, they were washed twice with PBS buffer (pH 7.4) and subsequently the COOH group was activated using a mixture of EDC/NHS at 100 mM in PBS. A magnetic field was used to separate the nanoparticles from the storage solution. Anti-cortisol antibody (10 μL) at 100 $\mu\text{g}/\text{mL}$ was added to the mixture and incubated with slow stirring at room temperature for 90 min. The nonreacted active carboxylic acid groups were blocked with bovine serum albumin (BSA) (0.1%) in PBS buffer for 30 min. The antibody-coated magnetic nanoparticles were then separated from the mixture, resuspended in 1 mL of PBS buffer, and used for the incubation. The procedure described above is summarized in Figure 4.




Figure 4. MNPs-COOH biofunctionalization with anti-cortisol antibody.

2.5. Characterization of HfO_2 Surface Using Contact Angle Measurement (CAM)

The HfO_2 substrate was characterized using the contact angle measurement system to follow each functionalization step by controlling the surface's hydrophilicity. The contact angle measurements were performed using ultrapure water. The dosing volume was $5 \mu\text{L}$ with a dosing rate of $5 \mu\text{L/s}$.

2.6. Electrochemical Measurements with HfO_2 /Silicon Structures

Electrochemical impedance spectroscopy (EIS) is an efficient technique for investigating interfacial properties on surface-modified working electrodes. Measurements were made on the HfO_2 /silicon structure with an aluminum backside contact (500 nm thick) as the working electrode, with a counter platinum electrode CE, and an Ag/AgCl reference electrode (Figure 5). All measurements were made with freshly prepared PBS solution that required 8 mL to fill the cell, while the analysis was performed inside a Faraday cage. A VMP3 multichannel potentiostat purchased from Biologic-EC-Lab was used. The preliminary plot established the required potential on the HfO_2 /silicon structure: -1.5 V versus Ag/AgCl, in the accumulation range for silicon. Starting from the lowest, increasing cortisol concentrations were added, incubated for 30 min at $4 \text{ }^\circ\text{C}$ before impedimetric measurements with the following conditions: EIS frequency ranging from 5 Hz to 100 kHz and a sinus amplitude of 25 mV with polarization potential of -1.5 V . The scan time was 26 s/scan. Data acquisition and analysis were accomplished using EC-Lab software V11.30. The obtained EIS data were modeled by the EC-Lab software using the Randomize + Simplex method. Here, randomize was fixed at 5000 iterations and the fit fixed at 5000 iterations.

Capacitive measurements were performed at 862 Hz, in a potential range from -0.8 V (accumulation range) to $+0.5 \text{ V}$ (inversion range).




Figure 5. Experimental setup.

3. Results

3.1. Surface Characterization

To assess the effectiveness of the functionalization, contact angle measurements were performed on bare HfO_2 , after activation by UV/Ozone, and after functionalization with TESUD. A contact angle of $53.9 \pm 3^\circ$ revealed a slightly hydrophilic nature on bare HfO_2 , which agrees with Lee's values [53]. After surface oxidation with UV/ozone activation, the HfO_2 surface became highly hydrophilic at $12.7 \pm 5^\circ$ due to the high amount of hydroxyl groups on the surface. After the functionalization process with the TESUD, the contact angle increased again to $80.3 \pm 2^\circ$. This hydrophobic character can be explained by the presence of TESUD hydrocarbon chains (see Figure 6).




Figure 6. Contact angle on (A) bare HfO_2 before surface activation; (B) bare HfO_2 after surface activation by UV/ozone; (C) following TESUD formation.

3.2. Biosensor Calibration by EIS Measurements

Bare HfO_2 substrates were cleaned and activated with TESUD, as previously described. The anti-cortisol antibody was immobilized in a conventional three-electrode cell overnight at 4°C , then rinsed with PBS. Following the immobilization of anti-cortisol antibody onto the substrate, the antibody-modified HfO_2 /silicon structure was incubated with cortisol at different increasing concentrations.

Impedance results were presented as a Nyquist plot, as shown in Figure 7A.

An attempt at signal enhancement was made by an additional incubation of the electrode with the MNPs functionalized with the complementary antibody. An SEM image, obtained using FEI Quanta FEG 250 (Thermo Fisher Scientific Inc, Waltham, MA USA) (Figure 8), shows a homogenous distribution of MNPs on the functionalized HfO_2 surface. Nevertheless, the effect of the enhancement of the signal was very disappointing.

The Nyquist plots of HfO_2 modified with anti-cortisol antibody, followed by increasing cortisol concentrations, were fitted with a Randles equivalent circuit (insert in Figure 7A). Here, the components are as follows: R_s corresponds to the resistance of the electrolyte solution; C_{dl} is the double-layer capacitance that is in parallel with R_{ct} , which is the charge-transfer resistance; and Z_w is the Warburg impedance. The values of the extracted parameters are presented in Table 1.




Figure 7. (A) Example of Nyquist plots for the Randles equivalent circuit model obtained by analyzing cortisol standard solutions in PBS (2, 10, 15, and 50 ng mL⁻¹). EIS frequency ranged from 5 Hz to 100 kHz, and a sinus amplitude of 25 mV with polarization potential of -1.5 V; (B) sensitivity curves obtained by analyzing standard solution containing cortisol or other HF biomarkers (e.g., TNF-α and NT-proBNP) in the concentration range 2–50 ng mL⁻¹ using the HfO₂ substrate functionalized with anti-cortisol antibody.




Figure 8. SEM image of functionalized TESUD substrates with Ab-Ag-Ab-MNP (magnetic nanoparticles) biorecognition: left side: SEM image; right side: scheme.

Table 1. Fitting parameters obtained from the equivalent circuit of the cortisol-based HfO₂ immunosensor.

Cortisol Concentration (ng mL ⁻¹)	R _s (Ω)	Q ₂ (nF.s ^(a-1))	R _{ct} (MΩ)	s2 (kΩ.s ^{-1/2})	χ ²
0	9181	10.91	0.206	0.547	1.754 × 10 ⁻³
2	27,031	10.69	0.302	1.88	0.514 × 10 ⁻³
10	29,128	10.87	0.370	2.925	1.845 × 10 ⁻³
15	33,652	10.84	0.450	4.419	1.050 × 10 ⁻³
50	44,199	12.3	0.620	3.633	1.436 × 10 ⁻³

An increase in R_{ct} (charge-transfer resistance) can be observed from R₀ (no cortisol) at 206 kΩ to 302 kΩ for a cortisol concentration of 2 ng/mL. The change in R_{ct} demonstrates the biorecognition of the cortisol by the anti-cortisol antibody grafted on the HfO₂ surface. The R_{ct} increased: 370 kΩ for 10 ng/mL; 450 kΩ for 15 ng/mL; 620 kΩ for 50 ng/mL. The ~104 kΩ variations between the anti-cortisol antibody modified electrode and the first concentration of cortisol at 2 ng/mL demonstrate that the quantification limit of the immunosensor is 2 ng/mL. In Figure 7B (black), the relative variation of R_{ct} versus the cortisol concentration plot produced a linear relationship ranging from 2 ng/mL to 50 ng/mL, with R² = 0.9722 with a slope of 0.03 (ng/mL)⁻¹. A cross-selectivity study was performed to assess the level of nonspecific binding. For this purpose, two other HF biomarkers were used, namely TNF-α and NT-proBNP, using the same conditions

and concentrations as for cortisol. From Figure 7B, it can be seen that the biosensor is demonstrated to be highly selective towards cortisol when compared to both TNF- α and NT-proBNP. The sensitivity for TNF- α (green) is $0.0005 \text{ (ng/mL)}^{-1}$, and for NT-proBNP (Red) the sensitivity is $0.0026 \text{ (ng/mL)}^{-1}$. The immunosensor is 200 times more sensitive to cortisol than TNF- α , and 12 times more sensitive to cortisol than NT-proBNP.

3.3. Biosensor Calibration by Capacitive Measurements

Additionally, capacitive measurements were performed to characterize the semi-conducting behavior of HfO₂/silicon structures for each cortisol concentration. The biosensor was maintained in the electrochemical cell and incubated in PBS containing cortisol at different concentrations. The biosensor was then rinsed with PBS to remove any adsorbed proteins and analyzed afterwards by capacitance measurements using PBS as an electrolyte. This procedure of immunosensor incubation was carried out for all cortisol concentrations (2 ng/mL to 50 ng/mL). The detection of cortisol at various concentrations is shown in Figure 9A. Here, the C_s/C_{max}-voltage curves show a shift in the positive direction with increasing cortisol concentrations, which confirms a flat-band voltage variation due to the increase in electrical charge at the surface of the HfO₂/silicon structure.




Figure 9. (A) Capacitance–voltage plots for cortisol detection using the capacitance biosensor; (B) the calibration curves of the cortisol detection (black curve) and the two interferences: TNF- α (red curve) and NT-proBNP (blue curve).

Figure 9B shows the calibration curve of the HfO₂-based capacitive immunosensor in a linear range from 2 ng/mL to 50 ng/mL of cortisol. The calibration curve has been presented as the absolute value of potential shift $|E-E_0|$ as a function of cortisol concentration, where E is the potential for different cortisol concentrations and E₀ is the potential with no cortisol in the solution. The slope of the straight line is $0.0023 \text{ (ng/mL)}^{-1}$.

The specificity of the biosensor prepared was studied using other biomarkers of heart failure, such as TNF- α and NT-proBNP. Capacitance measurements were performed using the same experimental process (Figure 10A,B). TNF- α and NT-proBNP were detected in the same linear range from 2 to 50 ng/mL with sensitivities of $0.0001 \text{ (ng/mL)}^{-1}$ and $0.000001 \text{ (ng/mL)}^{-1}$, respectively. In this configuration of measurements, the immunosensor is 23 times more sensitive to cortisol than TNF- α , and 2300 times more sensitive to cortisol than NT-proBNP.




Figure 10. (A) Capacitance–voltage plots for TNF- α detection; (B) for NT-proBNP detection using the capacitive immunosensor.

The analytical performance of the HfO₂/silicon-based immunosensor was compared to those of the published electrochemical immunosensors for the detection of cortisol (Table 2). The detection limit of the HfO₂/silicon-based immunosensor is in the lower range, and its detection range falls in the range of concentrations of cortisol for a healthy person and for a person with HF.

Table 2. Comparison of the analytical performance of the HfO₂/silicon-based immunosensor to those of the published electrochemical immunosensors for the detection of cortisol.

Technique	Electrode	Immobilizing Biomolecules	Analyte	Linear Range (ng mL ⁻¹)	Limit of Detection (LOD) (ng mL ⁻¹)	References
EIS	-	Anti-cortisol antibody	Human Tears	0.05–200	21.66	[56]
SWV	Graphite	Anti-cortisol antibody	Human saliva	0.5–55.5	1.7	[57]
Amperometry	Reduced graphene oxide	Anti-cortisol antibody	Human saliva and sweat	0.1–200	0.1	[58]
EIS	Au	Anti-cortisol antibody + BSA	Fish plasma	1440–2170	2750	[59]
EIS	Palladium + MoS ₂	Anti-cortisol antibody	Human sweat	1–500	1	[60]
EIS/capacitance	HfO ₂	Anti-cortisol antibody	Real saliva	2–50	0.66	This work

3.4. Cortisol Quantification in Saliva Samples by Standard Addition Method (SAM)

To simulate the saliva sample analysis with an unknown cortisol concentration, three aliquots (450 μ L) of the pooled saliva sample (PSS) were spiked with different volumes of 100 ng mL⁻¹ cortisol standard solution, obtaining a final concentration of 2, 5, and 7 ng mL⁻¹ of cortisol, and then named “unknown sample” A, B, and C. SAM was carried out by first preparing SAM samples, where a constant volume (50 μ L) of the “unknown sample” was added to each of a quartet of 1.5 mL Eppendorf Lo-bind centrifuge tubes (Sigma-Aldrich, Saint-Quentin-Fallavier, France). A total of 950 μ L of PBS was added to the first tube up to obtain sample C0. Then, an increasing volume of the 100 ng mL⁻¹ cortisol standard solution was added to each subsequent tube, before rounding the solutions in each tube to 1 mL with PBS, thus obtaining three samples, C1, C2, and C3, with known SAM concentrations, where they were analyzed by EIS.

Data obtained from SAM analysis are summarized in Table 3. The precision of our method was calculated by estimating the difference between the cortisol concentrations determined using SAM and their expected values. Based on these results, our biosensor shows good precision to determine the unknown concentration of cortisol in real saliva. These results confirmed that our biosensor was highly sensitive to the slight variation of the unknown concentration of cortisol.

Table 3. Data obtained from the analysis of three saliva samples (corresponding to three aliquots of PSS spiked with different cortisol) by SAM.

Sample Name	Added Concentration (ng/mL)	Dilution Factor	Calculated Concentration	Bias
Sample A	2	19.9	$2.4 \pm 0.3 \text{ ng mL}^{-1}$ (CV = 1%)	14%
Sample B	5	19.5	$4.7 \pm 0.7 \text{ ng mL}^{-1}$ (CV = 5%)	18%
Sample C	7	19.7	$7.1 \pm 0.2 \text{ ng mL}^{-1}$ (CV = 1%)	15%

4. Conclusions

In this study, we present the development of a label-free, highly sensitive, accurate, fast biosensor using EIS/capacitive measurement for cortisol detection in PBS/saliva while also showing good selectivity toward cortisol in the presence of other HF biomarkers (TNF- α and NT-proBNP). Tests performed in PBS showed a linear relationship with the increase in cortisol concentration and the resistance/capacitance of our transistor (R^2 was always > 0.97), showcasing the capability of our biosensor to quantitatively detect cortisol with a detection limit of 0.66 ng mL^{-1} and a dynamic range from 2 to 50 ng mL^{-1} . In addition, our biosensor showed high precision for detecting cortisol in unknown samples using the standard addition method and high sensitivity to small variations of cortisol concentration. This biosensor thus represents a promising bioanalytical tool for accurate quantification of cortisol in saliva to monitor symptoms of inflammation in HF patients. Moreover, the use of materials from microelectronics will allow their integration into silicon lab-on-a-chip devices.

Author Contributions: Writing—original draft preparation, formal analysis, methodology, validation, H.B.H.; formal analysis, N.Z.; validation, J.B.; writing—review and editing, N.J.-R.; supervision, A.E. All authors have read and agreed to the published version of the manuscript.

Funding: The authors acknowledge the financial support of the European Union’s Horizon 2020 research and innovation program, Project NMBP–13-2017 KardiaTool (Grant agreement No. 768686). The authors acknowledge financial support from the POC4Allergies project (Grant Agreement No. 768686), which received funding from ERA PerMed ERA-NET, and financial support from the Bionanosens project (Grant agreement No. 951887), which received funding from the European Union’s Horizon 2020.

Institutional Review Board Statement: Not applicable.

Informed Consent Statement: Not applicable.

Data Availability Statement: The data that support the findings of this study are available from the corresponding author upon reasonable request.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Jessup, M.; Brozena, S. Medical Progress. *Heart Failure. N. Engl. J. Med.* **2003**, *348*, 2007–2018.
2. Benjamin, E.J.; Muntner, P.; Alonso, A.; Bittencourt, M.S.; Callaway, C.W.; Carson, A.P.; Chamberlain, A.M.; Chang, A.R.; Cheng, S.; Das, S.R.; et al. Heart Disease and Stroke Statistics-2019 Update A Report from the American Heart Association. *Circulation* **2019**, *139*, e56–e528. [[CrossRef](#)]
3. Zea, M.; Bellagambi, F.G.; Ben Halima, H.; Zine, N.; Jaffrezic-Renault, N.; Villa, R.; Gabriel, G.; Errachid, A. Electrochemical Sensors for Cortisol Detections: Almost There. *TrAC Trends Anal. Chem.* **2020**, *132*, 116058. [[CrossRef](#)]
4. Raff, H. Utility of Salivary Cortisol Measurements in Cushing's Syndrome and Adrenal Insufficiency. *J. Clin. Endocrinol. Metab.* **2009**, *94*, 3647–3655. [[CrossRef](#)]
5. Whitworth, J.A.; Williamson, P.M.; Mangos, G.; Kelly, J.J. Cardiovascular Consequences of Cortisol Excess. *Vasc. Health Risk Manag.* **2005**, *1*, 291–299. [[CrossRef](#)]
6. Güder, G.; Bauersach, J.; Frantz, S.; Weismann, D.; Allolio, B.; Ertl, G.; Angermann, C.E.; Störk, S. Mortality Risk Prediction by Cortisol and Aldosterone in Chronic Heart Failure. *Exp. Clin. Endocrinol. Diabetes* **2007**, *115*, OR09_5. [[CrossRef](#)]
7. Morey, J.N.; Boggero, I.A.; Scott, A.B.; Segerstrom, S.C. Current Directions in Stress and Human Immune Function. *Curr. Opin. Psychol.* **2015**, *5*, 13–17. [[CrossRef](#)]
8. Turpeinen, U.; Hämäläinen, E. Determination of Cortisol in Serum, Saliva and Urine. *Best Pract. Res. Clin. Endocrinol. Metab.* **2013**, *27*, 795–801. [[CrossRef](#)]
9. Stephens, M.A.C.; Wand, G. Stress and the HPA Axis: Role of Glucocorticoids in Alcohol Dependence. *Alcohol Res. Curr. Rev.* **2012**, *34*, 468–483.
10. De Kloet, E.R.; Joëls, M.; Holsboer, F. Stress and the Brain: From Adaptation to Disease. *Nat. Rev. Neurosci.* **2005**, *6*, 463. [[CrossRef](#)]
11. Taves, M.D.; Gomez-Sanchez, C.E.; Soma, K.K. Extra-Adrenal Glucocorticoids and Mineralocorticoids: Evidence for Local Synthesis, Regulation, and Function. *Am. J. Physiol.-Endocrinol. Metab.* **2011**, *301*, E11–E24. [[CrossRef](#)]
12. Gecaite-Stonciene, J.; Hughes, B.M.; Kazukauskienė, N.; Bunevicius, A.; Burkauskas, J.; Neverauskas, J.; Bellani, M.; Mickuviene, N. Cortisol Response to Psychosocial Stress, Mental Distress, Fatigue and Quality of Life in Coronary Artery Disease Patients. *Sci. Rep.* **2022**, *12*, 19373. [[CrossRef](#)]
13. Zubarev, A.; Cuzminschi, M.; Iordache, A.-M.; Iordache, S.-M.; Rizea, C.; Grigorescu, C.E.A.; Giuglea, C. Graphene-Based Sensor for the Detection of Cortisol for Stress Level Monitoring and Diagnostics. *Diagnostics* **2022**, *12*, 2593. [[CrossRef](#)]
14. Eto, K.; Mazilu-Brown, J.K.; Henderson-MacLennan, N.; Dipple, K.M.; McCabe, E.R.B. Development of Catecholamine and Cortisol Stress Responses in Zebrafish. *Mol. Genet. Metab. Rep.* **2014**, *1*, 373–377. [[CrossRef](#)]
15. Kaur, K.; Dhingra, S.; Slezak, J.; Sharma, A.K.; Bajaj, A.; Singal, P.K. Biology of TNF α and IL-10, and Their Imbalance in Heart Failure. *Heart Fail. Rev.* **2009**, *14*, 113–123. [[CrossRef](#)]
16. James, G.D.; Alfarano, A.S.; van Berge-Landry, H.M. Differential Circadian Catecholamine and Cortisol Responses between Healthy Women with and without a Parental History of Hypertension. *Am. J. Hum. Biol.* **2014**, *26*, 753–759. [[CrossRef](#)]
17. Liguori, I.; Russo, G.; Curcio, F.; Bulli, G.; Aran, L.; Della-Morte, D.; Gargiulo, G.; Testa, G.; Cacciatore, F.; Bonaduce, D.; et al. Oxidative Stress, Aging, and Diseases. *CIA* **2018**, *13*, 757–772. [[CrossRef](#)]
18. Mateen, S.; Moin, S.; Khan, A.Q.; Zafar, A.; Fatima, N. Increased Reactive Oxygen Species Formation and Oxidative Stress in Rheumatoid Arthritis. *PLoS ONE* **2016**, *11*, e0152925. [[CrossRef](#)]
19. Cenini, G.; Lloret, A.; Cascella, R. Oxidative Stress in Neurodegenerative Diseases: From a Mitochondrial Point of View. *Oxidative Med. Cell. Longev.* **2019**, *2019*, 2105607. [[CrossRef](#)]
20. Kattoor, A.J.; Pothineni, N.V.K.; Palagiri, D.; Mehta, J.L. Oxidative Stress in Atherosclerosis. *Curr. Atheroscler. Rep.* **2017**, *19*, 42. [[CrossRef](#)]
21. Byrne, J.A.; Grieve, D.J.; Cave, A.C.; Shah, A.M. Oxidative Stress and Heart Failure. *Arch. Mal. Coeur. Vaiss.* **2003**, *96*, 214–221.
22. Tsutsui, H.; Kinugawa, S.; Matsushima, S. Oxidative Stress and Heart Failure. *Am. J. Physiol.-Heart Circ. Physiol.* **2011**, *301*, H2181–H2190. [[CrossRef](#)]
23. Grieve, D. Oxidative Stress in Heart Failure More than Just Damage. *Eur. Heart J.* **2003**, *24*, 2161–2163. [[CrossRef](#)]
24. Sawyer, D.B. Oxidative Stress in Heart Failure: What Are We Missing? *Am. J. Med. Sci.* **2011**, *342*, 120–124. [[CrossRef](#)]
25. Yamaji, M.; Tsutamoto, T.; Kawahara, C.; Nishiyama, K.; Yamamoto, T.; Fujii, M.; Horie, M. Serum Cortisol as a Useful Predictor of Cardiac Events in Patients with Chronic Heart Failure: The Impact of Oxidative Stress. *Circ. Heart Fail.* **2009**, *2*, 608–615. [[CrossRef](#)]
26. Tzanis, G.; Dimopoulos, S.; Agapitou, V.; Nanas, S. Exercise Intolerance in Chronic Heart Failure: The Role of Cortisol and the Catabolic State. *Curr. Heart Fail. Rep.* **2014**, *11*, 70–79. [[CrossRef](#)]
27. Güder, G.; Bauersachs, J.; Frantz, S.; Weismann, D.; Allolio, B.; Ertl, G.; Angermann, C.E.; Störk, S. Complementary and Incremental Mortality Risk Prediction by Cortisol and Aldosterone in Chronic Heart Failure. *Circulation* **2007**, *115*, 1754–1761. [[CrossRef](#)]
28. Braun, T.P.; Marks, D.L. The Regulation of Muscle Mass by Endogenous Glucocorticoids. *Front. Physiol.* **2015**, *6*, 12. [[CrossRef](#)]
29. Canalis, E.; Mazziotti, G.; Giustina, A.; Bilezikian, J.P. Glucocorticoid-Induced Osteoporosis: Pathophysiology and Therapy. *Osteoporos. Int.* **2007**, *18*, 1319–1328. [[CrossRef](#)]
30. Ebrecht, M.; Hextall, J.; Kirtley, L.-G.; Taylor, A.; Dyson, M.; Weinman, J. Perceived Stress and Cortisol Levels Predict Speed of Wound Healing in Healthy Male Adults. *Psychoneuroendocrinology* **2004**, *29*, 798–809. [[CrossRef](#)]
31. Sephton, S.E.; Lush, E.; Dedert, E.A.; Floyd, A.R.; Rebholz, W.N.; Dhabhar, F.S.; Spiegel, D.; Salmon, P. Diurnal Cortisol Rhythm as a Predictor of Lung Cancer Survival. *Brain Behav. Immun.* **2013**, *30*, S163–S170. [[CrossRef](#)]

32. Chan, S.; Debono, M. Review: Replication of Cortisol Circadian Rhythm: New Advances in Hydrocortisone Replacement Therapy. *Ther. Adv. Endocrinol.* **2010**, *1*, 129–138. [[CrossRef](#)]
33. Bellagambi, F.G.; Degano, I.; Ghimenti, S.; Lomonaco, T.; Dini, V.; Romanelli, M.; Mastorci, F.; Gemignani, A.; Salvo, P.; Fuoco, R.; et al. Determination of Salivary α -Amylase and Cortisol in Psoriatic Subjects Undergoing the Trier Social Stress Test. *Microchem. J.* **2018**, *136*, 177–184. [[CrossRef](#)]
34. Aas, M.; Ueland, T.; Inova, A.; Melle, I.; Andreassen, O.A.; Steen, N.E. Childhood Trauma Is Nominally Associated with Elevated Cortisol Metabolism in Severe Mental Disorder. *Front. Psychiatry* **2020**, *11*, 391. [[CrossRef](#)]
35. Hamzavi, M.; Tadbir, A.A.; Rezvani, G.; Ashraf, M.J.; Fattahi, M.J.; Khademi, B.; Sardari, Y.; Jeirudi, N. Tissue Expression, Serum and Salivary Levels of IL-10 in Patients with Head and Neck Squamous Cell Carcinoma. *Asian Pac. J. Cancer Prev.* **2013**, *14*, 1681–1685. [[CrossRef](#)]
36. King, A.; Munisamy, G.; de Wit, H.; Lin, S. Attenuated Cortisol Response to Alcohol in Heavy Social Drinkers. *Int. J. Psychophysiol.* **2006**, *59*, 203–209. [[CrossRef](#)]
37. Cinar, V.; Mogulkoc, R.; Baltaci, A.K.; Polat, Y. Adrenocorticotrophic Hormone and Cortisol Levels in Athletes and Sedentary Subjects at Rest and Exhaustion: Effects of Magnesium Supplementation. *Biol. Trace Elem. Res.* **2008**, *121*, 215–220. [[CrossRef](#)]
38. Noreen, E.E.; Sass, M.J.; Crowe, M.L.; Pabon, V.A.; Brandauer, J.; Averill, L.K. Effects of Supplemental Fish Oil on Resting Metabolic Rate, Body Composition, and Salivary Cortisol in Healthy Adults. *J. Int. Soc. Sport. Nutr.* **2010**, *7*, 31. [[CrossRef](#)]
39. Uedo, N.; Ishikawa, H.; Morimoto, K.; Ishihara, R.; Narahara, H.; Akedo, I.; Ioka, T.; Kaji, I.; Fukuda, S. Reduction in Salivary Cortisol Level by Music Therapy during Colonoscopic Examination. *Hepatogastroenterology* **2004**, *51*, 451–453.
40. Field, T.; Hernandez-Reif, M.; Diego, M.; Schanberg, S.; Kuhn, C. Cortisol decreases and serotonin and dopamine increase following massage therapy. *Int. J. Neurosci.* **2005**, *115*, 1397–1413. [[CrossRef](#)]
41. Berk, L.S.; Tan, S.A.; Berk, D. Cortisol and Catecholamine Stress Hormone Decrease Is Associated with the Behavior of Perceptual Anticipation of Mirthful Laughter. *FASEB J.* **2008**, *22*, 946.11. [[CrossRef](#)]
42. Quiroga Murcia, C.; Bongard, S.; Kreutz, G. Emotional and Neurohumoral Responses to Dancing Tango Argentino: The Effects of Music and Partner. *Music. Med.* **2009**, *1*, 14–21. [[CrossRef](#)]
43. Kalman, B.A.; Grahn, R.E. Measuring Salivary Cortisol in the Behavioral Neuroscience Laboratory. *J. Undergrad. Neurosci. Educ.* **2004**, *2*, A41–A49.
44. Gozansky, W.S.; Lynn, J.S.; Laudenslager, M.L.; Kohrt, W.M. Salivary Cortisol Determined by Enzyme Immunoassay Is Preferable to Serum Total Cortisol for Assessment of Dynamic Hypothalamic–Pituitary–Adrenal Axis Activity. *Clin. Endocrinol.* **2005**, *63*, 336–341. [[CrossRef](#)]
45. Inder, W.J.; Dimeski, G.; Russell, A. Measurement of Salivary Cortisol in 2012—Laboratory Techniques and Clinical Indications. *Clin. Endocrinol.* **2012**, *77*, 645–651. [[CrossRef](#)]
46. Kataoka, H.; Matsuura, E.; Mitani, K. Determination of Cortisol in Human Saliva by Automated In-Tube Solid-Phase Microextraction Coupled with Liquid Chromatography–Mass Spectrometry. *J. Pharm. Biomed. Anal.* **2007**, *44*, 160–165. [[CrossRef](#)]
47. Matsui, F.; Koh, E.; Yamamoto, K.; Sugimoto, K.; Sin, H.-S.; Maeda, Y.; Honma, S.; Namiki, M. Liquid Chromatography–Tandem Mass Spectrometry (LC-MS/MS) Assay for Simultaneous Measurement of Salivary Testosterone and Cortisol in Healthy Men for Utilization in the Diagnosis of Late-Onset Hypogonadism in Males. *Endocr. J.* **2009**, *56*, 1083–1093. [[CrossRef](#)]
48. Turpeinen, U.; Välimäki, M.J.; Hämäläinen, E. Determination of Salivary Cortisol by Liquid Chromatography–Tandem Mass Spectrometry. *Scand. J. Clin. Lab. Investig.* **2009**, *69*, 592–597. [[CrossRef](#)]
49. Ben Halima, H.; Bellagambi, F.G.; Alcacer, A.; Pfeiffer, N.; Heuberger, A.; Hangouët, M.; Zine, N.; Bausells, J.; Elaissari, A.; Errachid, A. A Silicon Nitride ISFET Based Immunosensor for Tumor Necrosis Factor-Alpha Detection in Saliva. A Promising Tool for Heart Failure Monitoring. *Anal. Chim. Acta* **2021**, *1161*, 338468. [[CrossRef](#)]
50. Ben Halima, H.; Bellagambi, F.G.; Brunon, F.; Alcacer, A.; Pfeiffer, N.; Heuberger, A.; Hangouët, M.; Zine, N.; Bausells, J.; Errachid, A. Immuno Field-Effect Transistor (ImmunoFET) for Detection of Salivary Cortisol Using Potentiometric and Impedance Spectroscopy for Monitoring Heart Failure. *Talanta* **2022**, 123802. [[CrossRef](#)]
51. Ben Halima, H.; Bellagambi, F.G.; Hangouët, M.; Alcacer, A.; Pfeiffer, N.; Heuberger, A.; Zine, N.; Bausells, J.; Errachid, A. A Novel IMFET Biosensor Strategy for Interleukin-10 Quantification for Early Screening Heart Failure Disease in Saliva. *Electroanalysis* **2022**, elan.202200141, early view. [[CrossRef](#)]
52. Bellagambi, F.G.; Petersen, C.; Salvo, P.; Ghimenti, S.; Franzini, M.; Biagini, D.; Hangouët, M.; Trivella, M.G.; Di Francesco, F.; Paolicchi, A.; et al. Determination and Stability of N-Terminal pro-Brain Natriuretic Peptide in Saliva Samples for Monitoring Heart Failure. *Sci. Rep.* **2021**, *11*, 13088. [[CrossRef](#)]
53. Lee, M.; Zine, N.; Baraket, A.; Zabala, M.; Campabadal, F.; Caruso, R.; Trivella, M.G.; Jaffrezic-Renault, N.; Errachid, A. A Novel Biosensor Based on Hafnium Oxide: Application for Early Stage Detection of Human Interleukin-10. *Sens. Actuators B Chem.* **2012**, *175*, 201–207. [[CrossRef](#)]
54. Ben Halima, H.; Bellagambi, F.G.; Hangouët, M.; Alcacer, A.; Pfeiffer, N.; Heuberger, A.; Zine, N.; Bausells, J.; Elaissari, A.; Errachid, A. A Novel Electrochemical Strategy for NT-ProBNP Detection Using IMFET for Monitoring Heart Failure by Saliva Analysis. *Talanta* **2023**, *251*, 123759. [[CrossRef](#)]

55. Jamshaid, T.; Tenório-Neto, E.T.; Eissa, M.; Zine, N.; El-Salhi, A.E.; Kunita, M.H.; Elaissari, A. Preparation and Characterization of Submicron Hybrid Magnetic Latex Particles. *Polym. Adv. Technol.* **2015**, *26*, 1102–1108. [[CrossRef](#)]
56. Cardinell, B.A.; Spano, M.L.; La Belle, J.T. Toward a Label-Free Electrochemical Impedance Immunosensor Design for Quantifying Cortisol in Tears. *Crit. Rev. Biomed. Eng.* **2019**, *47*, 207–215. [[CrossRef](#)]
57. Kämäräinen, S.; Mäki, M.; Tolonen, T.; Palleschi, G.; Virtanen, V.; Micheli, L.; Sesay, A.M. Disposable Electrochemical Immunosensor for Cortisol Determination in Human Saliva. *Talanta* **2018**, *188*, 50–57. [[CrossRef](#)]
58. Tuteja, S.K.; Ormsby, C.; Neethirajan, S. Noninvasive Label-Free Detection of Cortisol and Lactate Using Graphene Embedded Screen-Printed Electrode. *Nano-Micro. Lett.* **2018**, *10*, 41. [[CrossRef](#)]
59. Pali, M.; Garvey, J.E.; Small, B.; Suni, I.I. Detection of Fish Hormones by Electrochemical Impedance Spectroscopy and Quartz Crystal Microbalance. *Sens. Bio-Sens. Res.* **2017**, *13*, 1–8. [[CrossRef](#)]
60. Kinnamon, D.; Ghanta, R.; Lin, K.-C.; Muthukumar, S.; Prasad, S. Portable Biosensor for Monitoring Cortisol in Low-Volume Perspired Human Sweat. *Sci. Rep.* **2017**, *7*, 13312. [[CrossRef](#)]