



Resonance frequency measurement with accuracy and stability at the 10 –12 level in a copper microwave cavity below 26 K by experimental optimization

Haiyang Zhang, Bo Gao, Wenjing Liu, Changzhao Pan, Dongxu Han, Mark Plimmer, Ercang Luo, Laurent Pitre

► To cite this version:

Haiyang Zhang, Bo Gao, Wenjing Liu, Changzhao Pan, Dongxu Han, et al.. Resonance frequency measurement with accuracy and stability at the 10 –12 level in a copper microwave cavity below 26 K by experimental optimization. *Measurement Science and Technology*, 2020, 31 (7), pp.075011. <10.1088/1361-6501/ab796e>. <hal-04021174>

HAL Id: hal-04021174

<https://hal.science/hal-04021174v1>

Submitted on 9 Mar 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

Resonance frequency measurement with accuracy and stability at the 10^{-12} level in a copper microwave cavity below 26 K by experimental optimization

Haiyang Zhang^{a,b,&}, Bo Gao^{a,b,&*}, Wenjing Liu^{a,b,e}, Changzhao Pan^{c,a}, Dongxu Han^{d,a}, Mark Plimmer^c, Ercang Luo^{a,b,e}, Laurent Pitre^{c,a}

^aTIPC-LNE Joint Laboratory on Cryogenic Metrology Science and Technology, Technical Institute of Physics and Chemistry, Chinese Academy of Science, Beijing 100190, China

^bKey Laboratory of Cryogenics, Technical Institute of Physics and Chemistry (TIPC), Chinese Academy of Sciences (CAS), Beijing 100190, China

^cLaboratoire national de métrologie et d'essais-Conservatoire national des arts et métiers (LNE-Cnam), La Plaine-Saint Denis F93210, France

^dSchool of Mechanical Engineering, Beijing Institute of Petrochemical Technology, Beijing 102617, China

^eUniversity of Chinese Academy of Sciences, Beijing 100490, China

*Corresponding author. Tel/Tax:86-10-82554512. E-mail: bgao@mail.ipc.ac.cn.

&These authors contributed to the work equally and should be regarded as co-first authors.

Abstract

Single-pressure refractive index gas thermometry (SPRIGT) is a novel primary thermometry technique, developed jointly by TIPC of CAS in China and LNE-Cnam in France. To help obtain a competitive uncertainty of 0.25 mK in thermodynamic temperature measurements, high-stability and low-uncertainty of microwave resonance frequency measurements better than 2 ppb have been demonstrated. This article describes how high-stability and low-uncertainty resonance frequency measurements were achieved using a copper microwave cavity. Microwave measurements were carried out under vacuum and isobarically at helium-4 pressures of (30, 60, 90, and 120) kPa over the temperature range of (5 to 26) K, with good consistency among microwave modes for the thermodynamic temperatures determined. Performance was optimized using an Allan variance analysis. In this way, with an integration time of 3 hours, a stability and accuracy at the 10^{-12} level were obtained, an almost 20 fold improvement upon our previous result (Zhang *et al. Sci. Bull* 2019; 64: 286-288).

Keywords

Stability; Microwave resonance frequency; Resonator; Allan deviation; Primary thermometry; Thermodynamic temperature.

1 Introduction

Single-pressure refractive index gas thermometry (SPRIGT), developed jointly by Technical Institute of Physics and Chemistry of Chinese Academy of Sciences (TIPC-CAS) in China and Laboratoire national de métrologie et d'essais-Conservatoire national des arts et métiers (LNE-Cnam) in France [1], is a type of thermometry methods based on gas polarizability. As a relative primary thermometry technique [2], SPRIGT measurements are conducted on single isobars rather than isotherms, which alleviates the need for accurate absolute pressure measurement and increases the measurement speed ten-fold. With state-of-the-art *ab initio* calculations of helium-4 properties [3-8], a competitive uncertainty of 0.25 mK^1 is expected for the measurement of thermodynamic temperature below the neon triple point at 24.5561 K, which is of important strategic significance for the development of both cutting-edge scientific researches and large scientific facilities, such as the Spallation Neutron Source in China and Large Hadron Collider in Europe [9].

Microwave measurements are widely used in the field of gas metrology [10-16]. SPRIGT allows the thermodynamic temperature to be determined from the ratio of gas refractive indices at a single pressure, one measured at the unknown temperature and the other at the reference temperature [1, 17], while the refractive index can be measured from the microwave resonance frequencies in a quasi-spherical high-conductivity resonator. The reference temperature can be the fixed point of neon or a known thermodynamic temperature measured using another absolute primary thermometry method, such as acoustic gas thermometry (AGT). In refractive index gas thermometry (RIGT) [2, 18], microwave measurements are used to determine the gas refractive index $n(p, T)$. From a knowledge of $n(p, T)$ and the pressure p , the thermodynamic temperature T is deduced. RIGT has been successfully implemented at

¹On May 20th, 2019, the Bureau International des Poids et Mesures announced a major revision to the SI, in which the base units, the kelvin, symbol K, was redefined by fixing the value of Boltzmann constant. The practical realizations of the kelvin by primary thermometry are indicated in the “Mise en pratique for the definition of the kelvin in the SI”, in which low-uncertainty primary thermometry is required to promote the realizations of the new kelvin and the spread of high-accuracy, low-temperature metrology. (see <https://www.bipm.org/utls/en/pdf/si-mep/SI-App2-kelvin.pdf>).

1 NRC by Rourke to measure the temperatures 24.5561 K (neon triple point), 54.3584 K
 2 (oxygen triple point) and 83.8058 K (argon triple point) which are three ITS-90
 3 defining fixed points [19]. Recent progress has also been achieved in the same
 4 laboratory on the measurement of the thermodynamic temperature of the triple point of
 5 xenon $T = 161.40596$ K by extrapolating the experimentally-determined
 6 compressibility at the triple point of water [20]. In a complementary application of
 7 RIGT, where the pressure is deduced from the refractive index and temperature
 8 measurements, a quantum standard for absolute pressure measurements is currently
 9 being developed jointly by LNE in France and INRiM in Italy in the range from 0.2 kPa
 10 to 20 kPa by a superconducting microwave cavity [21]. In acoustic gas thermometry
 11 [10, 13-16], microwave measurements are used to determine the dimensions or volume
 12 of the resonator so that the thermodynamic temperature can be deduced from acoustic
 13 measurements. Besides, microwave measurements have also been used in other precise
 14 measurements, such as density and critical phenomena of helium [22], frequency [23]
 15 and quantum-gas and gravitational physics [24]. These were successfully implemented
 16 by using high-quality factor niobium microwave cavities. The aforementioned gas
 17 metrology methods (SPRIGT, RIGT, AGT and absolute pressure in a quantum standard)
 18 and precise measurements would benefit from any improvement in microwave
 19 measurements.

20 The best current primary standards can produce the SI second with a relative
 21 standard uncertainty approaching 10^{-16} , while the relative uncertainty of secondary
 22 frequency standards based on optical clocks is at the level of 10^{-18} [25]. A convenient
 23 way to perform experiments is to use a servo loop to lock the oscillator frequency to
 24 the maximum of a resonance. The frequency stability then depends on the shape and
 25 sharpness of the resonance, the signal-to-noise ratio, and the type of noise that
 26 predominates at the time scale under consideration [26-28]. One of the most commonly
 27 used techniques for servo-locking is the sideband method due to Pound and its optical
 28 equivalent [28]. By contrast, in SPRIGT, it is essential to perform scans over the
 29 resonance lines to check for any change in the width and shape (see Section 3.1). These
 30 parameters are used both to measure electrical conductivity and monitor systematic

effects. A scan takes typically two minutes. To implement SPRIGT, a system has been built at TIPC-CAS in China, which includes three subsystems, namely a temperature control system, a pressure control system and a microwave system. The temperature and pressure control systems are designed to provide a very stable working environment for the microwave resonator and connecting cables. High-stability temperature control (0.2 mK), pressure control (4 ppm, $1 \text{ ppm} \equiv 10^{-6}$) and microwave resonance frequency (2 ppb, $1 \text{ ppb} \equiv 10^{-9}$) are required for thermodynamic temperature measurements with an uncertainty of 0.25 mK. A simple description of the three subsystems is presented in Section 2.

In this paper, the structure of the quasi-spherical resonator is described and the experimental optimization of microwave measurement is presented. The rest of the paper is structured as follows. In Section 2, we describe the experimental apparatus involving the main instrumentations for microwave, temperature and pressure measurements. Detailed descriptions of the temperature [29, 30] and pressure measurement [31] can be found elsewhere and lie beyond the scope of the present paper. In Section 3, we present how to realize high-stability and low-uncertainty of microwave resonance frequency measurement by optimizing the microwave background, the microwave emission power, mechanical stability, temperature control method, time reference, and microwave signal.

2 Experimental set-up

The apparatus of SPRIGT is shown in Fig. 1a, including microwave, temperature control and pressure control subsystems. For the microwave system, the core element is a copper (Cu-EPT) microwave cavity, *i.e.*, a quasi-spherical resonator (QSR), shown in Fig. 1b. The high-quality factor² QSR was built from two hemispheres whose inner surfaces were machined by precision diamond turning. The dimensions are the same as those of the resonator used at LNE-Cnam for a determination of the Boltzmann constant [11]. The inner shape is designed to be a triaxial ellipsoid defined by

²Quality factors ($Q = f/2g$) for the 4 microwave modes at 5 K ~26 K are: $Q(\text{TM}11) \approx 140\,000$, $Q(\text{TE}11) \approx 210\,000$, $Q(\text{TM}12) \approx 170\,000$, $Q(\text{TE}13) \approx 290\,000$.

$$\frac{x^2}{a^2} + \frac{y^2}{a^2(1+\varepsilon_2)^2} + \frac{z^2}{a^2(1+\varepsilon_1)^2} = 1 \quad (1)$$

with $a = 49.50$ mm, $\varepsilon_1 = 0.001$, and $\varepsilon_2 = 0.0005$, corresponding to the nominal semi-major axes of the tri-axial ellipsoid of 49.50 mm, 49.75 mm and 50.00 mm on X , Y and Z axes, respectively. The nominal shell thickness of the resonator is 10.0 mm. Two loop antennas are connected to a two-port vector network analyzer (Keysight Technologies N5241A PNA-X), one in each hemisphere used for emission and reception. The frequency reference of the vector network analyzer is a 10 MHz signal provided by a rubidium frequency standard (Stanford Research Systems FS725) or, for better stability, a GPS time and frequency system (Stanford Research Systems FS740, locked to GPS with OCXO or rubidium frequency standard). For FS740, two kinds of antenna were used, one indoor, the other outdoor.




Fig.1. The structure schematic diagram of SPRIGT. (a) Simplified system diagram. (b) 3D drawing of the quasi-spherical resonator.

1 The QSR has been closed successfully at room temperature using a microwave
2 method by monitoring the change of relative excess half-width [32]. The resonator was
3 designed to be used under vacuum and then with helium gas. For this reason, to avoid
4 contamination, Argon (purity of 99.999 %, supplied by Beijing AP BAIF Gases
5 Industry Co. Ltd.) was flowed through it at a continuous rate of 80 SCCM³ until the
6 resonator was coupled in the pressure vessel of the cryostat. After putting the resonator
7 into the pressure vessel of the cryostat, a preliminary test on the performance of the
8 quasi-spherical resonator under vacuum was implemented without using any
9 optimization methods involved in the present work. Frequency stability and uncertainty
10 of 0.02 ppb and 0.03 ppb were realized by using 0 dBm microwave emission power
11 with an integration time of 3 h [17], equivalently 0.063 ppb and 0.095 ppb for -10 dBm
12 microwave emission power [1].

13 For the temperature control system, a cryogen-free cryostat was developed for
14 SPRIGT in our previous work [29], using a two-stage GM type pulse tube cryocooler
15 as the cooling resource. Based on multi-layer radiation shields combined with the
16 thermal-resistance method, gas type heat switch and PID control method, the pressure
17 vessel temperature stabilities of 0.021 mK ~ 0.050 mK were realized with an integration
18 time of 0.8 s in the temperature range from 5 K to 25 K. Based on further investigation
19 on thermal characterization of the cryostat, the pressure vessel temperature stabilities
20 were controlled at 0.019 mK in the same range [30]. In this work, the temperature of
21 the resonator was measured by a rhodium iron thermometer (Tinsley, SN226242)
22 coupled to an AC resistance bridge (ASL F900). The 10 Ω standard reference resistance
23 (Tinsley 5685A, SN1580409) was placed in an oil bath (Aikom Instruments MR 5100-
24 L) with temperature stability better than 1 mK.

25 For the pressure control system, the detailed content is not emphasized in this work
26 since it has been presented in our previous work [31]. The uncertainty contribution of
27 pressure to measurements of thermodynamic temperature has also been reported in our

³SCCM = standard cubic centimeters per minute, corresponding to $1.6667 \times 10^{-8} \text{ m}^3 \cdot \text{s}^{-1}$ in the International System of Units (SI). We define the volumetric flow 1 sccm as the flow of 1 cubic centimeter per minute of argon at the pressure 101.325 kPa and temperature 20°C.

previous work [1] and Fig. 2 in the review paper of RIGT [2]. To fill helium-4 gas (purity of 99.9999 %, supplied by Air Liquide) into the resonator and keep a single pressure, a gas compensation loop was built at room temperature to compensate for the leak of the piston gauge. The gas pressure was measured by an absolute-pressure piston gauge (Fluke PG 7601) and conducted by maintaining the piston at a constant height using a laser interferometer (KEYENCE LK-G80). Pressure stabilities of 0.0032 Pa at 30 kPa and 0.002 Pa at 90 kPa were realized with an integration time of 1.4 s at room temperature, corresponding to relative stabilities of 0.1 ppm and 0.02 ppm, respectively [31].

3 Experimental optimization and discussions

To realize high-stability and low-uncertainty microwave resonance frequency measurements at low temperatures, several experimental optimizations were implemented by using Allan analysis of variance in the present work. Detailed results are listed as below.

3.1 Microwave background polynomial

In this work, the microwave resonance frequencies f_n and half-widths g_n were determined from non-linear least-squares fitting (LM method [33]) of the measured complex scattering parameters S_{21} as a function of frequency [34]. Here we rewrite equation 15 in the literature to the following form

$$S_{21} = \sum_n \frac{A_n f}{f^2 - (f_n + i g_n)^2} + \sum_{j=0}^J B_j (f - f_*)^j \quad (2)$$

where the fitting parameters were the complex constants A_n , B_j and the three complex resonance frequencies $(f_n + i g_n)$, one for each component of the triplet. In equation 2, J is the background polynomial order, f is the source frequency and f_* is an arbitrary constant, here we made the same choice as the literature [34], *i.e.*, $f_* = f_x$, to avoid numerical problems in the fitting program.

Fig.2 shows the fitting results of TM11 microwave mode for different background polynomial orders at 5 K and 24.5 K under vacuum. With the polynomial order

increasing, the average fitted triplet frequency $\langle f + g \rangle$ and the uncertainty $u(\langle f + g \rangle)$ have a little change, as shown in Fig. 2a and Fig. 2c; the maximum relative changes on $\langle f + g \rangle$ are about 2.0×10^{-10} and 1.9×10^{-10} for 5 K and 24.5 K, respectively. These changes are less than half of the relative standard uncertainty of $\langle f + g \rangle$, indicating that increasing the microwave background polynomial order cannot obviously reduce their influence on the average fitted triplet frequency and the uncertainty in our system. Considering other microwave modes and those measured at room temperatures, a 2nd-order background polynomial is competent in this work. Table 1 lists the fitted frequencies and half-widths of the triplet for 2nd-order background polynomial. Fig. 2b and Fig. 2d compare the results of the measured and calculated components of the scattering parameters S_{21} , where good agreement was found.




Fig.2. Fitting results of TM11 microwave mode for different background polynomial orders under vacuum: (a) $\langle f + g \rangle$ against microwave background polynomial order J at 5 K. (b) Real and imaginary components of the parameter S_{21} at 5 K, measured and fitted using 2nd-order background polynomial. (c) $\langle f + g \rangle$ against microwave background polynomial order J at 24.5 K. (d) Real and imaginary components of the parameter S_{21} at 24.5 K, measured and fitted values using 2nd-order background polynomial.

Table 1. Fitted values of the triplet with 2nd-order background polynomial.

T / K	f_1 / MHz	g_1 / MHz	f_2 / MHz	g_2 / MHz	f_3 / MHz	g_3 / MHz
5	2617.0688336(11)	0.0093815(11)	2617.6300911(15)	0.0093725(15)	2618.2397203(16)	0.0098880(16)
24.5	2617.0614026(12)	0.0094027(12)	2617.6226012(16)	0.0093916(16)	2618.2321434(17)	0.0099124(17)

Besides, the background noise from cables and feed-throughs also influences the microwave measurement. To reduce the noise as much as possible, one can use some low-loss cables and feed-throughs, covering the working frequency range. For further improvement, the optimal combination of cables and feed-throughs, even the working microwave modes can be selected by measuring and analyzing the S_{21} parameters.

3.2 Microwave emission power

In this work, the microwave signal was generated by a two-port N5241A vector network analyzer. The signal was transmitted from port 1 of the vector network analyzer, passing through the cables, feed-throughs, and resonator, later received by port 2. Since the relative loss of the cables and feed-throughs are nearly constant at stable operating conditions, the microwave emission power has a direct influence on the measurement inside the resonator. If the power is too small, the microwave signal may be drowned in the background noise and thus leads to a relatively larger uncertainty of the fitted frequency or even no triplet. On the contrary, the microwave signal will heat the resonator too much, this will increase the temperature control instability and in turn reduce the frequency stability. Thus, it is necessary to optimize the microwave emission power.

Fig. 3 shows the optimized results of the power at 5.0 K under vacuum based on the TM₁₁, TE₁₁, TM₁₂ and TE₁₃ microwave modes, which were employed in our preliminary measurements [17]. A bigger power is easier to increase the temperature instability, as shown in Fig. 3a and Fig. 3b, where the temperature stabilities are about 0.042 mK and 0.024 mK for microwave emission power $P_{\text{MW}} = -7$ dBm and $P_{\text{MW}} = -10$ dBm, respectively. Fig. 3c plots the temperature Allan standard deviation against the integration time for these two powers. It can be seen that there is an oscillation of the temperature Allan standard deviation for $P_{\text{MW}} = -7$ dBm, which is coupling with the measurements of the 4 microwave modes, measured in the TM₁₁-TE₁₁-TM₁₂-

TE13 loop. While for $P_{\text{MW}} = -10$ dBm, no oscillation was observed on the temperature Allan standard deviation, which means the microwave power has negligible disturbance effect on temperature stability. Thus the optimal microwave emission power was found to be $P_{\text{opt}} = -10$ dBm. In the present work, we use one single microwave emission power P_{opt} during all the low-temperature microwave frequency measurements no matter whether the resonator is under vacuum or filled with high-purity helium-4 gas.




Fig.3. Microwave emission power optimization at 5.0 K under vacuum. (a) Temperature evolution with $P_{\text{MW}} = -7$ dBm. (b) Temperature evolution with $P_{\text{MW}} = -10$ dBm. (c) Temperature Allan standard deviation σ_T as a function of integration time for $P_{\text{MW}} = -7$ dBm and $P_{\text{MW}} = -10$ dBm.

3.3 Mechanical stability, temperature control method, time reference, and signal intensity

In this section, mechanical stability, temperature control method, time reference and signal intensity were also studied in this work under vacuum unless otherwise specified.

3.3.1 Mechanical stability

When cooling the resonator from room temperature down to low temperatures (5 K to 25 K) for each independent run in this work, it is important to keep the resonator in good mechanical stability before doing the microwave measurements. A simple way to release the mechanical stress is to do temperature cycling in the objective temperature range, for example, cooling and warming between 25 K and 5 K in this work. Because releasing mechanical stress needs some time. Besides, it is also good for temperature and frequency measurements. At the same set point for temperature control, after the temperature cycling, it is easier to realize highly stable temperature measurements, which has been observed in our system. The temperature stability, in turn, will improve the frequency measurement as shown in Section 3.3.2 below. When the temperature of the resonator reached the objective temperatures (5 K and 24.5 K), temperature regulations together with microwave measurements were implemented for several hours. If the frequencies of each microwave mode had no obvious change anymore (changing within microwave measurement uncertainty) at the objective temperatures, then it can be considered that the mechanical stress of the resonator is completely released.

A cooling-warming temperature cycling for 55 hours was carried out in this work. The result is shown in Fig. 4a, where the cooling process and warming process are denoted as C and W, respectively, in the figure legend, and the behind number is the sequence of the temperature cycle. The average amplitude of $|A_p|$ in equation 2 of the cooling process and warming process, denoted as A , was used to indicate the frequency change with the temperature change. They do not completely overlap mainly because of a big difference in temperature change speed between them. This will be improved later in our work by controlling the cooling and warming speed automatically. Generally, after several cooling and warming cycles, microwave measurements can be implemented in the resonator when the temperature and the filled gas pressure are stable simultaneously. This is the case for the present work, where obvious resonance frequency changes were not observed for each microwave mode at the two objective

1 temperatures, respectively.




Fig.4. Optimized results of mechanical stability, temperature control, time reference and signal on TM11 microwave mode under vacuum unless otherwise specified. (a) Evolution of resonator temperature and average amplitude of $|A_p|$ in equation 2 during temperature cycle. (b) Frequency Allan standard deviation with two temperature control methods at $T = 15.0$ K and $p = 30$ kPa. (c) Two-days frequency stability at 24.5 K with different time references. (d) Seven-days frequency stability at 5.0 K for FS740 time reference locked to GPS with Rubidium atomic clock configured indoor antenna. (e) Real and imaginary components of $|S_{21}|$ of the triplet at 5.0 K with and without the two microwave amplifiers. (f) Frequency Allan standard deviation at 5.0 K using FS740 locked to GPS with Rubidium atomic clock configured outdoor antenna and amplifiers.

3.3.2 Temperature control method

Due to thermal expansion effect of the resonator, its dimensions will change if the

1 temperature is not stable. The microwave resonance frequency is proportional to the
2 reciprocal of the radius of the resonator, thus the temperature stability has a direct
3 influence on microwave measurements. Fig. 4b shows the influence on microwave
4 resonance frequency stability of TM11 mode for two temperature control methods at T
5 $= 15.0$ K and $p = 30$ kPa. Method 1 uses an 8½ digit multimeter (Keithley 2002) with
6 Cernox sensor as in our previous work [30], while method 2 uses an AC resistance
7 bridge (ASL F18) with a calibrated rhodium-iron sensor as in CCT-K1 [35]. The
8 temperature stabilities of the two methods are 38 μ K and 8 μ K with an integration time
9 of 33.6 s, respectively. The frequency stability of method 2 is better than that of method
10 1 by about an order of magnitude with an integration time $\tau = 3$ h. Good temperature
11 stability can improve the microwave resonance frequency stability. Thus, in our later
12 measurements, method 2 was adopted all the time to maintain a good frequency stability.

13 3.3.3 Time reference

14 Usually, the vector network analyzer is connected to an external time reference,
15 such as rubidium atomic clock or GPS clock, to realize high-accuracy and high-stability
16 frequency measurements. In this work, rubidium atomic clock (FS725) and GPS clock
17 (FS740 locked to GPS) were used. While the long-term stability of GPS is excellent,
18 its short-term stability is rather poor in comparison to modern oscillators [36]. So in
19 this work, two different oscillators, an OCXO and a rubidium atomic frequency
20 standard, were configured for FS740.

21 Fig. 4c shows the typically measured stability of microwave resonance frequency
22 of TM11 mode at 24.5 K in two days with the three kinds of time references, namely,
23 FS725 (rubidium atomic clock), FS740 locked to GPS with rubidium atomic clock and
24 with OCXO clock, with GPS configured indoor antenna. The frequency stability of
25 FS740 locked to GPS with rubidium atomic clock configured indoor antenna and that
26 of the FS725 are comparatively equivalent, both stabilities are better than that of OCXO
27 clock. A resonance frequency stability of 0.02 ppb was realized using FS740 locked to
28 GPS with rubidium atomic clock configured indoor antenna and FS725 with an
29 integration time $\tau = 3$ h in the two-days microwave measurements. The SPRIGT

measurement at different temperatures and pressures may take one or two months, the time reference stability of the reference should be tested at this level to check if any other effect may perturb the measurement.

Furthermore, long-term microwave measurements at 5 K in seven days were carried out by FS740 locked to GPS with rubidium atomic clock configured indoor antenna. The stability result is plotted in Fig. 4d. A good resonance frequency stability of 0.04 ppb was achieved with $\tau = 3$ h.

3.3.4 Signal intensity

Amplifier is usually used to enhance the microwave signal intensity as shown in literature [37], where the amplitude of $|S_{21}|$ was amplified by nearly ten times with the help of a set of two 10 dBm amplifiers (Model ZX60-14012L manufactured by Mini-Circuits, bandwidth from 0.3 MHz to 14 GHz) connected in series at room temperature. The uncertainty was reduced by a factor of 5 - 12 even with a low-cost vector network analyzer.

In this work, with the same method and two amplifiers like the literature [37], measurements were carried out at 5 K under vacuum. Fig. 4e shows the comparison results of the real and imaginary components of $|S_{21}|$ with and without the two amplifiers. With amplifiers, the amplitudes were amplified by about 10 times and in turn, the fitted relative standard uncertainty of resonance frequency was reduced by a factor of 10 (from 0.3 ppb to 0.031 ppb). This is because the amplifiers increase the resolution of the peaks of the triple. A similar situation appears in TE₁₁, TM₁₂, and TE₁₃ microwave modes, with the fitted relative standard uncertainty of resonance frequency reduced by a factor of 5, 5, and 10, respectively. It is a simple and effective way to solve signal problems, such as high-loss of microwave cable and feed-through, even low-cost vector network analyzer. Besides, low-temperature microwave amplifiers are expected to have a better performance than room-temperature amplifiers, and this will be one solution of ultra-high accuracy microwave measurements in the future work.

3.3.5 Overall performance

In Section 3.3.3, the microwave measurements were carried out by using the indoor

1 antenna, which was placed outside a window of the laboratory building, facing north
2 against another building. Actually, the FS740 was supplied with two types of antenna,
3 indoor antenna and outdoor antenna. According to the user's manual of the FS740 time
4 reference, best results can be realized if the antenna has a clear unobstructed view of
5 the sky. It is highly recommended to put the outdoor antenna on the roof of the building,
6 within which the FS740 is located. Because it improves the quality and reliability of
7 the GPS signals as more satellites will typically be visible and with a higher signal-to-
8 noise ratio. This can improve the long-term stability of the FS740 time reference by a
9 factor of three. This configuration was used for the following test described in this
10 Section.

11 Combining the above optimized results, namely, second-order background
12 polynomial, optimized power $P_{\text{opt}} = -10$ dBm, good mechanical stability, temperature
13 control using an AC resistance bridge and rhodium-iron resistance sensor, the time
14 reference FS740 locked to GPS with rubidium atomic clock but configured outdoor
15 antenna and two room-temperature amplifiers connected in series, preliminary stability
16 measurements on TM11 mode were implemented at 5 K under vacuum. Fig. 4f plots
17 the Allan standard deviation of the resonance frequency. With the same time reference,
18 FS740 locked to GPS with rubidium atomic clock, the stability for outdoor antenna
19 with amplifiers is about an order of magnitude better than that for only indoor antenna
20 as shown in Fig. 4d and Fig. 4f. High-stability of 3.7 parts per trillion (ppt, 1 ppt \equiv
21 $10^{-12} \equiv 10^{-3}$ ppb) and low-uncertainty of 5.2 ppt were realized with an integration
22 time of 3 h. The frequency stability and uncertainty in the present work are nearly 20
23 times better than those in our previous work [17], 0.064 ppb and 0.095 ppb with the
24 same integration time and $P_{\text{opt}} = -10$ dBm [1].

25 In SPRIGT, the thermodynamic temperature uncertainty has many influence
26 factors, frequency is one of the major factors as shown in our previous work [1] and the
27 RIGT review paper [2]. The improvement in frequency measurement accuracy can
28 reduce its influence on thermodynamic temperature measurement by a factor of 20,
29 which is good for realizing the high-accuracy measurement of thermodynamic
30 temperature by SPRIGT. The above optimization methods have the potential to be used

in gas metrology and other research fields, where high-stability and low-uncertainty microwave measurements are necessary.

3.4 Results with pressures

After many experimental optimizations under vacuum, preliminary microwave measurements were performed from 5 K to 26 K with the resonator filled by high-purity helium-4 at the following conditions: 1) fitting S_{21} with 2nd order background polynomial; 2) setting microwave emission power $P = P_{\text{opt}} = -10$ dBm; 3) doing temperature cycle to make the resonator with a good mechanical stability; 4) controlling temperature by AC resistance bridge and rhodium iron sensor; 5) using time reference FS725; and 6) without the two amplifiers. The measurements were implemented on isobars and the results are described in the Fig. 5.

Fig. 5a shows the total relative standard uncertainty of $\langle f+g \rangle$ of TM11 mode at temperatures from 5 K to 26 K and pressures up to 120 kPa with $\tau = 80$ s⁴. Most uncertainties are located within 0.9 ppb, except the p - T region with pressures near 120 kPa and temperatures near 5 K. The main reason is that at lower temperatures, the pressure stability inside the resonator becomes a little worse than that at higher temperatures, which disturbed the frequency stability by the change of helium-4 density inside the resonator. Fig. 5b plots a one-day microwave measurement of TM11 mode at 24.5 K with different pressures. The frequency stabilities are about (0.03, 0.02, 0.04, 0.05) ppb for (30, 60, 90, 120) kPa with $\tau = 3$ h, respectively, and the temperature stabilities are about (6.9, 8.3, 7.9, 6.6) μ K with $\tau = 33.6$ s. Due to the pressure stability, 60 kPa has the best frequency stability in the one-day stability measurement. These are still better than our previous work (0.063 ppb under vacuum provided with $P_{\text{opt}} = -10$ dBm) even the pressure instability increases the microwave frequency measurement

⁴It is usually used a constant integration time for all the microwave modes. Figure 5(a) only shows the uncertainty of TM11 microwave mode for different pressures and temperatures. However, for TE13 mode, the uncertainties are very near 2 ppb. Besides, the integration time was also chosen to be in the same “time” as the other instruments, such as the resistance bridge. The integration time of 20 000 s is used to check and characterize the stability of the measurements.

instability. The stability and uncertainty still have room for improvement, provided later using the time reference, FS740 locked to GPS with rubidium atomic clock configured outdoor antenna, and microwave amplifiers.

Thermodynamic temperatures calculated from different modes were determined for the isobar $p = 60$ kPa, based on the SPRIGT principle [1]. Fig. 5c and Fig. 5d presents the differences between mode thermodynamic temperatures T_{mode} and the average values T_{avg} , ($T_{\text{mode}} - T_{\text{avg}}$), at temperatures from 5 K to 26 K, where microwave measurements with $\tau = 80$ s were used. Good mode consistency was observed on ($T_{\text{mode}} - T_{\text{avg}}$) within ± 80 μK , and a small relative standard uncertainty component for the thermodynamic temperatures, from mode consistency, was realized with $u(T_{\text{mode}})$ values less than 60 μK ⁵.




Fig.5. Microwave measurements from 5 K to 26 K at pressures up to 120 kPa. (a) Total relative standard uncertainty of $\langle f+g \rangle$ for TM11 mode with $\tau = 80$ s. (b) One-day frequency stability measurements at 24.5 K for (30, 60, 90, 120) kPa. (c) Thermodynamic temperature differences

⁵Uncertainty budget for SPRIGT has been reported in our previous work [1] and the review paper of RIGT [2]. Since this work is focused on microwave measurements, it will be conducted in our future work.

($T_{\text{mode}} - T_{\text{avg}}$) for TM11, TE11 and TE13 microwave modes at 60 kPa. (d) Standard uncertainty component of SPRIGT thermodynamic temperature from mode consistency.

4 Conclusion

The object of this study was the optimization of microwave frequency measurements at low temperatures. The measurements in question concern the resonance frequencies of a high quality factor, tri-axial ellipsoidal copper resonator either under vacuum or else filled with high-purity helium-4 gas (at pressures from 30 kPa to 120 kPa). The ratio of frequencies with and without gas yields the refractive index from which temperature can be determined with high accuracy once the effects of pressure are taken into account. This is called refractive index gas thermometry (RIGT). The application here concerns a novel variant of the technique exploiting single isobars such that pressure-dependent effects almost cancel – single-pressure RIGT or SPRIGT. Microwave resonances are detected using pairs of antennas linked to a vector network analyzer.

The study dealt with both experimental parameters and data analysis. Frequency was characterized using the two-sample (Allan) variance widely used in time and frequency metrology. To minimize self-heating yet with an acceptable signal-to-noise ratio, a microwave power of -10 dBm was found to be optimal. A slight improvement of the quality of the fitted lineshape compared with our previous work [17] was obtained using a quadratic function for the background rather than a linear one. The use of temperature cycling between 5 K and 25 K turned out to improve the reproducibility of frequency measurements. Fine control of the resonator temperature is achieved using a calibrated rhodium-iron resistance thermometer read using an AC resistance bridge instead of an 8½ digit multimeter. The local frequency reference was a GPS referenced commercial rubidium clock. It was essential to use a roof-based outdoor antenna for the GPS system. Frequency resolution was also improved by the introduction of two microwave pre-amplifiers in series, albeit close to the network analyzer located at room temperature. Increased performance is expected with the installation of cryogenically cooled pre-amplifiers located closer to the microwave resonator.

1 In this way, for an integration time of three hours, microwave resonance frequency
2 measurements with a stability at the 10^{-12} level were demonstrated, which represents an
3 almost 20 fold enhancement compared with those of our previous work [17]. The
4 present observations could help in other high-stability, low-uncertainty frequency
5 measurements in copper microwave cavities, not only for improving thermodynamic
6 temperature measurement in SPRIGT, but also for other precise measurements, where
7 high-stability and low-uncertainty microwave measurements are necessary.

8 The final uncertainty achieved using SPRIGT in our laboratory is equal to or better
9 than 0.17 mK in the range 5 K to 25 K. An extensive description of the uncertainty
10 budget will be presented in a future publication. As far as frequency measurements are
11 concerned (*i.e.*, the subject of the present manuscript), had they not been possible at the
12 10^{-12} level, the subtle influence of microwave cables and feed-thoughts would perhaps
13 not have been detected as early as it was. The effect of thermal cycling is not one of
14 uncertainty but rather of potential mechanical instability.

15 Acknowledgments

16 This work is supported financially by the National Key R&D Program of China (Grant
17 No. 2016YFE0204200), the National Natural Science Foundation of China (Grant No.
18 51627809), the International Partnership Program of the Chinese Academy of Sciences
19 (Grant No. 1A1111KYSB20160017), and the EMRP project Real-K (No. 18SIB02).
20 Author Changzhao Pan was supported by the funding provided by the Marie
21 Skłodowska-Curie Individual Fellowships-2018 (834024). The authors gratefully
22 acknowledge Richard Rusby from National Physical Laboratory UK for sharing his
23 long experience and constant advice on temperature measurements. We are deeply
24 grateful to Wei Wu from City University of Hong Kong China for his kind reading and
25 helpful suggestions on the manuscript. We would also like to thank Kun Liang from the
26 National Institute of Metrology China and Chongxia Zhong from Beijing Institute of
27 Metrology China for their helpful discussion on time reference.

28 References

29 [1] Gao B, Pitre L, Luo E C, Plimmer M D, Lin P, Zhang J T, Feng X J, Chen Y Y,
30 Sparasci F 2017 Feasibility of primary thermometry using refractive index
31 measurements at a single pressure *Measurement* **103** 258-262.

- 1 [2] Rourke P M C, Gaiser C, Gao B, Madonna Ripa D, Moldover M R, Pitre L,
2 Underwood R J 2019 Refractive-index gas thermometry *Metrologia* **56** 032001.
- 3 [3] Bruch L, Weinhold F 2000 Diamagnetism of helium *J. Chem. Phys.* **113** 8667-8670.
- 4 [4] Rizzo A, Hättig C, Fernández B, Koch H 2002 The effect of intermolecular
5 interactions on the electric properties of helium and argon. III. Quantum statistical
6 calculations of the dielectric second virial coefficients *J. Chem. Phys.* **117** 2609-2618.
- 7 [5] Garberoglio G, Moldover M R, Harvey A H 2011 Improved first-principles
8 calculation of the third virial coefficient of helium *Journal of research of the National*
9 *Institute of Standards and Technology* **116** 729.
- 10 [6] Cencek W, Przybytek M, Komasa J, Mehl J B, Jeziorski B, Szalewicz K 2012
11 Effects of adiabatic, relativistic, and quantum electrodynamics interactions on the pair
12 potential and thermophysical properties of helium *J. Chem. Phys.* **136** 224303.
- 13 [7] Shaul K R, Schultz A J, Kofke D A 2012 Path-integral Mayer-sampling calculations
14 of the quantum Boltzmann contribution to virial coefficients of helium-4 *J. Chem. Phys.*
15 **137** 184101.
- 16 [8] Puchalski M, Piszczatowski K, Komasa J, Jeziorski B, Szalewicz K 2016
17 Theoretical determination of the polarizability dispersion and the refractive index of
18 helium *Phys. Rev. A* **93** 032515.
- 19 [9] Balle C, Casas-Cubillos J, Vauthier N, Thermeau J, Calibration of Cryogenic
20 Thermometers for the LHC, in: AIP Conference Proceedings, American Institute of
21 Physics, 2008, pp. 965-972.
- 22 [10] Pitre L, Moldover M R, Tew W L 2006 Acoustic thermometry: new results from
23 273 K to 77 K and progress towards 4 K *Metrologia* **43** 142–162.
- 24 [11] Pitre L, Sparasci F, Truong D, Guillou A, Risegari L, Himbert M E 2011
25 Measurement of the Boltzmann constant k_B using a quasi-spherical acoustic resonator
26 *Int J Thermophys* **32** 1825.
- 27 [12] Pitre L, Sparasci F, Truong D, Guillou A, Risegari L, Himbert M E 2011
28 Determination of the Boltzmann constant using a quasi-spherical acoustic resonator
29 *Phil. Trans. R. Soc. A* **369** 4014-4027.
- 30 [13] Moldover M R, Gavioso R M, Mehl J B, Pitre L, De Podesta M, Zhang J 2014
31 Acoustic gas thermometry *Metrologia* **51** R1.
- 32 [14] Moldover M R, Gavioso R M, Newell D B 2015 Correlations among acoustic
33 measurements of the Boltzmann constant *Metrologia* **52** S376.
- 34 [15] Pitre L, Sparasci F, Risegari L, Guianvarc'h C, Martin C, Himbert M E, Plimmer
35 M D, Allard A, Marty B, Giuliano Albo P A, Gao B, Moldover M, Mehl J B 2017 New

1 measurement of the Boltzmann constant k by acoustic thermometry of helium-4 gas
2 *Metrologia* **54** 856.

3 [16] Gavioso R M, Madonna Ripa D, Steur P P M, Dematteis R, Imbraguglio D 2019
4 Determination of the thermodynamic temperature between 236 K and 430 K from speed
5 of sound measurements in helium *Metrologia* **56** 045006.

6 [17] Zhang H Y, Liu W J, Gao B, Chen Y Y, Pan C Z, Song Y N, Chen H, Han D X, Hu
7 J F, Luo E C, Pitre L 2019 A high-stability quasi-spherical resonator in SPRIGT for
8 microwave frequency measurements at low temperatures *Sci. Bull.* **64** 286-288.

9 [18] Rourke P M C, Hill K D 2015 Progress Toward Development of Low-Temperature
10 Microwave Refractive Index Gas Thermometry at NRC *Int. J. Thermophys.* **36** 205-
11 228.

12 [19] Rourke P M C 2017 NRC Microwave Refractive Index Gas Thermometry
13 Implementation Between 24.5 K and 84 K *Int. J. Thermophys.* **38** 107.

14 [20] Rourke P M C 2019 Thermodynamic temperature of the triple point of xenon
15 measured by refractive index gas thermometry *Metrologia* **57** 024001.

16 [21] Gambette P, Gavioso R, Ripa D M, Plimmer M, Pitre L 2018 Towards a Quantum
17 Standard for Absolute Pressure Measurements in the Range 200 Pa to 20 kPa Based on
18 a Superconducting Microwave Cavity *Conference on Precision Electromagnetic*
19 *Measurements* paper no. 393.

20 [22] Yeh N-C, Jiang W, Strayer D, Asplund N 1996 Precise measurements of the density
21 and critical phenomena near the phase transitions in helium using high-Q niobium
22 microwave cavities *Czech. J. Phys.* **46** 181-182.

23 [23] Yeh N-C, Strayer D, Anderson V, Asplund N 2000 Superconducting-cavity-
24 stabilized oscillators (SCSO) for precise frequency measurements *Phys. B* **280** 557-558.

25 [24] Corcovilos T, Strayer D, Asplund N, Yeh N-C 2004 Mugti-frequency
26 Superconducting Cavity Stabilized Oscillators (SCSO) for Quantum-Gas
27 Measurements and Gravitational Physics *J. Low Temp. Phys.* **134** 431-436.

28 [25] <https://www.bipm.org/utls/en/pdf/si-mep/SI-App2-second.pdf>.

29 [26] Allan D W 1966 Statistics of atomic frequency standards *Proceedings of the IEEE*
30 **54** 221-230.

31 [27] Audoin C, Guinot B, The measurement of time: time, frequency and the atomic
32 clock, Cambridge University Press, 2001.

33 [28] Drever R, Hall J L, Kowalski F, Hough J, Ford G, Munley A, Ward H 1983 Laser
34 phase and frequency stabilization using an optical resonator *Appl. Phys. B* **31** 97-105.

35 [29] Gao B, Pan C Z, Chen Y Y, Song Y N, Zhang H Y, Han D X, Liu W J, Chen H,

- 1 Luo E C, Pitre L 2018 Realization of an ultra-high precision temperature control in a
2 cryogen-free cryostat *Rev. Sci. Instrum.* **89** 104901.
- 3 [30] Chen Y Y, Zhang H Y, Song Y N, Pan C Z, Gao B, Liu W, Chen H, Han D, Luo E,
4 Plimmer M D, Pitre L 2019 Thermal response characteristics of a SPRIGT primary
5 thermometry system *Cryogenics* **97** 1-6.
- 6 [31] Han D X, Gao B, Chen H, Gambette P, Zhang H Y, Pan C Z, Song Y N, Liu W J,
7 Hu J F, Yu B, Luo E C, Pitre L 2018 Ultra-stable pressure is realized for Chinese single
8 pressure refractive index gas thermometry in the range 30–90 kPa *Sci. Bull* **63** 1601-
9 1603.
- 10 [32] Liu W J, Pitre L, Gao B, Zhang H Y, Chen Y Y, Plimmer M D, Luo E C, Sparasci
11 F, Song Y N, Chen H, Pan C Z, Yu B, Hu J Y 2018 Microwave Method for Closure
12 of Quasi-spherical Resonator *Conference on Precision Electromagnetic Measurements*
13 *(CPEM)* paper no. 452.
- 14 [33] Moré J J, The Levenberg-Marquardt algorithm: implementation and theory, in:
15 Numerical analysis, Springer, 1978, pp. 105-116.
- 16 [34] May E F, Pitre L, Mehl J B, Moldover M R, Schmidt J W 2004 Quasi-spherical
17 cavity resonators for metrology based on the relative dielectric permittivity of gases
18 *Rev. Sci. Instrum.* **75** 3307-3317.
- 19 [35] Rusby R, Head D, Meyer C, Tew W, Astrov D 2006 Key comparison: Final Report
20 on CCT-K1: Realizations of the ITS90, 0.65 K to 24.5561 K, using rhodium iron
21 resistance thermometers *Metrologia* **43** 3002-03002.
- 22 [36] <https://www.thinksrs.com/downloads/pdfs/manuals/FS740m.pdf>.
- 23 [37] Yang I, Underwood R, De Podesta M 2018 Investigating the adequacy of a low-
24 cost vector network analyser for microwave measurements in quasispherical resonators
25 *Meas. Sci. Technol.* **29** 075013.
- 26