

# Wireless light-emitting device for the determination of chirality in real samples

Gerardo Salinas, Giorgia Bonetti, Roberto Cirilli, Tiziana Benincori,

Alexander Kuhn, Serena Arnaboldi

## ▶ To cite this version:

Gerardo Salinas, Giorgia Bonetti, Roberto Cirilli, Tiziana Benincori, Alexander Kuhn, et al.. Wireless light-emitting device for the determination of chirality in real samples. Electrochimica Acta, 2022, 421, pp.140494. 10.1016/j.electacta.2022.140494 . hal-04050571

# HAL Id: hal-04050571 https://hal.science/hal-04050571

Submitted on 29 Mar 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

1	Wireless light-emitting device for the determination of chirality in real samples
2 3	Gerardo Salinas, <sup>1</sup> Giorgia Bonetti, <sup>2</sup> Roberto Cirilli, <sup>3</sup> Tiziana Benincori, <sup>2</sup> Alexander Kuhn <sup>1</sup> , Serena Arnaboldi <sup>4*</sup>
4 5 7 8 9 10	<ul> <li><sup>1</sup>Univ. Bordeaux, CNRS, Bordeaux INP, ISM, UMR 5255, F-33607 Pessac, France</li> <li><sup>2</sup>Dip. di Scienza e Alta Tecnologia, Univ. degli Studi dell'Insubria, Como, Italy</li> <li><sup>3</sup>Istituto Superiore di Sanità, Centro Nazionale per il Controllo e la Valutazione dei Farmaci, Rome, Italy</li> <li><sup>4</sup>Dip. di Chimica, Univ. degli Studi di Milano, Via Golgi 19, 20133 Milano, Italy</li> <li>*Corresponding author contact: serena.arnaboldi@unimi.it</li> </ul>
11	Abstract
12	Bipolar electrochemistry can be employed in the context of chiral recognition in order to
13	obtain useful analytical readouts of enantiomeric analytes. Herein, we employ this concept
14	for the simultaneous determination of two enantiomers present in solution, and its possible
15	use as transduction mechanism for complex real matrices analysis. This approach is based
16	on the combination of the enantioselective electrooxidation of only one of the two
17	antipodes of a chiral analyte with the emission of light from light-emitting diodes (LEDs).
18	A double hybrid device was designed, using the enantiomers of an inherently chiral
19	oligomer and a bare gold wire as the anode and cathode of a green and red LED. By

20 applying an appropriate voltage, the wirelessly induced redox reactions trigger light 21 emission only when the probe with the right configuration is present in solution. This 22 device was used to simultaneously measure the ratio between L- and D-Tryptophan, both

present in the same solution and to quantify L-ascorbic acid in a commercial juice sample. The measurement correlates with the value reported on the sample specifications. These results illustrate the possible use of such light emitting bipolar devices as analytical tools for qualitative and quantitative measurements of enantiomeric excess, even in real samples.

### 28 Keywords

Bipolar electrochemistry, chirality, light emitting diodes, enantiomeric excess, real sampleanalysis.

#### 31 **1. Introduction**

32 Amino acids are one of the most important types of natural compounds due to their 33 involvement in essential and well-known physiological processes. For example, in biology, commonly the L-form of amino acids largely dominates, thus, high levels of certain D-34 35 amino acids have been attributed to bacterial activity in food or to different pathologies, such as chronic kidney disease and neurological disorders, like schizophrenia and 36 Alzheimer's disease. Therefore, simultaneous monitoring and quantification of D- and L-37 amino acids for biomedical applications, as a marker of quality, authenticity and food 38 safety, is highly needed [1]. Commonly, simultaneous detection of both enantiomers of 39 amino acids has been carried out mainly through high performance liquid chromatography, 40 41 which requires time-consuming protocols and expensive instrumentation. In addition, alternative procedures such as capillary electrophoresis [2], reactive extraction [3], or 42 43 ultrafiltration through an enantioselective membrane have been developed [4]. Recently a different approach to distinguish the enantiomers of two classes of amino acids in blood 44 and plasma has been proposed, involving "on the fly" bio-detection [5]. However, such thus 45 type of detection was only possible by analyzing the enantiomers in separate solutions, 46 limiting the analysis of mixtures, often present in real samples. An interesting alternative is 47 the development of fast and straightforward optical read-outs of different electroactive 48 49 analytes [6-9]. Recently electrochemistry-based light emitting systems have gained considerable attention due to the possible direct visualization of chemical information. 50

Different light emitting systems, based on electroluminescence [10] or fluorescence [11] 51 52 have been exploited for the quantification of different analytes of interest [12-16]. However, these methods require the use of additional chemicals, such as luminophores, co-53 reactants, fluorescent dyes, and often the light emission is triggered by complex reactions 54 55 pathways. A promising alternative is the use of light emitting diodes (LEDs) for the direct visualization of electrochemical processes. In such devices, the analytical information is 56 directly encoded by the electric current passing through the LED and the concomitant light 57 emission. Although light emission can be triggered by a direct electrical connection [17], an 58 interesting approach is to power these devices in a wireless manner, via coupling 59 60 thermodynamically spontaneous reactions to the terminals of the LED [18,19] or by using 61 bipolar electrochemistry (BPE) [20,21]. The wireless nature of the BPE approach has been proven to be very convenient for materials science and analytical applications [22-28]. 62 63 Recently BPE has been used for the electroanalytical determination of chiral information coupled with mechanical actuation or light emission [29-31]. These systems are based on 64 65 the enantioselective oxidation of a chiral probe on the surface of an inherently chiral oligomer, coupled with the reduction of a pristine polypyrrole strip, leading to a selective 66 electron-flow. The selectivity of these systems is based on the diastereomeric interactions 67 68 between the inherently chiral oligomer and the antipodes dissolved in solution. Therefore, the actuation of the polypyrrole strip or the light emission occurs only in the presence of the 69 right enantiomer. Herein, we take advantage of such a highly enantioselective light 70 71 emission to design a double light-emitting bipolar recognition device for the simultaneous 72 analysis of enantiomeric mixtures of amino acids in real samples. The light emitting bipolar device is designed by combining three main ingredients; (i) an inherently chiral oligomer, 73 74 which is the site of enantioselective oxidation; (ii) pristine gold wires, where the reduction

reaction occurs; and (iii) a micro-LED for the light emission (Scheme 1). This hybrid 75 76 device is placed in a solution containing enantiomers of two different chiral model analytes; Tryptophan (Tryp) and Ascorbic Acid (AA). Tryp is an essential amino acid used 77 78 commonly to balance the nitrogen content in adults and also involved in the growth of 79 infants. Tryp antipodes were chosen to be simultaneously detected and quantified by using the double LEDs bipolar device. Finally, AA was chosen for its high abundance in fresh 80 fruit juice, especially blackcurrants, citrus fruit and strawberries. In addition, L-ascorbic 81 82 acid is one of the most important water-soluble vitamins, since it acts as a free-radical scavenger of reactive oxygen and nitrogen species [32,33]. Due to its relatively low 83 oxidation potential, different classic electrochemical methods have been used for ascorbic 84 acid quantification in complex media [34]. However, a straightforward visual readout of the 85 enantiorecognition of such a chiral probe has not been explored so far. 86

#### 87 **2. Experimental**

LiClO<sub>4</sub> (Aldrich, 99%), anhydrous acetonitrile (Aldrich, 99.9%), Au wires (d = 0.25 mm, 88 Alfa Aesar, 99.9%), L-ascorbic acid (ACS reagent,  $\geq$  99%), L- and D-Tryptophan (Aldrich, 89  $\geq$  99%), agar silver paint (Agar scientific Ltd.) and miniaturized red (0603 SMD diode, 90 91 Wurth Elektronik, 1.60 mm  $\times$  0.8 mm) and green light-emitting diodes (LED, 0603 SMD) diode, OSRAM Opto Semiconductors, 1.70 mm  $\times$  0.8 mm) were used as received. 92 Enantiopure (*R*)- or (*S*)-2,2'-bis(2-(5,2'-bithienyl))-3,3'-bithianaphthene ((*R*)-BT<sub>2</sub>T<sub>4</sub> and (*S*)-93 94  $BT_2T_4$ , scheme 1a) were synthetized following a published methodology [35]. All solutions were prepared with deionized water (MilliQ Direct-Q<sup>®</sup>). 95

Enantiopure oligo-(R) and oligo-(S)-BT<sub>2</sub>T<sub>4</sub> films were obtained according to previous 96 97 reports [29-31]. Electrodeposition was carried out in a 0.1 M LiClO<sub>4</sub>/ACN solution, containing 1 mM of monomer by chronopotentiometry applying 1.5 mA for 25 minutes. A 98 99 gold wire was used as working electrode together with a platinum mesh and Ag/AgCl as 100 counter and reference electrodes, respectively. The double bipolar devices were assembled by connecting a wire modified with the corresponding oligo-BT<sub>2</sub>T<sub>4</sub> film and an unmodified 101 Au wire to the anode and the cathode of the LEDs, respectively (Scheme 1b). In brief, the 102 103 Au electrodes with the desired dimensions and the LED were attached to a glass support using double-sided adhesive tape. A small volume of agar silver paint was employed to 104 105 establish an electric connection between the electrodes and the LED. The total length of the 106 final bipolar electrode is 1.8 cm. For the bipolar recognition, the enantiopure oligo-(R)- and 107 oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au devices were fixed in parallel at the center of a bipolar cell. Two 108 graphite feeder electrodes were positioned at the extremities of the cell at a distance of 5 cm. The first recognition experiments were carried out in an aqueous 5 mM LiClO<sub>4</sub> 109 solution in the presence of different ratios of the Tryp enantiomers (90:10 and 10:90) by 110 varying the electric field from 0 V cm<sup>-1</sup> to 3.6 V cm<sup>-1</sup>. The experiments with different 111 concentrations of L-AA (2, 4, 6, 8 and 10 mM) were carried out at a constant electric field 112 (2 V cm<sup>-1</sup>). The standard addition method (from 2 to 10 mM) was used to carry out the 113 quantification of L-AA in a commercial lemon juice sample, at a constant electric field (2 V 114 cm<sup>-1</sup>). Experiments were monitored by using a CCD camera (CANON EOS 70D, Objective 115 116 Canon Macro Lens 100 mm 1:2.8). Images were processed with Image J software.



118 Scheme 1. (a) Chemical structures of the two enantiomers of  $BT_2T_4$  oligomers. (b) 119 Schematic illustration of the bipolar set-up used for the wireless quantification of chiral 120 probes based on the double light emission with a representation of the associated chemical 121 reactions. The orange part stands for the  $BT_2T_4$  oligomer. The distance between the feeder 122 electrodes is 5 cm and the length of the bipolar object is 1.8 cm.

#### 123 **3. Results and Discussion**

In order to evaluate the electric field required to switch on the LED integrated in each 124 125 bipolar electrode, a non-enantioselective green and red Au-LED-Au system was tested first. The double LED device was placed at the center of a bipolar cell, containing an aqueous 5 126 mM LiClO<sub>4</sub> solution. To induce redox reactions in such a set-up, an electric field, high 127 enough to sufficiently polarize the bipolar electrode at its extremities, was applied. Under 128 these conditions, the oxidation and reduction of water at the anode and cathode of the LED, 129 respectively, trigger the light emission of the diode. At low electric field values (below 2 V 130 cm<sup>-1</sup>), there is not enough driving force to induce these redox reactions, thus both LEDs 131 remains in the off-state (Figure S1). Once the electric field reaches a threshold value 132 (around 2 and 2.4 V cm<sup>-1</sup> for the green and red LED, respectively) the polarization potential 133

difference, with respect to the surrounding solution, induces the oxidation and reduction of
water at the extremities of both Au-LED-Au bipolar electrodes. This creates an electron
flow from the anodic to the cathodic side of the LED and thus a concomitant light emission.
As it can be seen from Figure S1, the emitted light intensity is a function of the electric
field value. However, the difference of threshold value (0.4 V cm<sup>-1</sup>) between the green and
red LED is attributed to the intrinsic resistance of the diode.

After the evaluation of the threshold potential of a non-enantioselective set up, the 140 141 influence of the inherently chiral oligomer on the threshold potential of the same system, in 142 the absence of chiral probe, was studied. At first, a device combining a red oligo-(S)-143  $BT_2T_4$ -LED-Au and green oligo-(R)-BT\_2T\_4-LED-Au system was placed at the center of a 144 bipolar cell containing only an aqueous 5 mM LiClO<sub>4</sub> solution. In the absence of the chiral 145 probe, water oxidation and reduction remain the redox reactions occurring at the 146 extremities of the bipolar electrode. Thus, a similar behavior of the light intensity as a function of the applied electric field was observed for the double functionalized device 147 (Figure 1) in comparison with the non-enantioselective system (Figure S1). 148



Figure 1. Maximum light intensity as a function of the applied electric field for the combination of a red oligo-(*S*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au and green oligo-(*R*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au system in an aqueous solution of 5 mM LiClO<sub>4</sub>.

Below 2 V cm<sup>-1</sup> both LEDs remains switched-off, whereas, around 2 and 2.4 V cm<sup>-1</sup> (for the green and red LED, respectively) the electric field reaches the threshold value inducing the light emission. An identical response was observed for the inverted configuration, combining a red oligo-(R)-BT<sub>2</sub>T<sub>4</sub>-LED-Au and green oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au (Figure S2), with a similar difference of threshold potential (0.4 V cm<sup>-1</sup>) between the green and red device. Thus, in the presence of the inherently chiral oligomers and in the absence of a chiral probe, the threshold potential is governed by the intrinsic resistance of the LED.

160 The first step, before testing the device in the presence of a chiral probe, was to verify by 161 differential pulse voltammetry (DPV) the enantioselectivity of (R)- and (S)-BT<sub>2</sub>T<sub>4</sub> 162 oligomers deposited on a gold wire with respect to the enantiomers of Tryptophan, L-Tryp

and D-Tryp, in separate solutions (4.5 mM). The DPV shows a peak-to-peak separation of 163 approximately 700 mV on (S)-oligomer modified electrodes for the electrooxidation of the 164 Tryp enantiomers (Figure S3). This very high thermodynamic differentiation allows the 165 simultaneous detection of both Tryp enantiomers with a double light-emitting device. The 166 167 red oligo-(S) and green oligo-(R)- $BT_2T_4$ -LED-Au system was placed in the center of a 168 bipolar cell containing an aqueous 0.5 mM LiClO<sub>4</sub> solution. Theoretically, in the presence 169 of both chiral probes, by varying the electric field, both LEDs will switch on, and their light 170 intensity should correlate with the concentration of the corresponding enantiomer. For these 171 measurements, a ratio of 90:10 D:L-Tryp (4.5, 0.5 mM, respectively), and the specular one 10:90, have been chosen. 172



Figure 2. Maximum light intensity as a function of the applied electric field for a red oligo-(*S*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au and green oligo-(*R*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au system in (a) an aqueous solution of 5 mM LiClO<sub>4</sub> containing 4.5 mM D-Tryp and 0.5 mM L-Tryp and (b) an aqueous solution of 5 mM LiClO<sub>4</sub> containing 0.5 mM D-Tryp and 4.5 mM L-Tryp.

Under these conditions, the light intensity for the oligo-(*S*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au device is higher than the one related to the oligo-(*R*)- BT<sub>2</sub>T<sub>4</sub>-LED-Au when the double recognition device is placed in the 90 D : 10 L solution (Figure 2a). The specular response was obtained by

181 placing the same device in the 10 D : 90 L solution (Figure 2b). This is coherent with the

DPV measurements (Figure S3) showing a preferential oxidation of D-Tryp on the (*S*)oligomer modified wire. When comparing the light intensity for the oxidation of D- (250 a.u.) and L-Tryp (25 a.u.) at 2.2 V cm<sup>-1</sup>, the values correlate with the 90:10 D-Tryp/L-Tryp ratio present in solution. Similar results were obtained with the opposite ratio, D- (10 a.u.) and L-Tryp (210 a.u.) at 2.2 V cm<sup>-1</sup>,

187 In the next step, the threshold potential of the device in the presence of L-AA was evaluated. The diastereomeric interaction between the (S)-oligomer and L-AA is more 188 189 favorable from a thermodynamic point of view. Differential pulse voltammograms show a 190 peak-to-peak separation of approximately 100 mV for the electrooxidation of L-AA on (S)-191 and the (R)-oligomer modified electrodes, respectively (Figure S4). Although, this peak 192 separation is smaller in comparison with those observed for other chiral probes (e.g. 600 193 mV for DOPA enantiomers) [29-31], it is still high enough for bipolar recognition 194 experiments. Since for this set of experiments the enantiomeric probe was available in only 195 one configuration, the double device was built with only the green LEDs that are the ones having a lower intrinsic resistance (Figure S5a). The green oligo-(S) and oligo-(R)- $BT_2T_4$ -196 197 LED-Au system was placed at the center of a bipolar cell containing an aqueous 5 mM LiClO<sub>4</sub> solution in the presence of 10 mM L-AA. Under these conditions, the threshold 198 potential for generating light at the (S) bipolar electrode is lower (1.4 V cm<sup>-1</sup>), than the one 199 needed for the (*R*) device to switch on  $(1.8 \text{ V cm}^{-1})$ , thus allowing chiral discrimination. An 200 analog light emission behavior was observed when the red oligo-(S) and oligo-(R)-BT<sub>2</sub>T<sub>4</sub>-201 LED-Au system was used (Figure S5b), with a similar difference in threshold values (0.4 V 202 cm<sup>-1</sup>) between the enantiopure electrodes. 203



204

Figure 3. Maximum light emission as a function of L-AA concentration measured by the enantioselective oxidation of L-AA on a double green oligo-(R)- and oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device, in a 5 mM LiClO<sub>4</sub> solution at a constant applied electric field (2 V cm<sup>-1</sup>). Inset: optical pictures of the (S)- and (R)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device at a given L-AA concentration of 8mM.

After demonstrating chiral discrimination through an on-off light emission, the next step was to evaluate the possibility of using the concept for the quantitative analysis of L-AA. The light intensity was measured as a function of the enantiomer concentration. Five L-AA samples were analyzed (2, 4, 6, 8 mM and 10 mM) with a green oligo-(R)- and oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device by applying a constant electric field (2 V cm<sup>-1</sup>). For each concentration, the average light intensity over a period of 10 sec has been used, in order to eliminate small fluctuations of the light intensity during the experiment caused by the fast

consumption of the analyte. A plot of the light intensity as a function of L-AA 217 concentration reveals a linear correlation ( $R^2 = 0.983$ ) for the oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au 218 with a high sensitivity ( $m_{(S)} = 24.3$  a.u./mM), in contrast to the oligo-(R)-BT<sub>2</sub>T<sub>4</sub>-LED-Au, 219 220 which shows, as expected, almost no signal (Figure 3). This sensitivity and linearity is 221 comparable to the one obtained with classic electrochemical systems that require sophisticated electrode materials [36,37]. This illustrates the possible use of such a light 222 emitting bipolar device as an analytical tool for the qualitative and quantitative 223 224 measurement of enantiomers with relatively close oxidation potential.

Finally, in order to test the reliability of the results in real samples, the quantification of L-AA in commercial lemon juice was performed. Recently, different bipolar electrochemical light emitting systems, based on electrochemiluminescence, have been developed for the quantification of AA in aqueous solution [38,39]. However, the light response is based on indirect electrochemical mechanisms, without the possibility to directly distinguish between the enantiomers of the analyte.

Before starting the experiments, the sample had to be diluted (12.5%) due to the turbidity of the medium, which might interfere with the light response. First of all, light emission of the diluted juice was measured using a green oligo-(*S*)- and oligo-(*R*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device, at constant electric field (2 V cm<sup>-1</sup>). Results show, as expected, enantioselective onoff recognition of L-AA, with the green LED switching on only if connected to the (*S*)enantiomer of the oligomer (Figure 4).



Figure 4. Maximum light emission as a function of L-AA concentration measured by the enantioselective oxidation of L-AA on a double green oligo-(R)- and oligo-(S)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device, in 12.5% commercial lemon juice at a constant electric field (2 V cm<sup>-1</sup>). Error bars refer to measurements carried out three times on the same bipolar device. Inset: optical pictures of the (S)- and (R)-BT<sub>2</sub>T<sub>4</sub>-LED-Au bipolar device at a given L-AA concentration.

The standard addition method was selected for the quantification of L-AA in commercial lemon juice. Five solutions with different standard concentrations, from 2 to 10 mM, were prepared. Plotting the maximum light intensity as a function of concentration leads to a linear correlation ( $R^2 = 0.99$ ) for the oligo-(*S*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au, and again hardly any emission is recorded for the oligo-(*R*)-BT<sub>2</sub>T<sub>4</sub>-LED-Au (Figure 4). By using the regression equation y = 25x + 8.6, with y being the light intensity and x the concentration, a L-AA concentration of 2.7 mmol  $L^{-1}$  was calculated, which is in very good agreement with the L-AA content reported in the lemon juice specifications and in the literature (2.5 mmol  $L^{-1}$ ) [40].

253 4. Conclusions

254 A bipolar electrochemical device, modified with inherently chiral oligomers has been 255 designed for the simultaneous optical determination of the enantiomers of Tryptophan, present in solution, and for the quantification of L-ascorbic acid in real samples. The 256 concept of wireless light emission was successfully used for the determination of the 257 258 enantiomeric ratio between the two configurations of Tryp, both present in the same solution. The average light intensity was found to be directly proportional to the analyte 259 260 concentration and used for the direct quantification of L-ascorbic acid in commercial lemon juice through the standard addition technique. The calculated values are in perfect 261 agreement with those found in the literature and with the ones indicated on the lemon juice 262 specifications. The sensitivity and linearity achieved for L-AA quantification is comparable 263 to those obtained with classic electrochemical systems, that require sophisticated electrode 264 265 materials. The presented approach can thus be considered as an efficient and complementary alternative for chiral discrimination with respect to the common 266 267 spectroscopic methods. Moreover, since the same device has been employed for all the experiments presented herein, it can be used for multiple determinations without noticeable 268 269 fouling effects which might interfere with the measure of light emission. These results open up the possibility to use such double light emitting bipolar devices as straightforward 270 analytical tools for the qualitative and quantitative discrimination of enantiomers, even 271 when they have relatively close oxidation potentials on a given chiral selector layer. 272

15

### 273 5. Acknowledgments

- This work has been funded by the European Research Council (ERC) under the European
- Union's Horizon 2020 research and innovation program (grant agreement no 741251, ERC
- 276 Advanced grant ELECTRA).
- 277 6. References
- [1] E. Rosini, P. D'Antona, L. Pollegioni, Biosensors for D-amino acids: detection methods
  and applications. Int. J. Mol. Sci., 21, (2020), 4574.
- [2] M. Chiari, V. Desperati, E. Manera, R. Longhi, Combinatorial synthesis of highly
  selective cyclohexapeptides for separation of amino acid enantiomers by capillary
  electrophoresis, Anal. Chem., 70, (1998), 4967–4973.
- [3] D. Cascaval, C. Oniscu, A. I. Galaction, Selective separation of amino acids by reactive

extraction, Biochem. Eng., 7, (2001), 171-176.

- [4] T. Masawaki, M. Sasai, S. Tone, Optical resolution of an amino acid by an
  enantioselective ultrafiltration membrane, J. Chem. Eng. Jpn. 25 (1992) 33-39.
- [5] L. García-Carmona, M. Moreno-Guzmán, M. C. González, A. Escarpa, Class enzymebased motors for 'on the fly' enantiomer analysis of aminoacids. Biosens. Bioelectron., 96,
  (2017), 275-280.
- [6] S. Jansod, M. Cuartero, T. Cherubini, E. Bakker, Colorimetric readout for
  potentiometric sensors with closed bipolar electrodes. Anal Chem., 90, (2018), 6376-6379.

- [7] B. Roark, J. A. Tan, A. Ivanina, M. Chandler, J. Castaneda, H. S. Kim, S. JAwar, M.
- 293 Viard, S. Talic, K. L. Wustholz, Y. G. Yingling, M. Jones, K. A. Afonin, Fluorescence
- blinking as an output signal for biosensing, ACS Sens., 1, (2016), 1295-1300.
- [8] E. Villani, N. Shinda, S. Inagi, Electrogenerated chemiluminescence of luminol on
  wireless conducting polymer films, Electrochim. Acta, 389, (2021), 138718.
- [9] E. Villani, S. Inagi, Mapping the distribution of potential gradient in bipolar
  electrochemical systems through luminol electrochemiluminescence imaging, Anal. Chem.,
  93, (2021), 8152-8160.
- 300 [10] X. Ma, W. Gao, F. Du, F. Yuan, J. Yu, Y. Guan, N. Sojic, G. Xu, Rational design of
- electrochemiluminescent devices, Acc. Chem. Res., 54, (2021), 2936-2945.
- 302 [11] J. S. Stefano, F. Conzuelo, J. Masa, R. A. A. Munoz, W. Schuhmann, Coupling
  303 electrochemistry with a fluorescence reporting reaction enabled by bipolar
  304 electrochemistry, J. Electroanal. Chem., 872, (2020), 113921.
- [12] X. Wu, Y. Sun, T. He, Y. Zhang, G. J. Zhang, Q. Liu, S. Chen, Iron, Nitrogen-Doped
  Carbon Aerogels for Fluorescent and Electrochemical Dual-Mode Detection of Glucose,
  Langmuir, 37, (2021), 11309-11315.
- 308 [13] Y. Gu, Y. Hu, F. Zhang, L. Yi, D. Ren, Z. Ge, Electrochemiluminescence sensor based
  309 on cyclic peptides-recognition and Au nanoparticles assisted graphitic carbon nitride for
  310 glucose determination, Microchim. Acta, 188, (2021), 151.

- [14] L. Wang, X. H. Shi, Y. F. Zhang, A. A. Liu, S. L. Liu, Z. G. Wang, D. W. Pang,
  CdZnSeS quantum dots condensed with ordered mesoporous carbon for high-sensitive
  electrochemiluminescence detection of hydrogen peroxide in live cells, Electrochim. Acta,
  362, (2020), 137107.
- [15] W. Zhang, X. Sun, A. Zhou, M. Li, When Fluorescent Sensing Meets Electrochemical
  Amplifying: A Powerful Platform for Gene Detection with High Sensitivity and
  Specificity, Anal. Chem., 93, (2021), 7781-7786.
- 318 [16] Q. Zhang, Y. Tian, Z. Liang, Z. Wang, S. Xu, Q. Ma, DNA-Mediated Au-Au Dimer-
- 319 Based Surface Plasmon Coupling Electrochemiluminescence Sensor for BRCA1 Gene
- 320 Detection, Anal. Chem., 93, (2021), 3308-3314.
- [17] M. A. Schwarz, P. C. Hauser, Recent developments in detection methods for
  microfabricated analytical devices, Lab Chip, 1, (2001) 1-6.
- 323 [18] G. Salinas, A. L. Dauphin, C. Colin, E. Villani, S. Arbault, L. Bouffier, A. Kuhn,
- 324 Chemo- and magnetotaxis of self-propelled light-emitting chemo-electronic swimmers.
- 325 Angew. Chem. Int. Ed., 59, (2020), 7508-7513.
- 326 [19] G. Salinas, I. A. Pavel, N. Sojic, A. Kuhn, Electrochemistry-based light-emitting
- 327 mobile systems. ChemElectroChem, 7, (2020), 4853-4862.
- [20] R. Sharma, O. D. Velev, Remote steering of self-propelling microcircuits by
  modulated electric field, Adv. Funct. Mater., 25, (2015), 5512-5519.

- 330 [21] J. Roche, S. Carrara, J. Sanchez, J. Lannelongue, G. Loget, L. Bouffier, P. Fischer, A.
- Kuhn, Wireless powering of e-swimmers, Sci. Rep., 4, (2015) 6705.
- 332 [22] L. Bouffier, D. Zigah, N. Sojic, A. Kuhn, Bipolar (bio)electroanalysis, Annu. Rev.
- 333 Anal. Chem., 14, (2021), 65-86.
- [23] S. E. Fosdick, K. N. Knust, K. Scida, R. M. Crooks, Bipolar Electrochemistry, Angew.
  Chem. Int. Ed., 52, (2013), 10438-10456.
- 336 [24] L. Koefoed, S. U. Pedersen, K. Daasbjerg, Bipolar electrochemistry-A wireless
- approach for electrode reactions, Curr. Opin. Electrochem., 2, (2017), 13-17.
- [25] N. Karimian, P. Hashemi, A. Afkhami, H. Bagheri, The principles of bipolar
  electrochemistry and its electroanalysis applications, Curr. Opin. Electrochem., 17, (2019),
  30-37.
- [26] G. Salinas, S. Arnaboldi, L. Bouffier, A. Kuhn, Recent Advances in Bipolar
  Electrochemistry with Conducting Polymers, ChemElectroChem, 9, (2022), e202101234.
- [27] N. Shida, Y. Zhou, S. Inagi, Bipolar Electrochemistry: A Powerful Tool for
  Electrifying Functional Material Synthesis, Acc. Chem. Res., 52, (2019), 2598-2608
- 345[28] K. L. Rahn, R. K. Anand, Recent Advancements in Bipolar Electrochemical Methods
- of Analysis, Anal. Chem., 93, (2021), 103-123.
- [29] S. Arnaboldi, B. Gupta, T. Benincori, G. Bonetti, R. Cirilli, A. Kuhn, Absolute Chiral
  Recognition with Hybrid Wireless Electrochemical Actuators, Anal. Chem., 92, (2020),

```
349 10042-10047.
```

- [30] S. Arnaboldi, G. Salinas, G. Bonetti, R. Cirilli, T. Benincori, A. Kuhn, Bipolar
  Electrochemical Measurement of Enantiomeric Excess with Inherently Chiral Polymer
  Actuators, ACS Meas. Au, 1, (2021), 110-116.
- [31] G. Salinas, S. Arnaboldi, G. Bonetti, R. Cirilli, T. Benincori, A. Kuhn, Hybrid lightemitting devices for the straightforward readout of chiral information, Chirality, 33, (2021)
  875-882.
- [32] B. Gomez Ruiz, S. Roux, F. Courtois, C. Bonazzi, Spectrophotometric method for fast
  quantification of ascorbic acid and dehydroascorbic acid in simple matrix for kinetics
  measurements, Food Chem., 211 (2016), 583-589.
- [33] A. Domenech-Carbo, D. Dias, In situ electrochemical monitoring of ROS influence in
  the dynamics of ascorbic acid and polyphenolic compounds in apple fruits, Food Chem.,
  374, (2022), 131818
- [34] A. M. Pisoschi, A. Pop, A. I. Serban, C. Fafaneata, Electrochemical methods for
  ascorbic acid determination, Electrochim. Acta, 121, (2014), 443-460.
- [35] F. Sannicolò, S. Arnaboldi, T. Benincori, et al. Potential-driven chirality
  manifestations and impressive enantioselectivity by inherently chiral electroactive organic
  films. Angew. Chem. Int. Ed., 53, (2014), 2623-2627.
- 367 [36] X. Luo, L. Chen, J. Yang, S. Li, M. Li, Q. Mo, Y. Li, X. Li, Electrochemically
  368 simultaneous detection of ascorbic acid, sulfite and oxalic acid on Pt-Pd
  369 nanoparticles/chitosan/nitrogen doped graphene modified glassy carbon electrode: A
  370 method for drug quality control, Microchem. J., 169, (2021), 106623.

371	[37] F. Mazzara, B. Patella, G. Aiello, A. O'Riordan, C. Torino, A. Vilasi, R. Inguanta,
372	Electrochemical detection of uric acid and ascorbic acid using r-GO/NPs based sensors,
373	Electrochim. Acta, 388, (2021), 138652.

- 374 [38] Y. Hu, Y. He, Z. Peng, Y. Li, A ratiometric electrochemiluminescence sensing
  375 platform for robust ascorbic acid analysis based on a molecularly imprinted polymer
  376 modified bipolar electrode, Biosens. Bioelectron. 167, (2020), 112490
- [39] M. Bhaiyya, P. K. Pattnaik, S. Goel, Simultaneous detection of Vitamin B12 and
  Vitamin C from real samples using miniaturized laser-induced graphene based
  electrochemiluminescence device with closed bipolar electrode, Sens. Actuator A Phys.
  331, (2021), 112831.
- 381 [40] A. M. Pisoschi, A. Pop, G. P. Negulescu, A. Pisoschi, Determination of ascorbic acid
- content of some fruit juices and wine by voltammetry performed at Pt and carbon paste
- 383 electrodes, Molecules, 16, (2011), 1349-1365.