



HAL
open science

Lots of considerations when evaluating the FujiLAM assay – Author’s reply

Helena Huerga, Ankur Gupta-Wright, Winnie Muyindike, Catherine Hewison,
Martina Casenghi, Maryline Bonnet

► To cite this version:

Helena Huerga, Ankur Gupta-Wright, Winnie Muyindike, Catherine Hewison, Martina Casenghi, et al.. Lots of considerations when evaluating the FujiLAM assay – Author’s reply. *The Lancet global health*, 2023, 11 (4), pp.e502. 10.1016/S2214-109X(23)00115-8 . hal-04124063

HAL Id: hal-04124063

<https://hal.science/hal-04124063v1>

Submitted on 9 Jun 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L’archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d’enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial - NoDerivatives 4.0
International License

Lots of considerations when evaluating the FujiLAM assay

Author's reply

We thank Alberto Garcia-Basteiro and colleagues for their response to our research Article on the novel FujiLAM assay to detect tuberculosis in HIV-positive ambulatory patients in four African countries.¹

The evaluation of the inter-lot and intra-lot variability in diagnostic accuracy of an assay is usually not part of late-stage clinical studies but should be assessed during quality control by manufacturers. In this case, we were informed by FIND and the assay manufacturer about inter-lot variability of FujiLAM (Fujifilm, Tokyo, Japan) after our study was completed. Following discussions with our scientific advisory board, we added post-hoc analyses of diagnostic accuracy by test lot. Given the importance of the lot-to-lot variability, we presented pooled and by lot accuracy estimates as part of the main study results. A similar approach presenting pooled and by lot diagnostic accuracy estimates has been used by the FujiLAM study consortium.² Regarding the implications of these findings, as stated in the summary and discussion of our Article, this version of the FujiLAM assay cannot be recommended for clinical use until the problem of inter-lot variability is resolved.

We believe that sharing the results from a large and rigorous prospective clinical study is important to inform the scientific community, policy makers, end-users, and programme managers on the performance of an assay that has been considered a promising urine-based point-of-care test.³ We also hope that these data will be used by the manufacturer to inform root-cause analysis, identify solutions, and communicate these to key stakeholders. We think that providing

data in a transparent manner will not damage confidence in the currently recommended Alere Determine TB-LAM Ag test (Abbott, Waltham, MA, USA). Notable reasons for low uptake of this test are related to challenges with access, budget limitations, and difficulties in obtaining regulatory agency approvals.⁴ We believe that our conclusion is clear enough to prevent any false hope. We also think that we have an ethical obligation towards the study participants to publish these data.

Notably, the results on inter-lot variability reported by our group and others raise an important question on whether inter-lot and intra-lot variability should be systematically assessed as part of clinical diagnostic studies, and at which stage it is most appropriate to evaluate this. These considerations should be discussed when providing updated guidance for the evaluation of new diagnostics.⁵

Sputum-free point-of-care diagnostic tests for tuberculosis remain urgently needed. Despite these disappointing results for this version of the FujiLAM assay, we hope that research to develop more sensitive LAM-based point-of-care assays will continue.⁶

We declare no competing interests.

Copyright © 2023 The Author(s). Published by Elsevier Ltd. This is an Open Access article under the CC BY-NC-ND 4.0 license.

**Helena Huerga, Ankur Gupta-Wright, Winnie Muyindike, Catherine Hewison, Martina Casenghi, Maryline Bonnet*
helena.huerga@epicentre.msf.org

Department of Field Epidemiology, Epicentre, Paris 75019, France (HH); Institute for Global Health, University College London, London, UK (AG-W); Department of Medicine, Mbarara Regional Referral Hospital, Mbarara, Uganda (WM); Department of Medicine, Médecins Sans Frontières, Paris, France (CH); Department of Innovation and New Technology, Elizabeth Glaser Pediatric AIDS Foundation, Geneva, Switzerland (MC); Université de Montpellier, TransVIHMI, INSERM, IRD, Montpellier, France (MB)

1 Huerga H, Bastard M, Lubega AV, et al. Novel FujiLAM assay to detect tuberculosis in HIV-positive ambulatory patients in four African countries: a diagnostic accuracy study. *Lancet Glob Health* 2023; **11**: e126–35.

- 2 Székely R, Sossen B, Mukoka M, et al. Multicentre accuracy trial of FUJIFILM SILVAMP TB LAM test in people with HIV reveals lot variability. *medRxiv* 2022; published online Sept 9. <https://doi.org/10.1101/2022.09.07.22278961> (preprint).
- 3 Broger T, Nicol MP, Székely R, et al. Diagnostic accuracy of a novel tuberculosis point-of-care urine lipoarabinomannan assay for people living with HIV: a meta-analysis of individual in- and outpatient data. *PLoS Med* 2020; **17**: e1003113.
- 4 Singhroy DN, MacLean E, Kohli M, et al. Adoption and uptake of the lateral flow urine LAM test in countries with high tuberculosis and HIV/AIDS burden: current landscape and barriers. *Gates Open Res* 2020; **4**: 24.
- 5 Denking CM, Schumacher SG, Gilpin C, et al. Guidance for the evaluation of tuberculosis diagnostics that meet the World Health Organization (WHO) target product profiles: an introduction to WHO process and study design principles. *J Infect Dis* 2019; **220** (suppl 3): S91–98.
- 6 Sossen B, Meintjes G. Development of accurate non-sputum-based diagnostic tests for tuberculosis: an ongoing challenge. *Lancet Glob Health* 2023; **11**: e16–17.

