



**HAL**  
open science

## Improving dental epithelial junction on dental implants with bioengineered peptides

Ivan Panayotov, Attila Gergely Vegh, Marta Martin Fernandez, Boyan Vladimirov, Christian Larroque, Csilla Gergely, Frédéric Jg Cuisinier, Elias Estephan

### ► To cite this version:

Ivan Panayotov, Attila Gergely Vegh, Marta Martin Fernandez, Boyan Vladimirov, Christian Larroque, et al.. Improving dental epithelial junction on dental implants with bioengineered peptides. *Frontiers in Bioengineering and Biotechnology*, 2023, 11, pp.1165853. 10.3389/fbioe.2023.1165853 . hal-04139349

**HAL Id: hal-04139349**

**<https://hal.science/hal-04139349v1>**

Submitted on 12 Jul 2024

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License



## OPEN ACCESS

## EDITED BY

Changjiang Pan,  
Huaiyin Institute of Technology, China

## REVIEWED BY

Sara Ferraris,  
Polytechnic University of Turin, Italy  
Annabel Braem,  
KU Leuven, Belgium

## \*CORRESPONDENCE

Elias Estephan,  
✉ elias.estephan@umontpellier.fr

RECEIVED 14 February 2023

ACCEPTED 12 June 2023

PUBLISHED 20 June 2023

## CITATION

Panayotov IV, Végh AG, Martin M,  
Vladimirov B, Larroque C, Gergely C,  
Cuisinier FJG and Estephan E (2023),  
Improving dental epithelial junction on  
dental implants with  
bioengineered peptides.  
*Front. Bioeng. Biotechnol.* 11:1165853.  
doi: 10.3389/fbioe.2023.1165853

## COPYRIGHT

© 2023 Panayotov, Végh, Martin,  
Vladimirov, Larroque, Gergely, Cuisinier  
and Estephan. This is an open-access  
article distributed under the terms of the  
[Creative Commons Attribution License  
\(CC BY\)](https://creativecommons.org/licenses/by/4.0/). The use, distribution or  
reproduction in other forums is  
permitted, provided the original author(s)  
and the copyright owner(s) are credited  
and that the original publication in this  
journal is cited, in accordance with  
accepted academic practice. No use,  
distribution or reproduction is permitted  
which does not comply with these terms.

# Improving dental epithelial junction on dental implants with bioengineered peptides

Ivan V. Panayotov<sup>1,2</sup>, Attila G. Végh<sup>3</sup>, Marta Martin<sup>4</sup>,  
Boyan Vladimirov<sup>5</sup>, Christian Larroque<sup>6</sup>, Csilla Gergely<sup>4</sup>,  
Frédéric J. G. Cuisinier<sup>1,2</sup> and Elias Estephan<sup>1,7\*</sup>

<sup>1</sup>LBN, University Montpellier, Montpellier, France, <sup>2</sup>CSERD, CHU Montpellier, Montpellier, France, <sup>3</sup>Biological Research Centre, Institute of Biophysics, Eötvös Lóránd Research Network (ELKH), Szeged, Hungary, <sup>4</sup>L2C, University Montpellier, CNRS, Montpellier, France, <sup>5</sup>Department of Maxillofacial Surgery, Medical University of Plovdiv, Plovdiv, Bulgaria, <sup>6</sup>Department of Nephrology, CHU Montpellier, Hôpital Lapeyronie, IRMB, University of Montpellier, INSERM U1183, Montpellier, France, <sup>7</sup>Neuroscience Research Center, Faculty of Medical Sciences, Lebanese University, Beirut, Lebanon

**Introduction:** The functionalization of titanium (Ti) and titanium alloys (Ti6Al4V) implant surfaces via material-specific peptides influence host/biomaterial interaction. The impact of using peptides as molecular linkers between cells and implant material to improve keratinocyte adhesion is reported.

**Results:** The metal binding peptides (MBP-1, MBP-2) SVSVGKMKPSRP and WDPPTLKRVPSP were selected via phage display and combined with laminin-5 or E-cadherin epithelial cell specific peptides (CSP-1, CSP-2) to engineer four metal-cell specific peptides (MCSPs). Single-cell force spectroscopy and cell adhesion experiments were performed to select the most promising candidate. *In vivo* tests using the dental implant for rats showed that the selected bi functional peptide not only enabled stable cell adhesion on the trans-gingival part of the dental implant but also arrested the unwanted apical migration of epithelial cells.

**Conclusion:** The results demonstrated the outstanding performance of the bioengineered peptide in improving epithelial adhesion to Ti based implants and pointed towards promising new opportunities for applications in clinical practice.

## KEYWORDS

bioengineered peptide, phage display, implants, titanium surface functionalization, epithelial adhesion

## 1 Introduction

Pure titanium (Ti) (Brannemark et al, 1977) and titanium alloys (Ti6Al4V) (Katsikeris et al, 1987) have been the most successful and widespread metals used for dental and maxillofacial implants. Inflammation and destruction of soft and hard tissues surrounding dental implants are termed mucositis and peri-implantitis (Roos-Jansaker et al, 2003). Mucositis points out a bacteria-induced, reversible inflammatory process of the peri-implant soft tissue with reddening, swelling and bleeding on periodontal probing. Peri-implant mucositis can lead to peri-implantitis involving loss of marginal alveolar bone around a functioning oral implant (Zitzmann and Berglundh, 2008). In contrast to mucositis, peri-implantitis is a progressive and irreversible disease of implant-surrounding hard and soft tissues showing bone resorption, decreased osseointegration, increased pocket formation and abscesses (Smeets et al, 2014).

Peri-implantitis has a prevalence on the order of 10% of implants and 20% of patients 5–10 years after implant placement (Mombelli et al, 2012; Heitz-Mayfield and Mombelli, 2014). Peri-implant mucositis occurs in about 80% of patients and in about 50% of implant sites (Zitzmann and Berglundh, 2008). Several epidemiological and clinical (Heitz-Mayfield and Mombelli, 2014) studies have pointed out that untreated peri-implantitis may threaten general health by increasing the risk of cardiovascular diseases, preterm labor and pulmonary diseases (Pussinen et al, 2007). Cardiovascular diseases that occur in untreated periodontitis or peri-implantitis as a risk factor lead to mortality and disability (Mattila et al, 2005).

As oral mucosa is transfixed by the implant, the epithelial sealing has been identified as the critical factor to prevent peri-implant inflammation (Rompen et al, 2006). The adhesion of oral keratinocytes must be stable and resistant to external aggressors, including mechanical constraints and bacterial pathogens and toxins. Up to now, “platform switching” is the only clinically proven strategy for the prevention of peri-mucositis and peri-implantitis (Atieh et al, 2010; Baffone et al, 2012). This concept is based on the use of the transgingival part of the dental implant (implant abutment) with a smaller diameter than the intraosseous part of the implant which assures a sufficient dimension of peri-implant mucosa to control epithelial-junctional attachment (Atsuta et al, 2012; Cumbo et al, 2013).

Biomaterials surface modifications are usually performed to increase their biocompatibility and bioactivity. These modifications are habitually divided into two types: additive and subtractive method (Kazmierczak and Przekora, 2020). Additive surface modifications are represented by inorganics (Fu et al, 2020; Mumith et al, 2020; de Oliveira et al, 2021) and organics surface coatings which incorporates surface functionalization with different inorganic and organic molecules, proteins and peptides (Werner et al, 2009; Souza et al, 2019; Lallukka et al, 2022).

Titanium (Ti) and titanium alloy (Ti6Al4V) modification types (mechanical surface modifications, oxidative processes, sol-gel derived titania (TiO<sub>2</sub>) coatings and biofunctionalized surfaces) on nonkeratinized soft tissues were reviewed. It was shown that Ti implants with a roughness between 0.5 and 1.0 μm induce soft-tissue adhesion and that a fibroblast growth factor 2 apatite composite coating promoted soft-tissue attachment via Sharpey-like fibers (Zigterman et al, 2019). According to Panayotov et al, (2015), three main surface modification approaches have been used in order to improve the tissue-implant interface of Ti and Ti6Al4V: layer-by-layer deposition of polyelectrolyte films, phage display-selected surface binding peptides and self-assembled DNA monolayer systems. Van den Borre concludes that porous titanium coatings having large pores (>700 μm) support cell attachment and that nanostructured ceramic coatings are found to reduce the inflammatory response. In this latter review, a particular interest was conferred for biomolecule coatings so that a durable fixation of the implant can be ensured (Van den Borre et al, 2022). Anodization has gained special interest in surface modification, it was stated that the anodized titanium surface increases blood clot retention and nano-roughness, and aids osseointegration (Traini et al, 2018). Some reports have been published on combining specific peptides that can lead to osteointegration and bioactivity (Yazici et al, 2013; Rodriguez et al, 2017). Bifunctional chimeric peptides were also designed in order to prevent bacterial biofilm formation. The antibacterial potential of

these novel peptides was studied against a set of different bacterial strains (Yucesoy et al, 2015; Liu et al, 2016; Yazici et al, 2016; Geng et al, 2018; Zhang et al, 2018; Wisdom et al, 2019; Wisdom et al, 2020; Drexelius et al, 2023). In these later studies, bi-peptides were synthesized with two domains, the first one was for Titanium binding with a robust solid-surface coating and the second with antimicrobial properties. In order to promote bone regeneration, a bioinspired chimeric peptide was also designed to activate the canonical Wnt/β-catenin signaling pathway of stem cells (Zhou et al, 2015). This later peptide can enhance mineral deposition and osteogenesis. Zhao et al, (2021) used 3D printing technology and chimeric peptides in order to improve osseointegration on the implant–bone surface. According to Kumar Boda and Aparicio, (2022), a dual keratinocyte-attachment and an anti-inflammatory coating can help in reducing inflammation and promote permucosal/peri-implant soft tissue sealing. In this report, pristine and oxygen plasma pre-treated polished titanium was coated with conjugated linoleic acid and cationic cell adhesive peptides.

In the present investigation, we introduced metal-cell specific bifunctional peptides (MCSPs) designed to increase the gingival adhesion on the dental implant surface and inhibit the epithelial cell migration toward the apical part of the implant. This bifunctional peptide is a combined peptide composed of two parts (Brannemark et al, 1977): a specific peptide for the implant surface and (Katsikeris et al, 1987) a peptide with high affinity to endothelial cells. A set of physical and biological analyses was used to select the best candidates which were tested *in-vitro* and *in vivo* in a rat model.

## 2 Materials and methods

### 2.1 Substrate coating

#### 2.1.1 Substrate preparation

Ti and Ti6Al4V discs (d = 15 mm and thickness 1 mm) were polished on Silicon Carbide disks and on soft disks using diamond pastes (6 μm, 1 μm, and 0.25 μm) on a polishing machine (Escil, Lyon, France). Specimens were thoroughly cleaned in sodium dodecyl sulfate 0.1 M (Sigma Aldrich, St. Louis, United States), in hydrochloric acid 0.1% (Sigma Aldrich, St. Louis, United States) and finally in ultra-clean water (Milli-Q; Merck Millipore, Darmstadt, Germany) at 22°C in an ultrasound bath for 5 min. For cell incubation, all surfaces were cleaned in a 70% alcohol bath for 10 min. Finally, the samples were washed three times and stored in PBS (Gibco®, Invitrogen, Carlsbad, CA, United States) under sterile conditions. Polishing and cleanup with the same protocol procedures were repeated before phage display cycles and before each manipulation. After polishing and cleaning the substrate roughness were verified using AFM. A roughness of about 10 nm was obtained for the two type of substrates.

#### 2.1.2 Phage display selection of metal binding peptides

An M13 bacteriophage library (PhD-12 PD Peptide Library Kit™) supplied by New England Biolabs (Beverly, MA, United States) in phosphate-buffered saline solution containing 0.1% TWEEN-20 (PBST) was exposed to Ti and Ti6Al4V

samples. After rocking for 1 h at room temperature, the Ti and Ti-alloy surfaces were thoroughly washed with PBST to rinse off unbound phages. Bound phages were then eluted from the surface under acidic conditions (glycine-HCl pH 2.2, 10 min), which disrupt the interaction between the displayed peptide and the target. Before elution, target wells were changed to prevent the elution of phages bound to the plastic walls. After neutralization with Tris-HCl (pH 9.1), the eluted phages were infected into the bacterial host strain *Escherichia coli* ER2738 and thereby amplified. After three to six rounds of biopanning, monoclonal phage populations were selected and analyzed individually. Finally, ten phages were selected and amplified from each sample, followed by the extraction of their DNA to determine the genetic code of the expressed peptide.

## 2.2 Physicochemical characterization of the metal binding peptides

### 2.2.1 MALDI-TOF/TOF analysis and substrate coating with MBP

Ti and Ti6Al4V surfaces were incubated in a 100- $\mu$ M solution of SVSVGMPKPSRP peptide (MBP-1) or WDPPTLKRVPSP peptide (MBP-2) for 2 h. The samples were rinsed thoroughly, either with a hydrophilic solution: ultra clean water (Milli-Q; Merck Millipore, Darmstadt, Germany), a hydrophobic solution: acetonitrile 100% (Sigma—Aldrich St. Louis, MO, United States) or an ionic solution (NaCl, 1 M). The presence of the peptides on the dry surfaces was identified with MALDI TOF/TOF spectrometry. Samples were analyzed using a 4,800 Plus MALDI-tandem time-of-flight system (MALDI-TOF/TOF) Proteomics Analyzer (Applied Bio systems, Foster City, CA, United States) in positive reflector ion mode using a 20-kV acceleration voltage. The YAG laser was operated at a 200-Hz firing rate with a wavelength of 355 nm. Mass spectrometry spectra were acquired for each sample using 1,500 laser shots. All acquired spectra of the samples were processed using the 4,000 Series Explorer TM software (Applied Bio systems, Foster City, CA, United States) in default mode.

The peptide was identified by searching in the Swiss-Prot database using Protein Pilot TM 2.0 software (Applied Bio systems, Foster City, CA, United States) or Protein Prospector (<http://prospector.ucsf.edu/>). The ExPASy database ([www.expasy.org/tools/pi\\_tool.html](http://www.expasy.org/tools/pi_tool.html)) was used to calculate the mono isotopic theoretical mass of the peptide.

### 2.2.2 Atomic force microscopy

AFM measurements were performed using an Asylum MFP-3D head and controller (Asylum Research, Santa Barbara, CA, United States), mounted on an Olympus inverted microscope. Height images were recorded in tapping mode and in liquid at room temperature. Typically, 512  $\times$  512 points scans were taken at a scan rate of 1 Hz per line. Both trace and retrace images were recorded and compared.

#### 2.2.2.1 Force measurements by atomic force microscopy

Relative binding strengths of peptides onto Ti and Ti6Al4V surfaces were measured in contact mode and in a liquid medium (PBS pH 7.4) with a functionalized tip. Force measurements were

taken at constant loading rates (vertical piezo-velocity of 1  $\mu$ m/s). The spring constant of the tip was calibrated in the presence of PBS solution using the thermal fluctuation method and found to be approximately 18 pN/nm. For tip functionalization, the ultrasoft AFM cantilever tips (Bio lever-Olympus) were rinsed with copious amounts of Milli-Q water and then dried. In the next step, tip functionalization was performed. The AFM cantilever tips were incubated in 1  $\mu$ g/mL biotinylated bovine serum albumin (BSA; Sigma Aldrich, St. Louis, MO, United States) solution in PBST, pH 7.0, at room temperature overnight, and the tip was then incubated for 30 min in 100  $\mu$ g mL<sup>-1</sup> streptavidin in PBST and finally in BSA (1%) for 1 h to block the nonspecific binding sites. Thorough rinsing was performed between all steps. Biotinylated peptides were fixed on the tip prior to each measurement.

## 2.3 *In vitro* testing

### 2.3.1 Bi-functional peptides

Four bi-functional metal binding cell specific peptides (MCSPs) were synthesized (MilleGen, Toulouse, France) with a purity higher than 80%. Titanium and Ti6Al4V surfaces were incubated in 100  $\mu$ M PBS solutions of MCSPs for 2 h and were then washed three times with PBS (Invitrogen, Carlsbad, CA, United States) before cell incubation.

### 2.3.2 Oral keratinocyte cells

Cells from a non-tumoural, immortalized oral keratinocyte cell line, TERT-2 OKF-6 (BWH Cell Culture and Microscopy Core, United States) were cultivated in defined keratinocyte serum-free medium (KSF; Gibco<sup>®</sup>, Invitrogen, Carlsbad, CA, United States) supplemented with: CaCl<sub>2</sub> (Sigma Aldrich, St. Louis, MO, United States), 0.25  $\mu$ g Bovine Pituitary Extract (BPA), 0.2 ng/mL epithelial growth factor (EGF), and 0.3 mM, 10% Pen Strep X 100 (Penicillin-10,000 Unit/mL, Streptomycin-10,000 mg/mL) (Gibco<sup>®</sup>, Invitrogen, Carlsbad, CA, United States). The medium was changed every 2 days until cells were used. For all experiments, cells from passage number 7 to 9 were used. After reaching 90% confluence, the cells were detached with 0.05% Trypsin-EDTA (Gibco<sup>®</sup>, Invitrogen, Carlsbad, CA, United States) for 5 min.

### 2.3.3 Atomic force microscopy for *in vitro* studies

Tip-less cantilevers (MikroMasch, Tallinn, Estonia) were used for single cell-implant surface adhesion measurements. Determination of the spring constant for each cantilever was performed by the thermal calibration method implemented in the driving software (Hutter et al, 1993; Sader et al, 1999), resulting in a value of 0.03 N/m. Individual oral keratinocyte cells were attached to surface activated tip-less cantilevers using the protocol of Zhang et al, (2006) (41) to obtain concanavalin-A (Con A) mediated linkage. All measurements were conducted with the cantilever and the attached cell immersed in complemented KSF medium (Gibco<sup>®</sup>, Invitrogen, Carlsbad, CA, United States) at 30°C and within 3 h, allowing comparison of the adhesion forces for types of surfaces with the same cell. The measurements were taken at five different points on each surface. Force measurements were performed with a loading rate of 2  $\mu$ m/s and a load of 2 nN (see [Supplementary Information](#)).

### 2.3.4 Para-nitrophenyl-phosphate (pNPP) cell viability test

In order to analyze the protein phosphatase activity of the cells, the pNPP assay was performed for each of the studied Ti and Ti6Al4V surfaces with and without peptide functionalization at 4 h after cell incubation. Confluent cells were washed with PBS and detached with Trypsin-EDTA 0.5% (Gibco®, Invitrogen, Carlsbad, CA, United States) for 5 min at 37°C. Cells were then re-suspended in supplemented KSM, and  $5 \times 10^5$  cells per well were incubated with the bare and functionalized Ti and Ti6Al4V surfaces. The samples were incubated at 37°C under a 5% CO<sub>2</sub> humidified atmosphere for 4 hours. At the end of the incubation, the cells were washed three times with PBS and lysed with 500 µL of the acid phosphatase lysing buffer (0.1 M sodium acetate, 0.1% Triton X-100, pH 5.5), supplemented with 1 mg/mL of pNPP (para-Nitrophenyl Phosphate, Sigma Aldrich, St. Louis, MO, United States). After 1 h incubation at 37°C, the reaction was stopped by the addition of 50 µL of 1 N NaOH for 30 min at room temperature. The yellow colorimetric reaction was measured with a micro titer plate reader (EL-800 Universal Micro plate Reader, Bio Tec Instruments INC., VT, United States) at 405 nm. A linear relationship between the percentage of adhering cells and the light absorption due to para-nitrophenyl phosphate coloration was used to determine the concentration of adherent cells.

## 2.4 In vivo testing

The study was approved by the committee for animal welfare of Montpellier University with referral number 1143 15/03/2015.

### 2.4.1 Oral implantation

The oral implantation procedure was completed according to the immediate-implantation protocol, as described by (Ikeda et al 2000) (42). Eighteen 12-week-old Wistar rats (male, 300–320 g) were anesthetized with intra peritoneal pentobarbital sodium (50 mg/kg; ref), and the first right maxillary molar was extracted. Ti6Al4V transgingival implants were immediately implanted after extraction for 4 weeks period. Bare implant surface, MCSP-2 functionalized surface and MBP-1 functionalized surface (6 implants/group) were compared. (see [Supplementary Information](#)).

### 2.4.2 Tissue preparation

After 4 weeks, the rats from three groups were sacrificed and the samples were withdrawn, followed by 4% paraformaldehyde (Ph 7.4 in PBS) incubation, at 4°C for 4 h. Samples were demineralized in 5% EDTA, 4% sucrose in 0.01 m PB, Ph 7.4, for 4 days at 4°C. The oral mucosa surrounding the implant was then carefully removed from the implant. The gingiva around the left first molar was also removed from the tooth. All specimens were immersed in 20% sucrose in 0.1 m PBS at 4°C overnight for cryoprotection, and then embedded in O.C.T. compound. They were then quickly frozen in dry ice/isopentane, before cutting 10-mm bucco-palatal sections with a cryostat at 20°C. The cryo-sections were mounted on gelatin-coated glass slides.

### 2.4.3 Immunohistochemistry

Laminin-332  $\gamma$ 2 IHC was using the avidin-biotinylated peroxidase complex (ABC) kit (Vector Laboratories, Burlingame, CA, United States) was performed. After sectioning the samples

were washed in 0.01 M phosphate-buffered saline (PBS), pH 7.2, treated with 0.3% H<sub>2</sub>O<sub>2</sub> for 30 min to inhibit endogenous peroxidase activity, and blocked for 30 min with 10% normal goat serum in PBS. The tissue sections were then incubated overnight at 4°C with affinity-purified polyclonal rabbit IgG antibody to rat  $\gamma$ 2 chain of laminin-332 in PBS (1:100 or 1:50) for 48 h. The samples were then rinsed and incubated for 45 min with biotinylated goat anti-rabbit IgG in PBS (1:200), followed by a 60-min incubation with ABC dissolved in PBS (1:100). Immuno-positive staining was visualized by incubating the samples for 5 min in 0.02% diaminobenzidine-tetrahydrochloride (ABC kit) and then counterstaining them lightly with hematoxylin. The length of immunohistochemical coloration of laminin 332 indicate the epithelial adhesion on the implant surface was measured. This was observed like a dark-brown line at the implant-epithelium interface. All samples were observed with a light microscope Zeiss Axiolab comportsing Zeiss F40/0.65 dry objectives (Carl Zeiss Industrielle Messtechnik GmbH, Oberkochen, Germany) and photographed with a Sony  $\alpha$ 5100 camera (Sony corporation, 1-7-1 Konan Minato-ku, Tokyo, 108-0075 Japan). All measurements were made using ImageJ software (NIH, Bethesda, MD, United States, <http://rsb.info.nih.gov/ij/>), after image calibration.

## 2.5 Statistical analysis

Data analysis was performed using Stata software v14.2 (StataCorp LP 4905 Lake way Drive College Station, Texas 77845-4512. United States). Details on the statistical analysis are provided below separately for the different studies performed.

### 2.5.1 In vitro studies

#### 2.5.1.1 Single-cell force spectroscopy (SCFS) assay

Sample size calculation: SCFS measurement was performed at five different points of each of the studied surfaces. A minimum of 30 measurements were required to underline a minimal difference of  $1 \times 10^{-9}$ N between groups, with a standard deviation of  $1.5 \times 10^{-9}$ N and a power of 0.9.

The normality of each group of measurements was checked for the five different points by the Shapiro-Wilks test. For each of the five groups, the Gaussian distribution could not be rejected:

For titanium (Ti) surfaces:

Ti ( $p = 0.10$ ),

MCSP-1 ( $p = 0.07$ ),

MCSP-2 ( $p = 0.18$ ), MCSP-3 ( $p = 0.06$ ), MCSP-4 ( $p = 0.47$ ).

For titanium alloy (Ti6Al4V) surfaces: Ti6Al4V ( $p = 0.46$ ), MCSP-1 ( $p = 0.17$ ), MCSP-2 ( $p = 0.06$ ), MCSP-3 ( $p = 0.11$ ), MCSP-4 ( $p = 0.83$ ).


During the BSA inhibition essay: Ti ( $p = 0.06$ ), Ti6Al4V ( $p = 0.43$ ), Ti/MCSP-2 ( $p = 0.49$ ), Ti6Al4V/MCSP-2 ( $p = 0.09$ ), Ti/MCSP-2+P2+BSA ( $p = 0.58$ ), Ti6Al4V/MCSP-2+BSA ( $p = 0.99$ ).

Statistical tests: ANOVA with pairwise comparisons between groups was performed, considering the Bonferroni correction. A  $p$ -value of  $< 0.05$  was considered as significant.

#### 2.5.1.2 Para-nitrophenyl-phosphate (pNPP) assay

The sample size was computed for testing the row effect for a 5%-level test with 90% power. With an effect variance of 0.2, the number of





**FIGURE 1**  
Maldi-TOF/TOF spectra of MBP-1 (A,B) and MBP-2 (C,D) on Titanium and Ti-alloy (Ti6Al4V) surfaces, respectively, after acetonitrile rinsing.

measures per group was at minimum 9. Since the Shapiro-Wilks normality test showed that the data were not normally distributed, a root square transformation was applied to the data to obtain a Gaussian distribution Ti ( $p = 0.15$ ), Ti6Al4V ( $p = 0.35$ ). A two-way ANOVA was used to test the two main effects: material and peptide on the adhesion forces. A  $p$ -value of  $< 0.05$  was considered as significant.

### 2.5.2 *In vivo* study

Sample size calculation: since repeated measures were performed within the same animal, a cluster effect was considered. With a difference in epithelial height of 120  $\mu\text{m}$ , a standard deviation of 80  $\mu\text{m}$ , an intra class correlation coefficient of 0.03 and a power of 80%, the number of rats required was calculated to be equal at least to 6 per group, with 4 measurements per rat and a one-sided test ( $\alpha$  risk of 5%).

Normality was checked for the three groups by the Shapiro-Wilks test: bare Ti ( $p = 0.14$ ), SVSVGMPKSPRP (MBP-1) coated Ti ( $p = 0.10$ ) and WDPPTLKRVPSP (MBP-2) coated Ti ( $p = 0.76$ ).

Statistical tests: One-way ANOVA with pairwise comparisons between groups was performed, with a Bonferroni correction.


## 3 Results

### 3.1 Selection and affinity of metal binding peptides (MBP)

The metal binding peptides were selected via phage display using the M13 bacteriophage library and by performing four biopanning rounds against Ti6Al4V. SVSVGMPKSPRP (MBP-1) and

WDPPTLKRVPSP (MBP-2) were selected as two specific peptides for our materials. MBP-1 was expressed by 37.5% of the phages after the third round of bio panning while MBP-2 was manifested at a rate of 40% after the fourth round (see [Supplementary Information](#)). The affinity of both MBPs to Ti and Ti6Al4V was tested by mass spectrometry and force spectroscopy. Concerning the mass spectrometry after adsorption of MBP-1 and MBP-2 on the Ti and Ti- alloys substrate, the samples were rinsed with acetonitrile. [Figure 1](#) shows that the MBPs remain at the surfaces. Thus, these selected peptides present a high affinity to our substrates. The obtained mass-to-charge ratio ( $m/z$ ) for MBP-1 was 1,239.6, in agreement with the theoretical mass of the peptide, and on both metals, we identified a small peak at an  $m/z$  of 1,255.6, corresponding to the oxidized form of the peptide ([Figure 1A](#) for MBP-1-Ti and [Figure 1B](#) for MBP-1-Ti- alloy). The obtained mass-to-charge ratio ( $m/z$ ) for MBP-2 was 1,391.6, which corresponds to the theoretical mass of the protonated molecule  $[\text{MH}]^+$ . Mass spectroscopy also reveals two other peaks at  $m/z$  1,406.6 and  $m/z$  1,423.6 suggesting the presence of oxygen ions in the peptide structure ([Figure 1C](#) for MBP-2-Ti and [Figure 1D](#) for MBP-2-Ti-alloy). The slight differences in the obtained peptide masses on the different materials were due to variations in the thickness of the samples, which induced small differences in the time of flight (see [Supplementary Information](#)).

To have a clearer view of the affinity of MBP-1 and MBP-2 to the metals, we conducted force spectroscopy using AFM. Monitoring the unbinding processes of adsorbed molecules under external stress allows the quantification of adhesion forces ([Gergely et al, 2000](#); [Gergely et al, 2002](#)). The adhesion forces measured between MBPs and Ti and Ti6Al4V surfaces are presented in [Figure 2](#). The MBP-2



peptide exhibited statistically significantly higher adhesion forces to both metals. The adhesion of MBP-2 on Ti6Al4V was almost two-fold stronger than the adhesion force of MBP-1. For comparison, the strength of the osteopontin and  $\alpha\beta 3$  integrin bond is  $50 \pm 2$  pN (Litvinov et al, 2003), that of the ICAM-1 and Anti-ICAM-1 antibody bond is  $100 \pm 50$  pN (Willemsen et al, 1998), and cadherin-mediated cell-cell interactions have a minimal binding force in the range of 50 pN (Panorchan et al, 2006; Sivasankar et al, 2009). The highly specific biotin-streptavidin interaction is the strongest non-covalent bond, ranging from 250 to 320 pN, according to the experimental conditions (Wong et al, 1999). The interactions of MBP-1 and MBP-2 with the two metal surfaces are in the range of antigen-antibody forces.

### 3.2 Design of bi-functional peptides for epithelial cell attachment to titanium surfaces


Four bifunctional, metal-cell specific peptides (MCSPs) were engineered by combining a 12-mer metal binding peptide (MBP-1, MBP-2) to a cell specific sequence (CSP-1, CSP-2) (Figure 3). Both

MBPs were selected by phage display on titanium alloy surfaces (see Supplementary Information). The CSP-1 was selected because of its affinity for the laminin-LG3 globular domain (14 residues) (Kim et al, 2005; Werner et al, 2009), whereas the CSP-2 for its affinity to N- and E-cadherin ectodomains (12 residues) (Devemy and Blaschuk, 2009). These two types of cell receptors are found in the major classes of epithelial extracellular cell adhesion proteins (Smeets et al, 2014). The four MCSPs were composed of 29 or 27 residues, depending on the length of the CSP, including a spacer of three glycine amino acids between the metal binding peptide and the cell-specific peptide (Figure 3).


### 3.3 *In vitro* cell adhesion studies on Ti and Ti6Al4V alloy modified with engineered metal-cell specific peptides

Single-cell force spectroscopy was applied to evaluate the ability of the M-CSPs to increase oral keratinocyte adhesion to Ti and Ti6Al4V. Oral keratinocyte cells were attached to tip-less cantilevers by means of a concanavalin-A mediated linkage (Zhang et al, 2006; Vegh et al, 2012). During the recorded force curves a dwell period of 5 s was introduced, when the cell and the substrate were in contact. This short period implies that no information was recorded about how the cell responds to the surface on a longer time scale, thus excluding the effects of, e.g., changes in protein expression (Weder et al, 2009) and limiting our study to the study of mechanical properties of the cell recording the effect of the proteins and protein complexes that were already present at the cell membrane. To address the efficacy of the surface modification cell adhesion forces were measured against the bare Ti and Ti6Al4V surfaces, and after their functionalization with the MCSPs (Figure 4). Surfaces coated with MCSP-2 and MCSP-4 exhibited higher cell adhesion forces than surfaces functionalized with MCSP-1 or MCSP-3. A significant difference in cell adhesions to MCSP-2- and MCSP-4-functionalized Ti (of about 2 nN) compared to the non-functionalized Ti surface (0.8 nN) has been found. The MCSP-1 and MCSP-3 peptides incorporate a common cell-specific peptide motif (CSP-2) with different metal binding part (MBP) (Figure 3), but they did show lower affinity in terms of adhesion force compared to MCSP-2 and MCSP-4. Combining MBP-1 and CPS-2 (cadherin binding peptide) in MCSP-2 appears to be the most successful peptide configuration both for Ti and Ti6Al4V functionalization to promote oral epithelial cell adhesion. The obtained keratinocyte cell adhesion forces are in the range of typical cell-surface interactions reported in the literature (Puech et al, 2005; Taubenberger et al, 2007; Taubenberger et al, 2010)–(Puech et al, 2005; Taubenberger et al, 2007; Taubenberger et al, 2010).

The surface roughness is a crucial factor that influences cell attachment and proliferation on implants (Baharloo et al, 2005). Therefore, atomic force microscopy (AFM) was used to investigate the topography of the Ti and Ti alloy modified with the engineered metal-cell specific peptides. The surface roughness was characterized by calculating the root-mean-square (RMS) of each recorded topography. Figure 5 shows the obtained topographies for bare Ti (Figure 5A) and Ti6Al4V (Figure 5C) surfaces, and surfaces coated with MCSP2, where the highest RMS modifications were observed



**FIGURE 3** Combinational design of metal-cell specific peptides (MCSP). (A) Metal-binding peptides (MBPs) with high affinity to Ti and Ti6Al4V. (B) Cell specific peptides (CSPs) with affinity to laminin-332 (CSP-1) and to E-cadherin's ectodomains (CSP-2). (C) The engineered metal-cell specific peptides (MCSPs). Three glycines (G) residues are used as a spacer between the two peptide sequences.




**FIGURE 4** Strength of keratinocyte cell binding to bare and modified Ti and Ti6Al4V surfaces. Adhesion forces between living keratinocytes and metal surfaces were evaluated using AFM in force mode. (A) Akeratinocyte-decorated cantilever was used to measure a set of forces on bare and functionalized Ti and (B) Ti6Al4V surfaces, respectively. Ti, Ti6Al4V, MCSP-1, MCSP-2, MCSP-3, and MCSP-4 correspond respectively to the bare surfaces and those functionalized with MCSP-1, MCSP-2, MCSP-3, and MCSP-4. Error bars for (A,B) represent the standard deviation of multiple experiments. \*: Significant differences correspond to  $p = 0.0001$ .

(Figures 5B, D, respectively). When modified with MCSP2 the surface roughness of bare Ti increases to about 21 nm from 9.6 nm (Figure 5E). Almost the same RMS increase was


observed for the MCSP-3 functionalized Ti surface (Figure 5E). The RMS of a bare Ti6Al4V surface practically did not change after the adsorption of MCSP2. The AFM






height images for both metal surfaces after functionalization with MCSP-2 demonstrated similar peptide deposition patterns with an agglomerates size of approximately 200 nm. This led to an increased variation of surface roughness after functionalization with MCSP-2 as indicated by the high values of RMS standard deviation. Functionalization with MCSP-1 and MCSP-4 did not increase the surface roughness of both metals (Figure 5E), however, the presence of MCSP-4 increased significantly cell adhesion on the surfaces (Figure 4). These results indicate that the higher cell adhesion forces measured for the MCSP-2 and MCSP-4 functionalized Ti and Ti6Al4V surfaces (Figure 4) cannot be explained solely by changes in the surface roughness.

The cell adhesion measurements were completed with an *in vitro* para-nitrophenyl-phosphate (pNPP) cell viability test monitoring the adhesion of a high number of cells on the whole surface of the sample (Humphries, 2001). The number



of adherent live cells was counted on functionalized and non-functionalized Ti and Ti6Al4V surfaces after 4-h incubation and elimination of non-adherent cells by rinsing. The results indicated a higher number of adherent cells on Ti surfaces (with or without functionalization) than on Ti6Al4V surfaces (Figure 6). Surfaces functionalized with MCSP-2 demonstrated a higher concentration of adherent cells which is statistically significant compared to the non-functionalized surfaces. The pNPP test revealed that MCSP-2 is the most effective bifunctional peptide for Ti and Ti6Al4V surface functionalization in view of increased oral keratinocyte adhesion. To test the robustness of the effect of MCSP-2, single-cell adhesion forces were compared before and after bovine serum albumin (BSA) adsorption on functionalized MCSP-2 surfaces. BSA is a widely used blocking agent to prevent nonspecific binding by inhibiting hydrophobic, ionic or electrostatic interactions (Gergely et al, 2004; Buchwalow et al, 2011) between proteins. Adhesion forces of a living oral keratinocyte cell against naked and MCSP-2-functionalized Ti and Ti6Al4V surfaces before and after BSA treatment were measured using AFM force spectroscopy (Figure 7). A statistically significant increase in the adhesion forces of living oral keratinocytes on MCSP2 functionalized Ti and Ti6Al4V surfaces was found compared to those measured against the bare surfaces. After BSA adsorption, cell adhesion to Ti6Al4V/MCSP-2 surfaces decreased drastically to the level of adhesion measured on the bare Ti6Al4V surface. Although cell adhesion forces decreased after BSA treatment in the case of Ti/MCSP-2 surfaces as well, they remained significantly higher than the adhesion measured against bare Ti.



### 3.4 *In vivo* study in a rat model

Finally, the capacity of MCSP-2 peptide to increase the epithelial cell attachment on implant (Ti6Al4V) surfaces *in vivo* was evaluated. Only one of the two *in vitro* tested surfaces was selected, because the Ti6Al4V alloy is the most commonly used material for transgingival implant abutments (Ritz et al, 2017). We expected difference in cell adhesion on functionalized and bare implants could be detected 4 weeks after implantation in the rat mouth, when the apical migration of epithelial cells could be evaluated. It is known that 4 weeks after the implantation the peri-implant junctional epithelium migrates further apically and occupies 40% of the total interface between the implant and the soft connective tissue that is rich in collagen fiber and fibroblasts (Sculean et al, 2014). This is an unwanted phenomenon that could influence the primary stabilization of the implant.

Therefore, histological slices of the peri-implant gingiva interface were compared after 4 weeks when the implant was coated with the bi-functional metal-cell binding MCSP-2 peptide (Figure 8A), for a bare implant (Figure 8B) and when only the metal-binding part of the peptide, MBP-1 was used for functionalization (Figure 8C). Laminin 332 was used for immunohistochemical staining that evidenced the epithelial adhesion on the implant surface (see Supplementary

Information). One can identify from the figures the main structural components, which are the connective tissue (CT), the oral epithelium (OE) and the epithelial junction (JE). The results demonstrated that the epithelial cell migration apically was successfully limited on the MCSP-2 functionalized implants (see arrows pointing to the JE-dark brown line in Figure 8A).

Apical epithelial migration on bare implant surfaces was obvious, creating a longer junctional epithelium (Figure 8B), compared to implants functionalized with the MCSP-2 bifunctional peptide where the epithelial-implant interface was much shorter (Figure 8A). It also demonstrated organized gingival conjunctive fibers (indicated by small arrows in the JE domain) as is usually observed around teeth. There was a statistically significant difference in epithelial attachment, measured as the length of JE, between the bare and the MCSP-2 functionalized implants (Figure 8D). Implants coated with the metal binding peptide MBP-1 were also included in the study (Figure 8C). In this case, the length of JE indicating epithelial adhesion was statistically different from that obtained for the bare implant, but the apical migration of the epithelial cells was not as limited as on the surfaces coated with MCSP-2.


It is evident that the proposed MCSP-2 bifunctional peptide gave the best results between compared surfaces by both increasing the gingival adhesion on the surface and successfully inhibiting the epithelial cell migration toward the apical implant part.

## 4 Discussion

The strategy to modify the Ti and Ti6Al4V substrates adopted in the present study was to use bi-functional peptides to mimic the physiological epithelial cell attachment on the normal tooth via internal basal lamina (IBL). The IBL differs significantly from a typical basement membrane in terms of its protein composition including laminin 332 (Larjava et al, 2011). Once epithelial cells have migrated to the surface of an implant, they adhere directly via basal lamina (Rompen, 2012).

Laminin-332 and E-cadherin specific sequences for designing biomimetic peptides were chosen that were able to form strong and stable adhesions with keratinocyte cells. To anchor these cell-specific peptides to Ti or Ti6Al4V 12-mer metal binding sequences were elaborated via phage display technology (Seker and Demir, 2011; Estephan et al, 2012). Amongst the four bi-functional peptides proposed in this work, only the combination resulting in MCSP-2, composed of the MBP-1 metal binding peptide and the cadherin derived peptide CSP-2, provided a favorable spatial configuration to allow double adhesion function to both Ti based metals and oral epithelial cells.

It was demonstrated that indeed the SVSVGMPKSPRP sequence (MBP-1) adhered strongly to both implant surfaces and was resistant to hydrophobic, hydrophilic and ionic rinsing procedures (see Supplementary Information). The cell binding part, H-SWELYPLRANL-NH<sub>2</sub> (CSP-2) was described by Devemy and Blaschuk, (2009) as the specific binding peptide of the E-cadherin ectodomain with a high binding affinity of 9.4 mM.



**FIGURE 8**

*In vivo* comparison of the epithelial adhesions on the peptide covered and bare implant surfaces, respectively. Histological slices of peri-implant gingiva in the fourth week after implantation are presented. The immunohistochemical coloration of laminin 332 indicates the epithelial adhesion on the implant surface. The dark arrows indicate cervical and apical points of measurement of the laminin 332 distribution, which is visualized like a dark brown line at the implant-epithelium interface; **(A)**: MCSP-2 covered Ti6Al4V implant; **(B)**: bare implant; **(C)**: MBP-1 covered implant; CT: connective tissue; OE: Oral Epithelium; JE: epithelial junction; Bare scale = 10 µm; **(D)** Statistical analysis of multiple measurements of the length of junctional epithelium for the 6 implanted animals. JE length after 4 weeks of healing demonstrates a statistically significant difference between the bi-functional peptides (MCSP-2) and the bare Ti6Al4V alloy ( $p < 0.05$ ). The length of JE indicating epithelial adhesion was smaller also for the MBP-1 coated implant, but the apical migration of the epithelial cells was not as limited as on the surfaces functionalized with MCSP-2.

Both the cell culture tests and single force spectroscopy results demonstrated that the bi-functional MCSP-2 peptide increased oral keratinocyte cell adhesion on Ti and Ti6Al4V surfaces. The MCSP-2 peptide generated cell adhesion stable 4 hours after incubation even after BSA adsorption.

A similar approach using a bi-functional peptide was reported for triggering the endothelialization of Ti6Al4V in endovascular prostheses, showing the relevance of the peptide route functionalization (Meyers et al, 2011). Yazici et al, (2013) used Cell surface display to select peptides with higher affinity to Ti surfaces and then a bi functional peptide was designed by combining this later peptide with an integrin recognizing peptide motif. Results showed that the functionalization of the Ti surface by the selected bi-peptides significantly enhanced the bioactivity of osteoblast and fibroblast cells on implant-grade materials. It was also demonstrated that the use of combined titanium-hydroxyapatite peptides based on titanium peptide binder (KKLPDA) and hydroxyapatite peptide binder (EEEEEEEE) can be successfully adsorbed onto Ti6Al4V and hydroxyapatite surfaces (Rodriguez et al, 2017). An Alternative way is also reported, it consists of treating the Ti6Al4V by activated vapor salinization and then covalent attaching the RGD oligopeptides to the surface to increase the attachment, spreading and rearrangement of mesenchymal stem and progenitor cells (Álvarez-López et al, 2022). Just to note that all these reports lack *in vivo* studies.

*In vivo* animal studies confirmed the capacity of our bi-functional peptide to ensure stable cell adhesion on the trans-gingival part of the dental implant and also to arrest the unwanted apical migration of the epithelial cells pointing toward promising medical applications of the proposed novel peptide. The value of the rat model confirmation of the effectiveness of the MCSP-2 bi functional peptide lies in the fact that the *in vivo* study indirectly reflects the problem with oral biofilm formation.

It is known that epithelial adhesion to implant surfaces is poorer (Fujiseki et al, 2003) and therefore conducive to clinical problems such as peri-implantitis (Koldslund et al, 2010). From the results, it can be assumed that the MCSP-2 bi functional peptide by enhancing epithelial attachment arrested biofilm plaque formation and apical migration.

We believe that the results of the present study provide serious evidence on molecular, cellular and animal level that coating Ti implants with such bioengineered peptide represents an effective method to improve mucosal adhesion on dental implants and to reduce peri-implantitis and mucositis prevalence. The method could rapidly be clinically evaluated before finding clinical applications.

## 5 Conclusion

In order to improve osteo-integration, biocompatibility and bioactivity of titanium implant and its alloys, numerous strategies for surface modifications are already known. In this work, we report on combining two peptides: one specific and selective for Ti and Ti6Al4V surfaces extracted by phage display technology and another one that can present affinity to epithelial cells. This latter bi peptide can result in powerful bioengineered peptides. Four bi functional peptides were synthesized and evaluated using single-cell force spectroscopy, surface roughness and cell viability tests. It was shown that MCSP-2 is the most effective one. Our MCSP-2 showed on cellular and animal levels its capacity to improve gingival adhesion and to diminish peri-implantitis and mucositis prevalence. This latter bi-peptide can be clinically evaluated to introduce it in the medical field where Ti and its alloys are used. i.e., manufacturing of hip, artificial knee joints and dental implant prostheses components.

## Data availability statement

The datasets presented in this study can be found in online repositories. The names of the repository/repositories and accession number(s) can be found in the article/Supplementary Material.

## Ethics statement

The animal study was reviewed and approved by the committee for animal welfare of Montpellier University with referral number 1143 15/03/2015.

## Author contributions

IP: Conceptualization, formal analysis, investigation, writing—original draft, writing—review and editing. AV: Investigation, writing—review and editing. MM: Investigation, writing—review and editing. BV: Formal analysis, methodology. CL: Formal analysis, methodology. CG: Conceptualization, formal analysis, methodology. FC: Conceptualization, formal analysis, methodology, supervision, visualization. EE: Conceptualization, investigation, writing—review and editing. All authors contributed to the article and approved the submitted version.

## References

- Álvarez-López, A., Colchero, L., Elices, M., Guinea, G. V., Pérez-Rigueiro, J., and González-Nieto, D. (2022). Improved cell adhesion to activated vapor silanization-biofunctionalized Ti-6Al-4V surfaces with ECM-derived oligopeptides. *Biomater. Adv.* 133, 112614. doi:10.1016/j.msec.2021.112614
- Atieh, M. A., Ibrahim, H. M., and Atieh, A. H. (2010). Platform switching for marginal bone preservation around dental implants: A systematic review and meta-analysis. *J. Periodontol.* 81, 1350–1366. doi:10.1902/jop.2010.100232
- Atsuta, I., Ayukawa, Y., Ogino, Y., Moriyama, Y., Jinno, Y., and Koyan, K. (2012). Evaluations of epithelial sealing and peri-implant epithelial down-growth around “step-type” implants. *Oral Impl Res.* 23, 459–466. doi:10.1111/j.1600-0501.2011.02163.x
- Baffone, G. M., Botticelli, D., Canullo, L., Scala, A., Beolchini, M., and Lang, N. P. (2012). Effect of mismatching abutments on implants with wider platforms – An experimental study in dogs. *Clin. Oral Implants Res.* 23, 334–339. doi:10.1111/j.1600-0501.2011.02320.x
- Baharloo, B., Textor, M., and Brunette, D. M. (2005). Substratum roughness alters the growth, area, and focal adhesions of epithelial cells, and their proximity to titanium surfaces. *J. Biomed. Mater. Res. A* 74, 12–22. doi:10.1002/jbm.a.30321
- Brannemark, P., Hansson, B., Adell, R., Breine, U., Lundstrom, J., Hallen, O., et al. (1977). Osseointegrated implants in the treatment of the edentulous jaw. Experience from a 10-year period. *Scand. J. Plast. Reconstr. Surg.* 11, 1–132.
- Buchwalow, I., Samoilo, V., Boecker, W., and Tiemann, M. (2011). Non-specific binding of antibodies in immunohistochemistry: Fakes and facts. *Nat. Preced* 28, 1–6. doi:10.1038/npre.2011.5892.1
- Cumbo, C., Marigo, L., Somma, F., Torre, G., Minicacchi, I., and D’Addona, A. (2013). Implant platform switching concept: A literature review. *Eur. Rev. Med. Pharmacol. Sci.* 17, 392–397.
- de Oliveira, P. G. F. P., de Melo Soares, M. S., Silveira E Souza, A. M. M., Taba, M., Palioto, D. B., Messoro, M. R., et al. (2021). Influence of nano-hydroxyapatite coating implants on gene expression of osteogenic markers and micro-CT parameters. An *in vivo* study in diabetic rats. *J. Biomed. Mater. Res. A* 109 (5), 682–694. doi:10.1002/jbm.a.37052
- Devemy, E., and Blaschuk, O. W. (2009). Identification of a novel dual E-and N-cadherin antagonist. *Peptides* 30, 1539–1547. doi:10.1016/j.peptides.2009.05.010

## Acknowledgments

Special acknowledgments to Professor Paul Tramini for his help in statistical analysis.

## Conflict of interest

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

## Publisher’s note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

## Supplementary material

The Supplementary Material for this article can be found online at: <https://www.frontiersin.org/articles/10.3389/fbioe.2023.1165853/full#supplementary-material>

Drexelius, M., Arnold, R., Meinberger, D., Wilhelm, M., Mathur, S., and Neundorff, I. (2023). Rational design of bifunctional chimeric peptides that combine antimicrobial and titanium binding activity. *J. Pept. Sci.* 2023, e3481. doi:10.1002/psc.3481

Stephan, E., Dao, J., Saab, M. B., Panayotov, I., Martin, M., Larroque, C., et al. (2012). Svsygmksppr: A broad range adhesion peptide. *Biomed. Tech. Eng.* 57, 481–489. doi:10.1515/bmt-2011-0109

Fu, X., Liu, P., Zhao, D., Yuan, B., Xiao, Z., Zhou, Y., et al. (2020). Effects of nanotopography regulation and silicon doping on angiogenic and osteogenic activities of hydroxyapatite coating on titanium implant. *Int. J. Nanomedicine* 15, 4171–4189. doi:10.2147/ijn.s252936

Fujiseki, M., Matsuzaka, K., Yoshinari, M., Shimono, M., and Inoue, T. (2003). An experimental study on the features of peri-implant epithelium: Immunohistochemical and electron-microscopic observations. *Bull. Tokyo Dent. Coll.* 44, 185–199. doi:10.2209/tdpublication.44.185

Geng, H., Yuan, Y., Adayi, A., Zhang, X., Song, X., Gong, L., et al. (2018). Engineered chimeric peptides with antimicrobial and titanium-binding functions to inhibit biofilm formation on Ti implants. *Mater. Sci. Eng. C* 82, 141–154. doi:10.1016/j.msec.2017.08.062

Gergely, C., Bahi, S., Szalontai, B., Flores, H., Schaaf, P., Voegel, J. C., et al. (2004). Human serum albumin self-assembly on weak polyelectrolyte multilayer films structurally modified by pH changes. *Langmuir* 20, 5575–5582. doi:10.1021/la049932x

Gergely, C., Hemmerle, J., Schaaf, P., Horber, J. K., Voegel, J. C., and Senger, B. (2002). Multi-Bead-and-Spring model to interpret protein detachment studied by AFM force spectroscopy. *Biophys. J.* 83, 706–722. doi:10.1016/s0006-3495(02)75202-8

Gergely, C., Voegel, J. C., Schaaf, P., Senger, B., Maaloum, M., Horber, J. K., et al. (2000). Unbinding process of adsorbed proteins under external stress studied by atomic force microscopy spectroscopy. *PNAS Biophys. Comput. Biol.* 97, 10802–10807. doi:10.1073/pnas.180293097

Heitz-Mayfield, L. J., and Mombelli, A. (2014). The therapy of peri-implantitis: A systematic review. *Int. J. Oral Maxillofac. Implants* 29, 325–345. doi:10.11607/jomi.2014suppl.g5.3

Humphries, M. J. (2001). *In: Current protocols in cell biology*. New Jersey, United States: John Wiley & Sons, Inc. doi:10.1002/0471143030.cb0901s00



- Hutter, J. L., Bechhoefer, J., Sader, J. E., Chon, J. W. M., and Mulvaney, P. (1993). Erratum: "Calibration of atomic-force microscope tips" [rev. Sci. Instrum. 64, 1868 (1993)]. *Rev. Sci. Instrum.* 64, 3342. doi:10.1063/1.1144449
- Ikedo, H., Yamaza, T., Yoshinari, M., Ohsaki, Y., Ayukawa, Y., Kido, M. A., et al. (2000). Ultrastructural and immunoelectron microscopic studies of the peri-implant epithelium-implant (Ti-6Al-4V) interface of rat maxilla. *J. Periodontol.* 71, 961–973. doi:10.1902/jop.2000.71.6.961
- Katsikeris, N., Listrom, R., and Symington, J. (1987). Interface between titanium 6, 4 alloy implants and bone. *Int. J. Oral Maxillofac. Surg.* 16, 473–476. doi:10.1016/s0901-5027(87)80087-5
- Kazmierczak, P., and Przekora, A. (2020). Osteoconductive and osteoinductive surface modifications of biomaterials for bone regeneration: A concise review. *Coatings* 10 (10), 971–1025. doi:10.3390/coatings10100971
- Kim, J. M., Park, W. H., and Min, B. M. (2005). The PPFLMLLKGSTR motif in globular domain 3 of the human laminin-5  $\alpha 3$  chain is crucial for integrin  $\alpha 3 \beta 1$  binding and cell adhesion. *Exp. Cell Res.* 304, 317–327. doi:10.1016/j.yexcr.2004.11.009
- Koldsland, O. C., Scheie, A. A., and Aass, A. M. (2010). Prevalence of peri-implantitis related to severity of the disease with different degrees of bone loss. *J. Periodontol.* 81, 231–238. doi:10.1902/jop.2009.090269
- Kumar Boda, S., and Aparicio, C. (2022). Dual keratinocyte-attachment and anti-inflammatory coatings for soft tissue sealing around transmucosal oral implants. *Biomater. Sci.* 10 (3), 665–677. doi:10.1039/d1bm01649k
- Lallukka, M., Gama, F., Gobbo, V. A., Prato, M., Najmi, Z., Cochis, A., et al. (2022). Surface functionalization of Ti6Al4V-eli alloy with antimicrobial peptide nisin. *Nanomaterials* 12 (23), 4332–4418. doi:10.3390/nano12234332
- Larjava, H., Koivisto, L., Hakkinen, L., and Heino, J. (2011). Epithelial integrins with special reference to oral epithelia. *J. Dent. Res.* 90, 1367–1376. doi:10.1177/0022034511402207
- Litvinov, R. I., Vilaire, G., Shuman, H., Bennett, J. S., and Weisel, J. W. (2003). Quantitative analysis of platelet  $\alpha v \beta 3$  binding to osteopontin using laser tweezers. *J. Biol. Chem.* 278, 51285–51290. doi:10.1074/jbc.m304581200
- Liu, Z., Ma, S., Duan, S., Xuliang, D., Sun, Y., Zhang, X., et al. (2016). Modification of titanium substrates with chimeric peptides comprising antimicrobial and titanium-binding motifs connected by linkers to inhibit biofilm formation. *ACS Appl. Mater. Interfaces* 8 (8), 5124–5136. doi:10.1021/acsami.5b11949
- Mattila, K. J., Pussinen, P. J., and Paju, S. (2005). Dental infections and cardiovascular diseases: A review. *J. Periodontol.* 76, 2085–2088. doi:10.1902/jop.2005.76.11.s.2085
- Meyers, S. R., Kenan, D. J., Khoo, X., and Grinstaff, M. W. (2011). Bioactive stent surface coating that promotes endothelialization while preventing platelet adhesion. *Biomacromolecules* 12, 533–539. doi:10.1021/bm101212k
- Mombelli, A., Muller, N., and Cionca, N. (2012). The epidemiology of peri-implantitis. *Clin. Oral Implants Res.* 23, 67–76. Suppl 6. doi:10.1111/j.1600-0501.2012.02541.x
- Mumith, A., Cheong, V. S., Fromme, P., Coathup, M. J., and Blunn, G. W. (2020). The effect of strontium and silicon substituted hydroxyapatite electrochemical coatings on bone ingrowth and osseointegration of selective laser sintered porous metal implants. *PLoS One* 15 (1), 02272322–0227319. doi:10.1371/journal.pone.0227232
- Panayotov, I. V., Vladimirov, B. S., Dutilleul, P. Y., Levallois, B., and Cuisinier, F. (2015). Strategies for immobilization of bioactive organic molecules on titanium implant surfaces—a review. *Folia Med. Plovdiv.* 57, 11–18. doi:10.1515/folmed-2015-0014
- Panorchan, P., Thompson, M. S., Davis, K. J., Tseng, Y., Konstantopoulos, K., and Wirtz, D. (2006). Single-molecule analysis of cadherin-mediated cell-cell adhesion. *J. Cell Sci.* 119, 66–74. doi:10.1242/jcs.02719
- Puech, P. H., Taubenberger, A., Ulrich, F., Krieg, M., Muller, D. J., and Heisenberg, C. P. (2005). Measuring cell adhesion forces of primary gastrulating cells from zebrafish using atomic force microscopy. *J. Cell Sci.* 118, 4199–4206. doi:10.1242/jcs.02547
- Pussinen, P. J., Alftan, G., Jousilahti, P., Paju, S., and Tuomilehto, J. (2007). Systemic exposure to Porphyromonas gingivalis predicts incident stroke. *Atherosclerosis* 193, 222–228. doi:10.1016/j.atherosclerosis.2006.06.027
- Ritz, U., Nusselt, T., Sewing, A., Ziebart, T., Kaufmann, K., Baranowski, A., et al. (2017). The effect of different collagen modifications for titanium and titanium nitride surfaces on functions of gingival fibroblasts. *Clin. Oral Investig.* 21, 255–265. doi:10.1007/s00784-016-1784-5
- Rodriguez, G. M., Bowen, J., Grossin, D., Ben-Nissan, B., and Stamboulisa, A. (2017). Functionalisation of Ti6Al4V and hydroxyapatite surfaces with combined peptides based on KKLPGA and EEEEEEE peptides. *Colloids Surf. B Biointerfaces* 160, 154–160. doi:10.1016/j.colsurfb.2017.09.022
- Rompen, E., Domken, O., Degidi, M., Pontes, A. E., and Piattelli, A. (2006). The effect of material characteristics, of surface topography and of implant components and connections on soft tissue integration: A literature review. *Clin. Oral Implants Res.* 17, 55–67. doi:10.1111/j.1600-0501.2006.01367.x
- Rompen, E. (2012). The impact of the type and configuration of abutments and their (repeated) removal on the attachment level and marginal bone. *Eur. J. Oral Implantol.* 5, S83–S90.
- Roos-Jansaker, A. M., Renvert, S., and Egelberg, J. (2003). Treatment of peri-implant infections: A literature review. *J. Clin. Periodontol.* 30, 467–485. doi:10.1034/j.1600-051x.2003.00296.x
- Sader, J. E., Chon, J. W. M., and Mulvaney, P. (1999). Calibration of rectangular atomic force microscope cantilevers. *Rev. Sci. Instrum.* 70, 3967–3969. doi:10.1063/1.1150021
- Sculean, A., Gruber, R., and Bosshardt, D. D. (2014). Soft tissue wound healing around teeth and dental implants. *J. Clin. Periodontol.* 41, S6–S22. doi:10.1111/jcpe.12206
- Seker, U. O., and Demir, H. V. (2011). Material binding peptides for nanotechnology. *Molecules* 16, 1426–1451. doi:10.3390/molecules16021426
- Sivasankar, S., Zhang, Y., Nelson, W. J., and Chu, S. (2009). Characterizing the initial encounter complex in cadherin adhesion. *Structure* 17, 1075–1081. doi:10.1016/j.str.2009.06.012
- Smeets, R., Henningsen, A., Jung, O., Heiland, M., Hammacher, C., and Stein, J. M. (2014). Definition, etiology, prevention and treatment of peri-implantitis—a review. *Head. Face Med.* 10 (1–13), 34. doi:10.1186/1746-160x-10-34
- Souza, J. C. M., Sordi, M. B., Kanazawa, M., Ravindran, S., Henriques, B., Silva, F. S., et al. (2019). Nano-scale modification of titanium implant surfaces to enhance osseointegration. *Acta Biomater.* 94, 112–131. doi:10.1016/j.actbio.2019.05.045
- Taubenberger, A., Cisneros, D. A., Friedrichs, J., Puech, P. H., Muller, D. J., and Franz, C. M. (2007). Revealing early steps of  $\alpha 3 \beta 1$  Integrin-mediated adhesion to collagen type I by using single-cell force spectroscopy. *Mol. Biol. Cell* 18, 1634–1644. doi:10.1091/mbc.e06-09-0777
- Taubenberger, A. V., Woodruff, M. A., Bai, H., Muller, D. J., and Hutmacher, D. W. (2010). The effect of unlocking RGD-motifs in collagen I on pre-osteoblast adhesion and differentiation. *Biomaterials* 31, 2827–2835. doi:10.1016/j.biomaterials.2009.12.051
- Traini, T., Murmura, G., Sinjari, B., Perfetti, G., Scarano, A., D'Arcangelo, C., et al. (2018). The surface anodization of titanium dental implants improves blood clot formation followed by osseointegration. *Coatings* 8 (7), 252. doi:10.3390/coatings8070252
- Van den Borre, C. E., Zigterman, B. G. R., Mommaerts, M. Y., and Braem, A. (2022). How surface coatings on titanium implants affect keratinized tissue: A systematic review. *J. Biomed. Mater. Res. B Appl. Biomater.* 110 (7), 1713–1723. doi:10.1002/jbm.b.35025
- Vegh, A. G., Fazakas, C., Nagy, K., Wilhelm, I., Molnar, J., Krizbai, I. A., et al. (2012). Adhesion and stress relaxation forces between melanoma and cerebral endothelial cells. *Eur. Biophys. J. EBJ* 41, 139–145. doi:10.1007/s00249-011-0765-5
- Weder, G., Voros, J., Giazson, M., Matthey, N., Heinzelmann, H., and Liley, M. (2009). Measuring cell adhesion forces during the cell cycle by force spectroscopy. *Biointerphases* 4, 27–34. doi:10.1116/1.3139962
- Werner, S., Huck, O., Frisch, B., Vautier, D., Elkaim, R., Voegel, J. C., et al. (2009). The effect of microstructured surfaces and laminin-derived peptide coatings on soft tissue interactions with titanium dental implants. *Biomaterials* 30, 2291–2301. doi:10.1016/j.biomaterials.2009.01.004
- Wisdom, E. C., Zhou, Y., Chen, C., Tamerler, C., and Snead, M. L. (2020). Mitigation of peri-implantitis by rational design of bifunctional peptides with antimicrobial properties. *ACS Biomater. Sci. Eng.* 6 (5), 2682–2695. doi:10.1021/acsbomaterials.9b01213
- Willemsen, O. H., Snel, M. M., Werf, K. O., Grooth, B. G., Greve, J., Hinterdorfer, P., et al. (1998). Simultaneous height and adhesion imaging of antibody-antigen interactions by atomic force microscopy. *Biophys. J.* 75, 2220–2228. doi:10.1016/s0006-3495(98)77666-0
- Wisdom, C., Chen, C., Yuca, E., Zhou, Y., Tamerler, C., and Snead, M. L. (2019). Repeatedly applied peptide film kills bacteria on dental implants. *JOM* 71 (4), 1271–1280. doi:10.1007/s11837-019-03334-w
- Wong, J., Chilkoti, A., and Moy, V. T. (1999). Direct force measurements of the streptavidin–biotin interaction. *Biomol. Eng.* 16, 45–55. doi:10.1016/s1050-3862(99)00035-2
- Yazici, H., Fong, H., Wilson, B., Oren, E. E., Amos, F. A., Zhang, H., et al. (2013). Biological response on a titanium implant-grade surface functionalized with modular peptides. *Acta Biomater.* 9, 5341–5352. doi:10.1016/j.actbio.2012.11.004
- Yazici, H., O'Neill, M. B., Kacar, T., Wilson, B. R., Oren, E. E., Sarikaya, M., et al. (2016). Engineered chimeric peptides as antimicrobial surface coating agents toward infection-free implants. *ACS Appl. Mater. Interfaces* 8 (8), 5070–5081. doi:10.1021/acsaami.5b03697
- Yucesoy, D. T., Hnilova, M., Boone, K., Arnold, P. M., Snead, M. L., and Tamerler, C. (2015). Chimeric peptides as implant functionalization agents for titanium alloy



implants with antimicrobial properties. *JOM* 67 (4), 754–766. doi:10.1007/s11837-015-1350-7

Zhang, X., Geng, H., Gong, L., Zhang, Q., Li, H., Zhang, X., et al. (2018). Modification of the surface of titanium with multifunctional chimeric peptides to prevent biofilm formation via inhibition of initial colonizers. *Int. J. Nanomedicine* 13, 5361–5375. doi:10.2147/ijn.s170819

Zhang, X., Wojcikiewicz, E. P., and Moy, V. T. (2006). Dynamic adhesion of T lymphocytes to endothelial cells revealed by atomic force microscopy. *Exp. Biol. Med.* 231, 1306–1312. doi:10.1177/153537020623100804

Zhao, Z., Ma, S., Wu, C., Li, X., Ma, X., Hu, H., et al. (2021). Chimeric peptides quickly modify the surface of personalized 3D printing titanium implants to promote

osseointegration. *ACS Appl. Mater Interfaces* 13 (29), 33981–33994. doi:10.1021/acsami.1c11207

Zhou, Y., Snead, M. L., and Tamerler, C. (2015). Bio-inspired hard-to-soft interface for implant integration to bone. *Nanomedicine Nanotechnol. Biol. Med.* 11 (2), 431–434. doi:10.1016/j.nano.2014.10.003

Zigterman, B. G. R., Van den Borre, C., Braem, A., and Mommaerts, M. Y. (2019). Titanium surface modifications and their soft-tissue interface on nonkeratinized soft tissues—a systematic review (review). *Biointerphases* 14 (4), 040802. doi:10.1116/1.5113607

Zitzmann, N. U., and Berglundh, T. (2008). Definition and prevalence of peri-implant diseases. *J. Clin. Periodontol.* 35, 286–291. doi:10.1111/j.1600-051x.2008.01274.x