



Uncertainty Estimation for the Brillouin Frequency Shift Measurement Using a Scanning Tandem Fabry–Pérot Interferometer

Patrice Salzenstein, Thomas Yong Wu

► To cite this version:

Patrice Salzenstein, Thomas Yong Wu. Uncertainty Estimation for the Brillouin Frequency Shift Measurement Using a Scanning Tandem Fabry–Pérot Interferometer. *Micromachines*, 2023, 14 (7), pp.1429. 10.3390/mi14071429 . hal-04228715

HAL Id: hal-04228715

<https://hal.science/hal-04228715>

Submitted on 4 Oct 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Uncertainty estimation for the Brillouin frequency shift measurement using a scanning tandem Fabry–Pérot interferometer

Patrice Salzenstein ^{1,*}, Thomas Y. Wu ^{2,*}

¹ Centre National de la Recherche Scientifique (CNRS), Franche-Comté Electronique Mécanique Thermique Optique Sciences et Technologies (FEMTO-ST) Institute, Université de Franche-Comté (UFC), Besançon, France

² National Metrology Centre (NMC), Agency for Science, Technology and Research (A*STAR), 8 CleanTech Loop, #01-20, Singapore 637145, Republic of Singapore

* Correspondence : patrice.salzenstein@cnrs.fr, yongusa@qq.com

Abstract: The expanded uncertainty of the measured Brillouin scattering shift frequencies is essential in assessing the measurements of parameters of various materials. We describe the general operation principles of a Brillouin light scattering (BLS) spectrometer with a high-power laser and a scanning tandem Fabry–Pérot interferometer (TFPI) for material characterization. Various uncertainty components have been analyzed for the BLS spectrometer following the Guide to the Expression of Uncertainty in Measurement (GUM). The expanded relative uncertainty in the measured Brillouin frequency shift of 15.70 GHz for polymethyl methacrylate (PMMA) was estimated to be 0.26%. The calculated Brillouin frequency shift (based on material properties of PMMA) was determined to be 15.44 GHz with expanded relative uncertainty of 2.13%. It was shown that the measured and calculated Brillouin frequency shifts for PMMA agree within their expanded uncertainties. The TFPI-based BLS spectrometer can be used to measure the longitudinal modulus of materials with an expanded uncertainty of 1.9%, which is smaller than that of the ultrasonic velocity-based method (estimated to be 2.9%).

Keywords: brillouin light scattering; high-power laser; tandem Fabry–Pérot interferometer; Brillouin spectroscopy; elastic property; speed of sound; measurement uncertainty analysis;

1. Introduction

Brillouin light scattering (BLS) technique is a non-contact, non-destructive method for studying the elastic properties of materials and it is gradually gaining popularity in various industrial applications and research laboratories. BLS is the inelastic scattering of light (photons) by thermally generated acoustic vibrations (acoustic phonons) [1–3]. The thermally excited sound waves in materials have very weak intensity which cannot be detected by ultrasonic methods. We would like to recall the main contributions to the detection of sound waves, by brothers Curie [4], and the theoretical prediction of inelastic light scattering in materials by thermally excited acoustic phonons, by Brillouin [1] and Mandelstam [2] in 1920s. The first BLS spectrum from acoustic phonons of liquids was observed and reported in 1930 by Gross [5–7]. BLS measurements have been reported in [8–10] in 1930s. Grimsditch and Ramdas have made precise measurements with Brillouin scattering on Diamond in 1975 [11].

Kojima showed that BLS techniques are very useful for studying material properties [12]. Magnons can be divided into surface-like excitations and bulk-like excitations [3,13]. The magnetic properties of materials can also be measured via their magnetic excitations (magnons) using Brillouin scattering [13]. Brillouin spectroscopy has become an essential tool for the study of acoustic phonons, magnons and elastic properties of materials [12,14–16]. Brillouin spectroscopy and microscopy have emerged as a non-destructive, non-contact and label-free method for probing the viscoelastic properties of biological samples

[17–20]. Brillouin scattering has also been used in atmospheric aerosol measurement, study of air molecules and profiling of aerosols' optical properties [21–28]

Both Raman scattering and Brillouin scattering arise from the inelastic scattering of light. Raman scattering is associated with the scattering by optical phonons and molecular vibrations [29,30]. Brillouin spectroscopy is complementary to the Raman spectroscopy in material characterisation [31]. There are some key differences between the two types of inelastic light-scattering spectroscopy. The Raman spectroscopy is based on scattering by optical phonons with frequency shift at THz range and determines the sample's chemical composition and molecular structure using an interferometer or a dispersive spectrometer. Whereas the Brillouin spectroscopy is based on scattering of acoustic phonons with frequency shift at GHz range and traditionally measures the elastic properties of materials using scanning tandem Fabry–Pérot interferometer (TFPI) [32–38]. Brillouin and Raman micro-spectroscopy have been combined to obtain the Brillouin and Raman spectra of biological samples simultaneously to assign their chemical specificity to mechanical properties [39–41]. The multimodal micro-spectroscopy developed in [42] is based on simultaneous detection of Brillouin and Raman scattering with spectral coverage up to 100 THz.

The measurement of Brillouin spectra would need appropriate instrument like the Fabry–Pérot interferometer (FPI) which can provide high contrast [37,38,43–52]. A double-pass FPI was implemented by Sandercock in 1970 to detect the BLS [33]. Improved methods using multi-pass TFPI had been reported by Lindsay, Anderson and Sandercock [53], Dil *et al.* [54] and Mock *et al.* [55] to achieve 150 dB high-contrast measurement of the Brillouin frequency shift. The scanning multi-pass TFPI technique had been further improved with automatic computer control [51,56].

The scanning TFPI would require a long acquisition time to measure the Brillouin spectra. The long acquisition time makes the traditional scanning TFPI unsuitable for high-throughput biomedical applications or dynamic measurement. To make a rapid spectrum measurement within 1 s, a non-scanning angular dispersive FPI (ADFPI) was developed using a solid etalon and a multichannel detector [57–59]. A virtually imaged phased array (VIPA) was proposed to achieve large angular dispersion [60], which can be used to build another type of ADFPI. A non-scanning Brillouin spectrometer employing a VIPA etalon and CCD camera was developed to acquire Brillouin spectra within just a few seconds [61]. Cascading three-stage VIPA etalons can provide an extinction ratio up to 80 dB to reduce the Rayleigh scattering background and crosstalk substantially [62]. To further suppress the high scattering background, molecular or atomic absorption cells were introduced as notch filter before a single-stage VIPA spectrometer to absorb the Rayleigh scattering [39,63]. In such a design, the Brillouin peak position and shape could be altered by the atomic-molecular absorption filters, therefore a customized least-square fitting algorithm had been proposed to retrieve the Brillouin shifts and linewidths with high accuracy [64].

Coker *et al.* [65] have compared two VIPA-based spectrometers (780 nm and 532 nm wavelength) with molecular or atomic absorption cell to a scanning 6-pass TFPI to assess their measurement accuracy. With a acquisition time of ~0.5 s, the three Brillouin spectrometers were used to measure the Brillouin frequency and linewidth for acetone. The results showed that the scanning TFPI gave smaller deviation in Brillouin frequency (from the theoretical value) and narrower linewidth. This study has explored the possibility of reducing frequency and linewidth measurement standard deviations through extending the acquisition times using TFPI and VIPA-based spectrometers. They showed that the 780 nm VIPA-based spectrometer can achieve the minimum standard deviation in Brillouin frequency and linewidth measurement (for acetone) using much shorter acquisition time than TFPI, and its linewidth measurement accuracy is decided by the laser stability and optical components' quality [65].

Stimulated Brillouin scattering (SBS) is a nonlinear process which has been applied in optical fibers and optoelectronic engineering [66–69]. SBS manifests itself through the creation of a backward propagating Stokes wave. Most of the input power is carried by

Stokes wave when the Brillouin threshold is reached. Impulsive and frequency-domain SBS spectroscopy and imaging systems have been developed to avoid the issue of strong Rayleigh scattering background and achieve high-speed microscopic imaging [70–73]. Ballmann et al. [74] have compared the Brillouin shift measurement accuracy of impulsive SBS versus a 780 nm VIPA spectrometer, and showed that the Allan variances for acetone measurement using the impulsive SBS are much lower and reduce to the minimum values using a much shorter acquisition time. This study showed that the impulsive SBS is superior to VIPA spectrometer in terms of Brillouin frequency measurements stability, since the measured frequency is independent of the frequency of the pump or the probe laser.

In this paper, we will show the main steps in using a scanning 6-pass TFPI [75,76] to measure the Brillouin frequency shift and estimate the speed of propagation of phononic waves in materials. The objective of this paper is to assess the uncertainties in Brillouin frequency shift measurement using a TFPI via a consistent metrological approach. The measurement experiment is described in section 2, with the results given in section 3. The various measurement uncertainty components associated with the Brillouin frequency shift measurement will be derived in section 4. In section 5, we give a comparison of measured and calculated Brillouin frequency shift for the test material. The Brillouin frequency shift can be used to estimate the speed of the corresponding phononic wave and elastic modulus of materials with high accuracy. The measurement uncertainty in material's longitudinal modulus derived by the TFPI is also derived. Finally we will discuss the limitations, potential solutions and future work in Brillouin spectroscopy in section 6.

2. Materials and Methods

We present a method for detection of the BLS from a material under test, i. e. the Device under Test (DUT). This material can be isotropic or anisotropic. One of the key instrument in the measurement system is the scanning TFPI. Detected spectrum peaks are shifted from the frequency of the incident laser. Those offset frequencies depend on the properties of the material of the DUT. The measured Brillouin frequency shift can be used to estimate the parameters of the material, such as phase velocity of transverse and longitudinal waves and elastic modulus.

A BLS spectrometer using a 532 nm powerful Class 4 laser (up to 600 mW) is efficient to reveal the spin wave or acoustic signals, at frequencies from a few GHz to more than 100 GHz. Fluctuations of refractive index in a medium enables the detection and analysis of scattered laser light via the BLS spectrometer [53,54]. The TFPI (TFP-2, The Table Stable Ltd. [75,76]) is shown in Figure 1. The general principle is to send the signal generated by the laser and focus it on the sample that we want to characterize. The photons arrive in the material or the thin layer, and interact with the lattice or material.



Figure 1. Photograph of the scanning 6-pass TFPI (TFP-2) used for the BLS measurement.

Light helps to create phonons in the material. These phonons propagate with speeds that may be different depending on whether the mode is transverse or longitudinal. It depends on the nature of the material, which can be isotropic or anisotropic. The phonons in turn create light, which are shifted in frequency relatively to the frequency of the incident laser. The BLS spectrometer precisely makes analysis of the light scattered by a material [53,56].

The TFPI produces spectrum peaks which are shifted from the frequency of the incident laser depending on the material. Figure 2 gives the typical setup used for the measurement, showing a backward scattering configuration (scattering angle $\theta = 180^\circ$) and a picture of the measurement system.

We calibrated the BLS spectrometer with part of the laser signal, used as the bench reference. Inside the commercial TFPI spectrometer (TFP-2), the light goes through two different Fabry–Pérot interferometers with six passes. Each pair of mirrors is precisely aligned during the calibration procedure.




Figure 2. (a): Typical setup for BLS spectrometer (a backward scattering configuration). The TFP-2 is a commercial TFPI developed by the Table Stable Ltd. DUT: device under test. M: mirror. FP: Fabry–Pérot interferometer. P: prism. PD: photodetector. E: electronics. CU: computer unit. **(b):** The commercial FPI (TFP-2) is inside the box on the right side of this picture.

It is necessary to calibrate the instrument accurately, which is sensitive to mechanical vibrations, temperature and humidity. Alignment process requires an alignment of the two cavities. Each of the two cavities consists of a pair of parallel mirrors. TFPI produces two series of absorption peaks with respect to a flat noisy intensity level. We then obtain a curve providing the number of absorbed photons versus frequency.

3. Measurement results

Experiments can reproduce known Brillouin light scattering peaks of some bulk materials and thin films. Typical Brillouin scattering stimulations reveal acoustic or spin waves frequencies in the range between 3 and 150 GHz (though generally limited to around 30 GHz). In this section, we provide an example of BLS spectrum with the number of detected photons versus the frequency shift for an isotropic material.

We have measured a bulk material: polymethyl methacrylate (PMMA), using the BLS spectrometer. This isotropic material exhibits well-defined Brillouin frequency shift peaks. The BLS peaks are produced by sound waves in materials and they can be analysed by means of a Damped Harmonic Oscillator function (DHO). From the BLS, we can deduce parameters of the material such as the phase velocity of longitudinal waves. Knowing the n (optical refractive index of the material), λ_0 (laser wavelength), θ (scattering angle) and v (phase velocity of longitudinal waves), the Brillouin frequency shift ν_B can also be calculated by :

$$v_B = \frac{2nv}{\lambda_0} \sin\left(\frac{\theta}{2}\right) \quad (1)$$

where the phase velocity of longitudinal waves in the material can be obtained from literature or calculated as $v = \sqrt{c_{11}/\rho}$, ρ is the density of the material, c_{11} is the longitudinal modulus.




Figure 3. BLS spectrum for PMMA with a measured Brillouin frequency shift at 15.70 GHz (longitudinal acoustic mode, with a Brillouin linewidth of 324 MHz). Frequency shift is expressed in GHz on horizontal axis. Vertical axis corresponds to the number of detected photons.

We have measured the v_B for PMMA as an example of an isotropic material, which is shown by the BLS spectrum in Figure 3. The measured Brillouin frequency shift is $v_B = 15.70$ GHz (longitudinal acoustic mode), with a Brillouin linewidth of 324 MHz. The measured spectrum for PMMA (backward scattering) is given in Figure 3. Based on the measured frequency shift v_B , the phase velocity of longitudinal waves in the test material can be derived as,

$$v = \frac{v_B \lambda_0}{2n \sin(\frac{\theta}{2})} \quad (2)$$

With $\lambda_0 = 532$ nm, $n = 1.4953$ for PMMA [77], $\theta = 180^\circ$, the phase velocity of longitudinal waves is derived as $v = 2792.9$ m/s.

Note that for anisotropic materials such as sapphire, the frequency shift peaks will depend on the orientation of the DUT sample to be measured. In this case, it would be useful to check the slowness curves in the wave-vector space, corresponding to the orientation of the sample with respect to the incident direction of laser signal sent to the DUT.

4. Measurement uncertainty of the Brillouin frequency shift

In this section, we aim to estimate the uncertainty for the Brillouin frequency shift measured by the scanning 6-pass TFPI. In the scientific community, it is important to underline that a debate exists as to whether there is a true value of the measurand. Von Clarmann *et al.* offered a critical discussion on the error concept versus the uncertainty concept [78]. Lee *et al.* [79] compared the realist view of true value measurement and its uncertainty versus the instrumentalist view of measurement (quantities are not natural attributes of the world). They have shown that we need to understand the two views, which is critical to follow the Guide to the expression of uncertainty in measurement (GUM) [80].

Estimation of the measurement uncertainty requires careful analysis of the contributions from various error sources. We have followed the modern way of performing the

estimation of uncertainty [81]. We used a similar method like those in optical metrology [82–84], microwave metrology [85,86] and aerosol metrology [87,88] which is based on the GUM delivered by the Bureau International des Poids et Mesures (BIPM) in [80]. The measurement uncertainties consist of several components, which are grouped into two main categories. These relative uncertainty terms have been normalized by the measured Brillouin frequency shift for PMMA (15.70 GHz).

4.1. Contributions evaluated by statistical methods

Following the GUM guidelines, the first category is called type A uncertainties. It corresponds to uncertainty contributions evaluated by statistical methods such as reproducibility and repeatability. The repeatability (denoted by A_1) is used to show the variation in measurements obtained by one person on the same test item, using the same procedure, and under the same conditions (repeated in a short period of time). The repeatability for measured Brillouin frequency shift of 15.70 GHz for the PMMA is estimated to be $A_1 = 6.64 \times 10^{-5}$.

The same operator performed the measurements, with no changes in operator behavior. All components and devices are dedicated to the BLS spectrometer, and none of them is changed. Thus the reproducibility term A_2 can be assumed to be zero.

4.2. Contributions evaluated by other means

The second family of uncertainty contributions is for those assessed by other means. They are called type B uncertainties and depend on various measurement system components and ambient conditions. They are determined by the theoretical calculation, experimental experiences, general knowledge of the behavior, properties of relevant materials or measurement instrument specifications.

Frequency references of the 5 MHz or 10 MHz type ensures the frequency traceability of the BLS measurement system to national metrology standards [89,90]. It is then possible to have the best reference in terms of frequency stability to connect them to additional measuring devices such as oscilloscopes or other means of frequency measurement.

When there is no instrument calibration certificates, we can refer to manufacturer's specifications, calibration data or other certificates, or measurement uncertainty assigned to reference data from handbooks. Such a uncertainty term is denoted by B_R . The BLS spectrometer was not calibrated by other metrology standard, as the method is intrinsic. So the data provided by calibration are not applicable and we can assume the uncertainty term $B_R = 0$.

The frequency resolution of the measurement system depends on the number of samples, i. e. the difference between two measurement frequency points along the horizontal axis. We used 2048 samples in a 0 - 30 GHz span in the BLS spectrum measurement. Thus, we have a frequency interval of 14.66 MHz. The characteristic peak of Brillouin scattering has a Lorentzian distribution [91], also known as a Cauchy distribution (a probability density function). The Lorentzian function versus the frequency shift of optical signal, is given in the following expression :

$$L(f) = \left(\frac{\gamma}{\pi}\right) * \left[\frac{1}{\gamma^2 + (f - f_0)^2}\right] = \left(\frac{1}{\pi\gamma}\right) * \frac{1}{1 + (f - f_0)^2/\gamma^2} \quad (3)$$

where γ is half of the frequency width at half maximum (FWHM): $\gamma = \text{FWHM}/2$, f_0 is the assumed true value for the Brillouin frequency shift. Figure 4 shows the Lorentzian function for a peak in BLS spectrum for PMMA, assuming $f_0 = 15.70$ GHz and $\gamma = 0.16$ GHz.




Figure 4. The Lorentzian function for a peak in BLS spectrum for PMMA, assuming $f_0 = 15.70$ GHz and $\gamma = 0.16$ GHz. The frequency shift is expressed in GHz on horizontal axis.

For a given true value with a peak of great smoothness, we will have 3 points (in the worst configuration), which will allow us to approximate a curve in the form of Lorentzian function. BLS on PMMA shows well-defined peaks for isotropic materials. The maximum relative frequency error caused by the resolution limitation is $(14.66/2)/15700 = 0.00047$. Assuming a rectangular distribution, the standard uncertainty due to frequency resolution is estimated as $BL_0 = 0.00047/\sqrt{3} = 2.7 \times 10^{-4}$. We can see the impact of the resolution on the uncertainty of a measured peak. There is also a risk of not detecting a peak if the sampling frequency is too low.

The uncertainty contribution of the alignment of Torus laser (Laser Quantum) includes mainly the uncertainty due to the geometrical error and the Abbe error [92,93]. According to the manufacturer's datasheet, the laser beam diameter is 1.7 ± 0.2 mm, the pointing direction's stability is less than $2 \mu\text{rad}/^\circ\text{C}$, and the beam angle is less than 1 mrad. There is a geometrical error in the double Fabry-Pérot interferometer, since some cosine error can occur. The laser beam and the axis of displacement are not completely parallel [94,95]. If we denote the angle between the two axis (beam axis and displacement axis) as A , we have an elementary term of error $e_A = L(\cos A - 1) \approx -LA^2/2$ as $A \ll 1$. For a 1 mm distance, A is up to 10^{-4} and the relative error $|e_A/L|$ is up to 5×10^{-9} , which is negligible.

The Abbe error corresponds to the magnification of angular error over distance [92,93]. The relative Brillouin frequency measurement error is proportional to the displacement error in TFPI [96]. The Abbe error is typically estimated to be about 1 nm for a Fabry-Pérot interferometer setup, which does not depend on the displacement [96]. In our BLS spectrometer, the mirror displacement range (scanning range) is up to $2.5 \mu\text{m}$ for the TFP-2. Thus the relative Abbe error (or relative frequency measurement error) is up to $0.001/2.5 = 0.0004$. This elementary term of Abbe error is a dominating term for errors caused by parallelism. Assuming a rectangular distribution for this error, the standard uncertainty in frequency shift due to Abbe error is estimated as $BL_1 = 0.0004/1.732 = 2.31 \times 10^{-4}$.

Contribution of the laser to the noise is denoted as BL_2 . The Relative Intensity Noise (RIN) of laser is defined as the ratio between the average of the square of the fluctuation optical power ($\delta\varphi$) and the square of the average optical power φ_0 :

$$\text{RIN}(\omega) = \langle |\delta\varphi|^2 \rangle / \varphi_0^2 \quad (4)$$

where ω is the angular frequency offset. RIN generally presents a noise floor until the Fourier frequency which is equal to the relaxation frequency of the laser. Beyond that frequency, the RIN decreases. This relaxation frequency is generally in the range of 1 MHz. Using a Fabry-Pérot interferometer (JRS Scientific Instruments), the Torus 532 nm laser (Laser Quantum Ltd.) typically showed high spectral purity with side bands < -110 dB compared with the central mode. This laser is set to operate in normal conditions between 15°C to 35°C . The datasheet of the Torus 532 nm laser indicates a RIN not worse than -

125 dB around the frequency offset of 16 GHz. The RIN noise in the BLS spectrum will cause the peak position to have a small shift of Δf which can be estimated using (3). This relative frequency error is derived to be up to 8.2×10^{-9} . We have estimated the standard uncertainty contribution due to laser's RIN as $BL_2 = 8.2 \times 10^{-9} / 1.732 = 4.74 \times 10^{-9}$, assuming a rectangular distribution.

The uncertainty contribution of photodetector (Hamamatsu H10682-210) in our BLS spectrometer is denoted as BL_3 . The Datasheet of Hamamatsu H10682-210 indicates the specification for photon counting sensitivity is typically 4.6×10^5 and $1.3 \times 10^5 \text{ s}^{-1} \text{ pW}^{-1}$, at wavelengths of 500 nm and 600 nm, respectively (5°C to 40°C). We can assume it does not affect photon detection during BLS measurements. This contribution has negligible effect on the Brillouin frequency shift. Thus the uncertainty contribution $BL_3 \approx 0$.

We have considered the uncertainty contribution of ambient temperature, denoted as BL_4 . Temperature variation in the laboratory is in the range $23 \pm 2^\circ\text{C}$, with the maximum variation of $\pm 2^\circ\text{C}$. It has been shown that a photomultiplier can have variation of 0.33% in detected peak power for 1°C change in temperature [97]. We assume the temperature change is within 1°C during the BLS measurement. Thus the 1°C temperature change has influence on the BLS spectrum, which is estimated to be a fluctuation up to $e_{\text{Temp}} = 10 \times \text{Log}(0.9934) = -0.029 \text{ dB}$. This will cause the BLS peak position to have a relative shift error up to 8.41×10^{-4} . The probability distribution of this error is assumed to be rectangular, and we can derive the standard uncertainty due to ambient temperature variation as $BL_4 = 8.41 \times 10^{-4} / 1.732 = 4.86 \times 10^{-4}$.

There is uncertainty contribution on the laser wavelength which is due to environmental conditions such as ambient pressure and humidity. We denote it as BL_5 . Under normal laboratory measurement conditions, the contribution of small pressure variations and relative humidity remain negligible. Our BLS measurements do not show any dependence on those changes. Thus this uncertainty term BL_5 is considered to be negligible.

The uncertainty contribution due to resolution of power meter is denoted as BL_6 with a rectangular distribution. It is determined by the voltmeter resolution and the value read on each voltmeter for power meter. The maximum relative error in the frequency shift due to voltmeter resolution is estimated be 5×10^{-7} , which is derived assuming a spectrum noise sideband effect (-89.3 dB). Thus the standard uncertainty due to resolution of power meter is derived as $BL_6 = 5 \times 10^{-7} / \sqrt{3} = 2.89 \times 10^{-7}$, assuming a rectangular distribution.

The uncertainty contribution of the linearity error in scanning of Fabry-Pérot Interferometer (denoted as BL_7) is a dominating uncertainty term in Brillouin frequency shift [98,99]. The scan control electronics in the TFPI (for automatic scanning stage control) can produce a linearity error in the mirror spacing which will lead to a frequency shift error. Based on the specification of TFP-2, the linearity error is up to 0.2% [76]. Thus the effect of the linearity error will cause relative frequency shift error up to 2.0×10^{-3} . The standard uncertainty due to linearity error in scanning is derived as $BL_7 = 0.002 / \sqrt{3} = 1.15 \times 10^{-3}$.

We have considered the uncertainty contribution of vibrations from the environment, which is denoted as BL_8 . We do not operate the system if there is known vibration source in the environment. The optical table is robust enough to prevent diffusion of vibration. Pneumatic legs are used to support the optical table. The TFPI spectrometer's operation is only isolated against building's vibrations and not against vibrations introduced directly into the table. We have operated the TFPI spectrometer in safe conditions and avoided any potential vibration due to components on the table. Therefore, this uncertainty term can be assumed to be negligible.

4.3. Estimation of the expanded measurement uncertainty

All of the uncertainty terms in measured Brillouin frequency shift have been given in table 1. The expanded uncertainty for the measured frequency shift with approximately 95% confidence (coverage factor $k = 2$) is calculated as follows:

$$U_m = 2 \sqrt{A_1^2 + A_2^2 + B_R^2 + \sum_i (BL_i)^2} \quad (5)$$

From (5), the expanded relative uncertainty of Brillouin frequency shift measurement is estimated to be $U_m = 0.26\%$. For a Brillouin frequency shift measured at 15.70 GHz for PMMA, the expanded frequency measurement uncertainty is 41 MHz.

Table 1 Uncertainty budget table for Brillouin frequency shift measurement using our scanning 6-pass TFPI spectrometer

Uncertainty component	Probability distribution	Standard relative uncertainty contribution
Repeatability	t-distribution	6.64E-05
Reproducibility	t-distribution	negligible
Frequency resolution	rectangular	2.69E-04
Geometrical error in the BLS spectrometer	rectangular	negligible
Abbe error in BLS spectrometer	rectangular	2.31E-07
Relative Intensity Noise (RIN) of laser	rectangular	4.74E-09
Photodetector's counting sensitivity	rectangular	negligible
Ambient temperature variation	rectangular	4.86E-04
Ambient pressure and humidity variation	rectangular	negligible
Resolution of voltmeter for power meter	rectangular	2.89E-07
Linearity error in scanning of TFPI	rectangular	1.15E-03
Vibration effect	rectangular	negligible
Combined relative uncertainty for frequency shift measurement		0.13%
Expanded relative uncertainty for frequency shift measurement ($k = 2$)		0.26%

The corresponding propagation speed of the longitudinal phononic wave in the PMMA can be estimated using the BLS spectrometer. Based on $v = v_B \lambda / (2n)$, the measured Brillouin frequency shift (15.70 GHz) and the refractive index of PMMA ($n=1.4953$) [77], the longitudinal phononic wave speed in PMMA can be derived as $v = 2792.9$ m/s with an expanded uncertainty of 12.0 m/s.

5. Comparison with calculated Brillouin frequency shift

5.1. Uncertainty of the calculated frequency shift

For BLS measurement of PMMA, the shift frequency has been theoretically calculated using $\nu_c = \frac{2}{\lambda_0} \frac{n}{v} \sin(\frac{\theta}{2})$ to be $\nu_c = 15.44$ GHz, assuming the following parameters: $n=1.4953$ (optical refractive index of PMMA [77]), $\lambda_0 = 532$ nm (laser wavelength), $\theta = 180^\circ$ (scattering angle), longitudinal phononic wave speed $v = 2746.3$ m/s (based on measurements reported in [100]).

Considering the mean bulk density of PMMA $\rho = 1180$ kg/m³ (with maximum error of ± 20 kg/m³ [101]) and the longitudinal phononic wave speed ($v = 2746.3$ m/s) as measured in [100], the longitudinal modulus of PMMA can be estimated to be $C_{11} = 8.8$ GPa.

The longitudinal ultrasonic-velocity in PMMA was measured with a repeatability of $u_1 = 0.31\%$ based on data of [100], which contributes to the uncertainty in calculated frequency shift (as the repeatability in Brillouin frequency shift). The longitudinal ultrasonic velocity measurements in PMMA showed up to 1.74% difference when using single-around or pulse-echo method, as reported in [100]. This would lead to a standard uncertainty contribution to calculated shift frequency of $u_2 = 1.74\% / \sqrt{3} = 1.0\%$ (due to variability in speed measurement). The refractive index of PMMA has maximum error of 0.3% based on the various measurement methods and data given in [102], which leads to a standard uncertainty of $u_3 = 0.3\% / \sqrt{3} = 0.17\%$ in the calculated shift frequency.

Considering these three uncertainty factors, the expanded relative uncertainty of calculated Brillouin frequency shift (based on material mechanical properties) is estimated to be

$$U_c = 2\sqrt{\sum_i (u_i)^2} \quad (6)$$

As shown in Table 2, $U_c = 2.13\%$ (coverage factor $k = 2$).

Table 2 Uncertainty budget table for calculated Brillouin frequency shift (using mechanical property of PMMA)

Uncertainty component	Probability distribution	Standard relative uncertainty contribution
Repeatability	t-distribution	0.31%
Refractive index of PMMA	rectangular	0.17%
Ultrasonic velocity measurement variation	rectangular	1.00%
Combined relative uncertainty for calculated frequency shift		1.06%
Expanded relative uncertainty for calculated frequency shift ($k=2$)		2.13%

5.2 Comparison of measured and calculated Brillouin frequency shift

The measured Brillouin frequency shift by the TFPI ($\nu_m = 15.70$ GHz) and calculated Brillouin frequency shift ($\nu_c = 15.44$ GHz) for PMMA has a deviation of 0.26 GHz. To validate the estimated uncertainty for measured Brillouin frequency shift and compare this two frequency shift values, the normalised error E_n [103] is derived as

$$E_n = \frac{|\nu_m - \nu_c|}{\sqrt{\nu_m^2 U_m^2 + \nu_c^2 U_c^2}} \quad (7)$$

where U_m and U_c is the expanded relative uncertainty of measured and calculated Brillouin frequency shift, respectively. This formula has been commonly used in proficiency testing or inter-laboratory comparison.

Table 3 Comparison of the measured Brillouin frequency shift (ν_m) and calculated Brillouin frequency shift (ν_c) for PMMA

Measured frequency shift, GHz	Expanded uncertainty of measured frequency shift, ($\nu_m U_m$), GHz	Calculated frequency shift, GHz	Expanded uncertainty of calculated frequency shift ($\nu_c U_c$), GHz	Deviation of frequency shift, GHz	E_n
15.70	0.041	15.44	0.33	0.26	0.8

The $E_n = 0.8$ as shown in Table 3. Given the $E_n < 1$, we can conclude that the measured Brillouin frequency shift (ν_m) and calculated Brillouin frequency shift (ν_c) for PMMA agree within their expanded uncertainties. This checking using E_n derived by (7) serves to validate the expanded uncertainty we have derived for the Brillouin frequency shift measurement in section 4.

5.3 Uncertainty in the longitudinal modulus derived by BLS spectrometer

The BLS spectrometer can be used to measure the longitudinal modulus of material using

$$c_{11} = \rho v^2 = \rho \left(\frac{v_B \lambda_0}{2n} \right)^2 \quad (8)$$

where v_B is the measured frequency shift. Considering the standard uncertainty of 0.85% in the density of PMMA, the standard uncertainty of 0.17% in refractive index of PMMA and the standard uncertainty of 0.13% in Brillouin frequency shift measurement, we have estimated the expanded uncertainty in the derived longitudinal modulus c_{11} to be 1.9% (coverage factor $k=2$) using (8).

The longitudinal modulus of material can also be estimated using the ultrasonic-velocity based method via $c_{11} = \rho v^2$, where v is the measured ultrasonic-velocity in the material. The standard uncertainty contribution from the longitudinal ultrasonic-velocity measurement repeatability (for PMMA) is estimated to be 0.62%. The standard uncertainty in density of PMMA can be estimated as 0.85% and the standard uncertainty due to ultrasonic velocity measurement variability has been estimated to be 1.0%. Thus the expanded uncertainty in c_{11} using the ultrasonic-velocity based method is estimated to be 2.9% (coverage factor $k=2$).

Comparing the expanded uncertainty in estimation of longitudinal modulus of material using the two methods, the BLS spectrometer based method has a smaller expanded uncertainty of 1.9%. Thus the Brillouin frequency shift can be measured to derive the longitudinal modulus of material with higher accuracy, as compared to the ultrasonic-velocity based method.

6. Discussion and Conclusion

The Brillouin spectroscopy is a non-intrusive measurement method for bulk materials and thin films. A scanning 6-pass TFPI has been described for BLS measurement. Following the GUM, we have made detailed analysis and estimation of the uncertainties in the Brillouin frequency shift measurement which is related to the speed of propagation of phononic waves in bulk materials. The expanded relative uncertainty in measured Brillouin frequency shift is estimated to be 0.26% (coverage factor $k=2$), which corresponds to an expanded uncertainty of 41 MHz for the measured frequency shift of 15.70 GHz in testing PMMA.

We have also estimated the expanded relative uncertainty of calculated Brillouin frequency shift (at 15.44 GHz based on PMMA's mechanical properties) to be 2.13% ($k=2$). It is shown that the measured and calculated Brillouin frequency shift for PMMA agree within their expanded uncertainties. The scanning 6-pass TFPI can be used to measure the longitudinal modulus of material with an expanded uncertainty of 1.9%, smaller than that of the ultrasonic-velocity based method (estimated to be 2.9%). In our future work, we will conduct uncertainty analysis for PMMA's Brillouin linewidth measurement using the 6-pass TFPI and assess the sample's temperature stabilization effect.

Although the scanning TFPI has high accuracy in Brillouin frequency and linewidth measurement, it has limitations in the complex system design, high cost and slow acquisition time. Hence, a scanning TFPI is not suitable for biomedical imaging or probing viscoelastic property of biological samples. The VIPA-based spectrometers with very fast acquisition time have been developed for biomedical imaging, but such spontaneous Brillouin scattering based spectrometers still face the issue of high Rayleigh scattering background which makes it difficult to detect the weak Brillouin signals. An impulsive SBS method [72] had been proposed to further reduce the acquisition time, reduce the standard deviations in BLS measurement and improve the spectral resolution in 2D biomedical imaging. However, the impulsive SBS microscopy still need to improve its data acquisition speed (using a detector array), detection sensitivity, spatial resolution, etc.

The SBS spectroscopy can provide high-intensity signals to improve the signal-to-noise ratio (SNR) in BLS measurement. One disadvantage of this method is that the high optical power could cause phototoxicity or thermal damage to biological samples. Therefore, the overall illumination dosage need to be controlled. The detection limit of the frequency-domain or impulsive SBS spectroscopy need to be further improved.

Quantum-correlated light (squeezed light) can be used to squeeze the amplitude noise below the vacuum-state (shot-noise) limit [104–108]. The quantum-enhanced sensing with single-mode or two-mode squeezing has been used to improve the SNR in gravitational wave detection, Raman spectroscopy, saturation spectroscopy, Raman microscopy, microparticle tracking, etc [109–115]. Li *et al.* [116] have shown that two-mode intensity-difference squeezed light (generated by the four-wave mixing process in atomic ^{85}Rb vapor) can be used to improve the SNR of SBS spectroscopy by 3.4 dB. This quantum-enhanced SBS spectroscopy can still measure the Brillouin frequency and linewidth of water with good accuracy, when the optical pump power was reduced to 7.5 mW. The quantum squeezed light has a narrow spectral width in the range of 10 MHz, which enables the improvement of SNR in the SBS spectroscopy. Quantum sensing using squeezed light is expected to further improve the quantum noise reduction in the future via reduction of optical loss in the sensors, novel detection techniques for quantum squeezing, and cross-correlation measurement for detecting a squeezed state [106,117–120].

Future metrology works could study how to obtain corrected Brillouin spectrum which represents the scattering intensity and linewidth of a given sample as a function of Brillouin frequency shift and temperature. Calibration standards would be needed to check the reproducibility of observed Brillouin frequency shift and linewidth. Luminescent intensity reference standards would also be needed to calibrate various Brillouin spectrometers and make correction for instrument response variation across a Brillouin spectrum for a specified temperature range.

Authors' Contributions: Conceptualisation and measurements: P.S.; manuscript writing: T.W., P.S.; uncertainty evaluation, T.W., P.S.; review and revision of manuscript: T.W. supervision, P.S.; project administration and funding acquisition, P.S. All authors have read and agreed to the published version of the manuscript.

Funding: This work has been supported by the French Investissements d'Avenir program, project ISITE-BFC (contract ANR-15-IDEX-03), and funded by Région Bourgogne Franche-Comté.

Acknowledgments: The authors would like to thank their colleagues A. Mosset and V. Laude.

Conflicts of Interest: The authors declare no conflict of interest.

References

- Brillouin, L. Diffusion de La Lumière et Des Rayons X Par Un Corps Transparent Homogène. *Ann. Phys. (Paris)*. **1922**, *9*, 88–122, doi:10.1051/anphys/192209170088.
- Mandelstam, L. Light Scattering by Inhomogeneous Media. *Zh. Russ. Fiz. Khim. Ova* **58**, 381.
- Blachowicz, T.; Grimsditch, M. Scattering, Inelastic: Brillouin. *Encycl. Condens. Matter Phys.* **2005**, 199–205, doi:10.1016/B0-12-369401-9/00643-4.
- Curie, J.; Curie, P. Développement Par Compression de l'électricité Polaire Dans Les Cristaux Hémiedres à Faces Inclinées. *Bull. la Société minéralogique Fr.* **1880**, *3*, 90–93, doi:10.3406/bulmi.1880.1564.
- Gross, E. Change of Wave-Length of Light Due to Elastic Heat Waves at Scattering in Liquids. *Nature* **1930**, *126*, 201–202.
- Gross, E. The Splitting of Spectral Lines at Scattering of Light by Liquids. *Nature* **1930**, *126*, 400.
- Gross, E. Über Änderung Der Wellenlänge Bei Lichtzerstreuung in Kristallen. *Zeitschrift für Phys.* **1930**, *63*, 685–687.
- Gross, E. Modification of Light Quanta by Elastic Heat Oscillations in Scattering Media. *Nature* **1932**, *129*, 722–723.
- Raghavendra Rao, B. V. Examination of Molecularly Scattered Light with a Fabry-Perot Etalon - Part I. Liquid Benzene.

- Proc. Indian Acad. Sci. - Sect. A* **1934**, *1*, 261–268, doi:10.1007/BF03035565. 524
10. Raghavendra Rao, B. V. Examination of Molecularly Scattered Light with a Fabry-Perot Etalon - Part II. Liquids: Toluene and Carbon Tetrachloride. *Proc. Indian Acad. Sci. - Sect. A* **1935**, *1*, 473–483, doi:10.1007/BF03035595/METRICS. 525
526
 11. Grimsditch, M.H.; Ramdas, A.K. Brillouin Scattering in Diamond. *Phys. Rev. B* **1975**, *11*, 3139–3148, doi:10.1103/PhysRevB.11.3139. 527
528
 12. Kojima, S. 100th Anniversary of Brillouin Scattering: Impact on Materials Science. *Materials (Basel)*. **2022**, *15*, 3518, doi:10.3390/ma15103518. 529
530
 13. Borovik-Romanov, A.S.; Kreines, N.M. Brillouin-Mandelstam Scattering from Thermal and Excited Magnons. *Phys. Rep.* **1982**, *81*, 351–408, doi:10.1016/0370-1573(82)90118-1. 531
532
 14. Kargar, F.; Balandin, A.A. Advances in Brillouin–Mandelstam Light-Scattering Spectroscopy. *Nat. Photonics* **2021**, *15*, 720–731, doi:10.1038/s41566-021-00836-5. 533
534
 15. Merklein, M.; Kabakova, I. V.; Zarifi, A.; Eggleton, B.J. 100 Years of Brillouin Scattering: Historical and Future Perspectives. *Appl. Phys. Rev.* **2022**, *9*, doi:10.1063/5.0095488. 535
536
 16. Singaraju, A.B.; Bahl, D.; Stevens, L.L. Brillouin Light Scattering: Development of a Near Century-Old Technique for Characterizing the Mechanical Properties of Materials. *AAPS PharmSciTech* **2019**, *20*, 1–16, doi:10.1208/s12249-019-1311-5. 537
538
 17. Palombo, F.; Fioretto, D. Brillouin Light Scattering: Applications in Biomedical Sciences. *Chem. Rev.* **2019**, *119*, 7833–7847, doi:10.1021/acs.chemrev.9b00019. 539
540
 18. Prevedel, R.; Diz-Muñoz, A.; Ruocco, G.; Antonacci, G. Brillouin Microscopy: An Emerging Tool for Mechanobiology. *Nat. Methods* **2019**, *16*, 969–977, doi:10.1038/s41592-019-0543-3. 541
542
 19. Meng, Z.; Traverso, A.J.; Ballmann, C.W.; Troyanova-Wood, M.A.; Yakovlev, V. V Seeing Cells in a New Light: A Renaissance of Brillouin Spectroscopy. *Adv. Opt. Photon.* **2016**, *8*, 300–327, doi:10.1364/AOP.8.000300. 543
544
 20. Antonacci, G.; Beck, T.; Bilenca, A.; Czarske, J.; Elsayad, K.; Guck, J.; Kim, K.; Krug, B.; Palombo, F.; Prevedel, R.; et al. Recent Progress and Current Opinions in Brillouin Microscopy for Life Science Applications. *Biophys. Rev.* **2020**, *12*, 615–624, doi:10.1007/s12551-020-00701-9. 545
546
547
 21. Chitanvis, S.M.; Cantrell, C.D. Simple Approach to Stimulated Brillouin Scattering in Glass Aerosols. *JOSA B, Vol. 6, Issue 7, pp. 1326–1331* **1989**, *6*, 1326–1331, doi:10.1364/JOSAB.6.001326. 548
549
 22. Boyarchuk, K.A.; Lyakhov, G.A.; Svirko, Y.P. Microwave Driven Anti-Stokes Brillouin Scattering in an Ionized Medium. The Method of Remote Sensing of Atmospheric Aerosols. *J. Phys. D: Appl. Phys.* **1993**, *26*, 1561–1565, doi:10.1088/0022-3727/26/10/004. 550
551
552
 23. Hair, J.W.; Hostetler, C.A.; Cook, A.L.; Harper, D.B.; Ferrare, R.A.; Mack, T.L.; Welch, W.; Izquierdo, L.R.; Hovis, F.E. Airborne High Spectral Resolution Lidar for Profiling Aerosol Optical Properties. *Appl. Opt.* **2008**, *47*, 6734–6753, doi:10.1364/AO.47.006734. 553
554
555
 24. Esselborn, M.; Wirth, M.; Fix, A.; Tesche, M.; Ehret, G. Airborne High Spectral Resolution Lidar for Measuring Aerosol Extinction and Backscatter Coefficients. *Appl. Opt.* **2008**, *47*, 346–358, doi:10.1364/AO.47.000346. 556
557
 25. Gu, Z.; Witschas, B.; Van De Water, W.; Ubachs, W. Rayleigh-Brillouin Scattering Profiles of Air at Different Temperatures and Pressures. *Appl. Opt.* **2013**, *52*, 4640–4651, doi:10.1364/AO.52.004640. 558
559
 26. Witschas, B. Light Scattering on Molecules in the Atmosphere. In *Atmospheric Physics, Research Topics in Aerospace*; Springer-Verlag Berlin Heidelberg, 2012; pp. 69–83. 560
561
 27. Gu, Z.; Witschas, B.; Ubachs, W. Temperature Retrieval from Rayleigh-Brillouin Scattering Profiles Measured in Air. *Opt. Express, Vol. 22, Issue 24, pp. 29655–29667* **2014**, *22*, 29655–29667, doi:10.1364/OE.22.029655. 562
563
 28. Rank, D.H.; Wiggins, T.A.; Wick, R. V.; Eastman, D.P.; Guenther, A.H. Stimulated Brillouin Effect in High-Pressure Gases*. *J. Opt. Soc. Am.* **1966**, *56*, 174, doi:10.1364/josa.56.000174. 564
565
 29. Raman, C. V.; Krishnan, K.S. A New Type of Secondary Radiation [11]. *Nature* **1928**, *121*, 501–502, doi:10.1038/121501c0. 566
 30. Raman, C.V. A New Radiation. *Indian J. Phys.* **1928**, *2*, 387–398. 567
 31. Hendra, P.J.; Stratton, P.M. Laser-Raman Spectroscopy. *Chem. Rev.* **1969**, *69*, 325–344. 568
 32. Sandercock, J.R. Some Recent Developments in Brillouin Scattering. *R.C.A. Rev.* **1975**, *36*, 89–107. 569

33. Sandercock, J.R. Brillouin Scattering Study of SbSI Using a Double-Passed, Stabilised Scanning Interferometer. *Opt. Commun.* **1970**, *2*, 73–76, doi:10.1016/0030-4018(70)90047-7. 570
571
34. Sandercock, J.R. Light Scattering From Thermally Excited Surface Phonons and Magnons. In Proceedings of the Proc. 7th Int. Conf on Raman Spectroscopy; Murphy, W.F., Ed.; North-Holland: Amsterdam, 1980; pp. 364–367. 572
573
35. Sandercock, J.R. Trends in Brillouin Scattering: Studies of Opaque Materials, Supported Films, and Central Modes. In *Topics in Applied Physics*; Springer: Berlin, 1982; Vol. 51, pp. 173–206. 574
575
36. Vacher, R.; Boyer, L.; Boissier, M. Measurement of the Elastic Constants of Lithium Acetate by Means of the Brillouin Effect. *Phys. Rev. B* **1972**, *6*, 674, doi:10.1103/PhysRevB.6.674. 576
577
37. Sussner, H.; Vacher, R. High-Precision Measurements of Brillouin Scattering Frequencies. *Appl. Opt.* **1979**, *18*, 3815, doi:10.1364/ao.18.003815. 578
579
38. Vacher, R.; Sussner, H.; Schickfus, M. V. A Fully Stabilized Brillouin Spectrometer with High Contrast and High Resolution. *Rev. Sci. Instrum.* **1980**, *51*, 288–291, doi:10.1063/1.1136202. 580
581
39. Traverso, A.J.; Thompson, J. V.; Steelman, Z.A.; Meng, Z.; Scully, M.O.; Yakovlev, V. V. Dual Raman-Brillouin Microscope for Chemical and Mechanical Characterization and Imaging. *Anal. Chem.* **2015**, *87*, 7519–7523, doi:10.1021/acs.analchem.5b02104. 582
583
584
40. Mattana, S.; Alunni Cardinali, M.; Caponi, S.; Casagrande Pierantoni, D.; Corte, L.; Roscini, L.; Cardinali, G.; Fioretto, D. High-Contrast Brillouin and Raman Micro-Spectroscopy for Simultaneous Mechanical and Chemical Investigation of Microbial Biofilms. *Biophys. Chem.* **2017**, *229*, 123–129, doi:10.1016/j.bpc.2017.06.008. 585
586
587
41. Meng, Z.; Bustamante Lopez, S.C.; Meissner, K.E.; Yakovlev, V. V. Subcellular Measurements of Mechanical and Chemical Properties Using Dual Raman-Brillouin Microspectroscopy. *J. Biophotonics* **2016**, *9*, 201–207, doi:10.1002/jbio.201500163. 588
589
42. Scarponi, F.; Mattana, S.; Corezzi, S.; Caponi, S.; Comez, L.; Sassi, P.; Morresi, A.; Paolantoni, M.; Urbanelli, L.; Emiliani, C.; et al. High-Performance Versatile Setup for Simultaneous Brillouin-Raman Microspectroscopy. *Phys. Rev. X* **2017**, *7*, 31015, doi:10.1103/PhysRevX.7.031015. 590
591
592
43. Aschauer, R.; Asenbaum, A.; Geri, H. Fabry-Perot Interferometer with Personal Computer Control. *Appl. Opt.* **1990**, *29*, 953, doi:10.1364/ao.29.000953. 593
594
44. Błachowicz, T.; Bukowski, R.; Kleszczewski, Z. Fabry-Perot Interferometer in Brillouin Scattering Experiments. *Rev. Sci. Instrum.* **1996**, *67*, 4057–4060, doi:10.1063/1.1147550. 595
596
45. Hernandez, G. *Fabry-Perot Interferometers*; Cambridge University Press, 1988; 597
46. Perot, A.; Fabry, C. On the Application of Interference Phenomena to the Solution of Various Problems of Spectroscopy and Metrology. *Astrophys. J.* **1899**, *9*, 87, doi:10.1086/140557. 598
599
47. Fabry, C.; Pérot, A. Théorie et Applications d'une Nouvelle Méthode de Spectroscopie Interférentielle. *Ann. Chim. Phys.* **1899**, *16*, 115–144. 600
601
48. Vacher, R.; Boyer, L.; Boissier, M. Measurement of the Elastic Constants of Lithium Acetate by Means of the Brillouin Effect. *Phys. Rev. B* **1972**, *6*, 674, doi:10.1103/PhysRevB.6.674. 602
603
49. Heiman, D.; Hamilton, D.S.; Hellwarth, R.W. Brillouin Scattering Measurements on Optical Glasses. *Phys. Rev. B* **1979**, *19*, 6583–6592, doi:10.1103/PhysRevB.19.6583. 604
605
50. Asenbaum, A. Computer-Controlled Fabry-Perot Interferometer for Brillouin Spectroscopy. *Appl. Opt.* **1979**, *18*, 540, doi:10.1364/ao.18.000540. 606
607
51. Salvato, G.; Ponterio, R.C.; Aliotta, F. New Automatic System for Multipass Fabry-Prot Alignment and Stabilization. *Rev. Sci. Instrum.* **2006**, *77*, doi:10.1063/1.2369641. 608
609
52. Błachowicz, T. The Use of Pressure Controlled Fabry-Pérot Interferometer with Linear Scanning of Data for Brillouin-Type Experiments. *Rev. Sci. Instrum.* **2000**, *71*, 2988–2991, doi:10.1063/1.1304872. 610
611
53. Lindsay, S.M.; Anderson, M.W.; Sandercock, J.R. Construction and Alignment of a High Performance Multipass Vernier Tandem Fabry-Perot Interferometer. *Rev. Sci. Instrum.* **1981**, *52*, 1478–1486, doi:10.1063/1.1136479. 612
613
54. Dil, J.G.; van Hijningen, N.C.J.A.; van Dorst, F.; Aarts, R.M. Tandem Multipass Fabry-Perot Interferometer for Brillouin Scattering. *Appl. Opt.* **1981**, *20*, 1374, doi:10.1364/ao.20.001374. 614
615
55. Mock, R.; Hillebrands, B.; Sandercock, R. Construction and Performance of a Brillouin Scattering Set-up Using a Triple-Pass 616

- Tandem Fabry-Perot Interferometer. *J. Phys. E.* **1987**, *20*, 656, doi:10.1088/0022-3735/20/6/017. 617
56. Hillebrands, B. Progress in Multipass Tandem Fabry-Perot Interferometry: I. A Fully Automated, Easy to Use, Self-Aligning Spectrometer with Increased Stability and Flexibility. *Rev. Sci. Instrum.* **1999**, *70*, 1589–1598, doi:10.1063/1.1149637. 618
57. Itoh, S.I.; Yamana, T.; Kojima, S. Quick Measurement of Brillouin Spectra of Glass-Forming Material Trimethylene Glycol by Angular Dispersion-Type Fabry-Perot Interferometer System. *Japanese J. Appl. Physics, Part 1 Regul. Pap. Short Notes Rev. Pap.* **1996**, *35*, 2879–2881, doi:10.1143/jjap.35.2879. 620
58. Ko, J.H.; Kojima, S. Nonscanning Brillouin Spectroscopy Applied to Solid Materials. *Rev. Sci. Instrum.* **2002**, *73*, 4390–4392, doi:10.1063/1.1516847. 621
59. Ike, Y.; Tsukada, S.; Kojima, S. High-Resolution Brillouin Spectroscopy with Angular Dispersion-Type Fabry-Perot Interferometer and Its Application to a Quartz Crystal. *Rev. Sci. Instrum.* **2007**, *78*, 3–5, doi:10.1063/1.2753593. 622
60. Shirasaki, M. Large Angular Dispersion by a Virtually Imaged Phased Array and Its Application to a Wavelength Demultiplexer. *Opt. Lett.* **1996**, *21*, 366, doi:10.1364/ol.21.000366. 623
61. Scarcelli, G.; Yun, S.H. Confocal Brillouin Microscopy for Three-Dimensional Mechanical Imaging. *Nat. Photonics* **2008**, *2*, 39–43, doi:10.1038/nphoton.2007.250. 624
62. Scarcelli, G.; Yun, S.H. Multistage VIPA Etalons for High-Extinction Parallel Brillouin Spectroscopy. *Opt. Express* **2011**, *19*, 10913–10922, doi:10.1364/OE.19.010913. 625
63. Meng, Z.; Traverso, A.J.; Yakovlev, V. V. Background Clean-up in Brillouin Microspectroscopy of Scattering Medium. *Opt. Express* **2014**, *22*, 5410, doi:10.1364/oe.22.005410. 626
64. Meng, Z.; Yakovlev, V. V. Precise Determination of Brillouin Scattering Spectrum Using a Virtually Imaged Phase Array (VIPA) Spectrometer and Charge-Coupled Device (CCD) Camera. *Appl. Spectrosc.* **2016**, *70*, 1356–1363, doi:10.1177/0003702816654050. 627
65. Coker, Z.; Troyanova-Wood, M.; Traverso, A.J.; Yakupov, T.; Utegulov, Z.N.; Yakovlev, V. V. Assessing Performance of Modern Brillouin Spectrometers. *Opt. Express* **2018**, *26*, 2400, doi:10.1364/oe.26.002400. 628
66. Cotter, D. Stimulated Brillouin Scattering in Monomode Optical Fiber. *J. Opt. Commun.* **1983**, *4*, 10–19, doi:10.1515/JOC.1983.4.1.10. 629
67. Garmire, E. Stimulated Brillouin Review: Invented 50 Years Ago and Applied Today. *Int. J. Opt.* **2018**, *2018*, doi:10.1155/2018/2459501. 630
68. Bai, Z.; Yuan, H.; Liu, Z.; Xu, P.; Gao, Q.; Williams, R.J.; Kitzler, O.; Mildren, R.P.; Wang, Y.; Lu, Z. Stimulated Brillouin Scattering Materials, Experimental Design and Applications: A Review. *Opt. Mater. (Amst.)* **2018**, *75*, 626–645, doi:10.1016/j.OPTMAT.2017.10.035. 631
69. Agrawal, G.P. Stimulated Brillouin Scattering. In *Nonlinear Fiber Optics*; 2019; pp. 355–399. 632
70. Ballmann, C.W.; Thompson, J. V.; Traverso, A.J.; Meng, Z.; Scully, M.O.; Yakovlev, V. V. Stimulated Brillouin Scattering Microscopic Imaging. *Sci. Rep.* **2015**, *5*, 1–10, doi:10.1038/srep18139. 633
71. Krug, B.; Koukourakis, N.; Czarske, J.W. Impulsive Stimulated Brillouin Microscopy for Non-Contact, Fast Mechanical Investigations of Hydrogels. *Opt. Express* **2019**, *27*, 26910, doi:10.1364/oe.27.026910. 634
72. Ballmann, C.W.; Meng, Z.; Traverso, A.J.; Scully, M.O.; Yakovlev, V. V. Impulsive Brillouin Microscopy. *Optica* **2017**, *4*, 124, doi:10.1364/optica.4.000124. 635
73. Remer, I.; Shaashoua, R.; Shemesh, N.; Ben-Zvi, A.; Bilencia, A. High-Sensitivity and High-Specificity Biomechanical Imaging by Stimulated Brillouin Scattering Microscopy. *Nat. Methods* **2020**, *17*, 913–916, doi:10.1038/s41592-020-0882-0. 636
74. Ballmann, C.W.; Meng, Z.; Yakovlev, V. V. Nonlinear Brillouin Spectroscopy: What Makes It a Better Tool for Biological Viscoelastic Measurements. *Biomed. Opt. Express* **2019**, *10*, 1750, doi:10.1364/boe.10.001750. 637
75. JRS Brillouin Scattering by Means of the JRS TFP-1 Tandem Multi-Pass Fabry-Pérot Interferometer; JRS Scientific Instruments; 638
76. TableStable Tandem Fabry-Pérot Spectrometers TFP-1 and TFP-2 HC (Operator Manual); The Table Stable Ltd.; 639
77. Beadie, G.; Brindza, M.; Flynn, R.A.; Rosenberg, A.; Shirk, J.S. Refractive Index Measurements of Poly(Methyl Methacrylate) (PMMA) from 0.4–1.6 Mm. *Appl. Opt.* **2015**, *54*, 139–143. 640
78. Von Clarmann, T.; Compennolle, S.; Hase, F. Truth and Uncertainty. A Critical Discussion of the Error Concept versus the Uncertainty Concept. *Atmos. Meas. Tech.* **2022**, *15*, 1145–1157, doi:10.5194/amt-15-1145-2022. 641

79. Lee, J.W.; Hwang, E.; Kacker, R.N. True Value, Error, and Measurement Uncertainty: Two Views. *Accredit. Qual. Assur.* **2022**, *27*, 235–242, doi:10.1007/s00769-022-01508-9. 664
665
80. ISO ISO/IEC Guide 98-3:2008 - Uncertainty of Measurement — Part 3: Guide to the Expression of Uncertainty in Measurement (GUM:1995); 2008; 666
667
81. Kacker, R.; Sommer, K.D.; Kessel, R. Evolution of Modern Approaches to Express Uncertainty in Measurement. *Metrologia* **2007**, *44*, 513–529, doi:10.1088/0026-1394/44/6/011. 668
669
82. Salzenstein, P.; Pavlyuchenko, E.; Hmima, A.; Cholley, N.; Zarubin, M.; Galliou, S.; Chembo, Y.K.; Larger, L. Estimation of the Uncertainty for a Phase Noise Optoelectronic Metrology System. *Phys. Scr.* **2012**, *T149*, 14025, doi:10.1088/0031-8949/2012/T149/014025. 670
671
672
83. Pavlyuchenko, E.; Salzenstein, P. Application of Modern Method of Calculating Uncertainty to Microwaves and Opto-Electronics. *Proc. - 2014 Int. Conf. Laser Opt. LO 2014* **2014**, 6886449, doi:10.1109/LO.2014.6886449. 673
674
84. Salzenstein, P.; Pavlyuchenko, E. Uncertainty Evaluation on a 10.52 Ghz (5 Dbm) Optoelectronic Oscillator Phase Noise Performance. *Micromachines* **2021**, *12*, 474, doi:10.3390/mi12050474. 675
676
85. Lee, W.K.; Yu, D.H.; Park, C.Y.; Mun, J. The Uncertainty Associated with the Weighted Mean Frequency of a Phase-Stabilized Signal with White Phase Noise. *Metrologia* **2010**, *47*, 24–32, doi:10.1088/0026-1394/47/1/004. 677
678
86. Salzenstein, P.; Wu, T.Y. Uncertainty Analysis for a Phase-Detector Based Phase Noise Measurement System. *Meas. J. Int. Meas. Confed.* **2016**, *85*, doi:10.1016/j.measurement.2016.02.026. 679
680
87. Wu, T.Y.; Murashima, Y.; Sakurai, H.; Iida, K. A Bilateral Comparison of Particle Number Concentration Standards via Calibration of an Optical Particle Counter for Number Concentration up to ~1000 Cm⁻³. *Measurement* **2022**, *189*, 110446, doi:10.1016/j.measurement.2021.110446. 681
682
683
88. Wu, T.Y.; Horender, S.; Tancev, G.; Vasilatou, K. Evaluation of Aerosol-Spectrometer Based PM2.5 and PM10 Mass Concentration Measurement Using Ambient-like Model Aerosols in the Laboratory. *Measurement* **2022**, *201*, 111761, doi:10.1016/j.measurement.2022.111761. 684
685
686
89. Salzenstein, P.; Kuna, A.; Sojdr, L.; Chauvin, J. Significant Step in Ultra-High Stability Quartz Crystal Oscillators. *Electron. Lett.* **2010**, *46*, 1433–1434, doi:10.1049/el.2010.1828. 687
688
90. Salzenstein, P.; Cholley, N.; Kuna, A.; Abbé, P.; Lardet-Vieudrin, F.; Šojdr, L.; Chauvin, J. Distributed Amplified Ultra-Stable Signal Quartz Oscillator Based. *Meas. J. Int. Meas. Confed.* **2012**, *45*, 1937–1939, doi:10.1016/j.measurement.2012.03.035. 689
690
91. Gough, W. The Graphical Analysis of a Lorentzian Function and a Differentiated Lorentzian Function. *J. Phys. A Gen. Phys.* **1968**, *1*, 704–709, doi:10.1088/0305-4470/1/6/309. 691
692
92. Köning, R.; Flügge, J.; Bosse, H. A Method for the in Situ Determination of Abbe Errors and Their Correction. *Meas. Sci. Technol.* **2007**, *18*, 476–481, doi:10.1088/0957-0233/18/2/S21. 693
694
93. Leach, R. Abbe Error/Offset. *CIRP Encycl. Prod. Eng.* **2014**, 1–4, doi:10.1007/978-3-642-35950-7_16793-1. 695
94. Howard, L.; Stone, J.; Fu, J. Real-Time Displacement Measurements with a Fabry-Perot Cavity and a Diode Laser. *Precis. Eng.* **2001**, *25*, 321–335, doi:10.1016/S0141-6359(01)00086-1. 696
697
95. Joo, K.N.; Ellis, J.D.; Spronck, J.W.; Munnig Schmidt, R.H. Design of a Folded, Multi-Pass Fabry-Perot Cavity for Displacement Metrology. *Meas. Sci. Technol.* **2009**, *20*, 107001, doi:10.1088/0957-0233/20/10/107001. 698
699
96. Zhu, M.; Wei, H.; Wu, X.; Li, Y. Fabry – Perot Interferometer with Picometer Resolution Referenced to an Optical Frequency Comb. *Opt. Lasers Eng.* **2015**, *67*, 128–134, doi:10.1016/j.optlaseng.2014.11.010. 700
701
97. Frach, T.; Prescher, G.; Degenhardt, C.; De Gruyter, R.; Schmitz, A.; Ballzany, R. The Digital Silicon Photomultiplier - Principle of Operation and Intrinsic Detector Performance. In Proceedings of the IEEE Nuclear Science Symposium Conference Record; 2009; pp. 1959–1965. 702
703
704
98. Atherton, P.D. The Scanning Fabry-Perot Spectrometer. *Int. Astron. Union Colloq.* **1995**, *149*, 50–59, doi:10.1017/S0252921100022661. 705
706
99. Lindsay, S.M.; Shepherd, I.W. Linear Scanning Circuit for a Piezoelectrically Controlled Fabry-Perot Etalon. *Rev. Sci. Instrum.* **1977**, *48*, 1228–1229, doi:10.1063/1.1135204. 707
708
100. Afifi, H.A. Ultrasonic Pulse Echo Studies of the Physical Properties of PMMA, PS, and PVC. *Polym. - Plast. Technol. Eng.* **2003**, *42*, 193–205, doi:10.1081/PPT-120017922. 709
710

101. Root, S.E.; Gao, R.; Abrahamsson, C.K.; Kodaimati, M.S.; Ge, S.; Whitesides, G.M. Estimating the Density of Thin Polymeric Films Using Magnetic Levitation. *ACS Nano* **2021**, *15*, 15676–15686. 711
712
102. Polyanskiy, M.N. Refractive Index Database. [https://refractiveindex.info/?shelf=organic&book=poly\(methyl_methacrylate\)](https://refractiveindex.info/?shelf=organic&book=poly(methyl_methacrylate)). 713
103. ISO/IEC ISO/IEC 17043, Conformity Assessment - General Requirements for Proficiency Testing 2010. 714
104. Caves, C.M. Quantum-Mechanical Noise in an Interferometer. *Phys. Rev. D* **1981**, *23*, 1693–1708, doi:10.1103/PhysRevD.23.1693. 715
716
105. Slusher, R.E.; Hollberg, L.W.; Yurke, B.; Mertz, J.C.; Valley, J.F. Observation of Squeezed States Generated by Four-Wave Mixing in an Optical Cavity. *Phys. Rev. Lett.* **1985**, *55*, 2409–2412, doi:10.1103/PhysRevLett.55.2409. 717
718
106. Lawrie, B.J.; Lett, P.D.; Marino, A.M.; Pooser, R.C. Quantum Sensing with Squeezed Light. *ACS Photonics* **2019**, *6*, 1307–1318, doi:10.1021/acsp Photonics.9b00250. 719
720
107. Giovannetti, V.; Lloyd, S.; Maccone, L. Quantum-Enhanced Measurements: Beating the Standard Quantum Limit. *Science* (80-.). **2004**, *306*, 1330–1336, doi:10.1126/science.1104149. 721
722
108. Vahlbruch, H.; Mehmet, M.; Danzmann, K.; Schnabel, R. Detection of 15 DB Squeezed States of Light and Their Application for the Absolute Calibration of Photoelectric Quantum Efficiency. *Phys. Rev. Lett.* **2016**, *117*, 1–5, doi:10.1103/PhysRevLett.117.110801. 723
724
725
109. Polzik, E.S.; Carri, J.; Kimble, H.J. Spectroscopy with Squeezed Light. *Phys. Rev. Lett.* **1992**, *68*, 3020–3023, doi:10.1103/PhysRevLett.68.3020. 726
727
110. Casacio, C.A.; Madsen, L.S.; Terrasson, A.; Waleed, M.; Barnscheidt, K.; Hage, B.; Taylor, M.A.; Bowen, W.P. Quantum-Enhanced Nonlinear Microscopy. *Nature* **2021**, *594*, 201–206, doi:10.1038/s41586-021-03528-w. 728
729
111. Yu, H.; McCuller, L.; Tse, M.; Kijbunchoo, N.; Barsotti, L.; Mavalvala, N.; Betzwieser, J.; Blair, C.D.; Dwyer, S.E.; Effler, A.; et al. Quantum Correlations between Light and the Kilogram-Mass Mirrors of LIGO. *Nature* **2020**, *583*, 43–47, doi:10.1038/s41586-020-2420-8. 730
731
732
112. Taylor, M.A.; Janousek, J.; Daria, V.; Knittel, J.; Hage, B.; Bachor, H.A.; Bowen, W.P. Biological Measurement beyond the Quantum Limit. *Nat. Photonics* **2013**, *7*, 229–233, doi:10.1038/nphoton.2012.346. 733
734
113. Schnabel, R.; Mavalvala, N.; McClelland, D.E.; Lam, P.K. Quantum Metrology for Gravitational Wave Astronomy. *Nat. Commun.* **2010**, *1*, 110–121, doi:10.1038/ncomms1122. 735
736
114. Tse, M.; Yu, H.; Kijbunchoo, N.; Fernandez-Galiana, A.; Dupej, P.; Barsotti, L.; Blair, C.D.; Brown, D.D.; Dwyer, S.E.; Effler, A.; et al. Quantum-Enhanced Advanced LIGO Detectors in the Era of Gravitational-Wave Astronomy. *Phys. Rev. Lett.* **2019**, *123*, 231107, doi:10.1103/PhysRevLett.123.231107. 737
738
739
115. de Andrade, R.B.; Kerdoncuff, H.; Berg-Sørensen, K.; Gehring, T.; Lassen, M.; Andersen, U.L. Quantum-Enhanced Continuous-Wave Stimulated Raman Scattering Spectroscopy. *Optica* **2020**, *7*, 470, doi:10.1364/optica.386584. 740
741
116. Li, T.; Li, F.; Liu, X.; Yakovlev, V. V.; Agarwal, G.S. Quantum-Enhanced Stimulated Brillouin Scattering Spectroscopy and Imaging. *Optica* **2022**, *9*, 959, doi:10.1364/optica.467635. 742
743
117. Danilishin, S.L.; Khalili, F.Y.; Miao, H. Advanced Quantum Techniques for Future Gravitational-Wave Detectors. *Living Rev. Relativ.* **2019**, *22*, 1–89, doi:10.1007/s41114-019-0018-y. 744
745
118. Dorfman, K.E.; Schlawin, F.; Mukamel, S. Nonlinear Optical Signals and Spectroscopy with Quantum Light. *Rev. Mod. Phys.* **2016**, *88*, 1–67, doi:10.1103/RevModPhys.88.045008. 746
747
119. Li, T.; Anderson, B.E.; Horrom, T.; Schmittberger, B.L.; Jones, K.M.; Lett, P.D. Improved Measurement of Two-Mode Quantum Correlations Using a Phase-Sensitive Amplifier. *Opt. Express* **2017**, *25*, 21301, doi:10.1364/oe.25.021301. 748
749
120. Taylor, M.A.; Bowen, W.P. Quantum Metrology and Its Application in Biology. *Phys. Rep.* **2016**, *615*, 1–59, doi:10.1016/j.physrep.2015.12.002. 750
751
752
753
754
755