



HAL
open science

Intrinsically quantitative 2D HSQC NMR: a tool for deciphering complex mixtures

Jonathan Farjon

► **To cite this version:**

Jonathan Farjon. Intrinsically quantitative 2D HSQC NMR: a tool for deciphering complex mixtures. Annual reports in NMR, 2023. <hal-04242500>

HAL Id: hal-04242500

<https://hal.science/hal-04242500v1>

Submitted on 15 Oct 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

Intrinsically quantitative 2D HSQC NMR: a tool for deciphering complex mixtures

Jonathan Farjon^{a*}

^aNantes Université, CNRS, CEISAM UMR 6230, F-44000 Nantes, France

*Corresponding author: jonathan.farjon@univ-nantes.fr

Contents

1. Introduction.....	4
2. Different strategies for quantifying with 2D NMR	6
3. Intrinsically quantitative 2D NMR: Q-HSQC	8
4. Improving sensitivity and resolution: QUIPU HSQC	12
5. Mimicking 1D ¹ H NMR with 2D HSQC ₀	14
6. Global comparison between HSQC ₀ and Q-HSQC	17
7. Extending the toolbox for intrinsically 2D quantitative NMR	20
7.1 Better solvent elimination.....	20
7.2 ¹ J _{CH} compensation.....	21
7.3 Tolerant of any factor (TAF) HSQC	21
7.4 Faster acquisitions.....	23
7.5 Improving processing procedures for more accurate quantities.....	23
7.6 Making quantitative analysis more accessible	25
7.7 Boosting sensitivity.....	25
8. Concluding remarks.....	27

Number of words including Figures and References: 10285

Abstract:

Analysis of complex mixtures is of main interest since they are ubiquitous in the nature. Thus, it becomes challenging to decipher their complexity by developing NMR methods able to identify but also to quantify molecules in the most accurate way. 1D ¹H NMR is highly utilized for quantifying molecules but suffers from many overlaps making this approach limited. 2D NMR is a great manner to improve the signal

dispersion over 2 dimensions and the ^1H - ^{13}C HSQC experiment is the best choice to improve the spectral resolution despite a lower sensitivity. However, due to its multiple pulse nature, the quantitativity of HSQC is impacted and biased absolute concentrations are derived. In this chapter, we highlight the potential of a class of methods called intrinsically quantitative 2D HSQC. They are generated by designing pulse sequences able for each correlation peak to maintain constant or to compensate J_{HH} , J_{CH} , off resonances and T_2 effects responsible for the degradation of the quantitativity. Pioneer experiments as Q-HSQC and HSQC₀ are detailed as well as their improved versions and applicative examples are given. Then, these major techniques are compared to show their performances and limits. Finally, recent methodological developments are proposed and discussed to enhance the potential of future intrinsically quantitative 2D-based complex mixture analyses.

Keywords: Complex mixtures, Intrinsically Quantitative NMR, bidimensional NMR, fast NMR, pure shift NMR, Q-HSQC, HSQC₀, QUIPU HSQC

Abbreviations

ASAP	Acceleration by Sharing Adjacent Polarization
BIRD	Bilinear Rotation Decoupling
COSY	Correlation Spectroscopy
CPMG	Carr-Purcell-Meiboom-Gill
D-DNP	Dissolution Dynamic Nuclear Polarization
FMLR	Fast Maximum Likelihood Reconstruction
HSQC	Heteronuclear Single Quantum Correlation
HSQC₀	Heteronuclear Single Quantum Correlation zero
INEPT	Insensitive Nuclei Enhancement Polarization transfer
NUS	Non Uniform Sampling
PHIP	Parahydrogen Induced Polarisation
QEC-HSQC	Quantitative Equal Carbon HSQC
Q-HSQC	Quantitative HSQC
Q-OCCAHSQC	Quantitative, offset-compensated, CPMG-adjusted HSQC
QQ-HSQC	Quick Quantitative HSQC
Q QUIPU HSQC	Quick QUantitative Perfected and pUre shifted
QUIPU HSQC	QUantitative Perfected and pUre shifted
SABRE	Signal Amplification By Reversible Exchange

SENSASS HSQC SENSitivity increAsed and resolution enhanced by Signal
Suppression

SNRs Signal-to-Noise Ratios

ST2-PT Single Transition-to-single Transition Polarization Transfer

T_1 Longitudinal relaxation time

T_2 Transverse relaxation time

TAF HSQC Tolerant of Any Factor HSQC

TOCSY Total Correlation Spectroscopy

UFCOSY Ultra Fast Correlation Spectroscopy

UHPLC Ultra High Performance Liquid Chromatography

1. Introduction

NMR is a unique ensemble of methods able to give diverse information on molecules even in complex mixtures, their identity and quantity can jointly be accessed in addition to dynamic data related to exchange and conformations. The determination of the molecular quantity per unit of matter or the related concentration is becoming essential in various fields including pharmaceutical analysis [1], food sciences [2], plants [3] and environmental investigations [4]. Even though the quantitativity by NMR is suitable for a wide range of applications like accessing the number of nuclei per site, the relative and absolute concentration but also isotopic ratios, the review will be dedicated to the determination of concentrations.

Thus, 1D quantitative NMR or so called as qNMR has been highly exploited using well established protocols [5]. Although ^1H 1D NMR is a highly repeatable method its

application to complex mixture is limited due to many peak overlaps making quantification inaccurate or impossible. As a good alternative 2D NMR has the main advantage to extend the spectral dispersion over 2 dimensions. Among them homonuclear like COSY with its ultra-fast version: UFCOSY have been intensively used [6] but also TOCSY [7]. Heteronuclear ^1H - ^{13}C HSQC [8] is recognized as one of the most resolutive approach due to the larger ^{13}C spectral range for improving correlations separation despite the lower intrinsic sensitivity of ^{13}C . Moreover, 2D HSQC experiments are made of multiple pulses schemes embedding delays with a peak-dependent analytical response of the nuclei. This is due to numerous different modulations occurring during the pulse sequence either coming from homonuclear / heteronuclear scalar couplings, offsets effects and from transverse relaxation evolutions impeding the quantitative potential of NMR. Under these conditions, how could it be possible to make intrinsically 2D NMR quantitative?

An intrinsically quantitative 2D HSQC is self-sufficient for quantifying any molecule having correlation signals on 2D maps. The goal to tackle with intrinsically quantitative 2D HSQC is to reach like with 1D ^1H NMR the possibility to derive quantities from the integration of signals by using either an internal or an external reference.

In this chapter, the main purpose is to show and explain the different HSQC-based strategies by designing pulse sequences for turning standard 2D ^1H - ^{13}C HSQC into an “intrinsically” quantitative version. In this framework, Q-HSQC and HSQC₀ as pioneering experiments will be compared as well as their improved versions making them more performant and robust to tackle analytical challenges met in various fields of application (Figure 1). Moreover, some interesting recent methodological advances

will also be added to the main discussion to extend the potential of future intrinsically quantitative experiments.




Fig. 1. Q-HSQC and HSQC₀ and their improved versions are the pillars of the intrinsically 2D quantitative NMR temple.

2. Different strategies for quantifying with 2D NMR

Quantitative NMR (qNMR) is suitable to complex mixture analysis and has main applications for food, metabolomics, natural products, plants, herbs, forensics and pharmaceuticals. 1D ¹H NMR is recognized as intrinsically quantitative and is the most sensitive and fast strategy to access concentrations [5]. The most particular property of qNMR comes from the proportionality between the signal integral (Int^{1D}), the number of nuclei (p), the sensitive volume of the coil (V_s) and the concentration [c],

$$\text{Int}^{\text{1D}} = p \cdot [c] \cdot V_s \quad (\text{Eq. 1})$$

When overlaps are impeding 1D NMR, one good alternative is HSQC offering an enhanced spectral dispersion. HSQC can be used for quantifying complex mixtures and integral for correlation peaks (Int^{2D}) are depending, like 1D, on relaxation T_1 and

T_2 times but also all J_{HH} , J_{CH} and chemical shifts are evolving during the whole HSQC pulse sequence duration. These additional modulations could be gathered as a k factor, a site-specific coefficient of response for each correlation peak (see Fig.2):

$$\text{Int}^{2D} = k \cdot p \cdot [c] \cdot V_s \quad (\text{Eq. 2})$$

Figure 2 shows the different approaches that have been implemented with this aim. The first one is based on the determination, for each peak, of the exact value of k, supported by theoretical considerations [9] where all scalar couplings, relaxation times shall be precisely determined (see right part of Figure 2). This approach is time-consuming since it requires all NMR interactions to be accurately measured to enable determination of the concentration in the best manner.




Fig. 2. Different approaches to extract concentrations data from 2D NMR maps like HSQC.

A second type of approach gives access to k values through conventional analytical methods such as calibration or standard additions [10,11,12] (bottom arrow under the spectrum on Figure 2). These strategies are not different from those used in previous decades in other analytical techniques such as gas and liquid chromatography and Infra-Red / Raman spectroscopies. In the external calibration approach, calibration curves are obtained by plotting the 2D peak volumes versus the corresponding concentration. Then, a 2D spectrum is recorded on the sample under study containing the analyte of interest to quantify. Its concentration is determined with the help of the regression curve equation. This approach offers a great solution to the problem of accuracy in 2D NMR but there is a need to commercially access or to synthesize each molecule to quantify with a known purity.

A third approach is based on intrinsically quantitative 2D NMR and related pulse sequence adaptations and developments to make them usable for reaching concentrations (see the left part of Figure 2). Among such intrinsically quantitative 2D NMR methods Q-HSQC and HSQC₀ were the pioneers to be implemented.

3. Intrinsically quantitative 2D NMR: Q-HSQC

Making 2D experiments a quantitative tool needs modifying and/or combining several NMR pulse schemes, and such developments have been mainly focussed on ¹H–¹³C HSQC pulse sequences since they are easier to manipulate for compensating site-specific interactions. They were firstly developed by Heikkinen and coworkers, who proposed a method to eliminate the effects of heteronuclear ¹J_{CH} couplings in the ¹H–¹³C HSQC pulse sequence [13]. In the conventional HSQC experiment, the 2D peak volumes are proportional to $\sin^2(\pi\Delta^1J_{CH})$ where Δ is the duration of the INEPT polarization transfer delay [14]. In 2003, Heikkinen *et al.* developed for compensating

$^1J_{\text{CH}}$ to sum four HSQC spectra with different and suitably selected Δ : three with a smaller and one with a bigger value [15]. As a result, in the combined spectrum, 2D peak volumes are uniformly affected by $^1J_{\text{CH}}$ couplings from 125 to 225 Hz range (see Figure 3) with only 2% of variation (bottom green curve in Figure 3) as compared to 50% variation with the standard HSQC (dashed blue curve in Figure 3). Moreover, due to $^1J_{\text{CH}}$ compensation the sensitivity of signal is 25% lower (see Figure 3) as compared to the standard HSQC.




Fig.3. $^1J_{\text{CH}}$ dependence on different HSQC schemes impeding the quantitativity.

The approach proposed by Heikkinen *et al.* [16,17] requires four times more scans than the standard HSQC for compensating $^1J_{\text{CH}}$ evolutions. In 2007 this drawback was compensated for by Peterson and Loening, who proposed to acquire from different slices of the sample the signals corresponding to the different INEPT delays [18]. This approach called quick, quantitative HSQC (QQ HSQC) is 4 times faster than the one described by Heikkinen for the same result.

However, the 2D peak volumes in the HSQC spectra are also modulated by homonuclear J_{HH} couplings [19]. As described in theory, the cross-peak intensity can be affected by a factor:

$$\prod_{i=1}^k \left(\frac{1}{2} (1 + \cos(2\pi J_{HHi} \Delta)) \right)^2 \quad (\text{Eq. 3})$$

In 2005, Heikkinen *et al.* implemented a compensation for the effect of J_{HH} in the Carr–Purcell–Meiboom–Gill (CPMG) pulse trains in the INEPT and retro-INEPT schemes (see Figure 4 C) of the HSQC pulse sequence [19]. They also included composite CPMG pulse trains to reduce off-resonance effects that are more pronounced especially on the edge of the ^{13}C chemical shift range. More recently, they introduced adiabatic pulses (see Figure 4 H) in the evolution period of their pulse sequence to improve the quantitativity of the experiment [19]. The resulting pulse sequence, called quantitative, offset-compensated, CPMG-adjusted HSQC (Q-OCCAHSQC), including all these improvements.

Lowered performances could be obtained since peak volumes also depend on T_2 transverse relaxation times, that is not considered in the Q-HSQC protocol. This means that the Q-OCCAHSQC protocol is more convenient to quantify molecules with close or similar T_2 values, as reported by Zhang and Gellerstedt who developed a protocol where cross peaks with the same T_2 relaxation profiles were analyzed as individual groups [20]. The analytical performance was evaluated for this method, and they reported errors of up to 17% on a strychnine sample, with a 7.3% standard deviation.

More recently Heikkinen *et al.* were improving the Q-HSQC experiment by developing the Quantitative Equal Carbon or QEC HSQC [21]. QEC HSQC like Q

HSQC is requiring 4 times more scans to average out $^1J_{CH}$ over different CH_n sites of the studied molecules. This approach is allowing for integrating any correlation peak to derive quantitative data with no need of normalizing integrals by the number of 1H (see p in Figure 2). A great advantage of equalizing CH_n responses allows for increasing the dynamic range and reducing t_1 noise especially for methyl groups. With QEC HSQC, it was possible to obtain relative quantification of all 1H sites for brucine and also cinchonidine with a good 5 to 7% accuracy range.

In 2020, Zhang and coworkers were proposing another version of the Q-HSQC that is J -compensated and using a CPMG-based Levitt's composite echo-based HSQC. This Q-HSQC version is more sensitive than the first one published by Heikkinen [15] since having the same sensitivity than the standard HSQC in a $^1J_{CH}$ range from 120 to 180 Hz (see Figure 3, the top orange curve). As previously pointed out, Q-HSQC is dependent on different T_2 values varying from site to site and that can be different from the internal / external used reference. This approach in combination with a T_2 normalization protocol was valorized for quantifying some drugs in cream formulations to determine their absolute concentration as well as their enantiomeric excess with the help of a chiral solvating agent [22].

In the family of intrinsically quantitative HSQC experiments, Guerrini *et al.* succeeded in obtaining the relative composition of different monosaccharidic units of heparines [23,24]. They showed that finely tuning INEPT delays in standard HSQC, with respect to specific sites like anomers allowed accurate relative quantities to be obtained.




Fig. 4. Improvements of Q-HSQC by combining different schemes of interest. Prior to the excitation for the elimination of solvent signal(s) like A) presaturation with a continuous wave (CW) or B) multiple selective schemes (MPsat). J_{HH} are eliminated during INEPTs with C) CPMG or D) perfect echoes. The indirect dimension could be accelerated by using Non Uniform Sampling (NUS) and / or spectral aliasing. For preventing frequency offset effects, hard 180° ^{13}C inversion / refocussing pulses could be replaced by G) composite and H) adiabatic pulses. The acquisition could be a standard full FID or a pure-shift homodecoupled one by concatenated BIRD and chunks of FID.

4. Improving sensitivity and resolution: QUIPU HSQC

One main limitation of ^1H - ^{13}C HSQC is due to its lower sensitivity intrinsically resulting from the low natural abundance of ^{13}C . Moreover, for highly complex mixture having structurally close molecules, there is a need of much more resolute methods. Under this context Farjon and coworkers designed in 2016 the QUIPU

HSQC experiment able to give access to absolute quantification like for 1D ^1H qNMR by using an internal / external – reference with an accuracy that could reach 10%. This technique includes $^1J_{\text{CH}}$ compensation like in Q-HSQC (see Section 3.) but with perfect INEPT (see Figure 4 D) and retro-INEPT [25] to replace CPMG-INEPT schemes (see Figure 3 C) to remove J_{HH} modulations. In addition, the acquisition is homodecoupled thanks to a BIRD based pure shift scheme [26] for highly disperse correlations over the 2D map. Thanks to J_{HH} collapsing, sensitivity gains up to 2.6 were obtained on vegetal extracts allowing for quantifying metabolites in the range of 100 μM [27].

However, quantitative 2D NMR is time-consuming due to its 2D and quantitative nature. This experiment was accelerated in 2017 by Farjon *et al.* as Q QUIPU HSQC [28]. This approach combines Non Uniform Sampling (NUS) (see Figure 4 E), spectral aliasing (see Figure 4 F) [29] and variable recycling time [30] to speed up by a factor from 6 to 9, in comparison with the QUIPU version and allowing for quantification of metabolites in the 100 μM range from breast cancer cell extracts (Figure 5) in 5 h [28]. In addition, Q QUIPU HSQC was successful in quantifying natural products in food supplements. In this application, quantifications were obtained in very good agreement with data derived from UHPLC-based methods [31].




Fig. 5. QQUIPU HSQC used to quantify cell breast extracts A) Aliphatics, B) aromatics. C) Diagram bars for quantifying different metabolites within cell breast extract by using lactate as an internal reference. Reprinted with permissions from [28] Copyright (2018) Elsevier

5. Mimicking 1D ^1H NMR with 2D HSQC₀

Another strategy was proposed by Markley and his group to obtain k values (see Figure 2) that are constant whatever the site [32] This approach relies on obtaining an HSQC spectrum where correlation signals are not dependent on the molecular site. To obtain such a spectrum called HSQC₀ there is a need to acquire 3 successive HSQC

pulse sequences having 1, 2 or 3 HSQC blocks (Figure 6). With the hypothesis that the signal reduction is a function of J -couplings, relaxation times and off-resonance effects (see Figure 2) proportional to the number of HSQC i with $i=1,2,3$ block, the signal volume $V_{i,m}$ for a given peak m follows the equation :

$$\ln(V_{i,m}) = \ln(V_{0,m}) + i \ln(k_{A,m}) \text{ (Eq. 4)}$$

where $k_{A,m}$ is the amplitude reduction factor for a 2D signal m . By plotting, for each 2D signal, the logarithm of the 2D volume $\ln(V_{i,m})$ as a function of i , it is possible to calculate a 2D volume signal as though there was no HSQC or zero HSQC: HSQC $_0$, $V_{0,m}$ considering no evolutions at all in the HSQC like a 1D 90° pulse. The HSQC $_0$ 2D signal volume is consequently not dependent on evolution parameters like T_2 , J -couplings, delays, thus leading to absolute concentrations needing a reference signal. Even if the trueness of this technique was not highlighted, a good precision was reported, with a standard deviation close to 7%.




Fig. 6. (Top) Pulse sequence used for extrapolating HSQC₀ from 1, 2 or 3 consecutive 2D ¹H-¹³C HSQC. Narrow and wide black bars indicate 90° and 180° pulses, respectively. The delays are $\tau = 3.3$ (or 3.4) ms ; $T = 5.3$ (or 5.8) ms. The phase cycling is as follows : $\phi_1 = -y, y$; $\phi_2 = y, -y$; $\phi_3 = 2[x], 2[-x]$; $\phi_{\text{rec}} = x, -x, -x, x$. All other radio frequency pulses are applied with phase x, except as indicated. Quadrature detection in the ¹³C (t_1) dimension is achieved using States-TPPI applied to the phase ϕ_3 . The duration and strength of the pulsed field gradients applied along the z-axis are as follows : (g_1) 3 ms, 15.9 G/cm ; (g_2) 0.4 ms, 10.6 G/cm ; (g_3) 1 ms, 18.6 G/cm ; (g_4) 1 ms, -26.5 G/cm, and (g_5) 0.6 ms, 13.25 G/cm, followed by a gradient recovery period of 200 μ s. The 180° pulse marked with an asterisk serves to refocus proton chemical shift during the gradient delays for g_3 and g_4 to achieve better t_1 noise suppression. Pulse sequence proposed here for time-zero extrapolated HSQC (HSQC₀) spectroscopy. Three data sets are collected: the first at the position indicated by HSQC₁, the second at the position indicated by HSQC₂, and the third at the position indicated by HSQC₃; each data collection proceeds as shown for HSQC₃. In all three pulse sequences, composite phase cycling of ϕ_1 and ϕ_2 is carried out in the HSQC unit that precedes data acquisition ; otherwise, $\phi_1 = -y$ and $\phi_2 = y$. $\phi_3 = 2[x], 2[-x]$; $\phi_{\text{rec},1} = x, -x, -x, x$. The constant time chemical shift evolution period T is replaced by the invariant constant time T in the second and third HSQC units, respectively for HSQC₂ and HSQC₃. Additional phase cycling for HSQC₂ is $\phi_4 = 4[x], 4[-x]$; $\phi_{\text{rec},2} = x, -x, -x, x, -x, x, x, -x$. Additional phase cycling for HSQC₃ is $\phi_5 = 8[x], 8[-x]$; $\phi_{\text{rec},3} = x, -x, -x, x, x, -x, x, x, -x, -x, x, x, -x, x, -x, -x, x$. (Bottom) Shown are examples of HSQC₁, HSQC₂, and HSQC₃ spectra following Fourier transformation to yield frequency domain spectra. Peak intensities in the virtual HSQC₀ spectrum (bordered by the dashed line) are derived from linear extrapolation of the peak intensities of the

corresponding peaks in the HSQC₁, HSQC₂, and HSQC₃ spectra. Reprinted and adapted with permission from [32] Copyright (2011) American Chemical Society

This approach was improved in 2011 by incorporating suitably chosen coherence selection gradients yielding cleaner spectra [33]. It was successfully applied to quantify molecules in model samples [34], complex mixtures of natural products [35] in a targeted metabolomics study [36] as well as mixtures of lignins [37]. In 2022, HSQC₀ was even combined with the EXSY experiment and kinetic studies to better understand conformational changes in a peptide called Nukacin ISK-1 [38].

6. Global comparison between HSQC₀ and Q-HSQC

As previously described, in the class of intrinsically 2D HSQC, the best candidates are HSQC₀ and Q-HSQC that have advantages and drawbacks (see Table 1).

Table 1. Detailed comparison for intrinsically quantitative HSQCs and performed improvements

	HSQC ₀	Improvements	Q-HSQC	Improvements
Acquisition time	+++ - At least 3x more time - HSQC ₃ could reacquire more scans	Use of gradients to select coherences to reduce t ₁ noise	++++ 4x more scans / ¹ J _{CH} compensation	+ * Q-OCCAHSQC J _{HH} compensation More tolerant to offset effects * QQ HSQC Time /4 * QUIPU HSQC More sensitive and resolved * QQUIPU HSQC Faster
Easiness to implement	++++	/	++	/
Accuracy of concentrations	++++ <5%	++++ <5%	++ 5 to 10% Due to uncompensated	T ₂ normalization

			T_2 effect	
T_2 effects	T_2 are compensated by extrapolation to zero HSQC but for HSQC ₂ and more for HSQC ₃ T_2 is affecting a lot SNRs, especially for bigger-sized compounds	/	Not compensated	T_2 normalization
Sensitivity considerations	++++HSQC ₁ ++HSQC ₂ +HSQC ₃		+++ -25% due to $^1J_{CH}$ compensation	++++ - QUIPU with improved SNRs +20 to 180% / no pure-shift version ++++ -CPMG-Jcomp the same sensitivity than standard HSQC
Postprocessing	Longer since 3 maps to process +		+++	/

Variations of HSQC₀ seem more worthwhile to consider than Q-HSQC, as it considers the effects of transverse relaxation and leads to more accurate quantitation, but it is not always easy to get a good sensitivity especially for HSQC₃ due to T_2 evolution and the associated loss in sensitivity. Moreover, for HSQC₀, the hypothesis is made that the NMR signal intensity is attenuated linearly as a function of the number of repetitions. Q-HSQC and HSQC₀ were jointly implemented to determine very similar accurate absolute concentrations for organic acids in mixture [27] as well as the relative amount of different structural units in wood lignin, a complex polymer (see Figure 7) [39].




Fig. 7. (A) Full range HSQC₁ obtained for acetylated Norway spruce MWL using a 600 MHz spectrometer equipped with a cryoprobe. (B) Zoom-ins for the aromatic region, and the aliphatic region of contour-diagrams obtained for interesting regions for the HSQC₁, HSQC₂, and HSQC₃. (D) Graphical comparison of the intensities of signals corresponding to the main interunit bonding motifs found in the Norway spruce MWL sample (as % C9); legend: **A** – $\beta\text{-O-4'}$ (H α), **B** – $\beta\text{-O-4'}$ (H β including oxidised derivative), **C** – $\beta\text{-O-4'}$ (C2-H' of oxidised derivative), **D** – $\beta\text{-5'}$ (H α), **E** – $\beta\text{-1'}$ (H β), **F** – 5,5'- $\alpha,\beta\text{-O-4'}$ (H'), **G** – $\beta\text{-}\beta'$ (H α), **H** – cinnamyl alcohol (H Γ). Figure adapted and reprinted from [39] under the terms of the Creative Commons CC-BY license.

These techniques do not need knowledge on the targeted analytes. They must be used together with a reference method either internal or external, as in 1D ^1H qNMR

experiments. Still, some development will be essential in the future to make these methods more performant.

7. Extending the toolbox for intrinsically 2D quantitative NMR

Some improvements were already published to compensate some weaknesses of Q-HSQC and HSQC₀ (see Section 6.). However, there is still a need for some improvements for enhancing accuracy and precision on concentrations by better compensating for all biasing quantitativity sources. In addition, the quest to improve the sensitivity and speed of acquisitions for high throughput studies remain essential for complex mixture analyses. Further, making intrinsically 2D quantitative NMR usable on compact NMR systems is also highly desirable when there is a need of more accessibility.

In this part, recent developments taken from literature are discussed and proposed to reach more performant intrinsically quantitative 2D HSQCs.

7.1 Better solvent elimination

For biological samples, which are mainly made of water, it remains essential to have good water elimination especially to improve the sensitivity through the optimisation of receiver gain. Moreover, the choice of the elimination scheme and its position within the HSQC is of particular interest. In this field, Gronwald *et al.* [40] compared the presaturation-based (see Figure 4 A) Q-HSQC with the QUIPU HSQC and showing that such water vanishing methods did not sufficiently remove the residual water signal and its sidebands due to the homodecoupled acquisition. Thus, authors showed that the trueness for signals closer to the water was worse with QUIPU HSQC as compared to the Q-HSQC. Under these conditions, there is a need of a better global

water elimination technique in particular during the pure-shift scheme such as using weak gradients during BIRD and 180° hard pulse as proposed by Luy *et al.* [41].

7.2 $^1J_{\text{CH}}$ compensation

In order to further compensate for $^1J_{\text{CH}}$ couplings in INEPT blocks (see Sections 3. and 4.), other strategies of interest are using J -compensation through perfect-echo in combination with optimally controlled shaped pulse sandwiches [42]. In addition, swept coherence transfer based on the co-incrementation of J -evolution delay in INEPT with 2D incremental time by using the time-resolved NUS method allowed for obtaining quantitative data [43]. More recently, spatially encoded INEPT using a frequency swept broadband pulse for homogeneously transferring $^1J_{\text{CH}}$ seems to give access to 2D integrals that are more usable for quantitative studies especially when combined with a perfect INEPT [44], it allows for refocusing J_{HH} over the INEPT time [45]. These different approaches to compensate for $^1J_{\text{CH}}$ evolutions deserve to be implemented and tested in intrinsically quantitative 2D HSQC in the near future.

7.3 Tolerant of any factor (TAF) HSQC

In the field of compensating for any source of mis-quantification (see Figure 2), to push over the possibilities of quantitative HSQC for larger molecules, Okamura *et al.* [46] have developed a family of HSQC that could implement a TROSY scheme as intrinsically quantitative. In order to compensate for the different modulations especially due to T_2 evolutions, the authors combined different HSQC allowing for normalizing all volumes within the 2D map. They analyzed a mixture of lignin dimer and curdlan to test their approach. Under these conditions, it was possible to reach in a relative quantitative manner all T_2 compensated volumes (see Figure 8). By

considering the transverse relaxation during INEPT (see Figure 8 C): The magnetization decreases to $\delta \cdot 2H_z C_x$ (b) with $0 < \delta < 1$, and become $\delta \cdot \epsilon \cdot H_\gamma C^\beta + \delta \cdot \epsilon \cdot H_\gamma C^\alpha$ (d) with $0 < \epsilon < 1$ the reduction of the magnetization due to Single Transition-to-single Transition Polarization Transfer or ST2-PT (see Figure 8 C). Inverting the phase of the second 90° ^1H pulse: ψ and combining spectra is leading to differentiate different T_2 evolutions ϵ and δ through the α factor:

$$\alpha = [\text{Peak volume } (\psi = -y) + \text{Peak volume } (\psi = y)] /$$

$$[\text{Peak volume } (\psi = -y) - \text{Peak volume } (\psi = y)]$$

α is proportional to $\delta \cdot (u/v)$. In order to consider the different T_2 evolutions along with HSQC sequences (see Figure 8 C), peaks volume shall be corrected by dividing them by α^2 . Under these conditions signal M from lignin and P from curdlan are reaching a ratio 1 : 1 for M/P (Figure 8 B) instead of 0.6 : 1 for the conventional HSQC (Figure 8 A).




Fig. 8. Calibration of volumes of HSQC correlation peaks to correctly determine the amount of each component. The volumes of each correlation peak of the ^1H - ^{13}C HSQC spectrum recorded with either A) the standard HSQC or B) the TROSY-based HSQC (B). The volume of the peak corresponding to each CH moiety was obtained. The volumes are normalized so that the average of the volumes for the peaks of the lignin dimer is 1. Calibrated volumes were calculated and plotted for HSQC/ α^2 . Error bars represent standard deviations of four independent experimental data sets. Adapted and reprinted from ref [46] under the terms of the Creative Commons CC-BY license.

7.4 Faster acquisitions

As previously reported (see Sections 3 for QQ HSQC and 4 for QQUIPU HSQC) for some applications especially for high throughput studies, requiring to analyze a high number of samples in a short time, there is a need of faster NMR methods. For speeding up 2D HSQC, different approaches were proposed in the literature like fast repetition NMR [47] especially SENSASS HSQC [48] and ASAP HSQC [49]. Moreover, NUS [50] is nowadays an easy way to accelerate the acquisition of 2D matrices. More recently, the Time-Resolved NUS version was able to monitor quicker events with 2D experiments [51].

7.5 Improving processing procedures for more accurate quantities

Designing sequences for intrinsically quantitative 2D HSQC is one key to reach the most accurate concentration. However, this is essential to be able to process spectra with the most reliable procedures for reaching the most accurate concentrations. Thus, 2D signal volumes apodization and integration have to follow some requirements.

Thus, preserving the 2D peak volume determined by the maximum amplitude of the FID is crucial [52]. Thus, the weighting function and the number of points along F_1

have to be finely chosen. They have to follow the complete decrease to zero at the end of the FID sampling and its maximum has to fit with the maximum of the NMR signal on the beginning of the FID. In addition, baseline distortions could also impact peak volumes and related concentration accuracy [53]. For 2D NMR two baseline corrections along each dimension can be applied to peak volumes. Baseline correction are most of the time applied in the frequency domain as a polynomial function with a degree lower than 3 [54,55]. When the polynomial order is higher certain peaks could be mixed with the baseline leading to distortions biasing the quantitative analysis. The most interesting approaches to perform a reproducible analysis of 2D NMR spectra are with a parametric model fitting to spectral deconvolution [56]. Some papers reported the use of such techniques to determine peak volumes from 2D NMR spectra. Chylla *et al.* [57] for a protein showed the application of a hybrid time-domain, frequency-domain maximum likelihood fitting. They developed an algorithm called fast maximum likelihood reconstruction (FMLR) that performs spectral deconvolution of 2D NMR spectra for the purpose of accurate signal quantification [58]. FMLR is very promising and allowed reaching greater accuracy (0.5– 5.0% error) than peak height analysis and peak integral analysis with an operator work (Figure 9).




Fig. 9. A) Evaluation of the precision (% variance) and B) accuracy (% variation from a known concentration) of the fast maximum likelihood reconstruction method (light blue bars) in comparison with manual peak integration (dark blue bars), for accessing absolute concentrations in a mixture. Reprinted with permission from [58] R.A. Chylla, K. Hu, J.J. Ellinger, J.L. Markley, Deconvolution of Two-Dimensional NMR Spectra by Fast Maximum Likelihood Reconstruction: Application to Quantitative Metabolomics, *Anal. Chem.* 83 (2011) 4871–4880. Copyright (2011) American Chemical Society.

7.6 Making quantitative analysis more accessible

For more than a decade, compact mobile NMR apparatus have emerged, they don not require any cryogenic fluids, are cheaper and have reduced running and maintenance costs [59]. The previous sections show the interest in methodological developments to overcome the general limitations of NMR in the context of quantitative investigations. These techniques (see Sections 7.1 to 7.5) could become highly beneficial when implemented on compact NMR spectrometers since leading to peak overlap with lower sensitivities. Benchtop spectrometers have many advantages for quantitative investigations, for quality control of food [60], in forensic sciences [61] for monitoring a bioprocess [62] and for point-of-care test capabilities [63].

7.7 Boosting sensitivity

A great challenge to face in quantitative NMR for complex mixtures is its intrinsically low sensitivity. There is a need to improve it to tackle new challenges in chemistry and biology. Among the different solutions, hyperpolarization methods such as dissolution

dynamic nuclear polarization (D-DNP) and parahydrogen induced polarization (PHIP) are certainly the most promising.

D-DNP holds the promise to amplify the sensitivity of liquid state NMR by several orders of magnitude [64]. However, due to the short decay time of the generated hyperpolarized signal, D-DNP has been mainly applied to ^{13}C NMR, opening the boulevard to sensitive ^{13}C NMR detection in complex mixtures of molecules. Even if D-DNP has complex instrumental sources of biasing, the D-DNP experiment has recently shown to be repeatable at 5% on complex metabolic mixtures [65, 66]. A second class of hyperpolarization methods allowing to increase the sensitivity of quantitative NMR are based on the incorporation of para-hydrogen (p-H₂) into a targeted analyte by a catalytic hydrogenation reaction [67]. This approach was initially limited to analytes that could be hydrogenated, but it has been extended to a higher number of molecules through the signal amplification by reversible exchange or SABRE [68]. In the SABRE approach, the hyperpolarization is transferred from p-H₂ to the substrate via the coordination sphere of a metal complex, extending the field of application for this hyperpolarization method. In this strategy, the reversible binding of p-H₂ and of the analytes to a metal complex is exploited to transfer hyperpolarization from p-H₂ to compounds under study. Combined with 2D NMR acquisition schemes, this technique can track metabolites at sub-micromolar concentrations [69]. Tessari's team showed that this approach could be applied to targeted quantitative analysis of nicotine and its metabolite like cotinine in urine by using a standard addition workflow (see Figure 10) [70,71]. Despite the fact that PHIP allows the detection of specific metabolites binding to the metal catalyst, it provides a very sensitive approach for quantitative investigations.




Fig. 10. Calibration curves for Nicotine (red triangles) and Cotinine (blue circles) based on the linear fitting of 10 data points (2D spectrum signal integrals). Three parallels were measured and averaged for every point. Error bars represent the standard deviation of the three parallels. Calibration curves are analyte-specific due to different iridium binding affinities. Reproduced with permission from Ref. [70], Copyright (2021) American Chemical Society

8. Concluding remarks

Up to now, the main experiments of the intrinsic quantitative 2D NMR toolbox are based on Q-HSQC and HSQC₀. The methods described in this chapter are very promising from the quantitative point of view because they give access to 2D peak volumes directly reflecting the concentration with accuracies lower than 10%. The main point of improvement is relying on the compensation of T_2 effects mainly responsible for the lack of accuracy and loss of sensitivity. Moreover, they suffer from other limits impeding their applications.

On one hand, they need specific acquisition and/or processing procedures that are not commercially available, which could limit their general use. On the other hand, they are focussed to the HSQC experiment, which could limit their application to more concentrated samples.

In order to improve the quantitative performances of Q-HSQC and HSQC₀ different newly reported *J*-compensated schemes could be compared in terms of accuracy of derived concentrations. Moreover, different approaches like fast repetition NMR and pure-shift acquisition schemes could be combined to improve the performances of 2D HSQC in terms of sensitivity and also spectral dispersion.

In addition, there is a need of quantitative homonuclear based 2D methods. Despite more difficulties to make them quantitative due to multiple J_{HH} modulations requiring compensation, homonuclear 2D NMR is more sensitive than HSQC-based approaches.

Achieving more sensitivity could also be accessed with the help of dedicated instrumentation. Of course, increasing the magnetic field of the spectrometer could be a solution but using micro probes and cryoprobe are more affordable for magnifying the sensitivity. To go on boosting the sensitivity, hyperpolarization techniques like Parahydrogen based methods are of primary interest to enhance SNRs by several orders of magnitudes. The combination of hyperpolarized methods and microcoil through a microfluidic system was a recent major evolution to make lie on the low sensitivity of NMR [72].

These improvements should open new applicative perspectives for intrinsically quantitative 2D NMR, in particular in fields such as metabolomics, forensics and environmental sciences, where the complexity of samples and their low concentrated analytes are really challenging.

Acknowledgments

The author acknowledges the CNRS, Universities of Paris-South, Paris-Saclay, and INRA of Versailles for funding. J. F. warmly thanks the INRA for funding the internship of Soumaya Khlifi in the master formation context. This study benefited from a grant (LabEx Saclay Plant Sciences-SPS, reference ANR-10-LABX-0040-SPS), managed by the French National Research Agency under an Investments for the Future program (reference no. ANR-11-IDEX-0003-02). The authors from the CEISAM lab acknowledge the French National Infrastructure for Metabolomics and Fluxomics MetaboHUB-ANR-11-INBS-0010 (www.metabohub.fr) and the Corsaire metabolomics core facility (Biogenouest). J.F. thanks Sandrine Bouchet for an unfailing assistance.

9. References

- [1] U. Holzgrabe, R. Deubner, C. Schollmayer, B. Waibel, Quantitative NMR spectroscopy—Applications in drug analysis, *Journal of Pharmaceutical and Biomedical Analysis*. 38 (2005) 806–812.
- [2] L. Laghi, G. Picone, F. Capozzi, Nuclear magnetic resonance for foodomics beyond food analysis, *TrAC Trends in Analytical Chemistry*. 59 (2014) 93–102.
- [3] J.W. Allwood, R.C.H. De Vos, A. Moing, C. Deborde, A. Erban, J. Kopka, R. Goodacre, R.D. Hall, Chapter sixteen - Plant Metabolomics and Its Potential for

Systems Biology Research: Background Concepts, Technology, and Methodology, in: D. Jameson, M. Verma, H.V. Westerhoff (Eds.), *Methods in Enzymology*, Academic Press, 2011: pp. 299–336.

[4] J.G. Bundy, M.P. Davey, M.R. Viant, Environmental metabolomics: a critical review and future perspectives, *Metabolomics*. 5 (2009) 3–21.

[5] C. Simmler, J.G. Napolitano, J.B. McAlpine, S.-N. Chen, G.F. Pauli, Universal quantitative NMR analysis of complex natural samples, *Current Opinion in Biotechnology*. 25 (2014) 51–59.

[6] C. Lhoste, B. Lorandel, C. Praud, A. Marchand, R. Mishra, A. Dey, A. Bernard, J.-N. Dumez, P. Giraudeau, Ultrafast 2D NMR for the analysis of complex mixtures, *Progress in Nuclear Magnetic Resonance Spectroscopy*. 130–131 (2022) 1–46.

[7] E. Martineau, J.-N. Dumez, P. Giraudeau, Fast quantitative 2D NMR for metabolomics and lipidomics: A tutorial, *Magnetic Resonance in Chemistry*. 58 (2020) 390–403.

[8] G. Bodenhausen, D.J. Ruben, Natural abundance nitrogen-15 NMR by enhanced heteronuclear spectroscopy, *Chemical Physics Letters*. 69 (1980) 185–189.

[9] R. K. Rai, P. Tripathi, N. Sinha. *Anal. Chem.* 2009, 81, 10232–10238

[10] E. Martineau, P. Giraudeau, I. Tea, S. Akoka, Fast and precise quantitative analysis of metabolic mixtures by 2D ¹H INADEQUATE NMR, *Journal of Pharmaceutical and Biomedical Analysis*. 54 (2011) 252–257.

[11] G.A.N. Gowda, F. Tayyari, T. Ye, Y. Suryani, S. Wei, N. Shanaiah, D. Raftery, Quantitative Analysis of Blood Plasma Metabolites Using Isotope Enhanced NMR Methods, *Anal. Chem.* 82 (2010) 8983–8990.

- [12] F. Hu, K. Furihata, Y. Kato, M. Tanokura, Nondestructive Quantification of Organic Compounds in Whole Milk without Pretreatment by Two-Dimensional NMR Spectroscopy, *J. Agric. Food Chem.* 55 (2007) 4307–4311.
- [13] A. Bax, R. Freeman. *J. Magn. Reson.* 1981, 44, 542–561
- [14] G.A. Morris, R. Freeman, Enhancement of nuclear magnetic resonance signals by polarization transfer, *J. Am. Chem. Soc.* 101 (1979) 760–762.
- [15] S. Heikkinen, M.M. Toikka, P.T. Karhunen, I.A. Kilpeläinen, Quantitative 2D HSQC (Q-HSQC) via Suppression of J-Dependence of Polarization Transfer in NMR Spectroscopy: Application to Wood Lignin, *J. Am. Chem. Soc.* 125 (2003) 4362–4367.
- [16] H. Koskela, O. Heikkilä, I. Kilpeläinen, S. Heikkinen, Quantitative two-dimensional HSQC experiment for high magnetic field NMR spectrometers, *Journal of Magnetic Resonance.* 202 (2010) 24–33.
- [17] H. Koskela, Chapter 1 quantitative 2D NMR studies, *Annu. Rep. NMR Spectrosc.* 66 (2009) 1.
- [18] D.J. Peterson, N.M. Loening, QQ-HSQC: a quick, quantitative heteronuclear correlation experiment for NMR spectroscopy, *Magnetic Resonance in Chemistry.* 45 (2007) 937–941.
- [19] H. Koskela, I. Kilpeläinen, S. Heikkinen. *J. Magn. Reson.* 2005, 174,237–244.

- [20] L. Zhang, G. Gellerstedt, Quantitative 2D HSQC NMR determination of polymer structures by selecting suitable internal standard references, *Magnetic Resonance in Chemistry*. 45 (2007) 37–45.
- [21] V. Mäkelä, J. Helminen, I. Kilpeläinen, S. Heikkinen, Quantitative, equal carbon response HSQC experiment, QEC-HSQC, *J. Magn. Reson.* 271 (2016) 34.
- [22] T. Wang, Q. Liu, M. Wang, J. Zhou, M. Yang, G. Chen, F. Tang, E. Hatzakis, L. Zhang, Quantitative Measurement of a Chiral Drug in a Complex Matrix: A J-Compensated Quantitative HSQC NMR Method, *Anal. Chem.* 92 (2020) 3636–3642.
- [23] L. Mauri, G. Boccardi, G. Torri, M. Karfunkle, E. Macchi, L. Muzi, D. Keire, M. Guerrini, Qualification of HSQC methods for quantitative composition of heparin and low molecular weight heparins, *Journal of Pharmaceutical and Biomedical Analysis*. 136 (2017) 92–105.
- [24] G.L. Sasaki, M. Guerrini, R.V. Serrato, A.P. Santana Filho, J. Carlotto, F. Simas-Tosin, T.R. Cipriani, M. Iacomini, G. Torri, P.A.J. Gorin, Monosaccharide composition of glycans based on Q-HSQC NMR, *Carbohydrate Polymers*. 104 (2014) 34–41.
- [25] L. Castañar, E. Sistaré, A. Virgili, R.T. Williamson, T. Parella, Suppression of phase and amplitude J(HH) modulations in HSQC experiments, *Magn. Reson. Chem.* 53 (2015) 115.
- [26] L. Paudel, R.W. Adams, P. Király, J.A. Aguilar, M. Foroozandeh, M.J. Cliff, M. Nilsson, P. Sándor, J.P. Waltho, G.A. Morris, Simultaneously Enhancing Spectral Resolution and Sensitivity in Heteronuclear Correlation NMR Spectroscopy, *Angewandte Chemie International Edition*. 52 (2013) 11616–11619.

- [27] S. Khlifi, C. Mauve, N. Glab, F. Gilard, M. Hodges, G. Mouille, J. Farjon, An easier analysis of complex mixtures with highly resolved and sensitivity enhanced 2D quantitative NMR: application to tracking sugar phosphates in plants, *Anal. Methods*. 9 (2017) 2328–2333.
- [28] J. Farjon, C. Milande, E. Martineau, S. Akoka, P. Giraudeau, The FAQUIRE Approach: FAst, QUantitative, hIghly Resolved and sEnsitivity Enhanced 1H, 13C Data, *Anal. Chem.* 90 (2018) 1845–1851.
- [29] D. Jeannerat, Rapid Multidimensional NMR: High Resolution by Spectral Aliasing, in: *EMagRes*, John Wiley & Sons, Ltd, 2011.
- [30] S. Macura, Accelerated Multidimensional NMR Data Acquisition by Varying the Pulse Sequence Repetition Time, *J. Am. Chem. Soc.* 131 (2009) 9606–960
- [31] P.M. Le, C. Milande, E. Martineau, P. Giraudeau, J. Farjon, Quantification of natural products in herbal supplements: A combined NMR approach applied on goldenseal, *Journal of Pharmaceutical and Biomedical Analysis*. 165 (2019) 155–161.
- [32] K. Hu, W.M. Westler, J.L. Markley, Simultaneous Quantification and Identification of Individual Chemicals in Metabolite Mixtures by Two-Dimensional Extrapolated Time-Zero 1H–13C HSQC (HSQC0), *J. Am. Chem. Soc.* 133 (2011) 1662–1665.
- [33] Hu, K.; Ellinger, J. J.; Chylla, R. A.; Markley, J. L. Measurement of Absolute Concentrations of Individual Compounds in Metabolite Mixtures by Gradient-Selective Time-Zero 1H–13C HSQC with Two Concentration References and Fast Maximum Likelihood Reconstruction Analysis. *Anal. Chem.* **2011**, 83 (24), 9352–9360.

- [34] F. Fardus-Reid, J. Warren, A. Le Gresley, Validating heteronuclear 2D quantitative NMR, *Anal. Methods*. 8 (2016) 2013–2019.
- [35] K. Hu, T.P. Wyche, T.S. Bugni, J.L. Markley, Selective Quantification by 2D HSQC0 Spectroscopy of Thiocoraline in an Extract from a Sponge-Derived *Verrucospora* sp., *J. Nat. Prod.* 74 (2011) 2295–2298.
- [36] S. Halouska, R.J. Fenton, D.K. Zinniel, D.D. Marshall, R.G. Barletta, R. Powers, Metabolomics Analysis Identifies d-Alanine-d-Alanine Ligase as the Primary Lethal Target of d-Cycloserine in *Mycobacteria*, *J. Proteome Res.* 13 (2014) 1065–1076.
- [37] M. Talebi Amiri, S. Bertella, Y.M. Questell-Santiago, J.S. Luterbacher, Establishing lignin structure-upgradeability relationships using quantitative ¹H–¹³C heteronuclear single quantum coherence nuclear magnetic resonance (HSQC-NMR) spectroscopy, *Chem. Sci.* 10 (2019) 8135–8142.
- [38] S. Hayashi, D. Kohda, The time-zero HSQC method improves the linear free energy relationship of a polypeptide chain through the accurate measurement of residue-specific equilibrium constants, *Journal of Biomolecular NMR*. 76 (2022) 87–94.
- [39] M. Sette, H. Lange, C. Crestini, QUANTITATIVE HSQC ANALYSES OF LIGNIN: A PRACTICAL COMPARISON, *Computational and Structural Biotechnology Journal*. 6 (2013) e201303016.
- [40] A. Weitzel, C. Samol, P.J. Oefner, W. Gronwald, Robust Metabolite Quantification from J-Compensated 2D ¹H-¹³C-HSQC Experiments, *Metabolites*. 10 (2020).

- [41] A. Bodor, J.D. Haller, C. Bougouchtoui, F.-X. Theillet, L. Nyitray, B. Luy, Power of Pure Shift $H\alpha Ca$ Correlations: A Way to Characterize Biomolecules under Physiological Conditions, *Anal. Chem.* 92 (2020) 12423–12428.
- [42] S. Ehni, M.R.M. Koos, T. Reinsperger, J.D. Haller, D.L. Goodwin, B. Luy, Concurrent J-evolving refocusing pulses, *Journal of Magnetic Resonance.* 336 (2022) 107152.
- [43] D. Gołowicz, M. Urbańczyk, A. Shchukina, K. Kazimierczuk, SCoT: Swept coherence transfer for quantitative heteronuclear 2D NMR, *Journal of Magnetic Resonance.* 294 (2018) 1–6.
- [44] T. Parella, Towards perfect NMR: Spin-echo versus perfect-echo building blocks, *Magnetic Resonance in Chemistry.* 57 (2019) 13–29.
- [45] B. Baishya, R. Verma, R. Parihar, Spatially encoded polarization transfer for improving the quantitative aspect of $1H-13C$ HSQC, *Journal of Magnetic Resonance Open.* 12–13 (2022) 100063.
- [46] H. Okamura, H. Nishimura, T. Nagata, T. Kigawa, T. Watanabe, M. Katahira, Accurate and molecular-size-tolerant NMR quantitation of diverse components in solution, *Scientific Reports.* 6 (2016) 21742.
- [47] J. Farjon, How to face the low intrinsic sensitivity of 2D heteronuclear NMR with fast repetition techniques: go faster to go higher!, *Magnetic Resonance in Chemistry.* 55 (2017) 883–892.
- [48] N. Prosa, M.-C. Scherrmann, D. Merlet, J. Farjon, SENSASS NMR: New NMR techniques for enhancing the sensitivity and the spectral resolution of polymer supported chemicals, *Journal of Magnetic Resonance.* 237 (2013) 63–72.

- [49] D. Schulze-Süninghausen, J. Becker, B. Luy, Rapid Heteronuclear Single Quantum Correlation NMR Spectra at Natural Abundance, *J. Am. Chem. Soc.* 136 (2014) 1242–1245.
- [50] K. Kazimierczuk, V. Yu. Orekhov, Accelerated NMR Spectroscopy by Using Compressed Sensing, *Angewandte Chemie International Edition.* 50 (2011) 5556–5559.
- [51] D. Gołowicz, P. Kasprzak, V. Orekhov, K. Kazimierczuk, Fast time-resolved NMR with non-uniform sampling, *Progress in Nuclear Magnetic Resonance Spectroscopy.* 116 (2020) 40–55.
- [52] D. Canet, J.-C. Boubel, E. Canet Soulas, *La RMN, Concepts, méthodes et applications*, Dunod ed, Paris, 2002.
- [53] R. Freeman, *A Handbook of Nuclear Magnetic Resonance*, Longman ed, Harlow, 1998
- [54] E. Martineau, P. Giraudeau, I. Tea, S. Akoka, Fast and precise quantitative analysis of metabolic mixtures by 2D ¹H INADEQUATE NMR, *Journal of Pharmaceutical and Biomedical Analysis.* 54 (2011) 252–257.
- [55] M.I. Miller, A.S. Greene, Maximum-likelihood estimation for nuclear magnetic resonance spectroscopy, *Journal of Magnetic Resonance* (1969). 83 (1989) 525–548.
- [56] R.A. de Graaf, G.M.I. Chowdhury, K.L. Behar, Quantification of High-Resolution ¹H NMR Spectra from Rat Brain Extracts, *Anal. Chem.* 83 (2011) 216–224.

- [57] R.A. Chylla, B.F. Volkman, J.L. Markley, Practical Model Fitting Approaches to the Direct Extraction of NMR Parameters Simultaneously from All Dimensions of Multidimensional NMR Spectra, *Journal of Biomolecular NMR*. 12 (1998) 277–297.
- [58] R.A. Chylla, K. Hu, J.J. Ellinger, J.L. Markley, Deconvolution of Two-Dimensional NMR Spectra by Fast Maximum Likelihood Reconstruction: Application to Quantitative Metabolomics, *Anal. Chem.* 83 (2011) 4871–4880.
- [59] T. Castaing-Cordier, D. Bouillaud, J. Farjon, P. Giraudeau, Chapter Four - Recent advances in benchtop NMR spectroscopy and its applications, in: G.A. Webb (Ed.), *Annual Reports on NMR Spectroscopy*, Academic Press, 2021: pp. 191–258.
- [60] B. Gouilleux, J. Marchand, B. Charrier, G.S. Remaud, P. Giraudeau, High-throughput authentication of edible oils with benchtop Ultrafast 2D NMR, *Food Chemistry*. 244 (2018) 153–158.
- [61] T. Castaing-Cordier, A. Benavides Restrepo, D. Dubois, V. Ladroue, F. Besacier, A. Buleté, C. Charvoz, A. Goupille, D. Jacquemin, P. Giraudeau, J. Farjon, Characterization of new psychoactive substances by integrating benchtop NMR to multi-technique databases, *Drug Testing and Analysis*. n/a (2022).
- [62] D. Bouillaud, D. Drouin, B. Charrier, C. Jacquemmoz, J. Farjon, P. Giraudeau, O. Gonçalves, Using benchtop NMR spectroscopy as an online non-invasive in vivo lipid sensor for microalgae cultivated in photobioreactors, *Process Biochemistry*. 93 (2020) 63–68.
- [63] B.C. Percival, M. Grootveld, M. Gibson, Y. Osman, M. Molinari, F. Jafari, T. Sahota, M. Martin, F. Casanova, M.L. Mather, M. Edgar, J. Masania, P.B. Wilson, Low-Field, Benchtop NMR Spectroscopy as a Potential Tool for Point-of-Care

Diagnostics of Metabolic Conditions: Validation, Protocols and Computational Models, High-Throughput. 8 (2019).

[64] B. Plainchont, P. Berruyer, J.-N. Dumez, S. Jannin, P. Giraudeau, Dynamic Nuclear Polarization Opens New Perspectives for NMR Spectroscopy in Analytical Chemistry, *Anal. Chem.* 90 (2018) 3639–3650.

[65] A. Bornet, M. Maucourt, C. Deborde, D. Jacob, J. Milani, B. Vuichoud, X. Ji, J.-N. Dumez, A. Moing, G. Bodenhausen, S. Jannin, P. Giraudeau, Highly Repeatable Dissolution Dynamic Nuclear Polarization for Heteronuclear NMR Metabolomics, *Anal. Chem.* 88 (2016) 6179–6183.

[66] A. Dey, B. Charrier, K. Lemaitre, V. Ribay, D. Eshchenko, M. Schnell, R. Melzi, Q. Stern, S.F. Cousin, J.G. Kempf, S. Jannin, J.-N. Dumez, P. Giraudeau, Fine optimization of a dissolution dynamic nuclear polarization experimental setting for ^{13}C NMR of metabolic samples, *Magnetic Resonance*. 3 (2022) 183–202.

[67] R.W. Adams, J.A. Aguilar, K.D. Atkinson, M.J. Cowley, P.I.P. Elliott, S.B. Duckett, G.G.R. Green, I.G. Khazal, J. López-Serrano, D.C. Williamson, Reversible Interactions with para-Hydrogen Enhance NMR Sensitivity by Polarization Transfer, *Science*. 323 (2009) 1708–1711.

[68] N. Eshuis, N. Hermkens, B.J.A. van Weerdenburg, M.C. Feiters, F.P.J.T. Rutjes, S.S. Wijmenga, M. Tessari, Toward Nanomolar Detection by NMR Through SABRE Hyperpolarization, *J. Am. Chem. Soc.* 136 (2014) 2695–2698.

[69] N. Eshuis, R.L.E.G. Aspers, B.J.A. vanWeerdenburg, M.C. Feiters, F.P.J.T. Rutjes, S.S. Wijmenga, M. Tessari, 2D NMR Trace Analysis by Continuous Hyperpolarization at High Magnetic Field, *Angewandte Chemie International Edition*. 54 (2015) 14527–14530.

[70] N. Reimets, K. Ausmees, S. Vija, I. Reile, Developing Analytical Applications for Parahydrogen Hyperpolarization: Urinary Elimination Pharmacokinetics of Nicotine, *Anal. Chem.* 93 (2021) 9480–9485.

[71] K. Ausmees, N. Reimets, I. Reile, Parahydrogen hyperpolarization of minimally altered urine samples for sensitivity enhanced NMR metabolomics, *Chem. Commun.* 58 (2022) 463–466.

[72] Eills, J.; Hale, W.; Utz, M. Synergies between Hyperpolarized NMR and Microfluidics: A Review. *Progress in Nuclear Magnetic Resonance Spectroscopy* **2022**, *128*, 44–69.