



HAL
open science

Extended-depth of field random illumination microscopy, EDF-RIM, provides super-resolved projective imaging

Lorry Mazzella, Thomas Mangeat, Guillaume Giroussens, Benoit Rogez, Hao Li, Justine Creff, Mehdi Saadaoui, Carla Martins, Simon Labouesse, Jérôme Idier, et al.

► **To cite this version:**

Lorry Mazzella, Thomas Mangeat, Guillaume Giroussens, Benoit Rogez, Hao Li, et al.. Extended-depth of field random illumination microscopy, EDF-RIM, provides super-resolved projective imaging. 2023. hal-04269066

HAL Id: hal-04269066

<https://hal.science/hal-04269066>

Preprint submitted on 3 Nov 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Extended-depth of field random illumination microscopy, EDF-RIM, provides super-resolved projective imaging

Lorry Mazzella^{1,+}, Thomas Mangeat^{2,+}, Guillaume Giroussens¹, Benoit Rogez¹, Hao Li³, Justine Creff³, Mehdi Saadaoui⁴, Carla Martins¹, Simon Labouesse², Jérôme Idier⁵, Frédéric Galland¹, Marc Allain¹, Anne Sentenac^{1,*}, and Loïc LeGoff^{1,*}

¹Aix Marseille Univ, CNRS, Centrale Marseille, Institut Fresnel UMR7249, Turing Center for Living Systems, Marseille, France

²LITC Core Facility, Centre de Biologie Integrative (CBI), CNRS, Université de Toulouse, UT3, Toulouse, France

³MCD, Centre de Biologie Intégrative (CBI), CNRS, Université de Toulouse, UT3, Toulouse, France

⁴Aix Marseille Univ, CNRS, IBDM UMR7288, Turing Centre for Living systems, Marseille, France

⁵LS2N, CNRS UMR 6004, F44321 Nantes Cedex 3, France

⁺These authors contributed equally to this work

^{*}Authors for correspondence : anne.sentenac@fresnel.fr, loic.le-goff@univ-amu.fr

ABSTRACT

The ultimate aim of fluorescence microscopy is to achieve high-resolution imaging of increasingly larger biological samples. Extended depth of field presents a potential solution to accelerate imaging of large samples when compression of information along the optical axis is not detrimental to the interpretation of images. We have implemented an Extended Depth of Field (EDF) approach in a Random Illumination Microscope (RIM). RIM uses multiple speckled illuminations and variance data processing to double the resolution. It is particularly adapted to the imaging of thick samples as it does not require the knowledge of illumination patterns. We demonstrate highly-resolved projective images of biological tissues and cells.

1 Introduction

Fluorescence imaging has transformed biology by facilitating the observation of molecular events within live cells and tissues. One challenge is now to observe large samples (such as embryos) at the highest possible resolution¹. Two major difficulties are faced when imaging large three-dimensional (3D) samples. First, the resolution over a wide field of view is challenged by optical aberrations. Second, imaging a large volume is slow, as it requires the sequential acquisition of many planes.

Among the various methods for achieving super-resolution, structured illumination microscopy (SIM) presents a favorable balance between spatio-temporal resolutions and phototoxicity²⁻⁸. In SIM, the super-resolved image is computationally generated⁹ using multiple low-resolution images obtained with a periodic illumination pattern that is translated and rotated. SIM is amenable for live imaging and has been implemented in a number of modalities: light sheet microscopy¹⁰⁻¹², total internal reflection microscopy¹³⁻¹⁷, 3D-microscopy^{18,19}. However, SIM data processing requires the knowledge of the illumination pattern with great accuracy^{20,21}, which makes the technique sensitive to aberrations and undermines its use in tissues. An active field of research, therefore, aims at improving the characterization of illumination patterns^{9,20,22}, and to develop adaptive optics approaches^{23,24}, or post-processing^{25,26}, to correct potential artifacts. Another strategy was pursued in a variant of SIM called Random Illumination Microscopy (RIM). RIM replaces the periodic patterns with random speckled illuminations. It uses a reconstruction procedure based on the second-order statistics of the raw images which avoids the knowledge of the different illumination patterns. RIM has been demonstrated to be robust against aberrations and to achieve a two-fold improvement in lateral resolution compared to standard microscopy, while also providing strong sectioning in the axial direction²⁷.

If some SIM approaches are becoming amenable to tissue imaging, the need to acquire multiple images of the sample under different illuminations at each observation plane still hinders the widespread use of SIM for large volume imaging. When the information from the biological sample can be compressed along the optical axis, extended depth of field (EDF) imaging becomes an interesting option for accelerating the imaging process through optical projection. Two main approaches are used for achieving extended depth of field. The first approach involves rapid scanning of the imaged plane with a tunable lens^{28,29}, acoustic lens³⁰ or a deformable mirror^{31,32}. The second approach involves point spread function (PSF) engineering. Excitation PSF in scanning microscopes can be extended using, for example, Bessel beams³³⁻³⁶. Detection PSF can also be elongated

using phase masks³⁷. EDF imaging is attractive because all the objects scattered along the optical axis appear in focus. On the other hand, because of the image projection, their footprint in focus is superimposed on all their defocused marks. As a result, EDF images are often affected by an important background and its related Poisson noise which dims the contrast of the illumination patterns. This loss of contrast together with the distortion of the illumination pattern induced by the possibly thick sample has hampered the implementation of classical SIM in an extended depth of field configuration. As a result, we are not aware of any super-resolved extended depth microscope adapted to live biological samples.

Here, for the first time, we combine a super-resolved technique with extended depth capacity using the RIM modality. The technique consists in illuminating the sample with multiple speckled illuminations and collecting light with a detection scheme featuring EDF capacity. The super-resolved image is formed from the raw images using the second-order statistics of the images and speckles. We demonstrate that RIM reconstruction process efficiently removes the background noise typically present in EDF images and improves the resolution beyond the diffraction limit in the lateral direction. We also demonstrate a strategy to partially compensate for the lost information along the optical axis through topographical estimation of the tissue on which the EDF images can be projected.

2 Theory

2.1 Introduction to Random illumination Microscopy (RIM)

Introducing ρ the sample fluorescence density, h the PSF of the optical system and S the illumination intensity, the intensity recorded at the camera plane of a fluorescence microscope (assuming a magnification of one) reads³⁸

$$I(\mathbf{r}_\perp, z_f) = \int \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') h(\mathbf{r}_\perp - \mathbf{r}'_\perp, z_f - z') d\mathbf{r}'_\perp dz' \quad (1)$$

where z_f indicates the position of the focal plane, and the subscript \perp indicates transverse coordinates.

In Random Illumination Microscopy (RIM), a series of random speckle illuminations S_1, \dots, S_M are generated to provide a stack of speckled images. Each image has a resolution limit strictly enforced by the diffraction of light but the speckle illuminations, like the periodic light grid of SIM², open a way toward super-resolution. It was demonstrated that a two-fold improvement in resolution compared to standard fluorescence microscopy could be achieved by processing the variance of the speckled images²⁷. The key point of RIM is that the variance of the speckled images depends only on the observation point spread function and on the speckle autocorrelation, which are known functions. Up to now, RIM reconstruction algorithm has been implemented in the two-dimensional (2D) case²⁷. The sample was assumed to be infinitely thin at the focal plane, $\rho(\mathbf{r}'_\perp, z') = \rho(\mathbf{r}'_\perp) \delta(z_f - z')$ and the out-of-focus fluorescence reaching the camera was considered as a background noise³⁹. The super-resolved reconstruction of the fluorescence density was achieved by estimating iteratively the sample so as to minimize the distance between the *experimental* standard deviation of the speckled images and its theoretical counterpart (see Methods for further details),

$$\hat{\rho}(\mathbf{r}_\perp) = \arg \min_{\rho} \|\hat{\sigma}(\mathbf{r}_\perp) - \sigma(\mathbf{r}_\perp; \rho)\|^2 \quad (2)$$

with $\hat{\sigma}(\mathbf{r}_\perp)$ the experimental standard-deviation and $\sigma(\mathbf{r}_\perp; \rho)$ the *theoretical* standard deviation derived from the expected variance that reads

$$\sigma^2(\mathbf{r}_\perp; \rho) = \iint \rho(\mathbf{r}'_\perp) \rho(\mathbf{r}''_\perp) \Gamma_S(\mathbf{r}''_\perp - \mathbf{r}'_\perp, 0) h(\mathbf{r}_\perp - \mathbf{r}'_\perp, 0) h(\mathbf{r}_\perp - \mathbf{r}''_\perp, 0) d\mathbf{r}'_\perp d\mathbf{r}''_\perp \quad (3)$$

with $\Gamma_S(\mathbf{r}''_\perp - \mathbf{r}'_\perp, 0) = \langle S(\mathbf{r}'_\perp, z_f) S(\mathbf{r}''_\perp, z_f) \rangle$ the autocorrelation of the speckle illuminations at the focal plane. Note that the 2D RIM reconstruction algorithm was able to form three-dimensional images of a sample by simply stacking 2D reconstructions obtained independently at the different focal planes.

2.2 Introduction to Extended Depth of Field imaging

For Extended-Depth of Field (EDF) imaging, it is necessary to consider ρ , h , and S as 3D functions. More precisely, the EDF image $I_\perp(\mathbf{r}_\perp) = \int I(\mathbf{r}_\perp, z_f) dz_f$ reads,

$$I_\perp(\mathbf{r}_\perp) = \int \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) d\mathbf{r}'_\perp dz' \quad (4)$$

with $h_\perp(\mathbf{r}_\perp) = \int h(\mathbf{r}, z) dz$.

We will consider two different implementations of RIM in an EDF configuration. The first one relies on the modification of the experimental system in order to generate speckles that are invariant along the optical axis. The second one uses standard three-dimensional speckles and does not require any modification of the experimental system but is only valid for a smooth surface-distributed sample.

2.3 EDF-RIM using speckles invariant along the optical axis

We first assume that the speckles are invariant along the optical axis, $S(\mathbf{r}_\perp, z) = S_B(\mathbf{r}_\perp)$. Such columnar speckles can be achieved experimentally by putting an annulus mask in the pupil plane to generate Bessel-type speckled illuminations⁴⁰. In this case (see Appendix A), the theoretical expressions for the raw EDF image and its variance read,

$$I_B(\mathbf{r}_\perp) = \int \rho_\perp(\mathbf{r}'_\perp) S_B(\mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) d\mathbf{r}'_\perp \quad (5)$$

$$\sigma_B^2(\mathbf{r}_\perp; \rho_\perp) = \iint \rho_\perp(\mathbf{r}'_\perp) \rho_\perp(\mathbf{r}''_\perp) \Gamma_B(\mathbf{r}''_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) d\mathbf{r}'_\perp d\mathbf{r}''_\perp \quad (6)$$

where $\rho_\perp(\mathbf{r}_\perp) := \int \rho(\mathbf{r}_\perp, z) dz$ is the projection of the fluorescence density onto the transverse plane.

In Eq. 6, we have introduced $\Gamma_B(\mathbf{r}''_\perp - \mathbf{r}'_\perp) = \langle S_B(\mathbf{r}'_\perp) S_B(\mathbf{r}''_\perp) \rangle$ the autocorrelation of the Bessel-speckle, which is invariant along z , *i.e.*. In this case the EDF theoretical variance, Eq. 6, has the same structure as that derived in 2D RIM in Eq. 3. Thus we can leverage the RIM reconstruction strategy to provide a super-resolved EDF image of the projected sample ρ_\perp , by changing h into h_\perp .

Bessel-type speckles provide a clear and canonical framework to derive an EDF-RIM strategy. Obviously, whether EDF-RIM can be produced with conventional 3D-speckles is a pivotal question.

EDF-RIM using standard speckles for observing fluorescent surfaces

To what extent can EDF-RIM be performed with conventional 3D speckles? We recall that the EDF image depends on $\int \rho(\mathbf{r}_\perp, z) \times S(\mathbf{r}_\perp, z) dz$ so that the projection of the sample and that of the illumination cannot be separated. However, when the sample is a surface, (more generally a 2D manifold), EDF-RIM can be used under some approximations. We consider a sample in which the fluorophores are distributed along a surface parametrized by the function $Z(\mathbf{r}_\perp)$:

$$\rho(\mathbf{r}_\perp, z) = \rho_\perp(\mathbf{r}_\perp) \delta(z - Z(\mathbf{r}_\perp)). \quad (7)$$

Furthermore, we assume that the fluorophores distribute on a smooth topography, implying that the surface does not vary much in Z over lateral distances of the order of $\frac{\lambda}{2\text{NA}}$ where λ is the wavelength of the illuminating light,

$$\|\mathbf{r}'_\perp - \mathbf{r}_\perp\| < \frac{\lambda}{2\text{NA}} \implies Z(\mathbf{r}'_\perp) \approx Z(\mathbf{r}_\perp). \quad (8)$$

With these simplifying assumptions, we demonstrate in Appendix A that the EDF image variance also follows a 2D canonical form :

$$\sigma_{\text{EDF}}^2(\mathbf{r}_\perp) \approx \iint \rho_\perp(\mathbf{r}'_\perp) \rho_\perp(\mathbf{r}''_\perp) \Gamma_{\text{EDF}}(\mathbf{r}''_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) d\mathbf{r}'_\perp d\mathbf{r}''_\perp \quad (9)$$

where $\Gamma_{\text{EDF}}(\mathbf{r}_\perp)$ is the ‘‘slice’’ at $z = 0$ of the 3D autocorrelation of the speckle illuminations, *i.e.*, $\Gamma_{\text{EDF}}(\mathbf{r}_\perp) := \Gamma_S(\mathbf{r}_\perp, z = 0)$. To conclude, we have shown that the variance of the extended depth RIM images can be simplified into the classical 2D-RIM expression of eq. 3 for any type of sample, by using columnar speckles (eq. 6), and for sparse and smoothed samples, when using regular 3D speckles (eq. 9).

In Appendix B, we simulate EDF-RIM image formation to explore the implications of our smooth surface approximation. We simulate EDF-RIM imaging of an object with different topographies: i- a flat object for which eq. 6 is correct with no approximation; ii- a smoothly varying surface, which satisfies the conditions for the EDF variance to align with the canonical 2D-RIM expression (eq. 9); iii- a random topography, which does not satisfy the necessary conditions for the approximation. These simulations show that with a 3D propagating speckle, EDF-RIM does indeed provide super-resolution with the flat and smoothly varying topographies (conditions *i* and *ii*, see Appendix B and Fig. S1). However, with the random topography (condition *iii*), reconstruction is thwarted by amplified noise at small regularization parameter of the inversion process. Simulating EDF-RIM with a Bessel speckle, which is invariant along the optical axis, we were able to achieve super-resolved images for all samples, including the random z configuration, for every value of the regularization parameter tested (Fig. S1B, lower panels). Our simulations thus confirm that 3D-speckles can be used to image sparse samples along the optical axis, such as surfaces, and Bessel speckles can be used for arbitrary samples.

In the following section, we demonstrate an experimental implementation of EDF-RIM. We place ourselves in the sparse object approximation, which allows us to use a simple 3D-speckle illumination for our extended-depth imaging. To reach an extended depth detection, we either use a phase mask for a three-fold extension of depth ($\sim 1.5\mu\text{m}$) or a focal sweep at much larger depth ($\sim 20\mu\text{m}$ and possibly more).

3 Experimental implementation

Principle

To perform EDF-RIM in the context of the aforementioned approximation, we use fully developed speckles as 3D illuminations. The speckled illumination is then combined with a detection scheme with an extended detection PSF. This is experimentally realized through two complementary strategies: i- for a moderate increase of depth, we use a phase mask that extends the PSF approximately three-fold; ii- for larger depths, we use an electrically tunable lens (ETL) to rapidly sweep the imaged plane in one camera integration time.

The two implementations share similar optical layouts, where a telecentric relay in the detection branch of the microscope is used to place the mask or ETL in a plane conjugated to the pupil of the objective lens (see Fig. 1A and Methods section). In the focal sweep configuration, we send ramps of electrical current to the ETL within a single camera exposure, effectively summing the fluorescence signal over multiple planes to achieve extended depth (Fig. 1B). An advantage of using the mask over the ETL is that no such z-scan is necessary to perform volumetric detection, making the approach potentially faster, and more efficient photon budget-wise. Masks, however, can only be used for shallow depths ($\sim 1.5\mu\text{m}$), while the focal sweep can be used for much larger depths ($20\mu\text{m}$ or more).

Figure 1C,D show respectively (x,y) and (x,z) sections of an excitation speckle. Figure 1E shows the EDF-PSF, $h_{\perp}(\mathbf{r}_{\perp})$. Insets in Fig. 1C,D provide respectively the autocorrelations $\Gamma_{\text{EDF}}(\mathbf{r}_{\perp}) := \Gamma_S(x,y,z=0)$ and $\Gamma_S(x,0,z)$.

We determine the experimental image-variance from a set of 100-200 EDF-images captured under different speckled illuminations. The super-resolved reconstruction is then obtained through a variance matching algorithm (eq. 2), as explained in the Theory and Methods sections.




Figure 1. (A) Optical layout of the experimental system. Speckles are generated either with a diffuser or a spatial light modulator. The detection branch comprises a 4f optical relay (f_{R1} , f_{R2}) which includes either a phase mask to extend the detection PSF or an electrically tunable lens to perform a rapid focal sweep –both placed in a pupil plane. (B) Temporal scheme of an acquisition sequence. the top line represents the z-plane conjugated to the camera, it is either rapidly scanned (red line) in one integration time when using the ETL or kept at a constant range when using the phase mask (yellow band). (C,D) Example of an illumination speckle in the x,y and x,z planes. The insets display the speckle autocorrelation function $\Gamma_{\text{EDF}}(\mathbf{r}_{\perp}) := \Gamma_S(x,y,0)$ and $\Gamma_S(x,0,z)$. (E) Comparison between the 2D PSF, $h_{2D}(r)$, of the wide field microscope and the EDF-PSF, $h_{\perp}(\mathbf{r}_{\perp})$, which is obtained from the integral of the 3D-PSF along the z-axis.

In subsequent sections, we investigate the use of EDF-RIM for super-resolved cell imaging. For molecular cell imaging of thin samples at the highest resolution, we use a stair-step mask which extends three-fold the detection PSF of the microscope. For thicker and larger samples, we use the focal sweep in the context of tissue imaging.

EDF-RIM with a phase mask for high-resolution molecular tracking

In the first approach, we insert a stair-step phase mask in a pupil plane of the detection branch of the microscope (Fig. 2A). In such a phase mask, the contributions from each concentric sub-aperture -which have axially extended PSF- add up incoherently

into a combined PSF which is also axially extended but remains narrow laterally³⁷. Figure 2B, represents the detection PSF with and without the phase mask, illustrating the axial extension in the presence of the mask.

With an approximately three-fold increase in the axial extent of the PSF ($\sim 1.5\mu m$) structures from the biological sample can now be captured in a single plane, where several planes are necessary with conventional RIM. Figure 2C,C' compares a single plane image of desmosomes from the mouse intestinal epithelium, with (left) and without the mask (right). Owing to the slightly curved topography of the sample, some structures visible with the EDF-RIM are clearly missed with the conventional 2D-RIM. The insets in between 2C and C' provide close up views to inspect image quality. Standard EDF-wide field fluorescence is marred by background signal and poor resolution. EDF-RIM and 2D-RIM yield very similar lateral resolution.

As further demonstration of the resolution capacity of EDF-RIM, we image Myosin-II in the *Drosophila* egg chamber. Myosin-II plays an important role in the organization of the actin cytoskeleton and tensile force generation⁴¹. In the *Drosophila* egg chamber, follicle cells assemble polarized stress fibers on their basal surface –structures which are rich in Myosin-II⁴². Figure 2D shows a snapshot of several such follicle cells using a Cherry protein fusion on the Myosin-II regulatory light chain (the *sqh* gene in *Drosophila*). The inset to the right provides a close-up. Supplementary movie-1 provides a time sequence of an equivalent acquisition (Myosin-II:Cherry protein fusion) where pulses of Myosin-II can be observed. Supplementary movie-2 shows a similar acquisition at an earlier developmental stage, as the Myosin-II-rich stress fibers are starting to assemble. In all Myosin-II acquisitions, the signal appears granular at a fine scale owing to our ability to resolve individual mini-filaments of Myosin-II. Indeed, as schematized in the cartoon to the right of Fig. 2D, Myosin-II often assembles into mini-filaments, unitary motile multimers of bi-polar Myosin-II. The fluorescence signal of a mini-filament looks like two bright fluorescent points rigidly linked at a distance of $\sim 300nm$. A high numerical aperture objective yields a resolution of less than 100nm with RIM²⁷. Combined with a temporal resolution of $\Delta t = 960ms$, we could thus spatio-temporally resolve movements of the mini-filaments throughout the field of view, observing, for example, their rotation (sequence in Fig. 2E).

EDF-RIM with a swept remote focus for large-volume imaging

With the stair-step phase mask, we could increase the depth of imaging up to $\sim 1.5\mu m$, which can be used for small and thin samples at very high resolution. For larger volumes, we propose as an alternative to rapidly sweep the imaged plane in one camera integration, to reach a depth of tens of micrometers. In the optical layout, the phase mask is simply replaced with an ETL (Fig. 3A). Ramps of currents are then sent to the ETL in one camera integration for extended-depth imaging. To demonstrate the effectiveness of this set-up, we begin by imaging fluorescence from the actin cytoskeleton (Phalloidin-alexa488) of a human cultured cell. In Figure 3B,C, we compare EDF-widefield (Fig. 3B) with EDF-RIM (Fig. 3C) for the same field of view. As expected, the EDF-widefield image is marred by background. In contrast, the EDF-RIM signal provides a sharp image with no background haze. Furthermore, the resolution is noticeably improved with EDF-RIM, as actin filaments from the cell cortex are better discerned. The background reduction and resolution improvement are particularly striking in the close-up views (insets). To evaluate the effectiveness of EDF-RIM in tissue imaging, we imaged tight junctions in developing chick neural tubes (anti-ZO-1 Alexa-488 staining). We performed EDF imaging on a $5.5\mu m$ deep volume (Fig. 3D). A comparison of EDF-wide field with EDF-RIM on a close-up region shows qualitatively the reduced background and increased resolution of the latter (Fig. 3E). Intensity profiles along short lines demonstrate a better separation of nearby cell-interfaces with EDF-RIM (Fig. 3F).

In EDF imaging, information pertaining to the topography of the tissue is lost. However, it may be necessary to retain some level of topographical information, even if not at a super-resolved level. To address this, we reach a tradeoff between speed and resolution and acquire one plane-by-plane scan of the sample volume using one speckled illumination and extract topographical data with it. To achieve this, we use a robust estimation strategy, which can handle noise and background inherent to the 3D data⁴³ (see Methods). The proposed approach primarily involves identifying bright points within the 3D sample and filtering out those that are not part of the most densely populated surface (Fig. 3G). Figure 3H,I shows an orthogonal view and 3D representation of the surface and associated inliers/outliers from the same neural tube sample than Fig. 3D. Once the surface is estimated, we can then proceed with the regular EDF-RIM imaging protocol and ultimately project the EDF-RIM super-resolved image onto this topography for a 3D rendering of the fluorescent signal (Fig. 3J).

4 Discussion

In this work, we combine for the first time SIM with an extended depth-of-field approach. For this, we combined speckled illumination, extended depth of field detection and RIM treatment. We demonstrated two complementary strategies to generate an EDF detection. First, for moderately increased depths (three-fold with respect to 2D-RIM), we used a stair-step phase mask in the detection arm of the microscope to extend the PSF axially while keeping it narrow in the lateral dimension. Then, for larger extent (up to several tens of μm), we used an ETL to perform a rapid focal sweep in one camera integration time.

The strength of EDF-RIM lies in the use of speckle as a 3D illumination function, whose statistics are invariant along the optical axis and insensitive to optical aberrations. Combined with the statistical treatment of the inversion process, this results in an insensitivity to aberrations as the system sweeps the volume of the sample —which would be difficult to achieve in a




Figure 2. EDF-RIM with a phase mask for molecular imaging. (A) Principle: a stair-step phase mask is inserted in a pupil plane in the detection branch of the microscope. (B) Experimental detection PSF of the microscope with and without the phase mask. (C) Comparison of a RIM image of desmosome from the intestinal mouse epithelium with (C) and without (C') the phase mask. The increased depth in (C) captures more structures (on the left and bottom of the image). The insets between (C) and (C') provide the same close-up with different approaches – top: EDF-wide field; middle: EDF-RIM; bottom: 2D-RIM. (D-F) Application to molecular tracking of Myosin-II mini-filaments in the *Drosophila* ovary chamber. (D) Image of Myosin-II:GFP in the basal surface of follicle cells from a *Drosophila* egg chamber. The inset to the right provides a close-up. The signal appears granular at a fine scale owing to the microscope resolving individual mini-filaments of Myosin-II. The schematic on the right shows the molecular structure of a typical mini-filament. The anti-polar arrangement leads to a fluorescent structure consisting of two fluorescent points, rigidly linked and spaced by approximately 300nm. (E) Close-up on a sequence, imaged with a 960 ms interval, showing the rotation of an individual Myosin-II mini-filament. Scales bars=300 nm when not specified.




Figure 3. EDF-RIM with a swept remote focus. (A) Principle: An ETL is placed in a pupil plane of the detection branch of the microscope to sweep the focus. (B,C) Phalloidin-alexa488 labeling of the actin cytoskeleton on a cultured cell imaged with EDF-wide-field (B) and EDF-RIM (C). Insets show a close-up view of the outlined region. (D-J) large-scale EDF imaging with topographical information. (D) Large field of view EDF-RIM of tight junctions from a chick neural tube (ZO-1 staining). (E) Comparison of EDF wide-field and EDF-RIM on a close-up view. (F) Intensity profile along a short line of the close-up. Nearby cell interfaces are discerned only with EDF-RIM. (G) Principle of topographical estimation for epithelia: one slice by slice stack is acquired for one speckled illumination. A consensual inlier/outlier classification of the bright points is then determined with RANSAC piece-wise fits (see methods). Inliers are then interpolated. (H) Orthogonal view of the surface, inliers and outliers for the chick neural tube data. (I) 3D view of surface, inliers and outliers. For sake of clarity, outliers are only shown on half of the surface. (J) The EDF-RIM image projected on the topography.

conventional SIM approach.

Classical extended-depth approaches allow an increase in acquisition speed when imaging thick samples, but they suffer from an increased background signal. In all our samples –both with the focal sweep and phase mask techniques– we demonstrated a considerable reduction in background and an increase in resolution with EDF-RIM (insets of Fig. 2C, Fig. 3.B,C,E,F).

Compared to 2D-RIM, EDF-RIM allows the capture of volumetric features within one acquisition. For flat samples, the moderate increase in depth brought by the phase mask is sufficient to image tight junctions from a segment of mouse intestinal epithelium, where several planes would be required with 2D-RIM (Fig. 2C,C'). Reducing the number of acquired slices allows for faster imaging. We could thus capture the rotational and clustering dynamics of Myosin-II mini-filaments in a *Drosophila* egg chamber. These dynamics would have been difficult to capture with an imaging system three to four times slower. The focal sweep allowed us to image even larger volumes –cells a few microns thick or even tissues over $\sim 20\mu\text{m}$ depth. At such a large scale, EDF approaches may yield an important loss of information on the biological sample under study –especially in the context of morphogenesis, where the global shape of the tissue is an essential feature. In these circumstances, the low-resolution topographical estimation that we propose (Fig. 3G-J), which requires one slice-by-slice acquisition, may represent a satisfactory compromise between speed of acquisition and proper description of the geometrical features of the biological sample.

What are the limits of our current approach? We made the assumption that the fluorophore density is sparse along the optical axis. This allows us to reach a simple model for image formation that we use in the inversion process (see Theory section). A typically good configuration is when fluorophores are distributed over a smoothly varying surface (or several surfaces). Conversely, a randomly distributed topography can lead to a greater sensitivity to noise of the reconstruction at low regularization parameter, as demonstrated in our simulations (Fig. S1B, top-right panels). Experiments, however, tell us that the restriction is not drastic. We could image at high resolution the actin cytoskeleton of an entire cell, which is $\sim 5\mu\text{m}$ thick, with a much-improved contrast and resolution over extended wide field (Fig. 3B,C).

One improvement in future implementations could be the use of an illumination that is truly invariant in z , not just statistically invariant. This could be done using Bessel speckles –achieved experimentally by blocking the light outside an annulus in the pupil plane, and maintaining a random phase within this annulus with a diffuser or spatial light modulator. This would improve the inversion process for samples that are not distributed over a surface (Fig. S1B, bottom-right panels). The approach could also improve resolution owing to the stronger content in high spatial frequencies of Bessel speckles⁴⁰.

To conclude, we have developed the first wide-field structured illumination microscope, which combines super-resolution with an extended depth capacity. For this, we build upon the framework of the random illumination microscope (RIM), combining the intrinsically 3D nature of illumination speckles with extended depth detection. EDF-RIM appears as an attractive alternative to the classical EDF-widefield when a high resolution and low background is necessary, for example in the context of high-resolution molecular tracking and tissue imaging.

Methods

Optical setup

The two approaches have been built on separate setups for sake of convenience, but share a similar optical layout. Essentially, the intermediate image created by the objective and tube lenses is relayed by f_{R1} and f_{R2} , arranged in a $4f$ configuration, onto the camera. An objective and a tube lens create a magnified image of the sample ($M = f_T/f_O$) in an intermediate image plane. The intermediate image is relayed by f_{R1} and f_{R2} , arranged in a $4f$ configuration, onto the camera. The ETL or phase mask are placed in a pupil plane in between f_{R1} and f_{R2} ^{44,45}. Excitation speckles are characterized (correlation function $\Gamma_S(x,y,z)$) by replacing the sample with a 50 nm gold film and using the remote focus to image the speckle in 3D.

Rapid focal sweep set-up: The microscope is custom built, made of optomechanical elements. f_O is an 100x objective (Nikon, N100X-PFO, nominal NA=1.3, effective NA=1.21), f_T, f_{R1}, f_{R2} are thorlabs TTL165-A, AC254-125-A-ML, AC254-125-A-ML. The exact position of the ETL and its associated offset lens (Optotune EL-10-30-Ci Series) is finely adjusted so as to optimize the detection PSF. We found that placing specifically the ETL in the fourier plane and not the offset lens or the mid-point between the offset lens and the ETL was optimal. The lenses f_O, f_T, f_{R1}, f_{R2} are arranged to ensure telecentricity of the imaging path. In the excitation branch of the set-up, light from a 488 nm diode laser (Oxxius, LBX-488-200-CSB), impinging on a diffuser, is relayed to the pupil of the objective by f_{R3} and f_{R4} (Thorlabs LA1986-A-ML and LA1131-A-ML). Successive speckles are triggered through rotation of the diffuser with a rotary stage (2-axis controller 8SMC4-USB-B9-2). The illumination beam is combined with the imaging path with a dichroic mirror (semrock, Di02-R488-25x36). Fluorescence is further filtered with a bandpass filter (Semrock, FF01-525/45-25) and imaged with a camera (Prime BSI Express Scientific CMOS Camera). A Matlab interface controls all instruments of the setup. Hardware synchronization of instruments is done using a National Instrument (NI USB-6351) board, though digital pulses and analog signal sent to the ETL driver.

Phase mask setup: An inverted microscope (TEi Nikon) is used with a 100x TIRF objective (CFI SR APO 100x NA 1.49 Nikon) We use a 488 nm diode laser (Oxxius, LBX-488-200-CSB) for GFP/Alexa-488 and a 561 nm solid-state laser (Oxxius,

LMX-561L-200-COL) for cherry imaging. A fast spatial light phase binary modulator (SLM, QXGA Fourth Dimension) placed in an image plane was illuminated by the collimated 8.8 mm TEM00 beam. Fluorescence, after the dichroic mirror (Semrock DI02-R488-25×3625×36) and Stop-Line notch filter (Semrock NF03-405/488/561/635E-25) is collected by an sCMOS camera (OrcaFlash fusion). The stair-step phase mask (DoubleHelix, PNDT6) is placed in a pupil plane of the optical really consisting of f_{R1}, f_{R2} (focal lengths: respectively 200 mm and 300mm). Two bandpasses filter the collected fluorescence (Semrock FF01-514/30-25 for green fluorescent protein and Semrock FF01-650/92 for cherry). User-interface and synchronization of all instruments are provided by the INSCOPER software and controller.

Image reconstruction

The reconstruction is essentially based on the algoRIM treatment^{46,47} developed for the analysis of 2D-RIM images, and which is accessible at <https://github.com/teamRIM/tutoRIM>. In brief, the N images taken under different speckled illumination ($N \sim 100 - 200$ in most cases) are first processed with a Wiener filter, which can be written in the Fourier domain:

$$\tilde{g} = \frac{\tilde{h}_\perp^*}{|\tilde{h}_\perp|^2 + \eta} \quad (10)$$

where \tilde{h}_\perp is the Fourier transform of the extended depth PSF, a^* stands for the complex conjugate of a , and η is a regularization parameter set at 10^{-4} . This filter enhances high frequencies inside the OTF and removes noise beyond the OTF cutoff. We then estimate empirically the standard deviation of these pre-filtered speckled images. Finally, an iterative, moment-matching algorithm is used to achieve a super-resolved reconstruction of the object. We use a Tikhonov regularization method such that the reconstructed fluorescence density is given by

$$\hat{\rho}(\mathbf{r}) = \arg \min_{\rho} (|\hat{\sigma}(\mathbf{r}) - \sigma(\mathbf{r}; \rho)|^2 + \mu \|\rho\|^2) \quad (11)$$

with $\hat{\sigma}(\mathbf{r})$ the experimental standard-deviation and $\sigma(\mathbf{r}; \rho)$ the theoretical standard-deviation derived from the expected variance. In this theoretical expression, we generate a correlation function of the speckle and PSF that matches the Fourier support of the experimental autocorrelation function of the speckle. The difference with eq. 2 stands in the presence of the regularization term $\mu \|\rho\|^2$ in which μ is the Tikhonov regularization parameter. Due to the preprocessing step, we modified the model of the theoretical variance (eq. 6 for columnar speckles, eq. 9 for 3D speckles) by replacing h_\perp with $h_\perp * g$. When set too high, the regularization parameter μ can lead to image blurring and prevent super-resolution. Conversely, when set too low, it may lead to the amplification of noise and the appearance of artifacts. For experiments, we used a value $\mu = 1e - 4$ (Fig. 3C) and $\mu = 1e - 3$ (Fig. 3D). In the simulations (Appendix B), we explore the link between the value of μ and performance of the reconstruction in situations where the model for EDF image formation is correct, and when it is not.

Estimation of surface topography from one full scan

To estimate surface topography, the remote focus is used to capture the fluorescence signal under one speckled illumination plane by plane (*ie* one image per plane). This is typically performed by acquiring one plane every $2 \mu m$.

We then employ a robust approach for estimating surface topography, based on the method presented by Abouakil et al.⁴³ and Meng et al.⁴⁸. To correct for the spatial inhomogeneities of the background signal, we introduce a normalized signal of the form:

$$\hat{r}[\Omega_0](x, y, z) = \frac{s(x, y, z) - \hat{a}[\Omega_0](x, y, z)}{\hat{a}[\Omega_0](x, y, z) \hat{\sigma}[\Omega_0]} \quad (12)$$

Here, $s(x, y, z)$ is the measured signal, $\hat{a}[\Omega_0](x, y, z)$ is an estimation of the spatial inhomogeneities of the background, and $\hat{\sigma}[\Omega_0]$ is an estimation of the standard deviation of $s(x, y, z) / \hat{a}[\Omega_0](x, y, z)$ on the background. We determine both $\hat{a}[\Omega_0](x, y, z)$ and $\hat{\sigma}[\Omega_0]$ using only a subvolume Ω_0 of the acquired volume, taking randomly a small fraction (1/1000) of the acquired voxels. This treatment allows for a faster surface estimation.

The resulting normalized signal has a heavy-tailed Gaussian-like distribution in the histogram, with the Gaussian representing the background and the heavy tail representing the signal of interest from the biological structure. We set a threshold based on a chosen probability of false alarm ($p = 0.01$) and interpolate the epithelial surface modeled as $z = Z_s(x, y)$ using the detected bright points. To estimate the surface over large regions of interest, we use quadratic polynomial fits in overlapping windows (typically 3x3 to 5x5 windows), then use the RANSAC technique to estimate the most consensual classification of inliers and outliers⁴⁹. We fuse the overlapping estimations by preserving the inliers in all of the overlapping quadratic fits while discarding the remaining inliers. Finally, we estimate the surface by interpolating the remaining inliers using a simple bi-cubic harmonic spline interpolation.

Simulations

For the simulations, the object was defined with a 512x512x128 grid that ensured Nyquist criterium on the super-resolved simulated reconstruction. The simulated object is the combination of a star-shaped lateral distribution $\rho(r, \theta)$, and a topography $Z(r, \theta)$ (Fig. S1A). The lateral distribution follow the equation $\rho(r, \theta) = 1 + \cos(40\theta)$, which generates a star pattern characterized by higher spatial frequencies towards the center of the patter. The topography follows three different shapes: i) a flat constant Z; ii) a dome-shaped topography corresponding to a right circular cone with its radius equals to $\frac{N_x}{2}$, its height equals to N_z for $r < \frac{N_x}{2}$ and constant a flat constant Z for $r > \frac{N_x}{2}$; ; iii) a random topography with a uniformly distributed probability over the entire z-range, and no correlation between neighboring (x, y) points.

To generate illumination speckles, we computed the propagated field with a mask in the pupil plane with random phase and binary amplitude in the form of a disc for 3D-speckles and an annulus for Bessel speckles. The propagated speckle is given a Gaussian envelope, which introduces a correlation length in the spatial distribution of the random phases. A 3D image is then formed by multiplying the squared amplitude of the field with the object and convolving with the 3D PSF corresponding to the numerical aperture of the pupil. The EDF image is generated by summing the 3D image over all z. Simulations were run in Matlab.

Samples preparation

Immunolabeling of mouse intestinal epithelium. The jejunum of a sacrificed mouse was fixed in 4% paraformaldehyde (Electron Microscopy Sciences), washed in PBS, cut in 200-300 μ m thick slices, permeabilized with %1 Triton X-100 for 1h. Desmoplakin immunolabeling was done through overnight primary antibody incubation (Progen, 651109, 1/50 dilution), followed by washes in PBS, overnight secondary antibody incubation (Goat anti-Mouse, Alexa Fluor-488, Invitrogen, A11029, 1/200 dilution), washes in PBS and mounting in Aqua-Poly/Mount, (Polysciences).

Drosophila egg chamber live imaging. We followed the protocol of Prasad *et al.*⁵⁰. Briefly, Drosophila egg chambers (stage 9 to stage 10A) were dissected from adult females and mounted in Schneider's insect medium supplemented with 20% FBS and adjusted to a pH of 6.9. Prior to imaging, the egg chambers were slightly flattened on the coverslip by gentle pressure. Myosin-II was imaged in a transgenic line that expresses a protein fusion of Cherry with the regulatory light chain⁵¹ (*squash* gene) sqh:Cherry

Cytoskeletal labeling of cultured cells. cells were fixed with 4% Paraformaldehyde HCHO (15714 Electron Microscopy Sciences) in 37°C-prewarmed cytoskeleton buffer (10 mM MES with pH 6.1 with NaOH, 150 mM NaCl, 5 mM EGTA, 5mM glucose, 5mM $MgCl_2$) for 15 minutes and rinsed with PBS. Cells were then permeabilized, exposed to blocking solution, incubated with primary and secondary antibodies together with 0.165 μ M Alexa Fluor 488-phalloidin (A12379 Thermo Fisher Scientific) for 2 h at RT. While cells were treated with primary and secondary immunolabels, in this work we only imaged 488-phalloidin for actin cytoskeleton labeling.

Tight junction staining of chick neural tube. Chick embryos were fixed for 1 h in ice-cold 4% formaldehyde in PBS. Then they were cut along their midline and permeabilized for 15 min in 0.3% Triton X-100 in PBS (PBT 0.3%), before a 1-h blocking step in PBT 0.3% with 10% Goat Serum. The primary antibody is a mouse anti-ZO1 (Invitrogen ZO1-1A12) at 1:250 dilution, and secondary antibody is a goat anti-Mouse-alexaFluor488 (ThermoFisher Scientific, A28180) used at 1:500 dilutions. Embryos were then mounted between slide and coverslip using Vectashield (Vector Laboratories).

Acknowledgments

We thank Sophie Brasselet, Sandro Heuke and Hervé Rigneault, for support and fruitful discussions on the project.

This work was funded by the following agencies: Agence Nationale de la Recherche (ANR-18-CE13-028, ANR-20-CE45-0024, ANR-22-CE13-0039, ANR-22-CE42-0010, ANR-22-CE42-0026); Institut Carnot star (3D-RIM).

This project is funded by the « France 2030 » investment plan managed by the French National Research Agency (ANR-16-CONV-0001, ANR-21-ESRE-0002), and from Excellence Initiative of Aix-Marseille University - A*MIDEX.

References

1. Gao, R. *et al.* Cortical column and whole-brain imaging with molecular contrast and nanoscale resolution. *Science* **363**, eaau8302 (2019).
2. Gustafsson, M. G. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. *J. microscopy* **198**, 82–87 (2000).
3. Müller, C. B. & Enderlein, J. Image scanning microscopy. *Phys. review letters* **104**, 198101 (2010).
4. York, A. G. *et al.* Resolution doubling in live, multicellular organisms via multifocal structured illumination microscopy. *Nat. methods* **9**, 749–754 (2012).

5. Heintzmann, R. & Huser, T. Super-resolution structured illumination microscopy. *Chem. reviews* **117**, 13890–13908 (2017).
6. Wu, Y. & Shroff, H. Faster, sharper, and deeper: structured illumination microscopy for biological imaging. *Nat. methods* **15**, 1011–1019 (2018).
7. Schermelleh, L. *et al.* Super-resolution microscopy demystified. *Nat. cell biology* **21**, 72–84 (2019).
8. Prakash, K., Diederich, B., Heintzmann, R. & Schermelleh, L. Super-resolution microscopy: a brief history and new avenues. *Philos. Transactions Royal Soc. A* **380**, 20210110 (2022).
9. Chen, X. *et al.* Superresolution structured illumination microscopy reconstruction algorithms: a review. *Light. Sci. & Appl.* **12**, 172 (2023).
10. Chen, B.-C. *et al.* Lattice light-sheet microscopy: imaging molecules to embryos at high spatiotemporal resolution. *Science* **346**, 1257998 (2014).
11. Chang, B.-J., Perez Meza, V. D. & Stelzer, E. H. csilsm combines light-sheet fluorescence microscopy and coherent structured illumination for a lateral resolution below 100 nm. *Proc. Natl. Acad. Sci.* **114**, 4869–4874 (2017).
12. Chen, B. *et al.* Resolution doubling in light-sheet microscopy via oblique plane structured illumination. *Nat. methods* **19**, 1419–1426 (2022).
13. Chung, E., Kim, D., Cui, Y., Kim, Y.-H. & So, P. T. Two-dimensional standing wave total internal reflection fluorescence microscopy: superresolution imaging of single molecular and biological specimens. *Biophys. journal* **93**, 1747–1757 (2007).
14. Fiolka, R., Beck, M. & Stemmer, A. Structured illumination in total internal reflection fluorescence microscopy using a spatial light modulator. *Opt. letters* **33**, 1629–1631 (2008).
15. Li, D. *et al.* Extended-resolution structured illumination imaging of endocytic and cytoskeletal dynamics. *Science* **349**, aab3500 (2015).
16. Guo, M. *et al.* Single-shot super-resolution total internal reflection fluorescence microscopy. *Nat. methods* **15**, 425–428 (2018).
17. Roth, J., Mehl, J. & Rohrbach, A. Fast tirf-sim imaging of dynamic, low-fluorescent biological samples. *Biomed. optics express* **11**, 4008–4026 (2020).
18. Gustafsson, M. G. *et al.* Three-dimensional resolution doubling in wide-field fluorescence microscopy by structured illumination. *Biophys. journal* **94**, 4957–4970 (2008).
19. Li, X. *et al.* Three-dimensional structured illumination microscopy with enhanced axial resolution. *Nat. Biotechnol.* 1–13 (2023).
20. Wicker, K., Mandula, O., Best, G., Fiolka, R. & Heintzmann, R. Phase optimisation for structured illumination microscopy. *Opt. express* **21**, 2032–2049 (2013).
21. Demmerle, J. *et al.* Strategic and practical guidelines for successful structured illumination microscopy. *Nat. protocols* **12**, 988–1010 (2017).
22. Qian, J. *et al.* Structured illumination microscopy based on principal component analysis. *eLight* **3**, 4 (2023).
23. Turcotte, R. *et al.* Dynamic super-resolution structured illumination imaging in the living brain. *Proc. Natl. Acad. Sci.* **116**, 9586–9591 (2019).
24. Lin, R., Kipreos, E. T., Zhu, J., Khang, C. H. & Kner, P. Subcellular three-dimensional imaging deep through multicellular thick samples by structured illumination microscopy and adaptive optics. *Nat. communications* **12**, 3148 (2021).
25. Thomas, B., Wolstenholme, A., Chaudhari, S. N., Kipreos, E. T. & Kner, P. Enhanced resolution through thick tissue with structured illumination and adaptive optics. *J. biomedical optics* **20**, 026006–026006 (2015).
26. Förster, R., Müller, W., Richter, R. & Heintzmann, R. Automated distinction of shearing and distortion artefacts in structured illumination microscopy. *Opt. express* **26**, 20680–20694 (2018).
27. Mangeat, T. *et al.* Super-resolved live-cell imaging using random illumination microscopy. *Cell Reports Methods* **1**, 100009 (2021).
28. Liu, S. & Hua, H. Extended depth-of-field microscopic imaging with a variable focus microscope objective. *Opt. express* **19**, 353–362 (2011).

29. Mac, K. D. *et al.* Fast volumetric imaging with line-scan confocal microscopy by electrically tunable lens at resonant frequency. *Opt. Express* **30**, 19152–19164 (2022).
30. Mermillod-Blondin, A., McLeod, E. & Arnold, C. B. High-speed varifocal imaging with a tunable acoustic gradient index of refraction lens. *Opt. letters* **33**, 2146–2148 (2008).
31. Shain, W. J., Vickers, N. A., Goldberg, B. B., Bifano, T. & Mertz, J. Extended depth-of-field microscopy with a high-speed deformable mirror. *Opt. letters* **42**, 995–998 (2017).
32. Xiao, S., Tseng, H.-a., Gritton, H., Han, X. & Mertz, J. Video-rate volumetric neuronal imaging using 3d targeted illumination. *Sci. reports* **8**, 7921 (2018).
33. Botcherby, E., Juškaitis, R. & Wilson, T. Scanning two photon fluorescence microscopy with extended depth of field. *Opt. communications* **268**, 253–260 (2006).
34. Thériault, G., De Koninck, Y. & McCarthy, N. Extended depth of field microscopy for rapid volumetric two-photon imaging. *Opt. express* **21**, 10095–10104 (2013).
35. Thériault, G., Cottet, M., Castonguay, A., McCarthy, N. & De Koninck, Y. Extended two-photon microscopy in live samples with Bessel beams: steadier focus, faster volume scans, and simpler stereoscopic imaging. *Front. cellular neuroscience* **8**, 139 (2014).
36. Lu, R. *et al.* Video-rate volumetric functional imaging of the brain at synaptic resolution. *Nat. neuroscience* **20**, 620–628 (2017).
37. Abrahamsson, S., Usawa, S. & Gustafsson, M. A new approach to extended focus for high-speed high-resolution biological microscopy. *Three-Dimensional Multidimens. Microsc. Image Acquis. Process. XIII* **6090**, 128–135 (2006).
38. Mertz, J. *Introduction to optical microscopy* (Cambridge University Press, 2019).
39. Labouesse, S., Idier, J., Sentenac, A., Mangeat, T. & Allain, M. Random illumination microscopy from variance images. In *2020 28th European Signal Processing Conference (EUSIPCO)*, 785–789 (IEEE, 2021).
40. Affannoukoué, K. *et al.* Super-resolved total internal reflection fluorescence microscopy using random illuminations. *Optica* **10**, 1009–1017 (2023).
41. Clarke, D. N. & Martin, A. C. Actin-based force generation and cell adhesion in tissue morphogenesis. *Curr. Biol.* **31**, R667–R680 (2021).
42. Cetera, M. *et al.* Epithelial rotation promotes the global alignment of contractile actin bundles during drosophila egg chamber elongation. *Nat. communications* **5**, 5511 (2014).
43. Abouakil, F. *et al.* An adaptive microscope for the imaging of biological surfaces. *Light. Sci. & Appl.* **10**, 210 (2021).
44. Qu, Y. & Hu, Y. Analysis of axial scanning range and magnification variation in wide-field microscope for measurement using an electrically tunable lens. *Microsc. Res. Tech.* **82**, 101–113 (2019).
45. Marti, M. *et al.* Fast axial-scanning widefield microscopy with constant magnification and resolution. *J. Disp. Technol.* **11**, 913–920 (2015).
46. Idier, J. *et al.* On the superresolution capacity of imagers using unknown speckle illuminations. *IEEE Transactions on Comput. Imaging* **4**, 87–98 (2017).
47. Labouesse, S., Idier, J., Sentenac, A., Allain, M. & Mangeat, T. Proof of the resolution-doubling of random illumination microscopy using the variance of the speckled images. In *2021 29th European Signal Processing Conference (EUSIPCO)*, 1159–1162 (IEEE, 2021).
48. Meng, H., Nuzhdin, D., Sison, M., Galland, F. & LeGoff, L. Adaptive scans allow 3d-targeted laser dissection to probe the mechanics of cell sheets. *The Eur. Phys. J. Plus* **138**, 1–11 (2023).
49. Fischler, M. A. & Bolles, R. C. Random sample consensus: a paradigm for model fitting with applications to image analysis and automated cartography. *Commun. ACM* **24**, 381–395 (1981).
50. Prasad, M., Jang, A. C., Starz-Gaiano, M., Melani, M. & Montell, D. J. A protocol for culturing drosophila melanogaster stage 9 egg chambers for live imaging. *Nat. protocols* **2**, 2467–2473 (2007).
51. Martin, A. C., Kaschube, M. & Wieschaus, E. F. Pulsed contractions of an actin–myosin network drive apical constriction. *Nature* **457**, 495–499 (2009).
52. Goodman, J. W. Some fundamental properties of speckle. *JOSA* **66**, 1145–1150 (1976).

Supplementary Material

Appendix A: Detailed theory of EDF-RIM

Appendix B: Simulations

Supplementary Figure

Appendix A: Detailed theory of EDF-RIM

Model for extend depth image formation

We write $\mathbf{r} := (\mathbf{r}_\perp, z) \in \mathbb{R}^3$ as a spatial coordinate in 3D, with $\mathbf{r}_\perp := (x, y)$ locating a position in the (transverse) plane perpendicular to the optical axis z . Additionally, if f is a function defined over of \mathbf{r} , we note f_\perp its projection along the optical axis

$$f_\perp(\mathbf{r}_\perp) := \int f(\mathbf{r}) dz. \quad (13)$$

For a given 3D speckle excitation, the EDF (2D) image corresponds to the sum of the contribution from all planes, which writes as follows:

$$I_{\text{EDF}}(\mathbf{r}_\perp) = \int \left(\iint \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') h(\mathbf{r}_\perp - \mathbf{r}'_\perp, z - z') d\mathbf{r}'_\perp dz' \right) dz \quad (14)$$

with h is the 3D PSF of the microscope, S the speckled illumination and ρ the fluorescence density (which is the product between the fluorophore concentration and their brightness). In the above equation, if we integrate along z first, we get

$$I_{\text{EDF}}(\mathbf{r}_\perp) = \iint \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) d\mathbf{r}'_\perp dz' \quad (15)$$

where we used the fact that $\int h(\mathbf{r}_\perp, z - z') dz = \int h(\mathbf{r}_\perp, z) dz = h_\perp(\mathbf{r}_\perp)$. Let us assume further that speckles are columnar functions. In such an instance, S is invariant along z and we write $S(\mathbf{r}_\perp, z) = S_B(\mathbf{r}_\perp)$, $\forall z$. The relation above has the following specific form

$$I_B(\mathbf{r}_\perp) = \iint \rho(\mathbf{r}'_\perp, z') S_B(\mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) d\mathbf{r}'_\perp dz' \quad (16)$$

and if we finally integrate along z' , we obtain a simple convolution model that involves projected quantities only

$$I_B(\mathbf{r}_\perp) = \int \rho_\perp(\mathbf{r}'_\perp) S_B(\mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) d\mathbf{r}'_\perp. \quad (17)$$

In summary, under the assumption of a Bessel illumination, the model for the microscope image is analogous to the one of a standard widefield microscope, but it involves the extended depth PSF, h_\perp and the projection of the object ρ_\perp .

Theoretical expression of the EDF variance

We derive now the theoretical variance of a EDF-RIM experiment. We start first by making no specific assumption about the illumination function $S(\mathbf{r}_\perp, z)$. More specifically, we write from Eq. 15:

$$\begin{aligned} \sigma_{\text{EDF}}^2(\mathbf{r}_\perp) &:= \langle I_{\text{EDF}}^2(\mathbf{r}_\perp) \rangle \\ &= \int \left\langle \int \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') dz' \int \rho(\mathbf{r}''_\perp, z'') S(\mathbf{r}''_\perp, z'') dz'' \right\rangle h_\perp(\mathbf{r} - \mathbf{r}'_\perp) h_\perp(\mathbf{r} - \mathbf{r}''_\perp) d\mathbf{r}'_\perp d\mathbf{r}''_\perp \end{aligned} \quad (18)$$

Assuming speckles that are invariant along z , we can write $S(\mathbf{r}_\perp, z) = S_B(\mathbf{r}_\perp)$ which allows the following simplification

$$\int \rho(\mathbf{r}'_\perp, z') S(\mathbf{r}'_\perp, z') dz' \int \rho(\mathbf{r}''_\perp, z'') S(\mathbf{r}''_\perp, z'') dz'' = \rho_\perp(\mathbf{r}'_\perp) \rho_\perp(\mathbf{r}''_\perp) S_B(\mathbf{r}'_\perp) S_B(\mathbf{r}''_\perp) \quad (19)$$

hence leading to the following approximation for the EDF-RIM variance

$$\sigma_{\text{EDF}}^2(\mathbf{r}_\perp) \approx \sigma_B^2(\mathbf{r}_\perp) := \iint \rho_\perp(\mathbf{r}'_\perp) \rho_\perp(\mathbf{r}''_\perp) \Gamma_B(\mathbf{r}''_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) d\mathbf{r}'_\perp d\mathbf{r}''_\perp \quad (20)$$

where we have introduced the 2D autocorrelation of the columnar speckle $\Gamma_B(\mathbf{r}''_\perp - \mathbf{r}'_\perp) = \langle S_B(\mathbf{r}'_\perp) S_B(\mathbf{r}''_\perp) \rangle$. The right-hand side in Eq. 20 has the structure of the variance expression in RIM given by Eq. 3. It should be noted, however, that the classical result $h_\perp = \Gamma_B(\mathbf{r}_\perp)$ will not be verified, unless we modify the observation PSF by putting a ring in the pupil (Fourier plane).

EDF-RIM variance with standard speckles: the case of a spatially smooth fluorescence surface

Bessel-speckles are rarely used as an illumination function. Here, we investigate under what conditions EDF-RIM can be performed with conventional 3D speckles. We thus return to the general case of an illumination speckle that is not invariant along z . Using Eq. 18, the EDF-RIM variance reads

$$\sigma_{\text{EDF}}^2(\mathbf{r}_\perp) = \iint h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) \iint \rho(\mathbf{r}'_\perp, z') \rho(\mathbf{r}''_\perp, z'') \Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, z'' - z') \, dz' dz'' \, d\mathbf{r}'_\perp \, d\mathbf{r}''_\perp \quad (21)$$

where we have introduced the 3D autocorrelation of the speckle $\Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, z'' - z') = \langle S(\mathbf{r}'_\perp, z') S(\mathbf{r}''_\perp, z'') \rangle$. Let us assume now that the sample is such that the fluorophores are distributed along a surface denoted $Z(\mathbf{r}_\perp)$:

$$\rho(\mathbf{r}_\perp, z) = \rho(\mathbf{r}_\perp) \delta(z - Z(\mathbf{r}_\perp)) \quad (22)$$

where δ is the Dirac distribution. Under this assumption, Eq. 21 simplifies into

$$\sigma_{\text{EDF}}^2(\mathbf{r}_\perp) = \iint \rho(\mathbf{r}'_\perp) \rho(\mathbf{r}''_\perp) \Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, Z(\mathbf{r}''_\perp) - Z(\mathbf{r}'_\perp)) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) \, d\mathbf{r}'_\perp \, d\mathbf{r}''_\perp$$

This expression is similar to the standard 2D RIM expression of the variance Eq. 9 except that it involves $\Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, Z(\mathbf{r}''_\perp) - Z(\mathbf{r}'_\perp))$ rather than $\Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, 0)$. To further simplify the speckle autocorrelation into a 2D expression, we hypothesize that the objects distribute on a smooth topography. This translates mathematically into the following assumption for Γ_S :

- $\|\mathbf{r}''_\perp - \mathbf{r}'_\perp\| < \frac{\lambda}{2\text{NA}} \implies Z(\mathbf{r}''_\perp) \approx Z(\mathbf{r}'_\perp) \implies \Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, Z(\mathbf{r}''_\perp) - Z(\mathbf{r}'_\perp)) \approx \Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, 0)$
- $\|\mathbf{r}''_\perp - \mathbf{r}'_\perp\| > \frac{\lambda}{2\text{NA}} \implies \Gamma_S(\mathbf{r}'_\perp - \mathbf{r}''_\perp, z'' - z') \approx 0$

Two assumptions have been made. The first one states that the surface does not vary much in z over lateral distances of the order of $\frac{\lambda}{2\text{NA}}$ (smoothness hypothesis). The second one neglects long-range correlations in the illumination, which is ensured because speckle decorrelates on length-scales similar to the lateral extent of the PSF, $\frac{\lambda}{2\text{NA}}$ ⁵². Following these assumptions, the expression of the variance of the intensity is simplified: it involves the 2D PSF h_\perp and the 2D correlation function $\Gamma_{\text{EDF}}(\mathbf{r}_\perp) := \Gamma_S(\mathbf{r}_\perp, 0)$:

$$\sigma_{\text{EDF}}^2(\mathbf{r}_\perp) \approx \iint \rho(\mathbf{r}'_\perp) \rho(\mathbf{r}''_\perp) \Gamma_{\text{EDF}}(\mathbf{r}'_\perp - \mathbf{r}''_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}'_\perp) h_\perp(\mathbf{r}_\perp - \mathbf{r}''_\perp) \, d\mathbf{r}'_\perp \, d\mathbf{r}''_\perp \quad (23)$$

To conclude, our study has shown that the variance of the extended depth RIM images can be simplified into the classical RIM 2D expression of Eq. 3 for any type of sample, by using columnar speckles (see Eq. 20), and for surface and smoothed samples, when using regular 3D speckles (see Eq. 23).

In the following section, we demonstrate an experimental implementation of EDF-RIM. We then investigate the performance of EDF-RIM and the limits of its numerical approach through simulations.

Appendix B: Simulations

We investigate the realms of application of EDF-RIM through simulations. More specifically, we examine the implications of the EDF variance expression in the case of the 3D speckle (eq. 21) only approximately conforming to the canonical 2D-RIM expression (eq. 3). In the simulations, the imaged object is characterized by its star-shaped lateral distribution $\rho(r)$, as shown in Fig. S1.A (left). We investigate three distinct topographies for the 3D object, as depicted in Fig. S1.A (right). These include a flat structure, a smoothly varying surface in shape of a cone, and a random distribution of z-position. The flat object serves as a reference for which our model is correct, with no approximation. The smoothly varying surface satisfies the approximation conditions for the EDF variance to align with the canonical 2D-RIM expression (eq 9), whereas the random configuration does not.

From a numerical standpoint, the inversion requires the deconvolution of the speckled images using a Tikhonov regularized inverse filter²⁷. The effectiveness of this inverse filter depends on its regularization parameter. When this parameter is too high, it can result in image blurring and prevent super-resolution. Conversely, setting it too low may lead to the amplification of noise and the appearance of artefacts. We compared the inversions for 2 different regularization parameters -one large (10^{-5}) and one low (10^{-10} respectively), and for the 3 aforementioned 3D objects, the idea being that an inaccurate model generally induces more artifacts for small regularization parameters.


We first examine the scenario of a 3D propagating speckle, which aligns with our experimental implementation. In Fig. S1.B, we present the reconstructed images for the three topographies, with both large ($\mu = 10^{-5}$) and small ($\mu = 10^{-10}$) Tikhonov regularization parameters. While all three results appear satisfactory at $\mu = 10^{-5}$, this setting is over-regularized, leading to a suboptimal resolution. At $\mu = 10^{-10}$, only the smooth surface compares well with the 2D case, while reconstruction of the random z-position is thwarted by amplified noise. Thus, it stands that conditions that permit the use of an exact or at least approximate expression for the image variance are a prerequisite for EDF-RIM.

As outlined in Appendix A, the EDF-variance expression can be transformed into a 2D-RIM expression for all types of samples when using a speckle that remains invariant along the optical axis. This condition is achieved experimentally by blocking light outside an annulus in the pupil plane (Bessel speckle, Fig. S1.C, right). We simulated these Bessel speckles by modifying the binary mask in the pupil plane which is no longer a disk but an annulus. This annulus filters out all frequencies smaller than $0.95 \frac{2NA}{\lambda}$ and bigger than $\frac{2NA}{\lambda}$. The generated speckle are effectively invariant along the optical axis (Fig. S1.C, right panels).


Simulating EDF-RIM with these Bessel speckles, we were able to achieve high-resolution images for all samples, including the random z configuration, for every value of the regularization parameter tested (Fig. S1.B, lower panels). This is consistent with our prediction that with columnar speckles, the reconstruction does not rely on the sparsity hypothesis.

In conclusion, our simulations confirm that EDF-RIM is a feasible imaging method, and support the limits outlined in the theory section. Specifically, we have demonstrated that 3D-speckles can be used to image sparse samples along the optical axis, such as surfaces, and Bessel speckles can be used for arbitrary samples.

A-Object



B-Reconstruction



C-Speckle used



Figure S1. Simulations. A) Simulated object: the lateral distribution $\rho(r)$ is depicted on the left, while three topographies are studied – a flat structure (left), a dome (center) and a random topography (right). B) Simulated reconstructions for the 3 topographies studied, with 2 different illumination-patterns (see below). For each condition, we consider the reconstruction using a large and a small regularization parameter in the inversion procedure (10^{-5} and 10^{-10} respectively). Images include guidelines at $\frac{\lambda}{2NA}$ (blue circle), $\frac{\lambda}{2.96NA}$ (cyan circle) and $\frac{\lambda}{4NA}$ (red circle). C) The three illumination patterns used in the simulation. The 3D-Speckle is obtained numerically by simulating the propagation of an electromagnetic field with uniform amplitude and random phase distribution in the back focal plane of the objective. The Bessel is obtained by replacing the homogeneous amplitude in the back focal plane of the lens with an annulus illumination.