

#### Automation of the timed water swallow test (TWST) using high-resolution cervical auscultation signals

Lila Gravellier, Maxime Le Coz, Jérôme Farinas, Julien Pinquier

#### ▶ To cite this version:

Lila Gravellier, Maxime Le Coz, Jérôme Farinas, Julien Pinquier. Automation of the timed water swallow test (TWST) using high-resolution cervical auscultation signals. 31st Dysphagia Research Society Annual Meeting (DRS 2022), Mar 2023, San Francisco, CA, United States. hal-04285442

#### HAL Id: hal-04285442 https://hal.science/hal-04285442v1

Submitted on 14 Nov 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



## CONTEXT

The VideoFluoroscopy Swallowing Study (VFSS) is considered as the **reference examination** to assess dysphagia. However it cannot be done consistently due to limited accessibility, invasiveness and cost. High Resolution Cervical Auscultation (HRCA) is a promising alternative when VFS is not feasible<sup>[1-3]</sup>.

The **Timed Water Swallow Test**<sup>[4]</sup> (TWST) is a validated bedside assessment to evaluate dysphagia severity. This protocol consists in timing the patient drinking a full glass of water and counting the number of swallows needed to do so. The average volume per swallow, time per swallow and swallowing capacity are the three resulting indicators to evaluate dysphagia severity.

This preliminary study introduces a method for automating the TWST using HRCA signals.



Representation of consecutive swallows signals during TWST

#### REFERENCES

[1] Dudik, J.M. et al. 2015. Dysphagia Screening: Contributions of Cervical Auscultation Signals and Modern Signal-Processing Techniques. IEEE Transactions on Human-Machine Systems. 45, 4 (Aug. 2015), 465–477. DOI:https://doi.org/10.1109/THMS.2015.2408615.

[2] Donohue, C. et al. 2020. A Preliminary Investigation of Whether HRCA Signals Can Differentiate Between Swallows from Healthy People and Swallows from People with Neurodegenerative Diseases. Dysphagia. (Sep. 2020). DOI:https://doi.org/10.1007/s00455-020-10177-0.

# **AUTOMATION OF THE TIMED WATER SWALLOW TEST USING HIGH RESOLUTION CERVICAL AUSCULTATION SIGNALS**

Lila GRAVELLIER<sup>1,2</sup>, Maxime LE COZ<sup>2</sup> Jérôme FARINAS<sup>1</sup>, Julien PINQUIER<sup>1</sup>

<sup>1</sup> IRIT, Université de Toulouse, CNRS, Toulouse, France <sup>2</sup> Swallis Medical lila.gravellier@irit.fr, m.lecoz@swallismedical.com

### **High Resolution Cervical Auscultation**



Throat sensors

Signals acquired





# **DATA AND MATERIALS**

### Scenario and subjects

44 healthy participants (25 females, 19 males) from 21 to 57 y.o. (mean 33+/-10 y.o) were asked to drink a 100mL glass of water in one go.



### Device

The **Swallis DSA**<sup>™</sup>, which includes a three-axis accelerometer and a microphone, was used to record vibroacoustic signals on the neck of each subject.

# **METHOD**

Automatic swallow detection was done by an algorithm trained on HRCA signals to detect pharyngolaryngeal activities. The TWST KPI (volume/swallow, time/swallow, volume/time) were calculated for each subject. Average values are given by gender (F/M) and compared with the corresponding healthy groups G1 and G2 from the literature<sup>[4]</sup> (18-54 y.o.).

> [3] Donohue, C. et al. 2021. Establishing Reference Values for Temporal Kinematic Swallow Events Across the Lifespan in Healthy Community Dwelling Adults Using High-Resolution Cervical Auscultation. Dysphagia. (May 2021). DOI:https://doi.org/10.1007/s00455-021-10317-0. [4] Hughes, T.A.T. and Wiles, C.M. 1996. Clinical measurement of swallowing in health and in neurogenic dysphagia. QJM. 89, 2 (Feb. 1996), 109–116. DOI:https:// doi.org/10.1093/qjmed/89.2.109.



Analysis

# **RESULTS AND DISCUSSION**

**Detection results:** 196/224 swallows detected, 28 missed and 3 false positives.

#### **TWST indices for each group:**

	Volume/Swallow (ml)		Time/Swallow (s)		Volume/Time (ml/s)	
	М	F	М	F	М	F
Predicted	24,4	22,3	1,1	1,3	22,7	19,1
Annotated	22,2	20,3	1,1	1,2	21,5	17,7
G1 G2 Reference Range <sup>[4]</sup>	21,4-50	13,6-30	1-1,4	1-1,7	17-41,4	8,8-24,6

TWST.

Remaining detection errors could be tackled by adapting the algorithm for multiple successive swallows, which slightly differ in shape and duration from isolated swallows in HRCA signals.

### **DETECTION ALGORITHM STRUCTURE**

#### **1. Activity detection**





#### The automatic predicted results match the healthy group reference values. Therefore this method could help practicioners performing

