

Improving the real-time control of a non-invasive Brain-Machine Interface using a finite-state machine design

Corentin Piozin, Lisa Bouarroudj, Jean-Yves Audran, Brice Lavrard, Catherine Simon, Florian Waszak, Selim Eskiizmirliler

▶ To cite this version:

Corentin Piozin, Lisa Bouarroudj, Jean-Yves Audran, Brice Lavrard, Catherine Simon, et al.. Improving the real-time control of a non-invasive Brain-Machine Interface using a finite-state machine design. Journées CORTICO 2023, May 2023, Paris, France. hal-04312833

HAL Id: hal-04312833 https://hal.science/hal-04312833v1

Submitted on 28 Nov 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

Improving the real-time control of a non-invasive **Brain-Machine Interface** using a finite-state machine design

C. Piozin¹, L. Bouarroudj¹, J.Y. Audran², B. Lavrard³, C. Simon³, F. Waszak¹, S. Eskiizmirliler¹

¹ Integrative Neuroscience and Cognition Center (INCC), CNRS UMR8002, Université Paris Cité ² Otto Bock France, Otto Bock, Courtaboeuf ³ Pôle Ambulatoire, Pôle appareillage du membre supérieur, Institut Robert Merle d'Aubigné

Correspondence : piozin.corentin@hotmail.fr

Methods

Highlights

• The implementation of a finite-state machine design on the control of the prosthesis led to an increase in the classification performance on both the able-bodied and the amputee while simplifying the decision-making process after the classification.

• In real-time control, able-bodied subjects obtained higher performance compared to the one amputee, it could be due to the difference in EEG systems used with the groups (64 ActiCHamp vs. 16 electrodes OpenBCI).

• The sensory feedback system did not improve the classification performance in able-bodied and for the amputee.

Background

• The goal of developing a Brain-Machine Interface (BMI) control system for medical prosthetics is to provide robust, comfortable and easy-to-use devices for the concerned user. They need to be fast and have high performance in estimating the user's intention in real-time.

• Yet, the transition from offline to online settings often proves to be challenging (Deo et al., 2023), with a drop in performance when testing the BMI system in the latter.

• In the literature, there are numerous examples of methods and techniques aimed at targeting this issue, here applied to the control of a prosthetic hand.





Université de Paris

Alignement of covariance Finite-state machine matrices • Real-time control has been performed with amputees and able-bodied subjects i = 1during two sessions without and with the use of a sensory feedback device for i = 2provided by a vibrating bracelet. $\gamma(\overline{C}_{i-1,\text{test}}, C_i, \frac{1}{i-1})$ $i \ge 3$ • The experiments consisted in doing a sequence of Grasping-Moving-Releasing close Close Open Start / End 🚄 $\gamma(\overline{C}_{i-1,\text{test}}, C_i, \frac{1}{i-1}) = \overline{C}_{i-1}^{-1/2} \times (\overline{C}_{i-1}^{-1/2} \times C_i \times \overline{C}_{i-1}^{-1/2})^{\frac{1}{i-1}} \times \overline{C}_{i-1}^{-1/2}$ three objects, using a modified Myobock[®] prosthesis controlled by EEG data. open • The effect of three different methods of feature extraction and decision-making 3 process was tested on the online control performance and compared to the offline. **B** Subset of electrodes Move Grip → Release Go Half second of EEG is recorded Prosthesis and sensory feedback used → Trained → Movement on SVMs → prosthesis Feature by (A) able-bodied and (B) amputees. Extraction

The use of the finite-state machine design can improve the performances during the grip and release phases for the two groups

While the best performances were obtained with finite-state machine for the **BB** (Box&Blocks) & **CS** (Clothespin) objects, results are more nuanced with the CUP.


A finite-state machine design leads to

Able-bodied are performing better than the amputee

The difference could be explained by the different EEG system used for each group.


Participants were performing better during the release phase compared to the grip phase

The Riemannian feature extraction pipeline, the alignment method and the selection of a subset of electrodes all increased this difference for able-bodied while only the Riemannian pipeline did for the amputee.




a simplification of the decision process

The simplification does not prevent the system to raise errors during classification, as reflected by the low performance obtained when using the CUP object in the grip phase.




More sessions are needed to see an effect of sensory feedback



Online performances are getting closer to offline ones

With an exception for the CUP object, especially for the grip phase.



Conclusion

• Online classification performances were obtained with 19 able bodied and 1 amputee subject who controlled the modified Myobock prosthesis by using the EEG data recorded in real-time and compared to the performances obtained offline.

• The effect of two different feature extraction methods and of a new decision-making process on the classification performances were studied for the same classification algorithm (SVM RBF and TS-SVM).

• The decoding of object release is always better than the object grip with both able-bodied and amputee subjects for all different combinations.

• Using a Feature Extraction method based on Riemannian Geometry and alignment of covariance matrices, and a Finite-State Machine design as decision-making process in online allows to reach classification performances obtained in offline except for the CUP object.

Acknowledgments

This work has been partially supported by the French National ResearchTechnology Association and Otto Bock France in the framework of the CIFRE project n°2018/1620.

To access my PhD thesis on the control of prosthetic hands using non-invasive BMIs Scan the QR code :

