

Challenges and solutions for accurate magnetization reversal and tipping outside the RWA

Geneviève Tastevin, Pierre-Jean Nacher, Christopher P. Bidinosti

▶ To cite this version:

Geneviève Tastevin, Pierre-Jean Nacher, Christopher P. Bidinosti. Challenges and solutions for accurate magnetization reversal and tipping outside the RWA. European Magnetic Resonance Meeting EUROMAR 2023, Jul 2023, Glasgow (Ecosse), United Kingdom. https://euromar2023.org/abstractbook/. hal-04334246

HAL Id: hal-04334246

https://hal.science/hal-04334246

Submitted on 10 Dec 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.





Challenges and solutions for accurate magnetization reversal and tipping outside the RWA

Tastevin Geneviève¹, NACHER Pierre-Jean ¹, Bidinosti Christopher P. ²

Laboratoire Kastler Brossel, Paris, France

Department of Physics, University of Winnipeg, Winnipeg, Canada

Introduction

In low static field, with a linear rather than circular polarized rf field, a breakdown of the rotating-wave approximation (RWA) may occur and lead to significant deviations from expected trajectories on the Bloch sphere. The effects depend strongly on several factors including detuning (via the static field or rf frequency) and pulse characteristics (phase, amplitude, shape, or transients).

Aims

Using simulations and experiments, we seek to explore the many combined effects that one must consider when using pulsed linear rf fields outside the RWA. We aim to develop practical strategies that can be employed in this regime to generate accurate tip angles.

Methods

Relaxation and diffusion were ignored in simulations. The Bloch equation was solved via numerical integration to compute the magnetization vector in the laboratory frame. Either nominal or recorded coil currents were used as inputs in the simulations. Experiments were performed in a home-made 3-mT MRI system¹ with wired rf coils², using laser-polarized ³He gas or thermally-polarized water samples.

Results

The desired terminus can be achieved through the application of a suitable phase-dependent Bloch-Siegert shift, notably different from that of cw NMR, and appropriate consideration of pulse timings². The figure displays selected simulation results obtained for rectangular Π pulses: rf-driven North-South trajectories, terminus points, and Bloch-Siegert shifts needed to reach the South pole. Experimental data are in good agreement with expectations. They are very sensitive to fine details in coil currents and rf fields.

Conclusions

For any pulse duration and start phase, a shift from resonance and a small change in rf amplitude can be used to compensate for inaccuracies induced by operation outside the RWA. Practical considerations arising from this work and prospects for further investigations will be described.

¹Nacher et.al., J. Magn. Reson 310 (2020) 106638 ²Bidinosti et.al., J. Magn. Reson 345 (2022) 107306

