

MiniRead: A simple and inexpensive do-it-yourself device for multiple analyses of micro-organism growth kinetics

Matthieu Falque, Aurélie Bourgais, Fabrice Dumas, Mickaël de Carvalho, Célian Diblasi

► To cite this version:

Matthieu Falque, Aurélie Bourgais, Fabrice Dumas, Mickaël de Carvalho, Célian Diblasi. MiniRead: A simple and inexpensive do-it-yourself device for multiple analyses of micro-organism growth kinetics. Yeast, 2024, 10.1002/yea.3932 . hal-04478675

HAL Id: hal-04478675 https://hal.science/hal-04478675

Submitted on 27 Feb 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

REVIEW ARTICLE

TechnoYeast



MiniRead: A simple and inexpensive do-it-yourself device for multiple analyses of micro-organism growth kinetics

Célian Diblasi 💿

Revised: 31 January 2024

Matthieu Falque 💿 🕴 Aurélie Bourgais 🕴 Fabrice Dumas 🕴 Mickaël de Carvalho 💿

GQE-Le Moulon, Université Paris-Saclay, INRAE, CNRS, AgroParisTech, Gif-sur-Yvette, France

Correspondence

Matthieu Falgue, GOE-Le Moulon, Université Paris-Saclay, INRAE, CNRS, AgroParisTech, 12 Route 128. Gif-sur-Yvette 91190. France. Email: matthieu.falque@inrae.fr

Funding information Agence Nationale de la Recherche; Labex Saclay Plant Sciences

Abstract

Fitness in micro-organisms can be proxied by growth parameters on different media and/or temperatures. This is achieved by measuring optical density at 600 nm using a spectrophotometer, which measures the effect of absorbance and side scattering due to turbidity of cells suspensions. However, when growth kinetics must be monitored in many 96-well plates at the same time, buying several 96-channel spectrophotometers is often beyond budgets. The MiniRead device presented here is a simple and inexpensive do-it-yourself 96-well temperature-controlled turbidimeter designed to measure the interception of white light via absorption or side scattering through liquid culture medium. Turbidity is automatically recorded in each well at regular time intervals for up to several days or weeks. Output tabulated text files are recorded into a micro-SD memory card to be easily transferred to a computer. We propose also an R package which allows (1) to compute the nonlinear calibration curves required to convert raw readings into cell concentration values, and (2) to analyze growth kinetics output files to automatically estimate proxies of growth parameters such as lag time, maximum growth rate, or cell concentration at the plateau.

KEYWORDS

do-it-yourself, fitness, growth kinetics, micro-organism, turbidimeter

1 | INTRODUCTION

Although the ramifications of the concept of fitness can be quite bushy (Krimbas, 2004), the individual absolute fitness of an organism may be defined in a simplified way as the number of descendants that this organism will produce in a given environment. In asexually reproducing unicellular micro-organisms, different components of fitness can be experimentally measured quite directly by characterizing quantitatively the kinetics of mitotic cell divisions in different controlled environments. This may require to repeat growth experiments on many different chemical (composition of the culture media) and/or physical (e.g., growth temperature) environmental conditions. Therefore, fitness measurements often need to carry out many growth assays in parallel, even if the number of different sample strains or populations is initially limited. In that context, 96-well plates are appropriate to carry out growth experiments, and many manufacturers propose optical plate readers adapted to that format, which allow automatic growth kinetics monitoring.

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made. © 2024 The Authors. Yeast published by John Wiley & Sons Ltd.

² WILEY-Yeast

However, most readers are designed to measure absorbance at precise wavelengths or even fluorescence, whereas monitoring growth kinetics does not require such sophisticated measurements. Growth experiments with micro-organisms are typically carried out on either solid-gelled (e.g., see Barton et al., 2018) or liquid culture media, using 96 or even 384-well plates (Jung et al., 2015). On solid medium, growth can be monitored via measuring either the size of colonies by image analysis, or the level of light interception through a small column of medium on top of which the colony is growing. In liquid medium, cell concentration can be simply proxied by measuring the turbidity of the broth for different kinds of micro-organisms (Hills, 2017; Luong et al., 2011; Nwoba et al., 2022), related to the amount of light which is side-scattered or absorbed when a light beam is passing through a column of cell suspension. When using flat-bottom wells, the same holds true even if the cells have sedimented and form a layer at the bottom of the well. This makes agitation not necessary provided there is enough gas exchanges for cells to grow under the desired conditions. Long ago already, that turbidity approach had given rise to a simple device to monitor growth (Watson et al., 1969). However, by that time, the state-ofthe-art of electronics offered limited possibilities to efficient homemade implementations, and since then, we are not aware of any other published simple device of the same type. Our objective here was to build on the same idea to design and assemble a small and inexpensive 96-well homemade turbidimeter, of which one could produce several units with a modest budget.

The main requirements for the MiniRead device were: (1) use $150-200 \,\mu$ L volume of liquid culture in 96-wells plates without agitation, (2) record light interception by the column of cell suspension in each well at regular time intervals, (3) accurately control the temperature of the bottom of the culture plate (up to 45° C), (4) control independently the temperature of the plate lid, which should be kept for example, two degrees warmer than the bottom to avoid mist condensation—that point is critical for optical readings while keeping sterile conditions in the wells, (5) store data autonomously without the need for a computer, (6) simple assembly within the reach of nonprofessional electronicians having good DIY soldering experience, and (7) low cost allowing to build five such readers within a total budget of 1000\$.

2 | MATERIALS AND METHODS

2.1 | Hardware description and working principle

The detailed user manual of the MiniRead turbidimeter is available in File S1. The device (Figure 1) is build around a widely used Arduino MEGA256 board (see specifications at https://store.arduino.cc/ collections/boards/products/arduino-mega-2560-rev3) which includes a C-programmable ATmega2560 micro-controller with a large number of digital and analog input/output pins. The MiniRead is power supplied by an external regulated power supply (desktop power adapter) 12 V DC 5 A 60 W from RS components (https://fr.rs-online.com).

Take-away

We present a simple and inexpensive do-it-yourself turbidimeter for cell suspensions. It records automatically micro-organisms growth kinetics in 96-well plates without computer. Temperature is controlled in plate bottom and lid independently to avoid mist on the lids. An associated R package allows automatic kinetics analysis to estimate growth parameters.

This 12 V input directly powers the two heating elements via Darlington transistors, while there is an additional 5 V regulator in the MiniRead to supply the Arduino and its external electronics.

For the light sensors, we used 96 cheap GL5516 light-dependent resistors (LDRs) from Aliexpress (https://fr.aliexpress.com/) but we quality-controlled each LDR by measuring its resistance under a fixed level of illumination. This procedure is recommended, since it allowed us to identify and discard some defective components with outlier resistance values. More details are available on our website at https://moulon.inrae.fr/materiel_labo/miniread/. We used also cheap 3 mm white light emitting diodes (LEDs) from the same provider as sources of illumination. We did only a qualitative quality control of each LED by checking its illumination under a constant 20 mA current, and found very few defective LEDs.

Analog signals from the 96 LDRs are multiplexed using twelve CD4051BE analog 8-to-1 multiplexers and their outputs are connected to 12 analog inputs of the Arduino. To avoid any well-to-well light contamination, the 96 wells are lit and read one at a time by the software. Each reading, which takes less than 1 s, is the average of 500 successive acquisitions to reduce noise. The LED matrix and the multiplexers are addressed by digital outputs of the Arduino. LED intensity can be adjusted from the user interface by using the pulse-width modulation functionalities of Arduino digital outputs.

Temperature measurements are achieved by two NTC thermistances ($10 k\Omega 1\%$, EPCOS ref B57861S0103F040) purchased from RS components (https://fr.rs-online.com). There is one NTC in contact with the bottom of the culture plate and the other one in contact with the lid. Top and bottom heating elements are made by soldering resistive wire (6 Ohm/m from RS components https://fr.rsonline.com) on a PCB plate in a serpentine design around all wells (see Supporting Information S4: Figure S1) to ensure homogeneous heating. For one plate, the length of the wire is about 160 cm. Under 12 V power supply, the heating power is thus approximately 15 W for each heating plate, and the total maximum consumption of the MiniRead is around 3 A (the current drained by the LEDs and the overall electronics is negligible as compared to the heating current).

The user interface of the MiniRead (Figure 1a) contains a 4 × 16 characters LCD display (MC41605A6W-FPTLW-V2 from RS components; https://fr.rs-online.com), six pushbuttons and a control multicolor LED, connected to digital inputs and outputs of the



FIGURE 1 General view of the MiniRead. (a) Top view with the user interface and the front cover. (b) Inside view shows the plate holder and the heated lid with the 96 light-emitting diodes.

Arduino. It allows a completely autonomous use of the MiniRead without a computer. Two additional LEDS indicate when heating is activated for the bottom and the lid of the culture plate. For additional details, see the user manual in File S1.

During data acquisition, raw values are written in a micro-SD memory card through a cheap standard card reader module for Arduino from Aliexpress (https://fr.aliexpress.com/). The card (FAT32-formatted 8 Gb) can easily be un/plugged from the side of the device box to transfer files to a computer.

2.2 | Printed circuit board (PCB)

To easily assemble the device, we designed different pieces of PCB to build the plate holder, with soldering pads allowing for a quick assembly without any gluing or screwing step. Additional pieces of PCB are used for the bottom and top plates with their corresponding heating elements. We used EasyEda editor (https://easyeda.com/) to draw the PCB and ordered five pieces from JLCPCB (https://jlcpcb. com/), but other export formats (Gerber fabrication files, SVG and PDF images) are available on https://forgemia.inra.fr/gqe-base/ MiniRead so it should be possible to have the PCB fabricated by any other company. We did not use any surface-mounted components, so soldering of the components can be done with a conventional soldering iron by anybody having experience in electronics.

2.3 | Software

The MiniRead contains an embedded software (firmware) written in the Arduino version of the C language. It controls sequential LED illumination, light intensity data acquisition and averaging, temperature regulation, writing in the SD card, and user interface (writing information on the LCD display, reading keys,...). This software also includes different programs allowing the user to easily carry out growth kinetics recording, calibration, or single plate acquisition. The firmware source code in Arduino C language is available at https:// forgemia.inra.fr/gqe-base/MiniRead. It can be compiled and flashed into the Arduino board via an USB cable using the free Arduino IDE software available at https://www.arduino.cc/en/software. Beyond updating the firmware, the USB port can also be used for bidirectional serial communication of the MiniRead with a computer for development or debugging purposes, or to display the reading values in real-time on a computer during the growth kinetics.

Once raw data from the MiniRead have been transferred to a computer via the SD card, we propose a dedicated R package to help analyzing them. The main functions provided in this package allow to compute and model calibration curves for each of the 96 wells and to analyze growth kinetics experiments to extract different proxies of growth parameters. The package file is available as File S2 and as a release of the MiniRead project at https://forgemia.inra.fr/gqe-base/MiniRead/-/releases.

2.4 | Housing case

MiniRead devices must be insulated from external light during operation. Therefore, we designed a plywood box with hinged front cover allowing for easy placing and removing of the plate, top window fitting the user interface, and side openings to access the SD card and the USB connector of the Arduino board, which is necessary to flash new versions of the firmware or to use serial communication. Dimensions of the plywood sheets required to build the housing are available at https://moulon.inrae.fr/materiel_labo/miniread/.



To study the response curves of LDRs in all wells of all MiniRead machines, we measured an empty plate with its cover, after placing under the lid zero to 13 layers of neutral optical filter sheet (Frost Diffusion White Lighting Filter Gel Sheets from https://cableandlight. co.uk/) covering all wells. The optical density (OD) of the filter stack is proportional to the number of filter layers, so we used this method to model the nonlinear relationship between the raw MiniRead output values and the corresponding number of filter layers, by using the *Calibrate()* function of the MiniRead R package (see tutorial in File S3) which uses the *smooth.spline()* function of R.

2.6 | Biological calibration

We tested the MiniRead devices with the yeast species *S. cerevisiae* (diploid SK1 and haploid S288c strains), *Candida glabrata*, and *Nakaseomyces delphensis* species (haploid), and with the bacteria *Escherichia coli* (DH5 α strain). Liquid cultures for growth kinetics were carried out in 200 µL standard YPD medium for yeasts and LB medium for bacteria, in flat-bottom 96-well plates (Sarstedt ref. 833924500). Initial inoculation was done at OD = 0.01 and growth kinetics were carried out for 48 h at constant temperature of 30°C (plate bottom) and 32°C (lid) for yeasts, and at 37°C (plate bottom) and 32°C (lid) for the bacteria. LED light intensity was set to 100 (arbitrary units are between 1 and 255) and the MiniReads were programmed to make one acquisition every five (bacteria) or 10 (yeasts) min (KINETICS program).

To compute calibration curves, we focused only on yeasts. An initial liquid cell culture (600 mL) at the plateau stage was centrifuged to remove the medium and re-suspended in 150 mL liquid YPD supplemented with 250 mM EDTA (to stop growth). This cell suspension was used as the highest point of a linear range of five diluted cell concentration standards; the lowest (zero) point was made with pure YPD + EDTA medium. Cell suspensions of each dilution point were carefully homogenized and uniformly distributed in the 96 wells of a flat-bottom culture plate (200 mL/well), then measured with the CALIBRATION program of the MiniRead in four technical replicates.

3 | RESULTS AND DISCUSSION

3.1 | Functionalities of the user interface

The MiniRead displays in real-time the temperatures of the bottom and lid of the culture plate and the turbidity value acquired by the LDR sensors for each well coordinate (Supporting Information S5: Figure S2). The display also shows the set values for the LEDs brightness, the time interval between two data acquisitions in the KINETICS program, and the current output file name (see details in user manual in File S1). File names are composed of a prefix depending on the program (SING for single plate acquisition, CALI for sensors calibration, and KINE for growth kinetics experiments) followed by a number that is automatically incremented before each new experiment. File name extension can be set to MRA, MRB, MRC, and so on... to distinguish data produced by different MiniRead units when several devices are used at the same time.

3.2 | Accuracy of temperature control

To ensure stable results over time, the MiniRead must be powered on several hours before starting the experiment, ideally the day before. Once powered on, it takes about 10–15 min to stabilize plate and lid temperatures to respectively 30°C and 32°C. The precision obtained on these temperatures in our five prototypes is better than 0.1°C (Figure 2).

3.3 | Calibration results

The MiniRead raw output values are in arbitrary units between 0 and 1023, which measure the amount of light intercepted before reaching the LDR sensors. However, (1) the response curve of the sensors is not linear and may vary from well to well, and (2) the relationship between measured OD and cell concentration is also nonlinear as soon as the concentration becomes high. So we performed two different calibrations: (1) a physical calibration, to model the response curve of each individual LED-LDR pair (see Materials and Methods). The 96 response curves (Figure 3a) were computed by the *Calibrate()* function of the MiniRead R package (see tutorial of the package in File S3). Using these curves, the *ApplyCalibration()* function can transform raw data from any well of any MiniRead unit into a predicted value of the number of filter



FIGURE 2 Temperature of the plate bottom (solid lines) and plate lid (dashed lines) during a 4-day growth kinetics. Black (respectively red) curves are temperatures measured just before (respectively after) each data acquisition (every 10th minute).



FIGURE 3 Physical calibration of the 96 light-emitting diodes (LED)-light-dependent resistor (LDR) pairs using the Calibrate() function of the MiniRead R package. (a) Nonlinear relationship observed between raw output values produced by the MiniRead (y-axis) when placing increasing numbers of layers of neutral optical filter sheet under the plate lid (x-axis). Black circles are the real values measured and the red curve is the result of a nonlinear fit using splines. Each individual curve corresponds to one well of the plate. b: comparison between real numbers of filter sheets (x-axis) and predicted number (y-axis) as computed by the ApplyCalibration() function from the raw output values, using the nonlinear relationship modeled by Calibrate(). Black circles are the experimental values and the red line is the result of linear regression.

layers (see Figure 3b) which is proportional to the OD. This value is comparable across all wells, including from different machines. (2) a biological calibration, to model directly the relationship between raw MiniRead output values and cell concentrations, based on standard dilutions of a yeast cell culture (see Section 2). As before, the Calibrate() function was used to compute calibration curves (Figure 4a), which can be applied to raw data with the function ApplyCalibration(). Figure 4b shows the quality of this biological calibration. As we can see in the y-axis values of Figure 4a, even with a four-times concentrated plateau yeast culture, we are still far below the saturation value of the MiniRead (1023) and the response curves are relatively close to linearity. On the other hand, we see in Figure 3a that when OD is very high (with 13 layers of optical filter), the response curve is clearly nonlinear for some wells, but nevertheless, Figure 3b shows that that nonlinearity is very well corrected by the ApplyCalibration() function of the package, so the MiniRead can be used accurately over its whole dynamic range.

In case some very dark or very transparent culture media were to be used, it may be useful to change the LED brightness (0-255), instead of the default value set to 100, to optimize the dynamic range.

3.4 Growth kinetics analysis

During growth kinetics, the evaporation was not negligible and resulted in a linear increase of the estimated cell concentration over time, even with noninoculated sterile YPD medium (see Supporting Information S6: Figure S3). This is not straightforward to understand since evaporation proportionally decreases the optical path length and increases the concentration, which should not change the OD. Our hypothesis is that this effect is purely geometric: our sources of light are neither lasers, nor perfectly collimated sources, so the light rays emitted are not parallel but more or less divergent. When the rays enter the surface of the liquid medium which has a much higher refractive index than the air, they tend to become more perpendicular to the liquid surface (i.e., more vertical) than when they were in the air. So the fraction of rays that are lost on the sides becomes lower in the liquid than above it. Therefore, if evaporation lowers the height of the liquid surface, the fraction of light rays that escape the sensor increases. We confirmed this hypothesis experimentally with increasing volumes of both pure water and YPD medium at constant number of molecules of solutes (see Supporting Information S7: Figure S4). The effect was very clear and strong for the wells located at the edges of the plate, but for the inner wells there seems to be also other phenomena making the curves nonmonotonic, most probably due to complex reflections on the plastic of the neighboring wells. Evaporation was expected during growth kinetics since we chose plates designed to ensure sterile gas exchanges allowing aerobic microorganisms to grow in acceptable conditions. Thus the MiniRead package uses the plateau phase of the cell growth kinetics to model the effect of evaporation in each particular well, to correct this effect when estimating growth parameters. Figure 5 shows a typical growth



-WILEY-Yeast

6

Calibration with dilutions of yeast cell suspensions MiniRead E



FIGURE 4 Biological calibration of the 96 wells with the same approach as for Figure 3, except that filter layers were replaced by 96-well culture plates containing a range of yeast cell suspensions of increasing concentrations in YPD medium, from zero to a four-times concentrated culture at the plateau phase (see Section 2).



FIGURE 5 Typical yeast growth kinetics measured with the MiniRead, after automatic analysis of the output file with the function *AnalyzeGrowth()* of the MiniRead R package. Left panel: analysis without correction for evaporation (argument flatPlateau = FALSE). Right panel: with correction for evaporation (default: argument flatPlateau = TRUE). Black circles: experimental measurements transformed into cell concentrations after calibration with titrated cell suspensions. Gray solid line: nonlinear fit (splines smoothing) of the experimental points. Black dashed line: linear regression of the plateau phase, used to correct for evaporation. Pink solid line (right panel only): growth curve corrected for evaporation based on the plateau slope. Thick red dashed line: tangent at the inflection point (maximum growth rate) of the uncorrected gray (left panel) or corrected pink (right panel) curve. The other dashed lines indicate: beginning of the plateau (thin blue), value of the plateau (thick blue), end of the lag phase (thick green), time of the inflection point (thin red). Values provided as text in the figures indicate the estimated values of the growth parameters. In the right panel, the values between parentheses refer to the analysis without correction for evaporation.

curve of S. cerevisiae over 48 h, obtained with the MiniRead and analyzed with the AnalyzeGrowth() function of the R package using all default parameters (right panel) or with argument flatPlateau = FALSE (left panel). Three different phases of growth are automatically identified based on the first and second derivatives of the curve. First, the lag phase from the inoculation until cell concentration reaches a given proportion of the value at the plateau, Then the exponential growth phase, with more and more saturation resulting in an inflection point. Finally, the plateau phase where the number of cells is not supposed to change over time. Of course, depending on the organism and conditions used, this does not reflect the complexity of all phenomena controlling the kinetics of cell concentration (e.g., possible switch from fermentation to respiration in yeasts). However, this simple model allowed us to easily extract efficient proxies of three useful parameters of cell growth: the maximum rate of cell division (at the inflection point), the duration of the lag phase, and the cell concentration at the plateau (proxying the carrying capacity). These proxies are automatically estimated from the data before (between parentheses) and after correction for evaporation (Figure 5 right panel). We first tried to use classical growth kinetics models to fit our curves, but the fit quality was low, with strong auto-correlation of the residuals, which we interpreted a being due to our nonagitating conditions. So we found it more robust to adopt a nonparametric approach that extracts proxies describing only the main local characteristics of the curves, without fitting the whole curve to a growth model. The AnalyzeGrowth() function also computes the rate of evaporation (drift) and the signal-to-noise ratio which gives a useful estimation of the physical quality of the results.

All parameters computed from the kinetics, without and with correction for evaporation, are also written in a tabulated file to be opened by any spreadsheet software.

3.5 | Reproducibility

To assess the technical reproducibility of the MiniRead, we placed five layers of gray homogeneous optical filter in a plate and compared the readings across wells. First without applying any calibration, we found a CV of 14.4% across wells, which confirms that the sensors have very different characteristics and definitely need to be calibrated. After applying calibration (see Figure 3), the CV went down to 0.44%, confirming that our calibration procedure satisfactorily corrects the variations across sensors. Now to assess the biological reproducibility of our growth analyses, we distributed the same inoculated medium in all wells of five plates distributed into five different MiniRead devices, and recorded the growth kinetics. After applying calibration with cell suspensions (Figure 4), we measured biological CVs in the five devices. The values of the CVs ranged from 12.5% to 18.6% for maximum growth rate, from 6.8% to 9.4% for Lag Time, and from 12.2% to 14.6% for the Plateau. To figure out local or edge effects, Figure 6 shows the parameter values as a function of the position on the plate in each device. We see that wells located on the border rows and columns tend to have cell growth slightly affected as compared to the inner wells. This is expected because of evaporation which acts more on edges where the air is locally more dry. This source of variation should be as much as possible taken into



FIGURE 6 Assessment of local effects on estimated growth parameters depending of the position on the plate. Cold (resp. hot) colors correspond to low (resp. high) values of the parameter considered. Each column shows the results of five different MiniRead devices (Device ID from a to e) for one of the five parameters written as text in Figure 5.

* WILEY- Yeast

account in the experimental designs with the MiniRead, as it is also the case with any other type of plate reader.

3.6 | Conclusion

The MiniRead provides an efficient and inexpensive solution to measure growth kinetics of yeasts and bacteria in several hundreds of samples in parallel. All required information to home-build these small machines, including associated firmware and a data analysis R package, are freely available under GPL license. So we hope this device can be useful to other research groups, as it is to us.

AUTHOR CONTRIBUTIONS

Matthieu Falque developed the hardware, firmware, and R package, built the devices, and wrote the manuscript. Aurélie Bourgais, Fabrice Dumas, Mickaël de Carvalho, Célian Diblasi produced wet-lab results and helped with the first R scripts. All authors performed yeast growth kinetics experiments and approved the manuscript.

ACKNOWLEDGMENTS

The authors thank Olivier Martin for his help in specifying the temperature regulation algorithm, Franck Gauthier for checking the R package, Cécile Fairhead, Monique Bolotin-Fukuhara, Youfang Zhou, Killian Métivier, and Xavier Raffoux for helpful discussions and comments on the project, and an anonymous reviewer who helped us to deepen our thinking on evaporation effects. The present work has benefited from funding from the EVOLREC project (ANR-20-CE13-0010). GQE—Le Moulon benefits from the support of the LabEx Saclay Plant Sciences-SPS (ANR-10-LABX-0040-SPS).

DATA AVAILABILITY STATEMENT

The data that support the findings of this study are openly available in ForgeMia at https://forgemia.inra.fr/gqe-base/MiniRead. The project includes (1): user manual (File S1), firmware, list of electronic components and printed circuit board manufacturing files for the MiniRead device, and (2) a release containing the MiniRead R package for calibration and data analysis (File S2) with its tutorial (File S3). Detailed device building instructions are available at https://moulon. inrae.fr/materiel_labo/miniread/.

ORCID

Matthieu Falque b http://orcid.org/0000-0002-6444-858X Mickaël de Carvalho b http://orcid.org/0000-0003-0059-8324 Célian Diblasi b http://orcid.org/0009-0001-6548-0661

REFERENCES

- Barton, D. B. H., Georghiou, D., Dave, N., Alghamdi, M., Walsh, T. A., Louis, E. J., & Foster, S. S. (2018). PHENOS: A high-throughput and flexible tool for microorganism growth phenotyping on solid media. *BMC Microbiology*, 18, 9.
- Hills, A. E. (2017). Spectroscopy in biotechnology research and development. In J. C. Lindon, G. E. Tranter, & D. W. Koppenaal (Eds.), *Encyclopedia of spectroscopy and spectrometry* (3rd ed., pp. 198–202). Academic Press. https://www.sciencedirect.com/science/article/pii/ B9780128032244000352
- Jung, P. P., Christian, N., Kay, D. P., Skupin, A., & Linster, C. L. (2015). Protocols and programs for high-throughput growth and aging phenotyping in yeast. *PLoS One*, 10, e0119807.
- Krimbas, C. B. (2004). On fitness. Biology & Philosophy, 19, 185-203.
- Luong, J. H. T., Mahmoud, K. A., & Male, K. B. (2011). Instrumentation and analytical methods. In M. Moo-Young (Ed.), *Comprehensive biotechnology* (2nd ed., pp. 829–838). Academic Press. https://www. sciencedirect.com/science/article/pii/B9780080885049001392
- Nwoba, E. G., Chuka-Ogwude, D., Vadiveloo, A., & Ogbonna, J. C. (2022). Process control strategies applied to microalgae-based biofuel production. In E. Jacob-Lopes, L. Q. Zepka, I. A. Severo, & M. M. Maroneze (Eds.), 3rd generation biofuels (pp. 105–134). Woodhead Publishing. https://www.sciencedirect.com/science/article/pii/ B9780323909716000310
- Watson, B. W., Gauci, C. L., Blache, L., & O'Grady, F. W. (1969). A simple turbidity cell for continuously monitoring the growth of bacteria. *Physics in Medicine & Biology*, 14, 555–558.

SUPPORTING INFORMATION

Additional supporting information can be found online in the Supporting Information section at the end of this article.

How to cite this article: Falque, M., Bourgais, A., Dumas, F., de Carvalho, M., & Diblasi, C. (2024). MiniRead: A simple and inexpensive do-it-yourself device for multiple analyses of micro-organism growth kinetics. *Yeast*, 1–8. https://doi.org/10.1002/yea.3932