

The CIPM list 'Recommended values of standard frequencies': 2021 update

H Margolis, G Panfilo, G Petit, C Oates, T Ido, S Bize

▶ To cite this version:

H Margolis, G Panfilo, G Petit, C Oates, T Ido, et al.. The CIPM list 'Recommended values of standard frequencies': 2021 update. Metrologia, 2024, 61 (3), pp.035005. 10.1088/1681-7575/ad3afc. hal-04564764

HAL Id: hal-04564764

https://hal.science/hal-04564764

Submitted on 5 May 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.





PAPER • OPEN ACCESS

The CIPM list 'Recommended values of standard frequencies': 2021 update

To cite this article: H S Margolis et al 2024 Metrologia 61 035005

View the <u>article online</u> for updates and enhancements.

You may also like

- Absolute frequency measurement of the ${}^{1}S_{0}$ ${}^{3}P_{0}$ transition of ${}^{1/1}$ Yb with a link to international atomic time Marco Pizzocaro, Filippo Bregolin, Piero Barbieri et al.
- On the evaluation of ultra-high-precision frequency ratio measurements: examining closed loops in a graph theory framework L Robertsson
- On a definition of the SI second with a set of optical clock transitions
 Jérôme Lodewyck

The CIPM list 'Recommended values of standard frequencies': 2021 update

H S Margolis^{1,*}, G Panfilo², G Petit², C Oates³, T Ido⁴ and S Bize⁵

- ¹ National Physical Laboratory, Hampton Road, Teddington TW11 0LW, United Kingdom
- ² Bureau International des Poids et Mesures, 92310 Sèvres, France
- ³ National Institute of Standards and Technology (NIST), Boulder, CO 80305, United States of America
- ⁴ National Institute of Information and Communications Technology, 4-2-1 Nukui-kitamachi, Koganei, Tokyo 184-8795, Japan
- ⁵ LNE-SYRTE, Observatoire de Paris, Université PSL, CNRS, Sorbonne Université, 61 avenue de l'Observatoire, 75014 Paris, France

E-mail: helen.margolis@npl.co.uk

Received 11 December 2023, revised 5 April 2024 Accepted for publication 5 April 2024 Published 30 April 2024



Abstract

This paper gives a detailed account of the analysis underpinning the 2021 update to the list of standard reference frequency values recommended by the International Committee for Weights and Measures. This update focused on a subset of atomic transitions that are secondary representations of the second (SRS) or considered as potential SRS. As in previous updates in 2015 and 2017, methods for analysing over-determined data sets were applied to make optimum use of the worldwide body of published clock comparison data. To ensure that these methods were robust, three independent calculations were performed using two different algorithms. The 2021 update differed from previous updates in taking detailed account of correlations among the input data, a step shown to be important in deriving unbiased frequency values and avoiding underestimation of their uncertainties. It also differed in the procedures used to assess input data and to assign uncertainties to the recommended frequency values, with previous practice being adapted to produce a fully consistent output data set consisting of frequency ratio values as well as absolute frequencies. These changes are significant in the context of an anticipated redefinition of the second in terms of an optical transition or transitions, since optical frequency ratio measurements will be critical for verifying the international consistency of optical clocks prior to the redefinition. In the meantime, the reduced uncertainties for optical SRS resulting from this analysis significantly increases the weight that secondary frequency standards based on these transitions can have in the steering of International Atomic Time.

Keywords: secondary representation of the second, recommended values of standard frequencies, absolute frequency, frequency ratio, redefinition of the second

Original Content from this work may be used under the terms of the Creative Commons Attribution 4.0 licence. Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

^{*} Author to whom any correspondence should be addressed.

1. Introduction

The historical development of the list of standard reference frequency values recommended by the International Committee for Weights and Measures (CIPM) has been well described by Riehle et al [1]. Today it contains recommended values of standard frequencies for applications that include both the practical realisation of the definition of the metre and secondary representations of the definition of the second (SRS). The 2021 update to the list, described in this paper, focused on 14 atomic transitions (table 1) which had either already been adopted as SRS, or which were considered to be potential candidates for becoming SRS. The analysis underpinning the update was performed under the auspices of the Working Group on Frequency Standards (WGFS), a joint working group of the Consultative Committee for Length (CCL) and the Consultative Committee for Time and Frequency (CCTF), who are charged by the CIPM with maintaining the list and making proposals for recommendations to the relevant consultative committee [2]. The updated recommended frequency values and uncertainties were approved by the CCTF at their 22nd meeting in March 2021 [3] and became active on 13 April 2022, following publication on the website of the International Bureau of Weights and Measures (BIPM) [4].

The list of recommended frequencies plays an important role in progress towards an anticipated redefinition of the SI second based on an optical transition or transitions. Optical frequency standards are already used as secondary frequency standards (SFS), alongside caesium primary frequency standards (PFS), for calibration of the scale interval of International Atomic Time (TAI) [5–7] and the number and frequency of these contributions is expected to increase significantly in the coming years. SFS contribute to TAI using the recommended frequency value and uncertainty of the SRS on which they are based, and so the lower these uncertainties are, the more benefit these contributions will bring to the stability and accuracy of TAI. The analysis underpinning the derivation of the recommended frequency values can also help to verify the consistency of optical frequency ratio measurements, another key prerequisite for a redefinition of the second.

Ratios between unperturbed atomic transition frequencies are dimensionless quantities given by nature. For a collection of frequency standards based on N_S different reference transitions with frequencies ν_k ($k = 1, 2, ..., N_S$), it is in principle possible to measure a total of $N_{\rm S}(N_{\rm S}-1)/2$ different frequency ratios, but only $N_{\rm S}-1$ of these are independent. Since 2015, the worldwide body of clock comparison data available to the WGFS has contained an increasing number of optical frequency ratio measurements, in addition to absolute frequency measurements relative to caesium primary frequency standards. This means that the clock comparison data set is over-determined, i.e. it is possible to derive some frequency ratios from the results of several different experiments. To make full use of the available data, new analysis methods were developed [8] and were applied for the first time to update the list of recommended values of standard frequencies in 2015. By the time of the next update in 2017, a second method was also available [9], providing important verification of

Table 1. Atomic transitions included in the 2021 least-squares adjustment.

Label	Atomic species	Reference transition
$\overline{\nu_1}$	¹¹⁵ In ⁺	5s ² ¹ S ₀ –5s5p ³ P ₀
ν_2	1 H	$1s^2S_{1/2}-2s^2S_{1/2}$
ν_3	¹⁹⁹ Hg	$6s^2 {}^1S_0 - 6s6p {}^3P_0$
ν_4	$^{27}Al^{+}$	$3s^2 {}^1S_0 - 3s3p {}^3P_0$
ν_5	$^{199}{\rm Hg}^{+}$	$5d^{10}6s^2S_{1/2}(F=0)-5d^96s^2^2D_{5/2}(F=2)$
ν_6	$^{171}{\rm Yb}^{+}$	$6s^2S_{1/2}$ $(F=0)-5d^2D_{3/2}$ $(F=2)$
ν_7	$^{171}{\rm Yb}^{+}$	$6s^2S_{1/2}$ $(F=0)$ $-4f^{13}6s^2{}^2F_{7/2}$ $(F=3)$
ν_8	¹⁷¹ Yb	$6s^2 {}^1S_0 - 6s6p {}^3P_0$
ν_9	⁴⁰ Ca	$4s^2 {}^1S_0 - 4s4p {}^3P_1$
ν_{10}	$^{88}\mathrm{Sr}^{+}$	$5s^2S_{1/2}-4d^2D_{5/2}$
ν_{11}	⁸⁸ Sr	$5s^2 {}^1S_0 - 5s5p {}^3P_0$
ν_{12}	⁸⁷ Sr	$5s^2 {}^1S_0 - 5s5p {}^3P_0$
ν_{13}	40 Ca $^+$	$4s^2S_{1/2}-3d^2D_{5/2}$
ν_{14}	⁸⁷ Rb	$5s^2S_{1/2} (F=1) - 5s^2^2S_{1/2} (F=2)$
ν_{15}	¹³³ Cs	$6s^2S_{1/2}$ $(F=3)-6s^2^2S_{1/2}$ $(F=4)$

the results. These methods were employed again in the 2021 update to the list described here, but with several significant differences:

- (i) Correlations among the input data were given detailed consideration and taken into account in the analysis. As previously pointed out in [8, 10], this is critical to ensure that the recommended frequency values are unbiased and that their uncertainties are properly estimated.
- (ii) Reflecting the increasing importance of optical frequency ratios for verifying the international consistency of optical clocks as progress is made towards a redefinition of the second, these were computed and are provided as an appendix to the list of recommended frequency values.
- (iii) In assessing the input data and considering the uncertainty to be assigned to each recommended frequency value, the rules and criteria set out in [1] were modified to ensure that the output data set, which consists of correlated absolute frequency measurements and frequency ratio measurements, is internally self-consistent.

The inclusion of correlations in the analysis, combined with the larger number of new measurements compared to previous updates, significantly increased the effort involved in computing the new recommended frequency values. We note that the work was done under strict time constraints, imposed by the date of the CCTF meeting in March 2021. Some choices therefore had to be made in the analysis that might, in retrospect, have differed in some respects had more time been available for extended interactions with the groups who made the measurements used as input.

The paper is organised as follows. A brief description of the analysis methods is given in section 2, while the new measurement data available for the 2021 update is presented in section 3. Section 4 describes the approach that was taken to compute correlation coefficients between the input data. In section 5 we summarise all modifications that were made to

the input data, and describe the rationale for these. The recommended values that result from the analysis are presented and discussed in section 6. We finish in section 7 with some conclusions and perspectives that may be relevant for future discussions within the WGFS and the CCTF, in the period leading up to an anticipated optical redefinition of the SI second.

2. Analysis methods

To ensure that the methods used to derive optimal values for the frequencies ν_k are robust, three independent calculations were carried out based on two different algorithms.

2.1. Algorithm 1

The first approach used is a least-squares adjustment based on the well-established method employed by the CODATA Task Group on Fundamental Constants to derive a self-consistent set of values for the fundamental physical constants [11]. The method is described in detail in [8] and hence is outlined only briefly here.

The input data to the least-squares adjustment are a set of N frequency ratio measurements q_i , together with their standard uncertainties u_i and correlation coefficients $r(q_i, q_j)$, from which their variances u_i^2 and covariances $u_{ij} = u_{ji}$ are computed. The input data includes optical frequency ratios, microwave frequency ratios and optical-microwave frequency ratios, but all are handled in a similar way, with absolute frequency measurements treated simply as a special case of frequency ratios involving a caesium primary frequency standard.

The measured frequency ratios are expressed as a function of one or more of a set of $M = N_S - 1$ independent adjusted frequency ratios z_j , yielding a set of N equations. It is the values of these adjusted frequency ratios that are optimized in the least squares adjustment.

In most cases, the equations relating the measured frequency ratios to the adjusted frequency ratios are non-linear. So that linear matrix methods can be employed, the equations are therefore linearised prior to the least-squares adjustment by using a Taylor expansion around initial estimates of the adjusted frequency ratios. This linear approximation means that the best estimates for the values of the adjusted frequency ratios (together with their variances and covariances) are not exact solutions to the original non-linear equations, but the values obtained from the least-squares adjustment are used as starting values for a new linear approximation and another leastsquares adjustment performed. This process is repeated until the values of the adjusted frequency ratios converge. Any other frequency ratio of interest (and its uncertainty) can then be calculated from the adjusted frequency ratios and their covariance matrix.

Two independent implementations of this algorithm were used for the computation, one written in MATLAB® and the other written in Mathematica®.

2.2. Algorithm 2

The second algorithm used to derive optimal values for the frequencies ν_k uses a different conceptual approach described

in [9], based on the examination of closed loops in a graph theory framework. By analysing figure 1, a number of closed loops can easily be identified. As in the first algorithm, the equations reporting the ratios are non-linear, so to simplify the treatment the logarithms of the frequencies are used, which converts the problem to a linear least squares one. The logarithms of all frequency ratios in each closed loop should add up to zero. This provides a set of conditions that are used in a Lagrange multiplier method to identify the basis vectors for the residual space in the least squares calculation. A projection on this subspace gives the corrections to the experimental ratio values

A single implementation of this algorithm, written in Matlab, was used for the computation.

2.3. Numerical precision

In all cases, due to the extremely high accuracy with which frequency ratios have been measured, care had to be taken to perform numerical calculations to sufficiently high precision (more than 18 significant figures). This was achieved using routines designed for high precision floating point arithmetic, the standard double-precision floating point format being insufficient.

3. New measurement data since 2017

The starting point for the collection of the input data was the input data file from the previous adjustment performed by the WGFS in 2017. That file included 69 input data points, of which 58 were absolute frequency measurements (table 2). The remaining 11 were frequency ratio measurements that did not involve Cs (table 3), most of them optical frequency ratio measurements, but two being optical-microwave frequency ratios.

The additional data since 2017 consisted of 15 new absolute frequency measurements (table 4) and 22 new frequency ratio measurements (table 5), all of which met the WGFS requirement of being published in a peer-reviewed journal by March 2021. However, one of the new frequency measurements (q_{90}) included the contribution from the data of q_{51} , and so q_{51} was (effectively) excluded from the calculation by multiplying its uncertainty by 10⁶ in the input data file. (This approach to exclusion enabled continuity of measurement labels with the previous adjustment to be maintained, and the exact multiplication factor was verified to have no effect on the output of the adjustment, as long as it was sufficiently large.) The input data to the 2021 least-squares adjustment thus includes a total of 105 clock comparison measurements, which break down into 72 absolute frequency measurements and 33 frequency ratio measurements not involving Cs. A pictorial representation of the complete body of data is shown in figure 1.

No new atomic transitions were involved in 2021 as compared to 2017, but some of the new frequency ratios reported had never previously been measured directly. It was noted that a majority of new absolute frequency measurements had been made using International Atomic

Table 2. Absolute frequency measurement data used in the 2017 least-squares adjustment. All frequency values and uncertainties listed are the published ones, before adjustments made by the WGFS (see section 5).

Clock transition	Atomic species	Measured frequency/Hz	Fractional uncertainty	Reference	Label
$\overline{\nu_1}$	¹¹⁵ In ⁺	1267 402 452 899 920(230)	1.8×10^{-13}	[12]	q_1
		1267 402 452 901 265(256)	2.0×10^{-13}	$[13]^{a}$	q_2
		1267 402 452 901 049.9(6.9)	5.9×10^{-15}	[14]	q_3
ν_2	¹ H	1233 030 706 593 517.5(5.0)	4.1×10^{-15}	[15]	q_4
		1233 030 706 593 509.0(5.5)	4.5×10^{-15}	[16]	q_5
ν_3	¹⁹⁹ Hg	1128 575 290 808 155.1(6.4)	5.7×10^{-15}	$[17, 18]^{b}$	q_6
	C	1128 575 290 808 154.62(0.41)	3.6×10^{-16}	[19]	q_7
ν_4	$^{27}Al^{+}$	1121 015 393 207 851(6)	5.4×10^{-15}	[20]	q_8
ν_5	$^{199}{\rm Hg}^{+}$	1064 721 609 899 145.30(0.69)	6.5×10^{-16}	[21]	q_9
ν_6	¹⁷¹ Yb ⁺ E2	688 358 979 309 308.0(2.14)	3.1×10^{-15}	[22]	q_{10}
o .		688 358 979 309 306.97(0.73)	1.1×10^{-15}	[23] ^c	q_{11}
		688 358 979 309 310(9)	1.3×10^{-14}	[24]	q_{12}
		688 358 979 309 307.82(0.36)	5.2×10^{-16}	[25]	q_{13}
		688 358 979 309 308.42(0.42)	6.1×10^{-16}	[26]	q_{14}
ν_7	¹⁷¹ Yb ⁺ E3	642 121 496 772 657(12)	1.9×10^{-14}	[27]	q_{15}
·		642 121 496 772 645.15(0.52)	8.1×10^{-16}	[28]	q_{16}
		642 121 496 772 646.22(0.67)	1.0×10^{-15}	[29]	q_{17}
		642 121 496 772 644.91(0.37)	5.8×10^{-16}	[26]	q_{18}
		642 121 496 772 645.36(0.25)	3.9×10^{-16}	[30]	q_{19}
ν_8	¹⁷¹ Yb	518 295 836 590 864(28)	5.4×10^{-14}	[31]	q_{20}
•		518 295 836 590 863.1(2.0)	3.9×10^{-15}	[32]	q_{21}
		518 295 836 590 865.2(0.7)	1.4×10^{-15}	[33]	q_{22}
		518 295 836 590 863.5(8.1)	1.6×10^{-14}	[34]	q_{23}
		518 295 836 590 863.59(0.31)	6.0×10^{-16}	[35]	q_{24}
		518 295 836 590 863.38(0.57)	1.1×10^{-15}	[36]	q_{25}
ν_9	⁴⁰ Ca	455 986 240 494 144.0(5.3)	1.2×10^{-14}	[37]	q_{26}
		455 986 240 494 135.8(3.4)	7.5×10^{-15}	[38, 39]	q_{27}
ν_{10}	$^{88}{ m Sr}^{+}$	444 779 044 095 484.6(1.5)	3.4×10^{-15}	[40]	q_{28}
		444 779 044 095 484(15)	3.4×10^{-14}	[41]	q_{29}
		444 779 044 095 485.5(0.9)	2.0×10^{-15}	[42]	q_{30}
		444 779 044 095 486.71(0.24)	5.3×10^{-16}	[43]	q_{31}
		444 779 044 095 485.27(0.75)	1.7×10^{-15}	[44]	q_{32}
ν_{11}	⁸⁸ Sr	429 228 066 418 009(32)	7.5×10^{-14}	[45]	q_{33}
		429 228 066 418 008.3(2.1)	4.9×10^{-15}	[46]	q_{34}
		429 228 066 418 007.3(2.9)	6.8×10^{-15}	[46]	q_{35}
ν_{12}	⁸⁷ Sr	429 228 004 229 874.0(1.1)	2.6×10^{-15}	[47]	q_{36}
		429 228 004 229 873.65(0.37)	8.6×10^{-16}	[48]	q_{37}
		429 228 004 229 873.6(1.1)	2.6×10^{-15}	[49]	q_{38}
		429 228 004 229 874.1(2.4)	5.6×10^{-15}	[49]	q_{39}
		429 228 004 229 872.9(0.5)	1.2×10^{-15}	[50]	q_{40}
		429 228 004 229 873.9(1.4)	3.3×10^{-15}	[51]	q_{41}
		429 228 004 229 872.0(1.6)	3.7×10^{-15}	[52]	q_{42}
		429 228 004 229 873.56(0.49)	1.1×10^{-15}	[53]	q_{43}
		429 228 004 229 873.7(1.4)	3.3×10^{-15}	[54]	q_{44}
		429 228 004 229 873.13(0.17)	4.0×10^{-16}	[55]	q_{45}
		429 228 004 229 873.10(0.13)	3.1×10^{-16}	[56]	q_{46}
		429 228 004 229 872.92(0.12)	2.8×10^{-16}	[57]	q_{47}
		429 228 004 229 872.97(0.16)	3.7×10^{-16}	[58]	q_{48}
		429 228 004 229 873.04(0.11)	2.6×10^{-16}	[58]	q_{49}
		429 228 004 229 872.97(0.40)	9.3×10^{-16}	[59]	q_{50}
		429 228 004 229 872.99(0.18)	4.3×10^{-16}	[60]	q_{51}

(Continued.)

Table 2. (Continued.)

Clock transition	Atomic species	Measured frequency/Hz	Fractional uncertainty	Reference	Label
$\overline{ u_{13}}$	⁴⁰ Ca ⁺	411 042 129 776 393.2(1.0)	2.4×10^{-15}	[61]	q_{52}
		411 042 129 776 398.4(1.2)	2.9×10^{-15}	[62]	q_{53}
		411 042 129 776 400.5(1.2)	2.9×10^{-15}	[63]	q ₅₄
		411 042 129 776 401.7(1.1)	2.7×10^{-15}	[63]	q_{55}
$ u_{14}$	⁸⁷ Rb	$6834682610.9043129(3.0 \times 10^{-6})$	4.4×10^{-16}	[64]	<i>q</i> ₅₆
		$6834682610.9043070(3.1 \times 10^{-6})$	4.5×10^{-16}	[65]	q ₅₇
		$6834682610.9043125(2.1\times10^{-6})$	3.1×10^{-16}	[66]	q_{58}

^aAnother published absolute frequency measurement of ν_1 [67] was omitted from the input data set, because it is based on data that is apparently identical to that used in [13], but reports a much lower uncertainty of 63 Hz without explanation being provided.

Table 3. Frequency ratio measurement data used in the 2017 least-squares adjustment.

Frequency ratio	Atomic species	Measured ratio	Fractional uncertainty	Reference	Label
${\nu_{3}/\nu_{12}}$	¹⁹⁹ Hg/ ⁸⁷ Sr	2.629 314 209 898 909 60(22)	8.4×10^{-17}	[68]	<i>q</i> ₅₉
,		2.629 314 209 898 909 15(46)	1.7×10^{-16}	[19]	q_{60}
ν_3/ν_{14}	¹⁹⁹ Hg/ ⁸⁷ Rb	165 124.754 879 997 258(62)	3.8×10^{-16}	[19]	q_{61}
$ u_4/ u_5$	$^{27}\text{Al}^{+}/^{199}\text{Hg}^{+}$	1.052 871 833 148 990 438(55)	5.2×10^{-17}	[69]	q_{62}
$ u_6/ u_7$	171 Yb $^{+}$ E2/ 171 Yb $^{+}$ E3	1.072 007 373 634 206 30(36)	3.4×10^{-16}	[26]	q_{63}
$ u_8/ u_{12}$	¹⁷¹ Yb / ⁸⁷ Sr	1.207 507 039 343 3412(17)	1.4×10^{-15}	$[70, 71]^a$	q_{64}
		1.207 507 039 343 337 76(29)	2.4×10^{-16}	[72]	q_{65}
		1.207 507 039 343 337 749(55)	4.6×10^{-17}	[73]	q_{66}
$ u_{11}/ u_{12} $	88 Sr / 87 Sr	1.000 000 144 883 682 777(23)	2.3×10^{-17}	[74]	q_{67}
$ u_{12}/ u_{13} $	$^{87} {\rm Sr} / ^{40} {\rm Ca}^{+}$	1.044 243 334 529 6416(25)	2.4×10^{-15}	[62]	q_{68}
$ u_{12}/ u_{14}$	⁸⁷ Sr / ⁸⁷ Rb	62 801.453 800 512 435(21)	3.3×10^{-16}	[57]	q_{69}

^aThe value q_{64} is the corrected value given in [71], originating from the experiment described in [70].

Table 4. New absolute frequency measurements included in the 2021 least-squares adjustment. All uncertainties listed are the published ones, before adjustments made by the WGFS (see section 5).

Clock transition	Atomic species	Measured frequency/Hz	Fractional uncertainty	Reference	Label
$\overline{ u_1}$	¹¹⁵ In ⁺	1267 402 452 901 040.1(1.1)	8.5×10^{-16}	[75]	<i>q</i> ₇₄
$ u_4$	$^{27}Al^{+}$	1121 015 393 207 859.50(0.36)	3.2×10^{-16}	[76]	q_{97}
$ u_7$	¹⁷¹ Yb ⁺ E3	642 121 496 772 645.14(0.26) 642 121 496 772 645.10(0.08)	$4.0 \times 10^{-16} \\ 1.3 \times 10^{-16}$	[77] [78]	q_{71}
$ u_8$	¹⁷¹ Yb	518 295 836 590 863.30(0.38)	7.3×10^{-16}	[79]	q_{98} q_{70}
		518 295 836 590 863.71(0.11) 518 295 836 590 863.61(0.13)	$2.1 \times 10^{-16} $ 2.6×10^{-16}	[80] [81]	q_{75} q_{76}
$ u_{12}$	⁸⁷ Sr	518 295 836 590 863.54(0.26) 429 228 004 229 873.1(0.5)	5.0×10^{-16} 1.0×10^{-15}	[82] [83]	$q_{89} \\ q_{72}$
		429 228 004 229 873.00(0.07) 429 228 004 229 873.082(0.076)	$1.5 \times 10^{-16} \\ 1.8 \times 10^{-16}$	[84] [85]	q ₇₃ q ₉₀
		429 228 004 229 873.13(0.40)	9.3×10^{-16}	[86]	q_{91}
$ u_{13}$	$^{40}\mathrm{Ca}^+$	429 228 004 229 873.19(0.15) 411 042 129 776 400.41(0.23)	$3.5 \times 10^{-16} $ 5.6×10^{-16}	[76] [87]	q_{96} q_{88}
		411 042 129 776 400.6(0.5)	1.2×10^{-15}	[87]	q_{105}

Time (TAI) as a reference, with fractional uncertainties reaching the low parts in 10^{16} level. However the absolute frequency measurement with lowest reported fractional uncertainty (1.3×10^{-16}) was performed against local

caesium fountain primary frequency standards [78]. In total, seven absolute frequency measurements had reported fractional uncertainties below 3×10^{-16} , compared to just two in 2017. The most accurately measured optical frequency

^bThe value of q_6 is the corrected value given in [18], originating from the experiment described in [17].

^eThe value of q_{11} is the value published in [23], but corrected for the blackbody radiation shift.

Table 5. New frequency ratio measurements included in the 2021 least-squares adjustment. All uncertainties listed are the published ones, before adjustments made by the WGFS (see section 5).

Frequency ratio	Atomic species	Measured ratio	Fractional uncertainty	Reference	Label
${\nu_{1}/\nu_{12}}$	¹¹⁵ In ⁺ / ⁸⁷ Sr	2.9527487498748633(23)	7.7×10^{-16}	[75]	q_{78}
ν_3/ν_8	¹⁹⁹ Hg/ ¹⁷¹ Yb	2.177 473 194 134 565 07(19)	8.8×10^{-17}	[88]	q_{79}
$ u_4/ u_8$	$^{27}Al^{+}/^{171}Yb$	2.162 887 127 516 663 703(13)	5.9×10^{-18}	[89]	q_{104}
ν_4/ν_{12}	$^{27}\text{Al}^{+}/^{87}\text{Sr}$	2.611 701 431 781 463 025(21)	8.0×10^{-18}	[89]	q_{103}
ν_6/ν_7	$^{171}\text{Yb}^{+}\text{E2}/^{171}\text{Yb}^{+}\text{E3}$	1.072 007 373 634 205 469(37)	3.5×10^{-17}	[78] ^a	q 99
ν_7/ν_{12}	¹⁷¹ Yb ⁺ E3/ ⁸⁷ Sr	1.495 991 618 544 900 976(494)	3.3×10^{-16}	[90]	q_{84}
, , 12		1.495 991 618 544 901 113(404)	2.7×10^{-16}	[90]	q_{85}
		1.495 991 618 544 900 644(524)	3.5×10^{-16}	[90]	q_{86}
		1.495 991 618 544 900 858(299)	2.0×10^{-16}	[90]	q_{87}
		1.495 991 618 544 900 537(38)	2.5×10^{-17}	[91]	q_{92}
		1.495 991 618 544 900 840(344)	2.3×10^{-16}	[90]	q_{100}
		1.495 991 618 544 900 459(404)	2.7×10^{-16}	[90]	q_{101}
$ u_8/ u_{12}$	¹⁷¹ Yb/ ⁸⁷ Sr	1.207 507 039 343 337 90(70)	5.8×10^{-16}	[92]	q_{80}
·,		1.207 507 039 343 338 41(34)	2.8×10^{-16}	[86]	q_{81}
		1.207 507 039 343 338 05(34)	2.8×10^{-16}	[93]	q_{82}
		1.207 507 039 343 337 38(30)	2.5×10^{-16}	[93]	q_{83}
		1.207 507 039 343 338 58(49)	4.1×10^{-16}	[94]	q_{93}
		1.207 507 039 343 337 82(75)	6.2×10^{-16}	[82]	q_{94}
		1.207 507 039 343 337 8482(82)	6.8×10^{-18}	[89]	q_{102}
$ u_8/ u_{14}$	¹⁷¹ Yb/ ⁸⁷ Rb	75 833.197 545 114 174(42)	5.5×10^{-16}	[82]	<i>q</i> 95
-,		75 833.197 545 114 192(33)	4.4×10^{-16}	[80]	q_{106}
$ u_{11}/ u_{12}$	88 Sr/ 87 Sr	1.000 000 144 883 682 831(28)	2.8×10^{-17}	[95]	q_{77}

^aThe value reported here is the inverse of the one published in [78].




Figure 1. Input data to the 2021 least-squares adjustment. Absolute frequency measurements are indicated by blue solid lines, optical-microwave frequency ratios not involving caesium primary standards by grey dashed-dotted lines, and optical frequency ratio measurements by red dashed lines. The numbers on each line indicate how many measurements were available in each case.

ratio measurement (q_{104}) had a fractional uncertainty of 5.9×10^{-18} [89], with two others determined in the same campaign (q_{102} and q_{103}) also reaching fractional uncertainties below 1×10^{-17} .

4. Correlation coefficients

In contrast to previous least-squares adjustments performed by the WGFS, detailed consideration was given to correlations between the individual frequency ratio measurements, following the guidelines in [96], to ensure that the frequency values obtained from the analysis were unbiased and that their uncertainties were not underestimated. In total, 483 correlation coefficients were calculated and included in the analysis. These fell into two categories: firstly correlations arising through the use of the same PFS or SFS to access the SI second and secondly others that were computed on an ad-hoc basis. These categories are considered in the following sections 4.1 and 4.2 respectively.

The distribution of the magnitude of the correlation coefficients is shown in figure 2. Of the 483 coefficients, 300 have a magnitude greater than 0.01, while 99 have a magnitude greater than 0.1.

4.1. Correlation coefficients arising through access to the SI second

In this section, we estimate the correlation between absolute frequency measurements due to using the same PFS or SFS to access the SI second [96]. A procedure was developed to cover the measurements accessing the SI second through the duration of the TAI scale interval, d_{TAI} , published monthly by the BIPM in Circular T, in which case correlation arises through




Figure 2. Distribution of the magnitude of the 483 correlation coefficients included in the 2021 least-squares adjustment. Note the non-linear scale used to display the magnitudes.

the primary and secondary frequency standards (PSFSs) common to the different $d_{\rm TAI}$ estimations. The same scheme can also determine the correlation between one measurement using $d_{\rm TAI}$ and one using a local PSFS (or an ensemble thereof) to access the SI second, provided that the local PSFS is operated in similar conditions when performing the local comparisons and when operating for TAI reports. It can similarly be used for the correlation between two measurements using the same local standards. However, in those cases it is likely that other, possibly more important, sources of correlation are present so that those cases were generally covered by a specific study (see section 4.2).

For practical reasons the procedure was applied starting in January 2014. This choice of start date ensured consistency in the set of primary standards participating in TAI and in the provision of information relative to $d_{\rm TAI}$. It also eased the task of obtaining necessary information from the operators of PSFS as well as from those performing the measurements. Furthermore, it met the need to cover the most accurate absolute frequency measurements. The most recent measurements considered were carried out in March 2020.

Correlation between two measurements originating from using the same standards to access the SI second arises from the part of the standards' systematic uncertainty that is correlated from month to month, here denoted $u_{\rm bS}$, when the two measurements are performed in different months. When the two measurements occur during the same month, correlation is from the total uncertainty of the standards. The stationary part $u_{\rm bS}$ typically includes uncertainty contributions which are dominated by a model or calibration, examples being the blackbody radiation shift and the gravitational redshift. It

Table 6. Values of the stationary part u_{bS} of the systematic uncertainty of the PTB and SYRTE frequency standards over the years 2014–2020. For the PTB standards these were provided by S. Weyers, and for the SYRTE standards by M Abgrall and L Lorini.

Standard	Period	$u_{\rm bS}/10^{-16}$
PTB-CSF1	January 2014–February 2019	3.1
	March 2019–December 2020	2.0
PTB-CSF2	January 2014–December 2020	$Min(2.1, u_b)$
SYRTE-FO1	January 2014–February 2018	2.5
	March 2018-August 2018	2.3
	September 2018–December 2020	1.7
SYRTE-FO2	January 2014–September 2015	1.8
	October 2015–February 2018	1.5
	March 2018–December 2020	1.2
SYRTE-FORb	January 2014–September 2015	2.3
	October 2015–February 2018	2.1
	March 2018–December 2020	1.8

cannot readily be determined from information available in the published literature and a specific study is therefore necessary for each standard.

Because the frequency standards at PTB and LNE-SYRTE typically contribute about 90% of d_{TAI} over the period of interest and are also used as the local reference for a majority of the measurements considered here, their operators were approached for specific determinations of u_{bS} (table 6). Note that, in the case of SYRTE-FORb, the recommended uncertainty of the secondary representation of the second u_{Srep} was added to the values in table 6 to compute u_{bS} . For all other PSFS not included in table 6, u_{bS} is taken to be u_{b} .

The general equation to compute the correlation between measurements q_x and q_y due to using common PSFS in accessing the SI second is

$$r(q_x, q_y) = \frac{\sum_{i_x, i_y} w_{i_x} w_{i_y} \sum_{k} w_{i_x, k} u_{bS_{i_x, k}} w_{i_y, k} u_{bS_{i_y, k}}}{u_x u_y}$$
(1)

where i_x and i_y index the months used to access the SI second for measurements q_x and q_y and w_{i_x} and w_{i_y} are the weights of month i_x and i_y respectively. The PSFS are labelled with the index k.

For measurements accessing the SI second through d_{TAI} $w_{i,k}$ is the weight of standard k in the estimation of d_{TAI} for month i. The set of weights $w_{i,k}$ for all standards for each month i over the period January 2014–March 2020 (MJD 56 659–58 939) is collected from the monthly files etyy.mm (where yy.mm identifies the month), available on the BIPM ftp server [97], which provide the fractional frequency of the free atomic time scale EAL (Échelle Atomique Libre) as estimated from PSFSs.

For measurements accessing the SI second through local standards, $w_{i,k}$ is the weight of each local PSFS used in the comparisons. In some cases, exact information about these weights was obtained from the groups that performed the measurements, otherwise they were estimated from the publication (e.g. a plot of residuals indicating dates and uncertainties).

Table 7. List of absolute frequency measurements with relevant information used to compute correlation from access to the SI second.

Label	Transition	Lab.	No. of months	Standards
$\overline{q_3}$	$ u_1$	NICT	2	TAI
q_7	$ u_3$	SYRTE	2	FO2
q_{14}	$ u_6$	NPL	2	CsF2
q_{18}	$ u_7$	NPL	2	CsF2
q_{25}	$ u_8$	KRISS	1	TAI
q_{32}	$ u_{10}$	NRC	2	TAI
q_{34}	$ u_{11}$	Torun	1	TAI
q_{35}	$ u_{11}$	Torun	1	TAI
q_{43}	$ u_{12}$	NMIJ	1	TAI
q_{44}	$ u_{12}$	NIM	2	NIM5
q_{46}	$ u_{12}$	SYRTE	1 ^a	FO1,FO2,FOM
q_{47}	$ u_{12}$	SYRTE	4	FO1,FO2,FOM
q_{48}	$ u_{12}$	PTB	1	CSF2
q_{49}	$ u_{12}$	PTB	1	CSF1,CSF2
q_{50}	$ u_{12}$	NICT	3	TAI
q_{55}	$ u_{13}$	Wuhan	4	TAI
q_{56}	$ u_{14}$	SYRTE	37^{a}	TAI
q_{58}	$ u_{14}$	SYRTE	1	FO1,FO2,
				PTB-CSF1,
				PTB-CSF2
q_{70}	$ u_8$	ECNU	1	TAI
q_{71}	$ u_7$	NPL	1	TAI
q_{72}	$ u_{12}$	NPL	1	TAI
q_{73}	$ u_{12}$	PTB	10	CSF1,CSF2
q_{74}	$ u_1$	NICT	1	TAI
q_{75}	$ u_8$	NIST	8	TAI
q_{76}	$ u_8$	INRIM	5	TAI
q_{88}	$ u_{13}$	Wuhan	1	TAI
q_{89}	$ u_8$	NMIJ	6	TAI
q_{90}	$ u_{12}$	NICT	14	TAI ^a
q_{91}	$ u_{12}$	INRIM	2	CsF2
q 96	$ u_{12}$	NIST	4	TAI
q_{97}	$ u_4$	NIST	5	TAI
q_{98}	$ u_7$	PTB	10	CSF1,CSF2
q_{105}	$ u_{13}$	NIM	1	NIM5

^aSee text for further details.

The total uncertainties of the two measurements are u_x and u_y . However, when $i_x = i_y = i$ (two measurements made in a single common month), correlation is through the total uncertainty $u_{i,k}$ of the standard k in the estimation of d_{TAI} for this month, not through $u_{bS_{i,k}}$. In all cases u_{bS} may vary with time so the month of operation needs to be specified.

This computation was carried out for 34 absolute frequency measurements, 33 of which took place in the period considered. Measurement q_{46} [56] was added even though it dates from 2010 to 2011 because of its small uncertainty and significant correlation with several more recent measurements, To compute correlation coefficients involving q_{46} , the u_{bS} values of January 2014 were used for the relevant SYRTE standards. Similarly some of the numerous measurements of q_{56} were taken before 2014, and were assigned to the first months of 2014 by giving a larger weight to those months. The access to the SI second was modelled as follows (see table 7 for more details):

- For 19 measurements accessing the SI second through $d_{\rm TAI}$, by specifying the list of months along with the weight assigned to each month in the determination. Of these, nine measurements correspond to the simple case where the measurement was performed in a single month.
- For 14 measurements accessing the SI second through local standards, by assigning the ensemble of individual comparisons to a set of months and estimating for each month a weight which was then shared between the local PSFS used during that month. Note that measurement q₅₈ (an absolute frequency measurement of the SYRTE Rb fountain) included comparisons to the remote PTB Cs fountains in addition to the local Cs fountains.
- In the specific case of q_{90} [85], the authors estimated 63 comparisons of their optical clock to eight individual PFS using data published by the BIPM. Because the PFS used in the comparisons are those providing the estimation of d_{TAI} and their contributions to the absolute frequency measurement

were adequately provided with monthly values in figure 4b of [85], this measurement was introduced as if accessing through 14 monthly d_{TAI} values with the monthly weights taken from figure 4(b).

The computation of correlation coefficients also took into account revised total uncertainties for some absolute frequency measurements, as detailed in section 5. This concerns q_{73} and q_{98} following a specific computation (section 5.1), and q_{74} , q_{88} and q_{105} for which the uncertainty was enlarged (see section 5.2). The computation yielded 561 coefficients, 399 of which were larger than 0.001 and which were used in the least-squares analysis. Two of them $(r(q_{14}, q_{18})$ and $r(q_{50}, q_{90}))$ were not used because a specific calculation provided a more accurate estimation (see appendices A.2 and A.7). The complete list of correlation coefficients may be accessed at [98].

4.2. Correlation coefficients computed on an ad-hoc basis

Based on a review of the input data, a total of 86 additional correlation coefficients were identified to be potentially significant and were therefore computed on an adhoc basis. These correlations originate from several different sources.

Firstly, significant correlations are likely to arise if an atomic clock participates in more than one frequency comparison during the same period. In this scenario, the correlation coefficient will typically have contributions coming from both the statistical and the systematic uncertainties of the common clock. Measurements involving the same atomic clock, but performed at different times, may also be correlated if the systematic uncertainty of the clock is not re-evaluated, or is only partially re-evaluated, between the two measurements. Although the uncertainty budgets of optical clocks typically evolve more rapidly than in the case of the caesium fountain primary frequency standards considered in section 4.1, some instances of this type of correlation were identified. For example the uncertainty evaluation of the Hg optical lattice clock at RIKEN was largely common between the ¹⁹⁹Hg/⁸⁷Sr frequency ratio measurement q_{59} , published in 2015 [68] and the 199 Hg/ 171 Yb frequency ratio measurement q_{79} , published in 2020 [88], resulting in an estimated correlation coefficient of 0.826 (table A1).

Secondly, correlations may arise between any clocks based on the same atomic transition, if the same theoretical or experimental values of atomic coefficients are used to correct for systematic frequency shifts such as Zeeman or blackbody radiation shifts. Finally, correlations associated with systematic corrections may also arise between measurements involving clocks that are based on different atomic transitions, but that are located within the same laboratory. An example would be the case of the gravitational redshift correction, which would be largely common to any remote comparisons involving clocks in that laboratory.

Further details about the computation of correlation coefficients arising from these sources are provided in appendix A.

Table 8. Changes made to the uncertainties of the input data to the least-squares adjustment, compared to published uncertainties. Note that these differ in some cases from the changes made in 2017. The rationales listed are described in more detail in sections 3 (exclude from fit), 5.1 (unphysical correlation coefficients)), 5.2 (improved consistency) and 5.3 (sparsity of data).

Measurement label	Change to published uncertainty	Rationale
	uncertainty	Kationaic
q_1	Multiplied by 3	Improved consistency
q_{31}	Multiplied by 1.5	Sparsity of data
q_{51}	Multiplied by 10 ⁶	Exclude from fit
q_{52}	Multiplied by 6	Improved consistency
q_{73}	Increased to 1.65×10^{-16}	Unphysical correlation
		coefficient
q_{74}	Multiplied by 3	Sparsity of data
q_{78}	Multiplied by 3	Sparsity of data
q_{88}	Multiplied by 2	Sparsity of data
q_{98}	Increased to 1.6×10^{-16}	Unphysical correlation
		coefficient
q_{105}	Multiplied by 2	Sparsity of data

5. Modifications to the input data

Review of the input data by the WGFS resulted in one modification to an input frequency value, compared to the published value, and several modifications to the uncertainties of particular input values.

In evaluating the correlation coefficients arising through access to the SI second, it was realised that the absolute frequency measurement of $^{115}\text{In}^+$ reported in [14] was carried out in three sessions, but that only two of these used TAI as a reference, with the value obtained in the last session being based on the evaluation of the NICT ^{87}Sr lattice clock (q_{50}) reported in [59]. The group that performed the measurements was therefore asked by the WGFS to recompute the $^{115}\text{In}^+$ frequency using data only referenced to TAI, and the resulting value of 1 267 402 452 901 049.8(7.5) Hz, rather than the published value listed in table 2, was used for q_3 in the least-squares adjustment.

Modifications were also made to the uncertainties of a few measurements in the input data set. As already stated in section 3, the uncertainty of q_{51} was multiplied by 10^6 to effectively exclude it from the least-squares adjustment, because the same data was used to contribute to q_{90} . However modifications to uncertainties were also made for several other reasons:

- To avoid otherwise unphysical values of correlation coefficients.
- (ii) To handle particular data points that are identified as outliers, making them statistically more consistent with the overall body of data.
- (iii) To take account of the sparsity of the input data for some atomic transitions included in the adjustment.

These adjustments are described in more detail in sections 5.1–5.3 below, and summarized in table 8.

5.1. Avoiding unphysical correlation coefficients

Unphysical correlation coefficients may arise if different uncertainties are used and reported for the same frequency standard during the same period, when that standard is used for different purposes. Initial runs to determine correlation coefficients as described in section 4.1 revealed several large values, with a few of them larger than 1. These large coefficients mostly involved measurements q_{73} and q_{98} , with the largest value for $r(q_{73}, q_{98})$. Both measurements correspond to the determination of an optical frequency with respect to PTB primary standards CSF1 and CSF2 from a few tens of comparisons carried out over the period 2017-2019. In both cases, the details of the comparisons are well documented, q_{73} through table II in [84] and q_{98} through specific files transmitted by the PTB group to the WGFS [98]. It was realized that the systematic uncertainty $u_{b,Cs}$ associated with PTB-CSF1 is in both cases significantly lower than the systematic uncertainty value $u_{\rm b}$ in the reports of the monthly evaluation of TAI frequency by PTB-CSF1 [99] and also significantly lower than the stationary part of its systematic uncertainty, u_{bS} , as defined in section 4.1. On the other hand, the treatment for CSF2 is quite in line with the values of u_b reported for TAI and with the values of u_{bS} that we used to compute correlations.

Because of the lower $u_{b,Cs}$ values for CSF1, it is not possible to use the standard equation (1) to compute correlations with q_{73} and q_{98} . However, these are the two most accurate determinations of absolute frequencies to date and need to be fully taken into account. We therefore resolved to increase the total uncertainty of both measurements to the values that would be obtained if replacing for PTB-CSF1 the $u_{b,Cs}$ values by the u_{bS} values as defined in section 4.1. The resulting total uncertainties, used in the final analysis, are 1.65×10^{-16} for q_{73} and 1.6×10^{-16} for q_{98} , rather than the published uncertainties of 1.5×10^{-16} and 1.3×10^{-16} respectively.

Note that the main effect of this change is a higher uncertainty for the CSF1 measurements, especially those taken before February 2019 when $u_{\rm bS}=3.1\times10^{-16}$. It also results in a lower weight for the CSF1 measurements in the determination of the absolute value of the optical frequencies, meaning that in principle the values could change. This change is estimated to be lower than 3×10^{-17} for both measurements and it was therefore decided not to change the absolute frequency values for the 2021 adjustment. Using the modified total uncertainties, all correlation coefficients computed have sensible values, with the largest coefficient $r(q_{73}, q_{98})$ being 0.729.

5.2. Improving consistency of the input data set

The input data set includes several measurements of some quantities, the consistency of which can be checked directly [1]. However a more complete and rigorous evaluation of the consistency of the input data was performed by carrying out a preliminary least-squares adjustment to identify outliers in the input data set. This was achieved through inspection of the normalised residuals $\rho_i = (q_i - \hat{q}_i)/u_i$, where the \hat{q}_i are the optimised frequency values obtained from the fit.

This analysis indicated that the biggest outlier in the complete input data set was q_{52} , with $\rho_{52} = -6.90$. To make q_{52} statistically more consistent with other measurements, its published uncertainty was therefore multiplied by six in the input data set.

The other significant (>3 σ) outlier identified in this way was q_1 , with $\rho_1 = -4.87$. For consistency with the previous 2017 adjustment, the published uncertainty of q_1 was therefore multiplied by three in the final input data file.

5.3. Accounting for sparsity of the input data

The WGFS also request and take into account information pertaining to comparison of frequency standards based on the same atomic transition, but such data is not always available. As discussed by Riehle et al [1], the WGFS have developed procedures which take a cautious approach to uncertainty estimation in the case that a particular recommended frequency value is determined by very few, or even a single, input frequency value. Historically, whilst the recommended frequency values were derived from only absolute frequency measurements, this involved applying enlargement factors to the published uncertainty (in the case of a single input value) or to the uncertainty of the weighted mean (in the case of two independent input values). Although these procedures were previously amended to some degree to take account of the availability of high accuracy direct frequency ratio measurements, making ad-hoc adjustments to the uncertainties of selected optimised frequencies from the least-squares adjustment is unsatisfactory because it leads to inconsistencies in the complete output data set, which includes frequency ratios between every pair of atomic transitions. For this reason we departed from prior practice and made selected adjustments to the input data instead, followed later by a global enlargement (by a factor of two) to the uncertainties of every ratio in the output data set (discussed in more detail in section 6). This change in procedure ensured that the output data set we obtain is internally self-consistent.

The adjustments made to account for sparsity of the input data are as follows:

(i) In the case of the ⁸⁸Sr⁺ optical clock, the only published data consisted of absolute frequency measurements, meaning that this transition is decoupled from the rest of the analysis. none of the available measurements were new since 2017. In this case the frequency obtained from the least-squares adjustment is equivalent to taking a weighted mean of the measurements, and is dominated by q_{31} , which has a fractional uncertainty of 5.4×10^{-16} , compared to the next most precise measurement q_{32} with fractional uncertainty 1.5×10^{-15} . For this reason, in 2017, the uncertainty of the value from the fit was increased by a factor of three. To achieve a similar result in the 2021 update to the list of recommended frequency values, taking into account the global uncertainty enlargement by a factor of two (discussed in section 6), we increased the uncertainty on the input data point q_{31} by a factor of 1.5.

(ii) The 115 In⁺ clock transition frequency obtained from the least-squares adjustment is determined almost entirely by two defining measurements q_{74} and q_{78} with fractional uncertainties far lower than the other measurements involving this standard $(q_1, q_2 \text{ and } q_3)$. These two defining measurements, an absolute frequency measurement and an optical frequency ratio measurement against 87 Sr were performed during the same period in a single laboratory (NICT), and are strongly correlated, with a correlation coefficient $r(q_{74}, q_{78}) = 0.859$. For this reason the published uncertainties of both q_{74} and q_{78} were multiplied by three in the final input data file, whilst keeping $r(q_{74}, q_{78})$ unchanged.

(iii) For the optical clock transition in 40 Ca $^+$, the frequency obtained from the final least-squares adjustment is dominated by two new absolute frequency measurements q_{88} and q_{105} originating from a single laboratory and from the same measurement period. The published uncertainties of these measurements were therefore multiplied by two in the input data set. A factor of two rather than three was used in this case, because published comparisons between two 40 Ca $^+$ optical clocks [63, 87] support the uncertainty evaluation for the clock used in these measurements.

No new data was available for either the 1S–2S transition in ¹H or the optical clock transition in ⁴⁰Ca. Since these two transitions are linked to the other input data only via Cs, their recommended frequencies and uncertainties should remain the same as in 2017. For convenience, in these two cases, we left the input data unchanged and simply removed the transitions from the output covariance matrix, after verifying that the least-squares adjustment did indeed give the same values for the transition frequencies as in 2017.

Although no other adjustments were made to the published uncertainties in the input data, we note the following points:

- The optimized frequency for the ¹⁹⁹Hg optical clock transition is mainly determined by frequency ratio measurements from RIKEN involving other better-known optical frequencies, especially that of the ⁸⁷Sr clock transition. However measurements from a second laboratory (LNE-SYRTE) were also available, and all normalised residuals were observed to have a magnitude less than one.
- In the case of the 27 Al⁺ optical clock transition, all measurements originate from a single laboratory (NIST) and the frequency obtained from the least-squares adjustment is mainly determined by measurements q_{103} and q_{104} which were made during the same measurement campaign and have a correlation coefficient $r(q_{103}, q_{104}) = 0.329$. However all normalised residuals except for that for q_8 , which has little weight in the adjustment, were observed to have a magnitude less than one.
- The optimized frequency for the 199 Hg⁺ optical clock transition is determined almost entirely by a single optical frequency ratio measurement q_{62} . The only absolute frequency measurement available, q_9 , has a normalised residual with

- magnitude >2 but has little weight in the least-squares adjustment.
- The optimized frequency for the E2 optical clock transition in ¹⁷¹Yb⁺ is determined almost entirely by a frequency ratio measurement against the E3 transition, q_{99} , which has an uncertainty an order of magnitude lower than any of the other measurements involving the E2 transition.
- Although three absolute frequency measurements of the optical clock transition in 88 Sr were available, the frequency obtained from the least-squares adjustment is essentially completely determined by two measurements of the ratio between the 88 Sr and 87 Sr clock transitions (q_{67} and q_{77}), which have similar fractional uncertainties (2.3×10^{-17} and 2.8×10^{-17} , respectively). These measurements were made in different groups, and differ by 1.5 times their combined relative uncertainties.

The focus in this analysis was on sparsity of input data that affects the recommended frequency values. Some optical frequency ratios are also determined by one or two input measurements, but no adjustments were made to input data in such cases, meaning that the uncertainties assigned to optical frequency ratio values may perhaps be less conservative than those assigned to the recommended frequency values.

6. Results from the least-squares adjustment

Once the input data had been finalised, the analysis software was run for a last time to calculate optimised values for each frequency ratio and absolute frequency value, and to check the self-consistency of the input data. The results obtained using the different algorithms and software were compared to verify the results. The results obtained using the two different implementations of the least-squares algorithms yielded absolute frequency values and frequency ratio values that differed at most by one in the least-significant (24) digit of the computation, while the uncertainties were identical to the four significant figures computed. Slightly larger differences were seen in some cases between the results obtained using the two different algorithms. However even in this case the maximum differences observed in absolute frequency values or frequency ratio values were 2 parts in 10²¹ while uncertainties differed by no more than 2 in the least significant digit of the four computed. The output correlation coefficients between the absolute frequency values computed using the two algorithms agreed to better than 1 part in 10⁵. This level of agreement between the different algorithms and software is several orders of magnitudes below the uncertainties on the output values. When appropriate truncation and rounding is applied to obtain recommended values and uncertainties, all computations give identical recommended values and uncertainties. The computation time for the two approaches was very similar, being of the order of 10 s.

The importance of accounting for correlations in the analysis is clearly visible in figures 3 and 4: including correlations




Figure 3. Uncertainty of the 2021 adjusted frequency values for ten optical frequency standards, plotted on a logarithmic scale, showing that the effect of neglecting correlations is to underestimate the uncertainty. Also shown for comparison are the uncertainties of the 2017 recommended frequency values, though it should be noted that these are typically larger than the uncertainty from the 2017 fit, due to the cautious approach to uncertainty estimation taken by the WGFS.




Figure 4. 2021 adjusted frequency values and uncertainties for ten optical frequency standards (green circles), showing that neglecting correlations (blue crosses) can bias the frequency obtained from the fit. Also shown are both the 2017 and 2021 recommended frequency values and uncertainties (grey squares and purple diamonds, respectively). All ten optical frequency standards are now secondary representations of the second. Note that small differences observed between some of the 2021 recommended frequency values and the values obtained from the 2021 least-squares adjustment are due to rounding.

leads to uncertainties up to 60% higher than if correlations are neglected and shifts the adjusted frequency values by up to 70% of their uncertainty. The Birge ratio for the final least-squares adjustment (including correlations) was 1.064 and the goodness-of-fit was estimated to be 0.18, providing an important check on the overall consistency of the set of input measurements (with the modifications described in section 5).

Bearing in mind the purpose and applications of the list of recommended frequency values, in particular the use of SRS for calibration of the scale interval of TAI, the WGFS is concerned to avoid any discontinuities in the values, i.e. to ensure that any changes between adjustments have a magnitude less than the combined uncertainties of the old and new values. This motivates a cautious approach to the derivation of uncertainties, based on the informed judgement of the membership, as previously elucidated in [1].

The least-squares adjustment procedure used to analyse the over-constrained input data set implicitly assumes that the probability density of the input data is well approximated by a Gaussian distribution, which is not well justified or easily tested in cases where the output values are determined by just one or two input measurements. Furthermore, it

Metrologia 61 (2024) 035005	H S Margolis <i>et al</i>
------------------------------------	---------------------------

		1 1	
Transition label	Atomic species	2021 recommended frequency value/Hz	Recommended fractional uncertainty
$\overline{ u_1}$	¹¹⁵ In ⁺	1267 402 452 901 041.3	4.3×10^{-15}
ν_3	¹⁹⁹ Hg	1128 575 290 808 154.32	2.4×10^{-16}
$ u_4$	$^{27}\text{Al}^{+}$	1121 015 393 207 859.16	1.9×10^{-16}
ν_5	$^{199} Hg^{+}$	1064 721 609 899 146.96	2.2×10^{-16}
ν_6	¹⁷¹ Yb ⁺ (E2)	688 358 979 309 308.24	2.0×10^{-16}
$ u_7$	$^{171}{\rm Yb}^{+}$ (E3)	642 121 496 772 645.12	1.9×10^{-16}
$ u_8$	¹⁷¹ Yb	518 295 836 590 863.63	1.9×10^{-16}
$ u_{10}$	$^{88}{ m Sr}^{+}$	444 779 044 095 486.3	1.3×10^{-15}
ν_{11}	⁸⁸ Sr	429 228 066 418 007.01	2.0×10^{-16}
ν_{12}	⁸⁷ Sr	429 228 004 229 872.99	1.9×10^{-16}
ν_{13}	40 Ca $^+$	411 042 129 776 400.4	1.8×10^{-15}
ν_{14}	⁸⁷ Rb	6834 682 610.904 3126	3.4×10^{-16}
17			

Table 9. The twelve recommended frequency values updated in March 2021 [3].

assumes that correlations between the input measurements are precisely known and taken into account. Although significant effort was devoted to identifying and estimating such correlation coefficients in the 2021 analysis, it remains the case that as-yet-unidentified sources of uncertainty or correlations may exist.

For this reason the WGFS maintained the previous cautious approach to uncertainty estimation, but departed from previous practice by applying a global expansion factor to the output covariance matrix rather than applying separate expansion factors to the uncertainties of selected recommended frequency values. This global expansion factor achieves the important outcome that the set of recommended frequency values and the ratios between them are internally self-consistent. On this occasion the expansion factor selected corresponded to a multiplication by two of the raw fit uncertainties. The recommended frequency value itself is the result obtained directly from the least-squares adjustment, rounded as appropriate given the magnitude of the recommended uncertainty. The uncertainties of the recommended frequency values should be understood and used as estimated standard uncertainties, corresponding to a coverage interval of 68.27%, i.e. the expansion factor applied is intended to allow for as-yet-unidentified sources of uncertainty or correlation, and is separate from any coverage factor used to calculate expanded uncertainties or confidence intervals as described in [100].

The global expansion factor by a factor of two was also noted by the WGFS to have the effect, considered desirable, of avoiding recommended frequency values with uncertainties significantly lower than the uncertainty of any individual realisation of the SI second so far. We acknowledge, however, that there remains some degree of arbitrariness in deciding on this expansion factor, and an important topic for future discussion by the WGFS would be possible approaches to put this on a more objective footing, perhaps through the use of the Student's t-distribution [1].

The updated recommended frequency values and uncertainties were approved at the 22nd meeting of the CCTF in

March 2021 [3], and are listed in table 9. At the same time, two additional reference transitions (in ⁸⁸Sr and ⁴⁰Ca⁺) were approved as optical SRS, bringing the total to ten. These additions reflected the improved uncertainties for their recommended frequency values and the fact that prospects for using standards based on these atomic transitions for future contributions to TAI were considered to be good.

For the ten optical SRS, the 2021 recommended frequency values are compared with the 2017 values in figure 4. The most significant changes in the recommended frequency values are for ²⁷Al⁺ and ¹⁹⁹Hg⁺, but in all cases the new values lie within the uncertainties on the 2017 values. It is worth noting that the uncertainty of the recommended frequency value for ¹⁹⁹Hg⁺ is significantly reduced even though there is no new data for this transition. This is because in 2017, the ²⁷Al⁺/¹⁹⁹Hg⁺ ratio essentially determined the absolute frequency of ²⁷Al⁺ and had almost no effect on the absolute frequency of ¹⁹⁹Hg⁺, while the new data available for ²⁷Al⁺ in 2021 means that the ²⁷Al⁺/¹⁹⁹Hg⁺ ratio now reduces the uncertainty of the ¹⁹⁹Hg⁺ frequency.

The effect of the increased number of optical frequency ratio measurements in the input data set, many of which have uncertainties significantly smaller than any measurement involving caesium primary frequency standards, is to create strong correlations between most of the recommended frequency values, as indicated in figure 5. With the exception of correlation coefficients involving ⁸⁸Sr⁺ and ⁴⁰Ca⁺, all correlation coefficients between recommended frequency values for optical SRS are greater than 0.65, and ten are greater than 0.95. To calculate any other frequency ratio from the recommended frequency values, it is necessary to take account of these correlations by using the output covariance matrix in order to compute the uncertainty of that frequency ratio correctly [100]. For completeness, we therefore list values and uncertainties for each frequency ratio in appendix B. The need to take correlations into account similarly applies to any other quantities depending on or combining these frequency ratios.




Figure 5. Visualisation of the correlation matrix between the 2021 recommended frequency values. The colour in the heat map indicates the size of the correlation coefficient between each pair of recommended frequency values.

7. Conclusion

Including correlations in the analysis underpinning the 2021 update to the list of recommended frequency values, whilst essential for a proper treatment of the data, significantly increased the effort required from the group performing the analysis on behalf of the WGFS. This is because the vast majority of the correlation coefficients were estimated by this group. In many cases the publications describing the measurements do not contain sufficient detail to identify all potential sources of correlation, and although efforts were made to gather this information when measurements were submitted for consideration by the WGFS, time constraints limited the amount of follow-up interaction that was possible. To prepare for future updates to the list, it would be beneficial to invest time in improving practices that will facilitate the computation [101], for example by encouraging groups to compute and submit correlation coefficients themselves, at least between different measurements performed within any given institution, or within coordinated comparison campaigns.

As a result of the 2021 update to the list of recommended frequency values, six optical standards now have recommended uncertainties of 1.9×10^{-16} to 2.0×10^{-16} , essentially the same as that of the best caesium fountain primary frequency standards. This establishes a solid link between optical frequency standards and the current definition of the second, which will be important to ensure continuity at the point of a

future redefinition. Prior to that redefinition, it also increases the weight that optical SRS can have in the steering of TAI. In recent years, the very best reported evaluations of the frequency of TAI with SFS have a total uncertainty (excluding $u_{\rm Srep}$) of order 2×10^{-16} . With the uncertainties from the 2017 adjustment their weight in estimating d_{TAI} was typically about 4% for ¹⁷¹Yb clocks and about 6% for ⁸⁷Sr clocks resulting, in the case of three evaluations in a month, in a typical total contribution to d_{TAI} of about 10%. The uncertainties from the 2021 adjustment were first used in April 2022 and the typical weight of individual contributions is now between 10% and 16% with the total weight for three evaluations per month being between 16% and 26%. The number of evaluations is also tending to increase so that SFS may eventually contribute as much as Cs fountains in estimating d_{TAI} . Note that, in a future adjustment of the standard frequency values, this large contribution of SFS to d_{TAI} will complicate the evaluation of correlations for the cases when d_{TAI} is used to access the SI second. In the present work we could neglect this effect as the contributions of SFS were quite rare before March 2020 (with the notable exception of the SYRTE 87Rb fountain) and they very rarely contributed more than 10% in any given month.

Optical frequency ratio measurements have a vital role to play in verifying the international consistency of optical clocks at a level better than 5×10^{-18} , one of the criteria set out in the international roadmap towards the redefinition of the SI second [102]. In the 2021 update to the recommended frequency values, the rules and criteria previously employed by the WGFS were modified to ensure that the output data from the least-squares adjustment includes a complete set of frequency ratio values (including absolute frequencies as a special case) that are internally self-consistent. Two different options are currently envisaged for a redefinition of the second, which might be based on a single reference optical frequency in a similar way to today's caesium-based definition, or might instead be based on an ensemble of reference frequencies as proposed in [103]. Whichever option is eventually selected, one or more of the frequency ratios from a similar least-squares adjustment will be used to set the defining constant or constants appearing in the new definition, and the frequency ratios will also play an important role in the Mise en pratique for the new definition. Increasing scrutiny will therefore need to be paid to the evolution of these frequency ratio values in subsequent updates to the list of recommended frequency values, to ensure the stability of the new definition and realisation of the second.

Data availability statement

The data that support the findings of this study are openly available. The digital object identifier of the dataset is DOI:https://doi.org/10.59161/StdFreq202301.

Acknowledgment

We thank S Weyers, M Abgrall, L Lorini, M Pizzocaro, D B Hume, N Nemitz, H Hachisu, N Ohtsubo and T Kobayashi

for providing information to support the evaluation of correlation coefficients, as well as P Tavella, N Dimarcq and members of the CCL-CCTF WGFS for helpful discussions. H M and S B acknowledge funding from the European Metrology Programme for Innovation and Research (EMPIR) Project 18SIB05 ROCIT, co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme. H M also acknowledges funding from the UK Department for Science, Innovation and Technology as part of the National Measurement System Programme, and thanks the BIPM for supporting a one-month secondment.

Appendix A. Details of correlation coefficients computed on an ad-hoc basis

In this appendix, we provide details of the 86 correlation coefficients that were computed on an ad-hoc basis. The values of these correlation coefficients are listed in table A1.

A.1. Measurements involving PTB 171 Yb $^+$ and 87 Sr optical clocks

The largest correlation coefficient in the input data set is between q_{19} [30] and q_{45} [55]. These two absolute frequency measurements, of the $^{171}\mathrm{Yb^+}$ E2 and the $^{87}\mathrm{Sr}$ optical clock transitions, were performed at the same time, and the uncertainty arising from the local Cs fountains PTB-CSF1 and PTB-CSF2 completely dominates over the systematic uncertainty of either optical clock, resulting in a correlation coefficient $r(q_{19},q_{45})=0.981$ [96]. Other much smaller correlations exist between another set of measurements performed with the PTB $^{171}\mathrm{Yb^+}$ and $^{87}\mathrm{Sr}$ optical clocks (q_{73},q_{92},q_{98}) and q_{99} [78, 84, 91]), as the systematic uncertainty budgets of the clocks were common to the different measurements.

A.2. Measurements involving $^{171}{\rm Yb}^+$ trapped ion clocks at NPL and PTB

Measurements q_{14} , q_{18} and q_{63} were all performed in the same campaign involving the NPL ¹⁷¹Yb⁺ trapped ion optical clock [26]. The three correlation coefficients between these measurements are the only correlation coefficients that were included in the 2017 least-squares adjustment, and have contributions from both the statistical and systematic uncertainties of the clocks involved [96]. These measurements are also (less significantly) correlated with measurements q_{13} , q_{16} and q_{19} made at PTB [25, 28, 30] because the blackbody radiation shifts for the NPL ¹⁷¹Yb⁺ clock were calculated using experimental values for the differential polarizabilities of the atomic states determined at PTB [96].

A.3. Measurements performed in June 2015

Fourteen measurements in the input data set originate from June 2015, having been performed during a coordinated European clock comparison campaign. Five of these $(q_7, q_{47}, q_{49}, q_{58})$ and q_{71} are absolute frequency measurements of

¹⁹⁹Hg [19], ⁸⁷Sr [57, 58], ⁸⁷Rb [66] and ¹⁷¹Yb⁺ (E3) [77] atomic clocks. Three (q_{60} , q_{61} and q_{69}) are local frequency ratio measurements between clocks at LNE-SYRTE [19, 57], and the remainder (q_{84} – q_{87} , q_{100} and q_{101}) are remote optical frequency ratio measurements between ¹⁷¹Yb⁺ (E3) and ⁸⁷Sr optical clocks at NPL, LNE-SYRTE and PTB [90]. Correlations were largely considered to be due to systematic uncertainties that were common to different measurements, with the largest coming from the Rb fountain at LNE-SYRTE, closely followed by the their ¹⁹⁹Hg optical clock [96]. However contributions from statistical uncertainties were also estimated and included in some cases.

A.4. Optical frequency ratio measurements at NIST/JILA

The frequency ratio measurements q_{102} , q_{103} and q_{104} , which have the lowest uncertainties in the 2021 input data set, were performed using the ²⁷Al⁺, ¹⁷¹Yb and ⁸⁷Sr optical clocks at NIST and JILA. Correlations arise between pairs of ratios $(^{27}Al^{+}/^{171}Yb, ^{171}Yb/^{87}Sr \text{ and } ^{27}Al^{+}/^{87}Sr)$ due to the common atomic species and overlapping data. Correlations due to both systematic uncertainty and statistical noise were considered. The correlation due to statistical noise depends on the instability of each clock individually and the fraction of overlapping data. While the exact instability of each clock is unknown, the instabilities of the lattice clocks contribute negligibly to the correlation coefficients and were assumed to be equal. On the other hand, the ²⁷Al⁺ clock instability dominates the statistical noise for the ²⁷Al⁺ ratios so it can be determined from the measurements and contributes significantly to the correlation coefficient for those ratios. There is additional correlation due to fluctuations from the density shift in 87Sr (evaluated daily) and the observed excess scatter in the ²⁷Al⁺/¹⁷¹Yb and ¹⁷¹Yb/⁸⁷Sr ratios. Both were included as additional uncertainties acting day-to-day. The level of excess scatter was chosen to make the reduced chi-squared value equal to 1 for both ²⁷Al⁺/¹⁷¹Yb and ¹⁷¹Yb/⁸⁷Sr and is consistent with the histograms determined from Bayesian analysis in [89].

A.5. Measurements involving the INRIM 171 Yb lattice clock

Measurements q_{24} , q_{76} , q_{81} , q_{82} and q_{83} all involve the INRIM ¹⁷¹Yb lattice clock. Measurements q_{76} , q_{82} and q_{83} were obtained in the same period of time, and are therefore correlated through the systematic uncertainty of the ¹⁷¹Yb clock, as well as through extrapolations of maser noise over common periods of dead time. Measurements q_{82} and q_{83} are both measurements against the NICT ⁸⁷Sr optical lattice clock, and hence correlated through the systematic uncertainty of that clock. The systematic uncertainty of the INRIM ¹⁷¹Yb lattice clock was re-evaluated in 2019 so there are no significant correlations between earlier and later measurements.

A.6. Measurements involving optical clocks at RIKEN

Frequency ratio measurements q_{59} , q_{66} and q_{79} [68, 73, 88] involve the ⁸⁷Sr, ¹⁷¹Yb and ¹⁹⁹Hg optical clocks at RIKEN. The correlation coefficient between the two measurements

Table A1. Correlation coefficients computed on an ad-hoc basis.

		_		
Measurements involving	¹⁷¹ Yb ⁺ and ⁸⁷ Sr optical clock	cs at PTB		
$r(q_{19}, q_{45}) = 0.981$	$r(q_{73}, q_{92}) = -0.060$	$r(q_{92}, q_{98}) = 0.002$	$r(q_{92}, q_{99}) = -0.009$	$r(q_{98}, q_{99}) = -0.002$
Measurements involving	¹⁷¹ Yb ⁺ optical clocks at NPL	and PTB		
$r(q_{13}, q_{14}) = 0.030$ $r(q_{16}, q_{19}) = 0.006$	$r(q_{13}, q_{63}) = 0.060$ $r(q_{16}, q_{63}) = -0.007$	$r(q_{14}, q_{18}) = 0.680$ $r(q_{18}, q_{19}) = 0.009$	$r(q_{14}, q_{63}) = 0.507$ $r(q_{18}, q_{63}) = -0.018$	$r(q_{16}, q_{18}) = 0.004$ $r(q_{19}, q_{63}) = -0.015$
Measurements performed	d in June 2015			
	$r(q_7, q_{61}) = 0.221$ $r(q_{47}, q_{100}) = -0.026$ $r(q_{58}, q_{69}) = -0.713$ $r(q_{60}, q_{100}) = 0.041$ $r(q_{69}, q_{100}) = -0.022$ $r(q_{71}, q_{87})) = 0.151$ $r(q_{84}, q_{101}) = 0.100$ $r(q_{86}, q_{87})) = 0.178$	$r(q_{47}, q_{60}) = -0.033$ $r(q_{47}, q_{101}) = -0.022$ $r(q_{60}, q_{61}) = 0.456$ $r(q_{60}, q_{101}) = 0.035$ $r(q_{69}, q_{101}) = -0.019$ $r(q_{84}, q_{85}) = 0.155$ $r(q_{85}, q_{86}) = 0.128$ $r(q_{100}, q_{101}) = 0.027$	$r(q_{47}, q_{69}) = 0.018$ $r(q_{49}, q_{86}) = -0.004$ $r(q_{60}, q_{69}) = -0.028$ $r(q_{61}, q_{69}) = 0.597$ $r(q_{71}, q_{84}) = 0.092$ $r(q_{84}, q_{86}) = 0.105$ $r(q_{85}, q_{87}) = 0.224$	$r(q_{47}, q_{84}) = -0.018$ $r(q_{49}, q_{87}) = -0.007$ $r(q_{60}, q_{84}) = 0.028$ $r(q_{69}, q_{84}) = -0.015$ $r(q_{71}, q_{85}) = 0.112$ $r(q_{84}, q_{87}) = 0.183$ $r(q_{85}, q_{100}) = 0.027$
Optical frequency ratio n	neasurements at NIST/JILA ^a			
$r(q_{102}, q_{103}) = 0.615$	$r(q_{102}, q_{104}) = -0.207$	$r(q_{103}, q_{104}) = 0.329$		
Measurements involving	the INRIM ¹⁷¹ Yb lattice clock	k^b		
$r(q_{24}, q_{81}) = 0.088$	$r(q_{76}, q_{82}) = 0.191$	$r(q_{76}, q_{83}) = 0.175$	$r(q_{82}, q_{83}) = 0.386$	
Measurements involving	optical clocks at RIKEN ^c			
$r(q_{59}, q_{79}) = 0.826$	$r(q_{66}, q_{79}) = -0.009$			
Measurements involving	the ¹¹⁵ In ⁺ and ⁸⁷ Sr optical cl	ocks at NICT		
$r(q_3, q_{74}) = 0.026$ $r(q_{78}, q_{80}) = 0.006$	$r(q_3, q_{78}) = 0.028$ $r(q_{78}, q_{90}) = -0.021$	$r(q_{41}, q_{90}) = 0.001$ $r(q_{80}, q_{90}) = -0.026$	$r(q_{50}, q_{90}) = 0.061$	$r(q_{74}, q_{78}) = 0.859$
Measurements involving	the ¹⁷¹ Yb and ⁸⁷ Sr optical late	tice clocks at NMIJ ^d		
$ \overline{r(q_{78}, q_{94}) = 0.005} r(q_{93}, q_{94}) = 0.603 $	$r(q_{89}, q_{93}) = 0.748$ $r(q_{93}, q_{95}) = 0.680$	$r(q_{89}, q_{94}) = 0.672$ $r(q_{94}, q_{95}) = 0.611$	$r(q_{89}, q_{95}) = 0.860$	$r(q_{90}, q_{94}) = -0.033$
Other miscellaneous corr	relation coefficients			
$r(q_{31}, q_{57}) = 0.165$ $r(q_{61}, q_{106}) = 0.329$	$r(q_{53}, q_{68}) = -0.672$ $r(q_{69}, q_{106}) = 0.369$	$r(q_{81}, q_{91}) = -0.123$ $r(q_{95}, q_{106}) = 0.221$	$r(q_{56}, q_{106}) = -0.277$	$r(q_{58}, q_{106}) = -0.392$

^aThis subset of correlation coefficients was computed by D B Hume.

involving the ¹⁹⁹Hg optical lattice clock is particularly significant, with a value of 0.826, since the evaluation of systematic frequency shifts and uncertainty were determined in an identical evaluation campaign.

A.7. Measurements involving the ¹¹⁵In⁺ and ⁸⁷Sr optical clocks at NICT

The measurements q_3 , q_{41} , q_{50} , q_{74} , q_{78} , q_{80} and q_{90} [14, 51, 59, 75, 85, 92] involve the ¹¹⁵In⁺ and ⁸⁷Sr optical clocks at NICT. The largest correlation coefficient by far is between q_{74} (an absolute frequency measurement of the ¹¹⁵In⁺ clock

transition) and q_{78} (an $^{115}\mathrm{In}^+/^{87}\mathrm{Sr}$ optical frequency ratio measurement), with a value of $r(q_{74},q_{78})=0.859$. These two measurements were performed during the same campaign, with the uncertainty being dominated by that of the $^{115}\mathrm{In}^+$ optical clock (both systematics and statistics).

A.8. Measurements involving the ¹⁷¹Yb and ⁸⁷Sr optical lattice clocks at NMIJ

Measurements q_{89} , q_{93} , q_{94} and q_{95} involve the ¹⁷¹Yb and ⁸⁷Sr optical lattice clocks at NMIJ. The optical frequency ratio between the two clocks, q_{93} [94] was measured during the

^bThis subset of correlation coefficients was provided by M Pizzocaro. After approval of the 2021 recommended frequency values by the CCTF, an error in the calculation of $r(q_{24}, q_{81})$ was identified. However using the corrected value of 0.155 does not change the results at the relevant level of precision.

^eThis subset of correlation coefficients was computed by N Nemitz.

^dThis subset of correlation coefficients was computed by T Kobayashi.

last part of the campaign in which the absolute frequency q_{89} was measured [82], resulting in a significant correlation coefficient $r(q_{89},q_{93})=0.748$, arising mainly from the systematic uncertainty of the ¹⁷¹Yb lattice clock. The absolute frequency measurement was made via a comparison to TAI, during which period the NICT ⁸⁷Sr optical clock and the SYRTE ⁸⁷Rb fountain were amongst the standards used to calibrate TAI. This meant that it was possible to determine the ¹⁷¹Yb/⁸⁷Sr (q_{94}) and ¹⁷¹Yb/⁸⁷Rb ratios (q_{95}) in the same campaign. However the correlation coefficients $r(q_{89},q_{94})$, $r(q_{89},q_{95})$, $r(q_{93},q_{94})$ and $r(q_{93},q_{95})$ are also large, ranging from 0.603 to 0.860. Measurement q_{94} is correlated with q_{78} and q_{90} since NICT-Sr1 is a common standard. However these correlation coefficients are less significant due to a larger link uncertainty between Yb and Sr

A.9. Miscellaneous correlation coefficients

Absolute frequency measurements q_{31} [43] and q_{57} [65] were performed in a partly common period against the local caesium

fountain NPL-CsF1, and hence are correlated, mainly through the systematic uncertainty of the fountain. The two measurements q_{53} and q_{68} reported in [62], are correlated through the common systematic uncertainty of the NICT 40 Ca⁺ optical clock. Measurements q_{91} and q_{81} , both reported in [86], are correlated because the systematic uncertainty of the PTB transportable 87 Sr optical lattice clock is common to the two measurements. The 171 Yb/ 87 Rb ratio q_{106} was determined through measurements made against TAI, and hence is correlated with several other measurements (q_{56} , q_{58} , q_{61} , q_{69} and q_{95}) involving the LNE-SYRTE Rb fountain.

Appendix B. Frequency ratios

Table B1 lists the frequency ratio values consistent with the 2021 recommended frequency values, taking into account correlations between those values. Ratios involving ν_2 and ν_9 (1 H and 40 Ca) are excluded from the list, as these two recommended frequency values are not correlated with the others.

Table B1. Frequency ratios consistent with the 2021 recommended frequency values, taking into account the covariance of the output matrix. (ν_2 and ν_9 are linked to the other frequencies only via Cs (ν_{15}), and hence are not included in this table.).

Clock transitions	Atomic species	Frequency ratio	Fractional uncertainty
$\overline{\nu_1/\nu_3}$	¹¹⁵ In ⁺ / ¹⁹⁹ Hg	1.123 010 988 476 8743(49)	4.3×10^{-15}
ν_1/ν_4	$^{115}\text{In}^{+}/^{27}\text{Al}^{+}$	1.130 584 343 961 8487(49)	4.3×10^{-15}
ν_1/ν_5	$^{115}\text{In}^{+}/^{199}\text{Hg}^{+}$	1.190 360 410 756 6604(51)	4.3×10^{-15}
ν_1/ν_6	$^{115}\text{In}^{+}/^{171}\text{Yb}^{+}(\text{E2})$	1.841 194 044 091 2659(80)	4.3×10^{-15}
$ u_1/ u_7$	$^{115}\text{In}^{+}/^{171}\text{Yb}^{+}(\text{E3})$	1.973 773 591 557 2195(85)	4.3×10^{-15}
ν_1/ν_8	$^{115}In^{+}/^{171}Yb$	2.445 326 324 126 955(11)	4.3×10^{-15}
ν_{1}/ν_{10}	$^{115}\text{In}^{+}/^{88}\text{Sr}^{+}$	2.849 510 267 459 795(13)	4.5×10^{-15}
ν_{1}/ν_{11}	$^{115}In^{+}/^{88}Sr$	2.952748322069815(13)	4.3×10^{-15}
ν_{1}/ν_{12}	$^{115}In^{+}/^{87}Sr$	2.952748749874866(13)	4.3×10^{-15}
ν_1/ν_{13}	$^{115}\text{In}^{+}/^{40}\text{Ca}^{+}$	3.083 388 200 597 554(14)	4.7×10^{-15}
$ u_1/ u_{14}$	$^{115}\text{In}^{+}/^{87}\text{Rb}$	185 436.914 199 787 30(80)	4.3×10^{-15}
ν_3/ν_4	¹⁹⁹ Hg/ ²⁷ Al ⁺	1.006 743 794 640 198 49(15)	1.5×10^{-16}
ν_3/ν_5	¹⁹⁹ Hg/ ¹⁹⁹ Hg ⁺	1.059 972 184 574 196 57(19)	1.8×10^{-16}
ν_3/ν_6	199 Hg/ 171 Yb ⁺ (E2)	1.639 515 608 470 095 42(28)	1.7×10^{-16}
ν_3/ν_6 ν_3/ν_7	199 Hg/ 171 Yb ⁺ (E3)	1.757 572 821 468 313 31(27)	1.7×10^{-16} 1.5×10^{-16}
·	¹⁹⁹ Hg/ ¹⁷¹ Yb	2.177 473 194 134 564 88(32)	1.5×10^{-16} 1.5×10^{-16}
$ u_3/ u_8 $ $ u_3/ u_{10} $	¹⁹⁹ Hg/ ⁸⁸ Sr ⁺	2.537 384 136 663 3019(34)	1.3×10^{-15} 1.4×10^{-15}
	199 Hg / ⁸⁸ Sr	2.629 313 828 954 238 79(40)	1.4×10 1.5×10^{-16}
ν_3/ν_{11}	199 Hg / ⁸⁷ Sr		1.3×10^{-16}
ν_3/ν_{12}	199 Hg/ 40 Ca $^{+}$	2.629 314 209 898 909 56(39)	1.5×10^{-16}
ν_3/ν_{13}		2.745 643 838 071 0009(49)	1.8×10^{-15}
ν_3/ν_{14}	199 Hg/ 87 Rb	165 124.754 879 997 262(60)	3.6×10^{-16}
ν_4/ν_5	$^{27}\text{Al}^{+}/^{199}\text{Hg}^{+}$	1.052 871 833 148 990 45(11)	1.0×10^{-16}
ν_4/ν_6	$^{27}\text{Al}^+/^{171}\text{Yb}^+(\text{E2})$	1.628 533 115 573 902 39(14)	8.3×10^{-17}
$ u_4/ u_7$	$^{27}\text{Al}^{+}/^{171}\text{Yb}^{+}(\text{E3})$	1.745 799 508 102 709 104(84)	4.8×10^{-17}
$ u_4/ u_8$	$^{27}\text{Al}^{+}/^{171}\text{Yb}$	2.162 887 127 516 663 705(24)	1.1×10^{-17}
$ u_4/ u_{10} $	$^{27}\text{Al}^{+/88}\text{Sr}^{+}$	2.520 387 163 220 7488(34)	1.3×10^{-15}
$ u_4/ u_{11} $	$^{27}\text{Al}^{+}/^{88}\text{Sr}$	2.611 701 053 388 596 03(10)	3.9×10^{-17}
$ u_4/ u_{12}$	$^{27}\text{Al}^{+}/^{87}\text{Sr}$	2.611 701 431 781 463 019(39)	1.5×10^{-17}
ν_4/ν_{13}	$^{27}\text{Al}^{+}/^{40}\text{Ca}^{+}$	2.727 251 811 919 3078(48)	1.8×10^{-15}
$ u_4/ u_{14}$	$^{27}\text{Al}^{+}/^{87}\text{Rb}$	164 018.646 808 755 766(54)	3.3×10^{-16}
ν_5/ν_6	199 Hg ⁺ / 171 Yb ⁺ (E2)	1.546 753 426 486 100 05(21)	1.3×10^{-16}
$ u_5/ u_7$	199 Hg ⁺ / 171 Yb ⁺ (E3)	1.658 131 078 387 072 22(19)	1.1×10^{-16}
ν_5/ν_8	199 Hg ⁺ / 171 Yb	2.054 273 900 601 723 59(22)	1.0×10^{-16}
ν_{5}/ν_{10}	$^{199} Hg^{+} / ^{88} Sr^{+}$	2.393 821 435 684 7480(32)	1.3×10^{-15}
ν_5/ν_{11}	199 Hg $^{+}$ / 88 Sr	2.480 549 836 324 681 89(28)	1.1×10^{-16}
ν_5/ν_{12}	199 Hg $^{+}$ / 87 Sr	2.480 550 195 715 877 54(26)	1.1×10^{-16}
ν_5/ν_{13}	199 Hg $^{+}$ / 40 Ca $^{+}$	2.590 298 007 842 4970(46)	1.8×10^{-15}
ν_5/ν_{14}	199 Hg $^{+}$ / 87 Rb	155 782.158 516 102 797(54)	3.5×10^{-16}
ν_6/ν_7	$^{171}\text{Yb}^{+}(\text{E2})/^{171}\text{Yb}^{+}(\text{E3})$	1.072 007 373 634 205 473(73)	6.9×10^{-17}
$ u_6/ u_8$	$^{171}\text{Yb}^{+}(\text{E2})/^{171}\text{Yb}$	1.328 119 831 787 671 42(11)	8.3×10^{-17}
$ u_6/ u_{10} $	$^{171}\text{Yb}^{+}(\text{E2})/^{88}\text{Sr}^{+}$	1.547 642 561 958 3136(21)	1.3×10^{-15}
ν_{6}/ν_{11}	$^{171}\text{Yb}^{+}(\text{E2})/^{88}\text{Sr}$	1.603 713 813 623 139 52(14)	8.9×10^{-17}
ν_6/ν_{12}	$^{171}\text{Yb}^{+}(\text{E2})/^{87}\text{Sr}$	1.603 714 045 975 103 00(13)	8.2×10^{-17}
ν_6/ν_{13}	$^{171}\text{Yb}^{+}(\text{E2})/^{40}\text{Ca}^{+}$	1.674 667 703 001 0606(30)	1.8×10^{-15}
ν_6/ν_{14}	$^{171}\text{Yb}^{+}(\text{E2})/^{87}\text{Rb}$	100715.573567538329(34)	3.4×10^{-16}
ν_7/ν_8	$^{171}\text{Yb}^{+}(\text{E3})/^{171}\text{Yb}$	1.238 909 231 832 259 428(59)	4.7×10^{-17}
ν_7/ν_8	¹⁷¹ Yb ⁺ (E3)/ ⁸⁸ Sr ⁺	1.443 686 489 498 3514(19)	1.3×10^{-15}
$\nu_7/\nu_{10} = \nu_7/\nu_{11}$	$^{171}\text{Yb}^{+}(\text{E3})/^{88}\text{Sr}$	1.495 991 401 800 156 824(86)	5.8×10^{-17}
ν_7/ν_{12}	¹⁷¹ Yb ⁺ (E3)/ ⁸⁷ Sr	1.495 991 618 544 900 552(68)	4.6×10^{-17}
$ \frac{\nu_{1}/\nu_{12}}{\nu_{7}/\nu_{13}}$	$^{171}\text{Yb}^{+}(\text{E3})/^{40}\text{Ca}^{+}$	1.562 179 276 177 7189(28)	1.8×10^{-15}
	171 Yb ⁺ (E3)/ 87 Rb	93 950.448 518 001 415(31)	3.3×10^{-16}
ν_7/ν_{14}	171 Yb/ ⁸⁸ Sr ⁺	1.165 288 345 913 1553(16)	1.3×10^{-15}
$ u_8/\nu_{10} $	¹⁷¹ Yb/ ⁸⁸ Sr		3.8×10^{-17}
ν_8/ν_{11}	¹⁷¹ Yb/ ⁸⁷ Sr	1.207 506 864 395 296 327(46)	5.8 × 10 1.2 × 10-17
ν_8/ν_{12}	YD/ Sr	1.207 507 039 343 337 845(16)	1.3×10^{-17}

(Continued.)

Table B1. (Continued.

Clock transitions	Atomic species	Frequency ratio	Fractional uncertainty
ν_{8}/ν_{13}	¹⁷¹ Yb/ ⁴⁰ Ca ⁺	1.260 931 177 231 8993(22)	1.8×10^{-15}
ν_{8}/ν_{14}	¹⁷¹ Yb/ ⁸⁷ Rb	75 833.197 545 114 200(25)	3.3×10^{-16}
ν_{10}/ν_{11}	$^{88}Sr^{+}/^{88}Sr$	1.036 230 104 446 0007(14)	1.3×10^{-15}
ν_{10}/ν_{12}	$^{88}Sr^{+}/^{87}Sr$	1.036 230 254 578 8345(14)	1.3×10^{-15}
ν_{10}/ν_{13}	$^{88}\text{Sr}^{+}/^{40}\text{Ca}^{+}$	1.082 076 536 381 8990(24)	2.2×10^{-15}
$ u_{10}/ u_{14} $	$^{88}\text{Sr}^{+}/^{87}\text{Rb}$	65 076.766 459 625 929(88)	1.4×10^{-15}
ν_{11}/ν_{12}	⁸⁸ Sr/ ⁸⁷ Sr	1.000 000 144 883 682 799(36)	3.6×10^{-17}
ν_{11}/ν_{13}	88 Sr/ 40 Ca $^{+}$	1.044 243 485 823 4592(18)	1.8×10^{-15}
$ u_{11}/ u_{14} $	⁸⁸ Sr/ ⁸⁷ Rb	62 801.462 899 418 361(21)	3.3×10^{-16}
ν_{12}/ν_{13}	87 Sr/ 40 Ca $^{+}$	1.044 243 334 529 6392(18)	1.8×10^{-15}
ν_{12}/ν_{14}	87 Sr / 87 Rb	62 801.453 800 512 449(21)	3.3×10^{-16}
ν_{13}/ν_{14}	$^{40}\text{Ca}^{+}/^{87}\text{Rb}$	60 140.631 712 818 40(11)	1.8×10^{-15}

ORCID iDs

H S Margolis https://orcid.org/0000-0002-8991-3855 G Petit https://orcid.org/0000-0002-6837-3052

C Oates https://orcid.org/0000-0001-7842-8259

T Ido https://orcid.org/0000-0003-3500-6042

S Bize https://orcid.org/0000-0003-2483-5152

References

- [1] Riehle F, Gill P, Arias F and Robertsson L 2018 The CIPM list of recommended frequency standard values: guidelines and procedures *Metrologia* 55 188–200
- [2] Terms of reference for the CCL-CCTF Working Group on Frequency Standards (WGFS) (available at: www.bipm. org/en/committees/cc/cctf/wg/ccl-cctf-wgfs)
- [3] Consultative Committee for Time and Frequency 2021 Recommendation CCTF PSFS 2 (2021): updates to the CIPM list of standard frequencies (available at: www. bipm.org/en/committees/cc/cctf/22-_2-2021)
- [4] BIPM 2021 Recommended values of standard frequencies (available at: www.bipm.org/en/publications/mises-en-pratique/standard-frequencies.html)
- [5] Petit G, Arias F and Panfilo G 2015 International atomic time: status and future challenges C. R. Physique 16 480–8
- [6] Panfilo G and Arias F 2019 The Coordinated universal Time (UTC) Metrologia 56 042001
- [7] Primary and Secondary Frequency Standards participating to TAI (available at: https://webtai.bipm.org/ftp/pub/tai/data/PSFS_reports/)
- [8] Margolis HS and Gill P 2015 Least-squares analysis of clock frequency comparison data to deduce optimized frequency and frequency ratio values *Metrologia* 52 628–34
- [9] Robertsson L 2016 On the evaluation of ultra-high-precision frequency ratio measurements: examining closed loops in a graph theory framework *Metrologia* 53 1272–80
- [10] Margolis H S and Gill P 2016 Determination of optimized frequency and frequency ratio values from over-determined sets of clock comparison data *J. Phys.:* Conf. Ser. 723 012060
- [11] Mohr P J and Taylor B N 2000 CODATA recommended values of the fundamental physical constants: 1998 Rev. Mod. Phys. 72 351–495
- [12] von Zanthier J et al 2000 Absolute frequency measurement of the In⁺ clock transition with a mode-locked laser Opt. Lett. 25 1729–31

- [13] Wang Y H, Dumke R, Liu T, Stejskal A, Zhao Y N, Zhang J, Lu Z H, Wang L J, Becker T and Walther H 2007 Absolute frequency measurement and high resolution spectroscopy of $^{115} {\rm In}^+~5 {\rm s}^2~^1 {\rm S}_0\text{-}5 {\rm s}5 {\rm p}~^3 {\rm P}_0$ narrowline transition Opt. Commun.~273~526--31
- [14] Ohtsubo N, Li Y, Matsubara K, Ido T and Hayasaka K 2017 Frequency measurement of the clock transition of an indium ion sympathetically-cooled in a linear trap *Opt. Express* 25 11725–35
- [15] Parthey C G et al 2011 Improved measurement of the hydrogen 1S–2S transition frequency Phys. Rev. Lett. 107 203001
- [16] Matveev A et al 2013 Precision measurement of the hydrogen 1S–2S frequency via a 920 km fiber link Phys. Rev. Lett. 110 230801
- [17] McFerran J J, Yi L, Mejri S, Di Manno S, Zhang W, Guéna J, Le Coq Y and Bize S 2012 Neutral atom frequency reference in the deep ultraviolet with fractional uncertainty = 5.7×10^{-15} Phys. Rev. Lett. 108 183004
- [18] McFerran J J, Yi L, Mejri S, Di Manno S, Zhang W, Guéna J, Le Coq Y and Bize S 2015 Erratum: Neutral atom frequency reference in the deep ultraviolet with fractional uncertainty = 5.7 × 10⁻¹⁵ [Phys. Rev. Lett. 108, 183004 (2012)] Phys. Rev. Lett. 115 219901
- [19] Tyumenev R et al 2016 Comparing a mercury optical lattice clock with microwave and optical frequency standards New J. Phys. 18 113002
- [20] Rosenband T et al 2007 Observation of the ${}^{1}S_{0} \rightarrow {}^{3}P_{0}$ clock transition in ${}^{27}Al^{+}$ Phys. Rev. Lett. 98 220801
- [21] Stalnaker J E *et al* 2007 Optical-to-microwave frequency comparison with fractional uncertainty of 10^{-15} *Appl. Phys.* B **89** 167–76
- [22] Consultative Committee for Length 2005 Report of the 12th Meeting, Consultative Committee for Length (CCL) (15–16 September 2005) (available at: www.bipm.org/documents/20126/31697125/12th+meeting/058b3c8a-8914-18c1-3fe1-2e601dcbdfad)
- [23] Tamm C, Weyers S, Lipphardt B and Peik E 2009 Stray-field-induced quadrupole shift and absolute frequency of the 688-THz ¹⁷¹Yb⁺ single-ion optical frequency standard *Phys. Rev.* A 80 043403
- [24] Webster S, Godun R, King S, Huang G, Walton B, Tsatourian V, Margolis H, Lea S and Gill P 2010 Frequency measurement of the ²S_{1/2}-²D_{3/2} electric quadrupole transition in a single ¹⁷¹Yb⁺ ion *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* 57 592-9
- [25] Tamm C, Huntemann N, Lipphardt B, Gerginov V, Nemitz N, Kazda M, Weyers S and Peik E 2014 Cs-based optical

frequency measurement using cross-linked optical and microwave oscillators *Phys. Rev.* A **89** 023820

- [26] Godun R M, Nisbet-Jones P B R, Jones J M, King S A, Johnson L A M, Margolis H S, Szymaniec K, Lea S N, Bongs K and Gill P 2014 Frequency ratio of two optical clock transitions in ¹⁷¹Yb⁺ and constraints on the time variation of fundamental constants *Phys. Rev. Lett.* 113 210801
- [27] Hosaka K, Webster S A, Stannard A, Walton B R, Margolis H S and Gill P 2009 Frequency measurement of the ²S_{1/2}–²F_{7/2} electric octupole transition in a single ¹⁷¹Yb⁺ ion *Phys. Rev.* A **79** 033403
- [28] Huntemann N, Okhapkin M, Lipphardt B, Weyers S, Tamm C and Peik E 2012 High-accuracy optical clock based on the octupole transition in ¹⁷¹Yb⁺ Phys. Rev. Lett. 108 090801
- [29] King SA, Godun R M, Webster SA, Margolis H S, Johnson L A M, Szymaniec K, Baird P E G and Gill P 2012 Absolute frequency measurement of the ²S_{1/2}-²F_{7/2} electric octupole transition in a single ion of ¹⁷¹Yb⁺ with 10⁻¹⁵ fractional uncertainty *New J. Phys.* 14 013045
- [30] Huntemann N, Lipphardt B, Tamm C, Gerginov V, Weyers S and Peik E 2014 Improved limit on a temporal variation of m_p/m_e from comparisons of Yb⁺ and Cs atomic clocks Phys. Rev. Lett. 113 210802
- [31] Kohno T, Yasuda M, Hosaka K, Inaba H, Nakajima Y and Hong F-L 2009 One-dimensional optical lattice clock with a fermionic ¹⁷¹Yb isotope *Appl. Phys. Express* **2** 072501
- [32] Yasuda M *et al* 2012 Improved absolute frequency measurement of the ¹⁷¹Yb optical lattice clock towards a candidate for the redefinition of the second *Appl. Phys. Express* **5** 102401
- [33] Lemke N D, Ludlow A D, Barber Z W, Fortier T M, Diddams S A, Jiang Y, Jefferts S R, Heavner T P, Parker T E and Oates C W 2009 Spin-1/2 optical lattice clock *Phys. Rev. Lett.* 103 063001
- [34] Park C Y *et al* 2013 Absolute frequency measurement of ${}^{1}S_{0}$ (F=1/2) ${}^{-3}P_{0}$ (F=1/2) transition of 171 Yb atoms in a one-dimensional optical lattice at KRISS *Metrologia* 50 119–28
- [35] Pizzocaro M, Thoumany P, Rauf B, Bregolin F, Milani G, Clivati C, Costanzo G A, Levi F and Calonico D 2017 Absolute frequency measurement of the ¹S₀-³P₀ transition of ¹⁷¹Yb Metrologia 54 102
- [36] Kim H, Heo M-S, Lee W-K, Park C Y, Hong H-G, Hwang S-W and Yu D-H 2017 Improved absolute frequency measurement of the ¹⁷¹Yb optical lattice clock at KRISS relative to the SI second *Jpn. J. Appl. Phys.* 56 050302
- [37] Degenhardt C et al 2005 Calcium optical frequency standard with ultracold atoms: Approaching 10⁻¹⁵ relative uncertainty Phys. Rev. A 72 062111
- [38] Wilpers G, Oates C W and Hollberg L 2006 Improved uncertainty budget for optical frequency measurements with microkelvin neutral atoms: results for a high-stability ⁴⁰Ca optical frequency standard *Appl. Phys.* B 85 31–44
- [39] Wilpers G et al 2007 Absolute frequency measurement of the neutral ⁴⁰Ca optical frequency standard at 657 nm based on microkelvin atoms Metrologia 44 146–51
- [40] Margolis H S, Barwood G P, Huang G, Klein H A, Lea S N, Szymaniec K and Gill P 2004 Hertz-level measurement of the optical clock frequency in a single ⁸⁸Sr⁺ ion *Science* **306** 1355–8
- [41] Dubé P, Madej A A, Bernard J E, Marmet L, Boulanger J-S and Cundy S 2005 Electric quadrupole shift cancellation in single-ion optical frequency standards *Phys. Rev. Lett.* 95 033001

- [42] Madej A A, Dubé P, Zhou Z, Bernard J E and Gertsvolf M 2012 ⁸⁸Sr⁺ 445-THz single-ion reference at the 10⁻¹⁷ level via control and cancellation of systematic uncertainties and its measurement against the SI second *Phys. Rev. Lett.* 109 203002
- [43] Barwood G P, Huang G, Klein H A, Johnson L A M, King S A, Margolis H S, Szymaniec K and Gill P 2014 Agreement between two ⁸⁸Sr⁺ optical clocks to 4 parts in 10¹⁷ Phys. Rev. A 89 050501
- [44] Dubé P, Bernard J E and Gertsvolf M 2017 Absolute frequency measurement of the ⁸⁸Sr⁺ clock transition using a GPS link to the SI second *Metrologia* **54** 290–8
- [45] Baillard X, Fouché M, Le Targat R, Westergaard P G, Lecallier A, Le Coq Y, Rovera G D, Bize S and Lemonde P 2007 Accuracy evaluation of an optical lattice clock with bosonic atoms *Opt. Lett.* 32 1812–4
- [46] Morzyński P et al 2015 Absolute measurement of the ¹S₀-³P₀ clock transition in neutral ⁸⁸Sr over the 330 km-long stabilized fibre optic link Sci. Rep. 5 17495
- [47] Boyd M M, Ludlow A D, Blatt S, Foreman S M, Ido T, Zelevinsky T and Ye J 2007 ⁸⁷ Sr lattice clock with inaccuracy below 10⁻¹⁵ *Phys. Rev. Lett.* **98** 083002
- [48] Campbell G K et al 2008 The absolute frequency of the ⁸⁷Sr optical clock transition Metrologia 45 539–48
- [49] Baillard X *et al* 2008 An optical lattice clock with spin-polarized ⁸⁷Sr atoms *Eur. Phys. J.* D **48** 11–17
- [50] Falke S *et al* 2011 The ⁸⁷ Sr optical frequency standard at PTB *Metrologia* **48** 399
- [51] Yamaguchi A, Shiga N, Nagano S, Li Y, Ishijima H Hachisu H, Kumagai M and Ido T 2012 Stability transfer between two clock lasers operating at different wavelengths for absolute frequency measurement of clock transition in ⁸⁷Sr Appl. Phys. Express 5 022701
- [52] Akamatsu D, Inaba H, Hosaka K, Yasuda M, Onae A, Suzuyama T, Amemiya M and Hong F-L 2014 Spectroscopy and frequency measurement of the ⁸⁷Sr clock transition by laser linewidth transfer using an optical frequency comb *Appl. Phys. Express* 7 012401
- [53] Tanabe T *et al* 2015 Improved frequency measurement of the ${}^{1}S_{0}-{}^{3}P_{0}$ clock transition in ${}^{87}Sr$ using a Cs fountain clock as a transfer oscillator *J. Phys. Soc. Japan* **84** 115002
- [54] Lin Y-G, Wang Q, Li Y, Meng F, Lin B-K, Zang E-J, Sun Z, Fang F, Li T-C and Fang Z-J 2015 First evaluation and frequency measurement of the strontium optical lattice clock at NIM *Chin. Phys. Lett.* 32 090601
- [55] Falke S *et al* 2014 A strontium lattice clock with 3×10^{-17} inaccuracy and its frequency *New J. Phys.* **16** 073023
- [56] Le Targat R et al 2013 Experimental realization of an optical second with strontium lattice clocks Nat. Commun. 4 2109
- [57] Lodewyck J et al 2016 Optical to microwave clock frequency ratios with a nearly continuous strontium optical lattice clock Metrologia 53 1123
- [58] Grebing C, Al-Masoudi A, Dörscher S, Häfner S, Gerginov V, Weyers S, Lipphardt B, Riehle F, Sterr U and Lisdat C 2016 Realization of a timescale with an accurate optical lattice clock *Optica* 3 563–9
- [59] Hachisu H, Petit G and Ido T 2017 Absolute frequency measurement with uncertainty below 1×10^{-15} using international atomic time *Appl. Phys.* B **123** 34
- [60] Hachisu H, Petit G, Nakagawa F, Hanado Y and Ido T 2017 SI-traceable measurement of an optical frequency at the low 10⁻¹⁶ level without a local primary standard *Opt*. *Express* 25 8511–23
- [61] Chwalla M et al 2009 Absolute frequency measurement of the ⁴⁰Ca⁺ 4s ²S_{1/2}–3d ²D_{5/2} clock transition Phys. Rev. Lett. 102 023002

- [62] Matsubara K, Hachisu H, Li Y, Nagano S, Locke C, Nogami A, Kajita M, Hayasaka K, Ido T and Hosokawa M 2012 Direct comparison of a Ca⁺ single-ion clock against a Sr lattice clock to verify the absolute frequency measurement Opt. Express 20 22034–41
- [63] Huang Y, Guan H, Liu P, Bian W, Ma L, Liang K, Li T and Gao K 2016 Frequency comparison of two ⁴⁰Ca⁺ optical clocks with an uncertainty at the 10⁻¹⁷ level *Phys. Rev. Lett.* **116** 013001
- [64] LNE-SYRTE TAI data; MJD 55954-57867 (January 2012–April 2017) Request of the CCL-CCTF WGFS by Guéna J on 10 May 2017 (submitted)
- [65] Ovchinnikov Y B, Szymaniec K and Edris S 2015 Measurement of rubidium ground-state hyperfine transition frequency using atomic fountains *Metrologia* 52 595–9
- [66] Guéna J et al 2017 First international comparison of fountain primary frequency standards via a long distance optical fiber link Metrologia 54 348–54
- [67] Wang Y H, Liu T, Dumke R, Stejskal A, Zhao Y N, Zhang J, Lu Z H, Wang L J, Becker T and Walther H 2007 Improved absolute frequency measurement of the ¹¹⁵In⁺ 5s^{2 1}S₀-5s5p ³P₀ narrowline transition: progress towards an optical frequency standard *Laser Phys*. 17 1017-24
- [68] Yamanaka K, Ohmae N, Ushijima I, Takamoto M and Katori H 2015 Frequency ratio of ¹⁹⁹Hg and ⁸⁷Sr optical lattice clocks beyond the SI limit *Phys. Rev. Lett.* 114 230801
- [69] Rosenband T et al 2008 Frequency ratio of Al⁺ and Hg⁺ single-ion optical clocks; metrology at the 17th decimal place Science 319 1808–12
- [70] Akamatsu D, Yasuda M, Inaba H, Hosaka K, Tanabe T, Onae A and Hong F-L 2014 Frequency ratio measurement of ¹⁷¹Yb and ⁸⁷Sr optical lattice clocks *Opt. Express* 22 7898–905
- [71] Akamatsu D, Yasuda M, Inaba H, Hosaka K, Tanabe T, Onae A and Hong F-L 2014 Errata: Frequency ratio measurement of ¹⁷¹Yb and ⁸⁷Sr optical lattice clocks *Opt. Express* 22 32199
- [72] Takamoto M *et al* 2015 Frequency ratios of Sr, Yb and Hg based optical lattice clocks and their applications *C. R. Physique* **16** 489–98
- [73] Nemitz N, Ohkubo T, Takamoto M, Ushijima I, Das M, Ohmae N and Katori H 2016 Frequency ratio of Yb and Sr clocks with 5×10^{-17} uncertainty at 150 seconds averaging time *Nat. Photon.* **10** 258–61
- [74] Takano T, Mizushima R and Katori H 2017 Precise determination of the isotope shift of ⁸⁸Sr-⁸⁷Sr optical lattice clock by sharing perturbations Appl. Phys. Express 10 072801
- [75] Ohtsubo N, Li Y, Nemitz N, Hachisu H, Matsubara K, Ido T and Hayasaka K 2020 Frequency ratio of an ¹¹⁵In⁺ ion clock and a ⁸⁷Sr optical lattice clock *Opt. Lett.* 45 5950–3
- [76] Leopardi H *et al* 2021 Measurement of the ²⁷Al⁺ and ⁸⁷Sr absolute optical frequencies *Metrologia* **58** 015017
- [77] Baynham C \dot{F} A *et al* 2018 Absolute frequency measurement of the $^2S_{1/2}$ – $^2F_{7/2}$ optical clock transition in 171 Yb⁺ with an uncertainty of 4×10^{-16} using a frequency link to International Atomic Time *J. Mod. Opt.* **65** 585–91
- [78] Lange R, Huntemann N, Rahm J M, Sanner C, Shao H, Lipphardt B, Tamm C, Weyers S and Peik E 2021 Improved limits for violations of local position invariance from atomic clock comparisons *Phys. Rev. Lett.* 126 011102

- [79] Luo L et al 2020 Absolute frequency measurement of an Yb optical clock at the 10⁻¹⁶ level using International Atomic Time Metrologia 57 065017
- [80] McGrew W F et al 2019 Towards the optical second: verifying optical clocks at the SI limit Optica 6 448–54
- [81] Pizzocaro M, Bregolin F, Barbieri P, Rauf B, Levi F and Calonico D 2020 Absolute frequency measurement of the ¹S₀-³P₀ transition of ¹⁷¹Yb with a link to international atomic time *Metrologia* 57 035007
- [82] Kobayashi T, Akamatsu D, Hosaka K, Hisai Y, Wada M, Inaba H, Suzuyama T, Hong F-L and Yasuda M 2020 Demonstration of the nearly continuous operation of an ¹⁷¹Yb optical lattice clock for half a year *Metrologia* 57 065021
- [83] Hobson R, Bowden W, Vianello A Silva A, Baynham C F A, Margolis H S, Baird P E G, Gill P and Hill I R 2020 A strontium optical lattice clock with 1×10^{-17} uncertainty and measurement of its absolute frequency *Metrologia* 57 065026
- [84] Schwarz R, Dörscher S, Al-Masoudi A, Benkler E, Legero T, Sterr U, Weyers S, Rahm J, Lipphardt B and Lisdat C 2020 Long term measurement of the ⁸⁷Sr clock frequency at the limit of primary Cs clocks *Phys. Rev. Res.* 2 033242
- [85] Nemitz N, Gotoh T, Nakagawa F, Ito H, Hanado Y, Ido T and Hachisu H 2021 Absolute frequency of 87 Sr at 1.8×10^{-16} uncertainty by reference to remote primary frequency standards *Metrologia* **58** 025006
- [86] Grotti J *et al* 2018 Geodesy and metrology with a transportable optical clock *Nat. Phys.* **14** 437–41
- [87] Huang Y et al 2020 Geopotential measurement with a robust, transportable Ca⁺ optical clock Phys. Rev. A 102 050802
- [88] Ohmae N, Bregolin F, Nemitz N and Katori H 2020 Direct measurement of the frequency ratio for Hg and Yb optical lattice clocks and closure of the Hg/Yb/Sr loop Opt. Express 28 15112–21
- [89] Boulder Atomic Clock Optical Network (BACON) Collaboration 2021 Frequency ratio measurements at 18-digit accuracy using an optical clock network *Nature* 591 564–9
- [90] Riedel F et al 2020 Direct comparisons of European primary and secondary frequency standards via satellite techniques Metrologia 57 045005
- [91] Dörscher S, Huntemann N, Schwarz R, Lange R, Benkler E, Lipphardt B, Sterr U, Peik E and Lisdat C 2021 Optical frequency ratio of a ¹⁷¹Yb⁺ single-ion clock and a ⁸⁷Sr lattice clock *Metrologia* 58 015005
- [92] Fujieda M *et al* 2018 Advanced satellite-based frequency transfer at the 10⁻¹⁶ level *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **65** 973–8
- [93] Pizzocaro M *et al* 2021 Intercontinental comparison of optical atomic clocks through very long baseline interferometry *Nat. Phys.* 17 223–7
- [94] Hisai Y, Akamatsu D, Kobayashi T, Hosaka K, Inaba H, Hong F-L and Yasuda M 2021 Improved frequency ratio measurement with ⁸⁷Sr and ¹⁷¹Yb optical lattice clocks at NMIJ Metrologia 58 015008
- [95] Origlia S et al 2018 Towards an optical clock for space: compact, high-performance optical lattice clock based on bosonic atoms Phys. Rev. A 98 053443
- [96] Margolis H and Pizzocaro M 2020 Guidelines on the evaluation and reporting of correlation coefficients between frequency ratio measurements (available at: https://doi.org/10.59161/StdFreq202301)
- [97] BIPM Fractional frequency of EAL (available at: https://webtai.bipm.org/ftp/pub/tai/other-products/etoile/)

- [98] Margolis H S, Panfilo G, Petit G, Oates C, Ido T and Bize S 2023 Data associated with the 2021 adjustment of standard frequencies (available at: https://doi.org/10. 59161/StdFreq202301)
- [99] PSFS reports (available at: https://webtai.bipm.org/ftp/pub/tai/data/PSFS_reports/)
- [100] BIPM, IEC, IFCC, ILAC, ISO, IUPAC, IUPAP and OIML 2008 Evaluation of measurement data—guide to the expression of uncertainty in measurement *Joint Committee* for Guides in Metrology, JCGM vol 100 (available at: www.bipm.org/en/publications/guides/gum.html)
- [101] Consultative Committee for Time and Frequency 2021
 Recommendation CCTF PSFS 1 (2020):
 Recommendations for operating, comparing and reporting
 frequency standards to improve TAI and to prepare for a
 redefinition of the second by optical transitions (available
 at: www.bipm.org/en/committees/cc/cctf/22-_2-2021)
- [102] Dimarcq N et al 2024 Roadmap towards the redefinition of the second Metrologia 61 012001
- [103] Lodewyck J 2019 On a definition of the SI second with a set of optical clock transitions *Metrologia* **56** 055009