

Probe motion during mid-trimester fetal anomaly scan in the clinical setting: A prospective observational study

Maela Le Lous, Francisco Vasconcelos Junior, Chiara Di Vece, Brian Dromey, Raffaele Napolitano, Soojoeong Yoo, Eddie Edwards, Arnaud Huaulmé, Donald Peebles, Danail Stoyanov, et al.

▶ To cite this version:

Maela Le Lous, Francisco Vasconcelos Junior, Chiara Di Vece, Brian Dromey, Raffaele Napolitano, et al.. Probe motion during mid-trimester fetal anomaly scan in the clinical setting: A prospective observational study. European Journal of Obstetrics & Gynecology and Reproductive Biology, 2024, 298, pp.13-17. 10.1016/j.ejogrb.2024.042. hal-04574535

HAL Id: hal-04574535 https://hal.science/hal-04574535v1

Submitted on 2 Jul 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

1 Probe motion during mid-trimester fetal anomaly scan in the clinical setting: a 2 prospective observational study 3 Maela LE LOUS^{1,2,3,4}, Francisco VASCONCELOS⁴, Chiara DI VECE⁴, Brian DROMEY^{4, 5}, 4 5 Raffaele NAPOLITANO^{5,6}, Soojoeong YOO⁴, Eddie EDWARDS⁴, Arnaud HUAULME², Donald PEEBLES^{5,6}, Danail STOYANOV⁴, Pierre JANNIN². 6 7 8 9 10 11 12 13 14 15 16 Department of Obstetrics and Gynecology, University Hospital of Rennes, France. 1. Univ Rennes, INSERM, LTSI - UMR 1099, F35000, Rennes, France. 2. 3. CIC Inserm 1414, University Hospital of Rennes, University of Rennes 1, Rennes, France. 4. Department of Computer Science, Wellcome/EPSRC Centre for Interventional and Surgical Sciences (WEISS), University College London, London, United Kingdom 5. Elizabeth Garrett Anderson Institute for Women's Health, University College London, London, UK Fetal Medicine Unit, Elizabeth Garrett Anderson Wing, University College London Hospitals NHS Foundation Trust, 6. London, UK 17 18 **Corresponding author:** 19 Dr Maela Le Lous 20 Univ Rennes, INSERM, LTSI - UMR 1099, F35000, Rennes, France. 21 Phone: Mobile +33 6 95 02 38 05 22 E-mail: maela.le.lous@chu-rennes.fr 23 24 Conflict of interest: The authors report no conflict of interest. 25 26 **Funding:** 27 University Hospital of Rennes, France. 28 Institut de la Mère et de L'enfant, Rennes. 29 Wellcome/EPSRC Centre for Interventional and Surgical Sciences (WEISS) 30 31

32 Introduction

- 33 This study aims to investigate probe motion during full mid-trimester anomaly scans.
- 34

35 Methods

We undertook a prospective, observational study of obstetric sonographers at a UK University Teaching Hospital. We collected prospectively full-length video recordings of routine secondtrimester anomaly scans synchronized with probe trajectory tracking data during the scan. Videos were reviewed and trajectories analyzed using duration, path metrics (path length, velocity, acceleration, jerk, and volume) and angular metrics (spectral arc, angular area, angular velocity, angular acceleration, and angular jerk). These trajectories were then compared according to the participant level of expertise, fetal presentation, and patient BMI.

43

44 **Results**

A total of 17 anomaly scans were recorded. The average velocity of the probe was 12.9 ± 3.4 mm/s for the consultants versus 24.6 ± 5.7 mm/s for the fellows (p=0.02), the average acceleration 170.4 ± 26.3 mm/s² versus 328.9 ± 62.7 mm/s² (p=0.02), and the average jerk 7491.7 ± 1056.1 mm/s³ versus 14944.1 ± 3146.3 mm/s³ (p=0.02), the working volume 9.10^{6} $\pm4.10^{6}$ mm³ versus $29.10^{6}\pm11.10^{6}$ mm³ (p=0.03), respectively. The angular metrics were not significantly different according to the participant level of expertise, the fetal presentation, or to patients BMI.

52

53 Conclusion

Some differences in the probe path metrics (velocity, acceleration, jerk and working volume)
were noticed according to operator's level.

- 56
- 57 Keywords: Fetal ultrasound; anomaly scans; probe trajectory; trajectory; metrics; obstetrics;
 58 pregnancy; screening.
- 59

- 60 Introduction
- 61

62 Ultrasound is a dynamic, real-time imaging modality widely used for pregnancy screening, but it can be challenging depending on operator's level, fetal presentation, and 63 64 patient body mass index (BMI). In accordance with the latest International Society of 65 Ultrasound in Obstetrics and Gynecology (ISUOG) guidelines, routine mid-trimester fetal 66 ultrasound examination includes an evaluation of the following: cardiac activity, number of 67 fetuses, fetal size, basic fetal anatomy including head, face, heart, abdomen, limbs (skeletal), 68 umbilical cord, genitalia, placental appearance and location, and amniotic fluid volume [1]. 69 Although experienced sonographers usually follow a systematic approach, there is no universal 70 scanning protocol at this time [2]. Hence, each sonographer may conduct the examination 71 differently. In a previous study by Drukker et al., each anomaly scan was found to be unique in 72 its scanning sequence with a wide variation during routine second-trimester anomaly scans [3].

However, in some studies conducted in simulated settings, similarities between the trajectories of experts were found, suggesting a relationship between expertise and trajectory [4,5]. For example, differences in trajectory metrics such as path length, velocity acceleration, or dimensionless squared jerk could discriminate expertise. These studies demonstrated a link between sonographers' hand motion and their expertise during simulation sessions.

78 Repeating this experiment in a clinical environment allows us to determine if the skills 79 assessed in a simulated environment can be translated to clinical reality and how robust these 80 skills are when presented with the variability inherent to obstetric scanning. Specifically, the 81 degree to which fetal presentation and maternal body type impact the way the sonographers 82 move the probe remains unknown. To our knowledge, no study quantifying probe motion 83 during real-life ultrasound acquisition according to those specific conditions has been described 84 so far. We hypothesized that trajectory metrics might differ according to operator's level, fetal 85 presentation (e.g., cephalic or breech) or patient body mass index (BMI).

86 Therefore, this study aimed to quantitatively describe the probe trajectory during a
87 routine mid-trimester ultrasound scan according to operator's level, fetal position and patients'
88 BMI.

89

90 Methods

91 Setting

We undertook a prospective, observational study of sonographers at the Ultrasound
Screening Unit of University College London Hospital NHS Foundation Trust, London, UK
(UCLH).

95 Women attending the Ultrasound Screening Unit for a routine second-trimester anomaly scan on a Tuesday afternoon between February 1st and June 30th 2023 were offered the 96 97 opportunity to participate in the study. Each sonographer, a senior ultrasound fellow working 98 in that unit or a fetal medicine consultant was asked to conduct a routine second-trimester fetal 99 ultrasound scan using a VolusonE8® ultrasound machine (GE Healthcare, Chicago, Illinois). 100 Operators were allowed unlimited time to complete the scan, following the UK FASP 101 guidelines [6] and international guidelines, which requires the scan of the following structures: 102 head/brain, coronal face view (nose/lips), sagittal face (fetal profile), thorax, heart, abdomen, 103 spine, kidneys, bladder, femur, limbs, placenta, amniotic fluid, and Doppler imaging of the 104 maternal uterine artery [1]. The study was approved by the NHS Research Ethics Committee. 105 Informed consent was obtained from each participant and patient.

106

107 Acquisitions setting

108 We prospectively collected full-length video recordings of routine second-trimester 109 anomaly scans using the device's high-definition multimedia interface (HDMI) output; we 110 recorded them using an AV.io HD® graphic board (Epiphan videos, USA) (Figure 1a). The probe's position was tracked using the Aurora® electromagnetic tracking system (NDI Inc., 111 112 Waterloo, Ontario, Canada), whose field generator was placed under the patient. To track the 113 probe a 6 degrees-of-freedom (DOF) sensor was attached to the ultrasound probe by means of 114 a custom 3-D printed holder (Figure 1b). The probe tracking frequency was 40 Hz. The data 115 was synchronized, recorded, and gathered on an ODX® laptop using an in-house application 116 developed in LabVIEW, National Instruments (TM) (Figure 1c).

117

118 *Motion data analyzing*

119Trajectory analysis was performed on a dedicated computer using Intel® Xeon(R) W-1202245 CPU @ 3.90GHz with 64Go RAM using Matlab_R2018a. For each trajectory, the121following metrics were calculated:

- 122 123
- ACCEPTED MANUSCRIPT / CLEAN COPY 4

124	
125	• Duration (D) corresponds to the execution time of the second trimester
126	scans from the starting image until the end of the scan. It is measured in
127	minutes (min) using the video.
128	• Path metrics:
129	1. Path Length (PL) represents the total distance traveled by the probe during the
130	execution of the task (mm).
131 132	2. Average Velocity (AV) corresponds to the average linear speed of the probe. mm/s.
133	3. Average Acceleration (AA) : acceleration of the probe during the task (mm/s^2).
134	4. Average Jerk (AJ) derivative of the acceleration, also known as "smoothness"
135	measure (mm/s^3) .
136	5. Working volume (WV) represents the volume of the convex hull for each
137	trajectory (mm ³)
138	• Angular metrics:
139	6. Angular area (AA) is a size measurement for a region of possible directions
140	from a point, which is equivalent to regions on the surface of a sphere (rad^2) .
141	7. Angular spectral arc (SPARC) quantifies movement intermittencies (angular
142	smoothness) but is independent of its amplitude or duration (no unit).
143	8. Angular Velocity (AV) is the vector measure of the rotation rate, which refers
144	to how fast an object rotates or revolves relative to another point. It is expressed
145	in radians per second (rad/s).
146	9. Angular Acceleration (AAcc) is defined as the time rate of change of angular
147	velocity (rad/s ²).
148	10. Angular Jerk (AJ) is a derivation of angular acceleration (rad/s ³).

149 All the metrics were compared according to the fetal presentation (i.e., cephalic or 150 breech) and the patient BMI (i.e., <25 or ≥ 25).

151

152 Video analysis

The videos were manually annotated using CapCut®. The image sequence and the image capture frequency were collected, as well as the starting image and finishing image. The percentage of completion of the scan according to current international guidelines was calculated (number of standard planes achieved/ number of standard planes required x100). The number of images saved per anatomic region (brain, face, heart, abdomen, skeletal, spine, genitalia, placenta, cord) was also collected.

159 *Statistical analysis*

Qualitative variables are presented with numbers and percentages. Quantitative variables are expressed with mean \pm SD. Data were analyzed by conducting the Student's *t*-test or Mann– Whitney *U*-test depending on the distribution Gaussianity. Qualitative data associations were analyzed using a chi-squared test or Fisher's exact test. Statistical analyses were performed using R via BiostaTGV (https://biostatgv.sentiweb.fr/).

165 **Results**

166 **Demographics**

167 A total of 17 scans were recorded. Sonographers' and patients' demographics are 168 presented in **Table 1**. Mean gestational age at scanning was 20.4 ± 0.7 for the fellows and 169 19.6±1.5 for the consultants.

170

171 **Probe Motion data: trajectory and angular metrics.**

Firstly, the path metrics were different according to the level of the participant (consultant or fellow), except for the mean path length of the probe trajectory which did not reach significance ($16.6\pm4.9m$ for consultants versus $27.5\pm15.4m$ for the fellows, p=0.18). The other path metrics were all significantly different according to the level of the participant (**Table 2**).

177 Secondarily, the angular metrics were not different according to the participant's level 178 of expertise (**Table 2**), to the fetal presentation (cephalic or breech) (**Table 3**) nor to the 179 patient's BMI (**Table 4**).

180

181 Video Data

182 The mean number of images captured was 61.8 ± 23.9 . The image capture frequency was 183 3.8±1.6 per minute. The starting area was the heart for 14 patients (82.3%), or the head for 2 184 patients (11.7%). The ending image was most frequently the uterine arteries for 3 patients 185 (17.6%), fetal annexes for 4 patients (23.5%) (cord, amniotic fluid, placenta), or the profile for 186 2 patients (11.7%). The mean number of images saved for fellows 75 ± 22 was and the mean 187 number for consultants was 40±3. The mean number of images needed to complete the heart 188 examination was 18.3 ± 10.3 , 7.1 ± 5.1 for the spine, and 3.8 ± 1.6 for the brain (p<0.01). The 189 comparison of the volume of images saved according to fetal presentation is presented in Table 190 5. The total number of images captured was significantly higher for breech presentation, as 191 well as the number of images captured to secure the heart examination in this presentation.

192

- 193 Discussion
- 194

195 Main Findings

There was a high heterogeneity in the number of captured images (between 32 and 100), sequence, and path metrics: path length of the probe (from 11 m to 60 m), with consultants being shorter (18.5 min versus 21.3 min), slower (12.9 mm/s versus 24.6 mm/s), and smoother (jerk about 7500 versus 15000) than fellows.

However, angular metrics were similar among sonographers, meaning that the rotation of the probe was very homogenous among sonographers even under different clinical conditions such as different BMIs or fetal positions. Angular acceleration, velocity and jerk were similar across the different groups.

The breech position seems to be most challenging, with significantly more images captured compared to the cephalic presentation (72 ± 15 versus 43 ± 21 , p=0.01) and requiring on average one additional minute of scan time (19.8 min versus 18.9 min, p=0.10). No significant differences appeared comparing BMI above or below 25.

208

209 Comparison with existent literature

210 Most of the studies analyzing motion data during scans were performed in a simulated 211 setting which allows automatic recording of probe trajectories during training and evaluation. 212 However, findings from simulation training settings need to be verified in clinical settings. 213 First, the total path length of an anomaly second-trimester scan was between 11m and 60m, as 214 compared to simulated exercises of around 1m to 2m, which was a much longer task [4,5]. Considering the working volumes, they were also larger in the clinical setting (around 16.10^6 215 mm³ than in the simulated settings (around 1.10⁶ mm³), probably due to the more limited 216 217 simulated tasks and the reduced number of degrees of freedom of the simulator probe. Second, 218 the average velocity of the probe in the present study was 24.6 mm/s for the fellows versus 12.9 219 mm/s for the consultants (p=0.02). In the simulated setting, it was the opposite : experienced 220 sonographers were faster (around 29 mm/s) than novices (around 19mm/s), meaning that the 221 experienced sonographers move the probe slower and more accurate on a real patient, 222 examining a real baby, compared to completing a task on a simulator [5]. Similarly, the acceleration was lower in consultants (170.4 mm/s^2) than fellows (328.9 mm/s^2) on real patients 223 224 (p=0.02), as compared to the simulated setting : 228 mm/s^2 for the experts versus 137mm/s^2 for 225 the novices[5]. Considering the jerk, it was much higher in the clinical setting (around 12000 226 mm/s³) versus during simulated exercises (around 4528 mm/s³) for everybody. However,

considering the metrics by a whole, they were of the same order of magnitude in both thesimulated and clinical settings.

229 The angular metrics are a handy tool for representing rotations in space. Adjusting the 230 probe rotation, and therefore the view, is a major skill in fetal ultrasound, allowing the operator 231 to obtain sagittal views starting from a transversal one. Indeed, the transversal views are easily 232 obtained solely by translating the probe from the maternal pubic bone (head circumference) to 233 the abdomen and femur in a cephalic presentation, adding very small inclinations or rotations 234 to adapt. The sagittal views (fetal profile, spine, corpus callosum) are the most challenging to 235 obtain, with lateral translations on maternal sides, rotations, and extreme inclinations of the 236 probe often needed. We hypothesized that some differences in the rotation of the probe would 237 appear according to fetal presentations por the level of the participants. However, this was not 238 the case here, with very similar angular metrics all across the cases, meaning that, the manner 239 in which the probe is rotated was relatively constant, even in different difficulty levels such as 240 different fetal positions or augmented BMI in patients.

The manner of conducting the scans may also differ slightly according to the patient BMI. It has been found that maternal obesity increases the rate of suboptimal visualization for the fetal cardiac structures by 49.8% and for the craniospinal structures by 31% [10]. In the present study, we found no differences in the trajectory nor rotation metrics using a BMI threshold of 25, although the working volume seemed to be higher for higher BMI (p=0.06). More difficulties and therefore differences may arise for more extreme BMIs. Maybe a cut-off of 30 would have been more relevant.

248

249 Strengths and Limitations

The main strength of this study is the prospective collection of synchronized video data and trajectory metrics in the real clinical setting of a university teaching hospital. This is reflective of real life compared to our previous studies in simulated settings [4,6]. Real patients are more challenging than simulators because of changing conditions of the examinations (e.g., higher BMIs, moving fetuses, different fetal presentations).

One limitation of this study is the low number of participants, which was limited by the rooms' occupation, the study's acceptability for both fellows and patients and their availability. Performing a scan in front of stressed patients is very challenging and the stakes are so high that operators are not always keen to add research challenges on top of it. Another limit is that we did not record the communication or interactions with the patient, which could be interesting to highlight how it affects the workflow of the sonographer and the sequence of images[9]. For example, some couples may ask to see the face again or the sex of the baby in the middle of the scan. A further limitation of this study was that the tracking was not perfect; in some acquisitions, the pose metrics were not correctly tracked, and there was some intermittent tracking loss. The participants for whom there was more than 30% tracking loss were therefore excluded from the analysis, mostly for trajectory metrics. Despite this, our dataset of trajectories and ultrasound videos is the first dataset of this type, and it reflects the reality of ultrasound screening units in various clinical conditions.

- 268
- 269

Implications-Interpretation

270 This study emphasized the skills needed by a sonographer, able to capture a mean of 61 271 images per patient with a mean duration of 18 minutes, representing a workflow of 3 to 4 images 272 per minute, while moving the hand at a velocity of 11 mm/s. It also requires coordinating a 273 discussion with the patient and sometimes the partner, the correct interpretation of the images 274 and, at the same time, remembering the checklist of 35 structures and 21 images needed for 275 mid-trimester scans according to the ISUOG [1]. In a study about fetal ultrasound guidance, 276 Droste et al. were able to predict the next movement that an expert sonographer would perform, 277 confirming that the way to conduct fetal ultrasound is predictable in experts [7]. However, the 278 way to move the probe and rotate it to obtain the correct planes is challenging to teach and leads 279 to long learning curves [8], [9]. Although it is now possible to model the trajectory of the probe 280 held in the operator's hands to approach the ideal trajectory, we are still a long way from 281 standardized practice, a single ideal movement, an ultrasound applicable to all clinical 282 situations, which a robot could perform.

283 Secondarily, our findings reveal that the heart seemed to be the most challenging 284 structure to examine for sonographers. A mean of 18 images of the heart was captured per scan 285 and per patient when 5 only are minimally required in the international guidelines: heart 286 activity, 4 chamber view, right outflow tract, left outflow tract, and three vessel view [1]. The 287 fetal heart is a fast-moving structure in a moving fetus, in which the possibility of more than 288 one hundred conditions can be detected, most of them life-threatening [12], [13]. This may 289 explain why so many images were saved for this organ. The sonographers were willing to secure 290 the heart examination, looking for the best planes and repeating them, especially for trickier 291 fetal positions like breech. The other challenging structure was the spine, with a mean of 7 292 images captured per scan. This can be explained by the fact that sagittal planes are more difficult 293 to obtain, and the spine is a long and moving structure that cannot verify its integrity with a 294 single image. The international recommendations mention that the scan should include 295 transversal and sagittal spine views without further details [1]. The brain, interestingly, 296 appeared to be easily investigated at first glance when normal with a mean of 3 to 4 images 297 only. After the heart activity was checked, it was generally the first structure to be analyzed. 298 The total number of images needed to finalize the scan was 2 to 3 times higher than the essential 299 number according to the ISUOG protocol. This can be explained by higher expectations and 300 requirements in a university centers, and by the fellows' wish to make their fetal examination 301 absolutely safe. However, we should advise the trainees to take their time to get each plane 302 correct rather than taking a lot of uncorrect planes. Also, we should advise them to use the 303 systematic head-to-toe approach, which was used by the consultants in our study, to reach a 304 more standardized way of scanning.

The increasing number of images required in new guidelines, 21 in the ISUOG standards, makes it more difficult to remind the checklist and makes an omission more probable. Technologies such as scan assistants that automatically assess the completion of a scan could offer a solution to address this issue [11].

309

310 Conclusion

In a clinical setting, some differences in the path metrics (velocity, acceleration, jerk and working volume) were noticed according to participant level. However, angular metrics, i.e., the quantity of rotation applied to the probe, were not affected by operator's level, fetal presentations nor patient's BMI.

- 315
- 316

317 **References**

318

319 [1] Salomon LJ, Alfirevic Z, Berghella V, Bilardo C, Hernandez-Andrade E, Johnsen SL,
320 et al. Practice guidelines for performance of the routine mid-trimester fetal ultrasound scan.

321 Ultrasound Obstet Gynecol 2011;37:116–26. https://doi.org/10.1002/uog.8831.

322 [2] Tolsgaard MG, Ringsted C, Dreisler E, Klemmensen A, Loft A, Sorensen JL, et al.
323 Reliable and valid assessment of ultrasound operator competence in obstetrics and

324 gynecology. Ultrasound Obstet Gynecol 2014;43:437–43. https://doi.org/10.1002/uog.13198.

325 [3] Drukker L, Sharma H, Karim JN, Droste R, Noble JA, Papageorghiou AT. Clinical

workflow of sonographers performing fetal anomaly ultrasound scans: deep-learning-based analysis. Ultrasound Obstet Gynecol 2022;60:759–65. https://doi.org/10.1002/uog.24975.

[4] Dromey BP, Ahmed S, Vasconcelos F, Mazomenos E, Kunpalin Y, Ourselin S, et al.
 Dimensionless squared jerk: An objective differential to assess experienced and novice probe
 maximum at the abstatric ultrasound. Preset Discr 2020. https://doi.org/10.1002/u0g.24975.

330 movement in obstetric ultrasound. Prenat Diagn 2020. https://doi.org/10.1002/pd.5855.

- [5] Le Lous M, Despinoy F, Klein M, Fustec E, Lavoue V, Jannin P. Impact of Physician
 332 Expertise on Probe Trajectory During Obstetric Ultrasound: A Quantitative Approach for
- 333 Skill Assessment. Simul Healthc 2021;16:67–72.
- 334 https://doi.org/10.1097/SIH.000000000000465.

335 [6] Fetal anomaly screening programme handbook. GOVUK 2023.

https://www.gov.uk/government/publications/fetal-anomaly-screening-programme-handbook
 (accessed September 14, 2023).

Hendler I, Blackwell SC, Bujold E, Treadwell MC, Wolfe HM, Sokol RJ, et al. The impact of maternal obesity on midtrimester sonographic visualization of fetal cardiac and

340 craniospinal structures. Int J Obes Relat Metab Disord 2004;28:1607–11.

341 https://doi.org/10.1038/sj.ijo.0802759.

[8] Le Lous M, Despinoy F, Klein M, Fustec E, Lavoue V, Jannin P. Impact of Physician
Expertise on Probe Trajectory During Obstetric Ultrasound: A Quantitative Approach for

344 Skill Assessment. Simul Healthc 2020. https://doi.org/10.1097/SIH.0000000000465.

345 [9] Drukker L, Sharma H, Droste R, Alsharid M, Chatelain P, Noble JA, et al.
 346 Transforming obstetric ultrasound into data science using eye tracking, voice recording,

347 transducer motion and ultrasound video. Sci Rep 2021;11:14109.

348 https://doi.org/10.1038/s41598-021-92829-1.

349 [10] Droste R, Drukker L, Papageorghiou AT, Noble JA. Automatic Probe Movement

Guidance for Freehand Obstetric Ultrasound. Med Image Comput Comput Assist Interv
 2020;12263:583–92. https://doi.org/10.1007/978-3-030-59716-0_56.

352 [11] Taipale P, Ammälä M, Salonen R, Hiilesmaa V. Learning curve in ultrasonographic

353 screening for selected fetal structural anomalies in early pregnancy. Obstet Gynecol

354 2003;101:273-8. https://doi.org/10.1016/s0029-7844(02)02590-5.

Patel H, Chandrasekaran D, Myriokefalitaki E, Gebeh A, Jones K, Jeve YB, et al. The
Role of Ultrasound Simulation in Obstetrics and Gynecology Training: A UK Trainees'

- 357 Perspective. Simul Healthc 2016;11:340–4. https://doi.org/10.1097/SIH.00000000000176.
- 358 [13] Blyth M, Howe D, Gnanapragasam J, Wellesley D. The hidden mortality of
- 359 transposition of the great arteries and survival advantage provided by prenatal diagnosis.
- 360 BJOG 2008;115:1096–100. https://doi.org/10.1111/j.1471-0528.2008.01793.x.
- 361 [14] Dulgheroff FF, Peixoto AB, Petrini CG, Caldas TMR da C, Ramos DR, Magalhães

362 FO, et al. Fetal structural anomalies diagnosed during the first, second and third trimesters of

- 363 pregnancy using ultrasonography: a retrospective cohort study. Sao Paulo Med J 264 2010;127:201, 400, https://doi.org/10.1500/1516.2100.2010.026006002010
- 364 2019;137:391–400. https://doi.org/10.1590/1516-3180.2019.026906082019.
- 365 [15] Stirnemann JJ, Besson R, SpaggiarI E, Rojo S, Loge F, Peyro-Saint-Paul H, et al.
- 366 Development and clinical validation of real-time artificial intelligence diagnostic

- companion for fetal ultrasound examination. Ultrasound Obstet Gynecol 2023. https://doi.org/10.1002/uog.26242. 367
- 368 369



Fig. 1. The probe's position was tracked using the Aurora[®] electromagnetic tracking system (NDI Inc., Ontario, Canada), which table was placed under the patient (a-b). To track the probe a 6 degree-of-freedom (DOF) sensor was attached to the ultrasound probe by means of a custom 3-D printed holder (c). To record the video, an <u>AV.io</u> HD[®], Epiphan videos, USA was used (d). The data were synchronized and recorded using the Scanview Data Recorder software[®] (e).

 Table 1. Participant demographics.

Sonographers' background	N=17	Patients' demographics	N=17
Age		Age	
Mean±SD	38±5.6	Mean±SD	35.7±6.5
Years of experience		Gestational Age	
Mean±SD	6±5.6	Mean±SD	20.4±1.7
Less than 5 years N (%)	14		
More than 5 years N (%)	3		
Background N (%)		BMI	
OB/GYN	13	Mean±SD	26.9±6.3
Midwife	3	<25	10
Radiologist	1	≥25	7
		Fetal Presentation	
Ultrasound senior fellows	14	Cephalic	7
Fetal medicine Consultants	3	Breech	10

	Breech N=10	Cephalic N=7	р
% completion of the scan	98.8%±3.17	95.8%±9.3	0.71
Total number of images captured	72±15	43±21	0.01
Number of images per anatomical region			
Brain images	4±3	3±1	0.39
Face	6.6±4.0	4.8±2.3	0.39
Heart	23±9	12±8	0.04
Spine	7±6	5±2	0.46
Abdomen circumference	1.2±0.7	1.4±0.5	0.33
Genitalia	1.6±1.4	1.8±1.4	0.23
Skeletal (except spine)	9.8±3.7	9.0±3.1	0.46
Placenta	0.8±0.6	1.25±0.4	0.36
Umbilical cord	1.5±1.1	1.6±0.8	1

Table 2. Number of images captured per anatomic region according to the fetal presentation (breech or cephalic).