



HAL
open science

Biliary tract viability assessment and sequential hypothermic-normothermic perfusion in liver transplantation

Heithem Jeddou, Stylianos Tzedakis, Karim Boudjema

► **To cite this version:**

Heithem Jeddou, Stylianos Tzedakis, Karim Boudjema. Biliary tract viability assessment and sequential hypothermic-normothermic perfusion in liver transplantation. *Hepatobiliary Surgery and Nutrition*, 2024, 13 (3), pp.505-508. 10.21037/hbsn-24-144 . hal-04605130

HAL Id: hal-04605130

<https://hal.science/hal-04605130v1>

Submitted on 24 Sep 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Biliary tract viability assessment and sequential hypothermic-normothermic perfusion in liver transplantation

Heithem Jeddou^{1,2}, Stylianos Tzedakis^{3,4}, Karim Boudjema^{1,2}

¹Department of Hepatobiliary and Digestive Surgery, University Hospital, Rennes 1 University, Rennes, France; ²Inserm, EHESP, Irset (Institut de recherche en santé, environnement et travail) - UMR_S 1085, Rennes 1 University, Rennes, France; ³Department of Digestive, Pancreatic, Hepatobiliary and Endocrine Surgery, Cochin Hospital, Assistance Publique - Hôpitaux de Paris (AP-HP), Paris, France; ⁴Université Paris Cité, Paris, France

Correspondence to: Heithem Jeddou, MD. Department of Hepatobiliary and Digestive Surgery, University Hospital, Rennes 1 University, 2, Rue Henri Le Guilloux, 35033 Rennes Cedex 9, France; Inserm, EHESP, Irset (Institut de recherche en santé, environnement et travail) - UMR_S 1085, Rennes 1 University, Rennes, France. Email: heithem.jeddou@chu-rennes.fr.

Comment on: Mergental H, Laing RW, Kirkham AJ, *et al.* Discarded livers tested by normothermic machine perfusion in the VITTAL trial: Secondary end points and 5-year outcomes. *Liver Transpl* 2024;30:30-45.

Keywords: Biliary tract; viability assessment; normothermic perfusion; hypothermic perfusion; liver transplantation (LT)

Submitted Mar 10, 2024. Accepted for publication Apr 15, 2024. Published online May 23, 2024.

doi: 10.21037/hbsn-24-144

View this article at: <https://dx.doi.org/10.21037/hbsn-24-144>

We read with great interest the study of Mergental *et al.* (1) reporting the 5-year outcomes of the VITTAL trial (2,3). VITTAL was a prospective, non-randomized, single-arm trial that tested end-ischemic oxygenated normothermic machine perfusion (NMP) with a “back-to-base” strategy to evaluate, and potentially transplant, liver grafts declined by all liver transplantation (LT) centers in the United Kingdom. For a liver to be considered viable, it had to metabolize perfusate lactate to a concentration of ≤ 2.5 mmol/L within four hours from the start of perfusion, and meet at least two of the following criteria: bile production without a defined quantity; maintenance of perfusate pH above 7.3; glucose consumption in the perfusate; maintenance of stable arterial and portal flow above 150 and 500 mL/min, respectively; maintenance of graft suppleness and homogeneous perfusion (4). Thirty-one discarded human donor livers underwent viability testing by using end-ischemic NMP, of which 22 (71%) livers were subsequently transplanted. The primary outcome of the trial was graft survival rate at 90 days and it was 100%.

The authors reported the long-term outcomes of patients and liver grafts from the VITTAL trial. The graft and patient survival at 1, 3, and 5 years were respectively 91%, 82%, and 82%, and 100%, 91%, and 82%. Among the 22 grafts deemed transplantable, 12 grafts were harvested from donors after brain death (DBD) and

10 grafts were harvested from donors after circulatory death (DCD). Among the 10 DCD livers, three developed non-anastomotic biliary strictures (NAS) (30%) and required retransplantation. The fourth liver retransplantation in the series was performed in a patient who had received a DBD graft and was complicated by hepatic artery thrombosis.

Anastomotic biliary stricture occurred in two cases; each time treated by endoscopic retrograde stenting. Additionally, biliary tract irregularities on magnetic resonance cholangiopancreatography (MRCP), without liver function test alteration, were observed in three cases. At 5 years, four recipients had died, all of which had a functioning graft. Two of them had been retransplanted for NAS. The cause of death was recurrence of primary cancer in three cases and chronic rejection due to poor treatment adherence in one case.

Comments

This study presents the long-term results of liver grafts deemed non-transplantable by conventional criteria but transplanted after assessment via NMP. Authors should be appraised for this high-quality study. However, biliary complication incidence was elevated and two critical points require discussion: (I) the viability criteria applied in this trial; and (II) the exclusive application of NMP at the end of

the ischemic phase.

The viability criteria used in the VITTAL trial

During NMP, the liver is maintained in an almost physiological environment and recovers its metabolic functions. This leads to the restoration of aerobic metabolism, as evidenced by glucose consumption, lactate clearance, maintenance of acid-base balance, and bile production. There are two types of graft viability criteria during NMP: hepatocellular and cholangiocellular viability criteria (5-7). The quality of the graft can be assessed by evaluating: (I) its metabolic function (lactate clearance, pH maintenance, urea and coagulation factors production); (II) its excretory function (bile production, indocyanine green clearance) (8); (III) its appearance, consistency, and hemodynamics (flow, pressure, and resistance of the hepatic artery and portal vein) during perfusion. The function of the biliary epithelium can be assessed by analyzing bile composition (pH, glucose levels, and bicarbonate levels) (9). In the VITTAL trial, only hepatocellular viability was considered for evaluating liver transplantability during NMP although bile composition was studied, and biliary biopsies were performed. These two factors were not considered in the graft evaluation nor in the decision to proceed with transplantation or not and a posteriori, NAS developed in DCD livers that had a biliary pH <7.65 and a biliary bicarbonate level <25 mmol/L. Histological evaluation of these grafts revealed advanced biliary lesions and arterial media necrosis (1).

It has indeed been shown that bile composition during NMP can predict the occurrence of NAS (9,10). Transplanting livers from DCD donors without controlling bile composition during NMP exposes the recipient to NAS (5,6). This could explain the high incidence of NAS in the VITTAL trial (30% retransplantation rate for NAS in the DCD group). These biliary viability criteria were further refined in the DHOPE-COR-NMP trial (10). It had been shown that absolute values of bile pH, bicarbonate, and glucose during NMP may not be the most appropriate markers to test cholangiocellular viability and that the delta between bile and perfusate levels of pH, bicarbonate, and glucose should be used to identify bile alkalization and glucose reabsorption by the biliary epithelium (10).

Use of end-ischemic NMP alone

In the VITTAL trial, an end-ischemic NMP was chosen (2).

End-ischemic NMP exposes the graft to ischemia-reperfusion injury, to which the bile ducts are extremely sensitive (11). Since the beginning of the VITTAL trial in 2016, the Cambridge team shared a similar experience showing high percentages of biliary complications after NMP of livers from DCD donors (5,6). It has been demonstrated that a short period of hypothermic oxygenated perfusion (HOPE) reduces ischemia-reperfusion injuries (12,13). The Groningen group had shown a beneficial effect of the combination of arterial and portal hypothermic perfusion [Dual Hypothermic Oxygenated Perfusion (DHOPE)] and NMP (DHOPE-NMP) on cholangiocellular function. Biliary bicarbonate concentration was higher in the DHOPE-NMP group compared to the NMP group. Not only do biliary bicarbonates may serve as evidence of good cholangiocellular function but may also serve as protection to the biliary epithelium against the effects of toxic hydrophobic bile salts (14). This phenomenon has been described as the “bicarbonate umbrella” (14). In 2017, the same Groningen group launched the DHOPE-COR-NMP trial (combination of one-hour DHOPE, followed by one hour of controlled oxygenated rewarming, and subsequent NMP), aiming to reduce ischemia-reperfusion injuries at the beginning of NMP (6-8,10). All livers in the DHOPE-COR-NMP trial met the hepatocellular viability criteria used in the VITTAL study, despite the fact that all livers came from high-risk DCD donors whose median age was significantly higher than in the VITTAL study (10). With a median follow-up of 12 months, they observed only one case of post-transplant cholangiopathy after DHOPE-COR-NMP of high-risk DCD liver grafts, and no recipient was retransplanted (10). The same Groningen team has more recently reported its results using the DHOPE-NMP combination (15). One hundred and five procedures were performed to test grafts initially declined for transplantation, 69 livers were deemed transplantable and were transplanted. The graft and recipient survival at 1 and 3 years were respectively 93% and 91%, and 99% and 97%. Two patients (3%) developed NAS and required retransplantation. After these findings, the viability criteria were modified using the delta between bile and perfusate levels of pH, bicarbonate, and glucose rather than their absolute biliary values. Using these cholangiocellular viability criteria, no cases of NAS were observed (16).

Conclusions

NMP is an important tool for evaluating and selecting

donor livers which, based on classical criteria, were initially considered non-transplantable. The VITTAL trial has pushed the boundaries of using these high-risk liver grafts after NMP evaluation. However, end-ischemic NMP alone, as applied in the VITTAL study, does not seem to protect the bile ducts from ischemia-reperfusion injuries, particularly for livers from DCD donors. Therefore, a short period of DHOPE before NMP, along with the use of cholangiocellular viability criteria during NMP, could help reduce post-transplant morbidity and the risk of graft loss due to NAS.

Acknowledgments

Funding: None.

Footnote

Provenance and Peer Review: This article was commissioned by the editorial office, *Hepatobiliary Surgery and Nutrition*. The article did not undergo external peer review.

Conflicts of Interest: All authors have completed the ICMJE uniform disclosure form (available at <https://hbsn.amegroups.com/article/view/10.21037/hbsn-24-144/coif>). The authors have no conflicts of interest to declare.

Ethical Statement: The authors are accountable for all aspects of the work in ensuring that questions related to the accuracy or integrity of any part of the work are appropriately investigated and resolved.

Open Access Statement: This is an Open Access article distributed in accordance with the Creative Commons Attribution-NonCommercial-NoDerivs 4.0 International License (CC BY-NC-ND 4.0), which permits the non-commercial replication and distribution of the article with the strict proviso that no changes or edits are made and the original work is properly cited (including links to both the formal publication through the relevant DOI and the license). See: <https://creativecommons.org/licenses/by-nc-nd/4.0/>.

References

- Mergental H, Laing RW, Kirkham AJ, et al. Discarded livers tested by normothermic machine perfusion in the VITTAL trial: Secondary end points and 5-year outcomes. *Liver Transpl* 2024;30:30-45.
- Laing RW, Mergental H, Yap C, et al. Viability testing and transplantation of marginal livers (VITTAL) using normothermic machine perfusion: study protocol for an open-label, non-randomised, prospective, single-arm trial. *BMJ Open* 2017;7:e017733.
- Mergental H, Laing RW, Kirkham AJ, et al. Transplantation of discarded livers following viability testing with normothermic machine perfusion. *Nat Commun* 2020;11:2939.
- Mergental H, Perera MT, Laing RW, et al. Transplantation of Declined Liver Allografts Following Normothermic Ex-Situ Evaluation. *Am J Transplant* 2016;16:3235-45.
- Watson CJE, Kosmoliaptsis V, Pley C, et al. Observations on the ex situ perfusion of livers for transplantation. *Am J Transplant* 2018;18:2005-20.
- Watson CJE, Kosmoliaptsis V, Randle LV, et al. Normothermic Perfusion in the Assessment and Preservation of Declined Livers Before Transplantation: Hyperoxia and Vasoplegia-Important Lessons From the First 12 Cases. *Transplantation* 2017;101:1084-98.
- Brüggenwirth IMA, de Meijer VE, Porte RJ, et al. Viability criteria assessment during liver machine perfusion. *Nat Biotechnol* 2020;38:1260-2.
- Eshmuminov D, Becker D, Bautista Borrego L, et al. An integrated perfusion machine preserves injured human livers for 1 week. *Nat Biotechnol* 2020;38:189-98.
- Matton APM, de Vries Y, Burlage LC, et al. Biliary Bicarbonate, pH, and Glucose Are Suitable Biomarkers of Biliary Viability During Ex Situ Normothermic Machine Perfusion of Human Donor Livers. *Transplantation* 2019;103:1405-13.
- de Vries Y, Berendsen TA, Fujiyoshi M, et al. Transplantation of high-risk donor livers after resuscitation and viability assessment using a combined protocol of oxygenated hypothermic, rewarming and normothermic machine perfusion: study protocol for a prospective, single-arm study (DHOPE-COR-NMP trial). *BMJ Open* 2019;9:e028596.
- Teoh NC, Farrell GC. Hepatic ischemia reperfusion injury: pathogenic mechanisms and basis for hepatoprotection. *J Gastroenterol Hepatol* 2003;18:891-902.
- van Rijn R, van Leeuwen OB, Matton APM, et al. Hypothermic oxygenated machine perfusion reduces bile duct reperfusion injury after transplantation of donation after circulatory death livers. *Liver Transpl* 2018;24:655-64.
- van Leeuwen OB, de Vries Y, de Meijer VE, et al. Hypothermic machine perfusion before viability testing of previously discarded human livers. *Nat Commun*

- 2021;12:1008.
14. Westerkamp AC, Karimian N, Matton AP, et al. Oxygenated Hypothermic Machine Perfusion After Static Cold Storage Improves Hepatobiliary Function of Extended Criteria Donor Livers. *Transplantation* 2016;100:825-35.
 15. van Leeuwen OB, Bodewes SB, Porte RJ, et al. Excellent long-term outcomes after sequential hypothermic and normothermic machine perfusion challenges the importance of functional donor warm ischemia time in DCD liver transplantation. *J Hepatol* 2023;79:e244-5.
 16. van Leeuwen OB, Bodewes SB, Lantinga VA, et al. Sequential hypothermic and normothermic machine perfusion enables safe transplantation of high-risk donor livers. *Am J Transplant* 2022;22:1658-70.

Cite this article as: Jeddou H, Tzedakis S, Boudjema K. Biliary tract viability assessment and sequential hypothermic-normothermic perfusion in liver transplantation. *HepatoBiliary Surg Nutr* 2024;13(3):505-508. doi: 10.21037/hbsn-24-144