



HAL
open science

Benefits of intraoral cobotic scanning in the dental CAD/CAM chain

Rebecca Allport, Lebon Nicolas, Swaminath Venkateswaran, Floriane Laverne

► **To cite this version:**

Rebecca Allport, Lebon Nicolas, Swaminath Venkateswaran, Floriane Laverne. Benefits of intraoral cobotic scanning in the dental CAD/CAM chain. Journée de la Recherche en Odontologie, Jul 2024, Montrouge, France. , 2024. hal-04637456

HAL Id: hal-04637456

<https://hal.science/hal-04637456>

Submitted on 6 Jul 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

Introduction

- In the last years, the **modernisation** of the **digital impression** has been made possible thanks to **digital intra-oral scanners (IOS)** and **additive manufacturing** machines, which are increasingly used in dentistry, and enable the **quality** of **dental care** to be improved.
- STL files** obtained with IOS are the **first step** in digital manufacture of **dental prostheses**.
- Effective patient care are reflected in **precision**, **accuracy** and **fidelity** of CAD files [1-2].
- Many factors are involved, including the practitioner [3].

Objectives

- Define a **repeatable** and **reproducible** experimental measurement protocol with the cobot.
- Programming** the cobot to teach the practitioner the scanning gesture.
- Carrying out a **test campaign**.
- In vitro experiments :
Experiment 1: Scanning for an entire dental with spheres on the second molars and first premolars.
Experiment 2: Scanning for an entire dental arch.

Materials & Methods

Cobot UR5e HMI-MBS:

- Robotiq Wrist Camera
- Robotiq 2F-85 Adaptive Gripper

Reference object for Experiment 1:



Reference object for Experiment 2:

Frasaco and articulator.




Software:

- Creation of the trajectory on Solidworks.
- Simulation and adjustment on RoboDK.

Care Stream CS3800 IOS:


Using only the wireless side with its standard tip.

Tested trajectories:



Results

- Trajectory reproducibility
- Cobotic scan
- STL files
- Complete simulation environment for scanning
- In vitro scanning



Perspectives

- Comparison** of the **results** obtained (accuracy, precision and fidelity of CAD files) with those obtained by the practitioner.
- Vision** of cobots in dentistry in the future, in **autonomous** or **collaborative** form.

[1] Troizier Cheyne M, Lebon N, Laverne F. (2023). "Impact of Intra oral Scanner Sleeves on STL File Accuracy: A Metrological Approach". Computer Aided Design and Applications, 21(4), pp.567-580. 10.14733/cadaps.2024.567-580. hal-04284162.

[2] L. Dupagne, L. Tapie, N. Lebon, B. Mawussi, Comparison of the acquisition accuracy and digitizing noise of 9 intraoral and extraoral scanners: An objective method, The Journal of Prosthetic Dentistry, Volume 128, Issue 5, 2022, 1032-1040, <https://doi.org/10.1016/j.prosdent.2021.02.005>.

[3] Thomas AA, Jain RK. Influence of Operator Experience on Scanning Time and Accuracy with Two Different Intraoral Scanners - A Prospective Clinical Trial. Turk J Orthod. 2023; 36(1): 10-14.

[4] Crétot, M. (2003). L'arcade dentaire humaine: morphologie. Editions CdP-Groupe Liaisons.