



**HAL**  
open science

## A flexible instrument for robotic assisted minimally invasive esophagectomy

Paul Tucan, Iosif Birlescu, Alexandra Puscasu, Bogdan Gherman, Daniel Jucan, Tiberiu Antal, Calin Vaida, Adrian Pislă, Damien Chablat, Doina Pislă

### ► To cite this version:

Paul Tucan, Iosif Birlescu, Alexandra Puscasu, Bogdan Gherman, Daniel Jucan, et al.. A flexible instrument for robotic assisted minimally invasive esophagectomy. 9th European Conference on Mechanism Science, Sep 2024, Padova, Italy. hal-04644139

**HAL Id: hal-04644139**

**<https://hal.science/hal-04644139v1>**

Submitted on 10 Jul 2024

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

# A flexible instrument for robotic assisted minimally invasive esophagectomy

Paul Tucan<sup>1</sup>[0000-0001-5660-8259], Iosif Birlescu<sup>1</sup>[0000-0002-4026-2318], Alexandru Pusca<sup>1</sup>[0000-0002-5804-575X], Bogdan Gherman<sup>1</sup>[0000-0002-4427-6231], Daniela Jucan<sup>1</sup>[0009-0004-0219-9858], Tiberiu Antal<sup>1</sup>[0000-0002-0042-5258], Calin Vaida<sup>1</sup>[0000-0003-2822-9790], Adrian Pislă<sup>1</sup>[0000-0002-5531-6913], Damien Chablat<sup>1,2</sup>[0000-0001-7847-6162], Doina Pislă<sup>1,3\*</sup>[0000-0001-7014-9431]

<sup>1</sup> CESTER, Technical University of Cluj-Napoca, 400114 Cluj-Napoca, Romania

<sup>2</sup> École Centrale Nantes, Nantes Université, CNRS, LS2N, UMR 6004, F-44000 Nantes, France

<sup>3</sup> Technical Sciences Academy of Romania, 030167 - Bucharest, Romania

\*Corresponding author: [Doina.Pisla@mep.utcluj.ro](mailto:Doina.Pisla@mep.utcluj.ro)

**Abstract.** The paper presents a medical flexible instrument for esophagus stapling that can be integrated in robotic assisted esophagectomy. Starting with several considerations required for esophagectomy surgery, the forward kinematic model of the instrument is presented, followed by a numerical and a graphical simulation to validate the mathematical model of the instrument. A 3D printed mockup is developed in order to test the design of the instrument while simulating the stapling procedure.

**Keywords:** medical instrument, robotic-assisted surgery, kinematic modeling, esophagus stapling.

## 1 Introduction

Esophageal cancer ranks 9th globally among the most common types of cancer and 6th in terms of mortality rates. Annually, more than 500,000 cases of death are reported worldwide, with 16,640 cases in the United States alone. The mortality rate is reported to be 87.13%, affecting 14,500 individuals [1, 2].

One effective way to eliminate cancer, in addition to neoadjuvant therapy (chemotherapy and chemoradiotherapy), remains the resection of the diseased tissue [2-4].

The surgery for resection of esophageal cancer has evolved over time, with surgical interventions ranging from open procedures, where a large incision of up to 250 millimeters is made, to minimally invasive approaches. Minimally invasive techniques involve a minimal number (4-5) of incisions, typically ranging from 10 to 20 millimeters in diameter, and may also include robotic assisted surgery. These approaches aim to reduce the patient's hospitalization time, blood loss, postoperative trauma, pain, and result in a cosmetically superior outcome compared to classical surgery [5, 6].

To mitigate the disadvantages of classical surgery and improve the advantages of minimally invasive esophagectomy (MIE), a new technique was developed namely robot-assisted minimally invasive thoraco-laparoscopic esophagectomy (RAMIE). Based on the advantages presented in [7, 8] the RAMIE technique also come with a series of dis-

advantages like: poor intraoperative workspace for instruments, high price for surgery, poor pulmonary functioning generated by the thoracoscopic approach, less dexterity of the instruments, and the use of the rigid instruments. One of the major drawbacks during the MIE is the limited intraoperative workspace, where the surgeon has to manipulate several medical instruments with poor dexterity (current instruments used are stiff or have a low dexterity) in order to perform all the procedural steps of the surgery. To overcome this disadvantage and to increase the dexterity of the instrument and the workspace, this paper presents an innovative flexible instrument (patent pending [9]) used for esophagectomy. The medical instrument uses the RCM (Remote Center of Motion) principle, and its design includes several bends actuated using cable joints to allow bending of the tip of the instrument inside the patient to perform stapling of the esophagus before the anastomosis.

Following the introduction section, the paper is structured as follows: Section II, defines the surgical task and the instruments required for esophagectomy, Sections III and IV present the kinematic model and 3D modeling of the articulated instrument, and Sections V and VI present the workspace generation and the 3D printed prototype of the articulated instrument, concluding with the conclusions of this work.


## **2 Considerations regarding the esophagectomy procedure**

There are several methods to perform the MIE procedure:

- Transhiatal esophagectomy [10] -where the surgeon performs two incisions, one at the abdominal level and the other one at the cervical level.
- IvorLewis esophagectomy [11] -is a two-phase surgical procedure. The first phase is the abdominal approach where the esophagus is accessed via abdomen and the second phase implies stapling the esophagus for anastomosis using a thoracic incision. Currently the instruments used in this step are rigid or their dexterity is less, instrument presented in this work contribute to improving dexterity during tissues suture.
- McKeown esophagectomy [12]- involves mobilization of the thoracic esophagus with lymph node dissection (performed thoracoscopically or through an open approach), exploration of the abdomen, mobilization of the stomach with lymph node dissection, and finally, creation of a left cervical incision for anastomosis.
- Thoracoabdominal esophagectomy [13] - for many years, the left thoracoabdominal esophagectomy has served as the conventional method for resecting and reconstructing the gastroesophageal junction. This approach offers better visibility for dissection, placement of the esophageal substitute, and execution of the anastomosis.

For developing the flexible instrument, the Ivor Lewis esophagectomy approach was considered, being one of the common methods used currently in Romanian hospitals, unfortunately, only conventional procedure is performed, but esophagectomy can also be robotic assisted using the da Vinci Xi robot [14]. The medical protocol for robotic assisted Ivor Lewis esophagectomy approach is presented in [11] and the entry ports required for abdominal phase and thoracic phase are presented in Fig.1. Due to large number of instruments used during the procedure, the workspace is considerably limited, thus requiring specially designed instruments, to perform specific and personalized tasks, with


respect to the medical protocol, in order to reduce the time spent in the surgery, improve the outcomes of the surgery, reduce risks caused by instrument malfunctions (improper stitches, unsuccessful anastomosis) and allow fast post-operative recovery. In this scope the instrument presented within this paper, is specially designed to allow esophagus sectioning by stapling. Due to the fact that the stapler is inserted in the thoracic volume through an 8 mm port, the range of motion regarding the angular displacement of the instrument with respect to the RCM point is considerably limited by intercostal space that does not allow large amplitude motions, requiring a special instrument, able to bend beyond the point of RCM in order to allow stapling of the esophagus.



**Fig.1.** The ports placements during robotic assisted Iver Lewis esophagectomy.


### 3 Kinematics of the flexible instrument

Based on data presented in the previous section of the paper, Fig.2 presents the kinematic scheme of the flexible instrument designed for esophagus stapling before anastomosis. The instrument has 5 degrees of mobility. First motion (Motion 1) allows revolving the instrument around its longitudinal axis ( $OY$ ), this motion is required in order to change the position of the tip of the instrument after inserted through the trocar and is performed using  $q_1$  active joint. The second motion (Motion 2) represents the flexible part of the instrument, being a cable actuated mechanism that allows bending of the tip of the instrument using around revolute joint  $R_1$ . The motion is performed using active joint  $q_2$  that rotates the cable pulley for bending the instrument. The third motion (Motion 3) is performed around  $R_7$  revolute joint being also a cable mechanism actuated using  $q_3$  active joint. Motion 4 is the opening/closing motion of the stapler, also a cable mechanism actuated by  $q_4$ , and Motion 5 is the stapling motion, where a special mechanism slides inside the jaws of the stapler in order to perform the sectioning and the stapling of the esophagus, this motion is performed using  $q_5$  joint and a wire transmission.



**Fig. 2.** The kinematic scheme of the flexible instrument.

The parameters of the flexible instrument are illustrated in Fig.2 where  $l_1$  is length of the rigid segment,  $l_e$  is length of the bending elements,  $l_3$  is length of the segment between revolute joint  $R_7$  and the base of the stapler, and  $l_4$  is length of the stapler. Regarding the lengths of the active instruments used in robot-assisted surgery, they range between 420 - 470 mm. The maximum insertion of instrument into the patient's body is limited to 270 mm, and the flexible length of the instrument is between 60 - 130 mm ( $l_e - l_4$ ) [15-19]. To define the motion of each flexible element of the instrument the kinematic scheme from Fig. 3 is used. Point  $O_0$  represents the end of the rigid part of the instrument, and the flexible part of the instrument up to the last revolute joint before the gripper is composed of 6 flexible elements. The diameter of each element is 10 mm and the distance between two consecutive elements is constant. The bending of each element can be defined by two angles:  $\psi$  is rotation around the  $Z_0$  axis,  $\varphi$  is bending angle around revolute joint,  $\theta$  is the bending angle for the flexible part,  $L_1$  and  $L_2$  are the lengths of the flexible element for  $OX$  and  $OZ$  axis.



**Fig. 3.** Kinematic scheme of flexible instrument.

The forward geometric model of the bending section is computed using the coordinates of the point  $O_0$  in order to determine the coordinates of the last bending elements. The length between two consecutive elements is noted with  $l_e$  and the position of each  $k$ -element of the bending section can be obtained using Eq. 1.

$$\begin{cases} X_k = l_e \cdot \sin \varphi \cdot \cos \psi \\ Y_k = l_e \cdot \sin \varphi \cdot \sin \psi \\ Z_k = l_e \cdot \cos \varphi \end{cases}, k = 1 \rightarrow n \quad (1)$$

where:  $\varphi = 2 \cdot \theta / (n + 1)$  and  $\psi = \arctan(Y_k / X_k)$

The coordinates of point B (the revolute joint of the last bending element) are:

$$\begin{cases} X_B = L_2 \cdot \cos \psi ; \\ Y_B = L_2 \cdot \sin \psi ; \\ Z_B = L_1 \end{cases} \quad (2)$$

$$\text{and } L_1 = \frac{l_e}{n} \cdot \frac{\sin((n \cdot \varphi)/2) \cdot \cos((n+1) \cdot \varphi/2)}{\sin(\varphi/2)}, L_2 = \frac{l_e}{n} \cdot \frac{\sin((n+1) \cdot \varphi/2) \cdot \sin(n \cdot \varphi/2)}{\sin(\varphi/2)}$$


The second motion generated by the  $q_2$  is given by Eq.3.

$$q_2 = L_1 \cdot \tan^2(\theta) \quad (3)$$

where:  $\tan(\theta) = L_2 / L_1$

#### 4 CAD of the flexible medical instrument


The flexible instrument (Fig. 4) was modeled using the Siemens NX program and the material chosen for the instrument is ABS. With respect to the medical protocol for Iver Lewis robotic assisted esophagectomy the diameter of the instrument is 10 mm to be inserted into a trocar of 12 mm and the overall length as 465 mm.




**Fig. 4.** Modeling of the articulated instrument with the revolute joints and the stapling mechanism.

The bending elements of the instrument are connected through revolute passive joints and are driven by cables running along designated paths from the initial element to the last. Each element includes a section with a ten-degree angled surface to facilitate the bending of the instrument. In the right part of fig. 5 the cable paths are presented. Path  $W1$  and  $W3$  are for instrument bending (joint  $q_2$ ), path  $W2$  and  $W4$  are for stapler actuation and stapling operation ( $q_4$  and  $q_5$ ) and center path  $W5$  is for the stapler orientation (joint  $q_3$ ). Fig. 6 presents the workspace boundary of the flexible instrument, the workspace is almost half a sphere of radius  $R$ , where  $R$  represents the length of the instrument

from  $R_2$  to the tip of the instrument. Along the bending of the flexible elements of the instrument, additional dexterity is given by revolute joint  $q_3$  which allows various orientation of the stapler within the intraoperative workspace.




**Fig. 5.** Bending elements of the instrument.



**Fig. 6.** Workspace boundary of the instrument

## 5 Simulation of the flexible instrument tip

To validate both the mathematical model and the design of the instruments, several simulations were conducted. This involved a MATLAB simulation employing the mathematical model of the medical instrument, the simulation material has not been defined in MATLAB, alongside a graphical simulation using the CAD representation of the instrument. Both simulations used same input data namely the coordinates of the initial point of the tip of the instrument  $C_1$  ( $X_{C1}=57.30$  mm,  $Y_{C1}=-0.9$  mm,  $Z_{C1}=-554.61$ mm,  $\varphi_{C1}=-41.82^\circ$ ,  $\psi_{C1}=10.19^\circ$ ) and the coordinates of the final point  $C_2$  ( $X_{C2}=-17.62$  mm,  $Y_{C2}=121.51$  mm,  $Z_{C2}=-430.81$ mm,  $\varphi_{C2}=36.77^\circ$ ,  $\psi_{C2}=89.86^\circ$ ). The trajectory of the tip of instrument between the two points is illustrated in Fig. 7.




**Fig. 7.** Simulated trajectory of the tip of the instrument


Data extracted from these two simulations (Siemens NX and MATLAB) were parsed to a MATLAB program where were graphically represented in Fig. 9.

Fig. 8 illustrates the displacement, velocity, and acceleration of the tip of the instrument. Additionally, data from the graphical simulation is overlaid on the same graph using a black dotted line. While there is not a perfect overlap between the two sets of data, the variances are minimal, and both simulations exhibit a similar curve profile, confirming the efficiency of the mathematical model and the design of the instrument.

Also, for validation purposes, a prototype of the flexible instrument was manufactured using Stratasys Fortus 380mc 3D printer, the 3D printing being made with ABS. For the prototype a 2:1 scale was used in order to check the functionality of the instrument. The prototype of the instrument and the mock-up used to test the instrument functionality is presented in Fig. 9. The functionality of the instrument was tested by inserting it into a testing mockup, and the grasping and manipulation of the tissue were performed using its tips. The instrument behaved accordingly, encouraging further development on a 1:1 scale, adapting the instrument tip to the actual stapler, and laboratory test using the actuated instrument.



**Fig.8.** Time history diagrams of the coordinates, velocity, and acceleration of the instrument tip



**Fig.9.** Testing the functionality of the proposed instrument

## 6 Conclusion

This paper presents an innovative flexible medical instrument designed for minimally invasive esophageal surgery. By incorporating several cable-actuated bending elements, the access for the esophagus stapling procedure is improved, allowing bending of the stapler around two axes after passing through the RCM. The workspace definition of the instrument was achieved based on kinematic modeling and CAD design, generating promising results that encourage further developments. The validation of kinematic modeling and the functionality of the instrument were conducted through simulations in MATLAB and Siemens NX. Future studies will focus on prototyping the instrument using biocompatible materials (the instrument will be printed using Stratasys J5 MediJet 3D printer) and adapting it with the tip of a real stapler, developing a control system for the final prototype, and testing the instrument under laboratory conditions.

## Acknowledgements

This work was supported by the project New smart and adaptive robotics solutions for personalized minimally invasive surgery in cancer treatment - ATHENA, funded by European Union – NextGenerationEU and Romanian Government, under National Re-



covery and Resilience Plan for Romania, contract no. 760072/23.05.2023, code CF 116/15.11.2022, through the Romanian Ministry of Research, Innovation and Digitalization, within Component 9, investment I8.

## References

1. J. Straatman P J M et al (2016) Systematic review of patient-reported outcome measures in the surgical treatment of patients with esophageal cancer, *Diseases of the Esophagus*, Volume 29, Issue 7, 1 October 2016, Pages 760–772.
2. Sheikh M et al (2023) Current Status and Future Prospects for Esophageal Cancer. *Cancers* 2023, 15, 765.
3. Tagkalos E et al (2020) Robot-assisted minimally invasive esophagectomy (RAMIE) compared to conventional minimally invasive esophagectomy (MIE) for esophageal cancer: a propensity-matched analysis. *Dis Esophagus*. 2020 Apr 15;33(4).
4. Seto Y et al (2017) Robotic surgery for esophageal cancer: Merits and demerits. *Ann Gastroenterol Surg*. 2017 Aug 14;1(3):193-198.
5. Ruurda J P et al (2015) Robot-assisted minimally invasive esophagectomy for esophageal cancer: A systematic review. *J Surg Oncol*. 2015 Sep;112(3):257-65.
6. Pisla D et al (2021) New approach to hybrid robotic system application in Single Incision Laparoscopic Surgery. *Acta Technica Napocensis* 2021, 64, 369–378.
7. Vaida C et al (2020) A new robotic system for minimally invasive treatment of liver tumours. *Proc. Rom. Acad. Ser. A Math. Phys. Tech. Sci. Inf. Sci.* 2020, 21, 271–363.
8. Van der Sluis P C et al (2018) Robot assisted minimally invasive esophagectomy (RAMIE) for esophageal cancer. *Best Pract Res Clin Gastroenterol*. 2018 ;36-37:81-83.
9. Pisla D et al Automated instrument for robotic assisted minimally invasive surgery. Patent pending No. A00142, State Office for Inventions and Trademarks, 29 March 2024.
10. Nottingham J M et al (2024) Transhiatal esophagectomy, In: StatPearls Treasure Island (FL); 2024 Available from: [www.ncbi.nlm.nih.gov/books/NBK559196/](http://www.ncbi.nlm.nih.gov/books/NBK559196/).
11. Harrington C et al (2021) Minimally invasive Ivor Lewis esophagectomy in 10 steps. *JTCVS Tech*. 2021 Aug 8;10:489-494.
12. Kono K et al (2023) *EndoLaparoscopic and Thoracoscopic Surgery*, Springer, Singapore.
13. Barron J et al (2023) Thoracoabdominal Esophagectomy: Then and Now, *The Annals of Thoracic Surgery*, In press, <https://doi.org/10.1016/j.athoracsur.2023.12.017>.
14. The da Vinci Xi robot, <https://www.intuitive.com/en-us/products-and-services/da-vinci> 2024/03/20.
15. Lu X et al (2022) A Flexible Surgical Instrument for Robot-Assisted Minimally Invasive Surgery. *Actuators* 2022, 11, 206.
16. Maddah M R et al (2023) A decision aid for the port placement problem in robot-assisted hysterectomy. *Laparoscopic, Endoscopic and Robotic Surgery*, 6(2), pp: 43-56.
17. Pisla D et al (2009) Kinematics and Design of two variants of a reconfigurable parallel robot. In *Proceedings of the 2009 ASME/IFTOMM International Conference on Reconfigurable Mechanisms and Robots*, London, UK, 22–24 June 2009; pp. 624–631.
18. Vaida C, Plitea N, Pisla D et al (2013) Orientation module for surgical instruments—a systematic approach. *Meccanica* 48, pp:145–158.
19. Nedezki C M (2021) Singularities study of the 3RTS manipulator, *Acta Technica Napocensis -Series: Applied Mathematics, Mechanics, and Engineering*, Vol. 64, Issue III, pp. 389-394.