



**HAL**  
open science

## **Robotic Assistance in Percutaneous Liver Ablation Therapies: A Systematic Review and Meta-Analysis**

Ana Karla Uribe Rivera, Barbara Seeliger, Laurent Goffin, Alain Garcia Vazquez, Didier Mutter, Mariano Eduardo Giménez

► **To cite this version:**

Ana Karla Uribe Rivera, Barbara Seeliger, Laurent Goffin, Alain Garcia Vazquez, Didier Mutter, et al.. Robotic Assistance in Percutaneous Liver Ablation Therapies: A Systematic Review and Meta-Analysis. *Annals of Surgery Open*, 2024, 10.1097/AS9.0000000000000406 . hal-04654742

**HAL Id: hal-04654742**

**<https://hal.science/hal-04654742v1>**

Submitted on 19 Jul 2024

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# Robotic Assistance in Percutaneous Liver Ablation Therapies

## A Systematic Review and Meta-Analysis

Ana K. Uribe Rivera, MD,\* Barbara Seeliger, MD, PhD,\*†‡§¶||¶¶ Laurent Goffin, MSc,§¶¶#  
Alain García-Vázquez, MD,\* Didier Mutter, MD, PhD,\*†‡ Mariano E. Giménez, MD, PhD\*†\*\*

**Objective:** The aim of this systematic review and meta-analysis is to identify current robotic assistance systems for percutaneous liver ablations, compare approaches, and determine how to achieve standardization of procedural concepts for optimized ablation outcomes.

**Background:** Image-guided surgical approaches are increasingly common. Assistance by navigation and robotic systems allows to optimize procedural accuracy, with the aim to consistently obtain adequate ablation volumes.

**Methods:** Several databases (PubMed/MEDLINE, ProQuest, Science Direct, Research Rabbit, and IEEE Xplore) were systematically searched for robotic preclinical and clinical percutaneous liver ablation studies, and relevant original manuscripts were included according to the Preferred Reporting Items for Systematic Reviews and Meta-Analyses guidelines. The endpoints were the type of device, insertion technique (freehand or robotic), planning, execution, and confirmation of the procedure. A meta-analysis was performed, including comparative studies of freehand and robotic techniques in terms of radiation dose, accuracy, and Euclidean error.

**Results:** The inclusion criteria were met by 33/755 studies. There were 24 robotic devices reported for percutaneous liver surgery. The most used were the MAXIO robot (8/33; 24.2%), Zerobot, and AcuBot (each 2/33, 6.1%). The most common tracking system was optical (25/33, 75.8%). In the meta-analysis, the robotic approach was superior to the freehand technique in terms of individual radiation (0.5582, 95% confidence interval [CI] = 0.0167–1.0996, dose-length product range 79–2216 mGy.cm), accuracy (0.6260, 95% CI = 0.1423–1.1097), and Euclidean error (0.8189, 95% CI = –0.1020 to 1.7399).

**Conclusions:** Robotic assistance in percutaneous ablation for liver tumors achieves superior results and reduces errors compared with manual applicator insertion. Standardization of concepts and reporting is necessary and suggested to facilitate the comparison of the different parameters used to measure liver ablation results. The increasing use of image-guided surgery has encouraged robotic assistance for percutaneous liver ablations. This systematic review analyzed 33 studies and identified 24 robotic devices, with optical tracking prevailing. The meta-analysis favored robotic assessment, showing increased accuracy and reduced errors compared with freehand technique, emphasizing the need for conceptual standardization.

**Keywords:** accuracy, Euclidean error, minimally invasive liver ablation, navigation systems, robotic percutaneous liver ablation

## INTRODUCTION

Image-guided techniques have revolutionized the treatment of liver tumors. Minimally invasive percutaneous applicator-based<sup>1</sup> procedures are increasingly common among interventional radiologists and surgeons. Ablation therapies have the advantages of repeatability and rapid recovery, they spare liver parenchyma and

are mainly used for tumors measuring less than 3 cm. Procedural success depends on accurate planning and precise applicator placement to ensure adequate coverage of the tumor volume, including for larger tumors. The postprocedural confirmation should show an ablated margin of peritumoral liver tissue.<sup>2–5</sup>

There are several ablation modalities, including radiofrequency ablation (RFA), microwave ablation (MWA), cryoablation (Cryo),

From the \*IHU-Strasbourg, Institute of Image-Guided Surgery, Strasbourg, France; †Department of Visceral and Digestive Surgery, University Hospitals of Strasbourg, Strasbourg, France; ‡IRCAD, Research Institute Against Digestive Cancer, Strasbourg, France; §ICube, UMR 7357 CNRS, INSERM U1328 RODIN, University of Strasbourg, Strasbourg, France; ¶Inserm U1110, Institute for Viral and Liver Diseases, Strasbourg, France; ¶Trustworthy AI Lab, Centre National de la Recherche Scientifique (CNRS), France; #Computational Surgery SAS, Schiltigheim, France; \*\*DAICIM Foundation (Training, Research and Clinical Activity in Minimally Invasive Surgery), Buenos Aires, Argentina.

**SDC** Supplemental digital content is available for this article. Direct URL citations appear in the printed text and are provided in the HTML and PDF versions of this article on the journal's Web site ([www.annalsofsurgery.com](http://www.annalsofsurgery.com)).

This work was supported by French state funds managed within the "Plan Investissements d'Avenir" and by the ANR (reference ANR-10-IAHU-02). The French "Fondation Force pour la recherche et l'innovation en santé" supports the ARPA project as part of its 2023 awards under reference 2023/5. The funders were not involved in the study design, data collection and analysis, and manuscript preparation for publication.

Disclosure: B.S. has a research and education consulting agreement with CMR Surgical and Intuitive Surgical. L.G. is the president of Computational Surgery

SAS. M.E.G. is a consultant to Siemens, Medtronic, Boston Scientific and AngioDynamics. The other authors declare that they have nothing to disclose

The data used to support the findings of this systematic review and meta-analysis are included in the article. Further information and reprints are available upon request. Please contact the corresponding author.

The research and analysis plan were not preregistered in any independent, institutional registry.

Reprints: Ana Karla Uribe Rivera, MD, Image-Guided Surgery, University of Strasbourg, IHU-Strasbourg, Institute of Image-Guided Surgery, 1, Place de l'Hôpital, 67000 Strasbourg, France. Email: [anakarla.uriberivera@ihu-strasbourg.eu](mailto:anakarla.uriberivera@ihu-strasbourg.eu).

Copyright © 2024 The Author(s). Published by Wolters Kluwer Health, Inc. This is an open-access article distributed under the terms of the Creative Commons Attribution-Non Commercial-No Derivatives License 4.0 (CCBY-NC-ND), where it is permissible to download and share the work provided it is properly cited. The work cannot be changed in any way or used commercially without permission from the journal.

Annals of Surgery Open (2024) 2:e406

Received: 29 September 2023; Accepted 19 February 2024

Published online 26 March 2024

DOI: 10.1097/AS9.0000000000000406

and irreversible electroporation (IRE). The most commonly used for liver tumors are RFA and MWA.<sup>2,6</sup> Conventional (freehand) applicator insertion techniques require a long learning curve, and the outcomes depend on the physician's expertise and experience.<sup>7</sup> Consequently, navigation and robotic assistance systems were developed to enhance the planning and execution of applicator placement to reduce the need for readjustments. Strategies to improve tumor targeting include the use of advanced imaging and rigid or elastic fusion of different imaging modalities, including two-dimensional (2D) ultrasound (US), multiplanar computed tomography (CT), magnetic resonance imaging (MRI), and three-dimensional (3D) image reconstructions. Depending on the tumor location, challenges arise that may require oblique (out-of-plane) access and trajectory readjustments to minimize targeting errors, which are categorized as longitudinal, lateral, angular, and Euclidean errors.<sup>8-10</sup>

When compared with manual applicator insertion, robotic assistance has the potential to optimize procedural accuracy, resulting in adequate ablation volumes and optimal oncologic outcomes, while reducing the training period to reach proficiency.<sup>3,8</sup> A classification of 6 levels of autonomy was proposed for medical robotics, from no autonomy (level 0) to full automation where the robot performs a procedure (level 5).<sup>11-13</sup> Current robotic surgical systems have either no autonomy, or correspond to manual control with robotic assistance (level 1) or operator-initiated task autonomy (level 2), such as automatic suturing.<sup>14</sup> Telemanipulation, mechanical guidance, and task autonomy are not yet at the level of automation but already provide valuable assistance.

The purpose of the present systematic review is to identify available robotic devices and key parameters for standardized reporting of approaches, ablation margins, and results. In the meta-analysis, robotic and freehand approaches are compared in terms of radiation dose, procedural accuracy, and Euclidean error.

## METHODS

### Eligibility Criteria

All articles in the medical/surgical and engineering literature reporting robotic assistance with or without the use of a navigation system for percutaneous liver ablation were considered for inclusion. Inclusion criteria were original research articles published in peer-reviewed journals in English, excluded were review articles, case reports, and conference abstracts, as well as articles on organs other than the liver. This systematic review and meta-analysis was completed following the Preferred Reporting Items for Systematic Reviews and Meta-Analyses (PRISMA) guidelines.<sup>15</sup> For the systematic review, original preclinical/clinical articles reporting the use of a percutaneous robotic approach for liver ablations were included. Inclusion in the meta-analysis required that articles report the comparison of robotic and freehand percutaneous liver ablation techniques and at least one of the following endpoints: dose-length product (DLP) in patients, accuracy, and Euclidean error.

### Search Strategy

A systematic literature search was conducted in the medical/surgical and engineering literature from database inception to January 18, 2023. Detailed search terms and databases are listed in Supplementary Table 1, <http://links.lww.com/AOSO/A299>. References of published studies were searched to ensure the identification of all relevant articles.

### Study Selection

Two authors (A.K.U.R. and B.S.) independently reviewed the eligibility of each publication by reviewing titles, abstracts,

and full text as specified in the PRISMA flowchart (Fig. 1). Discrepancies during title/abstract screening were resolved at the full-text stage by consensus between the two reviewers.

### Data Extraction and Risk of Bias Assessment

To address the key points for the systematic review and meta-analysis, the relevant technical and clinical parameters were extracted and grouped into tables to facilitate comparison among studies and standardization of definitions related to needle placement and accuracy. If reported, quantitative data regarding the number of targets, errors, and readjustments were extracted according to the technique and needle type used.

The risk of bias in individual studies was assessed using the QUADAS-2<sup>16</sup> tool, classifying each study into low, high, or unclear risk of bias, and applicability problems in 4 domains: patient selection, index test, reference standard, and flow and timelines. Risk categories have to be defined according to the target study. Consequently, in the first domain (patient selection), we classified as high risk the use of phantom models, ex vivo models, live models with insufficient details of the setting, and clinical case reports with inadequate information about the inclusion and exclusion criteria. Applicability problems in this domain were related to reproducibility in a clinical setting. For the second domain (index test: robotic equipment), a high risk was determined when device descriptions lacked adequate technical detail, tests were performed without clear methodology, prototype devices were used outside of clinical trials, and when extensive engineering support was required that could limit their applicability. The third domain (reference standard: freehand technique) was based on adequate specifications of the study results in the context of the current state of the art, and concerns about its applicability focused on assessing the relevance and utility of the results in the clinical setting. In the last domain (flow and timelines), the assessment considered technical issues and decision-making timelines that could impact the accuracy of the results. Supplementary Table 2, <http://links.lww.com/AOSO/A300> provides a structured and detailed assessment of these aspects for the included studies.

### Definitions

To foster agreement on terminology and thus standardization and comparability between studies, a summary of relevant nomenclature is proposed with key points in the description of robotic equipment and its function/benefit in the setting of liver ablations (Table 1).

### Meta-analysis

Subgroup meta-analysis was performed to compare the freehand and robotic techniques in terms of individual radiation (DLP), ablation accuracy, and Euclidean error, as well as for a comparison of studies that used the same device. The presence or absence of heterogeneity was assessed with the Q-statistic ( $\chi^2$ ) of homogeneity, and the extent of heterogeneity of effects among studies was quantified with the  $I^2$  index (an  $I^2 > 60\%$  representing high heterogeneity). Sensitivity analysis was performed by comparing the results of the subgroup analyses against the overall results and the random-effects analyses against the fixed-effects analyses. Egger's test was used to assess potential publication bias ( $P$  value  $< 0.05$  is the cutoff point related to the possible publication bias). The R 4.3.1 meta-package was used for all meta-analyses and graphics design. Statistical significance was set with a  $P$  value  $< 0.05$ .

## RESULTS

Among the 755 references retrieved, 33 preclinical and clinical trials met the inclusion criteria for the systematic review




FIGURE 1. PRISMA flow chart.<sup>15</sup>

and 14/33 for the meta-analysis. The PRISMA flowchart of the study selection process is presented in Figure 1, and the selected references are listed in Supplementary Table 3,<sup>23-29</sup> <http://links.lww.com/AOSO/A301>. An overview of the characteristics of different studies regarding the device used and the ablation workflow phases is listed in Supplementary Table 4, <http://links.lww.com/AOSO/A302>.

**Systematic Review**

The QUADAS-2 tool was used to assess all 33 articles included in the systematic review. It revealed varying degrees of risk of bias and concerns regarding applicability in the different domains, as listed in Supplementary Table 2, <http://links.lww.com/AOSO/A300>. The first domain (patient selection) showed 11 studies with low risk (33.3%), 21 with high risk (63.7%), and 1 with unclear risk (3%). The second domain (index test)

revealed 29 studies classified as low risk (87.9%) and 4 as high risk (12.1%). The third domain (reference standard) showed 27 studies with low risk (81.8%), 5 with high risk (15.2%), and 1 with unclear risk (3%). The fourth domain (flow and time) revealed 29 studies with low risk (87.9%), 3 with high risk (9.1%), and 1 with unclear risk (3%). In terms of concerns for clinical applicability, in patient selection, 21 studies showed a low risk of bias (63.6%), 9 high risk (27.3%), and 3 unclear risk (9.1%). In the index test, 27 studies showed low risk (81.8%) and 6 high risk (18.2%). In the reference standard, 28 studies showed low risk (84.9%), 4 high-risk (12.1%), and 1 unclear risk (3%). In 3 of the 4 risk of bias domains and in all domains relevant for clinical applicability, the majority of included studies presented a low risk.

There were 27 prospective and 6 retrospective studies included (81.8 and 18.2%, respectively). Among the 13 clinical trials involving 482 patients, 7 were led by interventional

**TABLE 1.**  
**Concept Summary Based on the Systematic Review and Meta-analysis**

Concept	Description	References
Applicator	Generic term for energy-based devices (needles, probes, antennas, etc.).	1
Planning	Refers to the set of imaging techniques used (such as US, CT, MRI, or PET-CT) to assess the suitability of the procedure for the patient. These imaging modalities provide critical information about the size, shape, location, and number of tumors in the tissue and their relationship to blood vessels and adjacent structures.	1
Execution		1,17
Targeting	The step in which an applicator is placed within the tumor to be treated. Image-guided targeting techniques (US, CT, among others) allow for precise localization and delineation of the tumor and surrounding anatomy. Clear visualization, multiplanar capability, and interactive image functions are necessary. Transtumoral: applicators pass through the tumor to reach the target point. Peritumoral: applicators are arranged around the tumor to cover the target area.	
Monitoring	The process of observing and evaluating the treatment effects during the procedure under different imaging techniques.	
Intraprocedural modification	Refers to the tools and techniques used during the procedure that allow to assess the need for real-time modification of the ablation site by repositioning the applicator until the appropriate safety margin is obtained. Repositioning: reentry of the applicator through the organ capsule Readjustment (s): Withdrawal and/or advancement of the applicator without exiting the organ capsule.	
Confirmation	Immediate evaluation of the treatment response using imaging studies upon completion of the procedure. This evaluation aims to assess the efficacy and confirm whether the final objective has been achieved, demonstrating that the ablation zone covers the target tumor, with an adequate safety halo.	1
Target plane		8,18
Axial	The target is in the same axial plane as the applicator insertion point.	
Oblique	The target is in a different axial plane than the applicator insertion point.	
Image fusion	Imaging technique that combines information from several images (US, CT, MRI, PET-CT) of the same scene into a single image that ideally contains all the important features of each original image (e.g., sensor coils, internal references, anatomical landmarks).	19,20,21
Rigid	Fusion of 2 data sets in a predefined region of interest. One of the images is used as a reference to the geometric transformation that is applied in the other images (source images) and is linked to static anatomic structures.	
Elastic	Can fuse 2 data sets accurately over. Thereby compensate for different patient positioning in the preoperative and the intraoperative setting, with a median error below 1.34 mm.	
Accuracy	Refers to how closely a set of measurements aligns with their true value. It is assessed by measuring the error between the predicted "optimal" applicator position and the "actual" final position of the applicator. Ideally, the error should be as close as possible to zero, which indicates that the current location of the applicator is aligned with the intended position.	10,22
Precision	Refers to the closeness of measurements to each other (consistency or reproducibility) at the same target site during repeated procedures under similar conditions. It measures the ability of the procedure to consistently achieve the same target with a high level of accuracy.	10
Error		8–10
Longitudinal	Depth error. The applicator is in the correct axis and has to be advanced or retracted on the same axis.	
Lateral	Lateral distance of the target measured at a 90° angle to the insertion line.	
Euclidean	Also known as total error, is the physical distance between the applicator tip and the target, in a 3D space.	
Angular	Deviation between planned trajectory and the applicator axis.	

3D indicates three-dimensional; CT, computed tomography; MRI, magnetic resonance imaging; PET-CT, positron emission tomography–computed tomography, US, ultrasonography.

radiologists (53.8%), 3 by surgeons (23.1%), and 3 were interdisciplinary (23.1%).

The technical parameters of the included studies are detailed in Supplementary Table 5, <http://links.lww.com/AOSO/A303>. According to the IDEAL framework,<sup>30</sup> the devices were in pre-clinical (stage 0; 17 studies; 51.5%), first-in-human (stage I; 12 studies; 36.4%), prospective developmental (stage II; 3 studies; 9.1%), or larger randomized controlled or equivalent (stage III; 1 study; 3%) stages, and none was a long-term monitoring and registry (stage IV). According to the 6 levels of autonomy for medical robotics,<sup>11</sup> the devices had either no autonomy (level 0; 30 studies; 90.9%) or provided robotic assistance during continuous control by the physicians (level 1; 2 studies, 6.1%) or task autonomy during discrete control by the physicians (level 2; 1 study; 3%). None of the reported devices had conditional or high autonomy or provided full automation. The tracking systems were mainly optical (25/33, 75.8%), electromagnetic (4/33, 12.1%), or not specified (4/33, 12.1%). Overall, the use of 24 robotic devices was reported. The most common were the MAXIO robot (8/33; 24.2%), Zerobot and AcuBot (each 2/33, 6.1%), and prototypes (13/33; 39.4%) and other robots (8/33, one different robot per study; 3% each).

The ablation workflow and clinical parameters of the included studies are detailed in Supplementary Table 6, <http://links.lww.com/AOSO/A304>. In all studies, rigid image fusion was used for planning and execution; the use of elastic fusion was not

reported. Imaging for planning and final confirmation was predominantly achieved via multiplanar CT scan (20/33; 60.6%, and 22/33; 66.7%, respectively). Image planning was performed either via multiplanar CT scan alone (20/33; 60.6%) or combined 3D-CT and 3D-CT with 2D-US (4/33; 12.1% each) and much less with 3D-US alone (2/33; 6.1%). The use of MRI and MRI-3D reconstructions was rare (1/33; 3% each). During the execution phase, multiple applicators were used in most trials (23/33; 69.7%). The applicator was inserted by physicians in 29/33 (87.9%) or by a robotic arm in 1/33 (3%); insertion was not specified in 3/33 (9.1%). The publication by Chang et al,<sup>31</sup> despite being categorized as a review, was included as the authors reported original data. The robot (Transcutaneous Robot-assisted Ablation-device Insertion Navigation System: TRAINS) was programmed to execute the applicator insertion in an ex vivo model.

By definition, the applicator trajectory is either axial, with the target in the same axial plane as the insertion point (in-plane), or oblique, with the target in a different axial plane than the insertion point (out-of-plane). Only 1 study reported the insertion angle relative to the CT scan acquisition planes.<sup>8</sup>

The data to assess overall ablation accuracy and accuracy relative to technique, model, targets, learning curve, and follow-up are summarized in Supplementary Table 7, <http://links.lww.com/AOSO/A305>. RFA was the most common ablation modality, either alone (14/33; 42.4%) or combined with MWA (2/33;



6.1%), or with MWA and Cryo (1/33; 3%) with a total of 4 different applicator types used.<sup>32</sup> To a lesser extent, MWA (7/33; 21.2%) or IRE (1/33; 3%) alone and not specified data (8/33; 24.2%) were reported.

Three studies compared the accuracy results with the robotic approach between experts and novices. The number of operators involved in the studies reporting operator experience ranged between 1 and 6. No statistically significant difference in the accuracy of ablations was found in relation to the physician's experience ( $P < 0.01$ ,<sup>33</sup>  $P = 0.41$ ,<sup>34</sup> and  $P = 0.44$ <sup>35</sup>).

A total of 6 studies reported follow-up data, 3 with 6-week follow-up and 3 with 6-month follow-up. Five studies focused on lesion outcomes, while 1 evaluated lesion and patient outcomes, indicating an overall survival of 90% and a disease-free survival of 83.3%. Four studies reported technical success rates (76–100%) for the freehand and robotic approaches (88–100%). These success rates were considered the “primary efficacy of the technique” as determined by CT and MRI studies.

### Meta-analysis

The meta-analysis included 14 comparative studies of freehand and robotic techniques and assessed individual radiation, accuracy, and Euclidean error. For each study, the sample size ( $n$ ), effect size ( $d$ ), and 95% confidence interval (95% CI) are presented in Table 2 with the corresponding forest plots expressed in standard deviation units. For a subgroup meta-analysis of studies using the same commercially available device, the MAXIO was chosen as the most reported one (Table 3).

### Dose-length product

Nine out of the 14 included studies assessed the individual radiation measured as DLP, the ranges were 68–7025 versus 79–2216 mGy.cm for freehand and robotic evaluation, respectively. The DerSimonian and Laird procedure yielded a Q-statistic ( $\chi^2$ ) value of 55.5155,  $P < 0.05$ , and  $I^2$  of 85.58%, demonstrating a significant heterogeneity. The coefficient of variation of 1.9626 indicated a high variability between the studies. Consequently, the random pooled effects model was used and showed a random effect of 0.5582 in favor of the robotic technique (95% CI = 0.0167–1.0996). Egger's test showed no evidence of publication bias ( $P = 0.1666$ ).

### Accuracy

Eight studies reported the accuracy, comparing the techniques. There was statistical evidence for significant heterogeneity (Q-statistic ( $\chi^2$ ) value of 33.3511,  $P < 0.05$ , and  $I^2$  of 79.01%) and high variability between studies (coefficient of variation of 0.9061). A significant and positive effect on accuracy was found overall for the robotic technique (0.6260, 95% CI = 0.1423–1.1097). Egger's test indicated no evidence of publication bias ( $P = 0.7007$ ).

In the subgroup analysis of the 5 studies reporting the accuracy when using the MAXIO device and comparing the robotic and freehand techniques, the random-effects model demonstrated a significant and positive effect on overall accuracy in favor of the robotic technique (1.60, 95% CI = 0.42–2.77) (Table 3).

### Euclidean Error

Five studies reported the Euclidean error when comparing both techniques. There was high heterogeneity among studies (Q-statistic ( $\chi^2$ ) value of 51.3554,  $P < 0.05$ , and  $I^2$  of 92.21%), and high variability was identified between studies (coefficient of variation of 1.4541). The combined positive effect of the robotic

approach was statistically significant; the random pooled effects model was used and showed an effect of 0.8189 (95% CI = –0.1020 to 1.7399). Egger's test indicated no evidence of publication bias ( $P = 0.4115$ ).

## DISCUSSION

The systematic review and meta-analysis has revealed evidence that robotic-assisted percutaneous liver ablation is superior to the freehand approach in terms of individual radiation (such as DLP), accuracy, and Euclidean error. However, these results should be interpreted with caution, as heterogeneity between studies leads to greater uncertainty regarding the magnitude and direction of the observed effects. Despite the superiority of robotic assistance for these technical intraprocedural criteria, follow-up is crucial to assess the efficacy of each technique. Although the number of studies reporting follow-up is limited, the clinical results are also in favor of robotic assistance. Furthermore, robotic assistance is promising to lead the way for future therapeutic strategies by enhancing the learning curve aiming at reduced inter-operator variability and increased accuracy of image-guided liver ablations.

In one systematic review and meta-analysis of liver ablation techniques, 34 studies were included that used laparoscopic and/or percutaneous approaches.<sup>9</sup> The comparison of their efficacy with the manual technique demonstrated a significant improvement in treatment accuracy with stereotactic and/or robotic guidance, achieving 94% efficacy and low complication rates. It is important to emphasize that the study pooled both minimally invasive techniques, including 6 studies with robotic guidance, whereas our analysis focuses exclusively on the comparison between the robotic approach and the manual technique.

To facilitate agreement on terminology, standardization, and comparability between studies, a summary of the relevant nomenclature is proposed in Table 1. Overall, the number of studies comparing both approaches was low, and further comparative studies are anticipated. Among the 33 studies included in the systematic review, the most frequently reported device was the MAXIO system, which has stereotactic spatial positioning and includes software that registers current images with preoperative images; visualizes and edits estimated ablation volumes, includes a respiratory gating system, and verifies applicator placement by adapting the procedure according to the image recordings.<sup>48</sup>

Many of the devices used were in the prototype stage (13/33; 39.4%), with a majority of studies (29/33; 87.9%) reporting devices in the IDEAL innovation stages 0 (preclinical) and I (first-in-human).<sup>30</sup> In this context, it comes as no surprise that most of the 24 different devices used had no autonomy (90.9% of studies) or low autonomy with continuous or discrete physician control. Another key point of the present study is to provide definitions and concepts focusing exclusively on the robotic approach.

In a minimally invasive setting, precise placement of an ablation applicator is challenging. This includes involuntary operator movements,<sup>49,50</sup> breathing movements,<sup>18</sup> complex lesion management (e.g., small lesions, deep location, poor visibility, proximity to adjacent organs, and anatomical changes in the liver), and imaging studies for guidance, such as CT or MRI. Although MRI shows greater precision and target orientation, its use is limited by the nonferromagnetic materials or sensors required.<sup>51</sup>

The applicator insertion techniques can be conventional (freehand technique, most widely used) and hybrid (navigation guidance system with or without a robotic arm). Most complex cases require experience to improve accuracy,<sup>9,52</sup> the freehand technique may increase the number of readjustments or repositionings and thus complication rates.<sup>53</sup> The aim of local control is to completely ablate the target with a sufficient safety halo (distance

**TABLE 2.**


**s-analysis and Forest Plot Comparison of Freehand versus Robotic Targeting for Individual Radiation (Dose-length Product), Accuracy and Euclidean Error**

	Sample Size (n)	Effect Size (d)	CI (95%)
<b>DLP</b>			
Patriciu et al, 2005 <sup>36</sup>	14	-0.4406	-1.1018 0.2206
Abdullah et al, 2013 <sup>37</sup>	47	0.3722	-0.1052 0.8496
Abdullah et al, 2014 <sup>38</sup>	70	2.9318	1.4229 4.4407
Mbalisike et al, 2014 <sup>39</sup>	70	2.0531	1.2866 2.8195
Beyer et al, 2015 <sup>40</sup>	64	0.5402	0.0404 1.0400
Cornelis et al, 2015 <sup>41</sup>	48	0.7828	0.1672 1.3985
Beyer et al, 2016 <sup>42</sup>	40	-0.8075	-1.4365 -0.1784
Hiraki et al, 2017 <sup>43</sup>	36	-0.0312	0.5046 0.4422
Heerink et al, 2019 <sup>9</sup>	42	0.7679	0.1816 1.3542
Random effect	431	0.5582	0.0167 1.0996

Heterogeneity test ( $\chi^2$ ) = 55.5155;  $P < 0.05$ .

$I^2$  = 85.58%.

Egger test = 1.5438;  $P = 0.1666$ .




	n	d	CI (95%)
<b>Accuracy</b>			
Boctor et al, 2008 <sup>44</sup>	5	-0.3162	-0.9736 0.3412
Koethe et al, 2014 <sup>3</sup>	40	1.3849	0.8589 1.9110
Mbalisike et al, 2014 <sup>39</sup>	70	-0.3157	-2.1155 1.4842
Beyer et al, 2015 <sup>40</sup>	64	0.4012	-0.0460 0.8484
Cornelis et al, 2015 <sup>41</sup>	48	1.1740	0.6426 1.7054
Beyer et al, 2016 <sup>45</sup>	40	-0.1131	-0.6794 0.4531
Hiraki et al, 2017 <sup>43</sup>	36	0.8187	0.1727 1.4647
Guiu et al, 2021 <sup>35</sup>	79	1.3796	0.6900 2.0691
Random effect	382	0.6260	0.1423 1.1097

Heterogeneity test ( $\chi^2$ ) = 33.3511;  $P < 0.05$ .

$I^2$  = 79.01%.

Egger test = -0.4033;  $P = 0.7007$ .




	n	d	CI (95%)
<b>Euclidean error</b>			
Wallach et al, 2013 <sup>46</sup>	50	0.4012	-0.0406 0.8484
Mbalisike et al, 2014 <sup>39</sup>	70	2.8445	2.1766 3.5124
Heerink et al, 2019 <sup>9</sup>	42	-0.1219	-0.7423 0.4984
Levin et al, 2020 <sup>47</sup>	40	0.1982	-0.3575 0.7540
Guiu et al, 2021 <sup>35</sup>	79	0.8313	0.2009 1.4618
Random effect	281	0.8189	-0.1020 1.7399

Heterogeneity test ( $\chi^2$ ) = 51.3554;  $P < 0.05$ .

$I^2$  = 92.21%.

Egger test = 0.9515;  $P = 0.4115$ .



between the tumor margin and the ablation volume surface). To compensate for the inaccuracy of the ablation applicator (which differs between 5 and 10mm)<sup>1,54</sup> and the irregular shape of the tumors, the halo should measure 10mm of safety margin during

ablation; in contrast, during surgical resection, a safety margin of 1 mm is adequate in the absence of these limitations.

With tracking systems (optical, electromagnetic, and laser),<sup>2,3,55</sup> the information on location-position-orientation of

**TABLE 3.** Subgroup Meta-analysis and Forest Plot Comparison of Freehand versus Robotic Targeting for Accuracy in Advance-stage Device (MAXIO)

Study	Freehand			Robotic (MAXIO)			Mean Difference	95% CI	Mean Difference (MD)
	Total	Mean	SD	Total	Mean	SD			
				20	6.50	2.5000	9.30	(5.12–13.48)	
Koethe et al, 2014 <sup>3</sup>	20	15.80	-1.1018						
Mbalisike et al, 2014 <sup>39</sup>	40	4.10	-0.1052	30	2.10	0.7000	2.00	(1.39–2.61)	
Beyer et al, 2015 <sup>40</sup>	30	3.30	1.4229	34	1.60	1.3000	1.70	(0.98–2.42)	
Cornelis et al, 2015 <sup>41</sup>	24	4.50	0.1672	24	4.70	1.1000	-0.20	(-0.85–0.45)	
Beyer et al, 2016 <sup>45</sup>	19	3.10	-1.4365	21	2.20	1.0000	0.90	(0.21–1.59)	
Random effect							1.60	(0.42–2.77)	

Heterogeneity test ( $\chi^2$ ) = 1.4194;  $P < 0.05$ .  
 $I^2 = 90\%$ .

applicator-placement increases the accuracy of target management, and robotic navigation systems have increased the technical success<sup>9,55,56</sup> in association with ablation modalities (RFA, MWA, Cryo, and IRE) in the liver, reducing morbidity and mortality when compared with resection.<sup>57</sup> Furthermore, elastic image fusion creates an image of the organ, updating itself according to the data acquired in real-time, improving the evaluation, interpretation, and accuracy of the images. However, as shown in the present analysis, elastic image fusion was not used in any of the studies, and its worldwide diffusion is limited, which is probably related to the high costs of software development and the requirement of hybrid operating rooms.<sup>19,20</sup>

The “accuracy” is determined by the length of a special vector between the center of the target and the tip of the needle, using X-Y-Z coordinates, relevant for the different error types (Table 1, Figure 2).<sup>8,18,43,46</sup> The most important criterion for clinical success of tumor treatment with ablation is complete tissue destruction with a safety margin halo, a concept that goes beyond the multiple definitions described in the literature (conceived as “complete ablation without residual tumor”,<sup>40</sup> “primary efficacy”,<sup>54</sup> “treatment success”,<sup>45</sup> and “hit rate”,<sup>44</sup>) associated to image control.<sup>40,44,45,54</sup> However, this binary criterion lacks precision and does not allow the identification of weaknesses in applicator positioning systems. To gain insight into failed cases, it is useful to assess the percentage of tumor inclusion within the ablation zone. Clinicians can segment targets and the ablation zone, while specific software can calculate intersections and volumes. Furthermore, it is important to differentiate the tumor volume in the central ablation zone, the margin, and the outer area as 3 relevant criteria for the final result. One publication specifies the concepts of ablation coverage and overablation,<sup>4</sup> based on the ablation volume in relation to the tumor volume. Finally, secondary criteria are essential to evaluate the procedure in general: the number of adjustments (readjustments,<sup>37</sup> repositionings,<sup>8</sup> and invasiveness<sup>44</sup>), the number of applicators used, and the intervention time in relation to tumor parameters. Geometric criteria are no longer based on the actual ablation volume, but on the geometric position of the applicator relative to the target.<sup>8</sup> These criteria are often proposed in the literature and consist of measuring and calculating distances and residual angles. However, they must be performed on medical control images, and navigation systems alone do not guarantee a sufficient correlation with the quality of the final result.


Estimating the distance from the applicator tip to the tumor barycenter (Euclidean distance,<sup>8,52</sup> applicator deflection<sup>39</sup>) could

be equivalent to the volumetric criterion if the 2 volumes were isotropic (spherical), but this is not the case. Geometric methods do not consider the delivered power or the environment that influences the shape and size of the ablation zone. To be comparable, it would be necessary to contemplate the sphere enclosing the tumor and the minimum sphere enclosed in the ablation zone, which is very conservative and would minimize real effectiveness, but in the other case, the assessment of real results would be too optimistic.

Another criterion is the alignment of the applicator with the axis from the entry point to the center of the target. Ideally, this axis should be aligned with the longitudinal axis of the ellipsoid encompassing the tumor, so this criterion indicates the ability of the system to follow the ideal plan. It is used as an indicator of the overall ability of the system (mechanical, software, and human) to perform the procedure correctly; however, on its own, it is insufficient to assess the quality of the result due to possible lateral and longitudinal errors. Nevertheless, drawing reasonable conclusions from these criteria or creating a single score is challenging due to differences in units and mathematical relations between them. Thus, the ultimate oncologic effectiveness of tumor ablation and the decrease in the ablation margin of normal parenchyma around the tumor will be related to addressing these considerations, with the aim of achieving greater accuracy in real-world applications and improving oncologic outcomes, while minimizing the need for overablation.

In this systematic review, a few individual studies reported important concepts such as applicator positioning accuracy. One study<sup>58</sup> showed a mean accuracy of 3.5 mm in an in-vivo model with robotically inserted applicators, while others reported a range of 2.7–10.2 mm,<sup>18,22,32,35</sup> indicating that an acceptable target error should be less than 5 mm in a 10-mm lesion. Applicator specifications (quantity, angles, and readjustments) were explored in several studies. In a randomized controlled trial<sup>8</sup> the accuracy of freehand versus robotic technique was compared, and robotic guidance eliminated the repositioning need. In addition, for oblique (out-of-plane) targets, lateral accuracy improved from 16.1 to 5.6 mm. A study<sup>32</sup> comparing the applicator type energy device (RFA-MWA-Cryo) found a significant difference in target motion during insertion, and accuracy varied between the type and sharpness of the applicator. Several factors can influence the accuracy, such as procedural models, where insertion of the applicator into the target may move in in-vivo models, caused by several factors, like





**FIGURE 2.** Classification of errors and their illustration. Red dot: Indicates the target needle tip position from which the deviating distances and angles are measured. Target needle tip: The position is not in the center of the lesion, as its location depends on the size of the lesion, the number of applicators used, the ablation area, and the desired safety halo to obtain a complete ablation. A, Longitudinal error, measured in millimeters. B, Lateral error, measured in millimeters. C, Euclidean error, measured in millimeters. D, Angular error, measured in degrees. This image represents a 2D environment; for oblique/out-of-plane targets, a 3D error measure is required.

respiratory movement, displacement, and deformation of the tissue during puncture. Precision can be improved when these procedures are performed under general anesthesia, allowing respiratory motion to be controlled by several techniques (temporary disconnection of the endotracheal tube and high-frequency jet ventilation).<sup>43,55,59</sup>

Error concepts<sup>8,9</sup> are fundamental to understanding the 3D space in which the targets are located. Regarding radiation exposure, most of the procedures were performed under CT guidance. One study,<sup>22</sup> compared the need for confirmatory CT scans after applicator placement between freehand (requiring 6–7 scans) and robotic assistance (requiring 1–2 scans).

The findings demonstrated the advantages of robotics over the freehand technique, as synthesized by the present meta-analysis.

Currently, percutaneous robotic approaches and navigation systems are based on optical tracking. EM tracking uses magnetic fields, with promising results in various surgical fields, such as flexible and ultrasound-guided endoscopy. However, its accuracy in percutaneous robotic arms has not yet been investigated. Advantages over optical tracking features (e.g., not susceptible to line-of-sight obstructions, and simultaneous tracking of multiple applicators) would improve surgical workflow and efficiency, especially for the multiapplicator approach. In training programs, the neurosurgical<sup>58</sup> domain has shown that the use of robotic devices with navigation considerably shortened the learning curve, and further studies are required to quantify the training effect in liver ablation therapies.

Although robotic percutaneous ablation is an emerging technique, the current data are in favor of the robotic approach. While this overall analysis of the literature is limited by still low cohort sizes and heterogeneity of study populations, more randomized comparative studies will provide stronger evidence. When considering that most of the devices used are prototypes and studies focused on the feasibility of their use, further studies on accuracy and recognition of the error types and models used as influencing factors are eagerly awaited.

The percentage of studies classified as high risk in the domains of patient selection and index test underscores potential limitations that may affect the validity of the results.

Meanwhile, studies with low risk in the domains of reference standard and flow and timing suggest a more robust methodological approach in these aspects. Thus, applicability issues have to be anticipated in future study designs to ensure reproducibility in larger clinical cohorts and before potentially changing the current gold standard in favor of robotic assistance.

In conclusion, percutaneous robotic approaches enable a more precise management of liver tumors. In a minimally invasive setting, they decrease errors when compared with traditional techniques. Standardization of concepts and error reporting are necessary to ensure the comparability of results obtained with these systems. Error types are influenced by the study model (phantom, ex vivo or in-vivo model, static or dynamic model); the number and type of applicators, location of the target(s) (axial/in-plane or oblique/out-of-plane), the organ (respiratory motion of the liver), and the anesthesiologic management.

As a large number of devices are in a prototype stage, the present systematic review and meta-analysis can provide guidance for key points to address to enhance measurements and comparability of technical and future clinical results.

## ACKNOWLEDGMENTS

The authors are grateful to Dr. Cristian Díaz Vélez and Mickael Schaeffer for their assistance with the statistical analysis.

## Authors contribution

Conceptualization, formal analysis and writing—original draft preparation: A.K.U.R., B.S., M.E.G.; methodology A.K.U.R., B.S., L.G., M.E.G.; data curation and investigation: A.K.U.R., B.S., A.G.V.; writing—review and editing: A.K.U.R., B.S., L.G., A.G.V., D.M., M.E.G.; visualization: A.K.U.R., B.S.; supervision: D.M., M.E.G.; funding acquisition: A.K.U.R., B.S., D.M., M.E.G. All authors have read and agreed to the final version of the article.

## REFERENCES

- Ahmed M, Solbiati L, Brace CL, et al; International Working Group on Image-guided Tumor Ablation. Image- Robotic assisted radio-frequency ablation of liver tumors--randomized patient study Robotic assisted radio-frequency ablation of liver tumors--randomized patient study Robotic assisted radio-frequency ablation of liver tumors--randomized patient study guided tumor ablation: standardization of terminology and reporting criteria—a 10-year update. *Radiology*. 2014;273:241–260.
- de Baère T, Roux C, Deschamps F, et al. Evaluation of a new CT-guided robotic system for percutaneous needle insertion for thermal ablation of liver tumors: a prospective pilot study. *Cardiovasc Intervent Radiol*. 2022;45:1701–1709.
- Koethe Y, Xu S, Velusamy G, et al. Accuracy and efficacy of percutaneous biopsy and ablation using robotic assistance under computed tomography guidance: a phantom study. *Eur Radiol*. 2014;24:723–730.
- Liu P, Qin J, Duan B, et al. Overlapping radiofrequency ablation planning and robot-assisted needle insertion for large liver tumors. *Int J Med Robot*. 2019;15:e1952.
- Hendriks P, van Dijk KM, Boekestijn B, et al. Intraoperative assessment of ablation margins using computed tomography co-registration in hepatocellular carcinoma treatment with percutaneous ablation: IAMCOMPLETE study. *Diagn Intervent Imaging*. 2023;105:57–64.
- Frühling P, Seeliger B, Uribe Rivera AK, Freedman J, Gimenez ME. *Image-guided ablation for liver tumours – an addition to the armamentarium of multidisciplinary oncological and surgical approaches*. BJS Academy. 2023. Accessed July 31, 2023.
- Hildebrand P, Leibecke T, Kleemann M, et al. Influence of operator experience in radiofrequency ablation of malignant liver tumours on treatment outcome. *Eur J Surg Oncol*. 2006;32:430–434.
- Heerink WJ, Ruiters SJS, Pennings JP, et al. Robotic versus freehand needle positioning in CT-guided ablation of liver tumors: a randomized controlled trial. *Radiology*. 2019;290:826–832.
- Tinguely P, Paolucci I, Ruiters SJS, et al. Stereotactic and robotic minimally invasive thermal ablation of malignant liver tumors: a systematic review and meta-analysis. *Front Oncol*. 2021;11:713685.
- Accuracy, Precision, and Error. Oncology Medical Physics. Available at: <https://oncologymedicalphysics.com/quantifying-accuracy-precision-and-error/>. Accessed March 27, 2023.
- Yang G-Z, Cambias J, Cleary K, et al. Medical robotics—regulatory, ethical, and legal considerations for increasing levels of autonomy. *Sci Robot*. 2017;2:eaam8638.
- Attanasio A, Scaglioni B, De Momi E, et al. Autonomy in surgical robotics. *Annu Rev Control Robot Auton Syst*. 2021;4:651–679.
- Battaglia E, Boehm J, Zheng Y, et al. Rethinking autonomous surgery: focusing on enhancement over autonomy. *Eur Urol Focus*. 2021;7:696–705.
- Fiorini P, Goldberg KY, Liu Y, et al. Concepts and trends in autonomy for robot-assisted surgery. *Proc IEEE Inst Electr Electron Eng*. 2022;110:993–1011.
- Page MJ, McKenzie JE, Bossuyt PM, et al. The PRISMA 2020 statement: an updated guideline for reporting systematic reviews. *Rev Esp Cardiol (Engl Ed)*. 2021;74:790–799.
- Whiting PF, Rutjes AWS, Westwood ME, et al; QUADAS-2 Group. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. *Ann Intern Med*. 2011;155:529–536.
- Morikawa S, Naka S, Murakami K, et al. Preliminary clinical experiences of a motorized manipulator for magnetic resonance image-guided microwave coagulation therapy of liver tumors. *Am J Surg*. 2009;198:340–347.
- Engstrand J, Toporek G, Harbut P, et al. Stereotactic CT-Guided percutaneous microwave ablation of liver tumors with the use of high-frequency jet ventilation: an accuracy and procedural safety study. *AJR Am J Roentgenol*. 2017;208:193–200.
- Peterlík I, Courtecuisse H, Rohling R, et al. Fast elastic registration of soft tissues under large deformations. *Med Image Anal*. 2018;45:24–40.
- Negwer C, Hiepe P, Meyer B, et al. Elastic fusion enables fusion of intraoperative magnetic resonance imaging data with preoperative neuronavigation data. *World Neurosurg*. 2020;142:e223–e228.
- Schmidt FA, Mullally M, Lohmann M, et al. Elastic image fusion software to coregister preoperatively planned pedicle screws with intraoperative computed tomography data for image-guided spinal surgery. *Int J Spine Surg*. 2021;15:295–301.
- Wiles AD, Thompson DG, Frantz DD. Accuracy assessment and interpretation for optical tracking systems. In: *Medical Imaging 2004: Visualization, Image-Guided Procedures, and Display*. SPIE:421–432.
- Xu J, Jia Z, Song Z, et al. Three-dimensional ultrasound image-guided robotic system for accurate microwave coagulation of malignant liver tumours. *Int J Med Robot*. 2010;6:256–268.
- Yamada A, Tokuda J, Naka S, et al. Magnetic resonance and ultrasound image-guided navigation system using a needle manipulator. *Med Phys*. 2020;47:850–858.
- Won HJ, Kim N, Kim GB, et al. Validation of a CT-guided intervention robot for biopsy and radiofrequency ablation: experimental study with an abdominal phantom. *Diagn Interv Radiol*. 2017;23:233–237.
- Solomon SB, Patriciu A, Stoianovici DS. Tumor ablation treatment planning coupled to robotic implementation: a feasibility study. *J Vasc Interv Radiol*. 2006;17:903–907.
- Wen R, Tay W-L, Nguyen BP, et al. Hand gesture guided robot-assisted surgery based on a direct augmented reality interface. *Comput Methods Programs Biomed*. 2014;116:68–80.
- Liu S, Xia Z, Liu J, et al. Automatic multiple-needle surgical planning of robotic-assisted microwave coagulation in large liver tumor therapy. *PLoS One*. 2016;11:e0149482.
- Krücker J, Xu S, Glossop N, et al. Electromagnetic tracking for thermal ablation and biopsy guidance: clinical evaluation of spatial accuracy. *J Vasc Interv Radiol*. 2007;18:1141–1150.
- Marcus HJ, Bennett A, Chari A, et al. IDEAL-D framework for device innovation: a consensus statement on the preclinical stage. *Ann Surg*. 2022;275:73–79.
- Chang SK, Hlaing WW, Yang L, et al. Current technology in navigation and robotics for liver tumours ablation. *Ann Acad Med Singap*. 2011;40:231–236.
- Hiraki T, Matsuno T, Kamegawa T, et al. Robotic insertion of various ablation needles under computed tomography guidance: accuracy in animal experiments. *Eur J Radiol*. 2018;105:162–167.
- Ben-David E, Shochat M, Roth I, et al. Evaluation of a CT-guided robotic system for precise percutaneous needle insertion. *J Vasc Interv Radiol*. 2018;29:1440–1446.
- Kägebein U, Godenschweger F, Armstrong BSR, et al. Percutaneous MR-guided interventions using an optical Moiré phase tracking system: initial results. *PLoS One*. 2018;13:e0205394.
- Guiu B, De Baère T, Noel G, et al. Feasibility, safety and accuracy of a CT-guided robotic assistance for percutaneous needle placement in a swine liver model. *Sci Rep*. 2021;11:5218.
- Patriciu A, Awad M, Solomon SB, et al. Robotic assisted radio-frequency ablation of liver tumors--randomized patient study. *Med Image Comput Assist Interv*. 2005;8:526–533.
- Abdullah BJJ, Yeong CH, Goh KL, et al. Robot-assisted radiofrequency ablation of primary and secondary liver tumours: early experience. *Eur Radiol*. 2014;24:79–85.
- Abdullah BJJ, Yeong CH, Goh KL, et al. Robotic-assisted thermal ablation of liver tumours. *Eur Radiol*. 2015;25:246–257.
- Mbalisike EC, Vogl TJ, Zangos S, et al. Image-guided microwave thermoablation of hepatic tumours using novel robotic guidance: an early experience. *Eur Radiol*. 2015;25:454–462.
- Beyer LP, Pregler B, Niessen C, et al. Robot-assisted microwave thermoablation of liver tumors: a single-center experience. *Int J Comput Assist Radiol Surg*. 2016;11:253–259.
- Cornelis F, Takaki H, Laskhmanan M, et al. Comparison of CT fluoroscopy-guided manual and CT-guided robotic positioning system for in vivo needle placements in swine liver. *Cardiovasc Intervent Radiol*. 2015;38:1252–1260.
- Beyer LP, Pregler B, Michalik K, et al. Evaluation of a robotic system for irreversible electroporation (IRE) of malignant liver tumors: initial results. *Int J Comput Assist Radiol Surg*. 2017;12:803–809.
- Hiraki T, Kamegawa T, Matsuno T, et al. Robotically driven CT-guided needle insertion: preliminary results in phantom and animal experiments. *Radiology*. 2017;285:454–461.
- Boctor EM, Choti MA, Burdette EC, et al. Three-dimensional ultrasound-guided robotic needle placement: an experimental evaluation. *Int J Med Robot*. 2008;4:180–191.
- Beyer LP, Pregler B, Nießen C, et al. Stereotactically-navigated percutaneous irreversible electroporation (IRE) compared to conventional IRE: a prospective trial. *PeerJ*. 2016;4:e2277.
- Wallach D, Toporek G, Weber S, et al. Comparison of freehand-navigated and aiming device-navigated targeting of liver lesions. *Int J Med Robot*. 2014;10:35–43.
- Levin AA, Klimov DD, Nechunava AA, et al. The comparison of the process of manual and robotic positioning of the electrode performing radiofrequency ablation under the control of a surgical navigation system. *Sci Rep*. 2020;10:8612.

48. *PERFINT HEALTHCARE*. Available at: [http://www.perfinthealthcare.com/maxio\\_overview.php](http://www.perfinthealthcare.com/maxio_overview.php). Accessed August 4, 2023.
49. Ginoya T, Maddahi Y, Zareinia K. A historical review of medical robotic platforms. *Journal of Robotics*. 2021;2021:1–13.
50. Kettenbach J, Kronreif G. Robotic systems for percutaneous needle-guided interventions. *Minim Invasive Ther Allied Technol*. 2015;24:45–53.
51. Franco E, Ristic M, Rea M, et al. Robot-assistant for MRI-guided liver ablation: a pilot study. *Med Phys*. 2016;43:5347.
52. Scharll Y, Mitteregger A, Laimer G, et al. Comparison of a robotic and patient-mounted device for CT-guided needle placement: a phantom study. *J Clin Med*. 2022;11:3746.
53. Ciocan A, Elisei R, Graur F, et al. *Robot-Assisted Ablation of Liver Hepatocellular Carcinoma and Colorectal Metastases: A Systematic Review*. 2020:206–218.
54. Schaible J, Pregler B, Verloh N, et al. Improvement of the primary efficacy of microwave ablation of malignant liver tumors by using a robotic navigation system. *Radiol Oncol*. 2020;54:295–300.
55. Beyer LP, Wiggermann P. Planning and guidance: new tools to enhance the human skills in interventional oncology. *Diagn Interv Imaging*. 2017;98:583–588.
56. Bramhe S, Pathak SS. Robotic surgery: a narrative review. *Cureus*. 2022;14:e29179.
57. *Surgical approach to microwave and radiofrequency liver ablation or hepatocellular carcinoma and colorectal liver metastases less than 5 cm: a systematic review and meta-analysis - A SAGES Publication*. SAGES. Available at: <https://www.sages.org/publications/guidelines/surgical-approach-to-microwave-and-radiofrequency-liver-ablation/>. Accessed January 20, 2023.
58. Fong AJ, Stewart CL, Lafaro K, et al. Robotic assistance for quick and accurate image-guided needle placement. *Updates Surg*. 2021;73:1197–1201.
59. Musa MJ, Sharma K, Cleary K, et al. Respiratory compensated robot for liver cancer treatment: design, fabrication, and benchtop characterization. *IEEE/ASME Trans Mechatron*. 2022;27:268–279.