

An interlaboratory comparison to quantify oxidative potential measurement in aerosol particles: challenges and recommendations for harmonisation

Pamela A Dominutti, Jean-Luc Jaffrezo, Anouk Marsal, Takoua Mhadhbi, Rhabira Elazzouzi, Camille Rak, Fabrizia Cavalli, Jean-Philippe Putaud, Aikaterini Bougiatioti, Nikolaos Mihalopoulos, et al.

▶ To cite this version:

Pamela A Dominutti, Jean-Luc Jaffrezo, Anouk Marsal, Takoua Mhadhbi, Rhabira Elazzouzi, et al.. An interlaboratory comparison to quantify oxidative potential measurement in aerosol particles: challenges and recommendations for harmonisation. 2024. hal-04661735

HAL Id: hal-04661735 https://hal.science/hal-04661735v1

Preprint submitted on 25 Jul 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.





An interlaboratory comparison to quantify oxidative potential measurement in aerosol particles: challenges and recommendations for harmonisation

4 5

6 Pamela A. Dominutti¹, Jean-Luc Jaffrezo¹, Anouk Marsal¹, Takoua Mhadhbi¹, Rhabira Elazzouzi¹,
7 Camille Rak¹, Fabrizia Cavalli², Jean-Philippe Putaud², Aikaterini Bougiatioti³, Nikolaos
8 Mihalopoulos^{3,4}, Despina Paraskevopoulou³, Ian Mudway⁵, Athanasios Nenes^{6§}, Kaspar R.
9 Daellenbach⁷, Catherine Banach⁸, Steven J. Campbell⁹, Hana Cigánková¹⁰, Daniele Contini¹¹, Greg
10 Evans¹², Maria Georgopoulou⁶, Manuella Ghanem¹³, Drew A. Glencross⁵, Maria Rachele Guascito^{11,14},
11 Hartmut Herrmann¹⁵, Saima Iram¹⁶, Maja Jovanović¹⁷, Milena Jovašević-Stojanoić¹⁷, Markus
12 Kalberer¹⁸, Ingeborg M. Kooter^{19†}, Suzanne E. Paulson⁸, Anil Patel^{20#*}, Esperanza Perdrix²¹, Maria
13 Chiara Pietrogrande²², Pavel Mikuška¹⁰, Jean-Jacques Sauvain²³, Aikaterina Seitanidi⁶, Pourya
14 Shahpoury²⁴, Eduardo J. dos S. Souza¹⁵, Sarah Steimer^{20#}, Svetlana Stevanovic¹⁶, Guillaume Suarez²³,
15 P. S. Ganesh Subramanian²⁵, Battist Utinger¹⁸, Marloes F. van Os¹⁹, Vishal Verma²⁵, Xing Wang¹²,
16 Rodney J. Weber²⁶, Yuhan Yang²⁶, Xavier Querol²⁷, Gerard Hoek²⁸, Roy M. Harrison²⁹⁺, and Gaëlle
17 Uzu^{1*}.

18

19 ¹University Grenoble Alpes, CNRS, INRAE, IRD, INP-G, IGE (UMR 5001), 38000 Grenoble, France

20 ² European Commission, Joint Research Centre (JRC), Ispra, Italy

³ Institute for Environmental Research and Sustainable Development, National Observatory of Athens, Lofos
 Koufou, P. Penteli, Athens, 15236, Greece

23 ⁴Environmental Chemical Processes Laboratory, Department of Chemistry, University of Crete, Heraklion, 24 71003, Greece.

⁵ MRC Centre for Environment and Health, and the National Institute of Health Research, Health Protection
 Research Unit in Environmental Exposures and Health, Imperial College London, London, UK

²⁷ ⁶ Center for the Study of Air Quality and Climate Change, Institute of Chemical Engineering Sciences, Foundation
 ²⁸ for Research and Technology Hellas, Patras, Greece 26504

29 ⁷ Laboratory of Atmospheric Chemistry, Paul Scherrer Institute, 5232 Villigen PSI, Switzerland

⁸ Department of Atmospheric and Oceanic Sciences, University of California at Los Angeles, 520 Portola Plaza,
 Los Angeles, California, 90095, United States

32 ⁹MRC Centre for Environment and Health, Environmental Research Group, Imperial College London, 86 Wood 33 Lane, London W12 0BZ, UK.

¹⁰ Department of Environmental Analytical Chemistry, Institute of Analytical Chemistry, Czech Academy
 ³⁵ of Sciences, Veveří 97, 60200 Brno, Czech Republic

36 ¹¹Institute of Atmospheric Sciences and Climate, ISAC-CNR, Str. Prv. Lecce-Monteroni km 1.2, 73100 Lecce, Italy

37 12 Southern Ontario Centre for Atmospheric Aerosol Research, University of Toronto, Toronto, M5S 3E5, Canada





¹³ Department of Pollutant Metrology, Institut National de Recherche et de Sécurité (INRS), 54500 Vandœuvre ¹⁹ lès-Nancy, France

40 ¹⁴ Department of Environmental and Biological Sciences and Technologies (DISTEBA), University of 41 Salento, Lecce 73100, Italy

42¹⁵ Atmospheric Chemistry Department (ACD), Leibniz Institute for Tropospheric Research (TROPOS), 43 Permoserstraße 15, 04318 Leipzig, Germany

44 ¹⁶ School of Engineering, Deakin University, VIC 3216, Australia

45 ¹⁷ Vinča Institute of Nuclear Sciences – National Institute of the Republic of Serbia, University of Belgrade, 46 11 351 Belgrade, Serbia

47 ¹⁸ Department of Environmental Sciences, University of Basel, 4056 Basel, Switzerland

48 19 TNO Environmental Modelling, Sensing and Analysis, Princetonlaan 6-8, 3584 CB Utrecht, Netherlands

49 ²⁰ Department of Environmental Science, Stockholm University, Stockholm, 11418, Sweden

50²¹ IMT Nord Europe, Institut Mines-Télécom, Univ. Lille, Centre for Energy and Environment, F-59000 Lille, 51 France

52 ²² Department of Chemical, Pharmaceutical and Agricultural Sciences, University of Ferrara, Via Fossato 53 di Mortara 17/19, 44121 Ferrara, Italy

54 ²³ Center for Primary Care and Public Health (Unisanté), Department of Occupational and Environment Health
 55 (DSTE), University of Lausanne, Switzerland

56 ²⁴ Environmental and Life Sciences, Trent University, Peterborough, Canada

⁵⁷ ²⁵ Department of Civil and Environmental Engineering, University of Illinois at Urbana-Champaign, 205 North
 ⁵⁸ Mathews Avenue, Urbana, IL, 61801, United States

59 ²⁶ School of Earth and Atmospheric Sciences, Georgia Institute of Technology, Atlanta, Georgia 30332,60 United States

61 ²⁷ Institute of Environmental Assessment and Water Research (IDAEA-CSIC), 08034 Barcelona, Spain

62 ²⁸ Institute for Risk Assessment Sciences, Utrecht University, Utrecht, 3584CM, the Netherlands

⁶³ ²⁹ Division of Environmental Health and Risk Management, School of Geography Earth and Environmental
 ⁶⁴ Sciences, Edgbaston, Birmingham, UK, B15 2TT, United Kingdom

65 † Deceased

66 + Also at: Department of Environmental Sciences, King Abdulaziz University, Jeddah, Saudi Arabia

67 *Now at: Department of Atmospheric and Oceanic Sciences, University of California at Los Angeles, Los Angeles,
68 CA 90095-1565, USA

69 #Also at: Bolin Centre for Climate Research, Stockholm, 11418, Sweden

⁷⁰ SAlso at: Laboratory of Atmospheric Processes and their Impacts, Institute of Environmental Engineering, Ecole
 ⁷¹ Polytechnique Federale de Lausanne, Lausanne, Switzerland, 1015

72 *Corresponding author: gaelle.uzu@ird.fr





73 Abstract

74 This paper presents the findings from a collaborative interlaboratory comparison exercise designed to assess 75 oxidative potential (OP) measurements conducted by 20 laboratories worldwide. This study represents an 76 innovative effort as the first exercise specifically aimed at harmonising this type of OP assay, setting a new 77 benchmark in the field.

78 Over the last decade, there has been a noticeable increase in OP studies, with numerous research groups 79 investigating the effects of exposure to air pollution particles through the evaluation of OP levels. However, the 80 absence of standardised methods for OP measurements has resulted in variability in results across different 81 groups, rendering meaningful comparisons challenging. To address this issue, this study engages in an 82 international effort to compare OP measurements using a simplified method (with a dithiothreitol (DTT) 83 assay).

84 Here, we quantify the OP in liquid samples to focus on the protocol measurement itself, while future ILCs should 85 aim to assess the full-chain process, including the sample extraction. We analyse the similarities and 86 discrepancies observed in the results, identifying the critical parameters (such as the instrument used, the use 87 of a simplified protocol, the delivery and analysis time) that could influence OP measurements, and provide 88 recommendations for future studies and interlaboratory comparisons. Even if other crucial aspects, such as 89 sampling PM methods, sample storage, extraction methods and conditions, and the evaluation of other OP 90 assays, still need to be standardised. This collaborative approach enhances the robustness of the OP-DTT assay 91 and paves the way for future studies to build on a unified framework. This pioneering work concludes that 92 interlaboratory comparisons provide essential insights into the OP metric and are crucial to move toward the 93 harmonisation of OP measurements.

94

95 1. Introduction

96 Over the last decade, many studies demonstrated associations between exposure to ambient air pollution and 97 adverse human health outcomes (Hart et al., 2015; Laden et al., 2006; Lepeule et al., 2012; WHO, 2021b, a). 98 Adverse health effects attributable to particle matter (PM) are complex and diverse. Among environmental 99 factors, PM is considered to be the largest contributor to morbidity and mortality globally (WHO, 2017). The 100 casual mechanisms underpinning these adverse associations are diverse, with oxidative stress and 101 inflammation, genomic alterations, damage to the nervous system function, and epigenetic alterations, among 102 others, all cited as potential contributing pathways (Huang et al., 2022; Nicholson et al., 2022; Wilker et al., 103 2023; Zare Sakhvidi et al., 2022). Across these broad domains, the capacity of particles and particle-derived 104 chemicals to cause damaging biological oxidations appears to be a unifying mechanism, both through the 105 introduction of pro-oxidants and stable free radicals into the body, but also through secondary radical/oxidant 106 generation through altered metabolism and induction of inflammation (Li et al., 2008). By definition, oxidative





107 stress is a condition where excess production of reactive oxygen species (ROS) and nitrogen species (RNS) 108 overwhelm endogenous antioxidant defences (Shankar and Mehendale, 2014). Generally, ROS/RNS production 109 in the cells is regulated within physiological limits, through the actions of antioxidant enzymes (e.g. superoxide 110 dismutase, catalase, etc), low molecular weight water-soluble (e.g. glutathione, ascorbate) and fat-soluble 111 (vitamin E) antioxidants (Alkoussa et al., 2020). This antioxidant system plays a valuable key role by limiting 112 ROS/RNS damage, which is associated with cytotoxicity and the induction of inflammation due to changes in 113 the cellular redox balance (Cassee et al., 2013; Gao et al., 2020; Kelly and Fussell, 2017; Sies, 2018). The capacity 114 of PM to invoke biological oxidations has therefore been proposed as a proxy measure of their toxicity, and has 115 been referred to as their oxidative potential (OP); either that intrinsic to their possession of pro-oxidants, or 116 encompassing their capacity more wholistically to simulate ROS/RNS through interaction with cells (Ayres et 117 al., 2008; Bates et al., 2015; Cho et al., 2005; Sauvain et al., 2008; Uzu et al., 2011).

118 Consequently, the OP of PM is increasingly being studied as a potentially health-relevant metric to evaluate 119 effects due to exposure to PM, in addition to PM mass concentration, in multiple regions across the globe (Bates 120 et al., 2019; Bhattu et al., 2024; Daellenbach et al., 2017; Weichenthal et al., 2019). OP is a relatively simple 121 estimation of PM redox activity that reflects a complex interplay of all physico-chemical properties (chemical 122 composition, surface-area, solubility and particle size) contributing to the ROS/RNS generation and the 123 oxidation of target biomolecules, or probes. Implicit within this approach is the contention that not all 124 constituents of ambient PM are equally as harmful, and that those that drive damaging redox reactions, either 125 directly, or indirectly present a greater hazard. Thus, PM composition should be considered a factor more 126 directly linked to adverse health effects than PM mass concentration, highlighting the need to study additional 127 health-relevant metrics such as the OP (Park et al., 2018).

128 In last decades, there has been an increased interest in measuring and developing OP studies, applying different 129 in vivo or in vitro assays and aerosol characterisation techniques to estimate the main sources of OP related to 130 PM (Guascito et al., 2023), and attempting to integrate this metric into health studies. Several acellular chemical 131 methods have been applied for the estimation of the OP of atmospheric particles since these assays allow faster 132 measurement and are less labour-intensive than cell culture or in vivo methods (Bates et al., 2019). In addition, 133 these assays aim to mimic the interaction between PM and different lung antioxidants (e.g. glutathione, 134 ascorbate..., etc) during inhalation. The acellular assays which are most commonly applied include several 135 probe approaches based on antioxidants or surrogates, such as the dithiothreitol assay (DTT), ascorbic acid 136 assay (AA), glutathione assay (GSH), Ferric-Xylenol Orange assay (FOX), 9,10-bis (phenylethynyl) anthracene-137 nitroxide (BPEAnit) ROS assay and 2,7-dichlorofluorescein assay (DCFH) for bound-particles. Molecular probes 138 display variable sensitivities to PM components due to their unique redox potentials and chemical reaction 139 routes, contributing to aerosol OP values. Therefore, it may be necessary to use several assays simultaneously 140 for a broader assessment of the chemical species in PM potentially triggering oxidative stress and to evaluate 141 which of these probes might be most indicative and closely linked to health effects. Furthermore, one of the





142 main challenges within this rapidly expanding research field is the diversity of analytical methods and protocols143 used for OP each assay, which require standardised protocols to support synthesis across the evidence base144 (Ayres et al., 2008).

145 In 2008, a previous workshop gathered experts on OP and developed consensus statements addressing the 146 importance of standardised samples, the comparison between oxidative potential tests, the formulation of 147 consistent standard test protocols and the establishment of connections between OP tests and epidemiological 148 findings as reliable predictors of adverse health outcome (Ayres et al., 2008). Despite more than 15 years having 149 elapsed since that workshop, whilst protocols have matured, evolved, and proliferated worldwide, little 150 concrete work has been performed regarding the harmonization and standardization of these methods.

One of the main objectives of the RI-URBANS European project (https://riurbans.eu/) is to bring accessible service tools to enhance air quality monitoring networks, including evaluating air pollution exposure. As OP has been proposed and recommended as a parameter to be measured in the proposal for a new European Air Quality Directive (Council of the European Union, 2024), an international OP interlaboratory comparison (ILC) was launched to assess the consistency of OP measurements between participants that apply different OP DTT protocols, hindering comparison of results obtained worldwide. The main goal of the ILC was to identify potential discrepancies in results (obtained with the OP DTT assay, one of the most common acellular assays used for measuring the OP) that may arise due to differences in experimental procedures, equipment, or analytical techniques. This ILC constitutes a first step to identify potential sources of variability, resulting in the enhancement of the overall accuracy, reliability and comparability of OP measurements.

161 This paper presents the setup and results of this first large ILC study based on the dithiothreitol (DTT) assay, 162 with a large number of participants (20 groups). The first section includes a description of how a simplified 163 protocol was obtained, along with the coordination/management of the ILC. Subsequently, the results are 164 presented in the second section, combined with statistical analysis, both for the harmonized protocol and the 165 "home" protocols of each participant. Finally, some major findings and recommendations are presented.

166 2. Intercomparison strategy

167 This ILC was proposed within the RI-URBANS European Project framework to evaluate the discrepancies and 168 commonalities of OP measurements obtained by the different participating laboratories. The setup of the 169 protocol was led by a working group of laboratories with considerable experience in oxidative potential: FORTH 170 (The Foundation for Research and Technology – Hellas (FORTH, Greece), NOA (National Observatory of Athens, 171 Greece), ICL (Imperial College of London, United Kingdom), IGE (Institute of Environmental Geosciences, 172 France) and UoB (University of Birmingham, United Kingdom) (i.e. the "core group"). Considerations regarding 173 sampling techniques of PM filters or monitoring strategic approaches are beyond the scope of this exercise.

174 Multiple OP assays are available; however, following a literature review, it was decided to prioritise the DTT175 assay for this first ILC due to its widespread adoption and long-term application facilitating broad participation





176 from laboratories. The core group first produced a harmonised and simplified method, detailed in a 177 standardised operation procedure (SOP), that was integrated, implemented and tested by IGE, the organiser 178 for this ILC. This SOP is called the "RI-URBANS DTT SOP" in the following and is presented in detail in Section 179 2.2. Section 2.1 presents the selection procedure for some parameters of the DTT assay according to variations 180 observed in the literature. Sections 2.3 to 2.7 comprise different parts of the implementation of the ILC, along 181 with the procedure for data processing.

182

183 2.1 Testing the parameters for implementing the simplified RI-URBANS protocol

The simplified RI-URBANS DTT SOP was adapted from the original DTT protocols published in the early 2000s (SOP1: Li et al., 2003, 2009; SOP2: Cho et al., 2005; SOP3: Kumagai et al., 2002; called SOP1, SOP2 and SOP3, hereafter). The principle of the DTT assay relies on the production of superoxide radicals, with DTT acting as a surrogate for cellular reducing agents. This probe contains thiol groups similar to GSH and subjects to oxidation, forming stable cyclic disulphides by donating electrons to oxygen through intermediate redox-active species from PM. In the assay proposed by Kumagai et al. (2002), PM is incubated with DTT and its rate of oxidation over time is assessed through the use of 5,5'-dithiobis-(2-nitrobenzoic acid) (DTNB) with the reaction product 2-nitro-5-thiobenzoic acid (TNB), being detected at 412nm. Whilst DTT is in excess, the rate of DTT oxidation is proportional to the concentration of redox-active species in the PM (Cho et al., 2005; Sauvain et al., 2008). The DTT loss over time can then be expressed per unit concentration of PM (usually μg) or by volume of air (m³) to provide a measure of the intrinsic reactivity of the particles and assess human exposure (Cho et al., 2005), respectively.

196 A review of the pioneer DTT assay protocols revealed differences, and therefore to derive a simplified protocol,197 some variations of the parameters were examined by the ILC organiser laboratory. These results were198 evaluated by the RI-URBANS core group to obtain the final harmonized protocol.

Three key parameters were tested. First, the necessity for the inclusion of trichloroacetic acid (TCA) was evaluated. In the original DTT protocol by Kumagai et al. (2002), the reaction between samples and DTT is quenched at a specified incubation time by the addition of 1.0 ml of 10% TCA to the incubation mixture. Subsequently, 0.5 ml of the reaction mixture is extracted and combined with DTNB and tris–HCl buffer (pH 8.9). However, more recently, Li et al. (2009) demonstrated that DTNB rapidly reacts with DTT, with the absorption reaching its maximum value immediately and remaining stable for over two hours. In our initial tests, we found that this parameter was consistent showing no differences over all the samples tested with or without TCA (Figure S1). Thus, in the simplified OP RI-URBANS SOP, TCA addition was omitted, and we introduced the DTNB solution directly into the mixture at the prescribed incubation times and recorded the absorption after the 30minute reaction.





209 Second, EDTA is present in some OP DTT protocols, whether in the buffer of the incubating PM sample with 210 DTT (Kumagai et al., 2002; Li et al., 2009), or in the titration mixture of DTNB (Cho et al., 2005; Kumagai et al., 211 2002). EDTA is a strong chelating agent which is widely used in biological assays to prevent microbial 212 contamination and to facilitate cell lysis and the extraction of cellular components. Using EDTA is helpful in 213 purifying buffers at a low cost and decreasing a high rate of DTT loss in the blank by scavenging metal ions. 214 However, it can lead to artefacts during the assay. This is especially critical when used in the reaction mixture 215 with the PM, where it could induce complexation with redox changes. Moreover, this is particularly relevant for 216 iron where complexation increases solubility and the resultant EDTA-Fe complex is redox-active (Gao et al., 217 2024), or in the opposite, EDTA can chelate some metallic species, preventing their reactivity (Charrier and 218 Anastasio, 2012). In addition, EDTA also has antioxidant properties, which may compete in the solution with 219 DTT (Thbayh et al., 2023). Our results on different samples show that the presence of EDTA in the buffer leads 220 to underestimation of the OP DTT loss rates in comparison to without EDTA. An impact was observed in the 221 solutions, mainly the copper reference solution $(1 \ \mu M)$ and an ambient PM filter for testing both protocols 222 (Figure S2). That augmentation could also be related to an increase in the blank absorbance without EDTA due 223 to the impurities in the buffer, but this is controlled by subtracting the blank. Little or no impact was shown on 224 the 1,4-naphthoquinone solutions, an organic component. To prevent such undesirable interactions, many 225 laboratories have introduced Chelex® 100, a sodium-form resin to purify the buffers used in the OP DTT assay 226 (Calas et al., 2017, 2018; Charrier et al., 2015; Charrier and Anastasio, 2012), or used commercially available 227 high-purity (e.g. LC-MS/MS grade) water and buffer mixture to create the OP reaction medium (Shahpoury et 228 al., 2019, 2022). Nevertheless, the Chelex® commercial resin comes with a basic pH and requires a pre-229 treatment that could be an extra source of error for this current first international intercomparison. Finally, 230 EDTA was not included in the SOP to prevent complex ligand chemistries, and because the use of Chelex® was 231 complex to introduce for such first ICL; high-grade buffer powders were sent to all the participants instead.

Third, original protocols include the use of a Tris-HCl buffer at pH 8.9 in the solution of titration with DTNB, and this remains widely used. However, Li et al. (2009) showed that the pH of the solution drives both the catalytic redox reaction rate and also the molar extinction coefficient of the product of the reaction, TNB. This report confirmed the previously obtained results by Danehy et al. (1971) that showed that DTNB suffers from alkaline decomposition above pH 8, increasing absorbance values (Figure S3). As a result, for the simplified RI-URBANS DTT SOP, the core group selected a potassium phosphate buffer at a physiological pH of 7.4 to replace tris-HCl in the DTNB solution. This prevents the DTNB alkaline decomposition because the pH is in the favourable range of 5.5-8 where TNB is in the TNB²⁻ form and the DTT+DTNB system does not show significant pH-dependant changes in the absorbance values (Li et al., 2009).

241 Finally, the simplified RI-URBANS DTT SOP also includes a variation considering the instrument used for the 242 measurements. The SOP was elaborated and tested for both cuvette and plate reader spectrophotometers,





including the potential application of automatic samplers. However, the simplified SOP remains to be tested forother instruments (such as Liquid waveguide capillary cell instrument (LWCC)).

245

246 2.2 Simplified RI-URBANS DTT SOP and other measurements

247 The simplified RI-URBANS DTT SOP is proposed in SI-1. It contains a first step for the preparation of the 248 reagents needed for the analyses included in the ILC. A second step describes how to perform a calibration of 249 the analytical device, using a DTT calibration curve with at least 4 points for a concentration range between 0 250 and 60 μ M (titration with 1mM DTNB and reading of TNB formation at 412 nm). In the third step, the SOP 251 defines the performance of the measurements for the test samples provided, including the assessment in 252 triplicates of each test sample and control points (blanks). It should be noted that the DTT protocol also 253 integrates the analysis of control points (blanks), which allows quantifying the inherent DTT background 254 oxidation.

255 The duration of the analyses required for the ILC is variable depending on the instrument used. About 30 min 256 completion time is required, including the assessment of ILC test samples and control points, when performed 257 with a plate reader, and a similar time is needed to perform the calibration curve. When a cuvette-type 258 spectrometer is used, the total analysis time can be at least 2 hours to perform all the triplicates and the 259 corresponding calibration curve.

260 Furthermore, apart from the analytical equipment for monitoring the chemical reactions, the simplified 261 protocol also requires some standard laboratory equipment and conditions, such as access to ultrapure water 262 (18.2 Mohm cm⁻¹, TOC <5 ppb) for the preparation of reagents, the use of vortex for the homogenisation of 263 samples, refrigerated baths to conserve the reagents and transparent 96-wells plates and dark tubes. The 264 samples also need to be kept under agitation during the experiment and at a constant temperature of $37.4^{\circ}C$.

The calculation of the chemical reaction rate of OP DTT during this ILC involves a conversion using a calibration curve. Once the results are obtained, the kinetics of the DTT oxidation can be calculated by subtracting both the intrinsic absorption of each sample (absorption obtained from the samples before the addition of reagents to remove a potential matrix effect between samples) and the inherent DTT auto-oxidation rate (slope of a Control sample) from the DTT consumption rate in the presence of PM.

270 Participants were asked to perform additional OP measurements on the same samples, following the protocol 271 in use in their laboratory ("home protocol(s)") if they wanted to do so. In this case, they had to provide the 272 results in the particular units requested, depending on the applied assay.

273





274 2.3 Type of samples - test materials

275 The ILC was performed using three samples (SP1-SP3), including different concentrations of ambient PM and 276 commercial positive control (1,4-naphthoquinone, CAS [130-15-4], Sigma Aldrich), all extracted in ultra-pure 277 water and prepared by the ILC organiser. Providing ultra-pure water extracts - instead of filter fragments - to 278 the participants was selected for the current first ILC to avoid additional uncertainties associated with the use 279 of different procedures of sample extraction, different quality of the ultra-pure water for sample extractions, 280 and changes linked to the processing equipment available. More specifically, the samples sent to the 281 participants included SP1: 1,4-naphthoquinone solution (reference compound, 5 μg mL⁻¹), SP2: extract from a 282 PM sample influenced by biomass burning emission (obtained from a chamber experiment, at $25 \,\mu g \, m L^{-1}$), and 283 SP3: an urban PM extract highly influenced by traffic emission (from pooled roadside samples obtained from 284 TEOM - FDMS reference samplers, at 25 μg mL⁻¹). Additionally, a fourth sample, SP4: a sample extracted from a 285 blank/clean quartz filter was sent to the participants, but it was not included in the evaluation since the 286 measured values were close to the instrument limits of detection for most participants. For each of the 4 287 samples, all the sub-samples, distributed to participants, resulted from a unique 1L solution obtained from the 288 original sample substrate. For instance, SP1 and SP3 were powders that were solubilised and homogenised for 289 75 min by vortex agitation in ultra-pure water. SP2 and SP4 were quartz fibre filters subjected to a 75-minute 290 vortex extraction in ultrapure water.

291 Several 5 ml sample aliquots in dark polypropylene tubes were sent to each participant, according to their 292 needs, allowing them to perform triplicate measurements for the RI-URBANS DTT SOP and all the "home" OP 293 protocols implemented in their labs.

294 Solid potassium dihydrogen phosphate (CAS [7778-77-0], Roth), di-potassium hydrogen phosphate (CAS 295 [7758-11-4], Roth), 5,5'-Dithio-bis-(2-nitrobenzoic acid, CAS [69-78-3], Roth) and 1,4-Dithiothreitol (CAS 296 [3483-12-3], Roth) were also distributed to the participants, to prepare the solutions for the RI-URBANS DTT 297 SOP, including the DTT solution, the Dithio-bis-(2-nitrobenzoic acid) (DTNB) solution, and the potassium 298 phosphate buffer solution.

299 2.4 Transport of samples, ILC performance and duration

Test samples (SP1-SP4) were shipped to all participants on 17th January 2023 via courier in refrigerated and
isolated ice packs, and received as chilled liquid samples. The parcels were delivered between 18th January and
2nd February 2023. On average, the participants received the samples in 3±2.5 days and performed the analysis
in 14±10.5 days and up to 31 days after reception of samples. The recording of these parameters allows their
integration to the multiple linear models used in this work.

305 2.5 Reporting of the results

306 Participants were asked to report OP DTT results from the RI-URBANS SOP in nmol DTT min⁻¹ μ g⁻¹ and the % 307 DTT consumed in μ g⁻¹ min⁻¹, applying three decimal digits for all three replicates of test samples. An Excel 308 spreadsheet with all the calculations pre-included was prepared by the ILC organiser and shared with all the





participants to avoid calculation errors and to facilitate the standardisation of results. In addition, participants
were invited to report, under the same format, the values for other OP tests, such as OP DTT "home", and other
OP tests like AA, DCFH, OH, ESPR, GSH and RP (routinely applied by each participant) on the same samples.

312 2.6 Number of participants

313 It is worth noting that for the first time, a total of 20 research groups participated in the exercise: 14 of them 314 from Europe, including the United Kingdom, Italy, France, Switzerland, Greece, Germany, Serbia, the Czech 315 Republic, the Netherlands and Sweden, 3 participants from the United States, 2 participants from Canada and 316 1 from Australia. The participants were invited for their contribution to the RI-URBANS project, through a 317 public call to participate, or because they contacted the ILC organiser directly and were selected, due to their 318 active role in the OP scientific community. The ILC was performed using anonymous participation; thus, a 319 number was randomly assigned to each participant to present the results. Participant L5 cancelled his 320 participation in the ILC and two participants (L3 and L16) did not send their results for RI-URBANS DTT SOP.

321 2.7 Data evaluation

322 The ILC results were analysed by the European Commission Joint Research Centre (JRC), which provides 323 independent, evidence-based science and knowledge support for EU policies and in conducting ILC exercises. 324 The participation of an external independent evaluation was required following the International Global 325 Standard ISO 5725-2, related to the accuracy of measurement methods and results (Part 2: Basic method for 326 the determination of repeatability and reproducibility of a standard measurement method). The data 327 assessment includes statistical evaluation of homogeneity, ageing, and repeatability, as well as the 328 establishment of the assigned values and the related methods.

329 2.7.1 Estimation of the assigned value and participants' performance

330 Different methods can be applied to determine the assigned (consensus) value for a comparison exercise when 331 no reference or certified reference material is available. The first one would be to use a robust mean and 332 standard deviation including all participants, but this could become statistically ineffective if the number of 333 participants is below twenty. Further, the results we obtained were not normally distributed (Figure S4), 334 compromising the accuracy of the robust mean of the samples tested. Thus, three different alternative 335 approaches were evaluated in this ILC until meeting the requirements of Standard ISO 5725-2. The first one is 336 based solely on the results obtained by the ILC organiser, where the samples were prepared, and the 337 homogenisation tests were performed. The assigned values obtained from this approach include a coefficient 338 of variability of 11%, 7% and 6% for SP1, SP2 and SP3, respectively. The second approach applies the Q/Hampel 339 test, a robust algorithm useful for data sets with outliers to calculate the robust standard deviation and mean 340 of the data set. This method is largely applied to the statistical analysis of interlaboratory studies. The 341 Q/Hampel test results (integrating the average of the results from all the participants) present COVs of 60%, 342 61% and 70% for SP1, SP2 and SP3, respectively. However, the values obtained for the dataset were deemed





343 too high to guarantee the good performance of the Q/Hampel test calculation. The third approach relies on 344 calculating the assigned (consensus) value from the average and standard deviation for a specific reference 345 group, including 6 participants selected (without knowing their results) due to their previous pioneering 346 experience in developing OP protocols and measurements, taken as their strong expertise in the domain (>10 347 publications in the field). The participants selected were the L2, L4, L8, L12, L13 and L19. This approach leads 348 to coefficients of variability of 31%, 33%, and 47% for SP1, SP2 and SP3, respectively. Since the first two 349 methods showed a contrasted variability, one with too little and the other with too much variability, the 350 assigned values were calculated using the third approach, integrating the average and standard deviation of the 351 results from this selected experienced OP participants.

The individual performance of each group was further evaluated using z-scores, a metric indicating the deviation of each data point from the assigned value as compared with the standard deviation for proficiency assessment σ^* . For each laboratory and test sample, the z-score is computed using the formula: $z = (x_i - X)/\sigma^*$, where x_i is the result from participant *i*, X is the sample assigned value (0.53, 0.14, 0.07 nmol min⁻¹ µg⁻¹), and σ^* is calculated as the reproducibility standard deviation among the six selected participants (0.16, 0.06, 0.04 nmol min⁻¹ µg⁻¹). An "action signal" is triggered if a participant's entry produces a z-score exceeding +3 z or falling below -3 z, indicating a deviation of more than 3 standard deviations from the assigned value. Similarly, a "warning signal" is raised for a participant z-score above +2 z or below -2 z, representing a deviation between 2 and 3 standard deviations. A participant z-score between -2 z and +2 z signifies satisfactory performance concerning the standard deviation for proficiency assessment.

362

363 2.7.2 Analysis of the variability of the results

Additionally, the statistical distribution of results was evaluated using multiple linear regression models. In the first model, the effects of the protocol variables on the measured OP were investigated using linear regression models adjusted on the instrument (3-class variables: plate reader, cuvette and LWCC), delivery time (time between sample shipment and reception, continuous), and analysis time (time between reception and analysis, continuous). An additional model (M2) compared the RI-URBANS SOP and the DTT-home protocols was further adjusted based on the protocol (2-class variable: RI-Urbans, DTT-home). Finally, the evaluation of the average performances (3-class variables: low - 0<|z-score |<2, middle - 2<|z-score |<3, and high - |z-score |>3) was added in the M2 model, to assess whether performance affected DTT activity in the same direction (i.e. positive or negative), while considering other protocol variables. Each model was run separately for SP1, SP2 and SP3 samples. All analyses were conducted using R (version 4.2).





375 3. Results and Discussion

376 Out of the group with a total of 20 participants, 18 presented results obtained using the RI-URBANS DTT SOP.

377 Different instruments were used to apply the simplified SOP. Overall, 9 participants (47.5%) used the cuvette-378 type spectrophotometer, 8 (42%) used the plate reader-type spectrometer, and 2 (10.5%) implemented LWCC

379 measurements (one participant used two instruments).

380 3.1 Homogeneity of the samples

An initial assessment of the homogeneity of the OP measurements with the 3 test samples was performed by the ILC organiser, using a plate-reader type protocol, deriving the mean and the standard deviation of 10 replicate analyses performed on the same day. The results obtained from the coefficient of variations, showed the sample variabilities were up to 12%, 7%, and 9% for SP1, SP2 and SP3, respectively (Figure S5), showing a higher variability for the 1,4-naphthoquinone solution (SP1) compared to the two filter extracts (SP2 and SP3).

386 The overall uncertainty of the OP DTT assay has been evaluated to 18% for PM_{10} and 16.3% for $PM_{2.5}$ by Molina 387 et al., 2020. Despite some differences observed between our samples, the results are deemed acceptable, 388 presenting a variability of around 10%, which indicates a very good performance of the analysis.

389 3.2 Ageing of samples

To reduce the number of parameters affecting the preparation of the sample solutions liquid solutions of each sample were prepared and sent to the different participants. However, liquid samples can undergo ageing processes, impacting OP levels over time. For this purpose, an ageing test was performed by the ILC organiser to evaluate the potential changes over time. It consisted of regularly implementing the RI-URBANS DTT SOP to obtain the values of each test sample over time. Figure S6 shows these results, where SP2 and SP3 do not show a strong change over time, while sample SP1 presents a pronounced ageing effect. In routine tests of ICL organizer, the 1,4-naphthoquinone mother solution at such high concentration is usually stable in a glass container for weeks but here, potential interaction with the PP tubes' inner surface may have happened. Consequently, ageing could be a variable of importance for the participants who analysed the samples toward the end of the required period, and such parameter (date of analysis) was thus included in the parameters to 400 be tested for the research of critical parameters.

401

402 3.3 Statistical distribution of results: Participants' variability

403 In order to assess the intra-laboratory variability of the results, the coefficient of variation (COV = standard 404 deviation/mean * 100) of the results for each sample and each laboratory are presented in Figure 1, while the 405 standard deviations of the replicates reported for each sample are presented in Figure S7 and table S1. Overall, 406 higher COVs are observed for SP3 and SP1, where most participants (44.4 and 38.9%, respectively) presented 407 higher values for this sample compared to the SP2 sample. Specifically, high average COV values are observed





408 by L1, L12, L14, L15 and L21, with an average variation higher than 40%. Only a few participants (6 groups) 409 presented a variation lower than 10% for the three samples. This is the same pattern for the results obtained 410 by the ILC organiser during the homogeneity test (see 3.1.), where the COV for SP1 was larger than those for 411 SP2 and SP3 but with the highest COV below 15%. These findings confirm more homogenous results for 412 samples SP2 and SP3 compared to SP1 but could also indicate that some participants failed to achieve 413 repeatability observed by the ILC organiser. However, some groups (L2, L10, L19, L20) were able to produce 414 very homogeneous results with COV < 10% for all 3 samples.



415

Figure 1: Coefficients of variation of each participant (L1 to L21) for the three samples in triplicates
tested using the RI-URBANS DTT protocol and the median and mean repeatability for each laboratory.
Yellow-highlighted participants are the ones selected for the calculation of the assigned values. The
dashed line indicates the participant with COV lower than 20%.

420

421 3.4 Laboratory performances

422 The assessment of laboratory performances first presents the bias in results across participant groups 423 compared to the assigned values and their associated standard deviation for each sample. As illustrated in 424 Figure 2, SP1 exhibited the highest variances, ranging from 130% to -35%, with only five groups displaying 425 differences within $\pm 10\%$. The distribution of results for SP1 indicated a mix of overestimations and 426 underestimations. For SP2, differences are within a narrower range from 43% to -7%, primarily favouring 427 overestimations. For this sample, 12 participants returned results that were within $\pm 10\%$ of the assigned value 428 (see highlighted laboratory numbers in Figure 2). Finally, the results for SP3 demonstrated the least variation 429 among participants, with differences ranging from 30% to -6%, and 16 participants within $\pm 10\%$ of the 430 assigned value, again favouring overestimation compared to the assigned value. In total, 14 laboratories 431 obtained data with $\pm 10\%$ difference to the assigned value for SP2 and SP3 (see highlighted laboratory numbers 432 in Figure 2). These results show again that the reference samples with 1,4-naphthoquinone (SP1) most





433 probably present some characteristics leading to this variability and may be associated with a less stable 434 solution or led to saturation of some detectors regarding the relatively high concentrations, while the samples 435 from filter extractions do not. Additionally, it is interesting to note that there is apparently no systematic 436 pattern where a given participant would obtain out-of-range results for all samples. The results are really 437 diverse, and most participants can obtain "acceptable" results for one or two samples (SP) and larger variability 438 associated with one "unacceptable result" for one of them. In 2020, Molina et al. (2020) explored the total 439 uncertainty of OP DTT of a collection of samples and evaluated it to 18% for PM10 and 16.3% for PM2.5. The 440 leading factors identified were the DTT consumption rate (regression and repeatability of experimental data) 441 and the extraction volume operations (pipette). This underscores the need for further investigations on the 442 experimental causes of the variations observed, possibly in the next ICL.







444

445 Figure 2. Percentage differences from the assigned (consensus) value for each sample (SP1, SP2 and 446 SP3). The results compared the average of the triplicates reported by the participants. Yellow-447 highlighted participants are the ones selected for the calculation of the assigned values and underlined 448 the ones that obtained data into ±10% difference to the assigned value for SP2 and SP3.

449

450 The individual performance of each group was further evaluated using z-scores. The results are presented in 451 Figure 3. All underestimations fall within the acceptable range (lower than -2 z). Additionally, it is noteworthy 452 that no laboratory exhibits unsatisfactory performances across all three samples; for almost all participants,





453 while one sample can present poor results, it coexists with two acceptable ones. This has strong implications 454 for spatial and temporal analyses that are often performed for OP. Again, this calls for attributions that there 455 are no systematic biases in the analyses. While factors like sample inhomogeneity may be playing a role 456 (particularly for SP1), some other issues, including variability in the performance of the analysis, may have an 457 impact. Hence, participants exhibiting significant deviations (|z-scores| > 2) for some of their results should 458 thoroughly examine their procedures and possibly implement appropriate corrective actions to avoid similar 459 outcomes in future ILCs.

However, half of the participants achieved results within the acceptable limits of this test. Despite disparities,
these findings are really promising, especially considering that this is the first intercomparison of its kind. For
instance, such results are in the same range to those obtained for some of the first ILCs for PAHs (Grandesso et
al., 2012; Verlhac et al., 2014).



464

Figure 3. Z-scores were calculated to evaluate each participant's performance in the interlaboratory
comparison for each sample tested. Yellow-highlighted participants are the ones selected for the
calculation of the assigned values. Black and red horizontal lines indicate boundaries for triggering an
action signal a described in section 2.7.1.

469 To gain more knowledge about the factors causing the variability of the results, we first tried to perform a 470 cluster analysis using the Ward method. This grouped the participants into four clusters (see Figure S8), with 471 the main cluster (in yellow) including the 10 participants encompassing mainly the ones with satisfactory Z 472 scores. The clustering seemed independent of the instrument used and/or the time taken between the sample 473 delivery and analysis (i.e. near the delivery time or later in February).





474 In a second step, a multiple linear regression model was run to evaluate the associations of the results obtained 475 for the 3 samples, SP1, SP2 and SP3, considering a range of parameters, including the instrument used and the 476 delivery and analysis time (Figure 4, Table 2). The beta values are shown in Figure 4, representing the 477 association (effects) between the different parameters evaluated and the OP results obtained. In the model, the 478 reference variables were the RI-URBANS DTT SOP and the results obtained with the plate-reader instrument. 479 Regarding the instrument performance, the values provided by the cuvette-spectrometer were higher than 480 those obtained with plate readers in the case of SP1 and SP2 (showing significant overestimation in the case of 481 SP2 p-values <0.05), while the results for SP3 were quite similar. In the case of LWCC, higher variability is 482 observed when compared to both cuvette and plate reader for all the samples. SP1 LWCC results presented the 483 highest variability, and SP2 results were significantly overestimated (at a 95 % confidence level) when 484 compared with those obtained by the plate reader. The RI Urbans SOP was adapted for plate readers and 485 cuvettes, in order to perform the measurements in similar conditions of concentrations for the reagents. This 486 was not the case for the LWCC since we did not have all the necessary information concerning the specific 487 devices used by participants. Figure 4 suggests that the specific conditions of the reaction are probably 488 important factors for delivering an accurate value of OP. In Figure S6, we showed that SP1's OP activity 489 decreased over time during storage, but this ageing effect was not found to be significant in the model for either 490 delivery or analysis time. The storage effect remained consistent for SP2 and SP3, as there was a clear 491 association between OP and delivery or analysis time (Figure 4).









Figure 4. Associations (beta in nmol min⁻¹ μ g⁻¹) between OP DTT values for SP1, SP2 and SP3 using the RI-URBANS protocol and technical parameters, including the instrument used and the delivery and analysis time obtained by applying an adjusted multiple linear regression model. Full-colour dots represent the results with p-values <=0.05, and white dots represent the results with p-values >0.05.

498

499 A ranking of the samples is also proposed to evaluate the OP activity of the samples tested in this ILC and its 500 relative variability within the participants (only considering the results obtained with the simplified RI-501 URBANS protocol). For this purpose, SP1 was arbitrarily selected as the one with the highest OP activity with 502 an assigned value of 100, and SP2 and SP3 were evaluated in function of SP1. Figure S12 shows the results 503 obtained for the relative ranking of the samples. It can be noted that most of the participants presented similar 504 relative variability with SP1 > SP2 > SP3. Some exceptions were observed for L1, L17 and L19, which obtained 505 higher ranking for SP3 than SP2. A higher variability in the relative activity is obtained for SP2 than for SP3. 506 Within the participants showing a higher relative ranking for SP2 (higher than 50% compared to SP1), most of 507 them used either cuvette-type or LWCC instruments (except L20), suggesting some overestimation in the 508 results using these instruments. Overall, this similar ranking for the samples achieved by most of the groups is 509 noticeable and very encouraging. In fact, most of the data treatment performed on OP with atmospheric 510 variables or health data relies on associations and regressions where the relative variability of a time series is 511 of utmost importance, more than the absolute value.

512 3.5 Comparison with other OP tests provided

513 Participants were also invited to report results obtained using other OP assays. Since not all participants 514 submitted results from equivalent "home OP" tests, we exclusively focus on the outcomes obtained through the 515 "DTT-home" protocols involving 13 participants (Table S3, Figure S9). It is important to note that DTT-RI-516 URBANS protocol was simplified and does not include EDTA, Tris-HCL, TCA neither Chelex®, whereas DTT-517 "home" protocols are diverse and should exhibit at least one up to 3 of the last mentioned compounds at 518 different step of the reaction; however, this is challenging to evaluate because not all groups have submitted 519 the protocols related to their "home" results.

We have first evaluated the COV individually from the results obtained with protocols (RI-URBANS SOP versus all DTT-"home") for each test sample. Figure 5 shows that lower COVs are generally observed in the performance of the DTT-home protocols (more details can be found in Table S3). However, six out of 12 participants presented similar COVs (within 20%) for the two protocols. These results could indicate that the use of a simplified OP protocol needs some extent of training and guidance before its application. In addition, some of the participants presented higher COV values (L1 and L13 for DTT-home) when using the LWCC instrument. The lack of a simplified protocol for this instrument did not seem to be a major issue, as the application of the DTT-home protocols was also associated with high COV.







529

Figure 5: Coefficients of variation of each participant for the three samples tested from triplicates
using both the RI-URBANS DTT and the DTT-"home" protocols. The average and standard deviation of
each participant are also detailed in Table S3.

533

Another multiple linear regression model was run to evaluate the main differences in the results obtained between RI Urbans SOP and DTT-"home" protocols (Figure 6). For SP1, a significant overestimation was observed for the DTT-home protocol and the opposite was observed for SP3. In the case of SP2, there was no statistically significant difference between both protocols. More details on the concentration of DTT for each DTT method could provide more insight into this trend. Regarding the instrument performance, the LWCC presents poorer results compared to the cuvette and plate reader for all the samples, which is opposite to the trend observed with the results of the RI-URBANS protocol only (Figure 4). However, the results are significantly underestimated for SP3 only (p<0.05). For the cuvette-based measurements, the results are higher than those obtained with plate readers in the case of SP1 and SP2 (significantly overestimated for SP2) and similar for SP3, which is in line with the direction observed with the RI-URBANS protocol only. The delivery for SP3 (delivery time) but nothing significant lack of effect for SP2 (analysis time) and an underestimation for SP3 (delivery time) but nothing significant for SP1, although it had undergone ageing in the tests of the ILC organiser. Since the effects are very small compared to the effects of the protocol, or the instrument, these two variables (delivery and analysis time) may cause a greater impact on DTT values when protocols are harmonised, but not to date.

549 The results obtained by the participants (the z-scores evaluation) were also added to the former model to 550 evaluate the effect on OP values while adjusting the protocol variables (Figure S10). The results show a





significant OP overestimation of all the samples for the labs with poor performances in SP2 and SP3 samples,and also a significant underestimation in the OP value obtained for SP2 for the group with intermediateperformance.

Hierarchical cluster analysis was conducted, incorporating both the DTT "RI-Urbans" and DTT "home" 555 outcomes (Figure S11). Because some participants did not implement a DTT "home" protocol, the cluster 556 analysis involved a reduced set of OP values. The results reveal the presence of four primary clusters, with the 557 predominant cluster encompassing most participants (8 out of 12 for the DTT- "home"). The participants within 558 the green main cluster largely align with the results derived from the DTT "RI-Urbans" outcomes, encompassing 559 the groups in the two primary clusters (Figure S11). This assessment illustrates some consistency of results 560 obtained across various OP DTT protocols. Some of the participants with more reliable results for the RI-561 URBANS DTT SOP maintain their consistency regardless of the protocol used. However, some of those that did 562 not show an acceptable performance for the simplified protocol (i.e. L1 and L10) presented a better 563 performance for the DTT-"home", and the opposite was observed for the L19 (almost for SP1).



564

Figure 6. Associations (beta in nmol min⁻¹ μ g⁻¹) between OP DTT values obtained for SP1, SP2 and SP3 and the different parameters of the ILC, including the DTT protocol, the instrument used and the delivery and analysis time obtained by applying a multiple linear model. Full-colour dots represent the results with p-values <=0.05, and white dots represent the results with p-values >0.05.





Finally, to assess the performance of the participants in the DTT-"home" protocols, a comparable approach to
the simplified RI_URBANS SOP was employed for those participants who supplied OP results. The z-scores were
computed using the assigned values of each sample (SP1-SP3), obtained with RI-URBANS SOP application.
Figure 7 illustrates the z-scores of the OP results obtained through the application of the DTT-"home" protocols,
revealing a significant variation in the outcomes. Only five participants managed to produce satisfactory results
for all the tested samples. Despite the fact that the COV of the participants using DTT-"home" protocols showed
an improvement over the results of the simplified DTT SOP (Figure 5), the outcomes are still distant from the
consensus values of the samples obtained in this exercise. The results indicate a high degree of variability in
the OP activity using "home" OP methodologies, underscoring the pressing requirement for standardized
methods and harmonised protocols to ensure more reliable OP research.

580



581

Figure 7. Z-scores were calculated for the DTT-"home" protocol results to evaluate each participant's
performance in reference to the RI_URBANS assigned (consensus) values, for each sample tested. Black
and red horizontal lines indicate boundaries for triggering an action signal a described in section 2.7.1.

585

586 4. Strengths and limitations of this first intercomparison

587 The greatest strength of this ILC was the high number of participants (20) enhancing the comprehensiveness 588 and diversity of the study and allowing for a broader range of perspectives and expertise. This also allowed 589 comprehensive collaborative discussions during the preparation phase, promoting knowledge exchange and 590 consensus-building, and contributing to a more robust ILC design. These all show a willingness from the groups





591 to be actively part of the development of the intercomparison and to pursue a harmonisation on the OP 592 measurements.

593 The development of the first simplified OP DTT protocol (available in SI-1) also consolidates the experimental 594 experience of the participants, fostering methodological consistency across different research groups as a first 595 step toward method harmonisation. Finally, the collaboration with the JRC, an independent organisation for the 596 assessment of results, adds credibility and objectivity to the study, ensuring that findings are impartially 597 evaluated.

The sharing of liquid samples in this comparison comes with both advantages and limitations. On one hand, it eliminates biases associated with extraction methods and solvent purity. However, some samples exhibited signs of ageing during the interlaboratory comparison duration (though this was not identified as a critical parameter when identifying the main causes of variability). In addition, this approach introduced certain challenges with some of the "home" OP protocols which were designed originally to be used with solid samples. Finally, all liquid extracts should be provided with a similar PM concentration to limit known nonlinearities and to avoid potential saturation issues, as can be the case of LWCC instruments. Next, ICL should, in the future, include the whole chain assessment, including the extraction step.

606 The testing of three samples with different patterns makes it difficult to draw unilateral conclusions. A larger 607 sample size could support a more robust statistical analysis of the results, particularly for the factors 608 determining OP variability. Moreover, the inclusion of samples that are readily accessible worldwide, such as 609 standard reference materials, could facilitate the future adaptation of a simplified OP DTT protocol while 610 allowing comparison with an assigned (consensus) value.

611 The sole focus on the DTT method for the ILC could limit the broader evaluation of OP. The addition of other 612 OP assays, such as those included in previous inter-comparison studies (Ayres et al., 2008; Calas et al., 2018; 613 Shahpoury et al., 2022) could provide a more comprehensive understanding of the performances of the 614 different groups involved in the OP domain of research. There are contrasting reports about the relative 615 sensitivity of DTT assay to various organic and inorganic PM components, with some studies showing higher 616 reactivity towards the organic fraction. Therefore, additional consensus studies would be needed to assess this 617 aspect and the comparability of DTT to other OP metrics that rely on proper lung anti-oxidants and could be 618 considered more physiologically relevant. Such studies could support the identification of chemical species 619 which should be prioritized for future air quality management programs.

620 In conclusion, while this ILC of OP has highlighted considerable variability in the performance of the assay 621 between groups, it has notable strengths and provides a starting point towards the harmonisation of OP 622 measurements.

623

624 5. Recommendations for standardisation of OP protocols





Based on the findings of this ILC and also on general literature about OP, some recommendations for further standardising OP DTT measurements are proposed (Table 1). These include guidelines for sample and laboratory conditions, instrument type and calibration and the reporting of results. Additionally, a reference material (1,4-naphthoquinone, copper or other solution at a known concentration) should be proposed to facilitate future ILCs. Additionally, since some differences were observed in the results obtained from the use of the simplified SOP compared with the "home-developed" DTT protocols, harmonisation of procedures is needed to ensure data comparability. We describe below the crucial parameters that need to be considered in the move toward greater harmonisation of OP methodologies.

633 OP assay selection

634 To date, it remains unclear which oxidative potential (OP) assay is most effective at predicting health ٠ 635 outcomes related to oxidative stress. This uncertainty arises because different assays yield markedly different 636 results for the same particulate matter (PM) samples. Additionally, even OP values from the same assay are 637 often linked to various PM components and sources, depending on the different studies (He and Zhang, 2022). 638 Thus, based on current knowledge and epidemiological evidence, two complementary OP assays (a thiolbased probe (OPDTT or OPGSH) and another one (among OPOH, OPAA, or else)) are recommended to provide 639 640 a better picture of the potential oxidising damages from PM compounds and to strengthen the power of epidemiological studies. These aspects were previously discussed in a recent work integrating five different 641 OP assays (Dominutti et al., 2023). Finally, the final choice of the best OP test (or combination) must be based 642 643 on epidemiological evidence, which has begun only recently and needs more hindsight to be determined.

644 Sampling

• OP can be analysed in filter samples conventionally collected for air quality monitoring using small portions of these if adequately preserved (frozen). Pre-burn quartz filters or Teflon filters are appropriate and blank filters must be measured to remove the background induced by the matrix of the material. A previous study had shown no differences in the OPDTT values observed using Quartz or Teflon filters (Frezzini et al., 2022). However, it should be further evaluated when other OP assays are considered.

650 Sample storage

• Previous studies that evaluated the effect of storage time and conditions did not show a substantial effect on the OP DTT results (Frezzini et al., 2022). However, we recommend that PM samples should be immediately transported to the lab after sampling. The filters must be kept cold after sampling (at 4°C if the OP analysis is done within a few days after collection or -20°C if the analysis is delayed).

• The lifetime of the ROS may be very short, and measurements of OP on PM-extracted filters are likely affected both by the age of the samples, how they have been sampled and stored, and the nature of the extraction methodology. How all of these processes impact on the ageing of samples and the ultimate





- 658 quantification of OP needs to be addressed. Ageing studies should be performed for each OP assay in the long
- 659 term to define the maximum storage time of aerosol filters at low vs ambient temperature conditions.
- 660 Some OP components might be so short-lived that only online techniques are warranted.
- 661 Laboratory conditions
- OP assays are "trace" detection assays that require clean ambient conditions and high-quality reagents
 free of metal contamination. Considerations should be given to the use of certified clean rooms or proper
 laminar flow bench stations to prevent contamination of the samples.
- Use of clean material: vials, cones, and spatulas have to be washed before use (5% HNO3 bath to remove
 metals and rinse three times in ultra-pure water before drying in laminar flow).
- Control laboratory temperatures and light exposure by using dark polypropylene tubes at least for reactants
- 669 Extraction step

670 . The extraction step may be highly variable according to the procedures used, and several parameters are 671 known to impact OP results, such as the choice of the solvent, the concentration of buffer, the way of agitation, and the quantification of the final extracted mass. Notably, the ultrasonication of PM samples in aqueous 672 solutions generates ROS (Miljevic et al., 2014) and it could introduce artefacts in OP measurement. This effect 673 674 was also observed in the work of Frezzini et al. (2022), where different extraction methods were evaluated, with ultrasonic baths overestimating the results observed. 675 The effect of the solvent used was not evaluated in this ILC exercise. However, we recommend the use of 676 677 ultrapure water or simulated lining fluid for the sample extraction. Future ILC exercises should include the evaluation on the extraction conditions, including solvent use and methods. 678

679 Reaction step

• Several aspects in the reaction process affect the OP value, like the initial concentration of reactants (since the DTT test is mass-dependant (Charrier et al., 2016)), ratio of reactant/sample, time of reaction (some compounds present a non-linear reaction over time), the temperature of the reaction (which should be standardised to 37°C), agitation (mixing samples) and the type of measurements (kinetic or end-point value), etc.

Current literature mainly addresses extraction or reaction parameters separately. We advise that the
 whole chain factors should be evaluated together to quantify their relative impact on the results.

- 687 Development of a reference material with a certified "OP value."
- 688 The setup of reference material or in-house standard solutions (in collaboration with reference institutions
- 689 JRC or NIST, for instance) with a known OP value could help laboratories test and train themselves on the OP
- 690 protocol before testing the unknown ILC samples. This is something to be developed and tested in future ILCs.





691 Instrument calibration

- Investigate the optimal frequency for the calibration of spectrophotometers for such assays.
- 693 Report of results/units
- 694 The calculation of OP DTT activity during this ILC involved a conversion using a calibration curve. Since the
- 695 OP activity measures the rate of a chemical reaction and not a concentration, for future comparison exercises,
- 696 the possibility of exploring alternative methods for OP calculation should be tested. To date, results are mass
- 697 normalised in nmolAnti-oxidant min⁻¹ μg⁻¹, or volume normalised in nmolAnti-oxidant min⁻¹ m⁻³. The OP per
- 698 μg refers to the reactivity of one μg of the tested PM, whereas the OP per m³ refers to the exposure of one m³
- 699 of inhaled air.

700 Table1. Summary of recommendations for future OP measurements on filters

Condition or step	Recommendations
OP assay selection	 Two complementary OP assays (a thiol-based probe (OPDTT or OPGSH) and another one (among OPOH, OPAA, or else)) are recommended to provide a better picture of the potential oxidising damages from PM compounds
Sampling	 Pre-burn quartz filters or Teflon filters are appropriate and blank filters must be measured to remove the background induced by the matrix of the material.
Samples storage	• The PM filters must be kept cold after sampling (at 4°C if the OP analysis is done within a few days after collection or -18°C or -20°C if the analysis is delayed).
Laboratory conditions	 Clean conditions (including certified clean rooms or proper laminar flow bench stations) High-quality reagents free of metal contamination Use of clean material, which must be washed before use (5% HNO₃ bath to remove metals, rinse three times in ultra-pure water, and dry in laminar flow bench stations). Control laboratory temperatures and light exposure by using dark polypropylene tubes
Reaction step	 Several aspects in the reaction process affect the OP value, and to minimise their impact, standard conditions should be fixed as the initial concentration of reactants, ratio of reactant/sample, time of reaction, the temperature of the reaction (37°C), agitation (mixing samples) and the type of measurements (kinetic or end-point value), etc.
Instrument calibration	• Investigate the optimal frequency for the calibration of spectrophotometers for such assays.

701

702 6. Conclusions

703 This study represents an innovative effort as the first interlaboratory OP exercise specifically aimed at 704 harmonising this OP assay. This exercise provides the very first roadmap for refining interlaboratory





705 comparisons of OP, fostering greater confidence in the reliability of OP data and encouraging the scientific706 community to advance towards global OP harmonisation.

This first exercise focused on OP DTT, as it is widely used within the scientific community and has already shown positive associations with health outcomes (Bates et al., 2015; Borlaza et al., 2022; Dabass et al., 2018; Donaldson et al., 2001; Gao et al., 2020; Marsal et al., 2023; Weichenthal et al., 2016b, a, c). Even if there are several crucial points to be evaluated and harmonised in the whole chain of the determination of OP (sampling methods, sample storage, extraction conditions and methods) as well as the use of different OP assays, this first ILC engaging several research laboratories pay the way for future developments towards the standardisation of OP methods. Despite the need to evaluate and harmonize several crucial aspects throughout the entire process of OP measurements (such as sampling methods, sample storage, extraction conditions and methods, and the use of different OP assays), this initial ILC, which involves multiple research laboratories, paves the way for future advancements in the standardization of OP methods.

Our findings emphasise both the strengths and challenges associated with the use of the current OP DTT assay for driving a measurement of PM OP. Overall, half of the participants achieved results falling within a satisfactory range of z-scores for this test. The participating group performance levels are comparable to those observed in initial ILCs for PAHs in the 2010s (Grandesso et al., 2012; Verlhac et al., 2014). While notable agreement was observed in certain samples and between several groups, discrepancies and variability were also identified, emphasizing the need for harmonisation in the procedures and conditions. A number of factors may contribute to the underperformance observed in certain samples and participants. The main reasons are not clear, but the analysis conditions in the participating laboratories and the lack of experience in this type of metrological exercise are possible causes. Standardisation of protocols and harmonisation of procedures collaborative approach fosters a more robust OP science, facilitates data exchange and integration, and will ultimately contribute to a better understanding of the health impacts associated with PM exposure, allowing for more accurate exposure assessments and regulatory decisions.

730

731 Acknowledgements

The present work was supported by the European Union's Horizon 2020 research and innovation program under grant agreement 101036245 (RI-Urbans), including the Post-doc grant of Pamela Dominutti, and by a University Grenoble Alpes grant ACME IDEX (ANR-15-IDEX-02). Chemical analysis of the Air-O-Sol facility at IGE was made possible with the funding of some of the equipment by Labex OSUG@2020 (ANR10 LABX56). DP, AB and NM acknowledge support by the project REGENERATE funded by the Hellenic Foundation for Research and Innovation (HFRI) through the General Secretariat for Research and Technology (GSRT), Project No. 3232.





analyses in Serbia were supported by the EC Horizon Europe research and innovation program under grant
agreement GA 101060170 (WeBaSOOP). Analyses at ISAC-CNR were supported by the project PER-ACTRIS-IT,
funded by the Ministry of Education and Research (MIUR), Action II.1 of PON Research and Innovation 20142020. This study was also supported by the PROCOPE mobility program (0185-DEU-23-0008 LG1) for grant
0185-DEU-23-0008 LG1 to Eduardo Souza (ACD, TROPOS). PS acknowledges support from the Hazardous Air
Pollutant Laboratory of Environment and Climate Change Canada. BU and MK acknowledge funding by the
Swiss National Science Foundation, grant 200021_192192/1. IM received funding from the National Institute
for Health Research (NIHR) Health Protection Research Unit in Environmental Exposures and Health, a
partnership between the UK Health Security Agency and Imperial College, with IM receiving additional funding
from the MRC Centre for Environment and Health, which is funded by the Medical Research Council
(MR/S0196669/1, 2019-2024).

750 We also acknowledge Mark Diks (TNO), Alexandre Barth and Julian Resch (UNIBAS), Laurent Alleman (IMT-751 Nord) and Stephan Houdier (IGE) for providing help on the deployment of this ILC.

752 Data availability

753 All OP data is published in the SI.

754 Author contributions

GU designed the study and supervised the OP tests for the development of the simplified DTT RI-URBANS protocol. CR, TM, RE and PAD tested the different steps and prepared the samples and logistics of the intercomparison exercise. GH, R.M.H, A.N, I. M., K. B, N.M and G.U. evaluated and decided the samples to be compared and the protocol content. JPP and FC performed the data analysis and evaluated the performance of each group. PAD and AM developed the multiple regression model. PAD processed the data of the study and wrote the paper together with JLJ and GU before a first review by the RI-Urbans members. All authors participated in the interlaboratory comparison. All authors reviewed and edited the manuscript.

762 Competing interests

763 At least one of the (co)-authors is a member of the editorial board of Atmospheric Measurement Techniques.

764

765 7. References

766 767	Alkoussa, S., Hulo, S., Courcot, D., Billet, S., and Martin, P. J.: Extracellular vesicles as actors in the air pollution related cardiopulmonary diseases, Crit. Rev. Toxicol., 50, 402–423,
768	https://doi.org/10.1080/10408444.2020.1763252, 2020.
769	Ayres, J. G., Borm, P., Cassee, F. R., Castranova, V., Donaldson, K., Ghio, A., Harrison, R. M., Hider, R.,
770	Kelly, F., Kooter, I. M., Marano, F., Maynard, R. L., Mudway, I., Nel, A., Sioutas, C., Smith, S., Baeza-
771	Squiban, A., Cho, A., Duggan, S., and Froines, J.: Evaluating the toxicity of airborne particulate

matter and nanoparticles by measuring oxidative stress potential - A workshop report and





consensus statement, Inhal. Toxicol., 20, 75–99, https://doi.org/10.1080/08958370701665517,
2008.

- Bates, J. T., Weber, R. J., Abrams, J., Verma, V., Fang, T., Klein, M., Strickland, M. J., Sarnat, S. E.,
 Chang, H. H., Mulholland, J. A., Tolbert, P. E., and Russell, A. G.: Reactive Oxygen Species Generation
- Linked to Sources of Atmospheric Particulate Matter and Cardiorespiratory Effects, Environ. Sci.
 Tashaal 40, 12(05, 12(12) https://doi.org/10.1021/cas.act.Eb020(7, 2015)
- 778 Technol., 49, 13605–13612, https://doi.org/10.1021/acs.est.5b02967, 2015.

Bates, J. T., Fang, T., Verma, V., Zeng, L., Weber, R. J., Tolbert, P. E., Abrams, J. Y., Sarnat, S. E., Klein,
M., Mulholland, J. A., and Russell, A. G.: Review of Acellular Assays of Ambient Particulate Matter
Oxidative Potential: Methods and Relationships with Composition, Sources, and Health Effects,
Environ. Sci. Technol., 53, 4003–4019, https://doi.org/10.1021/acs.est.8b03430, 2019.

783 Bhattu, D., Tripathi, S. N., Bhowmik, H. S., Moschos, V., Lee, C. P., Rauber, M., Salazar, G., Abbaszade, 784 G., Cui, T., Slowik, J. G., Vats, P., Mishra, S., Lalchandani, V., Satish, R., Rai, P., Casotto, R., Tobler, A., 785 Kumar, V., Hao, Y., Qi, L., Khare, P., Manousakas, M. I., Wang, Q., Han, Y., Tian, J., Darfeuil, S., 786 Minguillon, M. C., Hueglin, C., Conil, S., Rastogi, N., Srivastava, A. K., Ganguly, D., Bjelic, S., Canonaco, 787 F., Schnelle-Kreis, J., Dominutti, P. A., Jaffrezo, J.-L., Szidat, S., Chen, Y., Cao, J., Baltensperger, U., 788 Uzu, G., Daellenbach, K. R., El Haddad, I., and Prévôt, A. S. H.: Local incomplete combustion 789 emissions define the PM2.5 oxidative potential in Northern India, Nat. Commun., 15, 3517, 790 https://doi.org/10.1038/s41467-024-47785-5, 2024.

- Borlaza, L. J. S., Uzu, G., Ouidir, M., Lyon-Caen, S., Marsal, A., Weber, S., Siroux, V., Lepeule, J.,
 Boudier, A., Jaffrezo, J.-L., Slama, R., Lyon-Caen, S., Siroux, V., Lepeule, J., Philippat, C., Slama, R.,
 Hofmann, P., Hullo, E., Llerena, C., Quentin, J., Pin, I., Eyriey, E., Licinia, A., Vellement, A., Morin, X.,
 Morlot, A., and the SEPAGES cohort study group: Personal exposure to PM2.5 oxidative potential
 and its association to birth outcomes, J. Expo. Sci. Environ. Epidemiol.,
- 796 https://doi.org/10.1038/s41370-022-00487-w, 2022.

Calas, A., Uzu, G., Martins, J. M. F., Voisin, D., Spadini, L., Lacroix, T., and Jaffrezo, J.-L.: The
importance of simulated lung fluid (SLF) extractions for a more relevant evaluation of the
oxidative potential of particulate matter, Sci. Rep., 7, 11617, https://doi.org/10.1038/s41598017-11979-3, 2017.

Calas, A., Uzu, G., Kelly, F. J., Houdier, S., Martins, J. M. F., Thomas, F., Molton, F., Charron, A.,
Dunster, C., Oliete, A., Jacob, V., Besombes, J.-L., Chevrier, F., and Jaffrezo, J.-L.: Comparison
between five acellular oxidative potential measurement assays performed with detailed
chemistry on PM10 samples from the city of Chamonix (France), Atmospheric Chem. Phys., 18,
7863–7875, https://doi.org/10.5194/acp-18-7863-2018, 2018.

- Cassee, F. R., Héroux, M.-E., Gerlofs-Nijland, M. E., and Kelly, F. J.: Particulate matter beyond mass:
 recent health evidence on the role of fractions, chemical constituents and sources of emission,
 Inhal. Toxicol., 25, 802–812, https://doi.org/10.3109/08958378.2013.850127, 2013.
- 809 Charrier, J. G. and Anastasio, C.: On dithiothreitol (DTT) as a measure of oxidative potential for
 810 ambient particles: evidence for the importance of soluble transition metals, Atmospheric Chem.
 811 Phys., 12, 9321–9333, https://doi.org/10.5194/acp-12-9321-2012, 2012.
- 812 Charrier, J. G., Richards-Henderson, N. K., Bein, K. J., McFall, A. S., Wexler, A. S., and Anastasio, C.:
- 813 Oxidant production from source-oriented particulate matter Part 1: Oxidative potential using
- the dithiothreitol (DTT) assay, Atmospheric Chem. Phys., 15, 2327–2340,
- 815 https://doi.org/10.5194/acp-15-2327-2015, 2015.





- 816 Charrier, J. G., McFall, A. S., Vu, K. K.-T., Baroi, J., Olea, C., Hasson, A., and Anastasio, C.: A bias in the
- 817 "mass-normalized" DTT response – An effect of non-linear concentration-response curves for
- 818 copper and manganese, Atmos. Environ., 144, 325-334, 819
- https://doi.org/10.1016/j.atmosenv.2016.08.071, 2016.
- 820 Cho, A. K., Sioutas, C., Miguel, A. H., Kumagai, Y., Schmitz, D. A., Singh, M., Eiguren-Fernandez, A., 821 and Froines, J. R.: Redox activity of airborne particulate matter at different sites in the Los Angeles
- 822 Basin, Environ. Res., 99, 40-47, https://doi.org/10.1016/j.envres.2005.01.003, 2005.
- 823 Council of the European Union: Proposal for a Directive of the European Parliament and of the 824 Council on ambient air quality and cleaner air for Europe, Brussels, 2024.
- 825 Dabass, A., Talbott, E. O., Rager, J. R., Marsh, G. M., Venkat, A., Holguin, F., and Sharma, R. K.: 826 Systemic inflammatory markers associated with cardiovascular disease and acute and chronic 827 exposure to fine particulate matter air pollution (PM2.5) among US NHANES adults with 828 metabolic syndrome, Environ. Res., 161, 485-491, https://doi.org/10.1016/j.envres.2017.11.042, 829 2018.
- 830 Daellenbach, K. R., Stefenelli, G., Bozzetti, C., Vlachou, A., Fermo, P., Gonzalez, R., Piazzalunga, A., 831 Colombi, C., Canonaco, F., Hueglin, C., Kasper-Giebl, A., Jaffrezo, J. L., Bianchi, F., Slowik, J. G.,
- 832 Baltensperger, U., El-Haddad, I., and Prévôt, A. S. H.: Long-term chemical analysis and organic
- 833 aerosol source apportionment at nine sites in central Europe: Source identification and
- 834 uncertainty assessment, Atmospheric Chem. Phys., 17, 13265-13282,
- 835 https://doi.org/10.5194/acp-17-13265-2017, 2017.
- 836 Danehy, J. P., Elia, V. J., and Lavelle, C. J.: Alkaline decomposition of organic disulfides. IV. Limitation on the use of Ellman's reagent. 2,2'-Dinitro-5,5'-dithiodibenzoic acid, J. Org. Chem., 36, 837 838 1003–1005, https://doi.org/10.1021/jo00806a036, 1971.
- 839 Dominutti, P. A., Borlaza, L. J. S., Sauvain, J.-J., Ngoc Thuy, V. D., Houdier, S., Suarez, G., Jaffrezo, J.-L., 840 Tobin, S., Trébuchon, C., Socquet, S., Moussu, E., Mary, G., and Uzu, G.: Source apportionment of 841 oxidative potential depends on the choice of the assay: insights into 5 protocols comparison and 842 implications for mitigation measures, Environ. Sci. Atmospheres, 10.1039.D3EA00007A, 843 https://doi.org/10.1039/D3EA00007A, 2023.
- 844 Donaldson, K., Stone, V., Seaton, A., and MacNee, W.: Ambient particle inhalation and the 845 cardiovascular system: Potential mechanisms, Environ. Health Perspect., 109, 523–527, https://doi.org/10.1289/ehp.01109s4523, 2001. 846
- 847 Frezzini, M. A., De Francesco, N., Massimi, L., and Canepari, S.: Effects of operating conditions on PM oxidative potential assays, Atmos. Environ., 268, 118802, 848
- 849 https://doi.org/10.1016/j.atmosenv.2021.118802, 2022.
- 850 Gao, D., Mulholland, J. A., Russell, A. G., and Weber, R. J.: Characterization of water-insoluble 851 oxidative potential of PM2.5 using the dithiothreitol assay, Atmos. Environ., 224, 117327, 852 https://doi.org/10.1016/j.atmosenv.2020.117327, 2020.
- 853 Gao, Y., Wang, P., Chu, Y., Kang, F., Cheng, Y., Repo, E., Feng, M., Yu, X., and Zeng, H.: Redox property of coordinated iron ion enables activation of O2 via in-situ generated H2O2 and additionally
- 854 855 added H2O2 in EDTA-chelated Fenton reaction, Water Res., 248, 120826,
- https://doi.org/10.1016/j.watres.2023.120826, 2024. 856





857 Grandesso, E., Kowalewski, K., and Perez Ballesta, P.: First EC-IRC PAHs inter-laboratory 858 comparison on PM10 quartz filters., European Commission. Joint Research Centre. Institute for 859 Environment and Sustainability., Luxembourg, 2012. 860 Guascito, M. R., Lionetto, M. G., Mazzotta, F., Conte, M., Giordano, M. E., Caricato, R., De Bartolomeo, 861 A. R., Dinoi, A., Cesari, D., Merico, E., Mazzotta, L., and Contini, D.: Characterisation of the 862 correlations between oxidative potential and in vitro biological effects of PM10 at three sites in 863 the central Mediterranean, J. Hazard. Mater., 448, 130872, 864 https://doi.org/10.1016/j.jhazmat.2023.130872, 2023. 865 Hart, J. E., Liao, X., Hong, B., Puett, R. C., Yanosky, J. D., Suh, H., Kioumourtzoglou, M.-A., Spiegelman, D., and Laden, F.: The association of long-term exposure to PM2.5 on all-cause mortality in the 866 867 Nurses' Health Study and the impact of measurement-error correction, Environ. Health, 14, 38, https://doi.org/10.1186/s12940-015-0027-6, 2015. 868 869 He, L. and Zhang, J. (Jim): Particulate matter (PM) oxidative potential: Measurement methods and 870 links to PM physicochemical characteristics and health effects, Crit. Rev. Environ. Sci. Technol., 53, 177-197, https://doi.org/10.1080/10643389.2022.2050148, 2022. 871 872 Huang, H., Yang, L., Liu, Y., Dong, G.-H., Chen, L., Li, S., Guo, Y., Xie, B., and Chen, G.: Long-term 873 ambient air pollution exposure and DNA methylation of peripheral brain-derived neurotrophic 874 factor promoter, Ecotoxicol. Environ. Saf., 244, 114061, https://doi.org/10.1016/j.ecoenv.2022.114061, 2022. 875 876 Kelly, F. J. and Fussell, J. C.: Role of oxidative stress in cardiovascular disease outcomes following 877 exposure to ambient air pollution, Free Radic. Biol. Med., 110, 345-367, https://doi.org/10.1016/j.freeradbiomed.2017.06.019, 2017. 878 879 Kumagai, Y., Koide, S., Taguchi, K., Endo, A., Nakai, Y., Yoshikawa, T., and Shimojo, N.: Oxidation of proximal protein sulfhydryls by phenanthraquinone, a component of diesel exhaust particles, 880 881 Chem. Res. Toxicol., 15, 483-489, https://doi.org/10.1021/tx0100993, 2002. 882 Laden, F., Schwartz, J., Speizer, F. E., and Dockery, D. W.: Reduction in Fine Particulate Air 883 Pollution and Mortality, Am. J. Respir. Crit. Care Med., 173, 667-672, 884 https://doi.org/10.1164/rccm.200503-4430C, 2006. 885 Lepeule, J., Laden, F., Dockery, D., and Schwartz, J.: Chronic Exposure to Fine Particles and 886 Mortality: An Extended Follow-up of the Harvard Six Cities Study from 1974 to 2009, Environ. 887 Health Perspect., 120, 965–970, https://doi.org/10.1289/ehp.1104660, 2012. 888 Li, N., Sioutas, C., Cho, A., Schmitz, D., Misra, C., Sempf, J., Wang, M., Oberley, T., Froines, J., and Nel, 889 A.: Ultrafine particulate pollutants induce oxidative stress and mitochondrial damage., Environ. 890 Health Perspect., 111, 455–460, https://doi.org/10.1289/ehp.6000, 2003. 891 Li, N., Xia, T., and Nel, A. E.: The role of oxidative stress in ambient particulate matter-induced lung 892 diseases and its implications in the toxicity of engineered nanoparticles, Free Radic. Biol. Med., 44, 893 1689–1699, https://doi.org/10.1016/j.freeradbiomed.2008.01.028, 2008. 894 Li, Q., Wyatt, A., and Kamens, R. M.: Oxidant generation and toxicity enhancement of aged-diesel 895 exhaust, Atmos. Environ., 43, 1037-1042, https://doi.org/10.1016/j.atmosenv.2008.11.018, 896 2009. 897 Marsal, A., Slama, R., Lyon-Caen, S., Borlaza, L. J. S., Jaffrezo, J.-L., Boudier, A., Darfeuil, S., Elazzouzi, 898 R., Gioria, Y., Lepeule, J., Chartier, R., Pin, I., Quentin, J., Bayat, S., Uzu, G., Siroux, V., and the





- SEPAGES cohort study group: Prenatal Exposure to PM2.5 Oxidative Potential and Lung Function
 in Infants and Preschool- Age Children: A Prospective Study, Environ. Health Perspect., 131,
 017004, https://doi.org/10.1289/EHP11155, 2023.
- 902 Miljevic, B., Hedayat, F., Stevanovic, S., Fairfull-Smith, K. E., Bottle, S. E., and Ristovski, Z. D.: To
- Sonicate or Not to Sonicate PM Filters: Reactive Oxygen Species Generation Upon Ultrasonic
 Irradiation, Aerosol Sci. Technol., 48, 1276–1284,
- 905 https://doi.org/10.1080/02786826.2014.981330, 2014.
- Molina, C., Andrade, C., Manzano, C. A., Richard Toro, A., Verma, V., and Leiva-Guzmán, M. A.:
 Dithiothreitol-based oxidative potential for airborne particulate matter: an estimation of the
 associated uncertainty, Environ. Sci. Pollut. Res., 27, 29672–29680,
- 909 https://doi.org/10.1007/s11356-020-09508-3, 2020.
- Nicholson, S., Baccarelli, A., and Prada, D.: Role of brain extracellular vesicles in air pollution related cognitive impairment and neurodegeneration, Environ. Res., 204, 112316,
 https://doi.org/10.1016/j.apures.2021.112216.2022
- 912 https://doi.org/10.1016/j.envres.2021.112316, 2022.
- Park, M., Joo, H. S., Lee, K., Jang, M., Kim, S. D., Kim, I., Borlaza, L. J. S., Lim, H., Shin, H., Chung, K. H.,
 Choi, Y.-H., Park, S. G., Bae, M.-S., Lee, J., Song, H., and Park, K.: Differential toxicities of fine
 particulate matters from various sources, Sci. Rep., 8, 17007, https://doi.org/10.1038/s41598018-35398-0, 2018.
- Sauvain, J. J., Deslarzes, S., and Riediker, M.: Nanoparticle reactivity toward dithiothreitol,
 Nanotoxicology, 2, 121–129, https://doi.org/10.1080/17435390802245716, 2008.
- Shahpoury, P., Harner, T., Lammel, G., Lelieveld, S., Tong, H., and Wilson, J.: Development of an
 antioxidant assay to study oxidative potential of airborne particulate matter, Atmospheric Meas.
 Tech., 12, 6529–6539, https://doi.org/10.5194/amt-12-6529-2019, 2019.
- Shahpoury, P., Zhang, Z. W., Filippi, A., Hildmann, S., Lelieveld, S., Mashtakov, B., Patel, B. R., Traub,
 A., Umbrio, D., Wietzoreck, M., Wilson, J., Berkemeier, T., Celo, V., Dabek-Zlotorzynska, E., Evans,
 G., Harner, T., Kerman, K., Lammel, G., Noroozifar, M., Pöschl, U., and Tong, H.: Inter-comparison of
- 925 oxidative potential metrics for airborne particles identifies differences between acellular
- 926 chemical assays, Atmospheric Pollut. Res., 13, 101596,
- 927 https://doi.org/10.1016/j.apr.2022.101596, 2022.
- 928 Shankar, K. and Mehendale, H. M.: Oxidative Stress, Third Edit., Elsevier, 735–737 pp.,
 929 https://doi.org/10.1016/B978-0-12-386454-3.00345-6, 2014.
- Sies, H.: On the history of oxidative stress: Concept and some aspects of current development,
 Curr. Opin. Toxicol., 7, 122–126, https://doi.org/10.1016/j.cotox.2018.01.002, 2018.
- 932 Thbayh, D. K., Palusiak, M., Viskolcz, B., and Fiser, B.: Comparative study of the antioxidant
- 933 capability of EDTA and Irganox, Heliyon, 9, e16064,
- 934 https://doi.org/10.1016/j.heliyon.2023.e16064, 2023.
- 935 Uzu, G., Sauvain, J. J., Baeza-Squiban, A., Riediker, M., Hohl, M. S. S., Val, S., Tack, K., Denys, S.,
- 936 Pradère, P., and Dumat, C.: In vitro assessment of the pulmonary toxicity and gastric availability of
- 937 lead-rich particles from a lead recycling plant, Environ. Sci. Technol., 45, 7888–7895,
- 938 https://doi.org/10.1021/es200374c, 2011.





- Verlhac, S., Leoz, E., Albinet, A., Cabillic, J., Lalere, B., and Fallot, C.: Comparaison interlaboratoires
 sur l'analyse des HAP dans l'air ambiant, Laboratoire Central de Surveillance de la Qualité de l'Air,
 2014.
- Weichenthal, S., Lavigne, E., Evans, G., Pollitt, K., and Burnett, R. T.: Ambient PM2.5 and risk of
 emergency room visits for myocardial infarction: impact of regional PM2.5 oxidative potential: a
 case-crossover study, Environ. Health, 15, 46, https://doi.org/10.1186/s12940-016-0129-9,
 2016a.
- Weichenthal, S., Crouse, D. L., Pinault, L., Godri-Pollitt, K., Lavigne, E., Evans, G., van Donkelaar, A.,
 Martin, R. V., and Burnett, R. T.: Oxidative burden of fine particulate air pollution and risk of
 cause-specific mortality in the Canadian Census Health and Environment Cohort (CanCHEC),
 Environ. Res., 146, 92–99, https://doi.org/10.1016/j.envres.2015.12.013, 2016b.
- 949 Environ. Res., 146, 92–99, https://doi.org/10.1016/j.envres.2015.12.013, 2016b.
- Weichenthal, S., Shekarrizfard, M., Traub, A., Kulka, R., Al-Rijleh, K., Anowar, S., Evans, G., and
 Hatzopoulou, M.: Within-City Spatial Variations in Multiple Measures of PM 2.5 Oxidative
- 952 Potential in Toronto, Canada, Environ. Sci. Technol., 53, 2799–2810,
- 953 https://doi.org/10.1021/acs.est.8b05543, 2019.

Weichenthal, S. A., Lavigne, E., Evans, G. J., Godri Pollitt, K. J., and Burnett, R. T.: Fine particulate
matter and emergency room visits for respiratory illness: Effect modification by oxidative
potential, Am. J. Respir. Crit. Care Med., 194, 577–586, https://doi.org/10.1164/rccm.20151224340C, 2016c.

- WHO: Review of evidence on health aspects of air pollution REVIHAAP Project- Technical Report,
 World Health Organization, Regional Office for Europe, 2021a.
- WHO: WHO Global air quality guidelines. Particulate matter (PM2.5 and PM10), ozone, nitrogen
 dioxide, sulfur dioxide and carbon monoxide, World Health Organization, 2021b.
- Wilker, E. H., Osman, M., and Weisskopf, M. G.: Ambient air pollution and clinical dementia:
 systematic review and meta-analysis, BMJ, e071620, https://doi.org/10.1136/bmj-2022-071620,
 2023.
- Zare Sakhvidi, M. J., Yang, J., Lequy, E., Chen, J., De Hoogh, K., Letellier, N., Mortamais, M., Ozguler,
 A., Vienneau, D., Zins, M., Goldberg, M., Berr, C., and Jacquemin, B.: Outdoor air pollution exposure
 and cognitive performance: findings from the enrolment phase of the CONSTANCES cohort,
 Lancet Planet. Health, 6, e219–e229, https://doi.org/10.1016/S2542-5196(22)00001-8, 2022.
- 969
- 970
- 971