



HAL
open science

IMPACT OF INTRAORAL SCANNER SLEEVES ON STL FILE ACCURACY: A METROLOGICAL APPROACH

Max Troizier Cheyne, Lebon Nicolas, Floriane Laverne

► **To cite this version:**

Max Troizier Cheyne, Lebon Nicolas, Floriane Laverne. IMPACT OF INTRAORAL SCANNER SLEEVES ON STL FILE ACCURACY: A METROLOGICAL APPROACH. 1ère Journée de Recherche en Odontologie, Jul 2023, Montrouge, France. , 2023. hal-04672964

HAL Id: hal-04672964

<https://hal.science/hal-04672964>

Submitted on 19 Aug 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

IMPACT OF INTRAORAL SCANNER SLEEVES ON STL FILE ACCURACY: A METROLOGICAL APPROACH

Max Troizier-Cheyne, Nicolas Lebon, Floriane Laverne

- Most accuracy studies rely on superimposition of a reference scan and a test scan using local best-fit algorithms
- Errors minimized, especially across larger objects
- Errors are always relative, and measure is not objective

Objective 1 Create a protocol to assess accuracy using an objective metrological approach.


- Intraoral scanner manufacturers have introduced different protective sleeves to limit cross-contamination
- Scanning sleeves must maintain scanning performance while satisfying rigorous cross-contamination protocols

Objective 2 Compare the quality of STL files produced using different scanning sleeves


Reference Object

8 mm Al₂O₃ Sphere
Grade 10 = max diameter deviation of 0.25 μm

3D printed support structure with spherical landmarks to facilitate scanning



Mesh Processing



- Creation of reference coordinate system in GOM Inspect
- Deletion of landmarks and support structure
- Parallel plane fitted to the surface and mesh deleted below a parallel plan 0.2 mm above it
- Isolated sphere
- Gaussian fitted sphere

Calculations

Trueness : Difference between the average of numerous tests and a true or accepted reference value


$$T = \bar{d} - d_R$$

Precision : Closeness of agreement between independent tests

$$P = \sqrt{\frac{1}{n-1} \sum_{i=1}^{n=24} (d_i - \bar{d})^2}$$

Digitizing Noise : Random artifacts of signal that are introduced during the recording

$$\sigma_N = \sqrt{\frac{1}{m-1} \sum_{j=1}^m (r_j - \bar{r})^2}$$



Materials and Methods

- Primescan intraoral scanner (Dentsply Sirona)
- 24 scans of the reference object per sleeve
- Mesh processing and calculations performed according to the metrological protocol



Results

Trueness

Precision

Digitizing Noise

	Standard (SS)	Autoclavable (AS)	Disposable (DS)
Trueness	-9.72 μm	2.55 μm	-9.66 μm
Precision	6.71 μm	7.91 μm	7.17 μm
Digitizing Noise	7.31 μm	6.37 μm	8.72 μm

- **Trueness** : Significant differences between AS-DS ($p < 0.001$) and AS-SS ($p < 0.001$)
- **Precision** : No significant differences between the sleeves for ($p = 0.623$)
- **Digitizing noise** : Significant differences between DS-AS ($p < 0.001$), DS-SS ($p = 0.001$), and SS-AS ($p = 0.001$)

Conclusion

- Significant differences between sleeves
- Errors appear to be within clinically acceptable limits
- Metrological approach is valid
- Further studies needed using reference objects with more measurable dimensions

