

A micro-electrochemical sensor based on a nanostructured carbon paste electrode as an original tool for selective detection of ultra-traces of 17α –ethinylestradiol in environmental water samples and in biologic fluids

Siham Amra, Saliha Bourouina Bacha, Mustapha Bourouina, Didier Hauchard

▶ To cite this version:

Siham Amra, Saliha Bourouina Bacha, Mustapha Bourouina, Didier Hauchard. A microelectrochemical sensor based on a nanostructured carbon paste electrode as an original tool for selective detection of ultra-traces of 17α –ethinylestradiol in environmental water samples and in biologic fluids. Microchemical Journal, 2024, 202, pp.110753. 10.1016/j.microc.2024.110753. hal-04690432

HAL Id: hal-04690432 https://hal.science/hal-04690432v1

Submitted on 10 Oct 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

A micro-electrochemical sensor based on a nanostructured carbon paste electrode as an original tool for selective detection of ultra-traces of 17α –ethiny-lestradiol in environmental water samples and in biologic fluids

Siham Amra, Saliha Bourouina-Bacha, Mustapha Bourouina, Didier Hauchard

PII:	S0026-265X(24)00865-8
DOI:	https://doi.org/10.1016/j.microc.2024.110753
Reference:	MICROC 110753
To appear in:	Microchemical Journal
Received Date:	2 March 2024
Revised Date:	4 May 2024
Accepted Date:	12 May 2024



Please cite this article as: S. Amra, S. Bourouina-Bacha, M. Bourouina, D. Hauchard, A micro-electrochemical sensor based on a nanostructured carbon paste electrode as an original tool for selective detection of ultra-traces of 17α –ethinylestradiol in environmental water samples and in biologic fluids, *Microchemical Journal* (2024), doi: https://doi.org/10.1016/j.microc.2024.110753

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

A micro-electrochemical sensor based on a nanostructured carbon paste electrode as an original tool for selective detection of ultra-traces of 17α – ethinylestradiol in environmental water samples and in biologic fluids.

Siham AMRA^{*ab*}, Saliha BOUROUINA-BACHA^{*c*}, Mustapha BOUROUINA^{*b*}, Didier HAUCHARD^{*a*}

^aUniversité Rennes, Ecole Nationale Supérieure de Chimie de Rennes, CNRS, ISCR (Institut des Sciences Chimiques de Rennes) – UMR 6226, F-35000 Rennes, France.

^bUniversité de Bejaia, Faculté des Sciences Exactes, Département de Chimie, 06000 Bejaia, Algeria.

^cDépartement de Génie des Procédés, Faculté de Technologie, Université de Bejaia, 06000 Bejaia, Algeria.

Abstract

A new electrochemical micro-sensor has been designed for the sensitive detection of ultratraces of 17α – ethinylestradiol (EE₂). A carbon paste cavity micro-electrode modified by carbon black nanoparticles (GC-nCB)-µCPE was used with an electrochemical preconcentration (EP) step of EE₂ on (GC-nCB)-µCPE at 0.0 V during 300 s in a 0.01 mol.L⁻¹ phosphate buffer solution (pH = 7.4) before voltammetric determination. The results demonstrate that two line calibrations were obtained in the concentration ranges of 0.005 to 0.2 µmol.L⁻¹ and 0.2 to 1 µmol.L⁻¹. The detection and quantification limits are estimated at 1.4 nmol.L⁻¹ and 4.9 nmol.L⁻¹, respectively. Moreover, a high sensitivity of 4.16 µA.mol⁻¹.L and good selectivity for the quantification of EE₂ in the presence of different interfering organic and inorganic compounds were observed. Based on the results (satisfactory recovery (less than 3.5%) and relationship standard deviation (less than 2.5%)), the designed microsensor can be successfully applied to the analysis of EE₂ in environmental water samples and in biologic fluids. The modified microsensor presented an efficient and inexpensive tool for the quantification of EE₂.

Keywords: Voltammetric micro-sensor, cavity micro-electrode, carbon paste, nanostructured materials, 17α -éthinylestradiol.

1. Introduction

In this work, It was chosen to quantify a synthetic estrogen, 17α -ethinylestradiol (EE₂) (Scheme 1), an active pharmaceutical compound introduced into aquatic environments by humans and animal excretions [1]. Due to its very high disruptive power, EE₂ is described as one of the most dangerous endocrine disruptors [2, 3]. This is why its quantification and effects on flora and fauna have been the subject of numerous studies in recent years [4-8]. Face with the necessity to dispose of quantification methods to assume the monitoring of water and wastewater, electrochemical sensors appear as reliable interesting detection technologies due to their capacity to be sensitive, selective, fast, inexpensive and suitable for on-site analysis. With conventional electrodes, it is difficult to quantify EE₂ at low concentrations by voltammetry due to the obtaining of a broad and poorly-defined irreversible EE₂ oxidation peak [9, 10]. Therefore, the design and manufacture of an efficient electrochemical microsensor for EE2 trace analysis constitute a major challenge.

Recently, nanotechnology has become very popular in the sensor field. Indeed, the incorporation of nano-sized materials into the electrode matrix can have beneficial effects on the performance of sensors [11, 12]. It has been shown that nano-sized materials possess novel and very interesting physical and chemical properties. In this context, carbon black (nCB) is a nanostructured material widely used in several industrial processes [13]. This nanomaterial provides excellent properties, including a large specific surface area and high thermal and electrical conductivity [14]. Moreover, nCB is unique in its ability to intercalate ions between graphite layers [15], which affect the conduction between the particles and the surface reaction [16]. In particular, the structural defects formed during the synthesis or functionalization processes can create several functional groups useful for the immobilization of various species, such as enzymes [17], genetic materials and antibodies [18] and nanoparticles [19]. Thanks to its highly interesting structural and electrical/electronic characteristics, nCB is an economically attractive material for the design and development of electrochemical devices [20]. Moreover, due to these fascinating properties, carbon paste electrode (CPE) has gained a great theoretical and practical reputation in the electroanalytical field [21, 22]. CPE offers a number of unique and valuable electrochemical performances, providing: rich and biocompatible surfaces, high conductivity, low background current, low ohmic resistance, an extended potential window, long-term stability which can last up to four years storage, and cheap repair [23, 24]. Furthermore, carbon paste electrodes can be modified by adding various types of modifiers (such as multi-walled carbon nanotube, graphène,...) to increase sensitivity. [25-29]. We therefore focused our work on the development of a microsensor for the analysis of EE₂ using a carbon paste modified with carbon black nanoparticles (GC-nCB)-CP as electrode material.

In recent years, several studies presented the use of cavity microelectrode (CME) successfully as electrochemical sensors [30, 31]. Indeed, the CME is an original tool for the study of powdered materials, essentially characterized by its rapid and ease of implementation, without resorting to the use of binder. It offers the possibility of carrying out studies using very small quantities of materials, thus reducing the limitations of the ohmic drop and capacitive effects on voltammograms even at high potential scan rates. Moreover, comparative studies of the CME with the usual carbon paste electrodes (millimeter size) have shown for the CME better discrimination of the peaks on voltammograms with an increase in the sensitivity of detection [32, 33].

Thanks to these advantages linked to the properties of CME, the objective of this study was to design an innovative and ultra-sensitive electrochemical micro-sensor for the analysis of ultra-trace amounts of EE₂ in water samples and urine. The development of this microsensor is based on the carbon paste cavity microelectrode modified by carbon black nanoparticles ((GC-nCB)- μ CPE). The voltammetry technique was applied to evaluate the electroanalytical performance of the prepared micro-sensor. As far as is known, ((GC-nCB)- μ CPE) has never been described before for the determination of EE₂ and was applied to EE₂ determination in real wastewater sample and in synthetic urine.

Scheme 1: Chemical structure of EE₂



2. Materials and methods

2.1. Reagents

All chemical reagents used in this study were of analytical grade (the highest purity available). Graphite powder ($\leq 4 \mu m$; 99,99 %) from Aldrich, France, carbon black ($\leq 100 nm$; 99,95%) from Cabot Corporation (Vulcan XC-72, average diameter 30 nm and SBET 254 m²/g), France. Ethanol (99%), 17 α -ethinylestradiol (analytical standard; 99.9%) were purchased from Sigma-Aldrich, France, and phosphate buffer EMSURE for analysis (> 98 %) from Supelco, France. The pH of the solutions was adjusted using NaOH (98 %, Sigma-Aldrich, France) and sulfuric acid (98 %, for analysis EMSURE, France). All aqueous solutions were prepared with ultrapure water at 18.2 M Ω cm (20 ° C) from the Millipore Direct-Q Water system. Stock solutions of EE₂ (0.01 M) were prepared with ethanol and subsequent dilutions were prepared daily with ultrapure water and stored at 4 ° C in the dark. All reagents were used as received, and all experiments in this study were performed at room temperature (20 ± 1 ° C).

2.2. Apparatus

All electrochemical techniques were carried out using a PGSTAT 20 (Metrohm) potentiostat/galvanostat in combination with NOVA 2.1 software. The measurements were carried out using a classical electrochemical cell composed of three electrodes: a carbon paste cavity microelectrode (CME) (microcavity: 50 μ m diameter; 32 μ m depth) (WE), an auxiliary Pt wire electrode (AE) and an Ag/AgCl (3M KCl) reference electrode (RE). The pH of solutions was measured using a pH- meter: Mettler Toledo FE20/EL20. Scanning electron microscopy (SEM) images were obtained using a JEOL JSM-7100F, EDS detector Oxford Instrument 50 mm² from SCANMAT (CMEBA plateform) at University of Rennes, France.

2.3. Preparation of carbon paste micro-sensor

Carbon black modified carbon paste composite microelectrodes (GC-nCB)- μ CPE were prepared by carefully mixing graphite carbon powder (GC) and carbon black nanoparticles (nCB). Indeed, in order to optimize the electrode matrix composition, various mass ratio of graphite and nCB have been used, respectively 40 to 100% and 0 to 60% for the preparation of the (GC-nCB)- μ CPE. All pastes were prepared manually as described in the literature [**34**] and consisted of mixing the ingredients in an agate mortar with a pestle until a homogeneous paste was obtained (during 30 min). The duration of preparation plays an essential role in order to have reproducible pastes. These different pastes were introduced and compacted into CME is essential and can be done in an ultrasonic bath by immersing the CME for 5 minutes successively in the following solutions: a nitric acid solution (HNO₃; 2M); then, in a solution of hydrogen peroxide (H₂O₂; 10%) and finally, in ultrapure water. Next, a control procedure was used to verify the emptying of the cavity that consists to plot cyclic voltammograms of the CME in an acidic medium (H₂SO₄ 0.5 M) allowing the observation of typical comportment of platinum electrode. The steps for the preparation of (GC-nCB)- μ CPE are illustrated in Figure 1.

Insert Fig 1

2.4. Real samples preparation

In this work, the aim was to study the micro-sensor developed in two local water samples (wastewater from WWTPs treated before being released into with a membrane process (WWTPs-M) and wastewater from WWTPs treated with a decantation process (WWTPs-D)) and in urine using synthetic urine. The synthetic urine used in this study was synthesized by using the method reported by N. Sarigul et al [35]. The individual samples were filtered through a syringe filter (0.22 μ m) and diluted for 10 times in the case of urine samples and spiked with tens or hundreds of nanomoles of EE₂ concentration and stored in the dark at 4 °C. The EE₂ concentration measurements were performed under optimal conditions using the standard addition method.

3. Results and discussion

3.1 Microscopic characterization of GC, nCB and (GC- nCB) - CP

The morphologies of the carbon graphite (CG) of 4 μ m in size, the carbon black (nCB) and their mixtures as carbon paste (GC-nCB) -CP in the modified microelectrode were analyzed by scanning electron microscopy (SEM). Figure 2 shows the SEM images obtained for the GC (A), nCB (B) and (C) and (GC-nCB)–CP (D) mixtures used for the preparation of the modified cavity microelectrodes. We can see on the SEM image (A) the typical morphology of the graphite particles, which appear in the form of dispersed sheets of size less than 4 μ m. (B) and (C) SEM images presents carbon black nanoparticles (nCB) that appears as spongy aggregates of small spherical particles with nanometric size (<100 nm). For the prepared mixture of GC and nCB particles (50/50, %w), the SEM image (D) shows that both materials are evenly dispersed in the GC sheets coated with nCB particles, avoiding graphite platelet stacking, promoting analyte accessibility in their surfaces and improving charge transfer in the paste. Globally, the (GC-nCB)–CP electrode material showed a uniform and nanostructured surface which could yield to large surface area for the microsensor.

Insert Fig 2

3.2. Electrochemical impedance spectroscopy (EIS) studies

EIS is a sensitive and efficient electrochemical technique to study and monitor the surface characteristics of modified electrodes which presents the electron transfer resistance (Rct) at the electrode surface [36]. This procedure provides valuable information on the change in conductance or impedance of an electrode/electrolyte interface during its modification. For this purpose, the microelectrodes: bare (GC)- μ CPE and modified (GC-nCB)- μ CPE, were studied by cyclic voltammetry (CV) and electrochemical impedance spectroscopy (EIS) in 5.0 mM Fe (CN)₆³⁻/Fe(CN)₆⁴⁻ equimolar mixture with 0.1 M KCl as supporting electrolyte. As

shown in Figure 3 (A), it is evident that the CV current of (GC-nCB)- μ CPE is much higher than that of (GC)- μ CPE (2.5 times higher), this is due to the increase in the specific surface area due to the nanostructuration of the carbon paste electrode by the nCBs. In order to determine the electroactive surface area of bare and modified electrodes, it was calculated the effective electrochemical surface area of (GC)- μ CPE and (GC-nCB)- μ CPE, using the Randles–Sevcik equation (Eq. (1)) [**37**]:

$$I_{pa} = 2.69 \times 10^5 \times A \times D^{1/2} \times n^{3/2} \times v^{1/2} \times C$$
(1).

Where I_{pa} (A) refers to the anodic peak current, n is the electron transfer number (n=1), D is the diffusion coefficient (6.70 × 10⁻⁶ cm² s⁻¹) [**38**], C is the concentration of Fe(CN)₆³⁻ in the solution (5 mmol.L⁻¹), v is the scan rate (0.05 V.s⁻¹) and A (cm²) is the effective surface area, which is estimated for the electrodes. Based on the results presented in Figure 3 (A) the value of the active electrode area was found to be 2.1×10^{-7} cm² and 1.4×10^{-6} cm² for (GC)-µCPE and (GC-nCB)-µCPE, respectively, which was nearly 7 times greater.

Figure 3 (B) shows the Nyquist plot of (GC)- μ CPE and (GC-nCB)- μ CPE obtained by EIS measurments using an applied potential of 0.2 V, amplitude of 10 mV and in a frequency range of 10⁻² to 10⁺⁵ Hz. As can be seen, in the (GC)- μ CPE, a semicircle was observed with résistance of charge transfert (Rct) of 148.0 K Ω obtained from Randles circuit. Once the microelectrode was modified with nCB, a much smaller semicircle was obtained with Rct of 60.6 K Ω , indicating that the subsequent nanostructuration with carbon black nanoparticles improved the charge transfer kinetics at the electrode-solution interface and reduced the Rct value by a factor of 2.5.

Insert Fig 3

3.3. Cyclic voltammetry study of µCPEs

3.3.1 Effect of paste composition on CV response of EE2

In order to determine the best composition of the carbon paste in term of sensitivity in EE₂ CV response (highest oxidation peak current), we have prepared various mass compositions of the carbon paste, comprising 40 to 100% graphite and 0 to 60% nCB: (i) graphite powder (100%) (GC)-CP-A, (ii) graphite powder (85%) and carbon black (15%) (GC-nCB)-CP-B, (iii) graphite powder (80%) and carbon black (20%) (GC-nCB)-CP-C, (iv) powder graphite (70%) and carbon black (30%) (GC-nCB)-CP-D, (v) graphite powder (60%) and carbon black (40%) (GC-nCB)-CP- E, (vi) graphite powder (50%) and carbon black (50%) (GC-nCB)-CP-F and (vii) graphite powder (40%) and carbon black (60%) (GC-nCB)-CP-J.

Voltammograms for EE₂ oxidation were recorded in an aqueous solution containing 0.01 M phosphate buffer (pH = 7.4) as the supporting electrolyte, explored potential 0.0 to 1.0 V/Ag/AgCl, scan rate 50 mV.s⁻¹and a low EE₂ concentration of 1.0×10^{-7} mol.L⁻¹ with an accumulation time of 5 minutes at 0.0 V/Ag/AgCl with stirring (300 rpm). By comparing the values of oxidation peak current obtained from these different (GC-nCB)– μ CPE (Figure 4), it is clear that the best current intensity was obtained with the compositions (vi) (GC-nCB)-CP-F and (vii) (GC-nCB)-CP-J of the modified microelectrode (GC-nCB)– μ CPE. This indicates that the addition of carbon black has significantly improved the electrochemical signal of EE₂ thanks to the exceptional properties of carbon black (nCB), such as its large surface area, its high thermal and electrical conductivity and it's very low cost. Several studies have shown

that the charge transfer and electroactive surface of modified nCB-based electrodes are comparable to those of carbon and graphene nanotubes [39]. These characteristics make nCB a promising candidate for the design of electrochemical sensors and biosensors [14, 40, 41]. For the two compositions (vi) and (vii) containing 50% and 60% nCB respectively, it was possible to record the same current intensity value as the best response (see figure 4). Consequently, using the composition containing 50%-50% as the optimum composition for the remainder of the experiments.

Insert Fig 4

3.3. 2 Effect of preconcentration time

The influence of the preconcentration time for the EP step at 0V/Ag/AgCl on the response of micro-sensor in linear sweep voltammetry was examined under the following operating conditions: 0.01 M phosphate buffer solution (pH = 7.4) in the presence of 2.5×10^{-7} mol.L⁻¹ of EE₂, scan rate 50 mV.s⁻¹, explored potential 0.0 to 1.0 V/Ag/AgCl and ET at 0V under stirring (300 rpm). As can be seen in insert Figure 5, the voltammograms present an oxidation peak of EE₂ at 0.48 V with a Gaussian shape that is characteristic of a diffusion less electrochemical behavior [42] corresponding to the oxidation of adsorbed EE₂. The adsorption of EE₂ occur for EP step with the polarization of (GC-nCB)-µCPE at the potential 0V/Ag/AgCl under stirring. The oxidation current peak intensity of adsorbed EE₂ increases with the accumulation time and reaches a maximum at 5 minutes. At longer preconcentration times up to 8 minutes, no change was observed, indicating that EE₂ absorption equilibrium is reached on (GC-nCB) - µCPE. Therefore, the accumulation time of 5 minutes was chosen as optimal condition for the subsequent experiments.

Insert Fig. 5

3.3. 3 Effect of sweep rate

In order to obtain information about the oxidation process of EE₂ on the (CG-nCB) $-\mu$ CPE electrode, the electrochemical response has been investigated for various scan rates 10 - 80 mV.s⁻¹. Voltammograms were recorded in 0.01 M phosphate buffer (pH = 7.4), in the presence of 1.5×10^{-7} mol.L⁻¹ of EE₂, an accumulation time of 5 minutes at 0.0 V/Ag/AgCl with stirring (300 rpm), and explored potential 0.0 to 1.0 V/Ag/AgCl (Figure 6 (A)). The results obtained show that the oxidation peak current of EE₂ increases with the potential scan rate and with a linear variation (correlation coefficient R² 0.999). This demonstrates diffusion less electrochemical process controlled by adsorbed electroactive species (EE₂) in the nanostructured carbon paste (GC-nCB-CP) in the cavity microelectrode [41, 42].

Additionally, the plot of oxidation peak potentials versus the logarithm of the scan rate allows us to determine the number of electrons involved in the electrochemical reduction of EE_2 . The results in Figure 6 (B) show that Epa as a function of log (v) is a straight line with a regression equation as follows:

$$Epa (V) = 0.0648 \log (v / mV. s^{-1}) + 0.397 (R^2 = 0.998)$$
(2).

As indicated by Laviron's theory [42], the slope value obtained corresponds to (- 2.3RT/ α nF); where R is the gas constant (8.314 J.mol⁻¹.K⁻¹), T is the ambient temperature; α is the transfer coefficient which in the case of a non-reversible process is equal to 0.5, n is the number of electrons transferred and F is the Faraday constant. Using the slope value, the number of

electrons transferred during the reaction of EE_2 was calculated to be equal to $1.822\approx 2$. Thus, it can be assumed that the global process of electro-oxidation of EE_2 on the (CG- nCB) - μCPE involves two electrons.

Insert Fig. 6

3.3. 4 Effect of pH

The influence of pH on the oxidation of EE₂ with the micro-sensor (GC-nCB) - μ CPE was studied by voltammetry in the pH range from 3.0 to 10.0. Operating parameters include a supporting electrolyte consisting of 0.01 M phosphate buffer solution with 1.0×10^{-7} mol.L⁻¹ EE₂, a scan rate of 50 mV.s⁻¹, a potential studied from 0.0 to 1.0 V/Ag/AgCl, an accumulation time of 5 minutes at 0.0 V/Ag/AgCl with stirring (300 rpm) and ET at 0 V. From the recorded voltammograms at different pH to the plots of Ip (A) and Ep (V) as a function of pH are presented respectively Figure 7 (A) and Figure 7 (B).

Insert Fig. 7

The pH has a direct influence on the peak current intensity of the electro-oxidation of EE₂, as can be seen in Figure 7 (A). From pH 3 to pH 8, the EE₂ oxidation peak intensity (Ip) increases linearly with the pH (slope of 0.07 μ A/pH unit), reaches a maximum and decreases for pH upper to 8. This decreasing is justified by the formation of the ionized form of EE₂ molecule (phenolate), which is anionic, since its pKa value is of about 10 [5]. The adsorption of this anionic form of EE₂ is not favorable during the electrochemical deposition step at 0V. In order to optimize the current intensity, a phosphate buffer at pH 7.4 was chosen for recording the voltammograms. By plotting the variation of the anodic peak potential (Epa) of EE₂ as a function of pH (Figure 7 (B)), it was observed that Ep shift to more negative values with increasing pH of the electrolyte. This demonstrates the involvement of proton in the electro-oxidation reaction of EE₂. Indeed, Epa for the oxidation of EE₂ on the electrode (GC-nCB) - μ CPE showed a linear relationship with the pH of the solution, according to the following equation:

$$Epa = -0.0592 \, pH + 0.918 \qquad (R^2 = 0.997) \tag{3}.$$

The value of the slope is equal to the theoretical one 0.059 V/pH (Nernstian behavior) [43], according to the Nernst equation; Ep (V) = 0.059 (m/n) / pH (n is the number of electrons while m is the number of protons). This suggests the participation of the same number of protons and electrons in the electrochemical EE₂ oxidation mechanism (n=m= 2) [44] which is consistent to the proposed reaction mechanism for electro-oxidation of EE₂ on (CG-nCB) – μ CPE presented in the scheme of Figure 8 [45].

Insert Fig. 8

3.4. Validation of the proposed analytical method with GC-nCB) –µCPE

3.4.1 Micro-sensor performance

The calibration curve was plotted under the optimal conditions previously determined: 0.01 M phosphate buffer solution (pH = 7.4) as supporting electrolyte, a potential range of 0.0 - 1.0 V/Ag/AgCl, a scan rate of 50 mV. s⁻¹, a preconcentration time of 5 min at 0V. Voltammograms were recorded with (GC-nCB)-µCPE by varying the EE₂ concentrations in the cell. The intensity of the EE₂ oxidation current peak increases with increasing

concentration (Figure 9). Two calibration lines were obtained for the lowest concentrations, in the range 5 -200 nmol.L⁻¹, and for highest concentrations from 0.2 to 1 μ mol.L⁻¹ (Figure 9); with respectively the following regression equations:

$$Ipa (\mu A) = 4.16 C + 4.47 \times 10^{-3} (mol.L^{-1}) (R^2 = 0.999)$$
(4).

And
$$Ipa(\mu A) = 1.232 C + 0.752(mol.L^{-1}) (R^2 = 0.993)$$
 (5).

Moreover, the sensitivity of the micro-sensor is 3.4 times larger for the lowest concentrations than those above 0.2 μ mol.L⁻¹. Furthermore, the LOD and LOQ were calculated, considering the linearity for the lowest concentrations, using the following equations: LOD = 3S/m and LOQ = 10S/m respectively [46, 47], where: S is the standard deviation of the blank and m is the slope of the calibration line. The calculated detection LOD and LOQ are 1.4 nmol.L⁻¹ and 4.9 nmol.L⁻¹, respectively.

Insert Fig 9

3.4.2. Repeatability and reproducibility of the proposed method

For the purpose of estimating the accuracy of the proposed analytical method, studies of reproducibility and repeatability were conducted and evaluated for three standard solutions with EE₂ concentrations: 5.0, 7.5 and 10.0×10^{-8} mol.L⁻¹. The voltammogrammes were recorded on (CG-nCB)–µCPE under optimal operating conditions (0.01 M phosphate buffer (pH = 7.4), an accumulation time of 5 minutes at 0.0 V/Ag/AgCl with stirring (300 rpm), a scan rate of 50 mV.s⁻¹, and explored potential 0.0 to 1.0 V/Ag/AgCl).

Indeed, the repeatability of (CG-nCB) – μ CPE was evaluated using the same microsensor and the reproducibility was studied using four micro-sensors prepared independently by the same procedure under the same operating conditions. The results obtained from the measured oxidation peak currents are collected in Table 1. For the different concentrations, the calculated relative standard deviations (RSD) do not exceed 3.8% for repeatability and 4.8% for reproducibility. These low RSD values show that the modified microelectrode (CG-nCB) – μ CPE present agood repeatability and reproducibility indicating that the filling of the microcavity by the (CG-nCB) carbon paste was very reproducible.

Insert Tab. 1

3.4.3 Effect of interference

The influence of various substances as compounds that could potentially interfere with the determination of EE₂ at (GC-nCB) - μ CPE in water or urine was investigated under the optimal conditions previously described. Voltammetry measurements were carried out on solutions containing EE₂ 1.0×10⁻⁷ mol.L⁻¹ in the presence of 1.0×10⁻⁷ mol.L⁻¹ of molecules commonly found in water and/or urine (sulfamethazine, amoxicillin, metronidazole, danofloxacin, ofloxacin, atrazine, ketoprophen, propazine, glyphosate, ampicillin, uric acid, urea and glucose). Some cations and anions species were also tested as potential interfering compounds (SO₄²⁻, K⁺, Cl⁻, CH₃COO⁻, HPO₄²⁻) but the results indicated that no modification of the EE₂ oxidation peak current occurred for these ionic species.

By exploiting the intensities of the oxidation peaks in the voltammograms recorded with (CG-nCB) – μ CPE, in the absence and in the presence of the organic compounds tested (Figure 10),

no Ipa variation was recorded for the majority of compounds. Only a few compounds (metronidazole, ketoprophen, propazine and ampicillin) leaded to, slight variations of less than 7.0 % of the initial oxidation current.

These results reveal that the micro-sensor (CG-nCB) $-\mu$ CPE with EP step at 0V provides a sufficient high selectivity for the EE₂ voltammetric detection in presence of interfering organic and inorganic compounds.

Insert Fig. 10

3.4.4 Analysis of real samples

To evaluate the validity of the proposed method, the (CG-nCB) – μ CPE was applied for the determination of the EE₂ in two local water samples collected after two different tertiary treatments before returning to environment (WWTPs -M and WWTPs -D) and in synthetic urine. Voltammetry measurements were carried out in an aqueous solution containing 0.01 M phosphate buffer (pH = 7.4) as supporting electrolyte, the potential range is 0 to 1 V versus Ag/AgCl, a scan rate of 50 mV.s⁻¹, and a preconcentration time of 5 min with stirring (300 rpm). In all the three samples, no trace of EE₂ was detected with the modified electrode (less than LOD 1.4 nmol.L⁻¹). Consequently, the analyses were carried out on water and urine samples enriched respectively with tens and hundreds of nanomolar of EE₂ concentration. The EE₂ concentration measurements were obtained by the standard addition method under the optimal operating conditions for the voltammetric method at (CG-nCB) – μ CPE.

The obtained results are summarized in Table 2. The values of the recoveries obtained for 4 replicates for the samples (WWTPs -M), (WWTPs -D) and urine are respectively of 101.3%, 101.1% and 103.3%, with RSD of 1.4%, 1.3% and 2.3%, indicating that there were no significant matrix interferences on the response of the (CG-nCB) – μ CPE. Therefore,

the CG-nCB) $-\mu$ CPE appears as a precise sensor for the EE₂ analysis in environmental water and urine samples.

Insert Tab. 2

3.4.5 Comparison with subsequent work

Table 3 lists some examples of electrochemical sensors intended for the analysis of the hormone 17α – ethinylestradiol (EE₂) and the comparison of analytical performances of the micro-sensor presented in this paper (the linearity range and the detection limit) with those of other electroanalytical methods reported recently in the literature. In this work, it was achieved some of the lowest LOD values (in the nanomolar range), similar to those obtained in AdSV at mercury electrodes (HMDE) [52, 53], with a large range of analyzable EE₂ concentrations. Therefore, these results confirm that the proposed micro-sensor (CG-nCB) – μ CPE could be a good alternative method for the quantification of ultra-traces of EE₂ in the environment and urine.

Insert Table.3

4. Conclusion

For the analysis of the hormone 17α -ethinylestradiol, it has been proposed a simple, easy-toprepare and economical voltammetric method by using a novel micro-sensor (GC-nCB)µCPE) based on microcavity electrode and nanostructured carbon paste. In fact, the combination of a carbon paste microsensor with nCB guarantees interesting sensitivity and selectivity of the method developed. Under optimum operating conditions (0.01 M phosphate buffer (pH = 7.4), scan rate of 50 mV. s⁻¹ and preconcentration time of 5 min at 0V under stirring), the proposed micro-sensor exhibits fast and precise response, and exceptional selectivity. In addition, (GC-nCB)-µCPE has a very low detection limit of around 1.5 nmol.L⁻¹, a wide linear concentration range (3 decades), good repeatability and reproducibility (overall the RSDs are less than 4.8%). (GC-nCB)-µCPE) was successfully applied, without interference, to the quantification of ultra-trace EE₂ in environmental water and urine samples (the recovery percentages for the spiked samples being very close to 100% with a maximum deviation of 3.3%). This method is also more sensitive than many other analytical techniques reported in the literature.

Acknowledgments

The authors would like to thank the Algerian Ministry of Education and Scientific Research and express their deep gratitude to F. Gouttefangeas of the CMEBA platform of ISCR-CNRS 6226 from University of Rennes 1(France) for the SEM analyses.

Declaration of Competing Interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to impact the work reported in this paper.

References

[1] A. L. Oropesa, D. Martín-Hidalgo, C. Fallola, M. C. Gil, Effects of exposure to 17- alphaethynylestradiol on sperm quality of tench (Tinca tinca), Ecotoxicol. Environ. Saf. 120 (2015) 318-325. https://doi: 10.1016/j.ecoenv.2015.06.016.

[2] S. Barreca and al., Determination of estrogenic endocrine disruptors in water at subng L^{-1} levels in compliance with Decision 2015/495/EU using offline-online solid phase extraction concentration coupled with high performance liquid chromatographytandem mass spectrometry, Microchem. J. 147 (2019) 1186-1191. https://doi:10.1016/j.microc.2019.04.030.

[3] K. L. Lopes, H. Leijoto de Oliveira, J. A. S. Serpa, J. A. Torres, F. G. E. Nogueira, V. A. A. de Freitas, K. B. Borges, M. C. Silva, Nanomagnets based on activated carbon/magnetite

nanocomposite for determination of endocrine disruptors in environmental water samples , Microchem. J. 168 (2021) 106366. https://doi: 10.1016/j.microc.2021.106366.

[4] Á. Torrinha, P. Carneiro, D. Dias, C. Delerue-Matos, S. Morais, The simpler the better: Highly sensitive 17α -ethinylestradiol sensor based on an unmodified carbon paper transducer, Talanta. 245 (2022) 123457. https://doi: 10.1016/j.talanta.2022.123457.

[5] P. M. Zagalo, P. A. Ribeiro, M. Raposo, Detecting Traces of 17α-Ethinylestradiol in Complex Water Matrices. Sensors. 20 (2020) 7324. https://doi.org/10.3390/s20247324.

[6] M. D. S. Monteiro and al., Reduced Graphene Oxide-based Sensor for 17α -Ethinylestradiol Voltammetric Determination in Wastewater, Tablets and Synthetic Urine Samples, Electroanalysis. 34 (2022) 1422-1430. https://doi:10.1002/elan.202100500.

[7] Y. Li and al., Cubic cobalt ferrite anchored on graphene sheets for nonenzymaticelectrochemical detection of 17α -ethinyl estradiol in environmental and biological samples, J. Environ. Chem. Eng. 12 (2024) 111612. https://doi:10.1016/j.jece.2023.111612.

[8] J. Klikarová, M. Chromá, M. Sýs, Simultaneous voltammetric determination of female hormones using different carbonaceous electrodes in a non-aqueous environment, Microchem. J. 193 (2023) 109219. https://doi: 10.1016/j.microc.2023.109219.

[9] H. Y. Zheng, O. A. Alsager, C. S. Wood, J. M. Hodgkiss, N. O. V. Plank, Carbon nanotube field effect transistor aptasensors for estrogen detection in liquids, J. Vac. Sci. Technol. B. 33 (2015) 06F904. https://doi: 10.1116/1.4935246.

[10] M. Nodehi, M. Baghayeri, R. Ansari, H. Veisi, Electrochemical quantification of 17α – Ethinylestradiol in biological samples using a Au/Fe3O4@TA/MWNT/GCE sensor, Mater. Chem. Phys. 244 (2020) 122687. https://doi:10.1016/j.matchemphys.2020.122687.

[11] M. Mekersi, M. Ferkhi, E. Kuyumcu Savan, Electrochemical biodetection of glucose using La0.6Sr0.4Co0.8Fe0.2O3 and La1,7Sr0,3CuO4 Nano-Particles modified with black carbon deposited on glassy carbon electrode, Microchem. J. 194 (2023) 109346. https://doi: 10.1016/j.microc.2023.109346.

[12] A. Sahragard, P. Varanusupakul, M. Miró, Nanomaterial decorated electrodes in flowthrough electrochemical sensing of environmental pollutants: A critical review, Trends Environ. Anal. Chem. 39 (2023) e00208. https://doi: 10.1016/j.teac.2023.e00208.

[13] S. Jiao and al., Emerging hydrovoltaic technology based on carbon black and porous carbon materials, A mini review, Carbon. 193 (2022) 339-355. https://doi:10.1016/j.carbon.2022.03.047.

[14] F. Arduini, S. Cinti, V. Mazzaracchio, V. Scognamiglio, A. Amine, D. Moscone, Carbon black as an outstanding and affordable nanomaterial for electrochemical (bio) sensor design, Biosens. Bioelectron. 156 (2020) 112033. https://doi.org/10.1016/j.bios.2020.112033.

[15] J.-P. Reboul, G. Moussalli, About Some D-C Conduction Processes in Carbon BlackFilled Polymers, Int. J. Polym. Mater. Polym. Biomater. 5 (1976) 133-146. https://doi:10.1080/00914037608072394. [16] I. Balberg, N. Wagner, Y. Goldstein, S. Z. Weisz, Tunneling and Percolation Behavior in Granular Metals , MRS Proc. 195 (2011) 195-233. https://doi: 10.1557/PROC-195-233.

[17] A. Koterwa and al., The role of electrolysis and enzymatic hydrolysis treatment in the enhancement of the electrochemical properties of 3D-printed carbon black/poly (lactic acid) structures, Appl. Surf. Sci. 574 (2022) 151587. https://doi:10.1016/j.apsusc.2021.151587.

[18] L. M. C. Ferreira and al., Using nanostructured carbon black-based electrochemical (bio)sensors for pharmaceutical and biomedical analyses: A comprehensive review, J. Pharm. Biomed. Anal. 221 (2022), 115032. https://doi: 10.1016/j.jpba.2022.115032.

[19] P. Batista Deroco, F. Campanhã Vicentini, O. Fatibello-Filho, An electrochemical sensor for the simultaneous determination of paracetamol and codeine using a glassy carbonelectrode modified with nickel oxide nanoparticles and carbon black, Electroanalysis. 27 (2015) 2214–2220. https://doi 10.1002/elan.201500156.

[20] D. N. da Silva, A. C. Pereira, An electrochemical sensor modified with a molecularly imprinted polymer and carbon black for 17- β -estradiol detection, Anal. Methods. 14 (2022) 1208-1213. https://doi: 10.1039/D1AY02062E.

[21] M.-S. Koshki, M. Baghayeri, M. Fayazi, Application of sepiolite/FeS2 nanocomposite for highly selective detection of mercury(II) based on stripping voltammetric analysis, J. Food Meas. Charact. 15 (2021) 5318-5325. https://doi: 10.1007/s11694-021-01097-0.

[22] M. Ghanei-Motlagh, M. A. Taher, M. Fayazi, M. Baghayeri, A. Hosseinifar, Non-Enzymatic Amperometric Sensing of Hydrogen Peroxide Based on Vanadium Pentoxide Nanostructures, J. Electrochem. Soc. 166 (2019) B367. https://doi: 10.1149/2.0521906jes.

[23] Aaryashree, A. K. Choudhary, Y. Yoshimi, Disposable Sensor Chips with Molecularly Imprinted Carbon Paste Electrodes for Monitoring Anti-Epileptic Drugs, Sensors. 23 (2023). https://doi: 10.3390/s23063271.

[24] M. Fayazi, M. Ghanei-Motlagh, C. Karami, Application of magnetic nanoparticles modified with L-cysteine for pre-concentration and voltammetric detection of copper(II), Microchem. J. 181 (2022) 107652. https://doi: 10.1016/j.microc.2022.107652.

[25] J. G. Manjunatha, B. E. K. Swamy, G. P. Mamatha, U. Chandra, E. Niranjana, B. S. Sherigara, Cyclic Voltammetric Studies of Dopamine at Lamotrigine and TX-100 Modified Carbon Paste Electrode, Int. J. Electrochem. Sci. 4 (2009) 187-196. 2009, https://doi: 10.1016/S1452-3981(23)15120-0.

[26] P. A Pushpanjali, J. G Manjunatha, G. Tigari, S. Fattepur, Poly(Niacin) Based Carbon Nanotube Sensor for the Sensitive and Selective Voltammetric Detection of Vanillin with Caffeine, Anal. Bioanal. Electrochem. 12 (2020) 553-568.

[27] J. G. Manjunatha, Surfactant modified carbon nanotube paste electrode for the sensitive determination of mitoxantrone anticancer drug, J. Electrochem. Sci. Eng. 7 (2017) 39-49. https://doi: 10.5599/jese.368.

[28] N. Hareesha, J. G. Manjunatha, Electro-oxidation of formoterol fumarate on the surface of novel poly(thiazole yellow-G) layered multi-walled carbon nanotube paste electrode, Sci. Rep. 11 (2021) 12797. https://doi: 10.1038/s41598-021-92099-x.

[29] J. G. Manjunatha, Electrochemical Polymerised Graphene Paste Electrode and Application to Catechol Sensing, Open Chem. Eng. J. 13 (2019) 81-87. https://doi: 10.2174/1874123101913010081.

[30] C. Cachet-Vivier, M. Keddam, V. Vivier, L. T. Yu, Development of cavity microelectrode devices and their uses in various research fields, J. Electroanal. Chem. 688 (2013) 12-19. https://doi: 10.1016/j.jelechem.2012.09.011.

[31] L. Taouri, M. Bourouina, S. Bourouina, D. Hauchard, A new highly sensitive microsensor for the ultra-traces analysis of paracetamol directly in water, Microchem. J. 178 (2022) 107380. https://doi: 10.1016/j.microc.2022.107380.

[32] A. Terbouche, C. A. Ramdane-Terbouche, D. Hauchard, S. Djebbar, Evaluation of adsorption capacities of humic acids extracted from Algerian soil on poly aniline for application to remove pollutants such as Cd(II), Zn(II) and Ni(II) and characterization with cavity microelectrode, J. Environ. Sci. 23 (2011) 1095-1103. https://doi: 10.1016/S1001-0742(10)60521-9.

[33] C. Ait-Ramdane-Terbouche A. Terbouche, M. Khalfaoui, C. L. Douib, H. Lakhdari, D. Lerari, K. Bachari, D. Mezaoui, J. P. Guegan, D. Hauchard, Synthesis, characterization and molecular modeling of binuclear Pd(II) and Ir(IV) complexes with bis [4-hydroxy-6- methyl-3-{(1E)-N-[2-(ethylamino)ethyl]ethanimidoyl}-2H-pyran-2-one]: Application to detection of glucose, J. Mol. Struct. 1170 (2018) 119-130. https://doi:10.1016/j.molstruc.2018.05.075.

[34] S. Shahamirifard, M. Ghaedi, A new electrochemical sensor for simultaneous determination of arbutin and vitamin C based on hydroxyapatite-ZnO-Pd nanoparticles modified carbon paste electrode, Biosens. Bioelectron. 141 (2019) 111474. https://doi: 10.1016/j.bios.2019.111474.

[35] N. Sarigul, F. Korkmaz, İ. Kurultak, A New Artificial Urine Protocol to Better Imitate Human Urine, Sci. Rep. 9 (2019) 20159. https://doi: 10.1038/s41598-019-56693-4.

[36] A. A. C. Riojas, Development of a new electrochemical sensor based on silver sulfide nanoparticles and hierarchical porous carbon modified carbon paste electrode for determination of cyanide in river water samples, Sensors and Actuators B: Chem. 287 (2019) 544-550. https://doi.org/10.1016/j.snb.2019.02.053.

[37] S. Hrapovic, Y. Liu, K. B. Male, J. H. T. Luong, Electrochemical Biosensing Platforms Using Platinum Nanoparticles and Carbon Nanotubes, Anal. Chem. 76 (2004) 1083-1088. https://doi: 10.1021/ac035143t.

[38] S. Khaliliazar, I. Öberg Månsson, A. Piper, L. Ouyang, P. Réu, M. M. Hamedi, Woven Electroanalytical Biosensor for Nucleic Acid Amplification Tests, Adv. Healthc. Mater. 10 (2021) 2100034. https://doi: 10.1002/adhm.202100034.

[39] T. A. Silva, F. C. Moraes, B. C. Janegitz, O. Fatibello-Filho, Electrochemical Biosensors Based on Nanostructured Carbon Black: A Review, Journal of Nanomaterials (2017). https://doi.org/10.1155/2017/4571614.doi.org/10.1155/2017/4571614.

[40] F. C. Vicentini, P. A. Raymundo-Pereira, B. C. Janegitz, S. A. Machado, O. FatibelloFilho, Nanostructured carbon black for simultaneous sensing in biological fluids , Sens. Actuators B Chem. 227 (2016) 610–618. https://doi.org/doi:10.1016/j.snb.2015.12.094.

[41] X. Kang, J. Wang, H. Wu, J. Liu, I. A. Aksay, Y. Lin, A graphene-based electrochemical sensor for sensitive detection of paracetamol, Talanta. 81 (2010) 754-759. https://doi: 10.1016/j.talanta.2010.01.009.

[42] E. Laviron, General expression of the linear potential sweep voltammogram in the case of diffusionless electrochemical systems, J. Electroanal. Chem. Interfacial Electrochem. 101 (1979) 19-28. https://doi: 10.1016/S0022-0728(79)80075-3.

[43] A. A. Ensafi, M. Izadi, H. Karimi-Maleh, Sensitive voltammetric determination of diclofenac using room-temperature ionic liquid-modified carbon nanotubes paste electrode, Ionics. 19 (2013) 137-144. https://doi: 10.1007/s11581-012-0705-0.

[44] N. C. Honakeri, S. J. Malode, R. M. Kulkarni, N. P. Shetti, Electrochemical behavior of diclofenac sodium at coreshell nanostructure modified electrode and its analysis in human urine and pharmaceutical samples, Sens. Int. 1 (2020) 100002. https://doi: 10.1016/j.sintl.2020.100002.

[45] M. M. Charithra, J. G. Manjunatha, Poly (1-Proline) modified carbon paste electrode as the voltammetric sensor for the detection of Estriol and its simultaneous determination with Folic and Ascorbic acid, Mater. Sci. Energy Technol. 2 (2019) 365-371. https://doi: 10.1016/j.mset.2019.05.002.

[46] A. Pavinatto, L. A. Mercante, C. S. Leandro, L. H. C. Mattoso, D. S. Correa, LayerbyLayer assembled films of chitosan and multi-walled carbon nanotubes for the electrochemical detection of 17α -ethinylestradiol, J. Electroanal. Chem. 755 (2015) 215-220. https://doi: 10.1016/j.jelechem.2015.08.002.

[47] C. Perez, F. R. Simões, L. Codognoto, Voltammetric determination of 17α ethinylestradiol hormone in supply dam using BDD electrode, J. Solid State Electrochem. 20 (2016) 2471-2478. https://doi: 10.1007/s10008-015-3091-5.

[48] C. Li, Voltammetric determination of ethinylestradiol at a carbon paste electrode in the presence of cetyl pyridine bromine, Bioelectrochemistry. 70 (2007) 263-268. https://doi: 10.1016/j.bioelechem.2006.10.001.

[49] J. Smajdor, R. Piech, M. Pięk, B. Paczosa-Bator, Sensitive Voltammetric Determination of Ethinyl Estradiol on Carbon Black Modified Electrode , J.Electrochem. Soc. 164 (2017) H885. https://doi: 10.1149/2.0851713jes.

[50] J. d. S. Fernandes, J. O. Fernandes, C. A. R. Bernardino, C. F. Mahler, B. F. Braz, R. E. Santelli, F. H. Cincotto, A New Electrochemical Sensor Based on Carbon Black Modified With Palladium Nanoparticles for Direct Determination of 17α-Ethinylestradiol in Real Samples, Electroanalysis, 34 (2022) 863-871. https://doi:10.1002/elan.202100474.

[51] V. O. d. Jesus, V. S. Ferreira, B. G. Lucca, Electrochemical study and simultaneous voltammetric determination of contraceptives ethinylestradiol and cyproterone acetate using silver nanoparticles solid amalgam electrode and cationic surfactant, Talanta. 210 (2020) 120610. https://doi: 10.1016/j.talanta.2019.120610.

[52] C. N. Nunes, L. E. Pauluk, M. L. Felsner, V. Egéa dos Anjos, S. P. Quináia, Rapid screening method for detecting ethinyl estradiol in natural water employing voltammetry, J. Anal. Methods Chem. 2016 (2016). https://doi.org/10.1155/2016/3217080.

[53] E. M. Ghoneim, H. S. El-Desoky, M. M. Ghoneim, Adsorptive cathodic strippingvoltammetric assay of the estrogen drug ethinylestradiol in pharmaceutical formulation and human plasma at a mercury electrode, J. Pharm. Biomed. Anal. 40 (2006) 255-261. https://doi: 10.1016/j.jpba.2005.07.024.

Figures

Fig 1













Fig 10

Figure captions

Figure. 1. Graphical presentation of (GC-nCB)-CP micro-sensor fabrication.

Figure 2. SEM images of (A) graphite carbon (GC), (B) and (C) carbon black. (nCB) and (D) the (GC- nCB)-CP mixture (graphite powder (50%) and carbon black (50%) –CP).

Figure 3. (A) CVs and (B) Nyquist plots of different electrodes for the unmodified (GC)- μ CPE and modified (GC-nCB)- μ CPE carbon paste microelectrode in a 0.1 M KCl solution containing 5 mmol L⁻¹ Fe (CN)₆ ^{3-/4-}.

Figure 4. Histogram illustrating the influence of CPE composition on micro-sensor voltammetric response, in 0.01 M phosphate buffer (pH = 7.4), EE₂ concentration of 1.0×10^{-7} mol L⁻¹, scan rate of 50 mV. s⁻¹, explored potential 0.0 to 1.0 V/Ag/AgCl.

Figure 5. Effect of preconcentration time on the EE₂ peak oxidation current intensity in voltammetry with (GC-nCB)- μ CPE; insert the corresponding voltammograms in 0.01 M phosphate buffer solution (pH = 7.4) in the presence of 2.5×10⁻⁷ mol.L⁻¹ of EE₂, scan rate 50 mV.s⁻¹ and ET at 0V under stirring (300 rpm).

Figure 6. Study of the effect of scan rate on (GC-NC)- μ CPE in 0.01 M phosphate buffer (pH = 7.4) containing 1.5×10^{-7} mol L⁻¹ of EE₂, with 5 min accumulation time at 0.0 V / Ag/AgCl for various scan rates 10 - 80 mV. s⁻¹: voltammograms and curve of variation of anodic peak currents as a function of scan rate (A) and variation of anodic peak potentials as a function of logarithms of scan rates (B).

Figure 7. Influence of pH (from 3.0 to 10.0) on the voltammetric response obtained for 1.0×10^{-7} mol.L⁻¹ of EE₂, with 5 min accumulation time at 0.0 V / Ag/AgCl, and a scan rate of 50 mV.s⁻¹: the variation of the anodic peak current intensity, Ip, (A) and the variation of the anodic peak potential, Ep, (B).

Figure 8. Probable mechanism of EE₂ electrooxidation at (GC-nCB)-µCPE.

Figure 9. Voltammograms and curve of variation of anodic peak currents as a function of EE₂ concentration, obtained at the (GC-nCB)- μ CPE electrode with 0.01 phosphate buffer (pH = 7.4) for different EE₂ concentrations (5 to 800 nmol L⁻¹), with 5 min accumulation time at 0.0 V / Ag/AgCl, ; v = 50 mV. s⁻¹.

Figure 10. Influence of potential interfering compounds on the EE₂ oxidation peak current in voltammetry at (GC-nCB)- μ CPE, the voltammetric response obtained for 1.0×10^{-7} mol.L⁻¹ of EE₂, with 5 min accumulation time at 0.0 V / Ag/AgCl, and a scan rate of 50 mV.s⁻¹

Tables

Tab 1

Conce	centration of EE ₂ (×10 ⁻⁸ mol L ⁻¹)		Repeatability of <i>I</i> p		Reproducibility of <i>I</i> p	
			KSD * (%)		KSD* (%)	
	5.0		3.8		4.8	
	7.5		1.1		2.9	
	10.0		1.3		3.7	
* n= (4)						
Гаb 2						
	Samples	Added	Found	RSD*	Recovery	
		[EE2] (µM)	[EE2] (µM)	(%)	(%)	
	(WWTPs -M)	non-doped	< LOD	-	-	
		0.050	0.051 ±0.001**	1.4	101.3	

(WWTPs -D)	non-doped	< LOD	-	-
	0.050	0.050 ±0.001**	1.3	101.1
urine	non-doped	< LOD	-	KC
	0.100	0.103 ±0.003**	2.3	103.3

* n= (3) ** Confidence interval for α = 0.05.

Tab 3

Technique / Electrode	Linear range	LOD	Ref
	(µmol L ⁻¹)	(µmol L ⁻¹)	
SWV/ (Chi/CNTs) ¹	0.05 - 20	0.09	[46]
SWV/ (BDD) ²	0.99 – 5.2	0.24	[47]
LSV /(CPE) ³	0.05 - 20	0.03	[48]
DPV/ (CBGC) ⁴	0.25 -3.0	0.13	[49]
DPV/ (CB- immobilized palladium) ⁵	0. 5 - 115	0.081	[50]
SCV /(AgNP-SAE) ⁶	0.64 - 7.8	0.103	[51]
DPAdSV /(HMDE) ⁷	0.0054 - 0.202	0.0016	[52]
SWAdSV /(HMDE) ⁸	0.0019 - 0.6	0.0006	[53]
CV / (CoFe2O4/RGO) ⁹	0.05 - 100.0	0.042	[7]
DPV/ (ErGO) ¹⁰	0.04 - 8.28	0.0068	[6]
DPV /(NTC/ aptamére) ¹¹	0.05 - 1.6	0.05	[9]

	Journal Pre-proofs			
DPV/ (MIP) ¹²	0.1 - 23.0	0.03	[20]	
CV/ ((CG-nCB)–µCPE)	0.005 - 0.8	0.0014	This worl	

(1) Square wave voltammetry with chitosan / Carbon Nanotube; (2) Square wave voltammetry/boron doped diamond electrode; (3) Linear sweep voltammetry/carbon paste electrode; (4) Differential pulse voltammetry/ glassy carbon black electrode; (5) Differential pulse voltammetry / carbon black containing immobilized palladium nanoparticles; (6) Staircase voltammetry / solid amalgam electrode made with silver nanoparticles; (7) Differential pulse adsorption stripping voltammetry / mercury drop suspended electrode; (8) Square wave adsorption denudation voltammetry/ suspended mercury drop electrode; (9) Cyclic voltammetry / cobalt ferric oxide with anti-spinel structure and reduced graphene oxide composite; (10) Differential pulse voltammetry / modified electrode with reduced graphene oxide; (11) Differential pulse voltammetry/ Carbon Nanotube aptasensor and (12) Differential pulse voltammetry / molecularly imprinted polymer.

Table caption

Table 1. Repeatability and reproducibility study of (GC-nCB)- μ CPE for different EE₂ concentrations (50, 75 and 100 nmol L⁻¹) in 0.01 mol L⁻¹ phosphate buffer) (pH = 7.4), v = 50 mV. s⁻¹.

Table 2. Analysis of WWTPs samples and urine enriched with EE₂ (respectively 0.05 and 0.1 μ mol.L⁻¹). Concentration determination at (GC-nCB) – μ CPE with standard addition method and recovery for spiked samples.

Table 3. Comparison of the analytical performances of different electrochemical sensors from literature for EE_2 analysis.

Declaration of interests

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

The authors declare the following financial interests/personal relationships which may be considered as potential competing interests:

Graphical abstract

HIGHLIGHTS

- Electrochemical micro-sensor for direct analysis of 17α ethinylestradiol.
- A carbon paste cavity micro-electrode modified and nanostructured by nCB (GC-nCB)-µCPE was used.
- Electrochemical preconcentration (EP) step was used to enhance the sensivity of EE₂ detection.
- The micro-sensor displayed a detection limit of 1.4 nmol.L⁻¹.
- The method developed was applied in environmental water and urine samples.