

Open-source tools for within-exam automatic brain segmentation for accurate and reproducible geometry planning

Delphin Aurélien, Pietras Johan, Gourieux Emmanuelle, Miquel Lucie, Chéchin David, Lamalle Laurent

▶ To cite this version:

Delphin Aurélien, Pietras Johan, Gourieux Emmanuelle, Miquel Lucie, Chéchin David, et al.. Opensource tools for within-exam automatic brain segmentation for accurate and reproducible geometry planning. ESMRMB 2024, Oct 2024, Barcelona, Spain. , 2024. hal-04725300

HAL Id: hal-04725300 https://hal.science/hal-04725300v1

Submitted on 8 Oct 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.











Open-source tools for within-exam automatic brain segmentation for accurate and reproducible geometry planning

Delphin Aurélien¹, Pietras Johan¹, Gourieux Emmanuelle^{1,2,3,4}, Miquel Lucie¹, Chéchin David⁵, Lamalle Laurent¹


- ¹ Univ. Grenoble Alpes, Inserm, US17, CNRS, UAR 3552, CHU Grenoble Alpes, IRMaGe, 38000 Grenoble, France ² Univ. Grenoble Alpes, Inserm, U1216, Grenoble Institut Neurosciences, GIN, 38000 Grenoble, France
- ³ Univ. Grenoble Alpes, CNRS LPNC UMR 5105, 38000 Grenoble, France
- ⁴ Univ. Grenoble Alpes, UAR GRICAD, 38000 Grenoble, France ⁵ Philips France Commercial, Suresnes, France


Motivation

Precise and reproducible imaging volume positioning is a must-have, but is often hindered by several difficulties:

- MRI scan planning sometimes targets very specific anatomic regions
- Limited automatic scan geometry planning on the MRI console
- Impossibility to use state-of-the-art open-source tools or canonical atlases [1, 2] at scan time on the MRI console
- → The Philips PRIDE Research tool allows to externally process data and reuse it in the console during an exam.
- → We illustrate the benefits of external data processing with two open-source tools to help with MRS voxel positioning.

Proposed workflow





Discussion

- The tool for Session 1 runs in around 15' and that for Session 2 runs in 2', which is reasonable research sessions with several scans to acquire during processing time. Note that for several sessions per patient, the segmentation is only needed once.
- The two proposed tools are available on GitHub, are low-dependency and have a user-friendly graphical interface.
- It should be straightforward to use a different segmentation algorithm to reduce the processing time or to obtain finer segmentations.
- We believe these tools can be precious to help the precision and reproducibility of imaging volume positioning, especially for users which are not expert in brain anatomy.

Bibliography: [1] Fonov et al., Neurolmage 2011 - [2] Fonov et al., Neurolmage 2009 - [3] Coupé et al., Neurolmage 2020 - [4] Boré et al., "dcm2bids" 2023 - [5] Jenkinson et al., Neurolmage 2002 [6] Woolrich et al., Neurolmage 2009 - [7] Woodcock et al., J Neuroimaging Psychiatry Neurol. 2018

