

FWI-based quantitative ultrasound computed tomography -perspective for imaging of musculoskeletal organs in children

Elise Doveri, Luis Espinosa, Hermann Konan, Vadim Monteiller, Régine Guillermin, Cécile Baron, Philippe Lasaygues

To cite this version:

Elise Doveri, Luis Espinosa, Hermann Konan, Vadim Monteiller, Régine Guillermin, et al.. FWI-based quantitative ultrasound computed tomography -perspective for imaging of musculoskeletal organs in children. 2024. hal-04741074

HAL Id: hal-04741074 <https://hal.science/hal-04741074v1>

Preprint submitted on 17 Oct 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

FWI-based quantitative ultrasound computed tomography - perspective for imaging of

musculoskeletal organs in children

3 Elise Doveri¹, Luis Espinosa², Kouakou Jean Hermann Konan¹, Vadim Monteiller¹, Régine

- Guillermin¹, Cécile Baron³ and Philippe Lasaygues^{1*}
-
- 1 Aix Marseille Univ, CNRS, Centrale Méditerranée, LMA, Marseille, France
- 2 Institut Clément Ader (ICA), Univ de Toulouse, Univ de Technologie de Tarbes, Tarbes, France
- 3 Aix Marseille Univ, CNRS, Centrale Méditerranée, IRPHE, Marseille, France
- * Corresponding Author,
- Lasaygues Philippe: Laboratory of Mechanics and Acoustics 4 impasse Nikola Tesla, CS 40006,
- 13013 Marseille, France, 33 (0)484 524 283, lasaygues@lma.cnrs-mrs.fr

1. Abstract

 B-mode ultrasound has long been a first-line examination tool for the diagnosis of many musculoskeletal diseases in children. Due to the high acoustic impedance contrast between echogenic bone structures and adjacent soft tissue, B-mode ultrasound can only see the outer surface of bone structures, not what lies inside. In this context, linear ultrasound computed tomography can visualize the different morphologies of small organs (bone structures and adjacent soft tissues), but does not provide quantitative, parametric images. This article proposes a non- linear approach to ultrasound computed tomography using a full waveform inversion algorithm, based on a complete numerical modeling of wave propagation in media and on the minimization, in 22 the L^2 -norm sense, of a functional based on the iterative solution of the inverse problem. Our study was performed in two dimensions, as justified by current conventional experimental setups in medical imaging. We used an acoustic modeling for simplification and computational cost reduction. The inverse problem was solved iteratively using a quasi-Newton method known as the memory-limited Broyden-Fletcher-Goldfarb-Shanno method. The gradient of the misfit function was obtained based on the adjoint state method, which required only two simulations of the wave propagation problem per source at each iteration. Experiments were conducted on a newborn arm phantom containing the humerus, radius and ulna, deep radial and ulnar veins, embedded in homogeneous adipose tissue. We show the images obtained for different configurations of initialization and *a priori* information on the medium: without any *a priori* information on the medium, *a priori* information on the initial map of mass densities, and ultrasonic wave velocities. Convergences were of the order of 10 iterations in practice for each frequency band used, typically 150 kHz to 600 kHz. The normalized error was limited to less than 11%.

2. Keywords

 Ultrasound computed tomography, Full waveform inversion, Musculoskeletal organ imaging, Children's limbs

3. Introduction

 Today, there is a growing interest in children's musculoskeletal health, firstly because children are concerned by musculoskeletal disorders and damages, and specific infantile osteopathologies (such as fractures, bone infections or tumors), and secondly because musculoskeletal health in childhood is of a great importance for musculoskeletal health in adulthood. Therefore, musculoskeletal development in children is a key issue. For pediatricians, radiologists and orthopedic surgeons, the knowledge of the musculoskeletal disorders and bone quality of the child's limbs remains a major objective, in order to anticipate the evolution of diseases, to guide diagnoses and thus define and follow optimal therapeutic strategies.

 B-mode ultrasound has long been a first-line examination tool for the diagnosis of many 48 musculoskeletal diseases in adults and children.¹ The other modalities, such as X-ray or magnetic resonance imaging, are often associated with inconveniences of variable importance (cost, irradiation, sedation, accessibility) that are constraining for pediatric applications. B-mode ultrasound is all the more interesting because the functional advantages associated with it are favorable to pediatric use, as the number of procedures can be increased without (known) harmful effects. But the ultrasonic waves are disturbed by the presence of bony structures and have difficulty penetrating bone. In its clinical use, this modality only makes it possible to see the outer surface of a bony structure, and not what lies within. Also, B-mode ultrasound is not a quantitative and parametric modality. The grayscale levels of the map are not linked to any significant physical parameters of the organs. However, the development of a quantitative ultrasound imaging modality, adapted to the musculoskeletal tissues and organs of children's limbs, as a complementary modality, remains a major challenge.

 The interactions between ultrasonic waves, hyper-echoic bony structures and adjacent hypo-echoic soft tissues, along with the associated propagation problems, have been studied for more than thirty 62 years.² In this context, linear ultrasound computed tomography (USCT) in the orthogonal plane, with several planar projections at different angles, provides images of musculoskeletal organs, at 64 the cost of adapted and time-consuming signal-processing algorithms.^{3,4} For example, linear reflection-mode USCT, with an adapted filtering and wavelet processing algorithm, makes it possible to reach, with a strong contrast-to-noise ratio, the different morphologies of small organs, 67 such as an artificial newborn arm or a chicken drumstick.⁵ With this method one can visualize small X-ray invisible blood vessels in the artificial newborn arm, and, more specifically, muscles, nerves, tibia and the much smaller fibula, fat and skin in the chicken drumstick. But the images obtained are only morphometric, better defined than B-mode ultrasound images, and non-parametric. To quantify the information, it is necessary to change the mathematical paradigm, and deal with the case of non-linear USCT.

 The approach proposed in this work consists in importing the highly sophisticated nonlinear full waveform inversion (FWI), based on the full numerical modeling of wave propagation in media 75 (the object to be imaged and the background). FWI is at the forefront of the field of geophysics, $6,7$ and has been used in medical imaging for a few years now with relevance (3D imaging, increased resolution, limited computation time, multi-parameterized images, deep learning, programming on *GPU* system). The goal of FWI is to iteratively reconstruct physical and elastic/acoustical parameters in the various media, without having to resort to any of the classical limiting assumptions currently made in USCT, for instance that tissue density is spatially constant. Indeed, FWI is an imaging method for heterogeneous media (objects and/or background) that is based on 82 the full numerical modeling of wave propagation in these media, and on the minimization, in the L^2 - norm sense, of the difference between recorded and simulated waveforms. The term "*full*" refers to 84 the use of the full-time series. $8-13$

 FWI has gained in popularity in the medical field because it achieves higher resolution and accuracy than methods that use only part of the information (e.g., the arrival times of the waves) or that are based on approximations (ray tracing, first-order Born approximation, uniform geometrical theory of diffraction, Bojarski's inverse theory, etc.), at the cost of an often high computational burden. The most widely studied application for nonlinear USCT in medicine today is female breast 90 imaging.^{14–16} In the presence of bones with surrounding soft tissues and marrow, solutions consist of iterative and adaptive schemes to integrate the induced nonlinear effects, either by retaining the Radon projection-like geometry and variation shapes of the higher-order Born approximation, or by extending the problem to FWI. Indeed, adapted iterative nonlinear methods have been proposed for objects whose acoustic impedance is strongly contrasted against a homogeneous reference (soft 95 tissues, water), such as the Kaczmarz-model based method, the Distorted Born Iterative 96 method,^{18–21} the Phase Shift Migration method,²² or the Full-Matrix Spectrum-Domain method,^{23–25}. Belanger *et al.* have adapted a hybrid algorithm for the mapping of ultrasound velocity in the 98 cortical bone.²⁶ These algorithms have been tested on numerical and experimental data on a bone- mimicking phantom and an *ex vivo* isolated single bone, without surrounding tissues, or by modeling a single layer (muscle tisue for example), and the results are very encouraging. J. Wiskin *et al.*,^{27–29} worked on a complete knee organ (skin, muscle, and bone) using a full wave 3D inverse

102 scattering transmission-based USCT. This approach is appropriate for the anatomical reconstruction 103 of the bone, but the images do not seem to correctly quantify the wave velocity in the bone.

 This article introduces, after the assumptions involved, the forward problem, the adjoint state method and the Broyden-Fletcher-Goldfarb-Shanno method. Then, the experimental results obtained on a newborn arm phantom studied with an ultrasound scanner are discussed, for measurements of mass density and ultrasonic wave velocity, for several frequencies ranging from 150 kHz to 600 kHz, and three configurations for the initial maps: without any *a priori* information about the phantom, with *a priori* information for the mass density and ultrasonic wave velocity 110 maps.

111 4. Full waveform inversion

112 4.1. Forward problem and adjoint-state method

113 Let us consider *xyz*-space \mathbb{R}^3 with the origin O, and a 2D (*xy*-plan) non-circular cylindrical object to 114 be imaged with generators parallel to the *z*-axis, immersed in a continuous surrounding 115 homogeneous water-like background. The acoustic pressure wave field $p(x; t)$ [x, 2D-space vector, 116 t , time variable] is described by the equation state system:

(1) 117

$$
\rho(x)\partial_t u_x(x;t) = \partial_x p(x;t),
$$

\n
$$
\rho(x)\partial_t u_y(x;t) = \partial_y p(x;t),
$$

\n
$$
\frac{1}{\kappa(x)}\partial_t p(x;t) = \partial_x u_x(x;t) + \partial_y u_y(x;t) + f_s(t),
$$

118 where $\rho(x)$ and $\kappa(x)$ are the spatially-varying mass density and bulk modulus, respectively. u_x and 119 u_y are the components of the particle velocity vector. f_s is the point source term.

120 The spatially variable parameters of the medium to be imaged using FWI are obtained by 121 minimizing the difference between signals recorded at certain positions (receivers) and the 122 corresponding modeled signals for one or more propagation events (sources) [\(Figure 1\)](#page-5-0).

123 **Figure 1: Approximately here**

124 Given R recorded signals, $p_r^{obs}(t) = p^{obs}(x_r; t)$, for $r = 1, ..., R$ receivers, and for a single 125 source, the cost function to minimize writes:

126 (2)
$$
C(m) = \sum_{r=1}^{R} \int_{0}^{T} [p_r^{obs}(t) - p_r^{num}(t; m)]^2 dt,
$$

127 with **m** the parameters of the object to be estimated, for instance, $\rho(x)$, $\kappa(x)$, on a spatial grid. 128 $p_r^{\text{num}}(t; m) = p^{\text{num}}(x_r; t; m)$ is the synthetic signal computed numerically in the inversion 129 process, and \overline{T} is the recorded time period (or time length of the signal).

130 The main classical difficulties encountered in the minimization of $C(m)$ are its significant 131 nonlinearity with respect to m , and thus the presence of numerous local minima. An iterative method based on the calculation of the gradient of the cost function is generally chosen to minimize $C(m)$, either a simple gradient descent method or a more sophisticated quasi-Newton technique, which uses information based on second-order derivatives in order to accelerate convergence. These methods make it possible to converge to the minimum of the function in a larger or smaller number of iterations, depending on the method used and on the complexity of the problem. Nevertheless, 137 the gradient of $C(m)$, i.e., its partial derivatives with respect to each element of m , cannot easily be calculated based on a standard and simple finite-difference approach, because this would require performing as many forward calculations of wave propagation as the number of parameters, and this for each iteration of the minimization algorithm. The adjoining field method, introduced in the 141 seismological domain in the $1970s³⁰$ makes it possible to overcome this difficulty. It can be shown that the gradient of the cost function can be obtained as the convolution product of the forward field with an adjoint field obtained by calculating the time-reversed wave propagation:

144 (3)
$$
r_r(t; m) = p_r^{obs}(-t) - p_r^{num}(-t; m),
$$

 with each receiver acting as a source. Two calculations of the propagation in each medium (object and background) are then sufficient to obtain the gradient (in the case of multiple sources, two calculations per source are needed).

 Minimization algorithms based on the gradient converge to the solution only if they are started close to the global minimum, i.e., provided that the initial model is already of good quality, which is of course rarely the case. In practice, this global minimum is unknown and there is therefore a risk of remaining trapped in a local minimum. In particular, if there is a difference in propagation time greater than half a period between the computed and real data, the algorithm will attempt to match different cycles of the signals and will not be able to converge (a problem known as "*cycle skipping*",¹³). A method to reduce this risk is to use an (incremental) "*frequency-hopping*" approach: first, a low-pass filter is used to retain only the lowest frequencies and then, once a minimum is 156 obtained, the maximum frequency is slightly increased and the process is run again, until all the 157 frequency information has been included. 31

158 The FWI method means that the model sought must minimize the mean squared difference between 159 the numerical synthetic time series and the observed basic time series to be fitted:

160 (4)
$$
\chi(m) = \sum_{s=1}^{S} \sum_{r=1}^{R} \int_{0}^{T} \frac{1}{2} ||p_{s,r}^{obs}(t) - p_{s,r}^{num}(t; m)||^{2} dt.
$$

161 In other words, the goal is to find a model of the object that can explain a large portion of the 162 recorded signals. This is basically the reason why drastically-enhanced pictures can be obtained. 163 This functional quantifies the L^2 difference between the observed waveforms, $p_{s,r}^{\text{obs}}$ 164 $p^{obs}(x_r, x_s; t)$ at receivers x_r , $r = 1, ..., R$, produced by sources at x_s , $s = 1, ..., S$, and the 165 corresponding synthetic time series, $p_{s,r}^{\text{num}}(t; m) = p^{\text{num}}(x_r, x_s; t; m)$ computed in model m. 166 In the vicinity of m , the misfit function can be expanded into a Taylor series:

167 (5)
$$
\chi(m + \delta m) \approx \chi(m) + g(m) \cdot \delta m + \delta m^{\dagger} \cdot H(m) \cdot \delta m,
$$

where $g(m) = \frac{\partial}{\partial m}$ $\frac{\chi(m)}{\partial m}$ is the gradient of the waveform misfit function, $H(m) = \frac{\partial^2 m}{\partial m^2}$ 168 where $g(m) = \frac{\partial \chi(m)}{\partial m}$ is the gradient of the waveform misfit function, $H(m) = \frac{\partial \chi(m)}{\partial m^2}$ the 169 Hessian, and \dagger denotes the transpose. The nearest minimum of χ in equation [\(5\)](#page-7-0) with respect to the 170 model perturbation δm is reached for $\delta m = -H^{-1}$. g. The local minimum of equation [\(4\)](#page-7-1) is thus 171 given by perturbing the model in the direction of the gradient preconditioned by the inverse 172 Hessian. A direct method to compute the gradient is to take the derivative of equation [\(4\)](#page-7-1) with 173 respect to model parameters:

174 (6)
$$
\frac{\partial \chi(m)}{\partial m} = -\sum_{s=1}^{S} \sum_{r=1}^{R} \int_{0}^{T} \frac{\partial p_{s,r}^{num}(t; m)}{\partial m} \left[p_{s,r}^{num}(t; m) - p_{s,r}^{obs}(t) \right] dt.
$$

175 This equation can be reformulated as the matrix-vector product, $g = -J^* \delta d$, where J^* is the adjoint of the Jacobian matrix of the forward problem that contains the Fréchet derivatives of the 177 data with respect to model parameters, δd is the vector that contains the data residuals. The 178 determination of *J* would require computing the Fréchet derivatives for each time step in the time window considered and for all the source-station pairs, which is completely prohibitive on current computers, even the latest and greatest ones. As mentioned above, it is however, and fortunately, possible to obtain this gradient without computing the Jacobian matrix explicitly. The idea is to resort to the adjoint state, which corresponds to the wave field transmitted and back-propagated

183 from the receivers The approach to determine the gradient without computing the Fréchet 184 derivatives was introduced in nonlinear optimization by Chavent et al. in the $1970s$.^{30,32} The 185 Lagrangian functional is the cost function augmented by the following constraint given by the 186 equation of state, 33 :

(7) 187 (7) $\mathcal{L}(u, u^a, \rho, \kappa) =$

$$
\sum_{s=1}^{S} \left\{ \sum_{r=1}^{R} \int_{0}^{T} \left[p_{s,r}^{num}(t; m) - p_{s,r}^{obs}(t) \right]^{2} dt - \int_{0}^{T} \int_{x} \int_{y} u^{a}(x;t) \left[\sigma(\rho; \kappa) \partial_{t} u - A \partial_{x} u - B \partial_{y} u - F_{s} \right] dy dx dt \right\}
$$

188 with
$$
\mathbf{u} = (u_x, u_y, p)^{\dagger}, \sigma(p; \kappa) = \begin{pmatrix} \rho & 0 & 0 \\ 0 & \rho & 0 \\ 0 & 0 & 1/\kappa \end{pmatrix}, \mathbf{A} = \begin{pmatrix} 0 & 0 & 1 \\ 0 & 0 & 0 \\ 1 & 0 & 0 \end{pmatrix}, \mathbf{B} = \begin{pmatrix} 0 & 0 & 0 \\ 0 & 0 & 1 \\ 0 & 1 & 0 \end{pmatrix},
$$

- 189 and $\mathbf{F}_s = (0,0,f_s)^{\dagger}$.
- 190 $u^a(x;t)$ is the Lagrange multiplier to be found by zeroing the partial derivative of the Lagrangian 191 functional, equation [\(7\)](#page-8-0), with respect to the wave field, \boldsymbol{u}^{34} .
- 192 By zeroing the misfit function, equation [\(7\)](#page-8-0), with respect to the model parameters, ρ , κ , the gradient 193 of the misfit function can then be found:

194 (8)
$$
\partial_{\rho,\kappa}\chi(\rho,\kappa) = \frac{1}{2}\sum_{s=1}^{S} \left\{ \int_0^T \int_x \int_y u_s^a(t) \partial_{\rho,\kappa}\sigma(\rho,\kappa) \partial_t u_s(t) \cdot (\delta \rho, \delta \kappa) \, dy dx dt \right\} = 0,
$$

195 which gives the expression of the gradient kernels for ρ and κ in terms of direct \boldsymbol{u} and the adjoint 196 $\boldsymbol{u}^{\text{a}}$ field:

197 (9)
$$
\nabla_{\rho} \chi(\rho, \kappa) = \sum_{s=1}^{S} \int_0^T \langle u_x^a(t) \partial_t u_x(t) + u_y^a(t) \partial_t u_y(t) \rangle_s dt,
$$

198 **(10)**
$$
\nabla_{\kappa} \chi(\rho, \kappa) = \sum_{s=1}^{S} \int_0^T \frac{p_s^a(t)}{\kappa(x)^2} \partial_t p_s(t),
$$

199 with p^a is the adjoint acoustic pressure wave field. As the above equations show, the principle of the adjoint-state method is to correlate two wave fields: the forward field propagating from the sources to the receivers, and the adjoint wavefield propagating from the receivers backwards in time. Calculating the gradient of the misfit function therefore requires only two simulations of the wave propagation problem per source.

 λ

204 4.2. The Limited-memory Broyden-Fletcher-Goldfarb-Shanno method

205 Gradient descent is the most commonly used iterative optimization algorithm for finding the local 206 minimum of a function:

207 (11) $m_{k+1} = m_k - \alpha g(m_k)$,

with \bm{m} ρ $\binom{\kappa}{\kappa_k}$ the parameters of the object at iteration k, $g(m_k)$ $\nabla_{\rho} \chi(m_k)$ 208 with $m_k = \binom{r_k}{k_k}$ the parameters of the object at iteration k , $g(m_k) = \binom{p_k}{\sum_{k} (m_k)}$ the gradient of 209 the cost function at iteration k, and $\alpha \in \mathbb{R}_+$ the step size.

 The parameter is updated by perturbing its value in the direction opposite to that of the gradient. 211 This method converges slowly and depends on the learning rate of the descent step α . The smaller the step size, the greater the chance of the algorithm converging, which considerably increases convergence time. This method is therefore not recommended for complex problems requiring high computing power to calculate the cost function and its gradient. In contrast to the gradient descent method, the Limited-memory Broyden-Fletcher-Goldfarb-Shanno (L-BFGS) method has a higher convergence speed, and the direction of perturbation is guided by the computation of the Hessian 217 matrix H , which is the matrix of partial second derivatives. This method is obtained from a second order Taylor series development.

219 The second-order Taylor series expansion of a function χ infinitely differentiable in m_k is:

220 (12)
$$
\chi(m) = \chi(m_k) + (m - m_k) * \nabla \chi(m_k) + \frac{1}{2}(m - m_k) * H * (m - m_k),
$$

221 where $*$ denotes the matrix multiplication. The derivative with respect to m is then written:

$$
222 \qquad (13) \qquad \qquad \nabla \chi(m) = \nabla \chi(m_k) + H * (m - m_k).
$$

223 An extremum is reached for $\nabla \chi(\mathbf{m}) = 0$. The final result is:

224 (14)
$$
m = m_k - H^{-1} \nabla \chi(m_k)
$$
.

225 And so, the numerical scheme of the L-BFGS method is defined by:

226 (15)
$$
m_{k+1} = m_k - H^{-1}(m_k)g(m_k),
$$

227 with H^{-1} the inverse of the Hessian matrix:

228 (16)
$$
H = \begin{bmatrix} \nabla_{\kappa^2} \chi(m_k) & \nabla_{\kappa} \nabla_{\rho} \chi(m_k) \\ \nabla_{\rho} \nabla_{\kappa} \chi(m_k) & \nabla_{\rho^2} \chi(m_k) \end{bmatrix}.
$$

229 Calculating the inverse of the Hessian matrix requires a large storage capacity (theoretical value,³⁵)

230 of around 600 GB in this work) and is time-consuming. To overcome these difficulties, the matrix 231 is calculated iteratively, resulting in the calculation of a single matrix, updated at each iteration as 232 follows:

233 (17)
$$
H_k = H_{k-1} + \frac{y_{k-1}y_{k-1}^{\dagger}}{y_{k-1}^{\dagger}d_{k-1}} - \frac{H_{k-1}d_{k-1}d_{k-1}^{\dagger}H_{k-1}}{d_{k-1}^{\dagger}H_{k-1}d_{k-1}}
$$

234 with $d_{k-1} = m_k - m_{k-1}, y_{k-1} = g(m_k) - g(m_{k-1}).$

235 Using this iterative method ensures that when an extremum is reached, it is a minimum, as it 236 preserves the symmetry and positivity properties of a matrix. (i.e., $\chi(x_0)$ is minimum of χ only if 237 $\nabla \chi(x_0) = 0$ and $H[\chi(x_0)]$ is a positive-definite function).

238 4.3. Discretization

239 For their numerical implementation, the above equations need to be discretized both in time and 240 space. We resort to a velocity-stress finite-difference approximation of the wave equation, with 241 explicit conditionally-stable time stepping in a staggered 2D Cartesian grid.³⁶ However, it is 242 important to note that other numerical approximations could be used, for instance the spectral-243 element method.³⁷

- 244 The time and space coordinates are discretized as follows:
- 245 (18) $t = l\Delta t$,
- 246 (19) $x = ih$,
- 247 **(20)** $y = jh$,

248 with Δt and h the time and space steps, respectively, and $l = 0, ..., L - 1, i = 0, ..., I - 1$, and 249 $j = 0, ..., J - 1$, temporal and spatial indices such as:

- 250 **(21)** $T = L\Delta t$,
- 251 **(22)** $\eta_x = Ih$,
- 252 **(23)** $n_v = Jh$.

253 In the spatial domain, the isotropic grid step h is chosen to obtain smooth, less pixelated images, 254 α according to the criterion of Bernard *et al.*³⁴:

$$
255 \qquad (24) \qquad \qquad \boldsymbol{h} \approx \frac{\lambda_{min}}{8},
$$

256 λ_{min} being the smallest wavelength obtained at the highest frequency. In the time domain, the step 257 Δt must respect the Courant-Friedrichs-Lewy (CFL) stability condition:

$$
258 \quad (25) \qquad \Delta t = \mathbf{CFL} \frac{h}{\sqrt{2}V_{max}},
$$

259 with CFL = 0.495 and V_{max} the highest velocity of ultrasonic waves in the object.

260 To avoid wave reflection on the domain contours, a Perfectly Matched absorbing Layer (PML), 38 is 261 considered around the simulation grid. It is used to model wave propagation in the infinite domain 262 by means of PML coefficients, which are zero inside the physical domain and non-zero outside it, 263 and an integer N_{PML} that is used to set the layer width [\(Figure 2\)](#page-11-0).

264 **Figure 2: Approximately here**

265 In this discretized framework, the cost function involves a sum over the time steps instead of a time 266 integral:

267 (26)
$$
\nabla_{\rho} \chi(m) = \sum_{s=1}^{S} \sum_{l=0}^{L-1} \langle u_x^a(l\Delta t) \partial_t u_x(l\Delta t) + u_y^a(l\Delta t) \partial_t u_y(l\Delta t) \rangle_s,
$$

268 (27)
$$
\nabla_{\kappa} \chi(m) = - \sum_{s=1}^{S} \sum_{l=0}^{L-1} \frac{p_s^a(l\Delta t)}{\kappa(x)^2} \partial_t p_s(l\Delta t),
$$

269 p_s , p_s^a , u_x , u_x^a , u_y and u_y^a are obtained by simulation of the forward problem at time $l\Delta t$ and points 270 (x, x_s) .

271 Using the trivially applied Chain Rule, the previous gradient can be expressed in terms of (ρ, v_n) , 272 with v_p the compression wave (ultrasonic wave) velocity (which from now on will be called US 273 velocity). The parameters to be achieved in this study are:

274 (28)
$$
\nabla_{\rho} \chi(\rho, v_p) = \nabla_{\kappa} \chi(\rho, \kappa) v_p^2(x) + \nabla_{\rho} \chi(\rho, \kappa)
$$

$$
275 \quad (29) \qquad \qquad \nabla_{v_p} \chi(\rho, v_p) = 2 \nabla_{\kappa} \chi(\rho, \kappa) \rho(x) v_p(x)
$$

276 $\nabla_{\kappa} \chi$, $\nabla_{\rho} \chi$ and $\nu_{\rm p}$ are respectively the bulk modulus gradient, the mass density gradient, and the US 277 velocity map.

278 4.4. Regularization

 The presence of electronic noise (or other noise, such as thermal noise) is handled adding a regularization term to the cost function. There are several types of regularization, depending on the assumptions made about the parameter to be optimized. For example, a Tikhonov-type regularization is suitable for smoothly varying signals, while a total variation (TV) regularization is suitable for rapidly varying signals. For the purposes of this project, total variation regulation has been adopted and the regulation term is formulated as follows:

285 (30)
$$
TV(m) = \sum_{i=0}^{I-1} \sum_{j=0}^{J-1} \sqrt{\frac{1}{h^2} \Big[\Big(\big[v_p \big]_{i+1,j} - \big[v_p \big]_{i,j} \Big)^2 + \Big(\big[v_p \big]_{i,j+1} - \big[v_p \big]_{i,j} \Big)^2 \Big] + \epsilon_{TV}}
$$

$$
+\sqrt{\frac{1}{h^2}\Big[\Big([\rho]_{i+1,j}-[\rho]_{i,j}\Big)^2+\Big([\rho]_{i,j+1}-[\rho]_{i,j}\Big)^2\Big]+\epsilon_{\text{TV}}}
$$

286 with $\epsilon_{\text{TV}} = 10^{-8.34}$ The formulation of the forward problem is defined as follows:

$$
287 \quad (31) \quad \chi'(m) = \chi(m) + \beta T V(m)
$$

288 with β a variable for adjusting the weight of the regularization term. After a number of tests, the 289 value retained for β is 10^{-15 34}

290 4.5. Precision of the reconstruction

291 To evaluate the precision of the reconstruction, the Normalized Root Mean Square Error (NRMSE) 292 is computed to quantify the difference between theoretical US velocity $v_p^{\text{th}}(x)$ and mass density 293 $\rho^{\text{th}}(x)$ maps, and estimated ones $[v_{\text{p}}(x)]$ and $\rho(x)$], as follows³⁹:

294 (32)
$$
NRMSE(v_p) = \sqrt{\frac{1}{lJ}\sum_{i=0}^{I-1}\sum_{j=0}^{J-1}\frac{([v_p^{th}]_{i,j} - [v_p]_{i,j})^2}{[v_p^{th}]_{i,j}^2}},
$$

295 (33)
$$
NRMSE(\rho) = \sqrt{\frac{1}{IJ} \sum_{i=0}^{I-1} \sum_{j=0}^{J-1} \frac{(\left[\rho^{th}\right]_{i,j} - [\rho]_{i,j})^2}{\left[\rho^{th}\right]_{i,j}^2}}
$$

5. Experimental setup and sample

5.1. USCT scanner

 The mechanical ultrasound scanner used consisted of two bistatic arms articulated in the horizontal plane [\(Figure 3\)](#page-13-0). Each arm carried two transverse carriages allowing the parallel translation of the transducers, for the correct adjustment of the ultrasonic beam to targets. The rotation of the target holder allowed an angular scan. In this work, (*S* x *R*) signals were recorded over 360°. All motions were insured by stepper motors sequentially driven by a programmable translator-indexer device fitted with a power multiplexer. The translator-indexer device and power multiplexer were integrated in a control rack that also included other remote controls, such as the distance adjustment of the arms, or the out-of-water setting system.

Figure 3: Approximately here.

5.2. Electro-acoustic devices

 The transducers were Ultran™ transducers with a center frequency of 648 kHz, and the 6dB- bandwidth is 515 kHz (from 375 kHz to 890 kHz) [\(Figure 4\)](#page-13-1). They were driven using a pulse- receiver generator (Sofranel 5077 PR, Olympus, Waltham, MA 02453, USA). The radio-frequency signals (RF-signals) were conveyed from the 12-bit oscilloscope (Lecroy HDO 6104, Teledyne Inc., Thousand Oaks, CA, USA) to a personal computer using a USB interface file transfer, and stored. 313 The signal and image processing algorithms were implemented using Matlab[®] (The MathWorks, Inc., Natick, MA, USA).

Figure 4: Approximately here

5.3. Newborn arm phantom

 The object studied was a phantom of a newborn baby's arm (True Phantom Solutions Inc., Windsor, Canada). This phantom, initially used for ultrasound guidance training for nurses and doctors, was composed of 3 different materials representing adipose soft tissue, artificial blood vessels and bone. All media were considered isotropic and homogeneous. Only compression waves 321 were propagated (no shear waves). Measurements on the phantom were taken at a distance $H = 10$ cm (\pm 0.2 cm) cm from the tip of the hand (index finger) [\(Figure 3\)](#page-13-0). The blood vessels were made of a flexible material containing a fluid that mimics blood. The bone part was made of a composite

 material of urethane reinforced with ceramic powder. It had a mass density similar to that of human bone, but US velocity was much lower and attenuation higher. Some of the properties of these materials are presented in [Table 1.](#page-14-0) Data for soft (adipose) tissue, vessels and blood were provided by the manufacturer. US velocity values in bone were determined using an interferometric 328 method,⁴⁰ on a block-sample (10 x 10 x 2 cm) of material identical to that constituting the phantom bone. Mass densities were measured based on Archimedes' principle with a micrometric balance and a density kit (Voyager 610 GX, Ohaus Corporation, Florham Park, NJ, USA).

331 **Table 1: Approximately here.**

332 5.4. Experimental and numerical parameters

 The positions of the transmitter and the receiver transducers (*S* x *R*) described two concentric circles with identical centers of rotation, corresponding to the position of the object studied, but with different diameters (since the transducers cannot be in the same place). The following parameters correspond to those of the experiment and the numerical simulation:

- 337 $S = 72$ sources; $R_s = 22.7$ cm
- 338 $R = 180$ receivers; $R_r = 36.2$ cm

$$
339 \qquad \bullet \quad \text{CFL} = 0.495
$$

- 340 $N_{\text{PML}} = 16$
- 341 $h = 0.14$ mm
- 342 $I = 5447; \eta_x = 2R_r + 4 = 76.4$ cm
- 343 $J = 5447; \eta_y = 2R_r + 4 = 76.4 \text{ cm}$

344 Adding 4 cm allows the model to fully recover the spatial domain.

345 •
$$
T = 560 \text{ }\mu\text{s}; \Delta t = 50 \text{ ns}
$$

- 346 $V_{\text{water}} = 1467 \text{ m/s}$
- 347 \bullet $\rho_{\text{water}} = 1000 \text{ kg/m}$
- 348 $Z_water = 1.47$ MRayls
- 349 In this work, $V_{\text{max}} = 1600 \text{ m/s} > V_{\text{water}} = 1467 \text{ m/s}$ and $V_{\text{min}} = V_{\text{bone}} = 1129 \text{ m/s}$ 350 which are respectively US velocity in water and in the material composing the bone.
- 351 $f_{\text{max}} = 1 \text{ MHz}$
- $\lambda_{\min} = \frac{V}{f}$ 352 • $\lambda_{\min} = \frac{v_{\min}}{f_{\max}} = 1.129$ mm

353 • For the incremental frequency-hopping method, the cut-off frequency of the filter was increased in several steps, starting at 150 kHz, then 250 kHz, 400 kHz and finally 600 kHz.

5.5. High performance computing resources

 The inversion was done at the Centre de Calcul Intensif d'Aix-Marseille (Mesocentre), using 32 nodes. Each 384 node was a Dell PowerEdge C6420 and contained 32 cores CPU Intel Xeon Gold 6142 (Sky Lake) 2.6 GHz and had 192 Gigabytes RAM. The computing was done using multiple cores by dividing the spatial domain into 1024 pieces (32386 x 32) and each piece of the domain was computed with one core. The computation took 12 hours for one 387 frequency, so 2 days in total.

6. Results

6.1. Methodology

The methodology adopted for this work is as follows:

- 1. The reference images, which were the objective, were defined by extracting the contours and binarizing the X-ray images of the object.
- 2. Initial maps were purely numerical, reconstructed from simulated data derived from previous reference maps.
- 3. For reconstructions from experimental data, three tests were carried out: without any *a priori* information on the object; by initializing the method with a mapping of *a priori* values for mass density, then another for US velocity.
- For each configuration, NRMSE were evaluated [\(Figure 9,](#page-16-0) [Figure 12,](#page-16-1) [Figure 16,](#page-17-0) [Figure 20\)](#page-17-1).

6.2. Reference images

 The first stage of the protocol involved performing X-ray micro-computed tomography (μCT) (EasyTom XL 150 "Mechanic Ultra", RX Solutions, Chavanod, France) of the newborn arm

- phantom [\(Figure 5\)](#page-15-0). Full details of the protocol are presented in Doveri et al..⁵
- **Figure 5: Approximately here**

 From the previous X-ray image, two theoretical maps were defined for each of the two properties 379 studied in this work, US velocity, v_p , and mass density, ρ , [\(Figure 6\)](#page-16-2). These images will be the expected reference images for FWI.

Figure 6: Approximately here

6.3. Reconstructions from simulated data

 In this section, data inversion was performed [\(Figure 7\)](#page-16-3) on simulated data by solving the forward problem for theoretical US velocity and mass density maps, without any initial *a priori* information 385 on the different media. Starting maps were initialized with measured values for water (i.e., $v_p =$ 386 1467 m/s and $\rho = 1000 \text{ kg/m}^3$). [Figure 8](#page-16-4) shows the gradient of the cost function at the last iteration with respect to the US velocity and the mass density.

- **Figure 7: Approximately here**
- **Figure 8: Approximately here**
- **Figure 9: Approximately here**
- 6.4. Reconstruction from experimental data
- $392 \rightarrow$ No *a priori* information on the different media

 In this section, data inversion was performed [\(Figure 10\)](#page-16-5) on experimental data without any signal processing, by solving the forward problem for theoretical US velocity and mass density maps, without any initial *a priori* information on the different media. Starting maps were initialized with 396 measured values for water (i.e., $v_p = 1467$ m/s and $\rho = 1000$ kg/m³). [Figure 11](#page-16-6) shows the gradient of the cost function at the last iteration with respect to the US velocity and the mass density.

- **Figure 10: Approximately here**
- **Figure 11: Approximately here**
- **Figure 12: Approximately here**
- *A priori* information on mass densities

 In this section, data inversion was performed [\(Figure 14\)](#page-17-2) on experimental data without any signal processing, by solving the forward problem for theoretical US velocity and mass density maps, starting from the theoretical map of mass densities [\(Figure 13\)](#page-17-3). Starting US velocity map was

- 405 initialized with measured values for water (i.e., $v_p = 1467$ m/s). [Figure 15](#page-17-4) shows the gradient of the
- cost function at the last iteration with respect to the US velocity and the mass density.
- **Figure 13: Approximately here**
- **Figure 14: Approximately here**
- **Figure 15: Approximately here**
- **Figure 16: Approximately here**
- 411 \rightarrow A priori information on US velocities

 In this section, data inversion was performed [\(Figure 18\)](#page-17-5) on experimental data by solving the forward problem for theoretical US velocity and mass density maps, starting from the theoretical US velocity map [\(Figure 17\)](#page-17-6). Starting mass density map was initialized with measured values for 415 water (i.e., $\rho = 1000 \text{ kg/m}^3$). [Figure 19](#page-17-7) shows the gradient of the cost function at the last iteration with respect to the US velocity and the mass density.

- **Figure 17: Approximately here**
- **Figure 18: Approximately here**
- **Figure 19: Approximately here**
- **Figure 20: Approximately here**

7. Discussion

 [Table 2](#page-17-8) shows the normalized root mean square errors (NRMSE) calculated for each of the images presented at the last inversion iteration.

Table 2: Approximately here

 The error is more important on the mass density maps than on the US velocity maps. It can be explained by the fact that the US velocity contrasts affect every wave (through time shifts for transmitted waves and through amplitude variations for diffracted waves) while the mass density contrasts only affect reflection coefficients at the interfaces. In linear USCT, this phenomenon is well known, and the transfer functions linked to physical parameters such as mass density, compressibility or acoustic impedance, are a function of the diffraction angle. For example, linear

431 ultrasonic impediography is reflection-mode USCT (backscatter imaging) 3.5 . In this work, the small 432 angles of observation (reflected waves) contain information on mass density contrasts ³⁴. This can be clearly seen on the US velocity and mass density maps obtained from the inversion of simulated data [\(Figure 6\)](#page-16-2), where the US velocity map is accurate whereas the mass density map is not, 435 especially in the bone part. The US velocity (~1115 m/s) is closer than the theoretical velocity 436 (1129 m/s) while the mass density $(\sim 1050 \text{ kg/m}^3)$ is half the theoretical mass density (2160 kg/m³). This can also be seen on the gradients of the cost function [\(Figure 8\)](#page-16-4), where the gradient with respect to the mass density (b) oscillates more than the gradient with respect to the US velocity (a).

 As explained in the theoretical section, gradient-based minimization algorithms only converge to the solution if they start close to the global minimum, i.e., provided the initial model is already of good quality. But we know that this is rarely the case. However, in this work, to improve the process, *a priori* information was used for the initial mass density map, which is easier to estimate from the numerous patient X-rays. We used a mass density starting map based on the image 444 obtained from the binarized X-ray μ CT. Starting the inversion process with a knowledge of the mass density greatly improves the result: the contours of the bone are well defined and the US velocity is close to the theoretical velocity (~1250 m/s on the external part of the bone), and the 447 mass density remains close to the theoretical mass density $\left(\sim 2000 \text{ kg/m}^3\right)$ in the bone part).

 However, starting the inversion with *a priori* information on the US velocity does not help much in improving the resulting maps. As the iterations progress, the US velocity map diverges from the theoretical velocity map (the NRMSE increases on [Figure 20\)](#page-17-1), while the mass density map remains close to the one obtained without any *a priori* information. It seems that the inversion falls in a local minimum: the cost function remains stable after 5 iterations [\(Figure 20\)](#page-17-1) and it stops decreasing at the last frequency (600 kHz).

 The difficulties encountered can also arise from the object in itself, especially the material composing the bone part of the newborn arm phantom used in this work. On the one hand, this material greatly attenuates the ultrasonic wave (see [Table 1\)](#page-14-0), and the attenuation is not taken into account in the forward modeling during the inversion. On the other hand, the US velocity contrast between the material composing the bone and the other materials composing the phantom is negative and below 1:

$$
460 \t(34) \t\t \t\t \frac{V_{bone}-V_{tissues}}{V_{tissues}} = \frac{1129}{1400} - 1 \approx -0.2,
$$

 while for a real bone it would be around 1.5 for a velocity around 3500 m/s. It is difficult to know to what extent this negative contrast is causing the algorithm to fail, but the inversion process remains stuck in a local minimum, where the US velocity is much higher than the theoretical velocity (1129 m/s).

8. Conclusion

 For several years now, linear ultrasound computed tomography (or, in clinical terms, B-mode ultrasound) has provided qualitative images of musculoskeletal tissue in children. To go one step further, and offer radiologists parametric images relating to the elastic properties of musculoskeletal tissue, a (diffraction-mode) acoustical imaging method based on a full waveform inversion (FWI) algorithm was introduced in this work. The full waveform inversion is an imaging method for heterogeneous media that is based on full numerical modeling of wave propagation in the heterogeneous medium, and on the minimization of the difference between recorded and simulated waveforms based on the iterative resolution of an inverse problem for the estimation of the parameters of the medium. The results presented in this work on a mimetic newborn arm phantom are quite encouraging, although this object was not initially intended for ultrasonic imaging and its material properties are far from those of a real newborn arm. The reconstruction error can be estimated at less than 11%. However, it would be interesting to test this algorithm on other objects, 478 with more realistic acoustic properties, such as the chicken drumstick used par Doveri et al.⁵. Future work should also involve viscoelastic materials, not only acoustic materials, because, especially in the case of bone, compressional and shear wave velocities are of the same order of magnitude. Then, it would be necessary to improve the forward model in order to take shear waves and attenuation into account. However, this would raise some issues regarding the computational cost and the interface hopping between fluid and solid media.

 Signal acquisition and processing could also be improved. In this work, we used few positions for receivers (72) and transmitters (180), and no processing of the recorded signals. More data would have improved the results by increasing redundant angular information. Similarly, only one frequency was used, using conventional piezoelectric transducers, whereas piezo-composite technology could improve the signal-to-noise ratio by over 30 dB. In parallel, suitable signal processing such as the wavelet-based coded excitation (WCE) method should further improve the initial observed data.

 When it can be made to work, full waveform inversion offers significantly improved resolution, compared with simpler methods based for example on travel times only or on a linearization of propagation in the object (Born approximation), at the cost of longer calculation time. But the computational cost issue tends to be less limiting nowadays, thanks to the continuous increase in the power of computers, easier access to high-performance computing systems, and significant progress in numerical methods, all of which making it possible to apply elastic or viscoelastic full waveform inversion to real seismic data, 41 . Moreover, the understanding of the advantages and disadvantages of the method (especially with regard to the convergence problems discussed above) has increased considerably in recent years.

9. References

1. Griffith, J.F. (2019). Diagnostic Ultrasound: Musculoskeletal 2nd Edition. (Amirsys Publishing).

 2. Laugier, P., and Grimal, Q. eds. (2022). Bone Quantitative Ultrasound: New Horizons, Springer International Publishing.

 3. Lasaygues, P., Espinosa, L., Bernard, S., Petit, P., and Guillermin, R. (2022). Ultrasound Computed Tomography. In Bone Quantitative Ultrasound: New Horizons Advances in Experimental Medicine and Biology., P. Laugier and Q. Grimal, eds. (Springer International Publishing), pp. 227–250.

 4. Shortell, M.P., Althomali, M.A.M., Wille, M.-L., and Langton, C.M. (2017). Combining Ultrasound Pulse- Echo and Transmission Computed Tomography for Quantitative Imaging the Cortical Shell of Long-Bone Replicas. Front. Mater. *4*, 40.

 5. Doveri, E., Sabatier, L., Long, V., and Lasaygues, P. (2021). Reflection-Mode Ultrasound Computed Tomography Based on Wavelet Processing for High-Contrast Anatomical and Morphometric Imaging. Appl. Sci. *11*, 9368.

6. Chen, P., and Lee, E.-J. (2015). Full-3D Seismic Waveform Inversion, Springer International Publishing.

 7. Fichtner, A. (2011). Full Seismic Waveform Modelling and Inversion (Springer Berlin Heidelberg) https://doi.org/10.1007/978-3-642-15807-0.

- 8. Fichtner, A., and Trampert, J. (2011). Resolution analysis in full waveform inversion: Resolution in full waveform inversion. Geophys. J. Int. *187*, 1604–1624.
- 9. Kan, L.-Y., Chevrot, S., and Monteiller, V. (2022). A consistent multiparameter Bayesian full waveform inversion scheme for imaging heterogeneous isotropic elastic media. Geophys. J. Int. *232*, 864–883.
- 10. Monteiller, V., Chevrot, S., Komatitsch, D., and Wang, Y. (2015). Three-dimensional full waveform inversion of short-period teleseismic wavefields based upon the SEM-DSM hybrid method. Geophys. J. Int. *202*, 811–827.
- 11. Tarantola, A. (1984). Inversion of seismic reflection data in the acoustic approximation. GEOPHYSICS *49*, 1259–1266.
- 12. Tromp, J., Komatitsch, D., and Liu, Q. (2008). Spectral-Element and Adjoint Methods in Seismology. Commun. Comput. Phys., 1–32.
- 13. Virieux, J., and Operto, S. (2009). An overview of full-waveform inversion in exploration geophysics. GEOPHYSICS *74*, WCC1–WCC26.
- 14. Pérez-Liva, M., Herraiz, J.L., Udías, J.M., Miller, E., Cox, B.T., and Treeby, B.E. (2017). Time domain reconstruction of sound speed and attenuation in ultrasound computed tomography using full wave inversiona). J. Acoust. Soc. Am. *141*, 1595–1604.
- 15. Sandhu, G.Y., Li, C., Roy, O., Schmidt, S., and Duric, N. (2015). Frequency domain ultrasound waveform tomography: breast imaging using a ring transducer. Phys. Med. Biol. *60*, 5381–5398.
- 16. Wang, K., Matthews, T., Anis, F., Li, C., Duric, N., and Anastasio, M. (2015). Waveform inversion with source encoding for breast sound speed reconstruction in ultrasound computed tomography. IEEE Trans. Ultrason. Ferroelectr. Freq. Control *62*, 475–493.
- 17. Salehi, L., and Schmitz, G. (2014). Nonlinear reconstruction of the speed of sound in soft tissues: A comparison between the simulation results applying Kaczmarz and Contrast Source Inversion methods. In (IEEE), pp. 2221–2224.
- 18. Guillermin, R., Lasaygues, P., and Rabau, G. (2015). Quantitative Ultrasonic Imaging of Bones. In The 22nd International Congress on Sound and Vibration IIAV. (IIAV), pp. 1–6.
- 19. Guillermin, R., Lasaygues, P., Rabau, G., and Lefebvre, J.-P. (2013). Quantitative non-linear ultrasonic imaging of targets with significant acoustic impedance contrast—An experimental study. J. Acoust. Soc. Am. *134*, 1001–1010.
- 20. Lasaygues, P., Guillermin, R., and Lefebvre, J.-P. (2006). Distorted Born diffraction tomography applied to inverting ultrasonic field scattered by noncircular infinite elastic tube. Ultrason. Imaging *28*, 211–229.
- 21. Lavarello, R., and Oelze, M. (2008). A study on the reconstruction of moderate contrast targets using the distorted born iterative method. IEEE Trans. Ultrason. Ferroelectr. Freq. Control *55*, 112–124.
- 22. Olofsson, T. (2010). Phase shift migration for imaging layered objects and objects immersed in water. IEEE Trans. Ultrason. Ferroelectr. Freq. Control *57*, 2522–2530.
- 23. Li, Y., Shi, Q., Liu, Y., Gu, M., Liu, C., Song, X., Ta, D., and Wang, W. (2021). Fourier-Domain Ultrasonic Imaging of Cortical Bone Based on Velocity Distribution Inversion. IEEE Trans. Ultrason. Ferroelectr. Freq. Control *68*, 2619–2634.
- 24. Wu, H., Chen, J., Yang, K., and Hu, X. (2016). Ultrasonic array imaging of multilayer structures using full matrix capture and extended phase shift migration. Meas. Sci. Technol. *27*, 045401.
- 25. Zhou, C., Xu, K., and Ta, D. (2023). Frequency-domain full-waveform inversion-based musculoskeletal ultrasound computed tomography. J. Acoust. Soc. Am. *154*, 279–294.
- 26. Falardeau, T., and Belanger, P. (2018). Ultrasound tomography in bone mimicking phantoms: Simulations and experiments. J. Acoust. Soc. Am. *144*, 2937–2946.
- 27. Wiskin, J., Malik, B., Borup, D., Pirshafiey, N., and Klock, J. (2020). Full wave 3D inverse scattering transmission ultrasound tomography in the presence of high contrast. Sci. Rep. *10*, 20166.
- 28. Wiskin, J.W., Malik, B., Natesan, R., Pirshafiey, N., Klock, J., and Lenox, M. (2019). 3D full inverse
- scattering ultrasound tomography of the human knee (Conference Presentation). In Medical Imaging 2019: Ultrasonic Imaging and Tomography, N. V. Ruiter and B. C. Byram, eds. (SPIE), p. 25.
- 29. Wiskin, J.W., Borup, D.T., Iuanow, E., Klock, J., and Lenox, M.W. (2017). 3-D Nonlinear Acoustic Inverse Scattering: Algorithm and Quantitative Results. IEEE Trans. Ultrason. Ferroelectr. Freq. Control *64*, 1161–1174.
- 30. Bamberger, A., Chavent, G., Hemon, Ch., and Lailly, P. (1982). Inversion of normal incidence seismograms. GEOPHYSICS *47*, 757–770.
- 31. Bunks, C., Saleck, F.M., Zaleski, S., and Chavent, G. (1995). Multiscale seismic waveform inversion. GEOPHYSICS *60*, 1457–1473.
- 32. Chavent, G. (2010). Nonlinear Least Squares for Inverse Problems Theoretical Foundations and Step-by-Step Guide for Applications (Springer Netherlands).
- 33. Plessix, R.-E. (2006). A review of the adjoint-state method for computing the gradient of a functional with geophysical applications. Geophys. J. Int. *167*, 495–503.
- 34. Bernard, S., Monteiller, V., Komatitsch, D., and Lasaygues, P. (2017). Ultrasonic computed tomography based on full-waveform inversion for bone quantitative imaging. Phys. Med. Biol. *62*, 7011–7035.
- 35. Steinbring, J., Pander, M., and Hanebeck, U. (2016). The Smart Sampling Kalman Filter with Symmetric Samples. J. Adv. Inf. Fusion *11*, 71–90.
- 36. Levander, A.R. (1988). Fourth‐ order finite‐ difference *P-SV* seismograms. GEOPHYSICS *53*, 1425–1436.
- 37. Cristini, P., and Komatitsch, D. (2012). Some illustrative examples of the use of a spectral-element method in ocean acoustics. J. Acoust. Soc. Am. *131*, EL229–EL235.
- 38. Komatitsch, D., and Martin, R. (2007). An unsplit convolutional perfectly matched layer improved at grazing incidence for the seismic wave equation. GEOPHYSICS *72*, SM155–SM167.
- 39. Espinosa, L., Doveri, E., Bernard, S., Monteiller, V., Guillermin, R., and Lasaygues, P. (2021). Ultrasonic Imaging of high-contrasted objects based on full-waveform inversion: limits under fluid modeling. Ultrason. Imaging *43*, 88–99.
- 40. Loosvelt, M., and Lasaygues, P. (2011). A Wavelet-Based Processing method for simultaneously determining ultrasonic velocity and material thickness. Ultrasonics *51*, 325–339.
- 41. Wang, Y., Chevrot, S., Monteiller, V., Komatitsch, D., Mouthereau, F., Manatschal, G., Sylvander, M., Diaz,
- J., Ruiz, M., Grimaud, F., et al. (2016). The deep roots of the western Pyrenees revealed by full waveform inversion of teleseismic P waves. Geology *44*, 475–478.
-

- Figure 1: Schematic diagram of the Full Waveform Inversion (FWI).
- Figure 2: Simulation range with absorbent layer.
- Figure 3: Transmission- and diffraction-mode USCT of a mimetic newborn arm phantom.
- Figure 4: Temporal and spectral representation of the impulse response of the transducer (65536 597 samples, $f_0 = 648$ kHz, $F_e = 20$ MHz). The dotted red curve on the temporal representation represents the envelope of the signal (i.e., the modulus of the Hilbert transform).
- 599 Figure 5: X-ray micro-tomography of the newborn arm phantom. Size of 1840 x 1456 pixels² and
- resolution of 30 μm. Marks No. 1 and 2 identify bone contours, No. 3, 4 and 5 blood vessels, and
- No. 6 adipose tissue.
- Figure 6: Theoretical maps defined from X-ray micro-tomography of the newborn arm phantom for 603 (a) the US velocity, v_p^{th} , and (b) the mass density, ρ^{th} .
- Figure 7: Numerical simulation of the non-linear USCT using FWI algorithm of the newborn arm phantom. (a) US velocity map, (b) mass density map. No a priori information on the different 606 media. Starting maps were initialized with measured values for water (i.e., $v_p = 1467$ m/s and $\rho =$ 1000 kg/m3).
- Figure 8: Gradient of the cost function at the last iteration with respect to (a) the US velocity and (b) the mass density, for the inversion of simulated data.
- Figure 9: (left) NRMSE for the different cut-off frequencies of the low pass filter used during inversion, as a function of the iterations, for the US velocity (in blue) and the mass density (in red); (right) convergence of the algorithm, for the inversion of simulated data.
- Figure 10: Experimental non-linear USCT using FWI algorithm of the newborn arm phantom. (a)
- US velocity map, (b) mass density map. No a priori information on the different media: starting
- 615 maps were initialized with measured values for water (i.e., $v_p = 1467$ m/s and $\rho = 1000 \text{ kg/m}^3$)
- Figure 11: Gradient of the cost function at the last iteration with respect to (a) the US velocity and
- (b) the mass density, for the inversion of experimental data without any a priori information.
- Figure 12: (left) NRMSE for the different cut-off frequencies of the low pass filter used during inversion, as a function of the iterations, for the US velocity (in blue) and the mass density (in red);

 (right) convergence of the algorithm, for the inversion of experimental data without any a priori information.

 Figure 13: Starting maps used for the first iteration of the non-linear USCT using FWI algorithm of the newborn arm phantom. (a) US velocity starting map initialized with US velocity measured in 624 water ($v_p = 1467$ m/s), (b) theoretical density starting map obtained from the binarized X-ray μ CT.

 Figure 14: Experimental non-linear USCT using FWI algorithm of the newborn arm phantom. (a) US velocity map, (b) mass density map. A priori information for the mass density starting map 627 obtained from the binarized X-ray μ CT.

 Figure 15: Gradient of the cost function at the last iteration with respect to (a) the US velocity and (b) the mass density, for the inversion of experimental data with a priori information for the mass 630 density starting map obtained from the binarized X-ray μ CT.

 Figure 16: (left) NRMSE for the different cut-off frequencies of the low pass filter used during inversion, as a function of the iterations, for the US velocity (in blue) and the mass density (in red); (right) convergence of the algorithm, for the inversion of experimental data with a priori 634 information for the mass density starting map obtained from the binarized X-ray μ CT.

Figure 17: Starting maps used for the first iteration of the non-linear USCT using FWI algorithm of

the newborn arm phantom. (a) theoretical US velocity starting map obtained from the binarized X-

637 and μ CT, (b) mass density map initialized with the mass density of water ($\rho = 1000 \text{ kg/m}^3$)

 Figure 18: Experimental non-linear USCT using FWI algorithm of the newborn arm phantom. (a) US velocity map, (b) mass density map. A priori information for the US velocity starting map 640 obtained from the binarized X-ray μ CT.

 Figure 19: Gradient of the cost function at the last iteration with respect to (a) the US velocity and (b) the mass density, for the inversion of experimental data with a priori information for the US 643 velocity starting map obtained from the binarized X-ray μ CT.

 Figure 20: (left) NRMSE for the different cut-off frequencies of the low pass filter used during inversion, as a function of the iterations, for the US velocity (in blue) and the mass density (in red); (right) convergence of the algorithm, for the inversion of experimental data with a priori 647 information for the US velocity starting map obtained from the binarized X-ray μ CT.

650 Table 1: Acoustic mean properties of water and materials in the newborn arm phantom (True

651 Phantom Solutions Inc., Windsor, Canada).

652 Table 2: Summary of the normalized root mean square errors (NRMSE) calculated for each of the

653 images presented at the last iteration.

654

649

10. Acknowledgment

 Centre de Calcul Intensif d'Aix-Marseille (Mesocentre) is acknowledged for granting access to its high-performance computing resources. The authors are grateful to Vincent Long and Philippe Bindzi from the Laboratory of Mechanics and Acoustics, for their help in setting up the experiments, and Laurent Sabatier for the X-ray µCT (Fédération Fabri de Peiresc, CNRS, Marseille) of the newborn arm phantom. The authors are grateful to Mustafa Ali Misri, Business Manager, Faiqa Khzir, Operations Officer, and Adrian Wydra, Chief Technology Officer from True Phantom Solutions in Windsor, Canada, for their assistance in the fabrication of the newborn arm phantom and the evaluation of some properties of the materials used.

11. Fundings

 The project leading to this publication has received funding from the Excellence Initiative of Aix- Marseille University - A*Midex, a French "Investissements d'Avenir programme" - Institut Mécanique et Ingénierie (IMI, AMX-19-IET-010). Elise Doveri's PhD thesis was funded by the French Ministry of Research and Kouakou Jean Hermann Konan's internship was funded by the A*Midex - Institut Marseille Imaging.